



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FHP

S/N 80-003412

Date In 2/11/16

Date Out 2/22/2016

Ship P/U H/D CMI EE

Intake Performed By <u>TR</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: _____ _____ _____	Quality Checks Performed By <u>JD</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>208</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>.148</u> (.139 - .169) 36mm <u>.164</u> (.156 - .190) 53mm <u>.234</u> (.228 - .278) 103mm <u>.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 07/11/2017</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201502G 02/24/2017</td> </tr> <tr> <td>0.20</td> <td>SD3933</td> <td>201505A 05/12/2017</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG51701 04/29/2017</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 07/11/2017	0.08	SD1011	201502G 02/24/2017	0.20	SD3933	201505A 05/12/2017	0.08 DGS	N/A	AG51701 04/29/2017	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547) Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____
Simulator	Serial #	Lot #/Exp															
0.05	SD1018	201507A 07/11/2017															
0.08	SD1011	201502G 02/24/2017															
0.20	SD3933	201505A 05/12/2017															
0.08 DGS	N/A	AG51701 04/29/2017															

RECEIVED
FEB 23 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																																													
0.000		N/A	N/A																																													
0.040																																																
0.100																																																
0.200																																																
0.400																																																
0.080 DGS	N/A																																															
Simulator	Serial Number	Lot Number	Expiration																																													
0.05																																																
0.08																																																
0.20																																																
0.08 DGS	N/A																																															

Department Inspection Performed By <u>JD</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1014</u> Gauge ID# <u>28427</u> <u>1013</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD122</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD3933</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	SD122	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD3933
Simulator	Serial Number											
0.00	SD122											
Interferent	SD1021											
0.05	SD1018											
0.08	SD1011											
0.20	SD3933											

Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____
--

Notes: OK-PWS

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy
Quality Control Review

0/22/16
Date

Stability Checks - Instrument # 80-003412 FHP 2/22/2016 88

WBB

2016

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003412
02/22/2016
Software: 8100.27

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003412
02/22/2016
Software: 8100.27

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003412
02/22/2016
Software: 8100.27

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003412
02/22/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:19
Control Test	0.050	12:20
Air Blank	0.000	12:20
Control Test	0.050	12:21
Air Blank	0.000	12:21
Control Test	0.051	12:22
Air Blank	0.000	12:22
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

Operator's Signature

Test	g/210L	Time
Air Blank	0.000	12:24
Control Test	0.080	12:24
Air Blank	0.000	12:25
Control Test	0.079	12:25
Air Blank	0.000	12:26
Control Test	0.080	12:27
Air Blank	0.000	12:27
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

Operator's Signature

Test	g/210L	Time
Air Blank	0.000	12:29
Control Test	0.201	12:29
Air Blank	0.000	12:30
Control Test	0.200	12:30
Air Blank	0.000	12:31
Control Test	0.200	12:32
Air Blank	0.000	12:32
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	

Operator's Signature

Test	g/210L	Time
Air Blank	0.000	12:34
Control Test	0.079	12:34
Air Blank	0.000	12:35
Control Test	0.079	12:35
Air Blank	0.000	12:35
Control Test	0.079	12:36
Air Blank	0.000	12:36
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

DGS