

INSTRUMENT PROCESSING SHEET

Agency Osceola CSO S/N 80-003935

Date In 6/28/16 Date Out 6/29/16 Ship P/U H/D CMI EE

RECEIVED
JUN 30 2016
FDLE
Alcohol Testing Program

Intake Performed By DP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
ok Case ok Handle
ok Dry Gas Holder ok Feet
ok Keyboard/Plug ok Back/Plugs
ok Screws tight ok Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By DP

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 130
 Flow Verification (L/s)
 Flow Column # A7P102
 32mm 143 (.139 - .169)
 36mm 106 (.156 - .190)
 53mm 235 (.228 - .278)
 103mm 496 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD3962	201507A 7/14/17
0.08	SD3964	201601F 01/26/18
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	18-605301 2/22/16

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DP

Barometric Pressure 1017 Gauge
 ID# 28427 1017 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2880
Interferent	C-28534
0.05	SD3962
0.08	SD3964
0.20	G4444

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QA/pc ok GDM

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
 Quality Control Review

6/30/16
 Date

80-001166
 Stability Checks
 6/29/16

JK

INTOXILYZER 8000
 Instrument Initialization
 06:48 06/29/2016

BREUARD CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001166
 06/29/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:27
Control Test	0.083	07:27
Air Blank	0.000	07:28
Control Test	0.082	07:28
Air Blank	0.000	07:29
Control Test	0.079	07:29
Air Blank	0.000	07:30
Control Test Stats		
Average	0.0813	
Std Dev	0.0021	
Rel Std Dev(%)	2.5594	

DGS

Operator's Signature

BREUARD CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001166
 06/29/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:37
Control Test	0.199	07:37
Air Blank	0.000	07:38
Control Test	0.198	07:38
Air Blank	0.000	07:39
Control Test	0.198	07:39
Air Blank	0.000	07:40
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

Operator's Signature

BREUARD CO SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-001166
 06/29/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:41
Control Test	0.078	07:42
Air Blank	0.000	07:42
Control Test	0.078	07:43
Air Blank	0.000	07:44
Control Test	0.079	07:44
Air Blank	0.000	07:45
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

Operator's Signature