

INSTRUMENT PROCESSING SHEET

Agency Osceola County SD S/N 80-003937
 Date In 6/28/16 Date Out 6/29/16 Ship P/U H/D CMI EE

RECEIVED
 JUN 30 2016
 FDLE
 Alcohol Testing Program

Intake Performed By DR

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By DR

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 215
 Flow Verification (L/s)
 Flow Column # A7P102
 32mm 150 (.139 - .169)
 36mm 179 (.156 - .190)
 53mm 261 (.228 - .278)
 103mm 508 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD3962	201507A 7/14/17
0.08	SD3964	201601F 1/26/18
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	AUG05301 2/22/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 43mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DR

Barometric Pressure 1012 Gauge
 ID# 28427 1012 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2880
Interferent	G-2834
0.05	SD3962
0.08	SD3964
0.20	G4444

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: Instrument came out of Ready Mode with no issue. Dept. Ins. Conducted.
In Compliance
DA/DR OK (PJM)

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Scott Kirkland
 Quality Control Review

6/30/16
 Date

80-003937
 STEPHEN LUCAS
 6/29/16

INTOXILYZER 8000
 Instrument Initialization
 06:50 06/29/2016

OSCEOLA COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003937
 06/29/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:29
Control Test	0.047	07:30
Air Blank	0.000	07:31
Control Test	0.048	07:31
Air Blank	0.000	07:32
Control Test	0.049	07:32
Air Blank	0.000	07:33
Control Test Stats		
Average	0.0480	
Std Dev	0.0010	
Rel Std Dev(%)	2.0833	

Operator's Signature

OSCEOLA COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003937
 06/29/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:34
Control Test	0.077	07:34
Air Blank	0.000	07:35
Control Test	0.078	07:35
Air Blank	0.000	07:35
Control Test	0.078	07:37
Air Blank	0.000	07:37
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

Operator's Signature

OSCEOLA COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003937
 06/29/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:39
Control Test	0.079	07:39
Air Blank	0.000	07:40
Control Test	0.079	07:40
Air Blank	0.000	07:41
Control Test	0.080	07:41
Air Blank	0.000	07:41
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DS

Operator's Signature

OSCEOLA COUNTY S.O.
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-003937
 06/29/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:43
Control Test	0.202	07:44
Air Blank	0.000	07:44
Control Test	0.204	07:45
Air Blank	0.000	07:45
Control Test	0.203	07:46
Air Blank	0.000	07:47
Control Test Stats		
Average	0.2030	
Std Dev	0.0010	
Rel Std Dev(%)	0.4926	

BSK

Operator's Signature