



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Osceola County SO S/N 80-004587
Date In 1/13/16 Date Out 2/1/16 Ship P/U H/D CMI EE

Intake Performed By TV

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 158
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 142 (.139 - .169)
 36mm 168 (.156 - .190)
 53mm 240 (.228 - .278)
 103mm 500 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201502G 2/24/17
0.20	SD1025	201505A 5/12/17
0.08 DGS	N/A	AG511701 4/22/17

Flow Calibration Performed By [Signature]

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 36mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
FEB 02 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1017 Gauge
 ID# 28427 1016 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: AT was unable to upload. Instrument
uploaded fine at FDLE.
QC-DMB

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy
Quality Control Review

2/2/16
Date

80-004587
 Stability checks
 2/1/16 *R*

Bus

DCS

QDM

INTOXILYZER 8000
 Instrument Initialization
 16:42 02/01/2016

OSCEOLA COUNTY S.C.
 Intoxilyzer - Alcotest Analyzer
 Model 8000 SN 80-004587
 02/01/2016
 Software: 8.00.27

Test	9:21:01	Time
Air Blank	0.000	09:19
Control Test	0.079	09:19
Air Blank	0.000	09:21
Control Test	0.079	09:21
Air Blank	0.000	09:21
Control Test	0.079	09:22
Air Blank	0.000	09:22
Control Test Stats		
Average	0.0799	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature *[Signature]*

OSCEOLA COUNTY S.C.
 Intoxilyzer - Alcotest Analyzer
 Model 8000 SN 80-004587
 02/01/2016
 Software: 8.00.27

Test	9:21:01	Time
Air Blank	0.000	09:23
Control Test	0.000	09:24
Air Blank	0.000	09:25
Control Test	0.000	09:25
Air Blank	0.000	09:25
Control Test	0.000	09:27
Air Blank	0.000	09:27
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature *[Signature]*

OSCEOLA COUNTY S.C.
 Intoxilyzer - Alcotest Analyzer
 Model 8000 SN 80-004587
 02/01/2016
 Software: 8.00.27

Test	9:21:01	Time
Air Blank	0.000	09:29
Control Test	0.079	09:29
Air Blank	0.000	09:30
Control Test	0.078	09:31
Air Blank	0.000	09:31
Control Test	0.078	09:31
Air Blank	0.000	09:32
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

Operator's Signature *[Signature]*

OSCEOLA COUNTY S.C.
 Intoxilyzer - Alcotest Analyzer
 Model 8000 SN 80-004587
 02/01/2016
 Software: 8.00.27

Test	9:21:01	Time
Air Blank	0.000	09:13
Control Test	0.198	09:14
Air Blank	0.000	09:15
Control Test	0.195	09:15
Air Blank	0.000	09:16
Control Test	0.200	09:17
Air Blank	0.000	09:17
Control Test Stats		
Average	0.1960	
Std Dev	0.0010	
Rel Std Dev(%)	0.5025	

Operator's Signature *[Signature]*