



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Hurlburt Field PD

S/N 80-004771

Date In 1/27/16

Date Out 1/27/16

Ship P/U R/D CMI EE

Intake Performed By PWS

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By PWS

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 170
 Flow Verification (L/s)
 Flow Column # ATP103
 32mm 144 (.139 - .169)
 36mm 1104 (.156 - .190)
 53mm 234 (.228 - .278)
 103mm 511 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201502G 2/24/17
0.20	SD1025	201505A 5/12/17
0.08 DGS	N/A	AG 511701 4/27/17

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
FEB 01 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By PWS

Barometric Pressure 1013 Gauge
 ID# 28427 1013 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: OK-BK

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy
Quality Control Review

2/1/16
Date

Stability Tests - Hurlburt Field PD # 80-004771 1/27/16

BSM

HURLBURT FIELD PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004771
01/27/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.080	13:08
Control Test	0.050	13:09
Air Blank	0.000	13:09
Control Test	0.051	13:10
Air Blank	0.000	13:10
Control Test	0.050	13:11
Air Blank	0.000	13:12
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

QNS

Operator's Signature

HURLBURT FIELD PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004771
01/27/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:19
Control Test	0.078	13:19
Air Blank	0.000	13:20
Control Test	0.080	13:21
Air Blank	0.000	13:21
Control Test	0.081	13:22
Air Blank	0.000	13:22
Control Test Stats		
Average	0.0797	
Std Dev	0.0015	
Rel Std Dev(%)	1.9174	

QNS

Operator's Signature

HURLBURT FIELD PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004771
01/27/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:23
Control Test	0.200	13:24
Air Blank	0.000	13:24
Control Test	0.203	13:25
Air Blank	0.000	13:26
Control Test	0.203	13:26
Air Blank	0.000	13:27
Control Test Stats		
Average	0.2020	
Std Dev	0.0017	
Rel Std Dev(%)	0.8574	

QNS

Operator's Signature

HURLBURT FIELD PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004771
01/27/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:28
Control Test	0.079	13:28
Air Blank	0.000	13:29
Control Test	0.079	13:29
Air Blank	0.000	13:30
Control Test	0.079	13:30
Air Blank	0.000	13:30
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

BK

DS

QNS

Operator's Signature