



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Okaloosa County, SD S/N 80-005057

Date In 8/15/16 Date Out 8/17/16 Ship P/U H/D CMI EE

Intake	Performed By <u>WOB</u>	Quality Checks	Performed By <u>PWS</u>	Flow Calibration	Performed By <u>PWS</u>															
<input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>193</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>AP102</u> 32mm <u>148</u> (.139 - .169) 36mm <u>167</u> (.156 - .190) 53mm <u>234</u> (.228 - .278) 103mm <u>511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> > 1% <input checked="" type="checkbox"/> Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>G11739</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>G8149</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>G11621</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG612405 5/3/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	G11739	201507A 7/14/17	0.08	G8149	201601F 1/26/18	0.20	G11621	201604C 4/5/18	0.08 DGS	N/A	AG612405 5/3/18		<input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # <u>Aug 19 2016</u> <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/min) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)	
Simulator	Serial #	Lot #/Exp																		
0.05	G11739	201507A 7/14/17																		
0.08	G8149	201601F 1/26/18																		
0.20	G11621	201604C 4/5/18																		
0.08 DGS	N/A	AG612405 5/3/18																		
		Maintenance	Performed By																	
		<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																		
		Suggested Service																		

Optical Bench Calibration	Performed By <u>PWS</u>	Department Inspection	Performed By <u>PWS</u>																																																												
<input type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge <u>1021</u> ID# <u>28427</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G16621</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>G2882</td> <td>16101</td> <td>2/2/18</td> </tr> <tr> <td>0.100</td> <td>G2078</td> <td>16001</td> <td>5/8/18</td> </tr> <tr> <td>0.200</td> <td>G2408</td> <td>15104</td> <td>5/22/17</td> </tr> <tr> <td>0.400</td> <td>G5358</td> <td>16102</td> <td>3/22/18</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>03415080A1</td> <td>3/5/17</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>G11739</td> <td>201507A</td> <td>7/14/17</td> </tr> <tr> <td>0.08</td> <td>G8149</td> <td>201601F</td> <td>1/26/18</td> </tr> <tr> <td>0.20</td> <td>G11621</td> <td>201604C</td> <td>4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG612405</td> <td>5/3/18</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	G16621	N/A	N/A	0.040	G2882	16101	2/2/18	0.100	G2078	16001	5/8/18	0.200	G2408	15104	5/22/17	0.400	G5358	16102	3/22/18	0.080 DGS	N/A	03415080A1	3/5/17	Simulator	Serial Number	Lot Number	Expiration	0.05	G11739	201507A	7/14/17	0.08	G8149	201601F	1/26/18	0.20	G11621	201604C	4/5/18	0.08 DGS	N/A	AG612405	5/3/18		<input checked="" type="checkbox"/> Barometric Pressure <u>1019</u> Gauge ID# <u>26932</u> <u>1021</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>G2879</td> </tr> <tr> <td>Interferent</td> <td>G8144</td> </tr> <tr> <td>0.05</td> <td>G11739</td> </tr> <tr> <td>0.08</td> <td>G8149</td> </tr> <tr> <td>0.20</td> <td>G11621</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	G2879	Interferent	G8144	0.05	G11739	0.08	G8149	0.20	G11621	
Simulator	Serial Number	Lot Number	Expiration																																																												
0.000	G16621	N/A	N/A																																																												
0.040	G2882	16101	2/2/18																																																												
0.100	G2078	16001	5/8/18																																																												
0.200	G2408	15104	5/22/17																																																												
0.400	G5358	16102	3/22/18																																																												
0.080 DGS	N/A	03415080A1	3/5/17																																																												
Simulator	Serial Number	Lot Number	Expiration																																																												
0.05	G11739	201507A	7/14/17																																																												
0.08	G8149	201601F	1/26/18																																																												
0.20	G11621	201604C	4/5/18																																																												
0.08 DGS	N/A	AG612405	5/3/18																																																												
Simulator	Serial Number																																																														
0.00	G2879																																																														
Interferent	G8144																																																														
0.05	G11739																																																														
0.08	G8149																																																														
0.20	G11621																																																														
		Attachments																																																													
		<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____																																																													

Notes: Calibrated Optical Bench due to Barometric Pressure being outside 1% of nominal. (PWS)

PA/OC OK PWS 8/18/16

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

Brett Whitford

Quality Control Review

8/19/16

Date

Stability Tests
Pre-Calibration

AK - Okaloosa CSD # 80-005057 8/17/16

AK

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005057
08/17/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:15
Control Test	0.049	09:16
Air Blank	0.000	09:16
Control Test	0.049	09:17
Air Blank	0.000	09:17
Control Test	0.050	09:18
Air Blank	0.000	09:19
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

QWS

Operator's Signature

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005057
08/17/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:20
Control Test	0.079	09:20
Air Blank	0.000	09:21
Control Test	0.078	09:22
Air Blank	0.000	09:22
Control Test	0.078	09:23
Air Blank	0.000	09:23
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

QWS

Operator's Signature

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005057
08/17/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:28
Control Test	0.195	09:29
Air Blank	0.000	09:29
Control Test	0.195	09:30
Air Blank	0.000	09:31
Control Test	0.195	09:31
Air Blank	0.000	09:32
Control Test Stats		
Average	0.1950	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

QWS

Operator's Signature

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005057
08/17/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:25
Control Test	0.081	09:25
Air Blank	0.000	09:26
Control Test	0.081	09:26
Air Blank	0.000	09:26
Control Test	0.081	09:27
Air Blank	0.000	09:27
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

QWS

Operator's Signature

QWS

Optical Bench
Calibration

OKalboson CSD
80-005057

8/17/16

OKALOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
08/17/2016
SN 80-005057
09:33:24

Auto Calibration
Max Power Res Value = 30
Auto Range Res Value = 16

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %***
Samples Taken = 4, Discarded = 1
Sum Io = 12568, Sum Io = 12710
<<<<< CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)
Sample #1 = 0.0430 (-0.0120)
Sample #2 = 0.0500 (0.0260)
Sample #3 = 0.0330 (0.0780)
Sample #4 = 0.0530 (0.0900)
Avg % Abs = 0.0453 (0.0647)
STD DEV = 0.0108 (0.0340)
REL STD DEV = 23.792 (52.608)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1040 (-0.0180)
Sample #2 = 0.0800 (0.0000)
Sample #3 = 0.0890 (0.0130)
Sample #4 = 0.1220 (0.0030)
Avg % Abs = 0.0997 (0.0053)
STD DEV = 0.0193 (0.0168)
REL STD DEV = 19.412 (127.629)

BM

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %***
Samples Taken = 4, Discarded = 1
Sum Io = 12552, Sum Io = 12705
<<<<< CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)
Sample #1 = 0.7700 (-0.0220)
Sample #2 = 0.7500 (0.0060)
Sample #3 = 0.7230 (0.0410)
Sample #4 = 0.7560 (0.0240)
Avg % Abs = 0.7430 (0.0237)
STD DEV = 0.0176 (0.0175)
REL STD DEV = 2.366 (73.954)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.4550 (-0.0180)
Sample #2 = 1.4560 (-0.0160)
Sample #3 = 1.4450 (-0.0030)
Sample #4 = 1.4730 (-0.0200)
Avg % Abs = 1.4580 (-0.0130)
STD DEV = 0.0141 (0.0089)
REL STD DEV = 0.968 (68.371)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %***
Samples Taken = 4, Discarded = 1
Sum Io = 12546, Sum Io = 12703
<<<<< CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)
Sample #1 = 1.8130 (-0.0200)
Sample #2 = 1.7990 (0.0020)
Sample #3 = 1.7980 (0.0110)
Sample #4 = 1.8270 (0.0060)
Avg % Abs = 1.8080 (0.0063)
STD DEV = 0.0165 (0.0045)
REL STD DEV = 0.911 (71.199)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.4950 (-0.0240)
Sample #2 = 3.4860 (-0.0150)
Sample #3 = 3.4970 (-0.0100)
Sample #4 = 3.5020 (-0.0290)
Avg % Abs = 3.4950 (-0.0180)
STD DEV = 0.0082 (0.0098)
REL STD DEV = 0.234 (54.716)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %***
Samples Taken = 4, Discarded = 1
Sum Io = 12541, Sum Io = 12704
<<<<< CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)
Sample #1 = 3.3220 (-0.0060)
Sample #2 = 3.5720 (-0.0080)
Sample #3 = 3.5980 (0.0000)
Sample #4 = 3.5680 (0.0200)
Avg % Abs = 3.5760 (0.0040)
STD DEV = 0.0106 (0.0144)
REL STD DEV = 0.296 (360.555)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.7710 (0.0060)
Sample #2 = 6.8180 (0.0060)
Sample #3 = 6.8360 (0.0010)
Sample #4 = 6.8440 (0.0160)
Avg % Abs = 6.8327 (0.0077)
STD DEV = 0.0133 (0.0076)
REL STD DEV = 0.195 (99.621)

Sol Value = 0.400 g/210L ***
Fit Value = 1.9148 mg/l %***
Samples Taken = 4, Discarded = 1
Sum Io = 12539, Sum Io = 12701
<<<<< CHANNEL 1 >>>>>

Sample % Abs (% Abs Ref)
Sample #1 = 6.7000 (-0.0040)
Sample #2 = 6.7050 (0.0180)
Sample #3 = 6.7380 (0.0000)
Sample #4 = 6.7600 (0.0020)
Avg % Abs = 6.7343 (0.0067)
STD DEV = 0.0277 (0.0099)
REL STD DEV = 0.411 (147.966)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 12.5570 (-0.0100)
Sample #2 = 12.5630 (0.0100)
Sample #3 = 12.6300 (0.0030)
Sample #4 = 12.6280 (-0.0070)
Avg % Abs = 12.6070 (0.0020)
STD DEV = 0.0381 (0.0085)
REL STD DEV = 0.302 (427.200)

**** AUTO CAL DATA ****
<<<<< CHANNEL 1 >>>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.045
Std Dev = 0.01 Rel Std Dev = 23.79
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.743
Std Dev = 0.02 Rel Std Dev = 2.37
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.808
Std Dev = 0.02 Rel Std Dev = 0.91
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.576
Std Dev = 0.01 Rel Std Dev = 0.30
Sol Val = 1.9148 mg/l or 0.400 g/210L
% Abs = 6.734
Std Dev = 0.03 Rel Std Dev = 0.41
Zero Order Coef = -56.14
First Order Coef = 2546.69
Second Order Coef = 42.69
Standard Deviation = 64.554642

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0013
0.040	0.039	0.0009
0.100	0.098	0.0015
0.200	0.202	-0.0016
0.400	0.400	0.0003

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %***
Samples Taken = 4, Discarded = 1
**** CHANNEL 1

Sample #1 = 3509.00
Sample #2 = 3459.00
Sample #3 = 3479.00
Sample #4 = 3462.00
Average Result = 3466.6667
STD DEV = 10.7858
REL STD DEV = 0.311

**** CHANNEL 2
Sample #1 = 3455.00
Sample #2 = 3455.00
Sample #3 = 3437.00
Sample #4 = 3454.00
Average Result = 3448.6667
STD DEV = 10.1160
REL STD DEV = 0.293

**** CHANNEL 1 >>>>>
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1021
3 um H2O Adjust (mg/l*10,000) = 343
9 um H2O Adjust (mg/l*10,000) = 361
**** AUTO CAL PASS

BK

Stability Tests
Post Calibration
BK

OKaloosa C80 # 80-005057 8/17/16

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005057
08/17/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:21
Control Test	0.049	10:22
Air Blank	0.000	10:22
Control Test	0.050	10:23
Air Blank	0.000	10:24
Control Test	0.049	10:24
Air Blank	0.000	10:25
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

gms

gms

Operator's Signature

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005057
08/17/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:35
Control Test	0.079	10:35
Air Blank	0.000	10:36
Control Test	0.079	10:37
Air Blank	0.000	10:37
Control Test	0.079	10:38
Air Blank	0.000	10:38
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

gms

Operator's Signature

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005057
08/17/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:26
Control Test	0.195	10:27
Air Blank	0.000	10:27
Control Test	0.197	10:28
Air Blank	0.000	10:29
Control Test	0.198	10:29
Air Blank	0.000	10:30
Control Test Stats		
Average	0.1967	
Std Dev	0.0015	
Rel Std Dev(%)	0.7767	

gms

Operator's Signature

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005057
08/17/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:31
Control Test	0.081	10:31
Air Blank	0.000	10:31
Control Test	0.081	10:32
Air Blank	0.000	10:32
Control Test	0.081	10:33
Air Blank	0.000	10:33
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

gms

gms

Operator's Signature