

INSTRUMENT PROCESSING SHEET

Agency Crestview P.D. S/N 80-005058

Date In 11/21/2016 Date Out 11/23/16 Ship P/U H/D CMI EE

Intake Performed By <u>TP</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____	Quality Checks Performed By <u>PGM</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>236</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32mm <u>160</u> (.139 - .169) 36mm <u>183</u> (.156 - .190) 53mm <u>246</u> (.228 - .278) 103mm <u>492</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>501018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>501011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>501025</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG626605 9/22/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	501018	201507A 7/14/17	0.08	501011	201601F 1/26/18	0.20	501025	201604C 4/5/18	0.08 DGS	N/A	AG626605 9/22/18	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	501018	201507A 7/14/17															
0.08	501011	201601F 1/26/18															
0.20	501025	201604C 4/5/18															
0.08 DGS	N/A	AG626605 9/22/18															
<div style="border: 1px solid black; padding: 5px; display: inline-block;"> RECEIVED NOV 28 2016 FDLE Alcohol Testing Program </div>																	
Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																	
Suggested Service _____ _____ _____																	

Optical Bench Calibration Performed By <u>PGM</u> <input type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge <u>1021</u> ID # <u>26932</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>501016</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>501024</td> <td>16101</td> <td>2/2/18</td> </tr> <tr> <td>0.100</td> <td>501013</td> <td>16001</td> <td>5/8/18</td> </tr> <tr> <td>0.200</td> <td>501012</td> <td>16103</td> <td>6/4/18</td> </tr> <tr> <td>0.400</td> <td>66621</td> <td>16102</td> <td>3/22/18</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>03415080A1</td> <td>3/5/17</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>501018</td> <td>201507A</td> <td>7/14/17</td> </tr> <tr> <td>0.08</td> <td>501011</td> <td>201601F</td> <td>1/26/18</td> </tr> <tr> <td>0.20</td> <td>501025</td> <td>201604C</td> <td>4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG626605</td> <td>9/22/18</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	501016	N/A	N/A	0.040	501024	16101	2/2/18	0.100	501013	16001	5/8/18	0.200	501012	16103	6/4/18	0.400	66621	16102	3/22/18	0.080 DGS	N/A	03415080A1	3/5/17	Simulator	Serial Number	Lot Number	Expiration	0.05	501018	201507A	7/14/17	0.08	501011	201601F	1/26/18	0.20	501025	201604C	4/5/18	0.08 DGS	N/A	AG626605	9/22/18	Department Inspection Performed By <u>PGM</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1021</u> Gauge ID# <u>28427</u> <u>1021</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-A</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>501019</td> </tr> <tr> <td>Interferent</td> <td>501021</td> </tr> <tr> <td>0.05</td> <td>501018</td> </tr> <tr> <td>0.08</td> <td>501011</td> </tr> <tr> <td>0.20</td> <td>501025</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	501019	Interferent	501021	0.05	501018	0.08	501011	0.20	501025
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000	501016	N/A	N/A																																																										
0.040	501024	16101	2/2/18																																																										
0.100	501013	16001	5/8/18																																																										
0.200	501012	16103	6/4/18																																																										
0.400	66621	16102	3/22/18																																																										
0.080 DGS	N/A	03415080A1	3/5/17																																																										
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05	501018	201507A	7/14/17																																																										
0.08	501011	201601F	1/26/18																																																										
0.20	501025	201604C	4/5/18																																																										
0.08 DGS	N/A	AG626605	9/22/18																																																										
Simulator	Serial Number																																																												
0.00	501019																																																												
Interferent	501021																																																												
0.05	501018																																																												
0.08	501011																																																												
0.20	501025																																																												
Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____																																																													

Notes: All values were within allowable range, cal adjusted to bring them closer to nominal.
QC OK SP

 Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett H. [Signature]
 Quality Control Review

11/28/16
 Date

CRESTVIEW PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005058
11/21/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:54
Control Test	0.050	09:55
Air Blank	0.000	09:56
Control Test	0.051	09:56
Air Blank	0.000	09:57
Control Test	0.051	09:58
Air Blank	0.000	09:58
Control Test Stats		
Average	0.0507	
Std Dev	0.0006	
Rel Std Dev(%)	1.1395	

CRESTVIEW PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005058
11/21/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:50
Control Test	0.081	09:51
Air Blank	0.000	09:51
Control Test	0.082	09:52
Air Blank	0.000	09:52
Control Test	0.081	09:53
Air Blank	0.000	09:53
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	

CRESTVIEW PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005058
11/21/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:59
Control Test	0.200	10:00
Air Blank	0.000	10:00
Control Test	0.200	10:01
Air Blank	0.000	10:02
Control Test	0.200	10:02
Air Blank	0.000	10:03
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

P Murphy

Operator's Signature

P Murphy

Operator's Signature

P Murphy

Operator's Signature

CRESTVIEW PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005058
11/21/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:18
Control Test	0.083	10:18
Air Blank	0.000	10:18
Control Test	0.084	10:19
Air Blank	0.000	10:19
Control Test	0.083	10:20
Air Blank	0.000	10:20
Control Test Stats		
Average	0.0833	
Std Dev	0.0006	
Rel Std Dev(%)	0.6928	

PRE-CALIBRATION - ADJUSTMENT

SP BK

DGS

P Murphy

Operator's Signature

CRESTVIEW PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005058
11/21/2016 10:21:41

Auto Calibration

Max Power Res Value = 37
Auto Range Res Value = 20

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12795, Sum Io = 12798

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0860 (0.0070)
Sample #2 = 0.1000 (0.0340)
Sample #3 = 0.0480 (0.1000)
Sample #4 = 0.0600 (0.1040)
Avg % Abs = 0.0693 (0.0793)
STD DEV = 0.0272 (0.0393)
REL STD DEV = 39.270 (49.551)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1410 (0.0120)
Sample #2 = 0.1640 (0.0000)
Sample #3 = 0.1490 (0.0150)
Sample #4 = 0.1420 (0.0220)
Avg % Abs = 0.1517 (0.0123)
STD DEV = 0.0112 (0.0112)
REL STD DEV = 7.411 (91.134)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12777, Sum Io = 12794

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8070 (-0.0120)
Sample #2 = 0.8180 (0.0190)
Sample #3 = 0.7720 (0.0540)
Sample #4 = 0.8030 (0.0420)
Avg % Abs = 0.7977 (0.0383)
STD DEV = 0.0235 (0.0178)
REL STD DEV = 2.941 (46.398)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5770 (-0.0060)
Sample #2 = 1.5640 (0.0140)
Sample #3 = 1.5630 (0.0230)
Sample #4 = 1.5580 (0.0280)
Avg % Abs = 1.5617 (0.0217)
STD DEV = 0.0032 (0.0071)
REL STD DEV = 0.206 (32.744)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12769, Sum Io = 12790

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.9530 (-0.0190)
Sample #2 = 1.9530 (-0.0030)
Sample #3 = 1.9090 (0.0330)
Sample #4 = 1.9040 (0.0460)
Avg % Abs = 1.9220 (0.0253)
STD DEV = 0.0270 (0.0254)
REL STD DEV = 1.403 (100.199)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6920 (-0.0020)
Sample #2 = 3.6870 (0.0030)
Sample #3 = 3.6840 (0.0140)
Sample #4 = 3.6610 (0.0310)
Avg % Abs = 3.6773 (0.0160)
STD DEV = 0.0142 (0.0141)
REL STD DEV = 0.387 (88.167)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12763, Sum Io = 12786

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.7210 (-0.0080)
Sample #2 = 3.7590 (-0.0060)
Sample #3 = 3.7470 (-0.0020)
Sample #4 = 3.7580 (0.0020)
Avg % Abs = 3.7547 (-0.0020)
STD DEV = 0.0067 (0.0040)
REL STD DEV = 0.177 (200.000)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 7.0650 (-0.0190)
Sample #2 = 7.0800 (0.0020)
Sample #3 = 7.0800 (0.0000)
Sample #4 = 7.1000 (-0.0020)
Avg % Abs = 7.0867 (0.0000)
STD DEV = 0.0115 (0.0020)
REL STD DEV = 0.163 (0.000)

Sol Value = 0.400 g/210L ***
Fit value = 1.9048 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12761, Sum Io = 12785

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 7.1110 (-0.0120)
Sample #2 = 7.1340 (-0.0030)
Sample #3 = 7.1060 (0.0340)
Sample #4 = 7.1230 (0.0170)
Avg % Abs = 7.1210 (0.0160)
STD DEV = 0.0141 (0.0185)
REL STD DEV = 0.198 (115.752)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 13.0810 (0.0090)
Sample #2 = 13.1010 (0.0230)
Sample #3 = 13.1440 (0.0140)
Sample #4 = 13.1210 (0.0190)
Avg % Abs = 13.1220 (0.0187)
STD DEV = 0.0215 (0.0045)
REL STD DEV = 0.164 (24.157)

***** AUTO CAL DATA *****

<<<< CHANNEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.069
Std Dev = 0.03 Rel Std Dev = 39.27
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.798
Std Dev = 0.02 Rel Std Dev = 2.94
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.922
Std Dev = 0.03 Rel Std Dev = 1.40
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.755
Std Dev = 0.01 Rel Std Dev = 0.18
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 7.121
Std Dev = 0.01 Rel Std Dev = 0.20
Zero Order Coef = -124.01
First Order Coef = 2460.22
Second Order Coef = 32.39
Standard Deviation = 44.500751

SP

<<<< CHANNEL 2 >>>>

Sol Val = 0.000 mg/l or 0.000 g/210L
% Abs = 0.152
Std Dev = 0.01 Rel Std Dev = 7.41
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.562
Std Dev = 0.00 Rel Std Dev = 0.21
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.677
Std Dev = 0.01 Rel Std Dev = 0.39
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 7.087
Std Dev = 0.01 Rel Std Dev = 0.16
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 13.122
Std Dev = 0.02 Rel Std Dev = 0.16
Zero Order Coef = -148.11
First Order Coef = 1267.21
Second Order Coef = 14.84
Standard Deviation = 47.065220

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0010
0.040	0.039	0.0010
0.100	0.099	0.0008
0.200	0.201	-0.0010
0.400	0.400	0.0002

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0009
0.040	0.039	0.0008
0.100	0.099	0.0010
0.200	0.201	-0.0011
0.400	0.400	0.0003

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l ****
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 3411.00
Sample #2 = 3401.00
Sample #3 = 3425.00
Sample #4 = 3411.00
Average Result = 3412.3333
STD DEV = 12.0554
REL STD DEV = 0.353

***** CHANNEL 2
Sample #1 = 3417.00
Sample #2 = 3378.00
Sample #3 = 3417.00
Sample #4 = 3414.00
Average Result = 3403.0000
STD DEV = 21.7025
REL STD DEV = 0.638

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1022
3 um H2O Adjust (mg/l*10,000) = 397
9 um H2O Adjust (mg/l*10,000) = 406
**** AUTO CAL PASS

CRESTVIEW PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005058
11/21/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:02
Control Test	0.050	11:03
Air Blank	0.000	11:04
Control Test	0.051	11:04
Air Blank	0.000	11:05
Control Test	0.050	11:05
Air Blank	0.000	11:06
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

POST CAL

P. Murphy
Operator's Signature

CRESTVIEW PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005058
11/21/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:14
Control Test	0.080	11:15
Air Blank	0.000	11:15
Control Test	0.080	11:16
Air Blank	0.000	11:16
Control Test	0.081	11:17
Air Blank	0.000	11:18
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

POST CAL

P. Murphy
Operator's Signature

CRESTVIEW PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005058
11/21/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:21
Control Test	0.197	11:22
Air Blank	0.000	11:23
Control Test	0.198	11:23
Air Blank	0.000	11:24
Control Test	0.198	11:24
Air Blank	0.000	11:25
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	

POST CAL

P. Murphy
Operator's Signature

CRESTVIEW PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005058
11/21/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:26
Control Test	0.080	11:26
Air Blank	0.000	11:27
Control Test	0.079	11:27
Air Blank	0.000	11:28
Control Test	0.080	11:28
Air Blank	0.000	11:28
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

POST CAL

DGS

P. Murphy
Operator's Signature

BK

SJP