

INSTRUMENT PROCESSING SHEET

Agency Hernando CSO

S/N 80-005249

Date In 8/25/16

Date Out 9/2/16

Ship P/U H/D CMI EE

Intake	Performed By <u>DB</u>	Quality Checks	Performed By <u>DB</u>	Flow Calibration	Performed By																							
<input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>ok</u> Case <u>ok</u> Handle <u>ok</u> Dry Gas Holder <u>ok</u> Feet <u>ok</u> Keyboard/Plug <u>ok</u> Back/Plugs <u>ok</u> Screws tight <u>ok</u> Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>227</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>D.156</u> (.139 - .169) 36mm <u>D.167</u> (.156 - .190) 53mm <u>D.234</u> (.228 - .278) 103mm <u>D.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check <u>968</u> INS Gauge ID # <u>28427</u> v. 1017 <u>609</u> <input checked="" type="checkbox"/> Stability Checks		<input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)																								
Notes: _____		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>A9162405 5/3/18</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	SD1025	201604C 4/5/18	0.08 DGS	N/A	A9162405 5/3/18	<table border="1"> <thead> <tr> <th>Maintenance</th> <th>Performed By</th> </tr> </thead> <tbody> <tr> <td><input type="checkbox"/> Battery Replacement</td> <td></td> </tr> <tr> <td><input type="checkbox"/> Dry Gas Regulator Replacement</td> <td></td> </tr> <tr> <td><input type="checkbox"/> Breath Tube Replacement</td> <td></td> </tr> <tr> <td><input type="checkbox"/> Other _____</td> <td></td> </tr> </tbody> </table>	Maintenance	Performed By	<input type="checkbox"/> Battery Replacement		<input type="checkbox"/> Dry Gas Regulator Replacement		<input type="checkbox"/> Breath Tube Replacement		<input type="checkbox"/> Other _____	
Simulator	Serial #	Lot #/Exp																										
0.05	SD1018	201507A 7/14/17																										
0.08	SD1011	201601F 1/26/18																										
0.20	SD1025	201604C 4/5/18																										
0.08 DGS	N/A	A9162405 5/3/18																										
Maintenance	Performed By																											
<input type="checkbox"/> Battery Replacement																												
<input type="checkbox"/> Dry Gas Regulator Replacement																												
<input type="checkbox"/> Breath Tube Replacement																												
<input type="checkbox"/> Other _____																												
				<table border="1"> <thead> <tr> <th>Suggested Service</th> </tr> </thead> <tbody> <tr> <td>_____</td> </tr> <tr> <td>_____</td> </tr> <tr> <td>_____</td> </tr> </tbody> </table>	Suggested Service	_____	_____	_____																				
Suggested Service																												

RECEIVED
 SEP 29 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration	Performed By <u>DB</u>	Department Inspection	Performed By <u>DB</u>																																								
<input type="checkbox"/> Optical Bench Calibration N/A <input checked="" type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge 1015 ID# <u>26932</u>		<input checked="" type="checkbox"/> Barometric Pressure <u>1016</u> Gauge ID# <u>28427</u> <u>1013</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2016-R</u>																																									
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>G4444</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>G2882</td> <td>16101</td> <td>2/2/18</td> </tr> <tr> <td>0.100</td> <td>G2078</td> <td>16001</td> <td>5/8/18</td> </tr> <tr> <td>0.200</td> <td>G2408</td> <td>16103</td> <td>4/14/18</td> </tr> <tr> <td>0.400</td> <td>G5358</td> <td>16102</td> <td>3/22/18</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td><u>154</u> AG612405</td> <td>7/5/17</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000	G4444	N/A	N/A	0.040	G2882	16101	2/2/18	0.100	G2078	16001	5/8/18	0.200	G2408	16103	4/14/18	0.400	G5358	16102	3/22/18	0.080 DGS	N/A	<u>154</u> AG612405	7/5/17		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1019</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	SD1019	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	SD1025	
Simulator	Serial Number	Lot Number	Expiration																																								
0.000	G4444	N/A	N/A																																								
0.040	G2882	16101	2/2/18																																								
0.100	G2078	16001	5/8/18																																								
0.200	G2408	16103	4/14/18																																								
0.400	G5358	16102	3/22/18																																								
0.080 DGS	N/A	<u>154</u> AG612405	7/5/17																																								
Simulator	Serial Number																																										
0.00	SD1019																																										
Interferent	SD1021																																										
0.05	SD1018																																										
0.08	SD1011																																										
0.20	SD1025																																										
<input checked="" type="checkbox"/> Post Calibration Stability Checks <u>15615080A2</u> <u>DB</u>		<table border="1"> <thead> <tr> <th>Attachments</th> </tr> </thead> <tbody> <tr> <td><input checked="" type="checkbox"/> Form 41</td> </tr> <tr> <td><input checked="" type="checkbox"/> Pre-Stability Tests</td> </tr> <tr> <td><input type="checkbox"/> Flow Calibration</td> </tr> <tr> <td><input checked="" type="checkbox"/> Optical Bench Cal</td> </tr> <tr> <td><input checked="" type="checkbox"/> Post-Stability Tests</td> </tr> <tr> <td><input type="checkbox"/> Other _____</td> </tr> </tbody> </table>	Attachments	<input checked="" type="checkbox"/> Form 41	<input checked="" type="checkbox"/> Pre-Stability Tests	<input type="checkbox"/> Flow Calibration	<input checked="" type="checkbox"/> Optical Bench Cal	<input checked="" type="checkbox"/> Post-Stability Tests	<input type="checkbox"/> Other _____																																		
Attachments																																											
<input checked="" type="checkbox"/> Form 41																																											
<input checked="" type="checkbox"/> Pre-Stability Tests																																											
<input type="checkbox"/> Flow Calibration																																											
<input checked="" type="checkbox"/> Optical Bench Cal																																											
<input checked="" type="checkbox"/> Post-Stability Tests																																											
<input type="checkbox"/> Other _____																																											
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A</td> <td>7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F</td> <td>1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C</td> <td>4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG612405</td> <td>5/3/18</td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.05	SD1018	201507A	7/14/17	0.08	SD1011	201601F	1/26/18	0.20	SD1025	201604C	4/5/18	0.08 DGS	N/A	AG612405	5/3/18																							
Simulator	Serial Number	Lot Number	Expiration																																								
0.05	SD1018	201507A	7/14/17																																								
0.08	SD1011	201601F	1/26/18																																								
0.20	SD1025	201604C	4/5/18																																								
0.08 DGS	N/A	AG612405	5/3/18																																								

Notes: Optical bench calibration performed to bring values closer to nominal; DB
DB-pms

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC
<input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC
<input checked="" type="checkbox"/> Return to/Place into Evidentiary Use
<input type="checkbox"/> Remain Out of Evidentiary Use
<input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use

Brett Kurland
 Quality Control Review

9/23/16
 Date

Pre-Car Stability Checks #80-005249 Hernando County S.D. 9/21/16 DMS

DMS

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
09/21/2016
Software: 8100.27

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
09/21/2016
Software: 8100.27

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
09/21/2016
Software: 8100.27

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005249
09/21/2016
Software: 8100.27

Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time
Air Blank	0.000	10:52	Air Blank	0.000	10:57	Air Blank	0.000	11:03	Air Blank	0.000	11:10
Control Test	0.049	10:53	Control Test	0.080	10:58	Control Test	0.197	11:04	Control Test	0.083	11:10
Air Blank	0.000	10:54	Air Blank	0.000	10:58	Air Blank	0.000	11:04	Air Blank	0.000	11:11
Control Test	0.049	10:54	Control Test	0.080	10:59	Control Test	0.197	11:05	Control Test	0.084	11:11
Air Blank	0.000	10:55	Air Blank	0.000	11:00	Air Blank	0.000	11:06	Air Blank	0.000	11:11
Control Test	0.049	10:55	Control Test	0.080	11:00	Control Test	0.198	11:06	Control Test	0.083	11:12
Air Blank	0.000	10:56	Air Blank	0.000	11:01	Air Blank	0.000	11:07	Air Blank	0.000	11:12
Control Test Stats			Control Test Stats			Control Test Stats			Control Test Stats		
Average	0.0490		Average	0.0800		Average	0.1973		Average	0.0833	
Std Dev	0.0000		Std Dev	0.0000		Std Dev	0.0006		Std Dev	0.0006	
Rel Std Dev(%)	0.0000		Rel Std Dev(%)	0.0000		Rel Std Dev(%)	0.2926		Rel Std Dev(%)	0.6928	

Find BK

DMS

Operator's Signature

DMS

Operator's Signature

DMS

Operator's Signature

DMS

Operator's Signature

Optical Bench Calibration Data #80-005249 Hernandez County S.O. 9/21/16 AUS

AMS

Solution Quadratic Fit Channel 1

HERNANDEZ County S.O. AUS
 Intoxilyzer Model 8000
 SN 80-005249
 09/21/2016 11:14:01

Auto Calibration
 Max Power Res Value = 58
 Auto Range Res Value = 36

Sample % Abs (% Abs Ref)
 Sample #1 = 1.4780 (-0.0160)
 Sample #2 = 1.5030 (-0.0180)
 Sample #3 = 1.4910 (-0.0060)
 Sample #4 = 1.5180 (-0.0070)
 Avg % Abs = 1.5040 (-0.0103)
 STD DEV = 0.0135 (0.0067)
 REL STD DEV = 0.899 (64.435)

Sample % Abs (% Abs Ref)
 Sample #1 = 6.8180 (-0.0190)
 Sample #2 = 6.7920 (0.0240)
 Sample #3 = 6.8300 (0.0080)
 Sample #4 = 6.8120 (0.0150)
 Avg % Abs = 6.8113 (0.0157)
 STD DEV = 0.0190 (0.0080)
 REL STD DEV = 0.279 (51.197)

Solution	Act	Fit	Residual
g/210L	g/210L	g/210L	g/210L
0.000	0.000	0.000	-0.0004
0.040	0.040	0.040	0.0000
0.100	0.099	0.099	0.0010
0.200	0.201	0.201	-0.0008
0.400	0.400	0.400	0.0001

SoI Value = 0.000 g/210L ***
 Fit value = 0.0000 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12600, Sum Io = 13672
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.0740 (-0.0030)
 Sample #2 = 0.0610 (0.0390)
 Sample #3 = 0.0250 (0.0620)
 Sample #4 = 0.0790 (0.0550)
 Avg % Abs = 0.0550 (0.0520)
 STD DEV = 0.0275 (0.0118)
 REL STD DEV = 49.992 (22.673)

SoI Value = 0.100 g/210L ***
 Fit value = 0.4762 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12583, Sum Io = 13665
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.8250 (-0.0140)
 Sample #2 = 1.8050 (-0.0010)
 Sample #3 = 1.8320 (0.0000)
 Sample #4 = 1.7900 (0.0140)
 Avg % Abs = 1.8090 (0.0043)
 STD DEV = 0.0213 (0.0084)
 REL STD DEV = 1.177 (193.535)

SoI Value = 0.400 g/210L ***
 Fit value = 1.9048 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12579, Sum Io = 13665
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 6.8180 (-0.0210)
 Sample #2 = 6.7970 (0.0310)
 Sample #3 = 6.8190 (0.0190)
 Sample #4 = 6.8520 (0.0240)
 Avg % Abs = 6.8227 (0.0247)
 STD DEV = 0.0277 (0.0060)
 REL STD DEV = 0.406 (24.437)

SoI Value = 0.400 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 <<<< CHANNEL 2 >>>>
 SoI Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.135
 Std Dev = 0.02 Rel Std Dev = 12.98
 SoI Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.504
 Std Dev = 0.01 Rel Std Dev = 0.90
 SoI Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.527
 Std Dev = 0.02 Rel Std Dev = 0.54
 SoI Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.811
 Std Dev = 0.02 Rel Std Dev = 0.28
 SoI Val = 1.9048 mg/l or 0.400 g/210L
 % Abs = 12.686
 Std Dev = 0.02 Rel Std Dev = 0.15
 Zero Order Coef = -146.31
 First Order Coef = 1326.93
 Second Order Coef = 14.60
 Standard Deviation = 33.5333

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0007
0.040	0.040	0.0005
0.100	0.099	0.0010
0.200	0.201	-0.0010
0.400	0.400	0.0002

<<<< CHANNEL 2 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.1720 (-0.0190)
 Sample #2 = 0.1200 (0.0180)
 Sample #3 = 0.1300 (0.0230)
 Sample #4 = 0.1540 (0.0140)
 Avg % Abs = 0.1347 (0.0183)
 STD DEV = 0.0175 (0.0045)
 REL STD DEV = 12.976 (24.596)

<<<< CHANNEL 2 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.5000 (-0.0080)
 Sample #2 = 3.5120 (0.0000)
 Sample #3 = 3.5480 (-0.0070)
 Sample #4 = 3.5200 (0.0090)
 Avg % Abs = 3.5267 (0.0007)
 STD DEV = 0.0189 (0.0080)
 REL STD DEV = 0.536 (1203.121)

<<<< CHANNEL 2 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 12.6540 (-0.0120)
 Sample #2 = 12.6660 (0.0520)
 Sample #3 = 12.6880 (0.0450)
 Sample #4 = 12.7040 (0.0400)
 Avg % Abs = 12.6660 (0.0457)
 STD DEV = 0.0191 (0.0060)
 REL STD DEV = 0.150 (13.199)

<<<< CHANNEL 2 >>>>
 SoI Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.135
 Std Dev = 0.02 Rel Std Dev = 12.98
 SoI Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.504
 Std Dev = 0.01 Rel Std Dev = 0.90
 SoI Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.527
 Std Dev = 0.02 Rel Std Dev = 0.54
 SoI Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.811
 Std Dev = 0.02 Rel Std Dev = 0.28
 SoI Val = 1.9048 mg/l or 0.400 g/210L
 % Abs = 12.686
 Std Dev = 0.02 Rel Std Dev = 0.15
 Zero Order Coef = -146.31
 First Order Coef = 1326.93
 Second Order Coef = 14.60
 Standard Deviation = 33.5333

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0007
0.040	0.040	0.0005
0.100	0.099	0.0010
0.200	0.201	-0.0010
0.400	0.400	0.0002

SoI Value = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12588, Sum Io = 13667
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.7290 (-0.0070)
 Sample #2 = 0.7730 (-0.0030)
 Sample #3 = 0.7540 (0.0000)
 Sample #4 = 0.7740 (0.0120)
 Avg % Abs = 0.7670 (0.0030)

SoI Value = 0.200 g/210L ***
 Fit value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12584, Sum Io = 13666
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.5830 (-0.0130)
 Sample #2 = 3.5340 (0.0340)
 Sample #3 = 3.6030 (0.0220)
 Sample #4 = 3.5390 (0.0520)

SoI Value = 0.400 g/210L ***
 Fit value = 1.9048 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12579, Sum Io = 13665
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 6.8180 (-0.0190)
 Sample #2 = 6.7920 (0.0240)
 Sample #3 = 6.8300 (0.0080)
 Sample #4 = 6.8120 (0.0150)
 Avg % Abs = 6.8113 (0.0157)
 STD DEV = 0.0190 (0.0080)
 REL STD DEV = 0.279 (51.197)

SoI Value = 0.400 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 <<<< CHANNEL 2 >>>>
 SoI Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.135
 Std Dev = 0.02 Rel Std Dev = 12.98
 SoI Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.504
 Std Dev = 0.01 Rel Std Dev = 0.90
 SoI Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.527
 Std Dev = 0.02 Rel Std Dev = 0.54
 SoI Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.811
 Std Dev = 0.02 Rel Std Dev = 0.28
 SoI Val = 1.9048 mg/l or 0.400 g/210L
 % Abs = 12.686
 Std Dev = 0.02 Rel Std Dev = 0.15
 Zero Order Coef = -146.31
 First Order Coef = 1326.93
 Second Order Coef = 14.60
 Standard Deviation = 33.5333

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.001	-0.0007
0.040	0.040	0.0005
0.100	0.099	0.0010
0.200	0.201	-0.0010
0.400	0.400	0.0002

Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1014
 3 um H2O Adjust (mg/l*10,000) = 396
 9 um H2O A.
 **** AUTO CRI PASS

Standard Deviation = 38.565609
 Standard Deviation = 38.565609

Avg % Abs = 3.5587 (0.0360)
 STD DEV = 0.0385 (0.0151)
 REL STD DEV = 1.081 (41.944)

STD DEV = 0.2113 (0.0079)
 REL STD DEV = 1.4669 (264.575)

9 um H2O Adjust (mg/l*10,000) = 396
 AMS

AMS

AMS

AMS

Post Calibration Stability Checks #80-005249

DES

Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time
Air Blank	0.000	12:32	Air Blank	0.000	13:04	Air Blank	0.000	12:18	Air Blank	0.000	12:18	Air Blank	0.000	12:27
Control Test	0.050	12:33	Control Test	0.078	13:05	Control Test	0.078	12:19	Control Test	0.194	12:19	Control Test	0.081	12:27
Air Blank	0.000	12:34	Air Blank	0.000	13:05	Air Blank	0.000	12:20	Air Blank	0.000	12:20	Air Blank	0.000	12:28
Control Test	0.049	12:34	Control Test	0.079	13:06	Control Test	0.079	12:20	Control Test	0.197	12:20	Control Test	0.080	12:28
Air Blank	0.000	12:35	Air Blank	0.000	13:07	Air Blank	0.000	12:21	Air Blank	0.000	12:21	Air Blank	0.000	12:29
Control Test	0.050	12:36	Control Test	0.079	13:07	Control Test	0.079	12:22	Control Test	0.197	12:22	Control Test	0.081	12:29
Air Blank	0.000	12:36	Air Blank	0.000	13:08	Air Blank	0.000	12:22	Air Blank	0.000	12:22	Air Blank	0.000	12:30
Control Test Stats			Control Test Stats			Control Test Stats			Control Test Stats			Control Test Stats		
Average	0.0497		Average	0.0787		Average	0.0787		Average	0.1960		Average	0.0807	
Std Dev	0.0006		Std Dev	0.0015		Std Dev	0.0006		Std Dev	0.0017		Std Dev	0.0006	
Rel Std Dev(%)	1.1625		Rel Std Dev(%)	1.9418		Rel Std Dev(%)	0.7339		Rel Std Dev(%)	0.8837		Rel Std Dev(%)	0.7157	

ASK

DES

DES

DES

DES

DES

Operator's Signature

Operator's Signature

Operator's Signature

Operator's Signature

Operator's Signature

Retested - Cold Counter

ms