



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Pinellas County SO S/N 80-005290  
Date In 2/2/16 Date Out 2/23/16  Ship  P/U  H/D  CMI  EE

**Intake** Performed By TP

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By DRB

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 182  
 Flow Verification (L/s)  
 Flow Column # ATP103  
 32mm 0.160 (.139 - .169)  
 36mm 0.179 (.156 - .190)  
 53mm 0.242 (.228 - .278)  
 103mm 0.500 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G11739	201507A 7/14/17
0.08	G8149	201502A 2/24/17
0.20	G11621	201505A 5/12/17
0.08 DGS	N/A	AG5117D1 4/27/17

**Flow Calibration** Performed By DRB

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 47mm  
 15L/min - 59mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service** \_\_\_\_\_

RECEIVED  
FEB 23 2016  
ALCOHOL TESTING PROGRAM  
FDLE

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By DRB

Barometric Pressure 1015 Gauge  
 ID# 26932 1014 Instrument

Mouth Alcohol Solution Lot # 2015-A  
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	G2879
Interferent	G8144
0.05	G11739
0.08	G8149
0.20	G11621

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration  
 Optical Bench Cal  
 Post-Stability Tests  
 Other Form 40

Notes: QC-BK

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy  
Quality Control Review

2/23/16  
Date

Stability Checks 80-005290 Pinellas County SO. 7/23/16 RMB

RMB

RMB

PINELLAS COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005290  
02/22/2016  
Software: 8100.27

PINELLAS COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005290  
02/22/2016  
Software: 8100.27

PINELLAS COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-005290  
02/22/2016  
Software: 8100.27

Test	9/21/0L	Time
Air Blank	0.000	10:39
Control Test	0.008	10:39
Air Blank	0.000	10:40
Control Test	0.049	10:41
Air Blank	0.000	10:41
Control Test	0.049	10:42
Air Blank	0.000	10:42
Control Test Stats		
Average	0.0487	
Std Dev	0.0085	
Rel Std Dev(%)	1.1863	

Test	9/21/0L	Time
Air Blank	0.000	10:43
Control Test	0.078	10:44
Air Blank	0.000	10:45
Control Test	0.078	10:45
Air Blank	0.000	10:46
Control Test	0.078	10:47
Air Blank	0.000	10:47
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	9/21/0L	Time
Air Blank	0.000	10:48
Control Test	0.201	10:49
Air Blank	0.000	10:50
Control Test	0.201	10:50
Air Blank	0.000	10:51
Control Test	0.201	10:52
Air Blank	0.000	10:52
Control Test Stats		
Average	0.2010	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	9/21/0L	Time
Air Blank	0.000	10:53
Control Test	0.079	10:54
Air Blank	0.000	10:54
Control Test	0.079	10:54
Air Blank	0.000	10:55
Control Test	0.079	10:55
Air Blank	0.000	10:56
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

BSK

RMB

Operator's Signature

RMB

Operator's Signature

RMB

Operator's Signature