

INSTRUMENT PROCESSING SHEET

Agency St. Petersburg PD S/N 80-005338
 Date In 2/29/16 Date Out 3/8/16 Ship P/U H/D CMI EE

Intake Performed By <u>TR</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>X</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u> Notes: <u>* Missing two feet</u>	Quality Checks Performed By <u>RDR</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>219</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP103</u> 32mm <u>0.144</u> (.139 - .169) 36mm <u>0.160</u> (.156 - .190) 53mm <u>0.234</u> (.228 - .278) 103mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
		Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service

RECEIVED
 MAR 09 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____ <input type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.400				0.080 DGS	N/A			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A			Department Inspection Performed By <u>RDR</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1022</u> Gauge ID# <u>28427</u> <u>1021</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2016-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td>SD1022</td> </tr> <tr> <td>Interferent</td> <td>SD1021</td> </tr> <tr> <td>0.05</td> <td>SD1018</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> </tr> <tr> <td>0.20</td> <td>G4444</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	Simulator	Serial Number	0.00	SD1022	Interferent	SD1021	0.05	SD1018	0.08	SD1011	0.20	G4444
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.00	SD1022																																																												
Interferent	SD1021																																																												
0.05	SD1018																																																												
0.08	SD1011																																																												
0.20	G4444																																																												
Notes: <u>QC-BK</u>	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																												

Patricia Murphy
 Quality Control Review

3/9/16
 Date

Stability Checks 80-005338 St. Petersburg P.D. 3/8/16 **DUB**

FBM

ES

ST. PETERSBURG PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005338
03/08/2016
Software: 8100.27

ST. PETERSBURG PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005338
03/26/2016
Software: 8100.27

ST. PETERSBURG PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005338
03/08/2016
Software: 8100.27

ST. PETERSBURG PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005338
03/08/2016
Software: 8100.27

Test	9/21/L	Time
Air Blank	0.000	12:34
Control Test	0.087	12:34
Air Blank	0.000	12:54
Control Test	0.081	12:55
Air Blank	0.000	12:55
Control Test	0.082	12:55
Air Blank	0.000	12:55
Control Test Stats		
Average	0.082	
Std Dev	0.000	
Rel. Std Dev(%)	1.2195	

Test	9/21/L	Time
Air Blank	0.000	12:48
Control Test	0.200	12:49
Air Blank	0.000	12:50
Control Test	0.200	12:50
Air Blank	0.000	12:51
Control Test	0.200	12:52
Air Blank	0.000	12:52
Control Test Stats		
Average	0.200	
Std Dev	0.000	
Rel. Std Dev(%)	0.0000	

Test	9/21/L	Time
Air Blank	0.000	12:40
Control Test	0.081	12:41
Air Blank	0.000	12:41
Control Test	0.081	12:42
Air Blank	0.000	12:42
Control Test	0.081	12:43
Air Blank	0.000	12:44
Control Test Stats		
Average	0.081	
Std Dev	0.000	
Rel. Std Dev(%)	0.0000	

Test	9/21/L	Time
Air Blank	0.000	12:33
Control Test	0.050	12:34
Air Blank	0.000	12:35
Control Test	0.052	12:35
Air Blank	0.000	12:36
Control Test	0.052	12:37
Air Blank	0.000	12:37
Control Test Stats		
Average	0.051	
Std Dev	0.000	
Rel. Std Dev(%)	2.2494	

ASK

DUB

DUB

DUB

DUB

Operator's Signature

Operator's Signature

Operator's Signature

Operator's Signature