

INSTRUMENT PROCESSING SHEET

Agency Ocala PD S/N 80-005454
Date In 5/23/16 Date Out 6/8/16 Ship P/U H/D CMI EE

RECEIVED
JUN 08 2016
FDLE
Alcohol Testing Program

Intake Performed By (Signature)

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By (Signature)

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 167
 Flow Verification (L/s)
Flow Column # ATP.102
32mm 148 (.139 - .169)
36mm 171 (.156 - .190)
53mm 246 (.228 - .278)
103mm 500 (.447 - .547)

Barometric Pressure Check
Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>62403</u>	<u>201507A</u> <u>7/14/17</u>
0.08	<u>503964</u>	<u>201601F</u> <u>1/26/18</u>
0.20	<u>64444</u>	<u>201505A</u> <u>5/12/17</u>
0.08 DGS	<u>N/A</u>	<u>AG600504</u> <u>01/05/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____

Post Calibration Verification (L/s)
Flow Column # _____
32mm _____ (.139 - .169)
36mm _____ (.156 - .190)
53mm _____ (.228 - .278)
103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		<u>N/A</u>	<u>N/A</u>
0.040			
0.100			
0.200			
0.400			
0.080 DGS	<u>N/A</u>		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	<u>N/A</u>		

Department Inspection Performed By (Signature)

Barometric Pressure 1007 Gauge
ID# 28427 1006 Instrument

Mouth Alcohol Solution Lot # 2016-A
Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>62880</u>
Interferent	<u>62884</u>
0.05	<u>62403</u>
0.08	<u>503964</u>
0.20	<u>64444</u>

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: (Signature) - PWS

Brett Whitland
Quality Control Review

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

6/8/16
Date