

INSTRUMENT PROCESSING SHEET

Agency Miami Dade Police Department S/N 80-005550
 Date In 1/8/2016 Date Out 1/8/2016 Ship P/U H/D CMI EE

Intake Performed By <u>DELR</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>ANTI STATIC BAG</u> Notes: _____ _____ _____	Quality Checks Performed By <u>DELL</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>155</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32mm <u>167</u> (.139 - .169) 36mm <u>187</u> (.156 - .190) 53mm <u>257</u> (.228 - .278) 103mm <u>537</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td><u>SD3967</u></td> <td><u>201507A</u> <u>07/14/2017</u></td> </tr> <tr> <td>0.08</td> <td><u>SD3968</u></td> <td><u>2015025</u> <u>02/24/2017</u></td> </tr> <tr> <td>0.20</td> <td><u>SD3969</u></td> <td><u>201505A</u> <u>05/12/2017</u></td> </tr> <tr> <td>0.08 DGS</td> <td><u>N/A</u></td> <td><u>96507503</u> <u>02/16/2017</u></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	<u>SD3967</u>	<u>201507A</u> <u>07/14/2017</u>	0.08	<u>SD3968</u>	<u>2015025</u> <u>02/24/2017</u>	0.20	<u>SD3969</u>	<u>201505A</u> <u>05/12/2017</u>	0.08 DGS	<u>N/A</u>	<u>96507503</u> <u>02/16/2017</u>	Flow Calibration Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	<u>SD3967</u>	<u>201507A</u> <u>07/14/2017</u>															
0.08	<u>SD3968</u>	<u>2015025</u> <u>02/24/2017</u>															
0.20	<u>SD3969</u>	<u>201505A</u> <u>05/12/2017</u>															
0.08 DGS	<u>N/A</u>	<u>96507503</u> <u>02/16/2017</u>															
Maintenance Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ Suggested Service _____ _____																	

RECEIVED
 JAN 21 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td><u>N/A</u></td> <td><u>N/A</u></td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.400</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td><u>N/A</u></td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td><u>N/A</u></td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	Lot Number	Expiration	0.000		<u>N/A</u>	<u>N/A</u>	0.040				0.100				0.200				0.400				0.080 DGS	<u>N/A</u>			Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	<u>N/A</u>			Department Inspection Performed By <u>DELR</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1012</u> Gauge ID# <u>28663</u> <u>1012</u> Instrument Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2015-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.00</td> <td><u>SD 3965</u></td> </tr> <tr> <td>Interferent</td> <td><u>SD 3966</u></td> </tr> <tr> <td>0.05</td> <td><u>SD 3967</u></td> </tr> <tr> <td>0.08</td> <td><u>SD 3968</u></td> </tr> <tr> <td>0.20</td> <td><u>SD 3969</u></td> </tr> </tbody> </table>	Simulator	Serial Number	0.00	<u>SD 3965</u>	Interferent	<u>SD 3966</u>	0.05	<u>SD 3967</u>	0.08	<u>SD 3968</u>	0.20	<u>SD 3969</u>
Simulator	Serial Number	Lot Number	Expiration																																																										
0.000		<u>N/A</u>	<u>N/A</u>																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.400																																																													
0.080 DGS	<u>N/A</u>																																																												
Simulator	Serial Number	Lot Number	Expiration																																																										
0.05																																																													
0.08																																																													
0.20																																																													
0.08 DGS	<u>N/A</u>																																																												
Simulator	Serial Number																																																												
0.00	<u>SD 3965</u>																																																												
Interferent	<u>SD 3966</u>																																																												
0.05	<u>SD 3967</u>																																																												
0.08	<u>SD 3968</u>																																																												
0.20	<u>SD 3969</u>																																																												
Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____																																																													

Notes: **E-MAILED** **APPROVED**
1-8-2016

QC-TSK

Patrick Murphy
 Quality Control Review

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

1/8/16
 Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-005550	Miami Dade Police Department	01/08/2016	<i>SK</i>

SK

0.05g/210L 0.047 to 0.053 <input checked="" type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>																																																																																																																																																
<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005550 01/08/2016 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:37</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:38</td></tr> <tr><td>Control Test</td><td>0.051</td><td>11:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:39</td></tr> <tr><td>Control Test</td><td>0.051</td><td>11:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:40</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0507</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1395</td><td></td></tr> </tbody> </table> <p><i>SK</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	11:37	Control Test	0.050	11:37	Air Blank	0.000	11:38	Control Test	0.051	11:39	Air Blank	0.000	11:39	Control Test	0.051	11:40	Air Blank	0.000	11:40	Control Test Stats			Average	0.0507		Std Dev	0.0006		Rel Std Dev(%)	1.1395		<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005550 01/08/2016 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:42</td></tr> <tr><td>Control Test</td><td>0.088</td><td>11:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:43</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:44</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:45</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table> <p><i>SK</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	11:42	Control Test	0.088	11:43	Air Blank	0.000	11:43	Control Test	0.080	11:44	Air Blank	0.000	11:44	Control Test	0.079	11:45	Air Blank	0.000	11:45	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247		<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005550 01/08/2016 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:47</td></tr> <tr><td>Control Test</td><td>0.201</td><td>11:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:48</td></tr> <tr><td>Control Test</td><td>0.200</td><td>11:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:49</td></tr> <tr><td>Control Test</td><td>0.200</td><td>11:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:50</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2003</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2882</td><td></td></tr> </tbody> </table> <p><i>SK</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	11:47	Control Test	0.201	11:48	Air Blank	0.000	11:48	Control Test	0.200	11:49	Air Blank	0.000	11:49	Control Test	0.200	11:50	Air Blank	0.000	11:50	Control Test Stats			Average	0.2003		Std Dev	0.0006		Rel Std Dev(%)	0.2882		<p>MIAMI DADE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005550 01/08/2016 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:52</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:53</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:54</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:55</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p><i>SK</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	11:52	Control Test	0.079	11:52	Air Blank	0.000	11:53	Control Test	0.079	11:53	Air Blank	0.000	11:54	Control Test	0.079	11:54	Air Blank	0.000	11:55	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:37																																																																																																																																																	
Control Test	0.050	11:37																																																																																																																																																	
Air Blank	0.000	11:38																																																																																																																																																	
Control Test	0.051	11:39																																																																																																																																																	
Air Blank	0.000	11:39																																																																																																																																																	
Control Test	0.051	11:40																																																																																																																																																	
Air Blank	0.000	11:40																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0507																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1395																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:42																																																																																																																																																	
Control Test	0.088	11:43																																																																																																																																																	
Air Blank	0.000	11:43																																																																																																																																																	
Control Test	0.080	11:44																																																																																																																																																	
Air Blank	0.000	11:44																																																																																																																																																	
Control Test	0.079	11:45																																																																																																																																																	
Air Blank	0.000	11:45																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:47																																																																																																																																																	
Control Test	0.201	11:48																																																																																																																																																	
Air Blank	0.000	11:48																																																																																																																																																	
Control Test	0.200	11:49																																																																																																																																																	
Air Blank	0.000	11:49																																																																																																																																																	
Control Test	0.200	11:50																																																																																																																																																	
Air Blank	0.000	11:50																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2003																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2882																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:52																																																																																																																																																	
Control Test	0.079	11:52																																																																																																																																																	
Air Blank	0.000	11:53																																																																																																																																																	
Control Test	0.079	11:53																																																																																																																																																	
Air Blank	0.000	11:54																																																																																																																																																	
Control Test	0.079	11:54																																																																																																																																																	
Air Blank	0.000	11:55																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		