



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Fort Myers PD

S/N 80-005655

Date In 9/30/16

Date Out 10/3/16

Ship P/U H/D CMI EE

Intake Performed By SP

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
ok Case ok Handle
ok Dry Gas Holder ok Feet
ok Keyboard/Plug ok Back/Plugs
ok Screws tight ok Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By SP

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 195
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 148 (.139 - .169)
 36mm 164 (.156 - .190)
 53mm 238 (.228 - .278)
 103mm 511 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7-14-17
0.08	SD1011	201601F 1-26-18
0.20	SD1025	201604C 4-5-18
0.08 DGS	N/A	AG612405 5-3-18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 50mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
OCT 03 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By SP

Barometric Pressure 1013 Gauge
 ID# 28427 1010 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SP201507A-SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other FORM 40

Notes: OK - PWS

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
Quality Control Review

10/3/16
Date

STABILITY CHECKS - INSTRUMENT #80-005655 - FORT MYERS PD - 10/13/16 SP

SP

DGS

FORT MEYERS PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005655
10/13/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:10
Control Test	0.050	10:11
Air Blank	0.000	10:11
Control Test	0.051	10:12
Air Blank	0.000	10:13
Control Test	0.051	10:13
Air Blank	0.000	10:14
Control Test Stats		
Average	0.0507	
Std Dev	0.0006	
Rel Std Dev(%)	1.1395	

SP

Operator's Signature

FORT MEYERS PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005655
10/13/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:21
Control Test	0.080	10:21
Air Blank	0.000	10:22
Control Test	0.081	10:22
Air Blank	0.000	10:23
Control Test	0.081	10:24
Air Blank	0.000	10:24
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

SP

Operator's Signature

FORT MEYERS PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005655
10/13/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:15
Control Test	0.199	10:16
Air Blank	0.000	10:16
Control Test	0.197	10:17
Air Blank	0.000	10:18
Control Test	0.197	10:18
Air Blank	0.000	10:19
Control Test Stats		
Average	0.1977	
Std Dev	0.0012	
Rel Std Dev(%)	0.5842	

SP

Operator's Signature

FORT MEYERS PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005655
10/13/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:26
Control Test	0.078	10:27
Air Blank	0.000	10:27
Control Test	0.077	10:28
Air Blank	0.000	10:28
Control Test	0.078	10:29
Air Blank	0.000	10:29
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

BK

SP

Operator's Signature