

INSTRUMENT PROCESSING SHEET

Agency Palm Beach County Sheriff's Office/IN 80-006029
Date In 5/16/2016 Date Out 5/16/2016 Ship P/U H/D CMI EE

Intake Performed By DELR

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By DELR

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 208
 Flow Verification (L/s)
 Flow Column # ATP 101
 32mm 148 (.139 - .169)
 36mm 171 (.156 - .190)
 53mm 234 (.228 - .278)
 103mm 496 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 281603

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD3967	201507A 07/14/2017
0.08	SD3968	201601F 01/26/2018
0.20	SD3969	201505A 05/12/2017
0.08 DGS	N/A	A66005D4 01/05/2018

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 103mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
MAY 25 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DELR

Barometric Pressure 1017 Gauge
 ID# 28199 1016 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD 3965
Interferent	SD 3966
0.05	SD 3967
0.08	SD 3968
0.20	SD 3969

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: **E-MAILED** **APPROVED**
5/16/2016
PR/DC OK (PCM)

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Kirkland
Quality Control Review

5/25/16
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-006029	Palm Beach County Sheriff's Office	05/16/2016	<i>WLL</i>

194

0.05g/210L <input checked="" type="checkbox"/>	0.08g/210L <input checked="" type="checkbox"/>	0.20g/210L <input checked="" type="checkbox"/>	DGS 0.08g/210L <input checked="" type="checkbox"/>
PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006029 05/16/2016 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:23 Control Test 0.050 09:24 Air Blank 0.000 09:25 Control Test 0.050 09:25 Air Blank 0.000 09:26 Control Test 0.050 09:26 Air Blank 0.000 09:27 Control Test Stats Average 0.0500 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature: <i>WLL</i>	PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006029 05/16/2016 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:33 Control Test 0.081 09:34 Air Blank 0.000 09:34 Control Test 0.080 09:35 Air Blank 0.000 09:35 Control Test 0.081 09:36 Air Blank 0.000 09:37 Control Test Stats Average 0.0807 Std Dev 0.0006 Rel Std Dev(%) 0.7157 Operator's Signature: <i>WLL</i>	PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006029 05/16/2016 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:39 Control Test 0.205 09:40 Air Blank 0.000 09:40 Control Test 0.205 09:41 Air Blank 0.000 09:42 Control Test 0.206 09:42 Air Blank 0.000 09:43 Control Test Stats Average 0.2053 Std Dev 0.0006 Rel Std Dev(%) 0.2812 Operator's Signature: <i>WLL</i>	PALM BEACH CO SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006029 05/16/2016 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:45 Control Test 0.082 09:46 Air Blank 0.000 09:46 Control Test 0.082 09:46 Air Blank 0.000 09:47 Control Test 0.082 09:47 Air Blank 0.000 09:48 Control Test Stats Average 0.0820 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature: <i>WLL</i>

WLL