

INSTRUMENT PROCESSING SHEET

Agency Martin County SO S/N 80-006168
 Date In 7/12/16 Date Out 7/28/16 Ship P/U H/D CMI EE

Intake Performed By [Signature]

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 Case OK Handle OK
 Dry Gas Holder OK Feet OK
 Keyboard/Plug OK Back/Plugs OK
 Screws tight OK Breath Hose OK

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 187
 Flow Verification (L/s)
 Flow Column # A7P102
 32mm 152 (.139 - .169)
 36mm 168 (.156 - .190)
 53mm 253 (.228 - .278)
 103mm 504 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G4444	201507A 7/14/17
0.08	G6621	201601F 1/26/18
0.20	SD1012	201505A 5/12/17
0.08 DGS	N/A	AG6612405 5/3/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 50mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/gram)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
JUL 29 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1016 Gauge
 ID# 28427 1015 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2835
Interferent	G2840
0.05	G4444
0.08	G6621
0.20	SD1012

Attachments

<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Optical Bench Cal
<input checked="" type="checkbox"/> Pre-Stability Tests	<input type="checkbox"/> Post-Stability Tests
<input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Other _____

Notes: OK / OK OK (GPM) 7/28/16

Brett Highland
 Quality Control Review

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

7/29/16
Date

80-006168
 Stability Checks
 7/28/16

INTOXILYZER 8000
 Instrument Initialization
 08:00 07/28/2016

AK

MARTIN COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006168
 07/28/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:53
Control Test	0.200	08:53
Air Blank	0.000	08:54
Control Test	0.201	08:54
Air Blank	0.000	08:55
Control Test	0.202	08:56
Air Blank	0.000	08:56
Control Test Stats		
Average	0.2010	
Std Dev	0.0010	
Rel Std Dev(%)	0.4975	

Operator's Signature

MARTIN COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006168
 07/28/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:57
Control Test	0.079	08:58
Air Blank	0.000	08:59
Control Test	0.079	08:59
Air Blank	0.000	09:00
Control Test	0.080	09:01
Air Blank	0.000	09:01
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Operator's Signature

MARTIN COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006168
 07/28/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:34
Control Test	0.080	08:35
Air Blank	0.000	08:35
Control Test	0.081	08:35
Air Blank	0.000	08:36
Control Test	0.080	08:36
Air Blank	0.000	08:36
Control Test	0.080	08:37
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

DOS

Operator's Signature

MARTIN COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006168
 07/28/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:03
Control Test	0.049	09:03
Air Blank	0.000	09:04
Control Test	0.050	09:05
Air Blank	0.000	09:05
Control Test	0.049	09:06
Air Blank	0.000	09:06
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

Operator's Signature

INSTRUMENT PROCESSING SHEET

Agency Martin County SO S/N 80-006168
 Date In 1/8/16 Date Out 1/25/16 Ship P/U H/D CMI EE

Intake Performed By JP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By [Signature]

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 188
 Flow Verification (L/s)
 Flow Column # A1P102
 32mm 150 (.139 - .169)
 36mm 162 (.156 - .190)
 53mm 239 (.228 - .278)
 103mm 498 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 26932

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1024	201507A 7/11/17
0.08	DR2035	201502G 2/24/17
0.20	DR3856	201505A 5/2/17
0.08 DGS	N/A	A6511701 4/27/17

Flow Calibration Performed By [Signature]

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____
 Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 JAN 25 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By [Signature]

Barometric Pressure 1024 Gauge
 ID# 26932 1024 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1076
Interferent	G2840
0.05	SD1024
0.08	DR2035
0.20	DR3856

Attachments

<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Optical Bench Cal'
<input checked="" type="checkbox"/> Pre-Stability Tests	<input type="checkbox"/> Post-Stability Tests
<input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Other _____

Notes: QC-TSK

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

[Signature]
 Quality Control Review

1/25/16
 Date

80-006168
 Stealing checks
 1/25/16

QAM

INTOXILYZER 8000
 Instrument Initialization
 08:10 01/25/2016

MARTIN COUNTY SO
 Intoxilyzer - Alcohol Analyzer SN 80-006168
 Model 8000
 01/25/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:48
Control Test	0.079	08:49
Air Blank	0.000	08:49
Control Test	0.079	08:50
Air Blank	0.000	08:50
Control Test	0.080	08:51
Air Blank	0.000	08:51
Control Test Stats		08:52
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Operator's Signature

MARTIN COUNTY SO
 Intoxilyzer - Alcohol Analyzer SN 80-006168
 Model 8000
 01/25/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:53
Control Test	0.050	08:54
Air Blank	0.000	08:54
Control Test	0.050	08:55
Air Blank	0.000	08:55
Control Test	0.051	08:56
Air Blank	0.000	08:56
Control Test Stats		08:57
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

Operator's Signature

MARTIN COUNTY SO
 Intoxilyzer - Alcohol Analyzer SN 80-006168
 Model 8000
 01/25/2016
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:58
Control Test	0.079	08:59
Air Blank	0.000	08:59
Control Test	0.080	09:00
Air Blank	0.000	09:00
Control Test	0.079	09:00
Air Blank	0.000	09:01
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

BK
 Operator's Signature

QCS

Operator's Signature