

INSTRUMENT PROCESSING SHEET

Agency Astoria PD S/N 80-006222
 Date In 3/10/16 Date Out 3/15/16 Ship P/U H/D CMI EE

Intake Performed By TR

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 Case OK Handle OK
 Dry Gas Holder OK Feet OK
 Keyboard/Plug OK Back/Plugs OK
 Screws tight OK Breath Hose OK

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: * Two feet missing

Quality Checks Performed By RWR

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 231
 Flow Verification (L/s)
 Flow Column # ATP103
 32mm 0.1160 (.139 - .169)
 36mm 0.175 (.156 - .190)
 53mm 0.250 (.228 - .278)
 103mm 0.515 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	AG605301 2/22/18

Flow Calibration Performed By RWR

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____
 Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 MAR 21 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Notes: QC-TSK

Department Inspection Performed By RWR

Barometric Pressure 1014 Gauge
 ID# 28427 1013 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	G4444

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other _____

Notes: _____

Patrick Murphy
 Quality Control Review

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Date 3/21/16

Stability Checks 80-006222 Astatula P.D. 3/15/16 RMB

RMB

RMB
 ASTATULA PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 03/15/2016
 Software: 8100.27
 SN 80-006222

ASTATULA PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 03/15/2016
 Software: 8100.27
 SN 80-006222

ASTATULA PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 03/15/2016
 Software: 8100.27
 SN 80-006222

ASTATULA PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 03/15/2016
 Software: 8100.27
 SN 80-006222

ASTATULA PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 03/15/2016
 Software: 8100.27
 SN 80-006222

Test	9/21/0L	Time
Air Blank	0.000	10:45
Control Test	0.048	10:46
Air Blank	0.000	10:47
Control Test	0.047	10:47
Air Blank	0.000	10:48
Control Test	0.049	10:49
Air Blank	0.000	10:49
Control Test Stats		
Average	0.0480	
Std Dev	0.0010	
Rel Std Dev(%)	2.0833	

Test	9/21/0L	Time
Air Blank	0.000	10:51
Control Test	0.077	10:51
Air Blank	0.000	10:52
Control Test	0.078	10:53
Air Blank	0.000	10:53
Control Test	0.078	10:54
Air Blank	0.000	10:54
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

Test	9/21/0L	Time
Air Blank	0.000	10:55
Control Test	0.199	10:55
Air Blank	0.000	10:55
Control Test	0.200	10:57
Air Blank	0.000	10:58
Control Test	0.200	10:58
Air Blank	0.000	10:59
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

Test	9/21/0L	Time
Air Blank	0.000	10:00
Control Test	0.079	10:00
Air Blank	0.000	10:01
Control Test	0.081	10:01
Air Blank	0.000	10:01
Control Test	0.079	10:02
Air Blank	0.000	10:02
Control Test Stats		
Average	0.0797	
Std Dev	0.0012	
Rel Std Dev(%)	1.4494	

ISK

RMB
 Operator's Signature

RMB
 Operator's Signature

RMB
 Operator's Signature

RMB
 Operator's Signature