

INSTRUMENT PROCESSING SHEET

Agency Astoria PD S/N 80-006235
 Date In 3/29/16 Date Out 4/4/16 Ship P/U H/D CMI EE

Intake Performed By JD

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By JD

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 255
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 0.156 (.139 -.169)
 36mm 0.171 (.156 -.190)
 53mm 0.238 (.228 -.278)
 103mm 0.507 (.447 -.547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201601F 1/26/18
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	AG605301 2/2/18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm
 R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 -.169)
 36mm _____ (.156 -.190)
 53mm _____ (.228 -.278)
 103mm _____ (.447 -.547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 APR 08 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge _____ ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By JD

Barometric Pressure _____ Gauge
 ID# 28427 _____ Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	G4444

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: QC JD 4/7/16

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Whitford
 Quality Control Review

4/8/16
 Date

Stability Checks 80-006235 Astatula P.D.

4/4/16 RMB BK

RMB

ASTATULA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006235
04/04/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	11:58
Control Test	0.049	11:59
Air Blank	0.000	11:59
Control Test	0.049	12:00
Air Blank	0.000	12:00
Control Test	0.049	12:01
Air Blank	0.000	12:02
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

RMB

Operator's Signature

ASTATULA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006235
04/04/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	12:03
Control Test	0.080	12:03
Air Blank	0.000	12:04
Control Test	0.080	12:05
Air Blank	0.000	12:05
Control Test	0.080	12:06
Air Blank	0.000	12:06
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

RMB

Operator's Signature

ASTATULA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006235
04/04/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	12:08
Control Test	0.204	12:09
Air Blank	0.000	12:10
Control Test	0.204	12:10
Air Blank	0.000	12:11
Control Test	0.205	12:11
Air Blank	0.000	12:12
Control Test Stats		
Average	0.2043	
Std Dev	0.0006	
Rel Std Dev(%)	0.2826	

RMB

Operator's Signature

ASTATULA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006235
04/04/2016
Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	12:14
Control Test	0.079	12:14
Air Blank	0.006	12:14
Control Test	0.080	12:15
Air Blank	0.000	12:15
Control Test	0.079	12:16
Air Blank	0.000	12:16
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

RMB

Operator's Signature