

Agency Palm Beach County Sheriff's Office/N 80-006238  
 Date In 2/11/2016 Date Out 2/11/2016  Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>DELL</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>OK</u> Case <u>OK</u> Handle <u>OK</u> Dry Gas Holder <u>OK</u> Feet <u>OK</u> Keyboard/Plug <u>OK</u> Back/Plugs <u>OK</u> Screws tight <u>OK</u> Breath Hose Other Equipment: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Other _____ Notes: _____ _____ _____	<b>Quality Checks</b> Performed By <u>DELL</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>231</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 101</u> 32mm <u>164</u> (.139 - .169) 36mm <u>179</u> (.156 - .190) 53mm <u>253</u> (.228 - .278) 103mm <u>531</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>503967</td> <td>201507A 07/14/2017</td> </tr> <tr> <td>0.08</td> <td>503968</td> <td>201502G 02/24/2017</td> </tr> <tr> <td>0.20</td> <td>503969</td> <td>201505A 05/12/2017</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>A6507503 03/16/2017</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.05	503967	201507A 07/14/2017	0.08	503968	201502G 02/24/2017	0.20	503969	201505A 05/12/2017	0.08 DGS	N/A	A6507503 03/16/2017	<b>Flow Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)
Simulator	Serial #	Lot #/Exp															
0.05	503967	201507A 07/14/2017															
0.08	503968	201502G 02/24/2017															
0.20	503969	201505A 05/12/2017															
0.08 DGS	N/A	A6507503 03/16/2017															
		<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Suggested Service</b> _____ _____															

**RECEIVED**  
 FEB 25 2016  
 FDLE  
 Alcohol Testing Program

<b>Optical Bench Calibration</b> Performed By _____																							
<input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____																							
Simulator	Serial Number	Lot Number	Expiration																				
0.000		N/A	N/A																				
0.040																							
0.100																							
0.200																							
0.400																							
0.080 DGS	N/A																						
<input type="checkbox"/> Post Calibration Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> <th>Lot Number</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.20</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial Number	Lot Number	Expiration	0.05				0.08				0.20				0.08 DGS	N/A		
Simulator	Serial Number	Lot Number	Expiration																				
0.05																							
0.08																							
0.20																							
0.08 DGS	N/A																						

<b>Department Inspection</b> Performed By <u>DELL</u>	
<input checked="" type="checkbox"/> Barometric Pressure ID# <u>28195</u>	<u>1023</u> Gauge <u>1025</u> Instrument
Mouth Alcohol Solution Lot # <u>2015-A</u> Acetone Stock Solution Lot # <u>2615-B</u>	
Simulator	Serial Number
0.00	503965
Interferent	503966
0.05	503967
0.08	503968
0.20	503969
<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Optical Bench Cal <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Other _____	

Notes:  E-MAILED 02/11/2016  **APPROVED**  
OC-TSK  
Patrick Murphy  
 Quality Control Review

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use  
2/25/16  
 Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-006238	Palm Beach County Sheriff's Office	02/11/2016	DeLL

*DeLL*

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>	0.194 to 0.206 <input checked="" type="checkbox"/>	0.077 to 0.083 <input checked="" type="checkbox"/>

PALM BEACH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-06238  
02/11/2016  
Software: 8100.27

Test g/210L Time

Air Blank 0.000 13:42  
Control Test 0.050 13:42  
Air Blank 0.000 13:42  
Control Test 0.052 13:42  
Air Blank 0.000 13:43  
Control Test 0.051 13:43  
Air Blank 0.000 13:43  
Control Test Stats  
Average 0.0513  
Std Dev 0.0019  
Rel Std Dev(%) 1.9608

Operator's Signature *DeLL*

PALM BEACH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-06238  
02/11/2016  
Software: 8100.27

Test g/210L Time

Air Blank 0.000 13:45  
Control Test 0.080 13:45  
Air Blank 0.000 13:46  
Control Test 0.081 13:46  
Air Blank 0.000 13:47  
Control Test 0.081 13:48  
Air Blank 0.000 13:48  
Control Test Stats  
Average 0.0807  
Std Dev 0.0006  
Rel Std Dev(%) 0.7157

Operator's Signature *DeLL*

PALM BEACH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-06238  
02/11/2016  
Software: 8100.27

Test g/210L Time

Air Blank 0.002 13:49  
Control Test 0.203 13:51  
Air Blank 0.000 13:51  
Control Test 0.205 13:51  
Air Blank 0.000 13:52  
Control Test 0.204 13:53  
Air Blank 0.000 13:53  
Control Test Stats  
Average 0.2040  
Std Dev 0.0018  
Rel Std Dev(%) 0.4902

Operator's Signature *DeLL*

PALM BEACH CO SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-06238  
02/11/2016  
Software: 8100.27

Test g/210L Time

Air Blank 0.000 13:56  
Control Test 0.081 13:57  
Air Blank 0.000 13:57  
Control Test 0.082 13:57  
Air Blank 0.000 13:58  
Control Test 0.081 13:58  
Air Blank 0.000 13:59  
Control Test Stats  
Average 0.0813  
Std Dev 0.0006  
Rel Std Dev(%) 0.7095

Operator's Signature *DeLL*

*DK*