



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Gainesville PD

S/N 80-006384

Date In 9/28/16 Date Out 10/5/16

Ship  P/U  H/D  CMI  EE

<b>Intake</b> Performed By <u>[Signature]</u> <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Return from CMI <input type="checkbox"/> Return from Enforcement Electronics <input type="checkbox"/> Other _____ Visual Inspection: <u>ok</u> Case <u>ok</u> Handle <u>ok</u> Dry Gas Holder <u>ok</u> Feet <u>ok</u> Keyboard/Plug <u>ok</u> Back/Plugs <u>ok</u> Screws tight <u>ok</u> Breath Hose Other Equipment: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Other <u>Static Bag</u>		<b>Quality Checks</b> Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>171</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32mm <u>.167</u> (.139 - .169) 36mm <u>.187</u> (.156 - .190) 53mm <u>.246</u> (.228 - .278) 103mm <u>.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks		<b>Flow Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Flow Calibration N/A <input type="checkbox"/> Flow Calibration Complete Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32mm _____ (.139 - .169) 36mm _____ (.156 - .190) 53mm _____ (.228 - .278) 103mm _____ (.447 - .547)																
Notes: _____ _____ _____		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.05</td> <td>SD1018</td> <td>201507A 7/14/17</td> </tr> <tr> <td>0.08</td> <td>SD1011</td> <td>201601F 1/26/18</td> </tr> <tr> <td>0.20</td> <td>SD1025</td> <td>201604C 4/5/18</td> </tr> <tr> <td>0.08 DGS</td> <td>N/A</td> <td>AG612405 5/3/18</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.05	SD1018	201507A 7/14/17	0.08	SD1011	201601F 1/26/18	0.20	SD1025	201604C 4/5/18	0.08 DGS	N/A	AG612405 5/3/18	<b>Maintenance</b> Performed By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <b>Suggested Service</b> _____ _____	
Simulator	Serial #	Lot #/Exp																		
0.05	SD1018	201507A 7/14/17																		
0.08	SD1011	201601F 1/26/18																		
0.20	SD1025	201604C 4/5/18																		
0.08 DGS	N/A	AG612405 5/3/18																		

RECEIVED  
OCT 06 2016  
FDLE  
Alcohol Testing Program

<b>Optical Bench Calibration</b> Performed By _____ <input checked="" type="checkbox"/> Optical Bench Calibration N/A <input type="checkbox"/> Optical Bench Calibration Complete Barometric Pressure Gauge ID # _____			
Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		
<input type="checkbox"/> Post Calibration Stability Checks			
Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

<b>Department Inspection</b> Performed By <u>[Signature]</u> <input checked="" type="checkbox"/> Barometric Pressure <u>1016</u> Gauge ID# <u>28427</u> <u>1011</u> Instrument Mouth Alcohol Solution Lot # <u>2016-A</u> Acetone Stock Solution Lot # <u>2016-B</u>	
Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025
<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Optical Bench Cal <input checked="" type="checkbox"/> Pre-Stability Tests <input type="checkbox"/> Post-Stability Tests <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Other _____	

Notes: QC: 8p  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_


<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
---

[Signature]  
Quality Control Review

10/6/16  
Date

GAINESVILLE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006384  
10/05/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:23
Control Test	0.049	10:24
Air Blank	0.000	10:24
Control Test	0.050	10:25
Air Blank	0.000	10:26
Control Test	0.050	10:26
Air Blank	0.000	10:27
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

  
-----  
Operator's Signature

GAINESVILLE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006384  
10/05/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:16
Control Test	0.079	10:17
Air Blank	0.000	10:18
Control Test	0.080	10:18
Air Blank	0.000	10:19
Control Test	0.080	10:20
Air Blank	0.000	10:20
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

  
-----  
Operator's Signature

GAINESVILLE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006384  
10/05/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:03
Control Test	0.198	10:04
Air Blank	0.000	10:04
Control Test	0.198	10:05
Air Blank	0.000	10:05
Control Test	0.199	10:06
Air Blank	0.000	10:07
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	

  
-----  
Operator's Signature

GAINESVILLE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006384  
10/05/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:31
Control Test	0.079	10:31
Air Blank	0.000	10:31
Control Test	0.079	10:32
Air Blank	0.000	10:32
Control Test	0.079	10:33
Air Blank	0.000	10:33
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP ASK

DG5

  
-----  
Operator's Signature