

INSTRUMENT PROCESSING SHEET

Agency Eschambia County SO S/N 80-00436
 Date In 5/23/16 Date Out 5/26/16 Ship P/U H/D CMI EE

Intake Performed By SD

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: _____

Quality Checks Performed By PW

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 226
 Flow Verification (L/s)
 Flow Column # AP105
 32mm 164 (.139 - .169)
 36mm 183 (.156 - .190)
 53mm 257 (.228 - .278)
 103mm 531 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>G2403</u>	<u>201507A</u> <u>7/14/17</u>
0.08	<u>SD3964</u>	<u>201601F</u> <u>1/26/18</u>
0.20	<u>G4444</u>	<u>201505A</u> <u>5/12/17</u>
0.08 DGS	<u>N/A</u>	<u>ABL00054</u> <u>1/5/18</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 30mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 MAY 31 2016
 Alcohol Testing Program
 FDLE

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		<u>N/A</u>	<u>N/A</u>
0.040			
0.100			
0.200			
0.400			
0.080 DGS	<u>N/A</u>		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	<u>N/A</u>		

Department Inspection Performed By PWS

Barometric Pressure
 ID# 28427 1020 Gauge
1021 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2885</u>
0.05	<u>G2403</u>
0.08	<u>SD3964</u>
0.20	<u>G4444</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration
 Optical Bench Cal
 Post-Stability Tests
 Other Form 40

Notes: DR/AC OK WPM

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Brett Hankland

Quality Control Review

5/31/16
Date

Stability Tests - Escambia C80

80-006436

5/26/14

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006436
05/26/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:29
Control Test	0.049	08:30
Air Blank	0.000	08:31
Control Test	0.049	08:31
Air Blank	0.000	08:32
Control Test	0.050	08:32
Air Blank	0.000	08:33
Control Test Stats		
Average	0.0493	
Std Dev	0.0005	
Rel Std Dev(%)	1.1703	

QMM

PWS

Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006436
05/26/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:34
Control Test	0.079	08:35
Air Blank	0.000	08:35
Control Test	0.079	08:36
Air Blank	0.000	08:37
Control Test	0.079	08:37
Air Blank	0.000	08:38
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

PWS

Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006436
05/26/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:22
Control Test	0.200	08:23
Air Blank	0.000	08:23
Control Test	0.200	08:24
Air Blank	0.000	08:25
Control Test	0.201	08:25
Air Blank	0.000	08:26
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2982	

PWS

Operator's Signature

ESCAMBIA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006436
05/26/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:40
Control Test	0.080	08:40
Air Blank	0.000	08:41
Control Test	0.080	08:41
Air Blank	0.000	08:42
Control Test	0.079	08:42
Air Blank	0.000	08:42
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

PWS

ASK

PWS

Operator's Signature