

INSTRUMENT PROCESSING SHEET

Agency Miami Police Department S/N 80-006450
 Date In 12/8/2016 Date Out 12/9/2016 Ship P/U H/D CMI EE

Intake Performed By DEER

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Anti static Bag

Notes: _____

Quality Checks Performed By DEER

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 248
 Flow Verification (L/s)
 Flow Column # ATP106
 32mm 140 (.139 - .169)
 36mm 167 (.156 - .190)
 53mm 234 (.228 - .278)
 103mm 496 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28663

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>SD3967</u>	<u>201507A</u> <u>07/14/2017</u>
0.08	<u>SD3968</u>	<u>201601F</u> <u>01/26/2018</u>
0.20	<u>SD3969</u>	<u>201505A</u> <u>05/12/2017</u>
0.08 DGS	N/A	<u>A5605301</u> <u>02/22/2018</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 DEC 14 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DEER

Barometric Pressure 1021 Gauge
 ID# 68639 1021 Instrument

Mouth Alcohol Solution Lot # 2016 A
 Acetone Stock Solution Lot # 2016 B

Simulator	Serial Number
0.00	<u>SD 3965</u>
Interferent	<u>SD 3966</u>
0.05	<u>SD 3967</u>
0.08	<u>SD 3968</u>
0.20	<u>SD 3969</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: **E-MAILED** **APPROVED**
12/9/2016
QA/PC OK BDM 12/13/2016

Brett Kunkel

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

12/14/16 Date

Quality Control Review

Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-006456	Miami Police Department	12/9/2016	<i>Bill</i>

PK

0.05g/210L <input checked="" type="checkbox"/>	0.08g/210L <input checked="" type="checkbox"/>	0.20g/210L <input checked="" type="checkbox"/>	DGS 0.08g/210L <input checked="" type="checkbox"/>
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083

MIAMI PD
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-006456
12/09/2016
Software: 8100.27

Test g/210L Time

Air Blank 0.000 08:10
Control Test 0.048 08:11
Air Blank 0.000 08:11
Control Test 0.049 08:12
Air Blank 0.000 08:12
Control Test 0.049 08:13
Air Blank 0.000 08:13
Control Test Stats
Average 0.0487
Std Dev 0.0006
Rel Std Dev(%) 1.1863

Bill
Operator's Signature

MIAMI PD
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-006456
12/09/2016
Software: 8100.27

Test g/210L Time

Air Blank 0.000 08:15
Control Test 0.078 08:15
Air Blank 0.000 08:16
Control Test 0.078 08:17
Air Blank 0.000 08:17
Control Test 0.079 08:18
Air Blank 0.000 08:18
Control Test Stats
Average 0.0783
Std Dev 0.0006
Rel Std Dev(%) 0.7370

Bill
Operator's Signature

MIAMI PD
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-006456
12/09/2016
Software: 8100.27

Test g/210L Time

Air Blank 0.000 08:20
Control Test 0.200 08:21
Air Blank 0.000 08:21
Control Test 0.201 08:22
Air Blank 0.000 08:23
Control Test 0.201 08:23
Air Blank 0.000 08:24
Control Test Stats
Average 0.2007
Std Dev 0.0006
Rel Std Dev(%) 0.2877

Bill
Operator's Signature

MIAMI PD
Intoxilizer - Alcohol Analyzer
Model 8000 SN 80-006456
12/09/2016
Software: 8100.27

Test g/210L Time

Air Blank 0.000 08:25
Control Test 0.080 08:25
Air Blank 0.000 08:26
Control Test 0.079 08:26
Air Blank 0.000 08:27
Control Test 0.080 08:27
Air Blank 0.000 08:27
Control Test Stats
Average 0.0797
Std Dev 0.0006
Rel Std Dev(%) 0.7247

Bill
Operator's Signature

g/g