

INSTRUMENT PROCESSING SHEET

Agency Manatee County SO S/N 80-006559
 Date In 12/22/15 Date Out 1/6/16 Ship P/U H/D CMI EE

RECEIVED
 JAN 07 2016
 FDLE
 Alcohol Testing Program

Intake Performed By TR

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other Static Bag

Notes: Extra panel cord/charger

Quality Checks Performed By PWS

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 199
 Flow Verification (L/s)
 Flow Column # ATP103
 32mm 164 (.139 - .169)
 36mm 183 (.156 - .190)
 53mm 250 (.228 - .278)
 103mm 519 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28462

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7/14/17
0.08	SD1011	201502B 2/24/17
0.20	SD1025	201505A 5/12/17
0.08 DGS	N/A	AG511701 4/27/17

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # Alcohol Testing Program
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

Optical Bench Calibration Performed By PWS

Optical Bench Calibration N/A
 Optical Bench Calibration Complete
 Barometric Pressure Gauge 1021 ID# 28421

Simulator	Serial Number	Lot Number	Expiration
0.000	G248D	N/A	N/A
0.040	SD3962	15108	8/18/17
0.100	SD3964	15001	5/20/17
0.200	SD3933	14104	6/25/16
0.400	G2403	15105	6/10/17
0.080 DGS	N/A	09014080A1	5/1/16

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05	SD1018	201507A	7/14/17
0.08	SD1011	201502B	2/24/17
0.20	SD1025	201505A	5/12/17
0.08 DGS	N/A	AG511701	4/27/17

Department Inspection Performed By PWS

Barometric Pressure 1021 Gauge
 ID# 28462 1020 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2015-B

Simulator	Serial Number
0.00	SD1022
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41 Optical Bench Cal
 Pre-Stability Tests Post-Stability Tests
 Flow Calibration Other Form 40
Reg. for Reg.

Notes: All stability values within range.
Calibrated optical bench to bring values
closer to nominal. (PWS)
QC - BK

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC
 Return to/Place Into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Patrick Murphy
 Quality Control Review

1/7/16
 Date

Stability Tests
Pre-Calibration

- Mandee CSO # 80-006559 1/6/16

BBM

MANATEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006559
01/06/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:15
Control Test	0.051	09:16
Air Blank	0.000	09:17
Control Test	0.051	09:17
Air Blank	0.000	09:18
Control Test	0.051	09:19
Air Blank	0.000	09:19
Control Test Stats		
Average	0.0510	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SMS

Operator's Signature

MANATEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006559
01/06/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:21
Control Test	0.079	09:21
Air Blank	0.000	09:22
Control Test	0.082	09:22
Air Blank	0.000	09:23
Control Test	0.082	09:24
Air Blank	0.000	09:24
Control Test Stats		
Average	0.0810	
Std Dev	0.0017	
Rel Std Dev(%)	2.1383	

SMS

Operator's Signature

MANATEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006559
01/06/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:26
Control Test	0.205	09:27
Air Blank	0.000	09:28
Control Test	0.207	09:28
Air Blank	0.000	09:29
Control Test	0.207	09:30
Air Blank	0.000	09:30
Control Test Stats		
Average	0.2063	
Std Dev	0.0012	
Rel Std Dev(%)	0.5596	

SMS

Operator's Signature

MANATEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006559
01/06/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:35
Control Test	0.081	09:35
Air Blank	0.000	09:35
Control Test	0.080	09:36
Air Blank	0.000	09:36
Control Test	0.080	09:37
Air Blank	0.000	09:37
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

SMS

Operator's Signature

AK

Optical Bench
Calibration
nanutek CSO
#80-006559
1/6/16
EPPM

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum to = 12509, Sum to = 13107
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.7230 (-0.0120)
Sample #2 = 0.7430 (-0.0200)
Sample #3 = 0.6940 (-0.0360)
Sample #4 = 0.7550 (-0.0140)
Avg % Abs = 0.7307 (0.0180)
STD DEV = 0.0323 (0.0183)
REL STD DEV = 4.423 (101.935)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.4870 (-0.0070)
Sample #2 = 1.5150 (-0.0080)
Sample #3 = 1.5010 (-0.0090)
Sample #4 = 1.5370 (-0.0040)
Avg % Abs = 1.5177 (0.0077)
STD DEV = 0.0181 (0.0087)
REL STD DEV = 1.196 (524.214)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum to = 12503, Sum to = 13105
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.7850 (-0.0100)
Sample #2 = 1.8260 (-0.0110)
Sample #3 = 1.8010 (-0.0220)
Sample #4 = 1.7840 (-0.0230)
Avg % Abs = 1.8037 (0.0133)
STD DEV = 0.0211 (0.0193)
REL STD DEV = 1.171 (170.715)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.5290 (-0.0020)
Sample #2 = 3.5860 (-0.0090)
Sample #3 = 3.5690 (0.0060)
Sample #4 = 3.5670 (0.0040)
Avg % Abs = 3.5720 (0.0003)
STD DEV = 0.0070 (0.0081)
REL STD DEV = 0.196 (2443.358)

MANATEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
SN 80-006559
13:56:52

Auto Calibration
Max Power Res Value = 100
Auto Range Res Value = 69

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum to = 12515, Sum to = 13107
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0200 (-0.0130)
Sample #2 = 0.0080 (-0.0040)
Sample #3 = 0.0480 (0.0040)
Sample #4 = 0.0450 (0.0100)
Avg % Abs = 0.0337 (0.0033)
STD DEV = 0.0223 (0.0070)
REL STD DEV = 66.174 (210.713)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1460 (-0.0270)
Sample #2 = 0.1240 (-0.0080)
Sample #3 = 0.1620 (-0.0230)
Sample #4 = 0.1760 (-0.0190)
Avg % Abs = 0.1540 (-0.0167)
STD DEV = 0.0269 (0.0078)
REL STD DEV = 17.472 (46.605)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum to = 12502, Sum to = 13104
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.5150 (-0.0040)
Sample #2 = 3.4710 (0.0320)
Sample #3 = 3.5190 (0.0160)
Sample #4 = 3.4710 (0.0400)
Avg % Abs = 3.4870 (0.0293)
STD DEV = 0.0277 (0.0122)
REL STD DEV = 0.795 (41.660)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.6390 (0.0020)
Sample #2 = 6.6510 (0.0050)
Sample #3 = 6.6910 (0.0010)
Sample #4 = 6.6620 (0.0220)
Avg % Abs = 6.6680 (0.0093)
STD DEV = 0.0207 (0.0112)
REL STD DEV = 0.310 (119.469)

Sol Value = 0.400 g/210L ***
Fit Value = 1.9048 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum to = 12496, Sum to = 13102
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.8010 (-0.0070)
Sample #2 = 6.8070 (0.0110)
Sample #3 = 6.8360 (-0.0100)
Sample #4 = 6.8300 (0.0140)
Avg % Abs = 6.8243 (0.0050)
STD DEV = 0.0153 (0.0131)
REL STD DEV = 0.224 (261.534)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 12.5940 (-0.0010)
Sample #2 = 12.6090 (0.0150)
Sample #3 = 12.6460 (0.0000)
Sample #4 = 12.6100 (0.0210)
Avg % Abs = 12.6217 (0.0120)
STD DEV = 0.0211 (0.0108)
REL STD DEV = 0.167 (90.139)

**** AUTO CAL DATA ****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.034
Std Dev = 0.02 Rel Std Dev = 66.17
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.731
Std Dev = 0.03 Rel Std Dev = 4.42
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.804
Std Dev = 0.02 Rel Std Dev = 1.17
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.487
Std Dev = 0.03 Rel Std Dev = 0.79
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 6.824
Std Dev = 0.02 Rel Std Dev = 0.22
Zero Order Coef = -92.67
First Order Coef = 2689.63
Second Order Coef = 16.96
Standard Deviation = 32.546326

<<<<< CHANNEL 2 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.154
Std Dev = 0.03 Rel Std Dev = 17.47
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.518
Std Dev = 0.02 Rel Std Dev = 1.20
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.572
Std Dev = 0.01 Rel Std Dev = 0.20
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.688
Std Dev = 0.02 Rel Std Dev = 0.31
Sol Val = 1.9048 mg/l or 0.400 g/210L
% Abs = 12.622
Std Dev = 0.02 Rel Std Dev = 0.17
Zero Order Coef = -224.00
First Order Coef = 1372.15
Second Order Coef = 12.31
Standard Deviation = 45.815792

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0000
0.040 0.040 0.0005
0.100 0.101 -0.0011
0.200 0.199 -0.0007
0.400 0.400 -0.0001

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0003
0.040 0.040 0.0004
0.100 0.102 -0.0015
0.200 0.199 0.0011
0.400 0.400 -0.0002

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
**** CHANNEL 1
Sample #1 = 3578.00
Sample #2 = 3689.00
Sample #3 = 3578.00
Sample #4 = 3633.00
Average Result = 3633.3333
STD DEV = 55.5008
REL STD DEV = 1.528

**** CHANNEL 2
Sample #1 = 3367.00
Sample #2 = 3410.00
Sample #3 = 3375.00
Sample #4 = 3377.00
Average Result = 3387.3333
STD DEV = 19.6554
REL STD DEV = 0.580

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1021
3 um H2O Adjust (mg/l*10,000) = 176
9 um H2O Adjust (mg/l*10,000) = 422
**** AUTO CAL PASS

Stability Tests
Post-Calibration

BBM

- Manatee CSO

#80-006559

1/6/16

MANATEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006559
01/06/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:06
Control Test	0.050	15:06
Air Blank	0.000	15:07
Control Test	0.049	15:07
Air Blank	0.000	15:08
Control Test	0.050	15:09
Air Blank	0.000	15:09
Control Test Stats	0.0497	
Average	0.0006	
Std Dev	1.1625	
Rel Std Dev(%)		

BBM

Operator's Signature

MANATEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006559
01/06/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:11
Control Test	0.078	15:12
Air Blank	0.000	15:12
Control Test	0.079	15:13
Air Blank	0.000	15:13
Control Test	0.078	15:14
Air Blank	0.000	15:15
Control Test Stats		
Average	0.0783	
Std Dev	0.0106	
Rel Std Dev(%)	0.7370	

BBM

Operator's Signature

MANATEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006559
01/06/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:16
Control Test	0.200	15:17
Air Blank	0.000	15:17
Control Test	0.200	15:18
Air Blank	0.000	15:19
Control Test	0.200	15:19
Air Blank	0.000	15:20
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

BBM

Operator's Signature

MANATEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006559
01/06/2016
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:21
Control Test	0.078	15:21
Air Blank	0.000	15:22
Control Test	0.079	15:22
Air Blank	0.000	15:23
Control Test	0.078	15:23
Air Blank	0.000	15:24
Control Test Stats		
Average	0.0783	
Std Dev	0.0106	
Rel Std Dev(%)	0.7370	

BBM

Operator's Signature