

INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol S/N 80-006626
 Date In 5/17/2016 Date Out 5/25/2016 Ship P/U H/D CMI EE

Intake Performed By DELL

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
OK Case OK Handle
OK Dry Gas Holder OK Feet
OK Keyboard/Plug OK Back/Plugs
OK Screws tight OK Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By DELL

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 204
 Flow Verification (L/s)

Flow Column #	<u>ATP105</u>	
32mm	<u>160</u>	(.139 - .169)
36mm	<u>179</u>	(.156 - .190)
53mm	<u>246</u>	(.228 - .278)
103mm	<u>511</u>	(.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>G2403</u>	<u>201507A</u> <u>07/14/2017</u>
0.08	<u>503964</u>	<u>201601F</u> <u>01/20/2018</u>
0.20	<u>G4444</u>	<u>201505A</u> <u>05/12/2017</u>
0.08 DGS	<u>N/A</u>	<u>A6600504</u> <u>01/05/2018</u>

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete

Flow Column # _____
 5L/min - 17mm
 15L/min - 43mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)

Flow Column #	
32mm	(.139 - .169)
36mm	(.156 - .190)
53mm	(.228 - .278)
103mm	(.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 MAY 26 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		<u>N/A</u>	<u>N/A</u>
0.040			
0.100			
0.200			
0.400			
0.080 DGS	<u>N/A</u>		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	<u>N/A</u>		

Department Inspection Performed By DELL

Barometric Pressure 1020 Gauge
 ID# 26932 1019 Instrument

Mouth Alcohol Solution Lot # 2016A
 Acetone Stock Solution Lot # 2016B

Simulator	Serial Number
0.00	<u>G2880</u>
Interferent	<u>G2834</u>
0.05	<u>G2403</u>
0.08	<u>503964</u>
0.20	<u>G4444</u>

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: AC. gms

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

5/26/16
Date

Bruce Whiteland
 Quality Control Review

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilites	80-006626	Florida Highway Patrol	05/24/2016	<i>PK</i>

PK

<p>0.05g/210L <input checked="" type="checkbox"/></p> <p>0.047 to 0.053 <input checked="" type="checkbox"/></p> <p>FHP Intoxilizer - Alcohol Analyzer Model 8000 SN 80-006626 05/25/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 12:03 Control Test 0.050 12:03 Air Blank 0.000 12:04 Control Test 0.050 12:04 Air Blank 0.000 12:05 Control Test 0.050 12:05 Air Blank 0.000 12:06 Control Test 0.000 12:06</p> <p>Control Test Stats Average 0.0500 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p><i>PK</i> Operator's Signature</p>	<p>0.08g/210L <input checked="" type="checkbox"/></p> <p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>FHP Intoxilizer - Alcohol Analyzer Model 8000 SN 80-006626 05/25/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 12:08 Control Test 0.079 12:09 Air Blank 0.000 12:10 Control Test 0.080 12:10 Air Blank 0.000 12:11 Control Test 0.080 12:11 Air Blank 0.000 12:12 Control Test 0.000 12:12</p> <p>Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247</p> <p><i>PK</i> Operator's Signature</p>	<p>0.20g/210L <input checked="" type="checkbox"/></p> <p>0.194 to 0.206 <input checked="" type="checkbox"/></p> <p>FHP Intoxilizer - Alcohol Analyzer Model 8000 SN 80-006626 05/25/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 12:14 Control Test 0.200 12:14 Air Blank 0.000 12:15 Control Test 0.200 12:15 Air Blank 0.000 12:16 Control Test 0.201 12:16 Air Blank 0.000 12:17 Control Test 0.000 12:17</p> <p>Control Test Stats Average 0.2003 Std Dev 0.0006 Rel Std Dev(%) 0.2882</p> <p><i>PK</i> Operator's Signature</p>	<p>DGS 0.08g/210L <input checked="" type="checkbox"/></p> <p>0.077 to 0.083 <input checked="" type="checkbox"/></p> <p>FHP Intoxilizer - Alcohol Analyzer Model 8000 SN 80-006626 05/25/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 12:20 Control Test 0.082 12:20 Air Blank 0.000 12:21 Control Test 0.082 12:21 Air Blank 0.000 12:22 Control Test 0.081 12:22 Air Blank 0.000 12:22</p> <p>Control Test Stats Average 0.0817 Std Dev 0.0006 Rel Std Dev(%) 0.7070</p> <p><i>PK</i> Operator's Signature</p>
---	---	---	--

SM