



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FHP S/N 80-006628
Date In 5/17/16 Date Out 5/25/2016 Ship P/U H/D CMI EE

Intake Performed By TP

Registration
 Annual
 Return from CMI
 Return from Enforcement
 Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By DELL

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 203
 Flow Verification (L/s)
 Flow Column # ATP.105
 32mm 164 (.139 - .169)
 36mm 183 (.156 - .190)
 53mm 250 (.228 - .278)
 103mm 523 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G2403	201507A 07/14/2017
0.08	SD3964	201101E 01/26/2018
0.20	G4444	201505A 05/12/2017
0.08 DGS	N/A	AG 600504 01/05/2018

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
MAY 26 2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DELL

Barometric Pressure 1020 Gauge
 ID# 26932 1018 Instrument

Mouth Alcohol Solution Lot # 2016-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2880
Interferent	G2834
0.05	G2403
0.08	SD3964
0.20	G4444

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other _____

Notes: GC-PW

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brett Hunkeler
Quality Control Review

5/26/16
Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-006628	Florida Highway Patrol	05/24/2016	<i>WLR</i>

WLR

0.05g/210L 0.047 to 0.053 <input checked="" type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>
<p>FHP Intoxilizer - Alcohol Analyzer Model 8000 SN 80-006628 05/25/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 12:13 Control Test 0.048 12:14 Air Blank 0.000 12:14 Control Test 0.049 12:15 Air Blank 0.000 12:16 Control Test 0.049 12:16 Air Blank 0.000 12:17 Control Test Stats Average 0.0487 Std Dev 0.0006 Rel Std Dev(%) 1.1853</p>	<p>FHP Intoxilizer - Alcohol Analyzer Model 8000 SN 80-006628 05/25/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 12:19 Control Test 0.077 12:20 Air Blank 0.000 12:21 Control Test 0.079 12:21 Air Blank 0.000 12:22 Control Test 0.079 12:22 Air Blank 0.000 12:23 Control Test Stats Average 0.0783 Std Dev 0.0012 Rel Std Dev(%) 1.4741</p>	<p>FHP Intoxilizer - Alcohol Analyzer Model 8000 SN 80-006628 05/25/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 12:25 Control Test 0.198 12:25 Air Blank 0.000 12:26 Control Test 0.199 12:27 Air Blank 0.000 12:27 Control Test 0.199 12:28 Air Blank 0.000 12:28 Control Test Stats Average 0.1987 Std Dev 0.0006 Rel Std Dev(%) 0.2906</p>	<p>FHP Intoxilizer - Alcohol Analyzer Model 8000 SN 80-006628 05/25/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 12:29 Control Test 0.080 12:30 Air Blank 0.000 12:30 Control Test 0.080 12:31 Air Blank 0.000 12:31 Control Test 0.080 12:31 Air Blank 0.000 12:32 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p>
Operator's Signature <i>WLR</i>	Operator's Signature <i>WLR</i>	Operator's Signature <i>WLR</i>	Operator's Signature <i>WLR</i>

WLR