



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency FHP S/N 80-006630
 Date In 5/17/16 Date Out 5/25/2016 Ship P/U H/D CMI EE

Intake Performed By TR

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case Handle
 Dry Gas Holder Feet
 Keyboard/Plug Back/Plugs
 Screws tight Breath Hose

Other Equipment:
 Power cord
 Printer Cable
 Other _____

Notes: _____

Quality Checks Performed By DELL

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 232
 Flow Verification (L/s)
 Flow Column # ATR105
 32mm 160 (.139 - .169)
 36mm 179 (.156 - .190)
 53mm 263 (.228 - .278)
 103mm 519 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G 2403	201507A 07/14/2017
0.08	SD 3964	201601F 11/26/2018
0.20	G 4444	201505A 05/12/2017
0.08 DGS	N/A	061600504 01/05/2018

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 47mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (Program)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
 MAY 26 2016
 FDLE
 Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By DELL

Barometric Pressure 1020 Gauge
 ID# 26932 1020 Instrument

Mouth Alcohol Solution Lot # 2016 A
 Acetone Stock Solution Lot # 2016 B

Simulator	Serial Number
0.00	G 2880
Interferent	G 2834
0.05	G 2403
0.08	SD 3964
0.20	G 4444

Attachments

<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Optical Bench Cal
<input checked="" type="checkbox"/> Pre-Stability Tests	<input type="checkbox"/> Post-Stability Tests
<input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Other _____

Notes: QA/QC OK (TR)

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use
 Conduct an Agency Inspection Before Evidentiary Use

Brith Hinkland
 Quality Control Review

5/26/16
 Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-006630	Florida Highway Patrol	05/25/2016	<i>MS</i>

YSK

0.05g/210L 0.047 to 0.053 <input checked="" type="checkbox"/>	0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>	0.20g/210L 0.194 to 0.206 <input checked="" type="checkbox"/>	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/>
<p>FHP Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006630 05/25/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:23 Control Test 0.049 08:23 Air Blank 0.000 08:24 Control Test 0.048 08:24 Air Blank 0.000 08:25 Control Test 0.049 08:26 Air Blank 0.000 08:26</p> <p>Control Test Stats Average 0.0487 Std Dev 0.0006 Rel Std Dev(%) 1.1863</p> <p>Operator's Signature <i>MS</i></p>	<p>FHP Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006630 05/25/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:30 Control Test 0.080 08:31 Air Blank 0.000 08:31 Control Test 0.080 08:32 Air Blank 0.000 08:32 Control Test 0.080 08:33 Air Blank 0.000 08:34</p> <p>Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature <i>MS</i></p>	<p>FHP Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006630 05/25/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:37 Control Test 0.202 08:38 Air Blank 0.000 08:38 Control Test 0.201 08:39 Air Blank 0.000 08:40 Control Test 0.202 08:40 Air Blank 0.000 08:41</p> <p>Control Test Stats Average 0.2017 Std Dev 0.0006 Rel Std Dev(%) 0.2863</p> <p>Operator's Signature <i>MS</i></p>	<p>FHP Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006630 05/25/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:42 Control Test 0.080 08:43 Air Blank 0.000 08:43 Control Test 0.079 08:44 Air Blank 0.000 08:44 Control Test 0.080 08:44 Air Blank 0.000 08:45</p> <p>Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247</p> <p>Operator's Signature <i>MS</i></p>

MS