



Alcohol Testing Program

### INSTRUMENT PROCESSING SHEET

Agency Manatee County SO S/N 80-00631  
 Date In 5/24/16 Date Out 5/26/16  Ship  P/U  H/D  CMI  EE

**Intake** Performed By TP

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Static Bag

Notes: AC Power Cord

**Quality Checks** Performed By DWS

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 201  
 Flow Verification (L/s)  
 Flow Column # ATP105  
 32mm 167 (.139 - .169)  
 36mm 187 (.156 - .190)  
 53mm 261 (.228 - .278)  
 103mm 535 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	G2403	201507A 7/14/18
0.08	SD3964	201601F 1/26/18
0.20	G4444	201505A 5/12/17
0.08 DGS	N/A	AG600504 1/5/18

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 17mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (Program)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service**

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete  
 Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By PND

Barometric Pressure 1020 Gauge  
 ID# 28427 1019 Instrument

Mouth Alcohol Solution Lot # 2015-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	G2880
Interferent	G2834
0.05	G2403
0.08	SD3964
0.20	G4444

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration  
 Optical Bench Cal  
 Post-Stability Tests  
 Other \_\_\_\_\_

Notes: QA/OC OK PGM

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC  
 Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use  
 Conduct an Agency Inspection Before Evidentiary Use

Scott Kurland  
 Quality Control Review

5/31/16  
 Date

Stability Tests - Manatee CSO #80-006631 5/26/16

MANATEE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006631  
05/26/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:35
Control Test	0.050	09:36
Air Blank	0.000	09:36
Control Test	0.050	09:37
Air Blank	0.000	09:37
Control Test	0.050	09:38
Air Blank	0.000	09:39
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

POA

*[Signature]*  
Operator's Signature

MANATEE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006631  
05/26/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:22
Control Test	0.078	09:23
Air Blank	0.000	09:23
Control Test	0.079	09:24
Air Blank	0.000	09:25
Control Test	0.079	09:25
Air Blank	0.000	09:26
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

*[Signature]*  
Operator's Signature

MANATEE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006631  
05/26/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:30
Control Test	0.200	09:31
Air Blank	0.000	09:31
Control Test	0.200	09:32
Air Blank	0.000	09:32
Control Test	0.201	09:33
Air Blank	0.000	09:34
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	

*[Signature]*  
Operator's Signature

MANATEE COUNTY SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006631  
05/26/2016  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:44
Control Test	0.080	09:44
Air Blank	0.000	09:45
Control Test	0.080	09:45
Air Blank	0.000	09:46
Control Test	0.080	09:46
Air Blank	0.000	09:46
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*[Signature]*  
Operator's Signature