



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Monroe County Sheriff's Office S/N 80-006693

Date In 10/11/2016 Date Out 10/12/2016  Ship  P/U  H/D  CMI  EE

**Intake** Performed By DEPR

Registration  
 Annual  
 Return from CMI  
 Return from Enforcement  
 Electronics  
 Other \_\_\_\_\_

Visual Inspection:  
OK Case OK Handle  
OK Dry Gas Holder OK Feet  
OK Keyboard/Plug OK Back/Plugs  
OK Screws tight OK Breath Hose

Other Equipment:  
 Power cord  
 Printer Cable  
 Other Anti Static Bag

Notes: \_\_\_\_\_

**Quality Checks** Performed By DEPR

Breath Tube Screen  
 Replace O-Rings  
 Instrument Set Up Verified  
 R-Value 280  
 Flow Verification (L/s)  
 Flow Column # ATA 101  
 32mm 156 (.139 - .169)  
 36mm 171 (.156 - .190)  
 53mm 250 (.228 - .278)  
 103mm 511 (.447 - .547)

Barometric Pressure Check  
 Gauge ID # 28663

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	<u>503967</u>	<u>201507A</u> <u>07/14/2017</u>
0.08	<u>503968</u>	<u>201601F</u> <u>01/26/2018</u>
0.20	<u>503969</u>	<u>201505A</u> <u>05/12/2017</u>
0.08 DGS	N/A	<u>A6600504</u> <u>01/05/2018</u>

**Flow Calibration** Performed By \_\_\_\_\_

Flow Calibration N/A  
 Flow Calibration Complete  
 Flow Column # \_\_\_\_\_  
 5L/min - 30mm  
 15L/min - 53mm  
 30L/min - 103mm

R-Value \_\_\_\_\_  
 Post Calibration Verification (L/s)  
 Flow Column # \_\_\_\_\_  
 32mm \_\_\_\_\_ (.139 - .169)  
 36mm \_\_\_\_\_ (.156 - .190)  
 53mm \_\_\_\_\_ (.228 - .278)  
 103mm \_\_\_\_\_ (.447 - .547)

**Maintenance** Performed By \_\_\_\_\_

Battery Replacement  
 Dry Gas Regulator Replacement  
 Breath Tube Replacement  
 Other \_\_\_\_\_

**Suggested Service** \_\_\_\_\_

RECEIVED  
OCT 18 2016  
FDLE Alcohol Testing Program

**Optical Bench Calibration** Performed By \_\_\_\_\_

Optical Bench Calibration N/A  
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # \_\_\_\_\_

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

**Department Inspection** Performed By DEPR

Barometric Pressure 1016 Gauge  
 ID# 28199 1016 Instrument

Mouth Alcohol Solution Lot # 2016-A  
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	<u>SD 3965</u>
Interferent	<u>SD 3966</u>
0.05	<u>SD 3967</u>
0.08	<u>SD 3968</u>
0.20	<u>SD 3969</u>

**Attachments**

Form 41  
 Pre-Stability Tests  
 Flow Calibration  
 Optical Bench Cal  
 Post-Stability Tests  
 Other REGISTRATION

Notes:  **E-MAILED**  
10/12/2016

PA/OC OK 10/18/2016

Brett Wickland

Instrument Complies with Chapter 11D-8, FAC  
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use  
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

10/18/16

Quality Control Review

Date

TYPE OF TEST	SERIAL NUMBER	AGENCY	DATE	PERFORMED BY
Stabilities	80-006693	Monroe County Sheriff's Office	10/12/2016	<i>WLL</i>

*WLL*

0.05g/210L <input checked="" type="checkbox"/>	0.08g/210L <input checked="" type="checkbox"/>	0.20g/210L <input checked="" type="checkbox"/>	DGS 0.08g/210L <input checked="" type="checkbox"/>
<p>MONROE COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-006693 10/12/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:17 Control Test 0.050 08:18 Air Blank 0.000 08:19 Control Test 0.050 08:19 Air Blank 0.000 08:20 Control Test 0.051 08:20 Air Blank 0.000 08:21 Control Test Status Average 0.0503 Std Dev 0.0005 Rel Std Dev(%) 1.1471</p> <p><i>WLL</i> Operator's Signature</p>	<p>MONROE COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-006693 10/12/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:22 Control Test 0.080 08:23 Air Blank 0.000 08:24 Control Test 0.080 08:24 Air Blank 0.000 08:25 Control Test 0.080 08:25 Air Blank 0.000 08:25 Control Test Status Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p><i>WLL</i> Operator's Signature</p>	<p>MONROE COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-006693 10/12/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:27 Control Test 0.201 08:28 Air Blank 0.000 08:29 Control Test 0.201 08:29 Air Blank 0.000 08:30 Control Test 0.201 08:31 Air Blank 0.000 08:31 Control Test Status Average 0.2010 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p><i>WLL</i> Operator's Signature</p>	<p>MONROE COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-006693 10/12/2016 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 08:33 Control Test 0.081 08:34 Air Blank 0.000 08:34 Control Test 0.080 08:35 Air Blank 0.000 08:35 Control Test 0.081 08:35 Air Blank 0.000 08:35 Control Test Status Average 0.0807 Std Dev 0.0005 Rel Std Dev(%) 0.7157</p> <p><i>WLL</i> Operator's Signature</p>

*WLL*