



Alcohol Testing Program

INSTRUMENT PROCESSING SHEET

Agency Manatee County

S/N 80-006694

Date In 9/28/16 Date Out 10/3/16

Ship P/U H/D CMI EE

Intake Performed By SP

Registration
 Annual
 Return from CMI
 Return from Enforcement Electronics
 Other _____

Visual Inspection:
 Case dk Handle dk
 Dry Gas Holder dk Feet dk
 Keyboard/Plug dk Back/Plugs dk
 Screws tight dk Breath Hose dk

Other Equipment:
 Power cord
 Printer Cable
 Other static Bag

Notes: _____

Quality Checks Performed By SP

Breath Tube Screen
 Replace O-Rings
 Instrument Set Up Verified
 R-Value 237
 Flow Verification (L/s)
 Flow Column # ATP102
 32mm 0.156 (.139 - .169)
 36mm 0.171 (.156 - .190)
 53mm 0.238 (.228 - .278)
 103mm 0.515 (.447 - .547)

Barometric Pressure Check
 Gauge ID # 28427

Stability Checks

Simulator	Serial #	Lot #/Exp
0.05	SD1018	201507A 7-14-17
0.08	SD1011	201601F 1-26-18
0.20	SD1025	201604C 4-5-18
0.08 DGS	N/A	AG1012405 5-3-18

Flow Calibration Performed By _____

Flow Calibration N/A
 Flow Calibration Complete
 Flow Column # _____
 5L/min - 17mm
 15L/min - 53mm
 30L/min - 103mm

R-Value _____
 Post Calibration Verification (L/s)
 Flow Column # _____
 32mm _____ (.139 - .169)
 36mm _____ (.156 - .190)
 53mm _____ (.228 - .278)
 103mm _____ (.447 - .547)

Maintenance Performed By _____

Battery Replacement
 Dry Gas Regulator Replacement
 Breath Tube Replacement
 Other _____

Suggested Service

RECEIVED
09/30/2016
FDLE
Alcohol Testing Program

Optical Bench Calibration Performed By _____

Optical Bench Calibration N/A
 Optical Bench Calibration Complete

Barometric Pressure Gauge ID # _____

Simulator	Serial Number	Lot Number	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.400			
0.080 DGS	N/A		

Post Calibration Stability Checks

Simulator	Serial Number	Lot Number	Expiration
0.05			
0.08			
0.20			
0.08 DGS	N/A		

Department Inspection Performed By SP

Barometric Pressure 1013 Gauge
 ID# 28427 1011 Instrument

Mouth Alcohol Solution Lot # 2015-A
 Acetone Stock Solution Lot # 2016-B

Simulator	Serial Number
0.00	SD1019
Interferent	SD1021
0.05	SD1018
0.08	SD1011
0.20	SD1025

Attachments

Form 41
 Pre-Stability Tests
 Flow Calibration

Optical Bench Cal
 Post-Stability Tests
 Other FORM 40

Notes: _____

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Burt Kirkland

Quality Control Review

10/3/16
Date

STABILITY CHECKS - INSTRUMENT # 80-006694 - MANATEE COUNTY SO ^{SP}
 10/3/16 ^{pub}

DAS

MANATEE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006694
 10/03/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	10:00
Control Test	0.048	10:01
Air Blank	0.000	10:01
Control Test	0.048	10:02
Air Blank	0.000	10:02
Control Test	0.048	10:03
Air Blank	0.000	10:04
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP
 Operator's Signature

MANATEE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006694
 10/03/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	10:11
Control Test	0.078	10:11
Air Blank	0.000	10:12
Control Test	0.079	10:12
Air Blank	0.000	10:13
Control Test	0.079	10:14
Air Blank	0.000	10:14
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

SP
 Operator's Signature

MANATEE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006694
 10/03/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	10:05
Control Test	0.194	10:05
Air Blank	0.000	10:06
Control Test	0.195	10:07
Air Blank	0.000	10:07
Control Test	0.196	10:08
Air Blank	0.000	10:08
Control Test Stats		
Average	0.1950	
Std Dev	0.0010	
Rel Std Dev(%)	0.5128	

SP
 Operator's Signature

MANATEE COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006694
 10/03/2016
 Software: 8100.27

Test	9/210L	Time
Air Blank	0.000	10:15
Control Test	0.079	10:16
Air Blank	0.000	10:16
Control Test	0.079	10:16
Air Blank	0.000	10:17
Control Test	0.079	10:17
Air Blank	0.000	10:18
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

ASK

SP
 Operator's Signature