

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15056 Method GC
 Analyst Name Adams (GCVOL.0), Susan

Your 15C1 proficiency test outcome is Satisfactory

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1086	0.072	-0.5	0.071	-1.0
Initial	B	2044	0.204	-0.3	0.208	1.0
Initial	C	3007	0.055	-1.0	0.056	-0.5
Initial	D	4116	0.140	1.0	0.138	0.0
Initial	E	5025	0.314	0.8	0.313	0.6

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15218 Method GC
 Analyst Name Adams (GCSOP1000.3), Susan

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1019	0.075	1.0	0.075	1.0
Initial	D	4044	0.139	0.5	0.139	0.5

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15075 Method GC
 Analyst Name Allard, Roger

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	C	3062	0.055	-1.0	0.056	-0.5
Initial	E	5050	0.306	-0.8	0.307	-0.6

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15097 Method GC
 Analyst Name Anderson, Carla

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1076	0.076	1.5	0.074	0.5
Initial	D	4006	0.139	0.5	0.138	0.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter 2nd Iter
A	0.067	0.079	0.073	0.002	136	0 0
B	0.196	0.214	0.205	0.003	134	1 1
C	0.051	0.063	0.057	0.002	134	0 0
D	0.132	0.144	0.138	0.002	130	1 2
E	0.295	0.325	0.310	0.005	140	1 3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle: 15C1 Receive Results By: 5/15/2015 Participation Number 15086 Method GC
 Analyst Name Ashfaque, Fauzia

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	D	4055	0.133	-2.5	0.134	-2.0
Initial	E	5013	0.313	0.6	0.314	0.8

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15171 Method GC
Analyst Name Baldwin, Jacquelyn

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	C	3051	0.056	-0.5	0.056	-0.5
Initial	E	5094	0.310	0.0	0.312	0.4

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15090 Method GC
Analyst Name Boland, Diane

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2022	0.203	-0.7	0.204	-0.3
Initial	E	5017	0.309	-0.2	0.305	-1.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter 2nd Iter
A	0.067	0.079	0.073	0.002	136	0 0
B	0.196	0.214	0.205	0.003	134	1 1
C	0.051	0.063	0.057	0.002	134	0 0
D	0.132	0.144	0.138	0.002	130	1 2
E	0.295	0.325	0.310	0.005	140	1 3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15082 Method GC
 Analyst Name Cangiano (GCVOL.0), Carol

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1041	0.071	-1.0	0.073	0.0
Initial	B	2057	0.209	1.3	0.209	1.3
Initial	C	3020	0.059	1.0	0.058	0.5
Initial	D	4008	0.138	0.0	0.140	1.0
Initial	E	5063	0.314	0.8	0.314	0.8

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15221 Method GC
 Analyst Name Cangiano (GCSOP1000.3), Carol

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	D	4058	0.139	0.5	0.138	0.0
Initial	E	5002	0.324	2.8	0.318	1.6

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter 2nd Iter
A	0.067	0.079	0.073	0.002	136	0 0
B	0.196	0.214	0.205	0.003	134	1 1
C	0.051	0.063	0.057	0.002	134	0 0
D	0.132	0.144	0.138	0.002	130	1 2
E	0.295	0.325	0.310	0.005	140	1 3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15080 Method GC
 Analyst Name Carey (GCVOL.0), Regina

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1069	0.072	-0.5	0.072	-0.5
Initial	B	2018	0.213	2.7	0.207	0.7
Initial	C	3036	0.056	-0.5	0.056	-0.5
Initial	D	4024	0.139	0.5	0.139	0.5
Initial	E	5052	0.311	0.2	0.315	1.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15215 Method GC
 Analyst Name Carey (GCSOP1000.3), Regina

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1067	0.075	1.0	0.075	1.0
Initial	B	2008	0.206	0.3	0.206	0.3

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15092 Method GC
Analyst Name Carswell (GCVOL.0), Katheryn

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1054	0.072	-0.5	0.072	-0.5
Initial	B	2001	0.205	0.0	0.206	0.3
Initial	C	3078	0.054	-1.5	0.054	-1.5
Initial	D	4099	0.135	-1.5	0.137	-0.5
Initial	E	5012	0.317	1.4	0.313	0.6

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15217 Method GC
 Analyst Name Carswell (GC SOP 1000.3), Katherine

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	D	4026	0.140	1.0	0.139	0.5
Initial	E	5008	0.308	-0.4	0.306	-0.8

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range	
						1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15117 Method GC
Analyst Name Catlin, Tara

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	C	3057	0.059	1.0	0.059	1.0
Initial	D	4082	0.139	0.5	0.140	1.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15014 Method GC
Analyst Name Chin, Jenna

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2051	0.207	0.7	0.207	0.7
Initial	C	3005	0.057	0.0	0.057	0.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range	
						1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15128 Method GC
 Analyst Name Connelly (AA-4), Gina

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1050	0.072	-0.5	0.072	-0.5
Initial	B	2003	0.204	-0.3	0.204	-0.3

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter 2nd Iter
A	0.067	0.079	0.073	0.002	136	0 0
B	0.196	0.214	0.205	0.003	134	1 1
C	0.051	0.063	0.057	0.002	134	0 0
D	0.132	0.144	0.138	0.002	130	1 2
E	0.295	0.325	0.310	0.005	140	1 3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 12129 Method GC
 Analyst Name Escobar-Jacobs, Monica

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2014	0.196	-3.0	0.202	-1.0
Initial	D	4076	0.132	-3.0	0.136	-1.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes:

Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15041 Method GC
 Analyst Name Fernandez, Denise

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	C	3006	0.057	0.0	0.058	0.5
Initial	D	4081	0.139	0.5	0.139	0.5

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15141 Method GC
Analyst Name Gagen, Sara

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1025	0.071	-1.0	0.070	-1.5
Initial	B	2007	0.204	-0.3	0.204	-0.3

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter 2nd Iter
A	0.067	0.079	0.073	0.002	136	0 0
B	0.196	0.214	0.205	0.003	134	1 1
C	0.051	0.063	0.057	0.002	134	0 0
D	0.132	0.144	0.138	0.002	130	1 2
E	0.295	0.325	0.310	0.005	140	1 3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15194 Method GC
Analyst Name Gonyea, Jennifer

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2056	0.205	0.0	0.205	0.0
Initial	D	4098	0.137	-0.5	0.138	0.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15034 Method GC
 Analyst Name Gottal, Josephine

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1004	0.074	0.5	0.074	0.5
Initial	E	5068	0.309	-0.2	0.310	0.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range	
						1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15032 Method GC
 Analyst Name Gray (GCVOL.0), Marda

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1018	0.073	0.0	0.074	0.5
Initial	B	2063	0.207	0.7	0.209	1.3
Initial	C	3046	0.056	-0.5	0.058	0.5
Initial	D	4005	0.139	0.5	0.141	1.5
Initial	E	5084	0.322	2.4	0.325	3.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15220 Method GC
 Analyst Name Gray (GCSOP1000.3), Marda

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2094	0.207	0.7	0.208	1.0
Initial	D	4063	0.141	1.5	0.141	1.5

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15022 Method GC
 Analyst Name Hays, Jeffrey

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	D	4079	0.140	1.0	0.137	-0.5
Initial	E	5018	0.311	0.2	0.311	0.2

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15Q1 Receive Results By: 6/12/2015 Participation Number 15039 Method GC
 Analyst Name Hime, George

Your 15Q1 proficiency test outcome is Unsatisfactory - This report supercedes all other reports for this proficiency cycle

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Re-Test	A	1080	0.066	-3.5	0.066	-3.5
Re-Test	B	2107	0.195	-3.3	0.195	-3.3
Re-Test	C	3114	0.048	-4.5	0.049	-4.0
Re-Test	D	4021	0.129	-4.5	0.129	-4.5
Re-Test	E	5106	0.297	-2.6	0.298	-2.4

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15039 Method GC
 Analyst Name Hime, George

Your 15C1 proficiency test outcome is **Unsatisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1022	0.067	-3.0	0.068	-2.5
Initial	E	5057	0.294	-3.2	0.293	-3.4

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range	
						1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15024 Method GC
 Analyst Name Hippolyte, Theresa

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2052	0.204	-0.3	0.204	-0.3
Initial	E	5019	0.308	-0.4	0.307	-0.6

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15191 Method GC
 Analyst Name Huene (HS110), Lauren

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	C	3065	0.056	-0.5	0.057	0.0
Initial	E	5100	0.312	0.4	0.311	0.2

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15192 Method GC
 Analyst Name Huene (HS70), Lauren

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1045	0.072	-0.5	0.073	0.0
Initial	D	4038	0.138	0.0	0.139	0.5

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15208 Method GC
 Analyst Name Huene (HS108), Lauren

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1095	0.073	0.0	0.072	-0.5
Initial	B	2028	0.203	-0.7	0.203	-0.7
Initial	C	3035	0.056	-0.5	0.057	0.0
Initial	D	4041	0.135	-1.5	0.137	-0.5
Initial	E	5056	0.307	-0.6	0.309	-0.2

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter 2nd Iter
A	0.067	0.079	0.073	0.002	136	0 0
B	0.196	0.214	0.205	0.003	134	1 1
C	0.051	0.063	0.057	0.002	134	0 0
D	0.132	0.144	0.138	0.002	130	1 2
E	0.295	0.325	0.310	0.005	140	1 3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15202 Method GC
 Analyst Name Humphrey, Susan

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1023	0.075	1.0	0.074	0.5
Initial	B	2046	0.204	-0.3	0.205	0.0
Initial	C	3084	0.058	0.5	0.057	0.0
Initial	D	4042	0.138	0.0	0.138	0.0
Initial	E	5098	0.305	-1.0	0.305	-1.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15200 Method GC
 Analyst Name Jardel, Danielle

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2097	0.202	-1.0	0.207	0.7
Initial	E	5049	0.306	-0.8	0.314	0.8
Initial			0.000	0.0	0.000	0.0
Initial			0.000	0.0	0.000	0.0
Initial			0.000	0.0	0.000	0.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	-0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15186 Method GC
 Analyst Name Jean (HS110), Kenson

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2015	0.207	0.7	0.207	0.7
Initial	E	5090	0.312	0.4	0.313	0.6

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter 2nd Iter
A	0.067	0.079	0.073	0.002	136	0 0
B	0.196	0.214	0.205	0.003	134	1 1
C	0.051	0.063	0.057	0.002	134	0 0
D	0.132	0.144	0.138	0.002	130	1 2
E	0.295	0.325	0.310	0.005	140	1 3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15188 Method GC
Analyst Name Jean (HS70), Kenson

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1027	0.075	1.0	0.074	0.5
Initial	C	3042	0.059	1.0	0.059	1.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter 2nd Iter
A	0.067	0.079	0.073	0.002	136	0 0
B	0.196	0.214	0.205	0.003	134	1 1
C	0.051	0.063	0.057	0.002	134	0 0
D	0.132	0.144	0.138	0.002	130	1 2
E	0.295	0.325	0.310	0.005	140	1 3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15209 Method GC
 Analyst Name Jean (HS108), Kenson

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1049	0.075	1.0	0.074	0.5
Initial	B	2075	0.206	0.3	0.207	0.7
Initial	C	3017	0.062	2.5	0.058	0.5
Initial	D	4009	0.137	-0.5	0.138	0.0
Initial	E	5036	0.306	-0.8	0.311	0.2

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15176 Method GC
 Analyst Name Jean, Wilsa

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	C	3037	0.054	-1.5	0.054	-1.5
Initial	E	5071	0.302	-1.6	0.302	-1.6

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15045 Method GC
 Analyst Name Jones (GCVOL.0), Carolyn

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1026	0.071	-1.0	0.070	-1.5
Initial	B	2047	0.201	-1.3	0.202	-1.0
Initial	C	3001	0.053	-2.0	0.054	-1.5
Initial	D	4065	0.136	-1.0	0.137	-0.5
Initial	E	5014	0.314	0.8	0.311	0.2

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15219 Method GC
 Analyst Name Jones (GCSOP1000.3), Carolyn

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2039	0.208	1.0	0.206	0.3
Initial	D	4083	0.140	1.0	0.139	0.5

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15149 Method GC
 Analyst Name Kahl, Joseph

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	D	4007	0.143	2.5	0.140	1.0
Initial	E	5072	0.305	-1.0	0.303	-1.4

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter 2nd Iter
A	0.067	0.079	0.073	0.002	136	0 0
B	0.196	0.214	0.205	0.003	134	1 1
C	0.051	0.063	0.057	0.002	134	0 0
D	0.132	0.144	0.138	0.002	130	1 2
E	0.295	0.325	0.310	0.005	140	1 3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15197 Method GC
 Analyst Name Kamandulis (AA-4r1), Linet

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2045	0.201	-1.3	0.202	-1.0
Initial	C	3069	0.054	-1.5	0.054	-1.5

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range	
						1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15198 Method GC
 Analyst Name Kamandulis (AA-5r1), Linet

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1043	0.072	-0.5	0.071	-1.0
Initial	E	5088	0.310	0.0	0.311	0.2

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range	
						1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15037 Method GC
 Analyst Name Kessler (GCVOL.0), Karen

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1005	0.070	-1.5	0.071	-1.0
Initial	B	2053	0.206	0.3	0.206	0.3
Initial	C	3016	0.054	-1.5	0.055	-1.0
Initial	D	4077	0.140	1.0	0.136	-1.0
Initial	E	5034	0.314	0.8	0.320	2.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15222 Method GC
Analyst Name Kessler (GCSOP1000.3), Karen

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1079	0.075	1.0	0.075	1.0
Initial	B	2093	0.205	0.0	0.205	0.3

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15152 Method GC
Analyst Name King (HS110), Joanna

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2055	0.208	1.0	0.208	1.0
Initial	C	3009	0.058	0.5	0.058	0.5

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15153 Method GC
 Analyst Name King (HS70), Joanna

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1075	0.076	1.5	0.076	1.5
Initial	D	4022	0.140	1.0	0.140	1.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15210 Method GC
 Analyst Name King (HS108), Joanna

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1014	0.077	2.0	0.076	1.5
Initial	B	2069	0.206	0.3	0.210	1.7
Initial	C	3090	0.060	1.5	0.061	2.0
Initial	D	4056	0.139	0.5	0.140	1.0
Initial	E	5048	0.308	-0.4	0.309	-0.2

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15157 Method GC
 Analyst Name Kirkland, Dominique

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	C	3052	0.055	-1.0	0.056	-0.5
Initial	D	4069	0.137	-0.5	0.138	0.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15181 Method GC
Analyst Name Kunsmann, Gary

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	C	3083	0.060	1.5	0.058	0.5
Initial	E	5016	0.304	-1.2	0.305	-1.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15195 Method GC
 Analyst Name Lewis (AA-4r1), Jennifer

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	C	3028	0.055	-1.0	0.056	-0.5
Initial	E	5091	0.306	-0.8	0.307	-0.6

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15196 Method GC
 Analyst Name Lewis (AA-5r1), Jennifer

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1062	0.073	0.0	0.073	0.0
Initial	C	3054	0.056	-0.5	0.056	-0.5

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15125 Method GC
Analyst Name Maragh, Sharin

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1074	0.072	-0.5	0.072	-0.5
Initial	B	2042	0.200	-1.7	0.202	-1.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15033 Method GC
 Analyst Name Mark (GCVOL.0), Amy

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1082	0.071	-1.0	0.071	-1.0
Initial	B	2065	0.209	1.3	0.208	1.0
Initial	C	3030	0.056	-0.5	0.057	0.0
Initial	D	4004	0.139	0.5	0.137	-0.5
Initial	E	5011	0.313	0.6	0.316	1.2

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15216 Method GC
 Analyst Name Mark (GCSOP1000.3), Amy

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.

Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	C	3048	0.058	0.5	0.058	0.5
Initial	E	5081	0.310	0.0	0.310	0.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15064 Method GC
Analyst Name Mazek (AA-4), Kathleen

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2064	0.202	-1.0	0.203	-0.7
Initial	D	4091	0.134	-2.0	0.134	-2.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15188 Method GC
 Analyst Name McKee, Stacy

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	C	3002	0.056	-0.5	0.055	-1.0
Initial	D	4048	0.133	-2.5	0.133	-2.5

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15201 Method GC
 Analyst Name Miller, Russell

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1083	0.074	0.5	0.074	0.5
Initial	D	4017	0.139	0.5	0.139	0.5
Initial			0.000	0.0	0.000	0.0
Initial			0.000	0.0	0.000	0.0
Initial			0.000	0.0	0.000	0.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15146 Method GC
 Analyst Name Montgomery (HS70), Lisa

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1012	0.075	1.0	0.077	2.0
Initial	E	5055	0.309	-0.2	0.309	-0.2

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15147 Method GC
 Analyst Name Montgomery (HS110), Lisa

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.

Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	C	3072	0.057	0.0	0.057	0.0
Initial	D	4088	0.139	0.5	0.139	0.5

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15211 Method GC
 Analyst Name Montgomery (HS108), Lisa

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1036	0.075	1.0	0.074	0.5
Initial	B	2079	0.205	0.0	0.204	-0.3
Initial	C	3060	0.058	0.5	0.058	0.5
Initial	D	4100	0.138	0.0	0.139	0.5
Initial	E	5024	0.311	0.2	0.310	0.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15144 Method GC
Analyst Name Mulcahy, Casey

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1030	0.071	-1.0	0.068	-2.5
Initial	B	2116	0.200	-1.7	0.201	-1.3
Initial	C	3049	0.056	-0.5	0.055	-1.0
Initial	D	4043	0.136	-1.0	0.144	3.0
Initial	E	5074	0.306	-0.8	0.306	-0.8

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15099 Method GC
Analyst Name Nash-Montgomery, Nadra

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1087	0.071	-1.0	0.071	-1.0
Initial	E	5062	0.296	-2.8	0.301	-1.8

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15175 Method GC
 Analyst Name Neufeld (GCVOL.0), Alison

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score	Number Out of Range 1st Iter 2nd Iter
Initial	A	1029	0.073	0.0	0.072	-0.5	0 0
Initial	B	2062	0.209	1.3	0.207	0.7	1 1
Initial	C	3047	0.056	-0.5	0.057	0.0	0 0
Initial	D	4085	0.142	2.0	0.140	1.0	2 2
Initial	E	5003	0.314	0.8	0.315	1.0	1 3

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter 2nd Iter
A	0.067	0.079	0.073	0.002	136	0 0
B	0.196	0.214	0.205	0.003	134	1 1
C	0.051	0.063	0.057	0.002	134	0 0
D	0.132	0.144	0.138	0.002	130	1 2
E	0.295	0.325	0.310	0.005	140	1 3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15225 Method GC
 Analyst Name Neufeld (GCSOP1000.3), Alison

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2087	0.205	0.0	0.205	0.0
Initial	D	4012	0.139	0.5	0.139	0.5

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range	
						1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15140 Method GC
 Analyst Name Nguyen, Linda

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1031	0.070	-1.5	0.070	-1.5
Initial	C	3019	0.053	-2.0	0.054	-1.5

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15142 Method GC
Analyst Name Nunez, William

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1017	0.070	-1.5	0.069	-2.0
Initial	E	5035	0.308	-0.4	0.306	-0.8

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15083 Method GC
Analyst Name Parsons, Bob

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1038	0.070	-1.5	0.070	-1.5
Initial	D	4051	0.132	-3.0	0.134	-2.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15029 Method GC
Analyst Name Patel (GCVOL.0), Minakshi

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1100	0.071	-1.0	0.073	0.0
Initial	B	2085	0.207	0.7	0.206	0.3
Initial	C	3053	0.055	-1.0	0.055	-1.0
Initial	D	4064	0.140	1.0	0.138	0.0
Initial	E	5020	0.310	0.0	0.317	1.4

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15224 Method GC
Analyst Name Patel(GCSOP1000.3), Minakshi

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	C	3077	0.059	1.0	0.058	0.5
Initial	E	5086	0.316	1.2	0.310	0.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 12130 Method GC
Analyst Name Ramsey, Shibaughn

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2068	0.202	-1.0	0.203	-0.7
Initial	E	5004	0.309	-0.2	0.310	0.0

Proficiency Cycle Statistics

Level	Lower Limit		Upper Limit		Mean (Average)	Std Dev	N	Number Out of Range	
	Limit		Limit					1st Iter	2nd Iter
A	0.067		0.079		0.073	0.002	136	0	0
B	0.196		0.214		0.205	0.003	134	1	1
C	0.051		0.063		0.057	0.002	134	0	0
D	0.132		0.144		0.138	0.002	130	1	2
E	0.295		0.325		0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15126 Method GC
 Analyst Name Reidy, Lisa

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1003	0.069	-2.0	0.070	-1.5
Initial	B	2088	0.199	-2.0	0.203	-0.7

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15203 Method GC
 Analyst Name Sanders (AA-5r2), Leeona

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score	Number Out of Range 1st Iter 2nd Iter
Initial	A	1068	0.073	0.0	0.073	0.0	0 0
Initial	B	2092	0.206	0.3	0.205	0.0	1 1
Initial	C	3024	0.057	0.0	0.057	0.0	0 0
Initial	D	4080	0.137	-0.5	0.137	-0.5	1 2
Initial	E	5001	0.311	0.2	0.307	-0.6	1 3

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N
A	0.067	0.079	0.073	0.002	136
B	0.196	0.214	0.205	0.003	134
C	0.051	0.063	0.057	0.002	134
D	0.132	0.144	0.138	0.002	130
E	0.295	0.325	0.310	0.005	140

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15204 Method GC
Analyst Name Sanders (AA-4r2), Leeona

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.

Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1042	0.071	-1.0	0.072	-0.5
Initial	B	2070	0.203	-0.7	0.203	-0.7
Initial	C	3098	0.055	-1.0	0.056	-0.5
Initial	D	4053	0.136	-1.0	0.136	-1.0
Initial	E	5039	0.313	0.6	0.308	-0.4

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 12137 Method GC
 Analyst Name Seagle (HS70), Carol

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1099	0.075	1.0	0.077	2.0
Initial	E	5007	0.310	0.0	0.311	0.2

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15124 Method GC
 Analyst Name Seagle (HS110), Carol

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2091	0.208	1.0	0.208	1.0
Initial	C	3003	0.058	0.5	0.058	0.5

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15213 Method GC
 Analyst Name Seagle (HS108), Carol

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1053	0.074	0.5	0.074	0.5
Initial	B	2005	0.206	0.3	0.206	0.3
Initial	C	3089	0.058	0.5	0.059	1.0
Initial	D	4027	0.137	-0.5	0.137	-0.5
Initial	E	5051	0.308	-0.4	0.309	-0.2

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15121 Method GC
 Analyst Name Shan (GC-FID/MS), Xiaogin

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1092	0.073	0.0	0.073	0.0
Initial	B	2066	0.203	-0.7	0.203	-0.7

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15173 Method GC
Analyst Name Shaw (AA-4), Kristie

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2011	0.202	-1.0	0.203	-0.7
Initial	C	3096	0.055	-1.0	0.055	-1.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15174 Method GC
 Analyst Name Shaw (AA-5), Kristie

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	D	4036	0.136	-1.0	0.137	-0.5
Initial	E	5099	0.306	-0.8	0.307	-0.6

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15193 Method GC
Analyst Name Shoff, Elisa

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.

Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	C	3073	0.053	-2.0	0.054	-1.5
Initial	E	5095	0.303	-1.4	0.304	-1.2

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15122 Method GC
 Analyst Name Tiscione (GC-FID/MS), Nicholas

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1020	0.071	-1.0	0.071	-1.0
Initial	B	2089	0.202	-1.0	0.202	-1.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15155 Method GC
 Analyst Name Titko, Justine

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1034	0.073	0.0	0.068	-2.5
Initial	E	5069	0.311	0.2	0.301	-1.8

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15060 Method GC
 Analyst Name Toth, Mary Lou

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	C	3081	0.060	1.5	0.059	1.0
Initial	D	4016	0.139	0.5	0.139	0.5

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter 2nd Iter
A	0.067	0.079	0.073	0.002	136	0 0
B	0.196	0.214	0.205	0.003	134	1 1
C	0.051	0.063	0.057	0.002	134	0 0
D	0.132	0.144	0.138	0.002	130	1 2
E	0.295	0.325	0.310	0.005	140	1 3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15085 Method GC
 Analyst Name Troeschler (GCVOL.0), Lisa

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1090	0.072	-0.5	0.073	0.0
Initial	B	2016	0.206	0.3	0.209	1.3
Initial	C	3074	0.055	-1.0	0.056	-0.5
Initial	D	4028	0.140	1.0	0.139	0.5
Initial	E	5053	0.311	0.2	0.315	1.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15223 Method GC
Analyst Name Troeschler (GCSOP1000.3), Lisa

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2099	0.207	0.7	0.206	0.3
Initial	C	3056	0.059	1.0	0.059	1.0

Proficiency Cycle Statistics

Level	Lower Limit		Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range	
							1st Iter	2nd Iter
A	0.067		0.079	0.073	0.002	136	0	0
B	0.196		0.214	0.205	0.003	134	1	1
C	0.051		0.063	0.057	0.002	134	0	0
D	0.132		0.144	0.138	0.002	130	1	2
E	0.295		0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15067 Method GC
Analyst Name Wagner, Michael

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2086	0.210	1.7	0.210	1.7
Initial	E	5022	0.316	1.2	0.313	0.6

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15051 Method GC
Analyst Name Warner (AA-4), Ryan

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1056	0.072	-0.5	0.072	-0.5
Initial	E	5010	0.310	0.0	0.310	0.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range	
						1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15139 Method GC
Analyst Name Warner (AA-5), Ryan

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	C	3038	0.057	0.0	0.056	-0.5
Initial	E	5076	0.310	0.0	0.308	-0.4

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15187 Method GC
 Analyst Name Yang (HS110), Michelle

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	B	2082	0.207	0.7	0.207	0.7
Initial	C	3008	0.058	0.5	0.058	0.5

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15190 Method GC
 Analyst Name Yang (HS70), Michelle

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1066	0.074	0.5	0.074	0.5
Initial	D	4018	0.139	0.5	0.138	0.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter 2nd Iter
A	0.067	0.079	0.073	0.002	136	0 0
B	0.196	0.214	0.205	0.003	134	1 1
C	0.051	0.063	0.057	0.002	134	0 0
D	0.132	0.144	0.138	0.002	130	1 2
E	0.295	0.325	0.310	0.005	140	1 3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15214 Method GC
 Analyst Name Yang (HS108), Michelle

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1073	0.075	1.0	0.075	1.0
Initial	B	2009	0.206	0.3	0.206	0.3
Initial	C	3066	0.058	0.5	0.059	1.0
Initial	D	4031	0.139	0.5	0.139	0.5
Initial	E	5044	0.310	0.0	0.307	-0.6

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15123 Method GC
 Analyst Name Yeatman (GC-FID/MS), Dustin

Your 15C1 proficiency test outcome is Satisfactory

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1057	0.071	-1.0	0.072	-0.5
Initial	B	2100	0.201	-1.3	0.202	-1.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15040 Method GC
Analyst Name Zaney, Mary Elizabeth

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	C	3044	0.056	-0.5	0.057	0.0
Initial	D	4095	0.139	0.5	0.141	1.5

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15205 Method GC
Analyst Name Zumaeta (AA-5r2), Edward

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1060	0.073	0.0	0.072	-0.5
Initial	B	2038	0.204	-0.3	0.204	-0.3
Initial	C	3094	0.055	-1.0	0.056	-0.5
Initial	D	4010	0.136	-1.0	0.137	-0.5
Initial	E	5054	0.307	-0.6	0.310	0.0

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.

Florida Department of Law Enforcement

Alcohol Testing Program

BLOOD ALCOHOL PROFICIENCY INITIAL TEST OUTCOME REPORT

Proficiency Cycle 15C1 Receive Results By: 5/15/2015 Participation Number 15206 Method GC
 Analyst Name Zumaeta (AA-4r2), Edward

Your 15C1 proficiency test outcome is **Satisfactory**

Evaluation Procedure

First iteration: For each level, the qualified results reported by participating analysts and reference labs are pooled, the mean, standard deviation and acceptable range are calculated using truncation, and results outside the acceptable range are excluded from the second iteration.
Second iteration: The mean, standard deviation and acceptable range are recalculated using truncation, every analyst's qualified results are compared to the recalculated acceptable range, and every analyst is assigned an outcome for the proficiency cycle. All analysts with Results Outside Acceptable Range receive retest samples. The retest results are compared to the acceptable range established by the initial test and the retested analyst's outcome for the proficiency cycle is assigned based on the second comparison.

Type Test	Level ID	Vial Number	Results #1 Reported (g/dL)	#1 Z Score	Results #2 Reported (g/dL)	#2 Z Score
Initial	A	1006	0.071	-1.0	0.072	-0.5
Initial	B	2058	0.203	-0.7	0.205	0.0
Initial	C	3027	0.055	-1.0	0.055	-1.0
Initial	D	4092	0.136	-1.0	0.136	-1.0
Initial	E	5041	0.308	-0.4	0.309	-0.2

Proficiency Cycle Statistics

Level	Lower Limit	Upper Limit	Mean (Average)	Std Dev	N	Number Out of Range 1st Iter	2nd Iter
A	0.067	0.079	0.073	0.002	136	0	0
B	0.196	0.214	0.205	0.003	134	1	1
C	0.051	0.063	0.057	0.002	134	0	0
D	0.132	0.144	0.138	0.002	130	1	2
E	0.295	0.325	0.310	0.005	140	1	3

Notes: Z score is your reported analytical result compared to the mean in terms of the standard deviation, calculated as (Result-Mean)/StdDev.
 The parameter n is the number of results reported which are used to calculate the second iteration mean and standard deviation.