

Florida Department of Law Enforcement Alcohol Testing Program

CERTIFICATE OF ASSURANCE

The Florida Department of Law Enforcement Alcohol Testing Program approves alcohol reference solution lots used to test the accuracy and precision of evidentiary breath test instruments. Alcohol reference solution lots are prepared by the Department or by a source approved by the Department, and their analysis is important to establish the reliability of alcohol breath test results.

Lot Number: 201302B
Simulated Breath Alcohol Concentration (g/210L): 0.05
Date of Approval: April 23, 2013

Source: Alcohol Countermeasure Systems, Inc.
Date of Manufacture: February 5, 2013
For Use Through: February 5 2015
Number of Bottles Manufactured: 1691

This certifies that ten sample bottles from the above lot were subjected to duplicate gas chromatographic analyses of their alcohol concentrations, that all the results were within the acceptable range, and that this lot of alcohol reference solution is in compliance with the requirements of Chapter 11D-8, F.A.C. and is approved for distribution and use in the State of Florida.

Patrick J. Murphy
Signature

Quality Assurance Manager
Title

PATRICK J MURPHY
Name (Printed or Typed)

4/23/2013
Date