



Florida Department of
Law Enforcement

Richard L. Swearingen
Commissioner

Criminal Justice Professionalism
Post Office Box 1489
Tallahassee, Florida 32302-1489
(850) 410-8600
www.fdle.state.fl.us

Rick Scott, *Governor*
Pam Bondi, *Attorney General*
Jeff Atwater, *Chief Financial Officer*
Adam Putnam, *Commissioner of Agriculture*

March 6, 2015

Mr. Christopher Wilson
Alcohol Countermeasure Systems, Inc.
60 International Boulevard
Toronto, Ontario M9W 612

Subject: Florida Acceptance of Alcohol Reference Solution Lot 201502G

Dear Mr. Wilson:

The Florida Department of Law Enforcement Alcohol Testing Program received bottles of alcohol reference solution, Lot 201502G, as samples for approval on February 27, 2015. Gas chromatographic analyses of ten (10) bottles from this lot was completed on March 5, 2015. This lot meets the requirements of Chapter 11D-8, Florida Administrative Code.

Lot 201502G is approved for use in Florida effective March 6, 2015.

If you have any questions regarding the analyses, please feel free to contact me at (850) 617-1280.

Sincerely,

Patrick J. Murphy
Program Manager
Alcohol Testing Program



Florida Department of
Law Enforcement

Gerald M. Bailey
Commissioner

Criminal Justice Professionalism
Post Office Box 1489
Tallahassee, Florida 32302-1489
(850) 410-8600
www.fdle.state.fl.us

Rick Scott, *Governor*
Pam Bondi, *Attorney General*
Jeff Atwater, *Chief Financial Officer*
Adam Putnam, *Commissioner of Agriculture*

MEMORANDUM

TO: Department Inspectors and Agency Inspectors

FROM: Patrick Murphy, Program Manager *PJM*

DATE: March 6, 2015

SUBJECT: Approved Lot of Alcohol Reference Solution – Lot 201502G

The following lot of alcohol reference solution has been approved for use in the State of Florida effective March 6, 2015:

<u>Lot Number</u>	<u>Concentration</u>	<u>Manufacture Date</u>	<u>Expiration Date</u>
201502G	0.080 g/210L	02/24/2015	02/24/2017

The official distributor of this lot is Alcohol Countermeasure Systems, Inc., 5818 Hoffner Avenue, Suite 903, Orlando, Florida 32822-4816. Telephone (877) 227-7112. Florida approved alcohol reference solution may also be obtained from other vendors in Florida.

If you have any questions concerning this approved lot, please contact me at (850) 617-1280.

PJM/pjm