

MEMORANDUM

TO: Department Inspectors and Agency Inspectors
FROM: Brett Kirkland, Program Manager *BK*
DATE: May 2nd, 2016
SUBJECT: Approved Lot of Alcohol Reference Solution – Lot 201604C

The following lot of alcohol reference solution has been approved for use in the State of Florida effective May 2nd, 2016:

<u>Lot Number</u>	<u>Concentration</u>	<u>Manufacture Date</u>	<u>Expiration Date</u>
201604C	0.200 g/210L	4/5/2016	4/5/2018

The official distributor of this lot is Alcohol Countermeasure Systems, Inc., 5818 Hoffner Avenue, Suite 903, Orlando, Florida 32822-4816. Telephone (877) 227-7112. Florida approved alcohol reference solution may also be obtained from other vendors in Florida.

If you have any questions concerning this approved lot, please contact me at (850) 617-1275.

BHK/bhk



Florida Department of
Law Enforcement

Richard L. Swearingen
Commissioner

Criminal Justice Professionalism
Post Office Box 1489
Tallahassee, Florida 32302-1489
(850) 410-8600
www.fdle.state.fl.us

Rick Scott, *Governor*
Pam Bondi, *Attorney General*
Jeff Atwater, *Chief Financial Officer*
Adam Putnam, *Commissioner of Agriculture*

May 2nd, 2016

Mr. Christopher Wilson
Alcohol Countermeasure Systems, Inc.
60 International Boulevard
Toronto, Ontario M9W 612

Subject: Florida Acceptance of Alcohol Reference Solution Lot 201604C

Dear Mr. Wilson:

The Florida Department of Law Enforcement Alcohol Testing Program received bottles of alcohol reference solution, Lot 201604C, as samples for approval on April 15, 2016. Gas chromatographic analyses of ten (10) bottles from this lot was completed on April 27th, 2016. This lot meets the requirements of Chapter 11D-8, Florida Administrative Code.

Lot 201604C is approved for use in Florida effective May 2nd, 2016.

If you have any questions regarding the analyses, please feel free to contact me at (850) 617-1275.

Sincerely,

A handwritten signature in cursive script that reads "Brett Kirkland".

Brett Kirkland, Ph.D.
Program Manager
Alcohol Testing Program