



Florida Department of
Law Enforcement

Richard L. Swearingen
Commissioner

Criminal Justice Professionalism
Post Office Box 1489
Tallahassee, Florida 32302-1489
(850) 410-8600
www.fdle.state.fl.us

Rick Scott, *Governor*
Pam Bondi, *Attorney General*
Jeff Atwater, *Chief Financial Officer*
Adam Putnam, *Commissioner of Agriculture*

April 10th, 2017

Mr. Christopher Wilson
Alcohol Countermeasure Systems, Inc.
60 International Boulevard
Toronto, Ontario M9W 612

Subject: Florida Acceptance of Alcohol Reference Solution Lot 201703A

Dear Mr. Wilson:

The Florida Department of Law Enforcement Alcohol Testing Program received bottles of alcohol reference solution, Lot 201703A, as samples for approval on March 14th, 2017. Gas chromatographic analyses of ten (10) bottles from this lot was completed on April 6th, 2017. This lot meets the requirements of Chapter 11D-8, Florida Administrative Code.

Lot 201703A is approved for use in Florida effective April 10th, 2017.

If you have any questions regarding the analyses, please feel free to contact me at (850) 617-1275.

Sincerely,

A handwritten signature in blue ink that reads "Brett Kirkland". The signature is written in a cursive, flowing style.

Brett Kirkland, Ph.D.
Program Manager
Alcohol Testing Program
Florida Department of Law Enforcement

MEMORANDUM

TO: Department Inspectors and Agency Inspectors
FROM: Brett Kirkland, Program Manager *BK*
DATE: April 10th, 2017
SUBJECT: Approved Lot of Alcohol Reference Solution – Lot 201703A

The following lot of alcohol reference solution has been approved for use in the State of Florida effective April 10th, 2017:

<u>Lot Number</u>	<u>Concentration</u>	<u>Manufacture Date</u>	<u>Expiration Date</u>
201703A	0.050 g/210L	3/7/2017	3/7/2019

The official distributor of this lot is Alcohol Countermeasure Systems, Inc., 5818 Hoffner Avenue, Suite 903, Orlando, Florida 32822-4816. Telephone (877) 227-7112. Florida approved alcohol reference solution may also be obtained from other vendors in Florida.

This lot was analyzed using section 2.14 of the ATP standard operating procedure, effective 12/14/2016.

If you have any questions concerning this approved lot, please contact me at (850) 617-1275.

BHK/bhk