



Florida Department of  
Law Enforcement

Richard L. Swearingen  
*Commissioner*

**Criminal Justice Professionalism**  
Post Office Box 1489  
Tallahassee, Florida 32302-1489  
(850) 410-8600  
www.fdle.state.fl.us

Ron DeSantis, *Governor*  
Ashley Moody, *Attorney General*  
Jimmy Patronis, *Chief Financial Officer*  
Nikki Fried, *Commissioner of Agriculture*

## MEMORANDUM

TO: Department Inspectors and Agency Inspectors

FROM: Brett Kirkland, Ph.D., Program Manager *BK*

DATE: September 1, 2020

SUBJECT: Approved Lot of Alcohol Reference Solution – Lot 202007A

---

The following lot of alcohol reference solution has been approved for use in the State of Florida effective September 1, 2020:

<u>Lot Number</u>	<u>Concentration</u>	<u>Manufacture Date</u>	<u>Expiration Date</u>
202007A	0.080 g/210L	7/21/2020	7/21/2022

The official distributor of this lot is Alcohol Countermeasure Systems, Inc., 5818 Hoffner Avenue, Suite 903, Orlando, Florida 32822-4816. Telephone (877) 227-7112. Florida approved alcohol reference solution may also be obtained from other vendors in Florida.

This lot was analyzed using section 4 of the ATP technical standard operating procedure, effective April 2020.

If you have any questions concerning this approved lot, please contact me at (850) 617-1290.

BHK