

Florida Department of Law Enforcement Alcohol Testing Program

CERTIFICATE OF ASSURANCE

The Florida Department of Law Enforcement Alcohol Testing Program approves alcohol reference solution lots used to test the accuracy and precision of evidentiary breath test instruments. Alcohol reference solution lots are prepared by the Department or by a source approved by the Department, and their analysis is important to establish the reliability of alcohol breath test results.

Lot Number: 202108D

Simulated Breath Alcohol Concentration (g/210L): 0.080 g/ 210L

Date of Approval: August 25, 2021


Source: Alcohol Countermeasure Systems, Inc.

Date of Manufacture: August 17, 2021

For Use Through: August 17, 2023

Number of Bottles Manufactured: 2255

This certifies that ten sample bottles from the above lot were subjected to duplicate gas chromatographic analyses of their alcohol concentrations, that all the results were within the acceptable range, and that this lot of alcohol reference solution is in compliance with the requirements of Chapter 11D-8, F.A.C. and is approved for distribution and use in the State of Florida.



Signature

T.J. Graham

Name (Printed or Typed)

Program Manager

Title

August 25, 2021

Date