

# Florida Department of Law Enforcement

## Alcohol Testing Program

---

### CERTIFICATE OF ASSURANCE

The Florida Department of Law Enforcement Alcohol Testing Program approves alcohol reference solution lots used to test the accuracy and precision of evidentiary breath test instruments. Alcohol reference solution lots are prepared by the Department or by a source approved by the Department, and their analysis is important to establish the reliability of alcohol breath test results.

Lot Number: 202201E  
Simulated Breath Alcohol Concentration (g/210L): 0.200 g/210L  
Date of Approval: January 26, 2022

Source: Alcohol Countermeasures Systems, Inc.  
Date of Manufacture: January 18, 2022  
For Use Through: January 18, 2024  
Number of Bottles Manufactured: 378

This certifies that ten sample bottles from the above lot were subjected to duplicate gas chromatographic analyses of their alcohol concentrations, that all the results were within the acceptable range, and that this lot of alcohol reference solution is in compliance with the requirements of Chapter 11D-8, F.A.C. and is approved for distribution and use in the State of Florida.



Signature

T.J. Graham

Name (Printed or Typed)

Program Manager

Title

January 26, 2022

Date