## **MEMORANDUM**

TO:

Department Inspectors and Agency Inspectors

FROM:

Brett Kirkland, Program Manager &K

DATE:

March 9<sup>th</sup>, 2017

SUBJECT:

Approved Lot of Alcohol Reference Solution – Lot 201702B

The following lot of alcohol reference solution has been approved for use in the State of Florida effective March 9<sup>th</sup>, 2017:

Lot Number	Concentration	<b>Manufacture Date</b>	<b>Expiration Date</b>
201702B	0.200 g/210L	2/23/2017	2/23/2019

The official distributor of this lot is Alcohol Countermeasure Systems, Inc., 5818 Hoffner Avenue, Suite 903, Orlando, Florida 32822-4816. Telephone (877) 227-7112. Florida approved alcohol reference solution may also be obtained from other vendors in Florida.

This lot was analyzed using section 2.14 of the ATP standard operating procedure, effective 12/14/2016.

If you have any questions concerning this approved lot, please contact me at (850) 617-1275.

BHK/bhk



Florida Department of Law Enforcement

Richard L. Swearingen Commissioner

Criminal Justice Professionalism Post Office Box 1489 Tallahassee, Florida 32302-1489 (850) 410-8600 www.fdle.state.fl.us

Rick Scott, Governor Pam Bondi, Attorney General Jeff Atwater, Chief Financial Officer Adam Putnam, Commissioner of Agriculture

March 9th, 2017

Mr. Christopher Wilson Alcohol Countermeasure Systems, Inc. 60 International Boulevard Toronto, Ontario M9W 612

Subject: Florida Acceptance of Alcohol Reference Solution Lot 201702B

Dear Mr. Wilson:

The Florida Department of Law Enforcement Alcohol Testing Program received bottles of alcohol reference solution, Lot 201702BB, as samples for approval on March 2<sup>nd</sup>, 2017. Gas chromatographic analyses of ten (10) bottles from this lot was completed on March 3<sup>rd</sup>, 2017. This lot meets the requirements of Chapter 11D-8, Florida Administrative Code.

## Lot 201702B is approved for use in Florida effective March 9th, 2017.

If you have any questions regarding the analyses, please feel free to contact me at (850) 617-1275.

Sincerely,

Brett Kirkland, Ph.D.

Program Manager

Alcohol Testing Program

Florida Department of Law Enforcement

Butt Kirkland