

Florida Department of Law Enforcement

Richard L. Swearingen Commissioner

Criminal Justice Professionalism Post Office Box 1489 Tallahassee, Florida 32302-1489 (850) 410-8600 www.fdle.state.fl.us

Rick Scott, *Governor*Pam Bondi, *Attorney General*Jeff Atwater, *Chief Financial Officer*Adam Putnam, *Commissioner of Agriculture*

December 16th, 2016

Mr. Christopher Wilson Alcohol Countermeasure Systems, Inc. 60 International Boulevard Toronto, Ontario M9W 612

Subject: Florida Acceptance of Alcohol Reference Solution Lot 201611B

Dear Mr. Wilson:

The Florida Department of Law Enforcement Alcohol Testing Program received bottles of alcohol reference solution, Lot 201611B, as samples for approval on November 28th, 2016. Gas chromatographic analyses of ten (10) bottles from this lot was completed on December 15th, 2016. This lot meets the requirements of Chapter 11D-8, Florida Administrative Code.

Lot 201611B is approved for use in Florida effective December 16th, 2016.

If you have any questions regarding the analyses, please feel free to contact me at (850) 617-1275.

Sincerely,

Brett Kirkland, Ph.D.

Program Manager

Alcohol Testing Program

Florida Department of Law Enforcement

Britt thirteland

MEMORANDUM

TO:

Department Inspectors and Agency Inspectors

FROM:

Brett Kirkland, Program Manager &K

DATE:

December 16, 2016

SUBJECT:

Approved Lot of Alcohol Reference Solution – Lot 201611B

The following lot of alcohol reference solution has been approved for use in the State of Florida effective Dec 16th, 2016:

Lot Number	Concentration	Manufacture Date	Expiration Date
201611B	0.080 g/210L	11/15/2016	11/15/2018

The official distributor of this lot is Alcohol Countermeasure Systems, Inc., 5818 Hoffner Avenue, Suite 903, Orlando, Florida 32822-4816. Telephone (877) 227-7112. Florida approved alcohol reference solution may also be obtained from other vendors in Florida.

If you have any questions concerning this approved lot, please contact me at (850) 617-1275.

BHK/bhk