

Florida Department of Law Enforcement

Richard L. Swearingen Commissioner Criminal Justice Professionalism Post Office Box 1489 Tallahassee, Florida 32302-1489 (850) 410-8600 www.fdle.state.fl.us Rick Scott, Governor Pam Bondi, Attorney General Jimmy Patronis, Chief Financial Officer Adam Putnam, Commissioner of Agriculture

August 14, 2017

Mr. Christopher Wilson Alcohol Countermeasure Systems, Inc. 60 International Boulevard Toronto, Ontario M9W 612

Subject: Florida Acceptance of Alcohol Reference Solution Lot 201707D

Dear Mr. Wilson:

The Florida Department of Law Enforcement Alcohol Testing Program received bottles of alcohol reference solution, Lot 201707D, as samples for approval on August 2nd, 2017. Gas chromatographic analyses of ten (10) bottles from this lot was completed on August 8th, 2017. This lot meets the requirements of Chapter 11D-8, Florida Administrative Code.

Lot 201707D is approved for use in Florida effective August 14th 2017.

If you have any questions regarding the analyses, please feel free to contact me at (850) 617-1275.

Sincerely,

Brett Kirkland, Ph.D.

Program Manager

Alcohol Testing Program

Brett Kirkland

Florida Department of Law Enforcement

MEMORANDUM

TO:

Department Inspectors

FROM:

Brett Kirkland, Program Manager

DATE:

August 14th, 2017

SUBJECT:

Approved Lot of Alcohol Reference Solution - Lot 201707D

The following lot of alcohol reference solution has been approved for use in the State of Florida effective August 14th, 2017:

Lot Number	Concentration	Manufacture Date	Expiration Date
201707D	0.050 g/210L	7/25/2017	7/25/2019

The official distributor of this lot is Alcohol Countermeasure Systems, Inc., 5818 Hoffner Avenue, Suite 903, Orlando, Florida 32822-4816. Telephone (877) 227-7112. Florida approved alcohol reference solution may also be obtained from other vendors in Florida.

This lot was analyzed using section 2.14 of the ATP standard operating procedure, effective May 2017.

If you have any questions concerning this approved lot, please contact me at (850) 617-1275.

BHK/bhk