

September 4, 2018

Mr. Christopher Wilson
Alcohol Countermeasure Systems, Inc.
60 International Boulevard
Toronto, Ontario M9W 612

Subject: Florida Acceptance of Alcohol Reference Solution Lot 201808E

Dear Mr. Wilson:

The Florida Department of Law Enforcement Alcohol Testing Program received bottles of alcohol reference solution, Lot 201808E, as samples for approval on August 27th, 2018. Gas chromatographic analyses of ten (10) bottles from this lot was completed on August 27th, 2018. This lot meets the requirements of Chapter 11D-8, Florida Administrative Code.

Lot 201808E is approved for use in Florida effective September 4th, 2018.

If you have any questions regarding the analyses, please feel free to contact me at (850) 617-1275.

Sincerely,



Brett Kirkland, Ph.D.
Program Manager
Alcohol Testing Program
Florida Department of Law Enforcement

MEMORANDUM

TO: Department Inspectors and Agency Inspectors

FROM: Brett Kirkland, Program Manager *BK*

DATE: September 4, 2018

SUBJECT: Approved Lot of Alcohol Reference Solution – Lot 201808E

The following lot of alcohol reference solution has been approved for use in the State of Florida effective September 4th, 2018:

<u>Lot Number</u>	<u>Concentration</u>	<u>Manufacture Date</u>	<u>Expiration Date</u>
201808E	0.080 g/210L	8/22/2018	8/22/2020

The official distributor of this lot is Alcohol Countermeasure Systems, Inc., 5818 Hoffner Avenue, Suite 903, Orlando, Florida 32822-4816. Telephone (877) 227-7112. Florida approved alcohol reference solution may also be obtained from other vendors in Florida.

This lot was analyzed using section 4 of the ATP standard operating procedure, effective October 2017.

If you have any questions concerning this approved lot, please contact me at (850) 617-1290.

BHK