



Florida Department of  
Law Enforcement

Richard L. Swearingen  
*Commissioner*

**Criminal Justice Professionalism**  
Post Office Box 1489  
Tallahassee, Florida 32302-1489  
(850) 410-8600  
[www.fdle.state.fl.us](http://www.fdle.state.fl.us)

Rick Scott, *Governor*  
Pam Bondi, *Attorney General*  
Jimmy Patronis, *Chief Financial Officer*  
Adam Putnam, *Commissioner of Agriculture*

December 4, 2018

Mr. Christopher Wilson  
Alcohol Countermeasure Systems, Inc.  
60 International Boulevard  
Toronto, Ontario M9W 612

Subject: Florida Acceptance of Alcohol Reference Solution Lot 201811C

Dear Mr. Wilson:

The Florida Department of Law Enforcement Alcohol Testing Program received bottles of alcohol reference solution, Lot 201811C, as samples for approval on November 20<sup>th</sup>, 2018. Gas chromatographic analyses of ten (10) bottles from this lot was completed on December 3<sup>rd</sup>, 2018. This lot meets the requirements of Chapter 11D-8, Florida Administrative Code.

**Lot 201811C is approved for use in Florida effective December 4<sup>th</sup>, 2018.**

If you have any questions regarding the analyses, please feel free to contact me at (850) 617-1275.

Sincerely,

A handwritten signature in blue ink that reads "Brett Kirkland".

Brett Kirkland, Ph.D.  
Program Manager  
Alcohol Testing Program  
Florida Department of Law Enforcement

**MEMORANDUM**

TO: Department Inspectors and Agency Inspectors  
FROM: Brett Kirkland, Program Manager *BK*  
DATE: December 4, 2018  
SUBJECT: Approved Lot of Alcohol Reference Solution – Lot 201811C

---

The following lot of alcohol reference solution has been approved for use in the State of Florida effective December 4<sup>th</sup>, 2018:

<u>Lot Number</u>	<u>Concentration</u>	<u>Manufacture Date</u>	<u>Expiration Date</u>
201811C	0.050 g/210L	11/13/2018	11/13/2020

The official distributor of this lot is Alcohol Countermeasure Systems, Inc., 5818 Hoffner Avenue, Suite 903, Orlando, Florida 32822-4816. Telephone (877) 227-7112. Florida approved alcohol reference solution may also be obtained from other vendors in Florida.

This lot was analyzed using section 4 of the ATP standard operating procedure, effective October 2017.

If you have any questions concerning this approved lot, please contact me at (850) 617-1290.

BHK