

Florida Department of Law Enforcement

Richard L. Swearingen Commissioner

Criminal Justice Professionalism Post Office Box 1489 Tallahassee, Florida 32302-1489 (850) 410-8600 www.fdle.state.fl.us

Ron DeSantis, Governor Ashley Moody, Attorney General Jimmy Patronis, Chief Financial Officer Nikki Fried, Commissioner of Agriculture

January 26, 2022

Mr. Christopher Wilson Alcohol Countermeasure Systems, Inc. 60 International Boulevard Toronto, Ontario M9W 612

Subject: Florida Acceptance of Alcohol Reference Solution Lot 202201D

Dear Mr. Wilson:

The Florida Department of Law Enforcement Alcohol Testing Program received bottles of alcohol reference solution, Lot 202201D, as samples for approval on January 24, 2022. Gas chromatographic analyses of ten (10) bottles from this lot was completed on January 25, 2022. This lot meets the requirements of Chapter 11D-8, Florida Administrative Code.

## Lot 202201D is approved for use in Florida effective January 26, 2022.

If you have any questions regarding the analyses, please feel free to contact me at (850) 617-1275.

Sincerely,

TJ Graham, D-ABFT-FA

Program Manager

Alcohol Testing Program

Florida Department of Law Enforcement

## **MEMORANDUM**

TO: Department Inspectors

FROM: TJ Graham, Program Manager

DATE: January 26, 2022

SUBJECT: Approved Lot of Alcohol Reference Solution – Lot 202201D

The following lot of alcohol reference solution has been approved for use in the State of Florida effective January 26, 2022:

Lot Number	Concentration	Manufacture Date	Expiration Date
202201D	0.080 g/210L	01/18/2022	01/18/2024

The official distributor of this lot is Alcohol Countermeasure Systems, Inc., 5818 Hoffner Avenue, Suite 903, Orlando, Florida 32822-4816. Telephone (877) 227-7112. Florida approved alcohol reference solution may also be obtained from other vendors in Florida.

This lot was analyzed using section 4 of the ATP technical standard operating procedure, effective Jan 2022.

If you have any questions concerning this approved lot, please contact me at (850) 617-1290.

John