

316 E. 9th St.
Owensboro, KY 42303
1-800-835-0690
Fax: 270-685-6678
www.alcoholtest.com

August 11, 2021

TJ Graham, Manager
Florida Department of Law Enforcement
Alcohol Testing Program
2331 Philips Road
Tallahassee, FL 32308

RE: Modification of the Intoxilyzer 8000


Dear Mr. Graham,

The following is to notify you of the anticipated hardware changes to the Intoxilyzer 8000 currently being utilized in the State of Florida:

1. Cell insulation (part number 470210) is now being supplied by an alternate vendor.

Rest assured these changes do not affect the accuracy or precision of the Intoxilyzer 8000 currently being utilized in the State of Florida. If you have any questions or concerns regarding the changes, please feel free to call.

Sincerely,



Jon Grantham
Engineering Manager





316 E. 9th St.
Owensboro, KY 42303
1-866-835-0690
Fax: 270-685-6678
www.alcoholtest.com

May 23, 2019

Florida Department of Law Enforcement
FDLE Off-Site Mail Center
813B Lake Bradford Road
FDLE Alcohol Testing Program
Tallahassee, FL 32304

Attn: Dr. Brett Kirkland
Alcohol Test Program Manager

RE: Modification of the Intoxilyzer 8000

Dear Dr. Kirkland:

CMI has updated the Intoxilyzer 8000 DSP firmware to 1362.16. New instruments shipped after May 21, 2019 will include this update. The Intoxilyzer 8000 instruments currently in use in the State of Florida do not require the updated firmware; however, any instruments that are sent in for repair may be updated.

This update addresses DSP diagnostic issues when the instrument transitions to Ready mode, after power on.

Rest assured these changes do not affect the accuracy or precision of the Intoxilyzer 8000 currently being utilized in the State of Florida. This update will not be noticeable in any way during the operation of the instrument. The Florida Intoxilyzer 8000 software version, 8100.27.00, remains the same. If you have any questions or concerns regarding the changes, please feel free to call.

Sincerely,

A handwritten signature in black ink, appearing to read "Jon Grantham", is written over a horizontal line.

Jon Grantham,
Manager, CMI Engineering

INTOXILYZER® ...so you can breathe easier



FDLE-ATP
2331 Phillips Rd
Tallahassee, FL 32301
Office: (850) 617-1290
Fax: (850) 921-3787
<http://www.fdle.state.fl.us>

Florida Department of
Law Enforcement



Fax

To: Jon Grantham	From: T. J. Graham
Fax: (270) 685-6525	Pages: 3
Phone:	Date: 3/5/19
Re: Form line break	cc:

Urgent For Review Please Comment Please Reply Please Recycle

Comments:

Per our telephone discussion.

PLEASE NOTE: Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your fax messages may, therefore, be subject to public disclosure.

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ESCAMBIA COUNTY SO
Time of Inspection: 06:33

Date of Inspection: 02/13/2019

Serial Number: 80-006436
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:770784 Exp: 04/27/2020
0.000	0.049	0.077	0.200	0.000 / 0.080
0.000	0.049	0.077	0.200	0.055 / 0.080
0.000	0.050	0.077	0.200	0.080 / 0.080

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance | GAS TUBE NOT CONNECTED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name JAMES A DANIELS

02/13/2019
Date

131962653463623203

Intoxilyzer 8000 Alcohol Analyzer
Instrument Serial Number: 80-006436
Instrument SW Version: 8100.27
Instrument Location: ESCAMBIA COUNTY SO
Event time: 06:33
Event date: 02/13/2019
Control Value: 0.080

Operator last name: DANIELS
Operator first name: JAMES
Operator middle initial: A
Inspection .05 Lot #: 201802B
Inspection .05 Expiry: 02/06/2020
Inspection .08 Lot #: 201709A
Inspection .08 Expiry: 09/12/2019
Inspection .20 Lot #: 201702B
Inspection .20 Expiry: 02/23/2019
Inspection Remarks: 08: Control Outside Tolerance
Inspection complies: GAS TUBE
Simulators used for inspection: 0
Inspection Standard Deviation: 5
Inspection Type:
Inspection test type: 770784
Test sequence #: 04/27/2020

Result value	Result exception	Inspection
Gas Lot #	Inspection Gas Expiry	

TRANSMISSION VERIFICATION REPORT

TIME : 03/05/2019 09:00
NAME : ALCOHOL TESTING
FAX : 8509213787
TEL : 8504107810
SER.# : BROB4J286861

DATE, TIME : 03/05 08:59
FAX NO./NAME : 612706856525
DURATION : 00:00:45
PAGE(S) : 03
RESULT : OK
MODE : STANDARD
ECM

FDLE-ATP
2331 Phillips Rd
Tallahassee, FL 32301
Office: (850) 617-1290
Fax: (850) 921-3787
<http://www.fdle.state.fl.us>

**Florida Department of
Law Enforcement**



Fax

To: Jon Grantham From: T. J. Graham
Fax: (270) 685-6525 Pages: 3
Phone: _____ Date: 3/5/19
Re: Form line break cc: _____

Urgent For Review Please Comment Please Reply Please Recycle

Comments:

Per our telephone discussion.

FDLE-ATP
2331 Phillips Rd
Tallahassee, FL 32301
Office: (850) 617-1290
Fax: (850) 921-3787
<http://www.fdle.state.fl.us>

Florida Department of
Law Enforcement



Fax

To: Jon Grantham	From: T. J. Graham
Fax: (270) 685-6525	Pages: 3
Phone:	Date: 3/5/19
Re: Form line break	cc:

Urgent For Review Please Comment Please Reply Please Recycle

Comments:

Per our telephone discussion.

PLEASE NOTE: Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your fax messages may, therefore, be subject to public disclosure.

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ESCAMBIA COUNTY SO
Time of Inspection: 06:33

Date of Inspection: 02/13/2019

Serial Number: 80-006436
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:201802B Exp: 02/06/2020	0.08g/210L Test (g/210L) Lot#:201709A Exp: 09/12/2019	0.20g/210L Test (g/210L) Lot#:201702B Exp: 02/23/2019	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:770784 Exp: 04/27/2020
0.000	0.049	0.077	0.200	0.000 / 0.080
0.000	0.049	0.077	0.200	0.055 / 0.080
0.000	0.050	0.077	0.200	0.080 / 0.080

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance | GAS TUBE NOT CONNECTED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name JAMES A DANIELS

02/13/2019
Date

131962653463623203

Intoxilyzer 8000 Alcohol Analyzer
Instrument Serial Number: 80-006436
Instrument SW Version: 8100.27
Instrument Location: ESCAMBIA COUNTY SO
Event time: 06:33
Event date: 02/13/2019
Control Value: 0.080

Operator last name: DANIELS
Operator first name: JAMES
Operator middle initial: A
Inspection .05 Lot #: 201802B
Inspection .05 Expiry: 02/06/2020
Inspection .08 Lot #: 201709A
Inspection .08 Expiry: 09/12/2019
Inspection .20 Lot #: 201702B
Inspection .20 Expiry: 02/23/2019
Inspection Remarks: 08: Control Outside Tolerance
Inspection complies: GAS TUBE
Simulators used for inspection: 0
Inspection Standard Deviation: 5
Inspection Type:
Inspection test type: 770784
Test sequence #: 04/27/2020

Result value	Result exception	Inspection
Gas Lot #	Inspection Gas Expiry	

TRANSMISSION VERIFICATION REPORT

TIME : 03/05/2019 09:00
NAME : ALCOHOL TESTING
FAX : 8509213787
TEL : 8504107810
SER.# : BROB4J286861

DATE, TIME : 03/05 08:59
FAX NO./NAME : 612706856525
DURATION : 00:00:45
PAGE(S) : 03
RESULT : OK
MODE : STANDARD
ECM

FDLE-ATP
2331 Phillips Rd
Tallahassee, FL 32301
Office: (850) 617-1290
Fax: (850) 921-3787
<http://www.fdle.state.fl.us>

**Florida Department of
Law Enforcement**



Fax

To: Jon Grantham From: T. J. Graham
Fax: (270) 685-6525 Pages: 3
Phone: _____ Date: 3/5/19
Re: Form line break cc: _____

Urgent For Review Please Comment Please Reply Please Recycle

Comments:

Per our telephone discussion.

TRANSMISSION VERIFICATION REPORT

TIME : 03/05/2019 09:04
NAME : ALCOHOLTESTING
FAX : 8509213787
TEL : 8504107810
SER.# : BROB4J286861

DATE, TIME : 03/05 09:03
FAX NO./NAME : 612706856525
DURATION : 00:00:47
PAGE(S) : 03
RESULT : OK
MODE : STANDARD
ECM

FDLE-ATP
2331 Phillips Rd
Tallahassee, FL 32301
Office: (850) 617-1290
Fax: (850) 921-3787
<http://www.fdle.state.fl.us>



Fax

To: Jon Grantham From: T. J. Graham
Fax: (270) 685-6525 Pages: 3
Phone: Date: 3/5/19
Re: Form line break cc:

Urgent For Review Please Comment Please Reply Please Recycle

Comments:

Per our telephone discussion.

From: [Graham, Thomas](#)
To: [Hagan, Pam](#)
Cc: [Murphy, Patrick](#)
Subject: Florida Intoxilyzer 8000s
Date: Tuesday, December 11, 2018 7:58:00 AM

Hi Pam,

We are nearing the end of 2018. I hope this year has gone well for you. As we wrap up our year here at FDLE, we need to confirm the location of a few instruments we believe are at CMI for repair. Could you send me a list of the Florida 8000s that are currently at CMI? I would specifically like to know if you still have the following:

80-000763
80-000958
80-000959
80-001232
80-001724
80-006472

Thank you in advance for any help you can provide.

TJ Graham, D-ABFT-FA
Quality Assurance Manager
Alcohol Testing Program
Florida Department of Law Enforcement
Office:850-617-1285
Cell: 850-445-7938

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.

From: [Graham, Thomas](#)
To: [Hagan, Pam](#)
Cc: [Murphy, Patrick](#)
Subject: FDLE 8000s
Date: Tuesday, December 11, 2018 8:24:00 AM

Pam,

Please don't waste your time looking for 80-000959 and 80-001724. These two instruments were deregistered a year ago.

TJ Graham, D-ABFT-FA
Quality Assurance Manager
Alcohol Testing Program
Florida Department of Law Enforcement
Office:850-617-1285
Cell: 850-445-7938

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.



316 E. 9th St.
Owensboro, KY 42303
1-800-835-0690
Fax: 270-685-6678
www.alcoholtest.com

FAX TRANSMISSION

TO: TJ (Company)
FROM: Pam Hagan (Fax Number) 850-921-3787
(Return Fax Number) (270) 685-6678
SUBJECT: Service Instructions COPIES TO: _____
DATE: _____ Number of Pages: 2

INTOXILYZER® ...so you can breathe easier



TOTAL P.02

80-000958
80-001232
80-000763
80-006472
80-001178
80-000785
80-000813
80-000781
80-000743
80-000202
80-007006
80-000811
80-000775
80-000948