

## **Executive Summary**

**Issue**                    **Reporting of Volume Not Met Results on the Intoxilyzer 8000 – Update  
9/25/2006 at 11:00 hours**

### **Background**

- One of the minimum breath sample requirements of the Intoxilyzer 8000 is a breath volume of at least 1.1 liters.
- Any breath sample that has a breath volume of less than 1.1 liters should be reported as Volume Not Met. Subsequent breath samples should be requested if this occurs on Breath Sample #1 or Breath Sample #2. If both Breath Sample #1 and Breath Sample #2 are less than 1.1 liters, the instrument will abort the test and operators are trained to begin another test.
- Research Studies were conducted by the Alcohol Testing Program on the Intoxilyzer 8000 using software version 8100.26 (the only software version to be used for evidentiary purposes) in January 2006, July 2006 and August 2006 and specifically tested the functionality of a breath sample not meeting the required volume to ensure the appropriate instrument response. During each Research Study, the Intoxilyzer 8000 responded correctly and appropriately to breath samples which were not at least 1.1 liters.

### **Discussion**

- The Intoxilyzer 8000 is occasionally and sporadically not identifying samples of less than 1.1 liters as Volume Not Met. Instead, it is accepting the breath sample as meeting the minimum breath sample requirements and reporting a numerical result.
- This issue was first identified by Stuart Hyman, defense counsel in Orange, Seminole and Brevard counties, at an Administrative Driver's License Hearing on Thursday, September 21, 2006.

### **Recommendation**

- The Alcohol Testing Program is currently reviewing the data from all breath tests conducted on evidentiary breath test instruments in Florida. This is approximately 10,000 breath tests.
- Breath samples which do not meet at least 1.1 liters and were not appropriately responded to by the instrument are being identified.
- After determining the number of tests and the instruments affected, the manufacturer, CMI, Inc., will be contacted and the issue discussed.
- Additional briefings with members of FDLE will subsequently occur.

## Executive Summary

**Issue**                    **Reporting of Volume Not Met Results on the Intoxilyzer 8000 – Update 9/29/2006 at 13:30 hours**

### Background

- See the Executive Summary: Reporting of Volume Not Met Results on the Intoxilyzer 8000 – Update 9/25/2006 at 11:00 hours.

### Discussion

- On 9/26/2006, the manufacturer of the Intoxilyzer 8000, CMI, Inc., was contacted regarding the breath samples which do not meet volume not being appropriately identified.
- On 9/27/2006, CMI, Inc. informed the Alcohol Testing Program that they had identified the problem. When a subject is providing a breath sample during and shortly after the three minutes allowed to provide the breath sample, the software was programmed to accept the breath sample but it was not programmed to look at the volume of the breath sample provided.
  - The software (version 8100.26) is working as programmed and written. A software change is needed to fix this “loophole” to write code to tell the instrument to look at the final volume when a subject is providing a breath sample past three minutes.
  - An example to further explain this is as follows: If the subject begins providing a breath sample at 2.58 minutes, continues to provide the sample until 3.02 minutes, and the breath sample at 3.02 minutes is less than 1.1 Liters, the instrument will accept the breath sample and calculate a result but it will not verify that the volume is at least 1.1 Liters. Therefore breath sample is not flagged as Volume Not Met.
- On 9/28/2006, CMI, Inc. notified the Alcohol Testing Program that the software code had been written and the engineering section was testing the revised software to ensure it is working properly.
- The Alcohol Testing Program has reviewed approximately 23,000 breath tests for any breath sample that is less than 1.1 Liters and is not flagged as Volume Not Met. We have identified 181 breath tests (which may or may not be valid tests) that meet this scenario. This review encompasses all data uploaded to the central database as of September 27, 2006.

### Recommendation

- The Alcohol Testing Program will draft a letter which will be sent to criminal justice agencies, State Attorney's Offices, the Department of Highway Safety and Motor Vehicles, and agency inspectors describing this issue.
- Notice of affected breath tests will be provided to the State Attorney's Office and the Department of Highway Safety and Motor Vehicles. The criminal justice agency who conducted the breath test will be copied on all of these notices. This is a Brady notice.
- The Alcohol Testing Program will continue to monitor each breath test conducted in Florida to determine if the breath sample volume is less than 1.1 Liters and not flagged as Volume Not Met. Brady notices will be filed with any subsequent breath tests found as well.
- The Alcohol Testing Program will beta test the revised software version on Monday, October 2, 2006 and provide the information to CMI, Inc. Based on the review of the revised software, additional timelines for evaluation and implementation of the new software version will subsequently occur.
- Additional briefings with members of FDLE will subsequently occur.



Florida Department of  
Law Enforcement

Gerald M. Bailey  
Interim Commissioner

Criminal Justice  
Professionalism Program  
Alcohol Testing Program

P.O. Box 1489  
Tallahassee, Florida 32302  
(850) 410-8600  
(850) 410-7816 Fax  
<http://www.fdle.state.fl.us>

October 5, 2006

NAME  
ADDRESS  
ADDRESS  
ADDRESS

SUBJECT: Intoxilyzer 8000

Dear NAME:

This letter is to inform you of an issue regarding the Intoxilyzer 8000 that the Florida Department of Law Enforcement Alcohol Testing Program has discovered. We have been in contact with the manufacturer of the Intoxilyzer 8000, CMI, Inc., and have determined there are missing instructions in the Intoxilyzer 8000 software version 8100.26.

The issue involves the instrument's ability to correctly identify the volume of a breath sample delivered past three minutes. The programming in software version 8100.26 was not written to address this very unique and specific scenario. This issue is affecting less than 1% of all breath tests conducted on the Intoxilyzer 8000.

A software revision to add these missing instructions is in the process of being completed and will be installed in your Intoxilyzer 8000 instruments in the near future. **I would like to assure you that software version 8100.26 is operating as written and programmed. The analytical functions of the instrument are not affected by this issue and the instrument continues to correctly analyze all breath samples.** It is only with this unique and specific scenario that the breath sample has not been determined to be reliable.

A detailed letter is being forwarded under separate cover to your Agency Inspector. Enclosed is a summary of the affected breath tests determined as of October 2, 2006. The Alcohol Testing Program is continuing to closely monitor every breath test result obtained in the State of Florida and will contact you regarding any additional affected breath tests conducted by your agency. If you have any questions, please feel free to contact me at (850) 410-7810.

Sincerely,

Gerald M. Bailey  
Interim Commissioner

Laura D. Barfield  
Alcohol Testing Program Manager

LDB/lb

Enclosure

Subject Last Name	Subject First Name	Subject's Date of Birth	Offense	Instrument Serial Number	Instrument Registered To	Arresting Officer Agency	Date of Test	Time of Test	Breath Sample Volume Results	Notes
Corredor	Miguel A	1/7/1986	DUI	80-000223	FDLE (Loaned to Indian River County Sheriff's Office)	Florida Highway Patrol	7/6/2006	22 58 02	First Sample Volume 1 933 L - OK Second Sample Volume 0 972L - No Flag	Not A Valid Test
Sedovic	Michael T	1/4/1976	DUI	80-000738	Fort Walton Beach PD	Okaloosa County SO	8/5/2006	4 54 39	No Second Sample - Purge Fail	Not A Valid Test
Pelola	Temilola A	12/24/1961	DUI	80-000739	Okaloosa County SO	Florida Highway Patrol	8/21/2006	0 10 12	First Sample Volume 1 066L - No Flag Second Sample Volume 1 164L - OK	Not A Valid Test
Fulford	Kenneth M	7/7/1951	DUI	80-000744	Escambia County SO	Pensacola PD	5/17/2006	2 04 10	First Sample Volume 1 550L - OK Second Sample Volume 0 605L - No Flag	Not A Valid Test
Burgos	Abel	7/11/1978	DUI	80-000744	Escambia County SO	Florida Highway Patrol	7/22/2006	3 58 14	First Sample Volume 0 664L - No Flag Second Sample Volume 2 105L - OK	Not A Valid Test
East	Easier K	3/21/1945	DUI	80-000744	Escambia County SO	Florida Highway Patrol	8/12/2004	2 04 13	First Sample Volume 1 082L - No Flag Second Sample - Subject Refused	Not A Valid Test
Rives	Lisa C	6/8/1959	DUI	80-000744	Escambia County SO	Florida Highway Patrol	9/28/2006	21 07 00	First Sample Volume 0 601L - No Flag Second Sample Volume 1 515L - OK	Not A Valid Test
Swanson	Amanda N	9/9/1984	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	5/18/2006	5 39 00	First Sample Volume 0 062L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Hidle	Brandt C	11/9/1972	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	5/25/2006	1 05 16	First Sample Volume 0 129L - Flagged Volume Not Met Second Sample Volume 0 347L - No Flag Third Sample Volume 0 127L - Flagged Volume Not Met	Not A Valid Test
Toney	Patrick S	3/10/1987	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	5/30/2006	1 29 54	First Sample Volume 1 000L - No Flag Second Sample Volume 1 550L - OK	Not A Valid Test
Boone	Michael J	11/9/1983	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	6/25/2006	2 30 07	First Sample Volume 0 265L - No Flag Second Sample Volume 0 054L - Flagged Volume Not Met Third Sample Volume 1 398L - OK	Not A Valid Test
Alicandri	Danielle M	8/1/1971	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	7/25/2006	2 36 03	First Sample Volume 0 742L - No Flag Second Sample Volume 0 855L - Flag Volume Not Met Third Sample Volume 0 214L - Flag Volume Not Met	Not A Valid Test
Jones	Elaine J	9/18/1969	DUI	80-000772	Santa Rosa County SO	Santa Rosa County SO	5/13/2006	2 38 47	First Sample Volume 0 476L - No Flag Second Sample Volume 1 738L - OK	Not A Valid Test
Bohannon	Lynette D	12/18/1972	DUI	80-000786	Broward County SO	Broward County SO	5/12/2006	0 10 42	First Sample Volume 0 652L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Shelton	Linda M	8/27/1940	DUI	80-000790	Vero Beach PD	Vero Beach PD	7/13/2006	3 05 22	First Sample Volume 0 761L - Flag Volume Not Met Second Sample Volume 1 292L - OK Third Sample Volume 0 612L - No Flag	Not A Valid Test
Vaught	Michael A	4/20/1981	DUI	80-000794	St Lucie County SO	St Lucie County SO	4/9/2006	2 49 38	First Sample Volume 1 527L - OK Second Sample Volume 0 261L - No Flag Third Sample Volume 1 371L - OK	Not A Valid Test
Lesko	Daryl R	5/10/1958	DUI	80-000794	St Lucie County SO	St Lucie County SO	6/5/2006	1 49 07	First Sample Volume 0 210L - No Flag Second Sample Volume 1 222L - OK Third Sample Volume 0 515L - No Flag	Not A Valid Test
Conerz	Javier O	1/14/1964	DUI	80-000794	St Lucie County SO	Fort Pierce PD	6/10/2006	0 36 52	First Sample Volume 1 570L - OK Second Sample Volume 0 300L - No Flag Third Sample Volume 2 429L - OK	Not A Valid Test
Schneider	Robert P	4/23/1962	DUI	80-000794	St Lucie County SO	Fort Pierce PD	6/19/2006	20 12 44	First Sample Volume 1 113L - OK Second Sample Volume 0 234L - No Flag Third Sample Volume 1 133L - OK	Not A Valid Test
Slater	Keith	6/24/1972	DUI	80-000794	St Lucie County SO	Fort Pierce PD	6/21/2006	9 25 13	First Sample Volume 0 515L - No Flag Second Sample Volume 0 812L - No Flag Third Sample Volume 1 207L - OK	Not A Valid Test
Malszewski	Dagmar E	8/5/1967	DUI	80-000794	St Lucie County SO	Port St Lucie PD	6/25/2006	23 16 05	First Sample Volume 1 343L - OK Second Control Test Outside Tolerance	Not A Valid Test
Diaz	Alejandro	10/29/1984	DUI	80-000794	St Lucie County SO	St Lucie County SO	8/19/2006	3 20 08	First Sample Volume 0 812L - No Flag No Second Sample - Subject Refused	Not A Valid Test
Jimenez	David Y	1/24/1977	DUI	80-000808	Hillsborough County SO	Hillsborough County SO	3/30/2006	23 07 03	First Sample Volume 0 816L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Piteasa	Petar J	11/25/1982	DUI	80-000808	Hillsborough County SO	Tampa PD	7/2/2006	3 58 14	First Sample Volume 1 054L - No Flag Second Sample Volume 1 261L - OK	Not A Valid Test

Roberts	Pamela R	7/23/1984	DUI	80-000808	Hillsborough County SO	Tampa PD	8/24/2006	0.30.17	First Sample Volume 1.070L - No Flag Second Sample Volume 1.679L - OK	Not A Valid Test
Reyes	Tony	6/11/1984	DUI	80-000808	Hillsborough County SO	Plant City PD	9/3/2006	0.04.38	First Sample Volume 0.363L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Clark	Michael C	4/9/1983	DUI	80-000808	Hillsborough County SO	Florida Highway Patrol	9/9/2006	5.52.04	First Sample Volume 0.246L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Hargrave	Evelyn M	7/20/1979	DUI	80-000808	Hillsborough County SO	Tampa PD	9/10/2006	3.08.59	First Sample Volume 0.925L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Diaz Gonzalez	Raul N	11/24/1957	DUI	80-000813	Marion County SO	Marion County SO	6/21/2006	14.35.11	First Sample Volume 0.476L - No Flag Second Sample Volume 2.738L - OK Purge Fail After Second Sample	Not A Valid Test
Merriott	Brady J	9/9/1984	DUI	80-000817	Citrus County SO	Citrus County SO	9/14/2006	4.23.44	First Sample Volume 0.398L - No Flag Second Sample Volume 0.179L - No Flag	Not A Valid Test
Welch	Andrei C	3/20/1966	DUI	80-000824	Leesburg PD	Leesburg PD	4/15/2006	0.19.50	First Sample Volume 0.324L - No Flag Second Sample Volume 1.117L - OK	Not A Valid Test
Garcia	Alfonso R	8/17/1980	DUI	80-000829	Hillsborough County SO	Tampa PD	4/12/2006	2.16.24	First Sample Volume 0.339L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Nilsen	Michael R	10/9/1961	DUI	80-000832	Martin County SO	Florida Highway Patrol	5/29/2006	3.47.51	First Sample Volume 1.660L - OK Second Sample Volume 0.972L - No Flag Third Sample Volume 1.539L - OK	Valid Test - First and Third Samples
Camacho	Rafael	4/3/1961	DUI	80-000832	Martin County SO	Stuart PD	6/22/2006	19.54.57	First Sample Volume 0.597L - No Flag Second Sample Volume 0.204L - Flagged Volume Not Met Third Sample Volume 0.219L - Flagged Volume Not Met	Not A Valid Test
Mendoza	Cecilio	11/22/1962	DUI	80-000832	Martin County SO	Martin County SO	7/9/2006	2.12.46	First Sample Volume 0.000L - Flagged Volume Not Met Second Sample Volume 2.226L - OK Third Sample Volume 0.183 - No Flag	Not A Valid Test
Schroeder	Brandon H	7/28/1987	DUI	80-000832	Martin County SO	Martin County SO	8/13/2006	2.45.47	First Sample Volume 0.468L - Flagged Volume Not Met Second Sample Volume 0.769L - No Flag Third Sample Volume 1.480L - OK	Not A Valid Test
Daher	Jabib	2/3/1948	DUI	80-000832	Martin County SO	FWC West Palm	8/26/2006	23.30.17	First Sample Volume 1.726L - OK Second Sample Volume 0.449L - No Flag	Not A Valid Test
Vandermark	Jeffrey W	6/16/1945	DUI	80-000832	Martin County SO	Sewall's Point PD	9/4/2006	2.04.32	First Sample Volume 0.808L - No Flag Second Sample Volume 3.394L - OK Third Sample Volume 1.691L - OK	Not A Valid Test
Michale	James M	2/27/1967	DUI	80-000832	Martin County SO	Stuart PD	9/4/2006	21.48.28	First Sample Volume 1.187L - OK Second Control Test Outside Tolerance	Not A Valid Test
Moretti	Derek P	6/2/1980	DUI	80-000833	Hillsborough County SO	Tampa PD	5/15/2006	1.09.00	First Sample Volume 0.742L - No Flag Second Sample Volume 0.196L - Flag Volume Not Met No Third Sample - Subject Refused	Not A Valid Test
Perez	Palermo	11/20/1972	DUI	80-000833	Hillsborough County SO	Hillsborough County SO	6/10/2006	19.51.30	First Sample Volume 0.105L - No Flag Second Sample Volume 0.212L - Flagged Volume Not Met Third Sample Volume 0.219L - Flagged Volume Not Met	Not A Valid Test
Pospischil	Ashley N	5/10/1983	DUI	80-000833	Hillsborough County SO	Tampa PD	7/3/2006	4.13.57	First Sample Volume 0.957L - No Flag Second Sample Volume 1.648L - OK	Not A Valid Test
Gandanilla	Petronio R	9/6/1972	DUI	80-000833	Hillsborough County SO	Plant City PD	7/21/2006	9.12.20	First Sample Volume 0.371L - No Flag Second Sample Volume 0.371L - No Flag	Not A Valid Test
Rojas	Oscar	8/27/1949	DUI	80-000833	Hillsborough County SO	Tampa PD	8/11/2006	2.46.47	First Sample Volume 0.625L - No Flag Second Sample - Subject Refused	Not A Valid Test
Rolon	Jose A	3/8/1961	DUI	80-000833	Hillsborough County SO	Florida Highway Patrol	8/13/2006	4.32.45	First Sample Volume 0.574L - No Flag Second Sample Volume 1.289L - OK	Not A Valid Test
Seymour	Patrick H	8/26/1967	DUI	80-000859	Cape Coral PD	Cape Coral PD	6/21/2006	20.35.16	First Sample Volume 0.429L - No Flag Second Sample Volume 0.194L - Flagged Volume Not Met Third Sample Volume 0.625L - No Flag	Not A Valid Test
Lisa	Denise A	12/9/1957	DUI	80-000859	Cape Coral PD	Cape Coral PD	7/10/2006	0.54.05	First Sample Volume 0.464L - No Flag Second Sample Volume 0.168L - Flagged Volume Not Met Third Sample Volume 0.582L - No Flag	Not A Valid Test
Hyland	Cara L	8/23/1986	DUI	80-000859	Cape Coral PD	Cape Coral PD	8/2/2006	2.37.55	First Sample Volume 1.035L - Flagged Volume Not Met Second Sample Volume 0.980L - No Flag Third Sample Volume 0.015L - Flagged Volume Not Met	Not A Valid Test
Velez	Ernesto	7/22/1966	DUI	80-000859	Cape Coral PD	Cape Coral PD	8/6/2006	1.33.11	First Sample Volume 0.097L - Flagged Volume Not Met Second Sample Volume 0.457L - No Flag Third Sample Volume 0.687L - Flagged Volume Not Met	Not A Valid Test
Jarrett	Donald	12/28/1955	DUI	80-000861	Sarasota County SO	Sarasota County SO	9/12/2006	21.54.31	First Sample Volume 0.035L - No Flag No Second Sample - Subject Refused	Not A Valid Test

Brown	Nathan	5/1/1979	DUI	80-000868	Monroe County SO	MA5 Key West	7/30/2006	2 11 35	First Sample Volume 0 250L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Morinere	Laurent L	6/24/1947	BUI	80-000869	Monroe County SO	Other State	4/15/2006	20 19 14	First Sample Volume 0 480L - No Flag Second Sample Volume 0 179L - Flagged Volume Not Met Third Sample Volume 1 804L - OK	Not A Valid Test
Clarke	Brian D	12/18/1961	BUI	80-000869	Monroe County SO	FWC	5/20/2006	21 23 30	First Sample Volume 0 687L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Cornejo	Jose	4/17/1980	DUI	80-000869	Monroe County SO	Monroe County SO	7/22/2006	4 48 56	First Sample Volume 1 078L - No Flag Second Sample Volume 1 246L - OK	Not A Valid Test
Kelly	Shane L	2/6/1981	DUI	80-000869	Monroe County SO	Monroe County SO	8/7/2006	5 16 12	First Sample Volume 0 449L - Flagged Volume Not Met Second Sample Volume 0 535L - No Flag Third Sample Volume 1 585L - OK	Not A Valid Test
La Paz	Gilda	11/20/1943	DUI	80-000871	Coral Gables PD	Coral Gables PD	6/4/2006	2 28 46	Second Sample Volume 0 149L - Flagged Volume Not Met Third Sample Volume 1 394L - OK	Not A Valid Test
Costa Moran	Enrique A	2/27/1975	DUI	80-000872	Miami PD	Miami PD	8/25/2006	3 03 40	First Sample Volume 1 578L - OK Second Sample Volume 0 328L - No Flag	Not A Valid Test
Mena	Hollman E	12/5/1984	DUI	80-000873	Miami PD	Miami PD	9/11/2006	9 26 14	First Sample Volume 0 808L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Mena	Hollman E	12/5/1984	DUI	80-000872	Miami PD	Miami PD	9/11/2006	9 52 59	First Sample Volume 1 078L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Diaz	Gerardo	7/30/1929	DUI	80-000873	Miami PD	Miami PD	4/1/2006	22 22 46	First Sample Volume 0 535L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Knowles	Patrick J	5/28/1966	DUI	80-000873	Miami PD	South Miami PD	9/16/2006	1 55 54	First Sample Volume 1 101L - OK Second Sample Volume 1 027L - No Flag	Not A Valid Test
Alonso	Ruben	11/22/1959	DUI	80-000876	Hialeah PD	Hialeah PD	5/1/2006	20 40 29	First Sample Volume 0 312L - No Flag Second Sample Volume 1 105L - OK	Not A Valid Test
Franco	Luis A	9/30/1967	DUI	80-000876	Hialeah PD	Hialeah PD	6/26/2006	21 54 09	First Sample Volume 0 718L - No Flag Second Sample Volume 1 246L - OK Third Sample Volume 1 148L - OK	Not A Valid Test
Bello	Farah T	10/14/1978	DUI	80-000879	Miami Beach PD	Miami Beach PD	9/16/2006	5 27 26	Second Control Test Outside Tolerance First Sample Volume 0 687L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Lozano	Cesar U	10/7/1975		80-000881	Miami Dade PD	Miami Dade PD	9/20/2006	2 09 52	First Sample Volume 0 304L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Martinez	Gilver O	9/2/1984	DUI	80-000884	Miami Dade PD	Miami Dade PD	4/9/2006	4 38 12	First Sample Volume 2 617L - OK Second Sample Volume 0 460L - No Flag	Not A Valid Test
Burgos-Donales	Myrna	8/9/1957	DUI	80-000885	Miami Dade PD	Miami Dade PD	5/6/2006	2 43 50	First Sample Volume 0 273L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Velasquez	Valerie A	3/1/1981	DUI	80-000885	Miami Dade PD	Miami Dade PD	6/10/2006	5 27 19	First Sample Volume 0 433L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Mena	Marianne	10/28/1983	DUI	80-000885	Miami Dade PD	Miami Dade PD	8/20/2006	6 53 05	First Sample Volume 0 117L - Flagged Volume Not Met Second Sample Volume 0 132L - Flagged Volume Not Met Third Sample Volume 0 976L - Flagged Volume Not Met	Not A Valid Test
Mena	Marianne	10/28/1983	DUI	80-000885	Miami Dade PD	Miami Dade PD	8/20/2006	7 13 49	First Sample Volume 0 117L - Flagged Volume Not Met Second Sample Volume 0 500L - No Flag Third Sample Volume 0 871L - Flagged Volume Not Met	Not A Valid Test
Winslon	Alan	4/17/1942	DUI	80-000886	North Miami Beach PD	North Miami Beach PD	4/22/2006	5 40 59	First Sample Volume 0 632L - No Flag Second Sample Volume 0 597L - No Flag	Not A Valid Test
Ghazarian	Sevak	9/9/1984	DUI	80-000888	Pinellas County SO	Pinellas County SO	6/2/2006	3 28 14	First Sample Volume 0 226L - No Flag Second Sample Volume 0 304L - Flagged Volume Not Met Third Sample Volume 0 101L - Flagged Volume Not Met	Not A Valid Test
Harvey	Raybeth	8/22/1959	DUI	80-000889	Pinellas County SO	Florida Highway Patrol	9/1/2006	21 20 50	First Sample Volume 1 203L - OK Second Sample Volume 0 089L - Flagged Volume Not Met Third Sample Volume 0 507L - No Flag	Not A Valid Test
Lee	Scott W	3/10/1973	DUI	80-000890	Pinellas County SO	Pinellas County SO	5/20/2006	4 31 38	First Sample Volume 0 777L - No Flag Second Sample Volume 1 074L - Flagged Volume Not Met Third Sample Volume 1 167L - OK	Not A Valid Test
Davis	Kevin M	5/20/1969	BUI	80-000892	Fish & Wildlife CC	FWC	5/20/2006	14 37 25	First Sample Volume 1 058L - No Flag No Second Sample - Improper Sample	Not A Valid Test
Lopez	Toribio M	10/5/1981	DUI	80-000907	Manatee County SO	Florida Highway Patrol	8/12/2006	0 51 18	First Sample Volume 1 441L - OK Second Sample Volume 0 554L - No Flag Third Sample Volume 2 554L - OK	Valid Test - First and Third Samples

Tamm	Trinity T.	6/28/1977	DUI	80-000912	Manatee County SO	Manatee County SO	4/19/2006	5.10.09	First Sample Volume 0.015L - Flagged Volume Not Met Second Sample Volume 0.082L - No Flag Third Sample Volume 0.296L - Flagged Volume Not Met	Not A Valid Test
Skoroo	Richard D	6/29/1962	DUI	80-000912	Manatee County SO	Florida Highway Patrol	8/28/2006	3.01.18	First Sample Volume 0.699L - No Flag Second Sample Volume 1.39L - OK	Not A Valid Test
Hurtado	Pablo E	3/22/1967	DUI	80-000939	Lee County SO	Lee County SO	4/27/2006	2.24.56	First Sample Volume 2.125L - OK Second Sample Volume 0.289L - Flagged Volume Not Met	Not A Valid Test
Fegno	Lillian	4/26/1962	DUI	80-000939	Lee County SO	Lee County SO	6/12/2006	15.27.43	First Sample Volume 0.347L - No Flag Second Sample Volume 1.496L - OK Third Sample Volume 1.496L - OK	Not A Valid Test
Moore	Shannon C	10/25/1982	DUI	80-000939	Lee County SO	Lee County SO	7/21/2006	0.18.42	First Sample Volume 0.375L - No Flag Second Sample Volume 1.871L - OK	Not A Valid Test
Conaway	Doreen M	11/15/1965	DUI	80-000939	Lee County SO	Lee County SO	7/23/2006	2.57.31	First Sample Volume 0.000L - Flagged Volume Not Met Second Sample Volume 0.000L - Flagged Volume Not Met Third Sample Volume - Subject Refused	Not A Valid Test Refusal
Daughtry	Amanda S	2/11/1972	DUI	80-000940	Brevard County SO	Brevard County SO	8/30/2006	3.25.19	First Sample Volume 1.605L - OK Second Sample Volume 0.230L - No Flag Third Sample Volume 0.757L - No Flag	Not A Valid Test
Cohen	Nikol A	6/6/1975	DUI	80-000943	Collier County SO	Naples PD	5/28/2006	3.00.50	First Sample Volume 0.613L - No Flag Second Sample Volume 1.816L - OK	Not A Valid Test
Lyach	Vladimir	2/27/1958	DUI	80-000943	Collier County SO	Collier County SO	8/7/2006	21.20.07	First Sample Volume 2.144L - OK Second Sample Volume 0.218L - No Flag Third Sample Volume 1.875L - OK	Not A Valid Test
Hall	Deborah L	11/25/1956	DUI	80-000949	Tallahassee PD	Tallahassee PD	7/26/2006	14.18.55	First Sample Volume 0.941L - No Flag Second Sample Volume 1.417L - OK	Not A Valid Test
Bravo-Neiro	Francisco	5/8/1978	DUI	80-000951	Hendry County SO	Hendry County SO	7/30/2006	2.44.40	First Sample Volume 0.164L - No Flag Second Sample Volume 2.507L - OK	Not A Valid Test
Adams	Rodney B	6/2/1961	DUI	80-000957	Leon County SO	Leon County SO	3/30/2006	19.14.15	First Sample Volume 1.316L - OK Second Sample Volume 0.410L - No Flag Third Sample Volume 1.570L - OK	Valid Test - First and Third Samples
Tucker	Leroy	10/2/1981	DUI	80-000958	Leon County SO	Tallahassee PD	4/15/2006	22.23.00	First Sample Volume 0.429L - No Flag Second Sample Volume 2.250L - OK	Not A Valid Test
Furman	Shawn A	10/21/1975	DUI	80-000958	Leon County SO	Tallahassee PD	9/9/2006	1.18.33	First Sample Volume 0.000L - Flagged Volume Not Met Second Sample Volume 2.382L - OK	Not A Valid Test
Childs	Kendra P	9/8/1969	DUI	80-000961	Wakulla County SO	Wakulla County SO	4/14/2006	23.27.48	First Sample Volume 0.675L - No Flag Second Sample Volume 0.015L - Flagged Volume Not Met Third Sample - Purge Fail	Not A Valid Test
Urostegui	Marcos	2/28/1982	DUI	80-000963	Orange County SO	Orlando PD	5/17/2006	4.51.48	First Sample Volume 1.453L - OK Second Sample Volume 0.183L - No Flag Third Sample Volume 1.453L - OK	Not A Valid Test
Altrabi	Fouad	4/2/1977	DUI	80-000963	Orange County SO	Orange County SO	5/29/2006	18.30.03	First Sample Volume 0.546L - No Flag Second Sample Volume 0.402L - Flagged Volume Not Met Third Sample Volume 3.941L - OK	Not A Valid Test
Miranda	Arnoldo	5/6/1983	DUI	80-000964	Orange County SO	Ocoee PD	9/24/2006	4.58.41	First Sample Volume 0.765L - No Flag Second Sample Volume 1.933L - OK	Not A Valid Test
Rodriguez	Candido J	8/8/1980	DUI	80-000966	Osceola County Corrections	Osceola County SO	4/22/2006	3.00.40	First Sample Volume 0.031L - No Flag Second Sample Volume 0.457L - No Flag Third Sample Volume 2.027L - OK	Not A Valid Test
Cannon	Heather M	3/3/1987	DUI	80-000999	Seminole County SO	Seminole County SO	9/22/2006	5.13.02	First Sample Volume 0.230L - No Flag Second Sample Volume 0.554L - Flagged Volume Not Met Third Sample Volume 0.554L - Flagged Volume Not Met	Not A Valid Test
Harazin	Linda S	8/24/1947	DUI	80-001003	Pinellas County SO	Pinellas Park PD	5/4/2006	4.47.35	First Sample Volume 0.394L - No Flag Second Sample Volume 0.429L - Flagged Volume Not Met Third Sample Volume 0.359L - Flagged Volume Not Met	Not A Valid Test
Chapman	Douglas R	12/25/1924	DUI	80-001004	Brevard County SO	Brevard County SO	6/25/2006	1.36.28	First Sample Volume 0.421L - No Flag Second Sample Volume 0.144L - Flagged Volume Not Met Third Sample Volume 1.250L - OK	Not A Valid Test
Irby	Howard J	10/7/1962	DUI	80-001005	Pinellas County SO	Pinellas County SO	4/11/2006	0.17.44	First Sample Volume 1.636L - OK Second Sample Volume 1.062L - No Flag	Not A Valid Test
Richmond	Aubrey L	2/9/1950	DUI	80-001005	Pinellas County SO	Pinellas County SO	4/15/2006	1.26.05	First Sample Volume 1.644L - OK Second Sample Volume 0.710L - No Flag	Not A Valid Test

Maggs	Robert F	9/29/1938	DUI	80-001005	Pinellas County SO	Pinellas County SO	7/26/2006	20 38 58	First Sample Volume 0 574L - No Flag Second Sample Volume 0 308L - Flagged Volume Not Met Third Sample Volume 0 261L - Flagged Volume Not Met	Not A Valid Test
Roundtree	Aaron C	8/22/1971	DUI	80-001005	Pinellas County SO	Pinellas County SO	8/20/2006	3 17 46	First Sample Volume 0 527L - No Flag Second Sample Volume 0 441L - Flagged Volume Not Met Third Sample Volume 1 152L - OK	Not A Valid Test
Smith	Heather L	10/10/1987	DUI	80-001009	Palm Bay PD	Palm Bay PD	8/28/2006	4 23 50	First Sample Volume 0 429L - No Flag Second Sample Volume 1 648L - OK Third Sample Volume 0 714L - Flagged Volume Not Met	Not A Valid Test
Thompson	Catherine S	2/18/1958	DUI	80-001009	Palm Bay PD	Palm Bay PD	9/11/2006	0 12 25	First Sample Volume 0 281L - No Flag Second Sample Volume 0 785L - Flagged Volume Not Met Third Sample Volume 0 109L - No Flag	Not A Valid Test
Zanchar	Roberto	1/30/1981	DUI	80-001043	Highlands County SO	Florida Highway Patrol	5/21/2006	0 31 19	No Second Sample - RFI First Sample Volume 1 878L - OK Second Sample Volume 0 687L - No Flag Third Sample Volume 1 402L - OK	Not A Valid Test
Cavender	Brenda A	9/23/1960	DUI	80-001043	Highlands County SO	Sebring PD	6/1/2006	1 56 15	First Sample Volume 0 382L - No Flag Second Sample Volume 1 640L - OK Third Sample Volume 1 753L - OK	Not A Valid Test
Marley	Lynne C	1/10/1959	DUI	80-001043	Highlands County SO	Florida Highway Patrol	6/10/2006	0 10 08	First Sample Volume 0 351L - No Flag Second Sample Volume 1 863L - OK Third Sample Volume 0 304L - No Flag	Not A Valid Test
Honorat	Hermane	1/9/1956	DUI	80-001044	Cocoon Creek PD	Cocoon Creek PD	6/26/2006	23 34 39	First Sample Volume 1 753L - OK Second Sample Volume 0 304L - No Flag Third Sample Volume 0 351L - No Flag	Not A Valid Test
Blair	Jason L	7/14/1975	DUI	80-001051	St. Petersburg PD	St. Petersburg PD	7/4/2006	19 14 37	No Third Sample - Subject Refused First Sample Volume 0 402L - No Flag Second Sample Volume 1 550L - OK Third Sample Volume 1 363L - OK	Valid Test - Second and Third Samples
Puett	Dennis R	12/4/1951	DUI	80-001051	St. Petersburg PD	St. Petersburg PD	8/18/2006	19 07 01	First Sample Volume 0 667L - No Flag Second Sample Volume 1 503L - OK Third Sample Volume 0 300L - No Flag	Not A Valid Test
Zuniga	Sherry C	7/11/1955	DUI	80-001055	Polk County SO	Florida Highway Patrol	7/29/2006	2 07 52	First Sample Volume 0 718L - Flagged Volume Not Met Second Sample Volume 1 125L - OK Third Sample Volume 1 105L - OK	Not A Valid Test
Addleman	Holly L	10/6/1977	DUI	80-001057	Lakeland PD	Lakeland PD	6/18/2006	3 23 44	First Sample Volume 0 699L - No Flag Second Sample Volume 1 656L - OK Third Sample Volume 0 109L - No Flag	Not A Valid Test
Rivera	Victor M	4/5/1950	DUI	80-001057	Lakeland PD	Lakeland PD	6/20/2006	3 12 02	First Sample Volume 0 109L - No Flag Second Sample Volume 1 488L - OK Third Sample Volume 0 738L - No Flag	Not A Valid Test
Godinez Perez	Alberto	1/11/1975	DUI	80-001063	Hollywood PD	Hollywood PD	7/1/2006	2 21 25	First Sample Volume 0 546L - Flagged Volume Not Met Second Sample Volume 0 738L - No Flag Third Sample Volume 1 105L - OK	Not A Valid Test
Buchanan	Bonne T	12/11/1940	DUI	80-001067	Broward County SO	Broward County SO	5/27/2006	19 03 55	First Sample Volume 0 718L - Flagged Volume Not Met Second Sample Volume 1 574L - OK Third Sample Volume 0 218L - No Flag	Not A Valid Test
Lanot	Jesufranc	1/11/1971	DUI	80-001067	Broward County SO	Broward County SO	7/1/2006	22 47 55	First Sample Volume 1 093L - No Flag Second Sample Volume 1 957L - OK Third Sample Volume 1 382L - OK	Not A Valid Test
Nelson	Stacy A	7/22/1963	DUI	80-001069	Jacksonville SO	Jacksonville SO	8/21/2006	4 30 22	First Sample Volume 0 703L - No Flag Second Sample Volume 0 671L - No Flag Third Sample Volume 0 949L - No Flag	Not A Valid Test
Doyle	Stephen V	7/24/1952	DUI	80-001069	Jacksonville SO	Jacksonville SO	9/5/2006	1 54 10	First Sample Volume 0 550L - No Flag Second Sample Volume 0 574L - Flagged Volume Not Met Third Sample Volume 1 457L - OK	Not A Valid Test
Buckley	Linda G	5/10/1967	DUI	80-001071	Clay County SO	Clay County SO	5/6/2006	2 11 12	First Sample Volume 0 671L - No Flag Second Sample Volume 0 574L - Flagged Volume Not Met Third Sample Volume 1 457L - OK	Not A Valid Test
Davis	Kim P	11/20/1957	DUI	80-001082	Plantation PD	Sunrise PD	5/9/2006	16 54 56	First Sample Volume 0 671L - No Flag Second Sample Volume 0 714L - Flagged Volume Not Met Third Sample Volume 0 179L - Flagged Volume Not Met	Not A Valid Test
Nguyen	Thai H	8/2/1963	DUI	80-001083	Jacksonville SO	Jacksonville SO	5/14/2006	22 56 41	First Sample Volume 0 949L - No Flag Second Sample Volume 0 714L - Flagged Volume Not Met Third Sample Volume 0 179L - Flagged Volume Not Met	Not A Valid Test
Gagne	Audra J	12/29/1966	DUI	80-001117	Pasco County SO	Pasco County SO	4/25/2006	2 54 49	No Second Sample - Purge Fail First Sample Volume 2 578L - OK Second Sample Volume 0 718L - No Flag Third Sample Volume 1 015L - No Flag	Not A Valid Test
Drake	Dawn D	6/9/1970	DUI	80-001117	Pasco County SO	Pasco County SO	4/27/2006	1 28 39	First Sample Volume 1 015L - No Flag Second Sample Volume 1 101L - OK Third Sample Volume 0 097L - Flagged Volume Not Met	Not A Valid Test
Brooks	Geoffrey D	4/20/1972	DUI	80-001131	Volusia County SO	Volusia County SO	7/4/2006	2 15 13	First Sample Volume 0 511L - Flagged Volume Not Met Second Sample Volume 1 261L - OK Third Sample Volume 0 972L - No Flag	Not A Valid Test
Brock	Lyn	7/20/1942	DUI	80-001137	Flagler County SO	Flagler County SO	7/22/2006	18 30 02	First Sample Volume 1 261L - OK Second Sample Volume 0 972L - No Flag Third Sample Volume 0 972L - No Flag	Not A Valid Test
Kenyon	John R	11/20/1979	DUI	80-001139	Broward County SO	Broward County SO	6/14/2006	1 48 17	First Sample Volume 0 574L - No Flag Second Sample Volume 0 308L - Flagged Volume Not Met Third Sample Volume 0 261L - Flagged Volume Not Met	Not A Valid Test



Jean Jacques	Patricio	5/10/1979	DUI	80-001139	Broward County SO	Broward County SO	7/8/2006	5:35:25	First Sample Volume 0 410L - No Flag Second Sample Volume 1 519L - OK Third Sample Volume 1 750L - OK	Valid Test - Second and Third Samples
Palacios	Miriam L	1/21/1978	DUI	80-001139	Broward County SO	Broward County SO	7/22/2006	3:46:52	First Sample Volume 0 414L - Flagged Volume Not Met Second Sample Volume 0 097L - Flagged Volume Not Met Third Sample Volume 1 406L - OK	Not A Valid Test
Hendry	Mark J	5/31/1960	DUI	80-001139	Broward County SO	Broward County SO	9/22/2006	1:11:06	First Sample Volume 0 531L - No Flag Second Sample Volume 0 421L - No Flag Third Sample Volume 0 191L - No Flag	Not A Valid Test
Barrera	Aristides E	8/21/1976	DUI	80-001153	Leon County SO	Leon County SO	4/13/2006	21:55:59	First Sample Volume 1 046L - No Flag Second Sample Volume 0 335L - Flagged Volume Not Met Third Sample Volume 1 199L - OK	Not A Valid Test
Gibson	Amelia C	7/30/1986	DUI	80-001153	Leon County SO	Leon County SO	8/23/2006	5:26:19	First Sample Volume 0 367L - No Flag Second Sample Volume 0 125L - Flagged Volume Not Met Third Sample Volume 1 199L - OK	Not A Valid Test
Waggoner	Nicholas C	5/4/1991	DUI	80-001158	Ponce Inlet PD	Daytona Beach PD	4/5/2006	23:12:41	First Sample Volume 2 066L - OK Second Sample Volume 2 066L - OK Third Sample Volume 1 281L - No Flag	Not A Valid Test
Currie II	Robert A	5/11/1955	DUI	80-001166	Brevard County SO	Brevard County SO	4/13/2006	20:11:17	First Sample Volume 1 058L - No Flag Second Sample Volume 0 828L - No Flag Third Sample Volume 2 519L - OK	Not A Valid Test
Zhang	Dong	6/24/1967	DUI	80-001190	Boynton Beach PD	Boynton Beach PD	6/17/2006	8:35:40	First Sample Volume 0 015L - Flagged Volume Not Met Second Sample Volume 2 066L - OK Third Sample Volume 0 726L - No Flag	Not A Valid Test
Ferranti	Edward J	5/26/1970	DUI	80-001190	Boynton Beach PD	Boynton Beach PD	7/22/2006	3:11:52	First Sample Volume 1 406L - OK Second Sample Volume 0 441L - No Flag Third Sample Volume 1 281L - No Flag	Not A Valid Test
Kaszak	Kenneth	4/16/1958	DUI	80-001190	Boynton Beach PD	Boynton Beach PD	8/13/2006	0:46:51	First Sample Volume 1 285L - OK Second Sample Volume 0 488L - No Flag Third Sample Volume 1 291L - OK	Not A Valid Test
Perez-Ortiz	Issac	10/23/1982	DUI	80-001235	West Palm Beach PD	West Palm Beach PD	4/8/2006	0:03:15	First Sample Volume 0 593L - No Flag Second Sample Volume 0 816L - No Flag Third Sample Volume 1 291L - OK	Not A Valid Test
Rucker	Curtis	3/12/1945	DUI	80-001248	Broward County SO	Broward County SO	5/7/2006	0:02:11	First Sample Volume 0 347L - No Flag Second Sample Volume 1 765L - OK Third Sample Volume 1 765L - OK	Not A Valid Test
Menner	Patricia	11/3/1968	DUI	80-001253	Brevard County SO	Brevard County SO	6/26/2006	22:40:39	First Sample Volume 0 468L - No Flag Second Sample Volume 1 785L - OK Third Sample Volume 0 644L - No Flag	Not A Valid Test
Brown	Joshua B	8/10/1976	DUI	80-001253	Brevard County SO	Brevard County SO	8/26/2006	23:45:57	First Sample Volume 0 644L - No Flag Second Sample Volume 1 242L - OK Third Sample Volume 1 242L - OK	Not A Valid Test
Doane	Richard G	10/26/1963	DUI	80-001255	Orange County SO	Orange County SO	4/7/2006	3:46:24	First Sample Volume 0 679L - No Flag Second Sample Volume 0 160L - Flagged Volume Not Met Third Sample Volume 1 628L - OK	Not A Valid Test
Apoite	Edgardo	10/9/1979	DUI	80-001255	Orange County SO	Orlando PD	9/9/2006	1:43:01	First Sample Volume 1 628L - OK Second Sample Volume 0 480L - No Flag Third Sample Volume 3 035L - OK	Not A Valid Test
Quiros	Ovalle	12/15/1978	DUI	80-001257	Orange County SO	Orlando PD	3/30/2006	4:45:18	First Sample Volume 0 746L - No Flag Second Sample Volume 1 128L - OK Third Sample Volume 1 351L - OK	Valid Test - Second and Third Samples
Meinken	Bradford G	11/3/1976	DUI	80-001257	Orange County SO	Orlando PD	8/5/2006	4:53:30	First Sample Volume 0 339L - No Flag Second Sample Volume 1 312L - OK Third Sample Volume 1 312L - OK	Not A Valid Test
Thomas	Jean C	7/21/1953	DUI	80-001258	Orange County SO	Florida Highway Patrol	5/6/2006	21:11:50	First Sample Volume 0 468L - No Flag Second Sample Volume 1 406L - OK Third Sample Volume 1 164L - OK	Not A Valid Test
McDaniel	Joseph N	10/6/1937	DUI	80-001260	Cocoa PD	Brevard County SO	7/15/2006	1:37:50	First Sample Volume 1 039L - No Flag Second Sample Volume 0 972L - No Flag Third Sample Volume 0 367L - No Flag	Not A Valid Test
Gause	Julia F	1/21/1969	DUI	80-001262	Melbourne PD	Melbourne PD	5/29/2006	22:12:15	First Sample Volume 0 296L - Flagged Volume Not Met Second Sample Volume 0 519L - No Flag Third Sample Volume 0 097L - Flagged Volume Not Met	Not A Valid Test
Young	Pamela	8/5/1958	DUI	80-001267	Palm Beach County So	Riveria Beach PD	5/30/2006	1:40:07	First Sample Volume 1 363L - OK Second Sample Volume 0 972L - No Flag Third Sample Volume 0 367L - No Flag	Not A Valid Test
Keenan	Maureen E	10/9/1970	DUI	80-001267	Palm Beach County So	Palm Beach County SO	7/19/2006	1:31:08	First Sample Volume 1 226L - OK Second Sample Volume 0 714L - No Flag Third Sample Volume 1 769L - OK	Not A Valid Test
Villaloro	Yener R	8/24/1984	DUI	80-001267	Palm Beach County So	Palm Beach County SO	8/6/2006	4:53:20	First Sample Volume 1 445L - OK Second Sample Volume 0 609L - No Flag Third Sample Volume 1 437L - OK	Valid Test - First and Third Samples
Suarez	Jean	3/31/1969	DUI	80-001269	Palm Beach County So	Florida Highway Patrol	5/5/2006	6:55:45	First Sample Volume 1 445L - OK Second Sample Volume 0 609L - No Flag Third Sample Volume 1 437L - OK	Valid Test - First and Third Samples
Goldberg	Paul E	3/18/1969	DUI	80-001269	Palm Beach County So	Florida Highway Patrol	8/24/2006	3:44:46	First Sample Volume 0 410L - No Flag Second Sample Volume 1 519L - OK Third Sample Volume 1 750L - OK	Valid Test - Second and Third Samples

Toft	Carole A	8/6/1964	DUI	80-001269	Palm Beach County So	Palm Beach County SO	9/21/2006	4 12 02	First Sample Volume 0 253L - No Flag Second Sample Volume 1 207L - OK	Not A Valid Test
Cornell	Daved E	2/2/1965	DUI	80-001270	Palm Beach County So	Delray Beach PD	7/23/2006	22 32 26	First Sample Volume 0 613L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Harris	Kareen R	6/3/1968	DUI	80-001272	Seminole County SO	Oviedo PD	4/2/2006	5 47 26	First Sample Volume 0 867L - No Flag Second Sample Volume 1 703L - OK	Not A Valid Test
Wells	Barry N	9/7/1968	DUI	80-001272	Seminole County SO	Lake Mary PD	4/10/2006	19 11 26	First Sample Volume 0 710L - No Flag Second Sample Volume 0 332L - Flagged Volume Not Met Third Sample Volume 0 785L - Flagged Volume Not Met	Not A Valid Test
Wilson	Glen M	9/15/1964	DUI	80-001272	Seminole County SO	Florida Highway Patrol	6/25/2006	2 12 47	First Sample Volume 0 675L - No Flag Second Sample Volume 0 796L - OK Second Control Test Outside Tolerance	Not A Valid Test
White	Leslie J	6/6/1960	DUI	80-001272	Seminole County SO	Oviedo PD	7/2/2006	2 46 00	Second Sample Volume 0 257L - Flagged Volume Not Met Third Sample Volume 0 679L - No Flag Second Control Test Outside Tolerance	Not A Valid Test
Newton	Justin B	10/21/1978	DUI	80-001272	Seminole County SO	Seminole County SO	7/21/2006	22 17 44	First Sample Volume 0 457L - No Flag Second Sample Volume 1 390L - OK	Not A Valid Test
Lopez-Hernandez	David	4/4/1978	DUI	80-001273	Pinellas County SO	Pinellas County SO	7/15/2006	3 03 14	First Sample Volume 0 050L - No Flag Second Sample Volume 1 414L - OK Third Sample Volume 1 613L - OK	Valid Test - Second and Third Samples
Portonondo	Enk J	12/11/1975	DUI	80-001276	Miami Dade PD	Miami Dade PD	5/12/2006	0 34 06	First Sample Volume 0 109L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Fandino	Marta L	6/18/1967	DUI	80-001276	Miami Dade PD	Miami Dade PD	5/27/2006	4 12 02	First Sample Volume 0 453L - No Flag Second Sample Volume 1 839L - OK Third Sample Volume - 1 292L - OK	Valid Test - Second and Third Samples
Fuhr	Alison R	10/31/1978	DUI	80-001282	University of Florida PD	University of FL PD	4/13/2006	3 34 42	First Sample Volume 0 867L - No Flag Second Sample Volume 0 152L - Flagged Volume Not Met Third Sample Volume 0 082L - Flagged Volume Not Met	Not A Valid Test
Johns	Cynthia J	5/18/1966	DUI	80-001284	Gilchrist County SO	Cedar Keys PD	6/15/2006	1 38 28	First Sample Volume 0 421L - No Flag Second Sample Volume 0 378L - Flagged Volume Not Met Third Sample Volume 0 816L - No Flag	Not A Valid Test
Campbell	Steven M	7/12/1974	DUI	80-001303	Santa Rosa County SO	Santa Rosa County SO	8/6/2006	20 30 09	First Sample Volume 0 289L - Flagged Volume Not Met Second Sample Volume 0 789L - No Flag Third Sample Volume 1 687L - OK	Not A Valid Test
Phillips	Darius K	4/9/1984	OTHER	80-001309	Bay County SO	Tyndall AFB	4/9/2006	13 22 01	First Sample Volume 0 628L - No Flag Second Sample Volume 0 578L - Flagged Volume Not Met No Third Sample - Subject Refused	Not A Valid Test
Cardwell	Curtis B	11/24/1981	DUI	80-001309	Bay County SO	Bay County SO	7/23/2006	23 35 34	First Sample Volume 0 511L - No Flag Second Sample Volume 0 000L - Flagged Volume Not Met Third Sample Volume 0 437L - Flagged Volume Not Met	Not A Valid Test
Miller	Chadrick L	4/24/1978	DUI	80-001309	Bay County SO	Parker PD	9/16/2006	22 34 35	First Sample Volume 0 437L - No Flag No Second Sample - Subject Refused	Not A Valid Test
Mings	Lonnie L	6/29/1977	DUI	80-001309	Bay County SO	Florida Highway Patrol	9/23/2006	3 41 36	First Sample Volume 0 097L - No Flag Second Sample Volume 1 925L - OK Third Sample Volume 1 925L - OK	Valid Test - First and Third Samples
Castillo-Vazquez	Jose	12/6/1960	DUI	80-001309	Bay County SO	Florida Highway Patrol	9/24/2006	1 50 55	First Sample Volume 0 906L - No Flag Second Sample Volume 2 613L - OK	Not A Valid Test
Clay	Susan S	12/9/1953	DUI	80-001310	Niceville PD	Crestview PD	6/12/2006	0 08 55	First Sample Volume 1 246L - OK Second Sample Volume 0 558L - No Flag Third Sample Volume 1 394L - OK	Valid Test - First and Third Samples
Diamond	Elnede F	11/1/1936	DUI	80-001328	Indian River County SO	Sebastian PD	8/30/2006	22 47 27	First Sample Volume 0 554L - No Flag Second Sample Volume 1 003L - No Flag First Sample Volume 1 972L - OK	Not A Valid Test
Angus	Anthony	6/1/1959	DUI	80-001328	Indian River County SO	Florida Highway Patrol	9/1/2006	8 34 49	Second Sample Volume 0 156L - Flagged Volume Not Met Third Sample Volume 0 882L - No Flag	Not A Valid Test
Kellowan	Shewon S	3/20/1979	DUI	80-001344	Sarasota County SO	Florida Highway Patrol	4/1/2006	4 55 54	First Sample Volume 0 093L - Flagged Volume Not Met Second Sample Volume 3 066L - OK Third Sample Volume 0 714L - No Flag	Not A Valid Test
Ryll	Frank L	10/27/1951	DUI	80-001344	Sarasota County SO	Sarasota County SO	4/25/2006	22 40 18	First Sample Volume 0 937L - No Flag Second Sample Volume 1 972L - OK	Not A Valid Test
Kerr	Julia P	2/11/1987	DUI	80-001344	Sarasota County SO	Sarasota County SO	5/3/2006	4 22 57	First Sample Volume 0 468L - No Flag Second Sample Volume 1 871L - OK Third Sample Volume 1 691L - OK	Not A Valid Test

Shapiro	Wayne G	3/25/1959	DUI	80-001344	Sarasota County SO	Sarasota PD	5/28/2006	3.58 0.11	First Sample Volume 0.656L - No Flag Second Sample Volume 1.683L - OK Third Sample Volume 1.800L - OK	Valid Test - Second and Third Samples
Quattrone	Nancy E	4/9/1957	DUI	80-001344	Sarasota County SO	Sarasota County SO	6/7/2006	0.27 3.0	First Sample Volume 0.324L - No Flag Second Sample Volume 1.921L - OK	Not A Valid Test
Velle	Victoreno	10/27/1979	DUI	80-001344	Sarasota County SO	Sarasota County SO	6/18/2006	3.26 4.0	First Sample Volume 1.625L - OK Second Sample Volume 0.636L - No Flag Third Sample Volume 2.160L - OK	Valid Test - First and Third Samples
Palmer	Scott D	7/13/1964	DUI	80-001344	Sarasota County SO	Sarasota PD	6/26/2006	1.53 5.0	First Sample Volume 1.101L - OK Second Sample Volume 0.804L - No Flag No Third Sample - Subject Refused	Not A Valid Test Refusal
Delahanty	Thomas L	5/15/1948	DUI	80-001344	Sarasota County SO	Sarasota County SO	7/8/2006	0.30 5.0	First Sample Volume 0.492L - No Flag Second Sample Volume 0.191L - Flagged Volume Not Met Third Sample Volume 0.378L - No Flag	Not A Valid Test
Germain	David M	4/8/1957	DUI	80-001367	Pinellas County SO	Florida Highway Patrol	7/16/2006	18.23 4.5	First Sample Volume 0.906L - No Flag Second Sample Volume 1.125L - OK Third Sample Volume 1.371L - OK	Valid Test - Second and Third Samples
Desimas	John R	2/1/1984	DUI	80-001415	Pinellas County SO	Pinellas Park PD	8/5/2006	3.02 0.7	First Sample Volume 0.130L - No Flag Second Sample Volume 0.593L - Flagged Volume Not Met Third Sample Volume 0.144L - Flagged Volume Not Met	Not A Valid Test
Dilmone	Linda J	3/13/1962	DUI	80-001416	Orange County SO	Orange County SO	6/6/2006	12.48 1.4	First Sample Volume 0.835L - No Flag Second Sample Volume 0.437L - Flagged Volume Not Met Third Sample Volume 0.070L - Flagged Volume Not Met	Not A Valid Test
Martinez	Noel I	9/22/1968	DUI	80-001416	Orange County SO	Orange County SO	6/10/2006	4.17 2.8	First Sample Volume 0.937L - No Flag Second Sample Volume 1.718L - OK	Not A Valid Test
Kars	Chad D	1/4/1967	DUI	80-001416	Orange County SO	Orlando PD	7/10/2006	1.34 3.9	First Sample Volume 1.308L - OK Second Sample Volume 0.792L - No Flag First Sample Volume 0.042L - Flagged Volume Not Met	Not A Valid Test
Reyes	Arturo	9/11/1965	DUI	80-001418	Orange County SO	Florida Highway Patrol	8/30/2006	1.34 2.4	First Sample Volume 0.804L - No Flag Second Sample Volume 2.421L - OK Third Sample Volume 0.703L - No Flag	Not A Valid Test
Poole	Nikki K	7/17/1975	DUI	80-001418	Orange County SO	Maitland PD	8/30/2006	3.38 2.0	First Sample Volume 0.261L - Flagged Volume Not Met Second Sample Volume 1.531L - OK Third Sample Volume 2.972L - OK	Not A Valid Test
Stephens	Patrick J	11/20/1984	DUI	80-001420	Orange County SO	Orange County SO	6/11/2006	4.28 5.4	First Sample Volume 0.394L - No Flag Second Sample Volume 1.789L - OK Third Sample Volume 1.464L - OK	Not A Valid Test
Buckley	Linda G	5/10/1967	DUI	80-001443	Clay County SO	Clay County SO	5/6/2006	1.46 1.8	First Sample Volume 0.648L - No Flag Second Control Test Outside Tolerance	Not A Valid Test

\*\* Notes are concerning the listed breath tests. Additional breath tests of the referenced subject may have been conducted and are not a part of this list.



Florida Department of  
Law Enforcement

Gerald M. Bailey  
*Interim Commissioner*

**Criminal Justice  
Professionalism Program  
Alcohol Testing Program**

P.O. Box 1489  
Tallahassee, Florida 32302  
(850) 410-8600  
(850) 410-7816 Fax  
<http://www.fdle.state.fl.us>

October 5, 2006

NAME  
ADDRESS  
ADDRESS  
ADDRESS

SUBJECT: Intoxilyzer 8000 Software Version 8100.26

Dear NAME:

This letter is to inform you of an issue regarding the Intoxilyzer 8000 that the Florida Department of Law Enforcement Alcohol Testing Program has discovered. We have been in contact with the manufacturer of the Intoxilyzer 8000, CMI, Inc., and have determined there are missing instructions in the Intoxilyzer 8000 software version 8100.26 dealing with the instrument's ability to correctly identify certain breath samples that do not meet a minimum volume of 1.1 Liters.

The Intoxilyzer 8000 uses the minimum breath sample requirements of time, volume and slope to ensure the breath sample is reliable. A subject has up to three (3) minutes to provide a breath sample. When the subject provides their breath sample within the three (3) minutes, the Intoxilyzer 8000 with software version 8100.26 is correctly identifying breath sample volumes of less than 1.1 Liters as "Volume Not Met". The instrument will either request subsequent breath samples or discontinue the breath testing process.

We have isolated the issue to a small number of breath samples where the subject provides a breath sample during and continues past the three (3) minute mark. The instrument accepts the breath sample and posts a result; however, the software is not instructed to verify the breath sample volume. Therefore, there is no "Volume Not Met" flag if the breath sample volume is less than 1.1 Liters. An example of this scenario is as follows: the subject begins providing a breath sample at 2.58 minutes and continues providing the breath sample until 3.02 minutes. The result is displayed. If the breath volume for the sample is less than 1.1 Liters, there will be no "Volume Not Met" flag.

**I would like to assure you that software version 8100.26 is operating as written and programmed. The analytical functions of the instrument are not affected by this issue and the instrument continues to correctly analyze all breath samples.** It is only with this unique and specific scenario that the breath sample has not been determined to be reliable; therefore the instrument should not report a result. A software revision to add these missing instructions is in the process of being completed and will be installed in your Intoxilyzer 8000 instruments in the near future.

A letter regarding this issue has been forwarded under separate cover to your Agency Head. Enclosed is a summary of the affected breath tests determined as of October 2, 2006. The Alcohol Testing Program is continuing to closely monitor every breath test result obtained in the State of Florida and will contact you regarding all affected breath tests. If you have any questions, please feel free to contact me at (850) 410-7810.

Sincerely,

Gerald M. Bailey  
Interim Commissioner

Laura D. Barfield  
Alcohol Testing Program Manager

LDB/lb

Enclosure

*Committed to  
Service - Integrity - Respect - Quality*

Subject Last Name	Subject First Name	Subject's Date of Birth	Offense	Instrument Serial Number	Instrument Registered To	Arresting Officer Agency	Date of Test	Time of Test	Breath Sample Volume Results	Notes
Corredor	Miguel A	1/7/1986	DUI	80-000223	FDLE (Loaned to Indian River County Sheriff's Office)	Florida Highway Patrol	7/6/2006	22 58 02	First Sample Volume 1.931 L - OK Second Sample Volume 0.972L - No Flag Third Sample Volume 0.226L - Flagged Volume Not Met	Not A Valid Test
Sedovic	Michael T	11/4/1976	DUI	80-000738	Fort Walton Beach PD	Okaloosa County SO	8/5/2006	4 54 39	No Second Sample - Purge Fail First Sample Volume 1.066L - No Flag Second Sample Volume 1.164L - OK	Not A Valid Test
Pelujola	Temilola A	12/24/1961	DUI	80-000739	Okaloosa County SO	Florida Highway Patrol	8/21/2006	0 10 12	First Sample Volume 1.550L - OK Second Sample Volume 0.605L - No Flag Third Sample Volume 1.164L - OK	Not A Valid Test
Fulkord	Kenneth M	7/7/1951	DUI	80-000744	Escambia County SO	Pensacola PD	5/17/2006	2 04 10	Second Sample Volume 0.664L - No Flag First Sample Volume 0.664L - No Flag Third Sample Volume 2.105L - OK	Not A Valid Test
Burgos	Abel	7/1/1978	DUI	80-000744	Escambia County SO	Florida Highway Patrol	7/22/2006	3 58 14	First Sample Volume 1.082L - No Flag Second Sample Volume 1.082L - No Flag Third Sample Volume 1.082L - No Flag	Not A Valid Test
East	Easter K	3/21/1945	DUI	80-000744	Escambia County SO	Florida Highway Patrol	8/12/2004	2 04 13	Second Sample - Subject Refused First Sample Volume 0.601L - No Flag Second Sample Volume 1.515L - OK	Not A Valid Test
Rives	Lisa C	6/8/1959	DUI	80-000744	Escambia County SO	Florida Highway Patrol	9/28/2006	21 07 00	First Sample Volume 0.062L - No Flag Second Sample - Purge Fail No Second Sample - Purge Fail	Not A Valid Test
Swanson	Amanda N	9/9/1984	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	5/18/2006	5 39 00	First Sample Volume 0.129L - Flagged Volume Not Met Second Sample Volume 0.347L - No Flag Third Sample Volume 0.127L - Flagged Volume Not Met	Not A Valid Test
Hidle	Brandi C.	1/9/1972	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	5/25/2006	1 05 16	First Sample Volume 1.000L - No Flag Second Sample Volume 1.000L - No Flag Third Sample Volume 1.500L - OK	Not A Valid Test
Toney	Patrick S.	3/10/1987	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	5/30/2006	1 29 54	First Sample Volume 0.265L - No Flag Second Sample Volume 0.054L - Flagged Volume Not Met Third Sample Volume 1.398L - OK	Not A Valid Test
Boone	Michael J	11/9/1983	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	6/25/2006	2 30 07	First Sample Volume 0.742L - No Flag Second Sample Volume 0.851L - Flag Volume Not Met Third Sample Volume 0.214L - Flag Volume Not Met	Not A Valid Test
Alicandr	Danielle M	8/1/1971	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	7/25/2006	2 36 03	First Sample Volume 0.476L - No Flag Second Sample Volume 1.738L - OK Third Sample Volume 0.652L - No Flag	Not A Valid Test
Jones	Elaime J.	9/18/1969	DUI	80-000772	Santa Rosa County SO	Santa Rosa County SO	5/13/2006	2 38 47	First Sample Volume 0.652L - No Flag Second Sample Volume 0.652L - No Flag Third Sample Volume 0.652L - No Flag	Not A Valid Test
Bohannon	Lynette D	12/18/1972	DUI	80-000786	Broward County SO	Broward County SO	5/12/2006	0 10 42	No Second Sample - Purge Fail First Sample Volume 0.324L - Flag Volume Not Met Second Sample Volume 0.632L - No Flag Third Sample Volume 0.761L - Flag Volume Not Met	Not A Valid Test
Shelton	Linda M.	8/27/1940	DUI	80-000790	Vero Beach PD	Vero Beach PD	7/13/2006	3 05 22	First Sample Volume 0.851L - Flag Volume Not Met Second Sample Volume 1.292L - OK Third Sample Volume 0.613L - No Flag	Not A Valid Test
Vaughl	Michael A	4/20/1981	DUI	80-000794	St. Lucie County SO	St. Lucie County SO	4/9/2006	2 49 38	First Sample Volume 1.527L - OK Second Sample Volume 0.261L - No Flag Third Sample Volume 1.371L - OK	Not A Valid Test
Lesko	Darryl R	5/10/1958	DUI	80-000794	St. Lucie County SO	St. Lucie County SO	6/5/2006	1 49 07	First Sample Volume 0.210L - No Flag Second Control Test Outside Tolerance Third Sample Volume 1.222L - OK	Not A Valid Test
Cortez	Javier O	1/14/1964	DUI	80-000794	St. Lucie County SO	Fort Pierce PD	6/10/2006	0 36 52	First Sample Volume 1.570L - OK Second Sample Volume 0.300L - No Flag Third Sample Volume 2.429L - OK	Not A Valid Test
Schender	Robert P	4/23/1962	DUI	80-000794	St. Lucie County SO	Fort Pierce PD	6/19/2006	20 12 44	First Sample Volume 1.113L - OK Second Sample Volume 0.234L - No Flag Third Sample Volume 0.515L - No Flag	Not A Valid Test
Slater	Keith	6/24/1972	DUI	80-000794	St. Lucie County SO	Fort Pierce PD	6/21/2006	9 25 13	Second Control Test Outside Tolerance First Sample Volume 0.515L - No Flag Second Sample Volume 1.207L - OK Third Sample Volume 0.812L - No Flag	Not A Valid Test
Maliszewski	Dagmar E	8/5/1967	DUI	80-000794	St. Lucie County SO	Port St. Lucie PD	6/25/2006	23 16 05	First Sample Volume 0.812L - No Flag Second Control Test Outside Tolerance First Sample Volume 1.207L - OK Second Sample Volume 0.812L - No Flag	Not A Valid Test
Diaz	Alejandro	10/29/1984	DUI	80-000794	St. Lucie County SO	St. Lucie County SO	8/19/2006	3 20 08	First Sample Volume 0.816L - No Flag Second Sample - Subject Refused Third Sample Volume 0.816L - No Flag	Not A Valid Test
Jimenez	David Y	1/24/1977	DUI	80-000808	Hillsborough County SO	Hillsborough County SO	3/30/2006	23 07 03	First Sample Volume 1.054L - No Flag Second Sample Volume 1.261L - OK	Not A Valid Test
Pileasa	Petar J	11/25/1982	DUI	80-000808	Hillsborough County SO	Tampa PD	7/2/2006	3 58 14	First Sample Volume 1.261L - OK Second Sample Volume 1.261L - OK	Not A Valid Test

Roberts	Pamela R	7/23/1984	DUI	80-000808	Hillsborough County SO	Tampa PD	8/24/2006	0.30.17	First Sample Volume 1.070L - No Flag Second Sample Volume 1.679L - OK First Sample Volume 0.363L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Reyes	Tony	6/11/1984	DUI	80-000808	Hillsborough County SO	Plant City PD	9/3/2006	0.04.38	First Sample Volume 0.246L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Clark	Michael C	4/9/1983	DUI	80-000808	Hillsborough County SO	Florida Highway Patrol	9/9/2006	5.52.04	First Sample Volume 0.925L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Hargrave	Evelyn M	7/20/1979	DUI	80-000808	Hillsborough County SO	Tampa PD	9/10/2006	3.08.59	First Sample Volume 2.738L - OK Second Sample Volume 0.476L - No Flag	Not A Valid Test
Diaz-Gonzalez	Raul N	11/24/1957	DUI	80-000813	Marion County SO	Marion County SO	6/21/2006	14.35.11	Purge Fail After Second Sample First Sample Volume 0.398L - No Flag Second Sample Volume 0.179L - No Flag	Not A Valid Test
Merritt	Brady J	9/9/1984	DUI	80-000817	Citrus County SO	Citrus County SO	9/14/2006	4.23.44	First Sample Volume 0.324L - No Flag Second Sample Volume 0.179L - No Flag	Not A Valid Test
Welch	Andre C	3/20/1966	DUI	80-000824	Leesburg PD	Leesburg PD	4/15/2006	0.19.50	First Sample Volume 1.117L - OK Second Sample Volume 0.339L - No Flag	Not A Valid Test
Garcia	Alfonso R	8/17/1980	DUI	80-000829	Hillsborough County SO	Tampa PD	4/12/2006	2.16.24	No Second Sample - Purge Fail First Sample Volume 1.660L - OK Second Sample Volume 0.972L - No Flag	Not A Valid Test
Nilsen	Michael R	10/9/1961	DUI	80-000832	Martin County SO	Florida Highway Patrol	5/29/2006	3.47.51	Third Sample Volume 1.539L - OK First Sample Volume 0.597L - No Flag Second Sample Volume 0.204L - Flagged Volume Not Met	Valid Test - First and Third Samples
Camacho	Rafael	4/3/1961	DUI	80-000832	Martin County SO	Stuart PD	6/22/2006	19.54.57	First Sample Volume 0.000L - Flagged Volume Not Met Second Sample Volume 0.219L - Flagged Volume Not Met Third Sample Volume 2.226L - OK	Not A Valid Test
Mendoza	Cecilio	11/22/1962	DUI	80-000832	Martin County SO	Martin County SO	7/9/2006	2.12.46	First Sample Volume 0.183 - No Flag Second Sample Volume 0.769L - No Flag Third Sample Volume 1.480L - OK	Not A Valid Test
Schroeder	Brandon H	7/28/1987	DUI	80-000832	Martin County SO	Martin County SO	8/13/2006	2.45.47	First Sample Volume 1.726L - OK Second Sample Volume 0.449L - No Flag	Not A Valid Test
Daher	Jabib	2/3/1948	DUI	80-000832	Martin County SO	FWC West Palm	8/26/2006	23.30.17	First Sample Volume 0.808L - No Flag Second Sample Volume 0.203L - OK Third Sample Volume 1.691L - OK	Not A Valid Test
Vandermark	Jeffrey W	6/16/1945	DUI	80-000832	Martin County SO	Sewall's Point PD	9/4/2006	2.04.32	Third Sample Volume 1.187L - OK Second Control Test Outside Tolerance	Not A Valid Test
Michale	James M	2/27/1967	DUI	80-000832	Martin County SO	Stuart PD	9/4/2006	21.48.28	First Sample Volume 0.742L - No Flag Second Sample Volume 1.96L - Flag Volume Not Met No Third Sample - Subject Refused	Not A Valid Test
Moretti	Derek P	6/2/1980	DUI	80-000833	Hillsborough County SO	Tampa PD	5/15/2006	1:09.00	First Sample Volume 0.105L - No Flag Second Sample Volume 0.212L - Flagged Volume Not Met Third Sample Volume 0.219L - Flagged Volume Not Met	Not A Valid Test
Perez	Palermo	11/20/1972	DUI	80-000833	Hillsborough County SO	Hillsborough County SO	6/10/2006	19.51.30	First Sample Volume 0.957L - No Flag Second Sample Volume 1.648L - OK	Not A Valid Test
Pospischl	Ashley N	5/10/1983	DUI	80-000833	Hillsborough County SO	Tampa PD	7/3/2006	4.13.57	First Sample Volume 0.371L - No Flag Second Sample - Subject Refused	Not A Valid Test
Gandanilla	Petronilo R	9/6/1972	DUI	80-000833	Hillsborough County SO	Plant City PD	7/21/2006	9.12.20	First Sample Volume 0.625L - No Flag Second Sample - Subject Refused	Not A Valid Test
Rojas	Oscar	8/27/1949	DUI	80-000833	Hillsborough County SO	Tampa PD	8/11/2006	2.46.47	First Sample Volume 0.574L - No Flag Second Sample Volume 1.289L - OK	Not A Valid Test
Rolon	Jose A	3/8/1961	DUI	80-000833	Hillsborough County SO	Florida Highway Patrol	8/13/2006	4.32.45	First Sample Volume 0.429L - No Flag Second Sample Volume 0.194L - Flagged Volume Not Met Third Sample Volume 0.625L - No Flag	Not A Valid Test
Seymour	Patrick H	8/26/1967	DUI	80-000859	Cape Coral PD	Cape Coral PD	6/21/2006	20.35.16	First Sample Volume 0.464L - No Flag Second Sample Volume 0.168L - Flagged Volume Not Met Third Sample Volume 0.582L - No Flag	Not A Valid Test
Lisa	Denise A	12/9/1957	DUI	80-000859	Cape Coral PD	Cape Coral PD	7/10/2006	0.54.05	First Sample Volume 1.035L - Flagged Volume Not Met Second Sample Volume 0.980L - No Flag Third Sample Volume 0.015L - Flagged Volume Not Met	Not A Valid Test
Hyland	Cara L	8/23/1986	DUI	80-000859	Cape Coral PD	Cape Coral PD	8/2/2006	2.37.55	First Sample Volume 0.097L - Flagged Volume Not Met Second Sample Volume 0.457L - No Flag Third Sample Volume 0.687L - Flagged Volume Not Met	Not A Valid Test
Veletz	Ernesto	7/22/1966	DUI	80-000859	Cape Coral PD	Cape Coral PD	8/6/2006	1.33.11	Second Control Test Outside Tolerance First Sample Volume 0.035L - No Flag No Second Sample - Subject Refused	Not A Valid Test
Jarrett	Donald	12/28/1955	DUI	80-000861	Sarasota County SO	Sarasota County SO	9/12/2006	21.54.31		Refusal

Brown	Nathan	5/1/1979	DUI	80-000868	Monroe County SO	NAS Key West	7/30/2006	2 11 35	First Sample Volume 0 250L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Morniere	Laurent L	6/24/1947	BUI	80-000869	Monroe County SO	Other State	4/15/2006	20 19 14	First Sample Volume 0 480L - No Flag Second Sample Volume 0 179L - Flagged Volume Not Met Third Sample Volume 1 804L - OK	Not A Valid Test
Clarke	Brian D	12/18/1961	BUI	80-000869	Monroe County SO	FWC	5/20/2006	21 23 30	First Sample Volume 0 687L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Cornejo	Jose	4/17/1980	DUI	80-000869	Monroe County SO	Monroe County SO	7/22/2006	4 48 56	First Sample Volume 0 449L - Flagged Volume Not Met Second Sample Volume 0 535L - No Flag Third Sample Volume 1 585L - OK	Not A Valid Test
Kelly	Shane L	2/6/1981	DUI	80-000869	Monroe County SO	Monroe County SO	8/7/2006	5 16 12	Second Sample Volume 0 149L - Flagged Volume Not Met Third Sample Volume 1 394L - OK	Not A Valid Test
La Paz	Guida	11/20/1943	DUI	80-000871	Coral Gables PD	Coral Gables PD	6/4/2006	2 28 46	First Sample Volume 1 578L - OK Second Control Test Outside Tolerance	Not A Valid Test
Costa Moran	Enrique A	2/27/1975	DUI	80-000872	Miami PD	Miami PD	8/25/2006	3 03 40	First Sample Volume 0 328L - No Flag Second Sample Volume 0 808L - No Flag	Not A Valid Test
Mena	Hollman E.	12/5/1984	DUI	80-000873	Miami PD	Miami PD	9/11/2006	9 26 14	First Sample Volume - Purge Fail No Second Sample - Purge Fail	Not A Valid Test
Mena	Hollman E.	12/5/1984	DUI	80-000872	Miami PD	Miami PD	9/11/2006	9 52 59	First Sample Volume 1 078L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Diaz	Gerardo	7/30/1929	DUI	80-000873	Miami PD	Miami PD	4/1/2006	22 22 46	First Sample Volume 1 101L - OK Second Sample Volume 0 312L - No Flag	Not A Valid Test
Knowles	Patrick J	5/28/1966	DUI	80-000873	Miami PD	South Miami PD	9/16/2006	1 55 54	First Sample Volume 1 101L - OK Second Sample Volume 0 312L - No Flag	Not A Valid Test
Alonso	Ruben	11/22/1959	DUI	80-000876	Hialeah PD	Hialeah PD	5/1/2006	20 40 29	First Sample Volume 0 718L - No Flag Second Sample Volume 1 148L - OK Third Sample Volume 1 246L - OK	Not A Valid Test
Franco	Luis A.	9/30/1967	DUI	80-000876	Hialeah PD	Hialeah PD	6/26/2006	21 54 09	First Sample Volume 0 687L - No Flag Second Control Test Outside Tolerance	Not A Valid Test
Bello	Farah T	10/14/1978	DUI	80-000879	Miami Beach PD	Miami Beach PD	9/16/2006	5 27 26	First Sample Volume 0 304L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Lozano	Cesar U	10/7/1975		80-000881	Miami Dade PD	Miami Dade PD	9/20/2006	2 09 52	First Sample Volume 2 617L - OK No Second Sample - Purge Fail	Not A Valid Test
Martinez	Giver O	9/2/1984	DUI	80-000884	Miami Dade PD	Miami Dade PD	4/9/2006	4 38 12	First Sample Volume 0 273L - No Flag Second Sample Volume 0 460L - No Flag	Not A Valid Test
Burgos-Donales	Myrna	8/9/1957	DUI	80-000885	Miami Dade PD	Miami Dade PD	5/6/2006	2 43 50	First Sample Volume 0 433L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Velasquez	Valerie A.	3/1/1981	DUI	80-000885	Miami Dade PD	Miami Dade PD	6/10/2006	5 27 19	First Sample Volume 0 339L - No Flag Second Sample Volume 0 132L - Flagged Volume Not Met Third Sample Volume 0 976L - Flagged Volume Not Met	Not A Valid Test
Mena	Marianne	10/28/1983	DUI	80-000885	Miami Dade PD	Miami Dade PD	8/20/2006	6 53 05	First Sample Volume 0 117L - Flagged Volume Not Met Second Sample Volume 0 500L - No Flag Third Sample Volume 0 871L - Flagged Volume Not Met	Not A Valid Test
Mena	Marianne	10/28/1983	DUI	80-000885	Miami Dade PD	Miami Dade PD	8/20/2006	7 13 49	First Sample Volume 0 632L - No Flag Second Sample Volume 0 597L - No Flag Third Sample Volume 0 597L - No Flag	Not A Valid Test
Winsion	Alan	4/17/1942	DUI	80-000886	North Miami Beach PD	North Miami Beach PD	4/22/2006	5 40 59	First Sample Volume 0 226L - No Flag Second Sample Volume 0 304L - Flagged Volume Not Met Third Sample Volume 0 101L - Flagged Volume Not Met	Not A Valid Test
Chazanan	Sevak	9/9/1984	DUI	80-000888	Pinellas County SO	Pinellas County SO	6/2/2006	3 28 14	First Sample Volume 1 203L - OK Second Sample Volume 0 089L - Flagged Volume Not Met Third Sample Volume 0 507L - No Flag	Not A Valid Test
Harvey	Raybeth	8/22/1959	DUI	80-000889	Pinellas County SO	Florida Highway Patrol	9/1/2006	21 20 50	First Sample Volume 0 777L - No Flag Second Sample Volume 1 074L - Flagged Volume Not Met Third Sample Volume 1 167L - OK	Not A Valid Test
Lee	Scott W	3/10/1973	DUI	80-000890	Pinellas County SO	Pinellas County SO	5/20/2006	4 31 38	First Sample Volume 1 058L - No Flag No Second Sample - Improper Sample	Not A Valid Test
Davis	Kevin M	5/20/1969	BUI	80-000892	Fish & Wildlife CC	FWC	5/20/2006	14 37 25	First Sample Volume 1 441L - OK Second Sample Volume 0 554L - No Flag Third Sample Volume 2 554L - OK	Valid Test - First and Third Samples
Lopez	Torbio M	10/5/1981	DUI	80-000907	Manatee County SO	Florida Highway Patrol	8/12/2006	0 51 18		

Tamm	Trinity T	6/28/1977	DUI	80-000912	Manatee County SO	Manatee County SO	4/19/2006	5.10.09	First Sample Volume 0.015L - Flagged Volume Not Met Second Sample Volume 0.082L - No Flag Third Sample Volume 0.296L - Flagged Volume Not Met	Not A Valid Test
Skloroo	Richard D	6/29/1962	DUI	80-000912	Manatee County SO	Florida Highway Patrol	8/28/2006	3.01.18	First Sample Volume 0.699L - No Flag Second Sample Volume 1.39L - OK Third Sample Volume 2.125L - OK	Not A Valid Test
Hurtado	Pablo E	3/22/1967	DUI	80-000939	Lee County SO	Lee County SO	4/27/2006	2.24.56	First Sample Volume 0.289L - Flagged Volume Not Met Second Sample Volume 0.550L - No Flag Third Sample Volume 1.496L - OK	Not A Valid Test
Ferro	Lillian	4/26/1962	DUI	80-000939	Lee County SO	Lee County SO	6/12/2006	15.27.43	First Sample Volume 0.347L - No Flag Second Sample Volume 1.871L - OK Third Sample Volume 0.347L - No Flag	Not A Valid Test
Moore	Shannon C	10/25/1982	DUI	80-000939	Lee County SO	Lee County SO	7/21/2006	0.18.42	First Sample Volume 0.375L - No Flag Second Sample Volume 1.871L - OK Third Sample Volume 0.757L - No Flag	Not A Valid Test
Conaway	Doreen M	11/15/1965	DUI	80-000939	Lee County SO	Lee County SO	7/23/2006	2.57.31	First Sample Volume 1.605L - OK Second Sample Volume 0.230L - No Flag Third Sample Volume 0.757L - No Flag	Not A Valid Test
Daughtry	Amanda S	2/11/1972	DUI	80-000940	Brevard County SO	Brevard County SO	8/30/2006	3.25.19	First Sample Volume 0.613L - No Flag Second Sample Volume 1.816L - OK Third Sample Volume 1.816L - OK	Not A Valid Test
Cohen	Nikol A	6/6/1975	DUI	80-000943	Collier County SO	Naples PD	5/28/2006	3.00.50	First Sample Volume 2.144L - OK Second Sample Volume 0.218L - No Flag Third Sample Volume 1.875L - OK	Not A Valid Test
Lyach	Vladimir	2/27/1958	DUI	80-000943	Collier County SO	Collier County SO	8/7/2006	21.20.07	First Sample Volume 0.941L - No Flag Second Sample Volume 0.410L - No Flag Third Sample Volume 0.941L - No Flag	Not A Valid Test
Hall	Deborah L	11/25/1956	DUI	80-000949	Tallahassee PD	Tallahassee PD	7/26/2006	14.18.55	First Sample Volume 0.164L - No Flag Second Sample Volume 1.417L - OK Third Sample Volume 0.164L - No Flag	Not A Valid Test
Bravo-Neitro	Francisco	5/8/1978	DUI	80-000951	Hendry County SO	Hendry County SO	7/30/2006	2.44.40	First Sample Volume 1.316L - OK Second Sample Volume 2.507L - OK Third Sample Volume 0.410L - No Flag	Not A Valid Test
Adams	Rodney B	6/2/1961	DUI	80-000957	Leon County SO	Leon County SO	3/30/2006	19.14.15	First Sample Volume 0.429L - No Flag Second Sample Volume 1.570L - OK Third Sample Volume 0.429L - No Flag	Not A Valid Test
Tucker	Leroy	10/2/1981	DUI	80-000958	Leon County SO	Tallahassee PD	4/15/2006	22.23.00	First Sample Volume 0.347L - No Flag Second Sample Volume 0.000L - Flagged Volume Not Met Third Sample Volume 2.250L - OK	Not A Valid Test
Furman	Shawn A.	10/21/1975	DUI	80-000958	Leon County SO	Tallahassee PD	9/9/2006	1.18.33	First Sample Volume 0.675L - No Flag Second Sample Volume 0.015L - Flagged Volume Not Met Third Sample Volume 2.382L - OK	Not A Valid Test
Childs	Kendra P.	9/8/1969	DUI	80-000961	Wakulla County SO	Wakulla County SO	4/14/2006	23.27.48	First Sample Volume 0.191L - Flagged Volume Not Met Second Sample Volume 1.453L - OK Third Sample Volume 0.191L - Flagged Volume Not Met	Not A Valid Test
Unostegu	Marcos	2/28/1982	DUI	80-000963	Orange County SO	Orlando PD	5/17/2006	4.51.48	First Sample Volume 0.546L - No Flag Second Sample Volume 0.402L - Flagged Volume Not Met Third Sample Volume 1.453L - OK	Not A Valid Test
Altrabi	Fouad	4/2/1977	DUI	80-000963	Orange County SO	Orange County SO	5/29/2006	18.30.03	First Sample Volume 0.765L - No Flag Second Sample Volume 1.933L - OK Third Sample Volume 3.941L - OK	Not A Valid Test
Miranda	Arnoldo	5/6/1983	DUI	80-000964	Orange County SO	Ocoee PD	9/24/2006	4.58.41	First Sample Volume 0.457L - No Flag Second Sample Volume 0.031L - Flagged Volume Not Met Third Sample Volume 1.453L - OK	Not A Valid Test
Rodriguez	Candido J	8/8/1980	DUI	80-000966	Osceola County Corrections	Osceola County SO	4/22/2006	3.00.40	First Sample Volume 0.230L - No Flag Second Sample Volume 0.031L - Flagged Volume Not Met Third Sample Volume 0.554L - Flagged Volume Not Met	Not A Valid Test
Cannon	Heather M	3/3/1987	DUI	80-000999	Seminole County SO	Seminole County SO	9/22/2006	5.13.02	First Sample Volume 0.394L - No Flag Second Sample Volume 0.429L - Flagged Volume Not Met Third Sample Volume 1.250L - OK	Not A Valid Test
Harazin	Linda S	8/24/1947	DUI	80-001003	Pinellas County SO	Pinellas Park PD	5/4/2006	4.47.35	First Sample Volume 0.421L - No Flag Second Sample Volume 0.359L - Flagged Volume Not Met Third Sample Volume 1.44L - Flagged Volume Not Met	Not A Valid Test
Chapman	Douglas R	12/25/1924	DUI	80-001004	Brevard County SO	Brevard County SO	6/25/2006	1.36.28	First Sample Volume 1.636L - OK Second Sample Volume 0.144L - Flagged Volume Not Met Third Sample Volume 1.250L - OK	Not A Valid Test
lby	Howard J	10/7/1962	DUI	80-001005	Pinellas County SO	Pinellas County SO	4/11/2006	0.17.44	First Sample Volume 1.641L - OK Second Sample Volume 1.062L - No Flag Third Sample Volume 0.710L - No Flag	Not A Valid Test
Richmond	Aubrey L	2/9/1950	DUI	80-001005	Pinellas County SO	Pinellas County SO	4/15/2006	1.26.05	First Sample Volume 0.710L - No Flag Second Sample Volume 0.710L - No Flag Third Sample Volume 1.641L - OK	Not A Valid Test



Maggs	Robert F	9/29/1938	DUI	80-001005	Pinellas County SO	Pinellas County SO	7/26/2006	20:38:58	First Sample Volume 0.574L - No Flag Second Sample Volume 0.308L - Flagged Volume Not Met Third Sample Volume 0.261L - Flagged Volume Not Met	Not A Valid Test
Roundtree	Aaron C	8/22/1971	DUI	80-001005	Pinellas County SO	Pinellas County SO	8/20/2006	3:17:46	First Sample Volume 0.429L - No Flag Second Sample Volume 1.648L - OK Third Sample Volume 1.152L - OK	Not A Valid Test
Smith	Heather L	10/10/1987	DUI	80-001009	Palm Bay PD	Palm Bay PD	8/28/2006	4:23:50	First Sample Volume 0.714L - Flagged Volume Not Met Second Sample Volume 0.281L - No Flag Third Sample Volume 0.785L - Flagged Volume Not Met	Not A Valid Test
Thompson	Catherine S	2/18/1958	DUI	80-001009	Palm Bay PD	Palm Bay PD	9/11/2006	0:12:25	First Sample Volume 0.109L - No Flag Second Sample - RFI	Not A Valid Test
Zarchoz	Roberto	1/30/1981	DUI	80-001043	Highlands County SO	Florida Highway Patrol	5/21/2006	0:31:19	First Sample Volume 1.878L - OK Second Sample Volume 0.687L - No Flag Third Sample Volume 1.402L - OK	Not A Valid Test
Cawnder	Brenda A	9/23/1960	DUI	80-001043	Highlands County SO	Sebring PD	6/12/2006	1:56:15	First Sample Volume 1.402L - OK Second Sample Volume 1.640L - OK Third Sample Volume 1.753L - OK	Not A Valid Test
Marley	Lynne C	1/10/1959	DUI	80-001043	Highlands County SO	Florida Highway Patrol	6/10/2006	0:10:08	First Sample Volume 0.304L - No Flag Second Sample Volume 0.351L - No Flag Third Sample Volume 1.863L - OK	Not A Valid Test
Honorat	Hermane	1/9/1956	DUI	80-001044	Cocoanut Creek PD	Cocoanut Creek PD	6/26/2006	23:34:39	First Sample Volume 1.753L - OK Second Sample Volume 0.304L - No Flag Third Sample Volume 0.351L - No Flag	Not A Valid Test
Blair	Jason L.	7/14/1975	DUI	80-001051	St Petersburg PD	St Petersburg PD	7/4/2006	19:14:37	First Sample Volume 0.402L - No Flag Second Sample Volume 1.550L - OK Third Sample Volume 1.363L - OK	Not A Valid Test
Pruett	Dennis R	12/4/1951	DUI	80-001051	St Petersburg PD	St Petersburg PD	8/18/2006	19:07:01	First Sample Volume 0.667L - No Flag Second Sample Volume 1.503L - OK Third Sample Volume 1.363L - OK	Not A Valid Test
Zuniga	Sherry C	7/11/1955	DUI	80-001055	Polk County SO	Florida Highway Patrol	7/29/2006	2:07:52	First Sample Volume 0.300L - No Flag Second Sample Volume 1.125L - OK Third Sample Volume 1.105L - OK	Not A Valid Test
Addleman	Holly L	10/6/1977	DUI	80-001057	Lakeland PD	Lakeland PD	6/18/2006	3:23:44	First Sample Volume 0.699L - No Flag Second Sample Volume 1.656L - OK Third Sample Volume 0.109L - No Flag	Not A Valid Test
Rivera	Victor M	4/5/1950	DUI	80-001057	Lakeland PD	Lakeland PD	6/20/2006	3:12:02	First Sample Volume 0.109L - No Flag Second Sample Volume 1.488L - OK Third Sample Volume 1.574L - OK	Not A Valid Test
Godinez Perez	Alberto	1/11/1975	DUI	80-001063	Hollywood PD	Hollywood PD	7/12/2006	2:21:25	First Sample Volume 0.546L - Flagged Volume Not Met Second Control Test Outside Tolerance Third Sample Volume 0.218L - No Flag	Not A Valid Test
Buchanan	Bonnie T	12/11/1940	DUI	80-001067	Broward County SO	Broward County SO	5/27/2006	19:03:55	First Sample Volume 0.718L - Flagged Volume Not Met Second Sample Volume 1.574L - OK Third Sample Volume 0.218L - No Flag	Not A Valid Test
Lanot	Jesufranc	11/17/1971	DUI	80-001067	Broward County SO	Broward County SO	7/11/2006	22:47:55	First Sample Volume 1.093L - No Flag Second Sample Volume 1.957L - OK Third Sample Volume 1.574L - OK	Not A Valid Test
Nelson	Stacy A	7/22/1963	DUI	80-001069	Jacksonville SO	Jacksonville SO	8/21/2006	4:30:22	First Sample Volume 1.382L - OK Second Sample Volume 0.703L - No Flag Third Sample Volume 0.571L - No Flag	Not A Valid Test
Doyle	Stephen V	7/24/1952	DUI	80-001069	Jacksonville SO	Jacksonville SO	9/5/2006	1:54:10	First Sample Volume 0.671L - No Flag Second Sample Volume 0.550L - No Flag Third Sample Volume 0.949L - No Flag	Not A Valid Test
Buckley	Linda G	5/10/1967	DUI	80-001071	Clay County SO	Clay County SO	5/6/2006	2:11:12	First Sample Volume 0.550L - No Flag Second Sample Volume 0.574L - Flagged Volume Not Met Third Sample Volume 1.457L - OK	Not A Valid Test
Davis	Kim P	11/20/1957	DUI	80-001082	Plantation PD	Sunrise PD	5/9/2006	16:54:56	First Sample Volume 0.671L - No Flag Second Sample Volume 0.714L - Flagged Volume Not Met Third Sample Volume 1.101L - OK	Not A Valid Test
Nguyen	Thai H	8/2/1963	DUI	80-001083	Jacksonville SO	Jacksonville SO	5/14/2006	22:56:41	First Sample Volume 0.949L - No Flag Second Sample - Purge Fail Third Sample Volume 2.578L - OK	Not A Valid Test
Gagne	Audra J	12/29/1966	DUI	80-001117	Pasco County SO	Pasco County SO	4/25/2006	2:54:49	First Sample Volume 0.714L - Flagged Volume Not Met Second Sample Volume 0.179L - Flagged Volume Not Met Third Sample Volume 0.949L - No Flag	Not A Valid Test
Drake	Dawn D	6/9/1970	DUI	80-001117	Pasco County SO	Pasco County SO	4/27/2006	1:28:39	First Sample Volume 1.015L - No Flag Second Sample Volume 1.101L - OK Third Sample Volume 0.097L - Flagged Volume Not Met	Not A Valid Test
Brooks	Geoffrey D	4/20/1972	DUI	80-001131	Volusia County SO	Volusia County SO	7/4/2006	2:15:13	First Sample Volume 0.097L - Flagged Volume Not Met Second Sample Volume 0.019L - No Flag Third Sample Volume 0.511L - Flagged Volume Not Met	Not A Valid Test
Brock	Lyn	7/20/1942	DUI	80-001137	Flagler County SO	Flagler County SO	7/22/2006	18:30:02	First Sample Volume 1.261L - OK Second Sample Volume 0.972L - No Flag	Not A Valid Test
Kenyon	John R	11/20/1979	DUI	80-001139	Broward County SO	Broward County SO	6/14/2006	1:48:17	First Sample Volume 0.574L - No Flag Second Sample Volume 0.308L - Flagged Volume Not Met Third Sample Volume 0.261L - Flagged Volume Not Met	Not A Valid Test

Jean-Jacques	Paholo	5/10/1979	DUI	80-001139	Broward County SO	Broward County SO	7/8/2006	5.35 25	First Sample Volume 0.410L - No Flag Second Sample Volume 1.519L - OK Third Sample Volume 1.750L - OK	Valid Test - Second and Third Samples
Palacios	Miriam L	1/21/1978	DUI	80-001139	Broward County SO	Broward County SO	7/22/2006	3.46 52	First Sample Volume 0.097L - Flagged Volume Not Met Second Sample Volume 0.414L - Flagged Volume Not Met Third Sample Volume 1.406L - OK	Not A Valid Test
Hendry	Mark J	5/31/1960	DUI	80-001139	Broward County SO	Broward County SO	9/22/2006	1.11 06	First Sample Volume 0.531L - No Flag Second Sample Volume 0.421L - No Flag Third Sample Volume 0.191L - No Flag	Not A Valid Test
Barera	Aristides E	8/21/1976	DUI	80-001153	Leon County SO	Leon County SO	4/13/2006	21.55 59	First Sample Volume 1.046L - No Flag Second Sample Volume 0.335L - Flagged Volume Not Met Third Sample Volume 1.199L - OK	Not A Valid Test
Gibson	Amelia C	7/30/1986	DUI	80-001153	Leon County SO	Leon County SO	8/23/2006	5.26 19	First Sample Volume 0.367L - No Flag Second Sample Volume 0.125L - Flagged Volume Not Met Third Sample Volume 1.097L - No Flag	Not A Valid Test
Waggoner	Nicholas C	5/4/1991	DUI	80-001158	Ponce Inlet PD	Daytona Beach PD	4/5/2006	23.12 41	First Sample Volume 2.066L - OK Second Sample Volume 1.058L - No Flag Third Sample Volume 0.828L - No Flag	Not A Valid Test
Currie II	Robert A	5/11/1955	DUI	80-001166	Brevard County SO	Brevard County SO	4/13/2006	20.11.17	First Sample Volume 2.519L - OK Second Sample Volume 0.726L - No Flag Third Sample Volume 1.406L - OK	Not A Valid Test
Zhang	Dong	6/24/1967	DUI	80-001190	Boynton Beach PD	Boynton Beach PD	6/17/2006	8.35 40	First Sample Volume 0.015L - Flagged Volume Not Met Second Sample Volume 2.066L - OK Third Sample Volume 0.726L - No Flag	Not A Valid Test
Ferranti	Edward J	5/26/1970	DUI	80-001190	Boynton Beach PD	Boynton Beach PD	7/22/2006	3.11 52	First Sample Volume 1.406L - OK Second Sample Volume 0.441L - No Flag Third Sample Volume 1.281L - OK	Not A Valid Test
Kaszak	Kenneth	4/16/1958	DUI	80-001190	Boynton Beach PD	Boynton Beach PD	8/13/2006	0.46 51	First Sample Volume 1.285L - OK Second Sample Volume 0.488L - No Flag Third Sample Volume 1.281L - OK	Not A Valid Test
Perez-Ortiz	Issac	10/23/1982	DUI	80-001235	West Palm Beach PD	West Palm Beach PD	4/8/2006	0.03 15	First Sample Volume 0.593L - No Flag Second Sample Volume 0.816L - No Flag Third Sample Volume 0.347L - No Flag	Not A Valid Test
Rucker	Curtis	3/12/1945	DUI	80-001248	Broward County SO	Broward County SO	5/7/2006	0.02 11	First Sample Volume 1.765L - OK Second Sample Volume 0.468L - No Flag Third Sample Volume 1.785L - OK	Not A Valid Test
Meener	Patricia	11/3/1968	DUI	80-001253	Brevard County SO	Brevard County SO	6/26/2006	22.40 39	First Sample Volume 0.644L - No Flag Second Sample Volume 1.242L - OK Third Sample Volume 1.242L - OK	Not A Valid Test
Brown	Joshua B	8/10/1976	DUI	80-001253	Brevard County SO	Brevard County SO	8/26/2006	23.45 57	First Sample Volume 0.679L - No Flag Second Sample Volume 0.160L - Flagged Volume Not Met Third Sample - Subject Refused	Not A Valid Test
Doane	Richard G	10/26/1963	DUI	80-001255	Orange County SO	Orange County SO	4/7/2006	3.46 24	First Sample Volume 1.628L - OK Second Sample Volume 0.480L - No Flag Third Sample Volume 3.035L - OK	Not A Valid Test
Aponte	Edgardo	10/9/1979	DUI	80-001255	Orange County SO	Orlando PD	9/9/2006	1.43 01	First Sample Volume 0.746L - No Flag Second Sample Volume 1.128L - OK Third Sample Volume 1.351L - OK	Not A Valid Test
Quiros	Ovalle	12/15/1978	DUI	80-001257	Orange County SO	Orlando PD	3/30/2006	4.45 18	First Sample Volume 1.184L - OK Second Sample Volume 1.039L - No Flag Third Sample Volume 0.296L - Flagged Volume Not Met	Not A Valid Test
Meinken	Bradford G	11/3/1976	DUI	80-001257	Orange County SO	Orlando PD	8/5/2006	4.53 30	First Sample Volume 0.339L - No Flag Second Sample Volume 1.312L - OK Third Sample Volume 1.406L - OK	Not A Valid Test
Thomas	Jean C	7/21/1953	DUI	80-001258	Orange County SO	Florida Highway Patrol	5/6/2006	21.11 50	First Sample Volume 1.184L - OK Second Sample Volume 1.039L - No Flag Third Sample Volume 0.296L - Flagged Volume Not Met	Not A Valid Test
McDaniel	Joseph N	10/6/1937	DUI	80-001260	Cocoa PD	Brevard County SO	7/15/2006	1.37 50	First Sample Volume 0.097L - Flagged Volume Not Met Second Sample Volume 1.363L - OK Third Sample Volume 0.972L - No Flag	Not A Valid Test
Gause	Julia F	1/21/1969	DUI	80-001262	Melbourne PD	Melbourne PD	5/29/2006	22.12 15	First Sample Volume 1.226L - OK Second Sample Volume 0.714L - No Flag Third Sample Volume 1.769L - OK	Not A Valid Test
Young	Pamela	8/5/1958	DUI	80-001267	Palm Beach County So	Riveria Beach PD	5/30/2006	1.40 07	First Sample Volume 0.972L - No Flag Second Sample Volume 1.363L - OK Third Sample Volume 0.367L - No Flag	Not A Valid Test
Keenan	Maureen E	10/9/1970	DUI	80-001267	Palm Beach County So	Palm Beach County SO	7/19/2006	1.31 08	First Sample Volume 1.226L - OK Second Sample Volume 0.714L - No Flag Third Sample Volume 1.769L - OK	Not A Valid Test
Vilatoro	Yener R	8/24/1984	DUI	80-001267	Palm Beach County So	Palm Beach County SO	8/6/2006	4.53 20	First Sample Volume 1.445L - OK Second Sample Volume 0.609L - No Flag Third Sample Volume 1.437L - OK	Not A Valid Test
Suarez	Jean	3/31/1969	DUI	80-001269	Palm Beach County So	Florida Highway Patrol	5/5/2006	6.55 45	First Sample Volume 1.445L - OK Second Sample Volume 0.609L - No Flag Third Sample Volume 1.437L - OK	Not A Valid Test
Goldberg	Paul E	3/18/1969	DUI	80-001269	Palm Beach County So	Florida Highway Patrol	8/24/2006	3.44 46	First Sample Volume 1.445L - OK Second Sample Volume 0.609L - No Flag Third Sample Volume 1.437L - OK	Not A Valid Test

Toft	Carole A	8/6/1964	DUI	80-001269	Palm Beach County So	Palm Beach County SO	9/21/2006	4 12 02	First Sample Volume 0 253L - No Flag Second Sample Volume 1 207L - OK First Sample Volume 0 613L - No Flag	Not A Valid Test
Cornell	Daved E	2/2/1965	DUI	80-001270	Palm Beach County So	Delray Beach PD	7/23/2006	22 32 26	First Sample Volume - Purge Fail First Sample Volume 0 867L - No Flag Second Sample Volume 1 703L - OK	Not A Valid Test
Harris	Kareem R	6/3/1968	DUI	80-001272	Seminole County SO	Oviedo PD	4/2/2006	5 47 26	First Sample Volume 0 710L - No Flag Second Sample Volume 0 332L - Flagged Volume Not Met Third Sample Volume 0 785L - Flagged Volume Not Met	Not A Valid Test
Wells	Barry N	9/7/1968	DUI	80-001272	Seminole County SO	Lake Mary PD	4/10/2006	19 11 26	First Sample Volume 0 675L - No Flag Second Sample Volume 0 796L - OK Second Control Test Outside Tolerance	Not A Valid Test
Wilson	Glen M	9/15/1964	DUI	80-001272	Seminole County SO	Florida Highway Patrol	6/25/2006	2 12 47	Second Sample Volume 0 257L - Flagged Volume Not Met Third Sample Volume 0 679L - No Flag	Not A Valid Test
White	Leslie J	6/6/1960	DUI	80-001272	Seminole County SO	Oviedo PD	7/2/2006	2 46 00	Second Control Test Outside Tolerance First Sample Volume 0 457L - No Flag Second Sample Volume 1 390L - OK	Not A Valid Test
Newton	Justin B	10/21/1978	DUI	80-001272	Seminole County SO	Seminole County SO	7/21/2006	22 17 44	First Sample Volume 0 050L - No Flag Second Sample Volume 1 414L - OK Third Sample Volume 1 613L - OK	Not A Valid Test
Lopez-Hernandez	David	4/4/1978	DUI	80-001273	Pinellas County SO	Pinellas County SO	7/15/2006	3 03 14	First Sample Volume 0 109L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Portuondo	Enk J	12/1/1975	DUI	80-001276	Miami Dade PD	Miami Dade PD	5/12/2006	0 34 06	First Sample Volume 0 453L - No Flag Second Sample Volume 1 839L - OK Third Sample Volume - 1 292L - OK	Not A Valid Test
Fandino	Maria L	8/18/1967	DUI	80-001276	Miami Dade PD	Miami Dade PD	5/27/2006	4 12 02	First Sample Volume 0 867L - No Flag Second Sample Volume 0 152L - Flagged Volume Not Met Third Sample Volume 0 082L - Flagged Volume Not Met	Not A Valid Test
Fuhr	Alison R	10/31/1978	DUI	80-001282	University of Florida PD	University of FL PD	4/13/2006	3 34 42	First Sample Volume 0 421L - No Flag Second Sample Volume 0 378L - Flagged Volume Not Met Third Sample Volume 0 816L - No Flag	Not A Valid Test
Johns	Cynthia J	5/18/1966	DUI	80-001284	Glitchri County SO	Cedar Keys PD	6/15/2006	1 38 28	First Sample Volume 0 289L - Flagged Volume Not Met Second Sample Volume 0 789L - No Flag Third Sample Volume 1 687L - OK	Not A Valid Test
Campbell	Steven M	7/12/1974	DUI	80-001303	Santa Rosa County SO	Santa Rosa County SO	8/6/2006	20 30 09	First Sample Volume 0 628L - No Flag Second Sample Volume 0 578L - Flagged Volume Not Met Third Sample Volume - Subject Refused	Not A Valid Test
Phillips	Darius K	4/9/1984	OTHER	80-001309	Bay County SO	Tyndall AFB	4/9/2006	13 22 01	First Sample Volume 0 511L - No Flag Second Sample Volume 0 000L - Flagged Volume Not Met Third Sample Volume 0 437L - Flagged Volume Not Met	Not A Valid Test
Cardwell	Curtis B	11/24/1981	DUI	80-001309	Bay County SO	Bay County SO	7/23/2006	23 35 34	First Sample Volume 0 437L - No Flag Second Sample Volume - Subject Refused	Not A Valid Test
Miller	Chadrick L	4/24/1978	DUI	80-001309	Bay County SO	Parker PD	9/16/2006	22 34 35	First Sample Volume 2 710L - OK Second Sample Volume 0 097L - No Flag Third Sample Volume 1 925L - OK	Valid Test - First and Third Samples
Mings	Lonnie L	6/29/1977	DUI	80-001309	Bay County SO	Florida Highway Patrol	9/23/2006	3 41 36	First Sample Volume 0 906L - No Flag Second Sample Volume 2 613L - OK	Not A Valid Test
Castillo-Vazquez	Jose	12/6/1960	DUI	80-001309	Bay County SO	Florida Highway Patrol	9/24/2006	1 50 55	First Sample Volume 1 246L - OK Second Sample Volume 0 558L - No Flag Third Sample Volume 1 394L - OK	Valid Test - First and Third Samples
Clay	Susan S.	12/9/1953	DUI	80-001310	Niceville PD	Crestview PD	6/12/2006	0 08 55	First Sample Volume 0 554L - No Flag Second Sample Volume 1 003L - No Flag	Not A Valid Test
Diamond	Elfriede F.	11/1/1936	DUI	80-001328	Indian River County SO	Sebastian PD	8/30/2006	22 47 27	First Sample Volume 1 972L - OK Second Sample Volume 0 156L - Flagged Volume Not Met Third Sample Volume 0 882L - No Flag	Not A Valid Test
Angus	Anthony	6/11/1959	DUI	80-001328	Indian River County SO	Florida Highway Patrol	9/11/2006	8 34 49	First Sample Volume 0 093L - Flagged Volume Not Met Second Sample Volume 3 066L - OK Third Sample Volume 0 714L - No Flag	Not A Valid Test
Kellowan	Shevron S	3/20/1979	DUI	80-001344	Sarasota County SO	Florida Highway Patrol	4/1/2006	4 55 54	First Sample Volume 0 937L - No Flag Second Sample Volume 1 972L - OK	Not A Valid Test
Ryll	Frank L	10/27/1951	DUI	80-001344	Sarasota County SO	Sarasota County SO	4/25/2006	22 40 18	First Sample Volume 0 468L - No Flag Second Sample Volume 1 871L - OK Third Sample Volume 1 691L - OK	Not A Valid Test
Kerr	Julia P	2/11/1987	DUI	80-001344	Sarasota County SO	Sarasota County SO	5/3/2006	4 22 57	First Sample Volume 0 253L - No Flag Second Sample Volume 1 207L - OK First Sample Volume 0 613L - No Flag	Not A Valid Test

Shapiro	Wayne G	3/25/1959	DUI	80-001344	Sarasota County SO	Sarasota PD	5/28/2006	3 58 01	First Sample Volume 0 656L - No Flag Second Sample Volume 1 683L - OK Third Sample Volume 1 800L - OK	Valid Test - Second and Third Samples
Quattrone	Nancy E	4/9/1957	DUI	80-001344	Sarasota County SO	Sarasota County SO	6/7/2006	0 27 30	First Sample Volume 0 324L - No Flag Second Sample Volume 1 921L - OK	Not A Valid Test
Velle	Victorino	10/27/1979	DUI	80-001344	Sarasota County SO	Sarasota County SO	6/18/2006	3 26 40	First Sample Volume 1 625L - OK Second Sample Volume 0 636L - No Flag Third Sample Volume 2 160L - OK	Valid Test - First and Third Samples
Palmer	Scott D	7/13/1964	DUI	80-001344	Sarasota County SO	Sarasota PD	6/26/2006	1 53 50	First Sample Volume 1 101L - OK Second Sample Volume 0 804L - No Flag No Third Sample - Subject Refused	Not A Valid Test Refusal
Delahanty	Thomas L	5/15/1948	DUI	80-001344	Sarasota County SO	Sarasota County SO	7/8/2006	0 30 50	First Sample Volume 0 492L - No Flag Second Sample Volume 0 191L - Flagged Volume Not Met Third Sample Volume 0 378L - No Flag	Not A Valid Test
German	David M	4/8/1957	DUI	80-001367	Pinellas County SO	Florida Highway Patrol	7/16/2006	18 23 45	First Sample Volume 0 906L - No Flag Second Sample Volume 1 125L - OK Third Sample Volume 1 371L - OK	Valid Test - Second and Third Samples
Desimas	John R	2/1/1984	DUI	80-001415	Pinellas County SO	Pinellas Park PD	8/5/2006	3 02 07	First Sample Volume 0 130L - No Flag Second Sample Volume 0 593L - Flagged Volume Not Met Third Sample Volume 0 144L - Flagged Volume Not Met	Not A Valid Test
Dillmone	Linda J	3/13/1962	DUI	80-001416	Orange County SO	Orange County SO	6/6/2006	12 48 14	First Sample Volume 0 835L - No Flag Second Sample Volume 0 437L - Flagged Volume Not Met Third Sample Volume 0 070L - Flagged Volume Not Met	Not A Valid Test
Martinez	Noel I	9/22/1968	DUI	80-001416	Orange County SO	Orange County SO	6/10/2006	4 17 28	First Sample Volume 0 937L - No Flag Second Sample Volume 1 718L - OK	Not A Valid Test
Kars	Chad D	1/4/1967	DUI	80-001416	Orange County SO	Orlando PD	7/10/2006	1 34 39	First Sample Volume 1 308L - OK Second Sample Volume 0 792L - No Flag	Not A Valid Test
Reyes	Arturo	9/11/1965	DUI	80-001418	Orange County SO	Florida Highway Patrol	8/30/2006	1 34 24	First Sample Volume 0 042L - Flagged Volume Not Met Second Sample Volume 0 804L - No Flag Third Sample Volume 2 421L - OK	Not A Valid Test
Poole	Nikki K	7/17/1975	DUI	80-001418	Orange County SO	Maitland PD	8/30/2006	3 38 20	First Sample Volume 0 703L - No Flag Second Sample Volume 0 261L - Flagged Volume Not Met Third Sample Volume 1 531L - OK	Not A Valid Test
Stephens	Patrick J	11/20/1984	DUI	80-001420	Orange County SO	Orange County SO	6/11/2006	4 28 54	First Sample Volume 2 972L - OK Second Sample Volume 0 394L - No Flag Third Sample Volume 1 789L - OK	Not A Valid Test
Buckley	Linda G	5/10/1967	DUI	80-001443	Clay County SO	Clay County SO	5/6/2006	1 46 18	First Sample Volume 1 464L - OK Second Sample Volume 0 648L - No Flag Third Sample Volume 0 648L - No Flag Second Control Test Outside Tolerance	Not A Valid Test

\*\* Notes are concerning the listed breath tests Additional breath tests of the referenced subject may have been conducted and are not a part of this list



Florida Department of  
Law Enforcement

Gerald M. Bailey  
Interim Commissioner

Criminal Justice  
Professionalism Program  
Alcohol Testing Program

P. O. Box 1489  
Tallahassee, Florida 32302  
(850) 410-8600  
(850) 410-7816 Fax  
<http://www.fdle.state.fl.us>

October 5, 2006

NAME  
ADDRESS  
ADDRESS  
ADDRESS

SUBJECT: Intoxilyzer 8000

Dear NAME:

This letter is to inform you of an issue regarding the Intoxilyzer 8000 that the Florida Department of Law Enforcement Alcohol Testing Program has discovered. We have been in contact with the manufacturer of the Intoxilyzer 8000, CMI, Inc., and have determined there are missing instructions in the Intoxilyzer 8000 software version 8100.26.

The issue involves the instrument's ability to correctly identify the volume of a breath sample delivered past three minutes. The programming in software version 8100.26 was not written to address this very unique and specific scenario. This issue is affecting less than 1% of all breath tests conducted on the Intoxilyzer 8000.

A software revision to add these missing instructions is in the process of being completed and will be installed in your Intoxilyzer 8000 instruments in the near future. **I would like to assure you that software version 8100.26 is operating as written and programmed. The analytical functions of the instrument are not affected by this issue and the instrument continues to correctly analyze all breath samples.** It is only with this unique and specific scenario that the breath sample has not been determined to be reliable.

A detailed letter is being forwarded under separate cover to Criminal Justice Agencies in your circuit. Enclosed is a summary of the affected breath tests determined as of October 2, 2006. The Alcohol Testing Program is continuing to closely monitor every breath test result obtained in the State of Florida and will contact you regarding any additional affected breath tests conducted by your agency. If you have any questions, please feel free to contact me at (850) 410-7810.

Sincerely,

Gerald M. Bailey  
Interim Commissioner

Laura D. Barfield  
Alcohol Testing Program Manager

LDB/lb

Enclosure

*Committed to  
Service - Integrity - Respect - Quality*

Subject Last Name	Subject First Name	Subject's Date of Birth	Offense	Instrument Serial Number	Instrument Registered To	Arresting Officer Agency	Date of Test	Time of Test	Breath Sample Volume Results	Notes
Corredor	Miguel A	1/7/1986	DUI	80-000223	FDLE (Loaned to Indian River County Sheriff's Office)	Florida Highway Patrol	7/6/2006	22 58 02	First Sample Volume 1 933 L - OK Second Sample Volume 0 972L - No Flag	Not A Valid Test
Sedovic	Michael T	1/14/1976	DUI	80-000738	Fort Walton Beach PD	Okaloosa County SO	8/5/2006	4 54 39	First Sample Volume 1 066L - No Flag Second Sample Volume 0 226L - Flagged Volume Not Met	Not A Valid Test
Pelullo	Temilola A	12/24/1961	DUI	80-000739	Okaloosa County SO	Florida Highway Patrol	8/21/2006	0 10 12	First Sample Volume 1 550L - OK Second Sample Volume 1 164L - OK	Not A Valid Test
Fulford	Kenneth M	7/7/1951	DUI	80-000744	Escambia County SO	Pensacola PD	5/17/2006	2 04 10	First Sample Volume 0 664L - No Flag Second Sample Volume 0 605L - No Flag	Not A Valid Test
Burgos	Abel	7/11/1978	DUI	80-000744	Escambia County SO	Florida Highway Patrol	7/22/2006	3 58 14	First Sample Volume 2 105L - OK Second Sample Volume 1 082L - No Flag	Not A Valid Test
East	Easter K	3/21/1945	DUI	80-000744	Escambia County SO	Florida Highway Patrol	8/12/2004	2 04 13	First Sample Volume 0 601L - No Flag Second Sample - Subject Refused	Not A Valid Test
Rives	Lisa C	6/8/1959	DUI	80-000744	Escambia County SO	Florida Highway Patrol	9/28/2006	21 07 00	First Sample Volume 1 515L - OK Second Sample Volume 0 062L - No Flag	Not A Valid Test
Swanson	Amanda N	9/9/1984	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	5/18/2006	5 39 00	No Second Sample - Purge Fail	Not A Valid Test
Hidle	Brandi C	11/9/1972	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	5/25/2006	1 05 16	First Sample Volume 0 129L - Flagged Volume Not Met Second Sample Volume 0 347L - No Flag	Not A Valid Test
Toney	Patrick S	3/10/1987	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	5/30/2006	1 29 54	First Sample Volume 1 000L - No Flag Second Sample Volume 1 550L - OK	Not A Valid Test
Boone	Michael J	11/9/1983	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	6/25/2006	2 30 07	First Sample Volume 0 742L - No Flag Second Sample Volume 0 054L - Flagged Volume Not Met	Not A Valid Test
Alicandri	Danielle M	8/1/1971	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	7/25/2006	2 36 03	First Sample Volume 0 476L - No Flag Second Sample Volume 1 738L - OK	Not A Valid Test
Jones	Elaine J	9/18/1969	DUI	80-000772	Santa Rosa County SO	Santa Rosa County SO	5/13/2006	2 38 47	First Sample Volume 0 652L - No Flag Second Sample Volume 1 738L - OK	Not A Valid Test
Bohannon	Lynette D	12/18/1972	DUI	80-000786	Broward County SO	Broward County SO	5/12/2006	0 10 42	No Second Sample - Purge Fail	Not A Valid Test
Shelton	Linda M	8/27/1940	DUI	80-000790	Vero Beach PD	Vero Beach PD	7/13/2006	3 05 22	First Sample Volume 0 851L - Flag Volume Not Met Second Sample Volume 1 292L - OK	Not A Valid Test
Vaught	Michael A	4/20/1981	DUI	80-000794	St Lucie County SO	St Lucie County SO	4/9/2006	2 49 38	First Sample Volume 1 527L - OK Second Sample Volume 0 613L - No Flag	Not A Valid Test
Lesko	Darryl R	5/10/1958	DUI	80-000794	St Lucie County SO	St Lucie County SO	6/5/2006	1 49 07	First Sample Volume 0 210L - No Flag Second Sample Volume 1 371L - OK	Not A Valid Test
Conerz	Javier O	1/14/1964	DUI	80-000794	St Lucie County SO	Fort Pierce PD	6/10/2006	0 36 52	First Sample Volume 1 570L - OK Second Sample Volume 0 300L - No Flag	Not A Valid Test
Scheider	Robert P	4/23/1962	DUI	80-000794	St Lucie County SO	Fort Pierce PD	6/19/2006	20 12 44	Third Sample Volume 2 429L - OK Second Control Test Outside Tolerance	Not A Valid Test
Slater	Keith	6/24/1972	DUI	80-000794	St Lucie County SO	Fort Pierce PD	6/21/2006	9 25 13	First Sample Volume 1 113L - OK Second Sample Volume 0 234L - No Flag	Not A Valid Test
Maliszewski	Dagmar E	8/5/1967	DUI	80-000794	St Lucie County SO	Port St Lucie PD	6/25/2006	23 16 05	Third Sample Volume 1 207L - OK Second Control Test Outside Tolerance	Not A Valid Test
Diaz	Alejandro	10/29/1984	DUI	80-000794	St Lucie County SO	St Lucie County SO	8/19/2006	3 20 08	First Sample Volume 0 812L - No Flag Second Sample - Subject Refused	Not A Valid Test
Jimenez	David Y	1/24/1977	DUI	80-000808	Hillsborough County SO	Hillsborough County SO	3/30/2006	23 07 03	First Sample Volume 0 816L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Pitasa	Petar J	11/25/1982	DUI	80-000808	Hillsborough County SO	Tampa PD	7/2/2006	3 58 14	First Sample Volume 1 054L - No Flag Second Sample Volume 1 261L - OK	Not A Valid Test

Roberts	Pamela R	7/23/1984	DUI	80-000808	Hillsborough County SO	Tampa PD	8/24/2006	0 30 17	First Sample Volume 1 070L - No Flag Second Sample Volume 1 679L - OK	Not A Valid Test
Reyes	Tony	6/11/1984	DUI	80-000808	Hillsborough County SO	Plant City PD	9/3/2006	0 04 38	First Sample Volume 0 363L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Clark	Michael C	4/9/1983	DUI	80-000808	Hillsborough County SO	Florida Highway Patrol	9/9/2006	5 52 04	First Sample Volume 0 246L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Hargrave	Evelyn M	7/20/1979	DUI	80-000808	Hillsborough County SO	Tampa PD	9/10/2006	3 08 59	First Sample Volume 0 925L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Diaz Gonzalez	Raul N	11/24/1957	DUI	80-000813	Marion County SO	Marion County SO	6/21/2006	14 35 11	First Sample Volume 0 476L - No Flag Second Sample Volume 0 476L - No Flag Purge Fail After Second Sample	Not A Valid Test
Merritt	Brady J	9/9/1984	DUI	80-000817	Citrus County SO	Citrus County SO	9/14/2006	4 23 44	First Sample Volume 0 398L - No Flag Second Sample Volume 0 179L - No Flag	Not A Valid Test
Welch	Andrei C	3/20/1966	DUI	80-000824	Leesburg PD	Leesburg PD	4/15/2006	0 19 50	First Sample Volume 0 324L - No Flag Second Sample Volume 1 117L - OK	Not A Valid Test
Garcia	Alfonso R	8/17/1980	DUI	80-000829	Hillsborough County SO	Tampa PD	4/12/2006	2 16 24	First Sample Volume 0 339L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Nilsen	Michael R	10/9/1961	DUI	80-000832	Martin County SO	Florida Highway Patrol	5/29/2006	3 47 51	First Sample Volume 1 660L - OK Second Sample Volume 0 972L - No Flag Third Sample Volume 1 539L - OK	Valid Test - First and Third Samples
Camacho	Rafael	4/3/1961	DUI	80-000832	Martin County SO	Stuart PD	6/22/2006	19 54 57	First Sample Volume 0 597L - No Flag Second Sample Volume 0 204L - Flagged Volume Not Met Third Sample Volume 0 219L - Flagged Volume Not Met	Not A Valid Test
Mendoza	Cecilio	11/22/1962	DUI	80-000832	Martin County SO	Martin County SO	7/9/2006	2 12 46	First Sample Volume 0 183 - No Flag Second Sample Volume 2 226L - OK Third Sample Volume 0 183 - No Flag	Not A Valid Test
Schroeder	Brandon H	7/28/1987	DUI	80-000832	Martin County SO	Martin County SO	8/13/2006	2 45 47	First Sample Volume 0 468L - Flagged Volume Not Met Second Sample Volume 0 769L - No Flag Third Sample Volume 1 480L - OK	Not A Valid Test
Daher	Jabib	2/3/1948	DUI	80-000832	Martin County SO	FWC West Palm	8/26/2006	23 30 17	First Sample Volume 1 726L - OK Second Sample Volume 0 449L - No Flag	Not A Valid Test
Vandermark	Jeffrey W	6/16/1945	DUI	80-000832	Martin County SO	Sewall's Point PD	9/4/2006	2 04 32	First Sample Volume 0 808L - No Flag Second Sample Volume 3 394L - OK	Not A Valid Test
Michale	James M	2/27/1967	DUI	80-000832	Martin County SO	Stuart PD	9/4/2006	21 48 28	First Sample Volume 0 203L - No Flag Second Sample Volume 1 691L - OK Third Sample Volume 1 187L - OK	Not A Valid Test
Moreth	Derek P	6/2/1980	DUI	80-000833	Hillsborough County SO	Tampa PD	5/15/2006	1 09 00	First Sample Volume 0 742L - No Flag Second Sample Volume 0 196L - Flag Volume Not Met No Third Sample - Subject Refused	Not A Valid Test
Perez	Palermo	11/20/1972	DUI	80-000833	Hillsborough County SO	Hillsborough County SO	6/10/2006	19 51 30	First Sample Volume 0 105L - No Flag Second Sample Volume 0 212L - Flagged Volume Not Met Third Sample Volume 0 219L - Flagged Volume Not Met	Not A Valid Test
Pospischil	Ashley N	5/10/1983	DUI	80-000833	Hillsborough County SO	Tampa PD	7/3/2006	4 13 57	First Sample Volume 0 957L - No Flag Second Sample Volume 1 648L - OK	Not A Valid Test
Gandanilla	Petronio R	9/6/1972	DUI	80-000833	Hillsborough County SO	Plant City PD	7/21/2006	9 12 20	First Sample Volume 0 371L - No Flag Second Sample - Subject Refused	Not A Valid Test
Rojas	Oscar	8/27/1949	DUI	80-000833	Hillsborough County SO	Tampa PD	8/11/2006	2 46 47	First Sample Volume 0 625L - No Flag Second Sample - Subject Refused	Not A Valid Test
Rolon	Jose A	3/8/1961	DUI	80-000833	Hillsborough County SO	Florida Highway Patrol	8/13/2006	4 32 45	First Sample Volume 0 574L - No Flag Second Sample Volume 1 289L - OK	Not A Valid Test
Seymour	Patrick H	8/26/1967	DUI	80-000859	Cape Coral PD	Cape Coral PD	6/21/2006	20 35 16	First Sample Volume 0 429L - No Flag Second Sample Volume 0 194L - Flagged Volume Not Met Third Sample Volume 0 625L - No Flag	Not A Valid Test
Lisa	Denise A	12/9/1957	DUI	80-000859	Cape Coral PD	Cape Coral PD	7/10/2006	0 54 05	First Sample Volume 0 464L - No Flag Second Sample Volume 0 168L - Flagged Volume Not Met Third Sample Volume 0 582L - No Flag	Not A Valid Test
Hyland	Cara L	8/23/1986	DUI	80-000859	Cape Coral PD	Cape Coral PD	8/2/2006	2 37 55	First Sample Volume 1 035L - Flagged Volume Not Met Second Sample Volume 0 980L - No Flag Third Sample Volume 0 015L - Flagged Volume Not Met	Not A Valid Test
Veletz	Ernesto	7/22/1966	DUI	80-000859	Cape Coral PD	Cape Coral PD	8/6/2006	1 33 11	First Sample Volume 0 097L - Flagged Volume Not Met Second Sample Volume 0 457L - No Flag Third Sample Volume 0 687L - Flagged Volume Not Met	Not A Valid Test
Jarrett	Donald	12/28/1955	DUI	80-000861	Sarasota County SO	Sarasota County SO	9/12/2006	21 54 31	First Sample Volume 0 035L - No Flag Second Sample - Subject Refused	Not A Valid Test

Brown	Nathan	5/1/1979	DUI	80-000868	Monroe County SO	NAS Key West	7/30/2006	2 11 35	First Sample Volume 0 250L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Morinere	Laurent L	6/24/1947	BUJ	80-000869	Monroe County SO	Other State	4/15/2006	20 19 14	First Sample Volume 0 480L - No Flag Second Sample Volume 0 179L - Flagged Volume Not Met Third Sample Volume 1 804L - OK	Not A Valid Test
Clarke	Brian D	12/18/1961	BUJ	80-000869	Monroe County SO	FWC	5/20/2006	21 23 30	First Sample Volume 0 687L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Cornejo	Jose	4/17/1980	DUI	80-000869	Monroe County SO	Monroe County SO	7/22/2006	4 48 56	First Sample Volume 1 078L - No Flag Second Sample Volume 1 246L - OK	Not A Valid Test
Kelly	Shane L	2/6/1981	DUI	80-000869	Monroe County SO	Monroe County SO	8/7/2006	5 16 12	First Sample Volume 0 449L - Flagged Volume Not Met Second Sample Volume 0 535L - No Flag Third Sample Volume 1 585L - OK	Not A Valid Test
La Paz	Gilda	11/20/1943	DUI	80-000871	Coral Gables PD	Coral Gables PD	6/4/2006	2 28 46	Second Sample Volume 0 149L - Flagged Volume Not Met Third Sample Volume 1 394L - OK	Not A Valid Test
Costa Moran	Enrique A	2/27/1975	DUI	80-000872	Miami PD	Miami PD	8/25/2006	3 03 40	First Sample Volume 1 578L - OK Second Sample Volume 0 328L - No Flag	Not A Valid Test
Mena	Hollman E	12/5/1984	DUI	80-000873	Miami PD	Miami PD	9/11/2006	9 26 14	First Sample Volume 0 808L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Mena	Hollman E	12/5/1984	DUI	80-000872	Miami PD	Miami PD	9/11/2006	9 52 59	First Sample Volume 1 078L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Diaz	Gerardo	7/30/1929	DUI	80-000873	Miami PD	Miami PD	4/11/2006	22 22 46	First Sample Volume 0 535L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Knowles	Patrick J	5/28/1966	DUI	80-000873	Miami PD	South Miami PD	9/16/2006	1 55 54	First Sample Volume 1 101L - OK Second Sample Volume 1 027L - No Flag	Not A Valid Test
Alonso	Ruben	11/22/1959	DUI	80-000876	Hialeah PD	Hialeah PD	5/11/2006	20 40 29	First Sample Volume 0 718L - No Flag Second Sample Volume 1 105L - OK	Not A Valid Test
Franco	Luis A	9/30/1967	DUI	80-000876	Hialeah PD	Hialeah PD	6/26/2006	21 54 09	First Sample Volume 0 687L - No Flag Second Control Test Outside Tolerance	Not A Valid Test
Bello	Farah T	10/14/1978	DUI	80-000879	Miami Beach PD	Miami Beach PD	9/16/2006	5 27 26	No Second Sample - Purge Fail	Not A Valid Test
Lozano	Cesar U	10/7/1975		80-000881	Miami Dade PD	Miami Dade PD	9/20/2006	2 09 52	First Sample Volume 0 304L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Martinez	Giver O	9/2/1984	DUI	80-000884	Miami Dade PD	Miami Dade PD	4/9/2006	4 38 12	First Sample Volume 2 617L - OK Second Sample Volume 0 460L - No Flag	Not A Valid Test
Burgos-Donates	Myrna	8/9/1957	DUI	80-000885	Miami Dade PD	Miami Dade PD	5/6/2006	2 43 50	First Sample Volume 0 273L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Velasquez	Valerie A	3/1/1981	DUI	80-000885	Miami Dade PD	Miami Dade PD	6/10/2006	5 27 19	First Sample Volume 0 433L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Mena	Marianne	10/28/1983	DUI	80-000885	Miami Dade PD	Miami Dade PD	8/20/2006	6 53 05	First Sample Volume 0 117L - Flagged Volume Not Met Second Sample Volume 0 500L - No Flag Third Sample Volume 0 976L - Flagged Volume Not Met	Not A Valid Test
Mena	Marianne	10/28/1983	DUI	80-000885	Miami Dade PD	Miami Dade PD	8/20/2006	7 13 49	First Sample Volume 0 871L - Flagged Volume Not Met Second Sample Volume 0 500L - No Flag Third Sample Volume 0 101L - Flagged Volume Not Met	Not A Valid Test
Winston	Alan	4/17/1942	DUI	80-000886	North Miami Beach PD	North Miami Beach PD	4/22/2006	5 40 59	First Sample Volume 0 597L - No Flag Second Sample Volume 0 226L - No Flag Third Sample Volume 0 304L - Flagged Volume Not Met	Not A Valid Test
Ghazarian	Sevak	9/9/1984	DUI	80-000888	Pineellas County SO	Pineellas County SO	6/21/2006	3 28 14	First Sample Volume 0 101L - Flagged Volume Not Met Second Sample Volume 0 507L - No Flag Third Sample Volume 0 777L - No Flag	Not A Valid Test
Harvey	Raybein	8/22/1959	DUI	80-000889	Pineellas County SO	Florida Highway Patrol	9/11/2006	21 20 50	First Sample Volume 1 203L - OK Second Sample Volume 0 089L - Flagged Volume Not Met Third Sample Volume 0 507L - No Flag	Not A Valid Test
Lee	Scott W	3/10/1973	DUI	80-000890	Pineellas County SO	Pineellas County SO	5/20/2006	4 31 38	Second Sample Volume 1 074L - Flagged Volume Not Met Third Sample Volume 1 167L - OK	Not A Valid Test
Davis	Kevin M	5/20/1969	BUJ	80-000892	Fish & Wildlife CC	FWC	5/20/2006	14 37 25	First Sample Volume 1 058L - No Flag No Second Sample - Improper Sample	Not A Valid Test
Lopez	Torbio M	10/5/1981	DUI	80-000907	Manatee County SO	Florida Highway Patrol	8/12/2006	0 51 18	First Sample Volume 1 411L - OK Second Sample Volume 0 554L - No Flag Third Sample Volume 2 554L - OK	Valid Test - First and Third Samples



Tam	Trinity T	6/28/1977	DUI	80-000912	Manatee County SO	Manatee County SO	4/19/2006	5 10 09	First Sample Volume 0 015L - Flagged Volume Not Met Second Sample Volume 0 082L - No Flag Third Sample Volume 0 296L - Flagged Volume Not Met	Not A Valid Test
Skloroo	Richard D	6/29/1962	DUI	80-000912	Manatee County SO	Florida Highway Patrol	8/28/2006	3 01 18	First Sample Volume 0 699L - No Flag Second Sample Volume 1 339L - OK Third Sample Volume 2 125L - OK	Not A Valid Test
Hurtado	Pablo E	3/22/1967	DUI	80-000939	Lee County SO	Lee County SO	4/27/2006	2 24 56	First Sample Volume 0 289L - Flagged Volume Not Met Second Sample Volume 0 550L - No Flag Third Sample Volume 1 496L - OK	Not A Valid Test
Ferro	Lillian	4/26/1962	DUI	80-000939	Lee County SO	Lee County SO	6/12/2006	15 27 43	First Sample Volume 0 347L - No Flag Second Sample Volume 1 871L - OK Third Sample Volume 0 375L - No Flag	Not A Valid Test
Moore	Shannon C	10/25/1982	DUI	80-000939	Lee County SO	Lee County SO	7/21/2006	0 18 42	First Sample Volume 0 000L - Flagged Volume Not Met Second Sample Volume 0 000L - Flagged Volume Not Met Third Sample Volume - Subject Refused	Not A Valid Test Refusal
Conaway	Doreen M	11/15/1965	DUI	80-000939	Lee County SO	Lee County SO	7/23/2006	2 57 31	First Sample Volume 1 605L - OK Second Sample Volume 0 230L - No Flag Third Sample Volume 0 757L - No Flag	Not A Valid Test
Daughly	Amanda S	2/11/1972	DUI	80-000940	Brevard County SO	Brevard County SO	8/30/2006	3 25 19	First Sample Volume 0 613L - No Flag Second Sample Volume 1 816L - OK Third Sample Volume 2 141L - OK	Not A Valid Test
Cohen	Nikol A	6/6/1975	DUI	80-000943	Collier County SO	Naples PD	5/28/2006	3 00 50	First Sample Volume 1 316L - OK Second Sample Volume 0 410L - No Flag Third Sample Volume 1 570L - OK	Valid Test - First and Third Samples
Lyach	Vladimir	2/27/1958	DUI	80-000943	Collier County SO	Collier County SO	8/7/2006	21 20 07	First Sample Volume 0 941L - No Flag Second Sample Volume 1 417L - OK Third Sample Volume 0 75L - OK	Not A Valid Test
Hall	Deborah L	11/25/1956	DUI	80-000949	Tallahassee PD	Tallahassee PD	7/26/2006	14 18 55	First Sample Volume 0 164L - No Flag Second Sample Volume 1 417L - OK Third Sample Volume 2 507L - OK	Not A Valid Test
Bravo-Neitro	Francisco	5/8/1978	DUI	80-000951	Hendry County SO	Hendry County SO	7/30/2006	2 44 40	First Sample Volume 1 316L - OK Second Sample Volume 0 429L - No Flag Third Sample Volume 1 570L - OK	Not A Valid Test
Adams	Rodney B	6/2/1961	DUI	80-000957	Leon County SO	Leon County SO	3/30/2006	19 14 15	First Sample Volume 0 347L - No Flag Second Sample Volume 0 000L - Flagged Volume Not Met Third Sample Volume 2 382L - OK	Not A Valid Test
Tucker	Leroy	10/2/1981	DUI	80-000958	Leon County SO	Tallahassee PD	4/15/2006	22 23 00	First Sample Volume 0 429L - No Flag Second Sample Volume 2 250L - OK Third Sample Volume 0 347L - No Flag	Not A Valid Test
Furman	Shawn A	10/21/1975	DUI	80-000958	Leon County SO	Tallahassee PD	9/9/2006	1 18 33	First Sample Volume 0 675L - No Flag Second Sample Volume 0 015L - Flagged Volume Not Met Third Sample Volume 1 453L - OK	Not A Valid Test
Childs	Kendra P	9/8/1969	DUI	80-000961	Wakulla County SO	Wakulla County SO	4/14/2006	23 27 48	No Third Sample - Purge Fail First Sample Volume 0 191L - Flagged Volume Not Met Second Sample Volume 0 183L - No Flag Third Sample Volume 1 453L - OK	Not A Valid Test
Uroslegui	Marcos	2/28/1982	DUI	80-000963	Orange County SO	Orlando PD	5/17/2006	4 51 48	First Sample Volume 0 546L - No Flag Second Sample Volume 0 402L - Flagged Volume Not Met Third Sample Volume 3 941L - OK	Not A Valid Test
Altrabi	Fouad	4/2/1977	DUI	80-000963	Orange County SO	Orange County SO	5/29/2006	18 30 03	First Sample Volume 0 765L - No Flag Second Sample Volume 1 933L - OK Third Sample Volume 0 394L - No Flag	Not A Valid Test
Miranda	Arnoldo	5/6/1983	DUI	80-000964	Orange County SO	Osceola PD	9/24/2006	4 58 41	First Sample Volume 1 453L - OK Second Sample Volume 0 554L - Flagged Volume Not Met Third Sample Volume 2 027L - OK	Not A Valid Test
Rodriguez	Candido J	8/8/1980	DUI	80-000966	Osceola County Corrections	Osceola County SO	4/22/2006	3 00 40	First Sample Volume 0 230L - No Flag Second Sample Volume 0 031L - Flagged Volume Not Met Third Sample Volume 0 554L - Flagged Volume Not Met Second Control Test Outside Tolerance	Not A Valid Test
Cannon	Heather M	3/3/1987	DUI	80-000999	Seminole County SO	Seminole County SO	9/22/2006	5 13 02	First Sample Volume 0 394L - No Flag Second Sample Volume 0 429L - Flagged Volume Not Met Third Sample Volume 0 359L - Flagged Volume Not Met	Not A Valid Test
Harazin	Linda S	8/24/1947	DUI	80-001003	Pinellas County SO	Pinellas Park PD	5/4/2006	4 47 35	First Sample Volume 0 427L - No Flag Second Sample Volume 0 144L - Flagged Volume Not Met Third Sample Volume 1 250L - OK	Not A Valid Test
Chapman	Douglas R	12/25/1924	DUI	80-001004	Brevard County SO	Brevard County SO	6/25/2006	1 36 28	First Sample Volume 1 636L - OK Second Sample Volume 1 662L - No Flag Third Sample Volume 1 644L - OK	Not A Valid Test
Itby	Howard J	10/7/1962	DUI	80-001005	Pinellas County SO	Pinellas County SO	4/11/2006	0 17 44	First Sample Volume 1 644L - OK Second Sample Volume 0 710L - No Flag	Not A Valid Test
Richmond	Aubrey L	2/9/1950	DUI	80-001005	Pinellas County SO	Pinellas County SO	4/15/2006	1 26 05	First Sample Volume 0 015L - Flagged Volume Not Met Second Sample Volume 0 082L - No Flag Third Sample Volume 0 296L - Flagged Volume Not Met	Not A Valid Test

Maggis	Robert F	9/29/1938	DUI	80-001005	Pinellas County SO	Pinellas County SO	7/26/2006	20 38 58	First Sample Volume 0.574L - No Flag Second Sample Volume 0.308L - Flagged Volume Not Met Third Sample Volume 0.261L - Flagged Volume Not Met	Not A Valid Test
Roundtree	Aaron C	8/22/1971	DUI	80-001005	Pinellas County SO	Pinellas County SO	8/20/2006	3 17 46	First Sample Volume 0.527L - No Flag Second Sample Volume 0.441L - Flagged Volume Not Met Third Sample Volume 1.152L - OK	Not A Valid Test
Smith	Heather L	10/10/1987	DUI	80-001009	Palm Bay PD	Palm Bay PD	8/28/2006	4 23 50	First Sample Volume 0.429L - No Flag Second Sample Volume 1.648L - OK	Not A Valid Test
Thompson	Catherine S	2/18/1958	DUI	80-001009	Palm Bay PD	Palm Bay PD	9/11/2006	0 12 25	First Sample Volume 0.714L - Flagged Volume Not Met Second Sample Volume 0.281L - No Flag Third Sample Volume 0.785L - Flagged Volume Not Met	Not A Valid Test
Zanchez	Roberto	1/30/1981	DUI	80-001043	Highlands County SO	Florida Highway Patrol	5/21/2006	0 31 19	First Sample Volume 0.109L - No Flag No Second Sample - RFI	Not A Valid Test
Cavinder	Brenda A	9/23/1960	DUI	80-001043	Highlands County SO	Sebring PD	6/1/2006	1 56 15	First Sample Volume 1.878L - OK Second Sample Volume 0.687L - No Flag Third Sample Volume 1.402L - OK	Valid Test - First and Third Samples
Marley	Lynne C	1/10/1959	DUI	80-001043	Highlands County SO	Florida Highway Patrol	6/10/2006	0 10 08	First Sample Volume 0.382L - No Flag Second Sample Volume 1.640L - OK	Not A Valid Test
Honorat	Hernane	1/9/1956	DUI	80-001044	Coconut Creek PD	Coconut Creek PD	6/26/2006	23 34 39	First Sample Volume 1.753L - OK Second Sample Volume 0.304L - No Flag First Sample Volume 0.351L - No Flag Second Sample Volume 1.863L - OK	Not A Valid Test
Blair	Jason L	7/14/1975	DUI	80-001051	St Petersburg PD	St Petersburg PD	7/14/2006	19 14 37	No Third Sample - Subject Refused	Valid Test - Second and Third Samples
Pruett	Dennis R	12/4/1951	DUI	80-001051	St Petersburg PD	St Petersburg PD	8/18/2006	19 07 01	First Sample Volume 0.402L - No Flag Second Sample Volume 1.550L - OK Third Sample Volume 1.363L - OK	Not A Valid Test
Zuniga	Shery C	7/11/1955	DUI	80-001055	Polk County SO	Florida Highway Patrol	7/29/2006	2 07 52	First Sample Volume 0.667L - No Flag Second Sample Volume 1.503L - OK	Not A Valid Test
Addleman	Holly L	10/6/1977	DUI	80-001057	Lakeland PD	Lakeland PD	6/18/2006	3 23 44	First Sample Volume 0.300L - No Flag Second Sample Volume 1.125L - OK	Not A Valid Test
Rivera	Victor M	4/5/1950	DUI	80-001057	Lakeland PD	Lakeland PD	6/20/2006	3 12 02	First Sample Volume 0.699L - No Flag Second Sample Volume 1.656L - OK	Not A Valid Test
Godinez Perez	Alberto	1/11/1975	DUI	80-001063	Hollywood PD	Hollywood PD	7/11/2006	2 21 25	First Sample Volume 0.109L - No Flag Second Sample Volume 1.488L - OK Second Control Test Outside Tolerance	Not A Valid Test
Buchanan	Bonnie T	12/11/1940	DUI	80-001067	Broward County SO	Broward County SO	5/27/2006	19 03 55	First Sample Volume 0.546L - Flagged Volume Not Met Second Sample Volume 0.738L - No Flag Third Sample Volume 1.105L - OK	Not A Valid Test
Lanol	Jesufranc	11/17/1971	DUI	80-001067	Broward County SO	Broward County SO	7/11/2006	22 47 55	First Sample Volume 0.218L - No Flag Second Sample Volume 0.574L - OK Third Sample Volume 0.218L - No Flag	Not A Valid Test
Nelson	Stacy A	7/22/1963	DUI	80-001069	Jacksonville SO	Jacksonville SO	8/21/2006	4 30 22	First Sample Volume 1.093L - No Flag Second Sample Volume 1.957L - OK	Not A Valid Test
Doyle	Stephen V	7/24/1952	DUI	80-001069	Jacksonville SO	Jacksonville SO	9/5/2006	1 54 10	First Sample Volume 1.382L - OK Second Sample Volume 0.703L - No Flag	Not A Valid Test
Buckley	Linda G	5/10/1967	DUI	80-001071	Clay County SO	Clay County SO	5/6/2006	2 11 12	First Sample Volume 0.671L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Davis	Kim P	11/20/1957	DUI	80-001082	Plantation PD	Sunrise PD	5/9/2006	16 54 56	First Sample Volume 0.550L - No Flag Second Sample Volume 0.574L - Flagged Volume Not Met Third Sample Volume 1.457L - OK	Not A Valid Test
Nguyen	Thai H	8/2/1963	DUI	80-001083	Jacksonville SO	Jacksonville SO	5/14/2006	22 56 41	First Sample Volume 0.671L - No Flag Second Sample Volume 0.714L - Flagged Volume Not Met Third Sample Volume 0.179L - Flagged Volume Not Met	Not A Valid Test
Gagne	Audra J	12/29/1966	DUI	80-001117	Pasco County SO	Pasco County SO	4/25/2006	2 54 49	First Sample Volume 0.949L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Drake	Dawn D	6/9/1970	DUI	80-001117	Pasco County SO	Pasco County SO	4/27/2006	1 28 39	First Sample Volume 2.578L - OK Second Sample Volume 0.718L - No Flag	Not A Valid Test
Brooks	Geoffrey D	4/20/1972	DUI	80-001131	Volusia County SO	Volusia County SO	7/4/2006	2 15 13	First Sample Volume 1.015L - No Flag Second Sample Volume 1.101L - OK	Not A Valid Test
Brock	Lyn	7/20/1942	DUI	80-001137	Flagler County SO	Flagler County SO	7/22/2006	18 30 02	First Sample Volume 0.097L - Flagged Volume Not Met Second Sample Volume 0.019L - No Flag Third Sample Volume 0.511L - Flagged Volume Not Met	Not A Valid Test
Kenyon	John R	11/20/1979	DUI	80-001139	Broward County SO	Broward County SO	6/14/2006	1 48 17	First Sample Volume 1.261L - OK Second Sample Volume 0.972L - No Flag	Not A Valid Test

Jean-Jacques	Paholo	5/10/1979	DUI	80-001139	Broward County SO	Broward County SO	7/8/2006	5:35:25	First Sample Volume 0.410L - No Flag Second Sample Volume 1.519L - OK Third Sample Volume 1.750L - OK	Valid Test - Second and Third Samples
Palacios	Miriam L	1/21/1978	DUI	80-001139	Broward County SO	Broward County SO	7/22/2006	3:46:52	First Sample Volume 0.519L - No Flag Second Sample Volume 0.097L - Flagged Volume Not Met Third Sample Volume 0.414L - Flagged Volume Not Met	Not A Valid Test
Hendry	Mark J	5/31/1960	DUI	80-001139	Broward County SO	Broward County SO	9/22/2006	1:11:06	First Sample Volume 1.406L - OK Second Sample Volume 0.531L - No Flag	Not A Valid Test
Barra	Aristides E	8/21/1976	DUI	80-001153	Leon County SO	Leon County SO	4/13/2006	21:55:59	First Sample Volume 1.046L - No Flag Second Sample Volume 0.335L - Flagged Volume Not Met Third Sample Volume 1.199L - OK	Not A Valid Test
Gibson	Amelia C	7/30/1986	DUI	80-001153	Leon County SO	Leon County SO	8/23/2006	5:26:19	First Sample Volume 0.367L - No Flag Second Sample Volume 0.125L - Flagged Volume Not Met Third Sample Volume 1.199L - OK	Not A Valid Test
Waggoner	Nicholas C	5/4/1991	DUI	80-001158	Ponce Inlet PD	Daytona Beach PD	4/5/2006	23:12:41	First Sample Volume 2.066L - OK Second Sample Volume 1.058L - No Flag Third Sample Volume 1.058L - No Flag	Not A Valid Test
Currie II	Robert A	5/11/1955	DUI	80-001166	Brevard County SO	Brevard County SO	4/13/2006	20:11:17	First Sample Volume 0.828L - No Flag Second Sample Volume 2.519L - OK Third Sample Volume 2.519L - OK	Not A Valid Test
Zhang	Dong	6/24/1967	DUI	80-001190	Boynton Beach PD	Boynton Beach PD	6/17/2006	8:35:40	First Sample Volume 0.015L - Flagged Volume Not Met Second Sample Volume 2.066L - OK Third Sample Volume 0.726L - No Flag	Not A Valid Test
Ferranti	Edward J	5/26/1970	DUI	80-001190	Boynton Beach PD	Boynton Beach PD	7/22/2006	3:11:52	First Sample Volume 1.406L - OK Second Sample Volume 0.441L - No Flag	Not A Valid Test
Kaszak	Kenneth	4/16/1958	DUI	80-001190	Boynton Beach PD	Boynton Beach PD	8/13/2006	0:46:51	First Sample Volume 1.285L - OK Second Sample Volume 0.488L - No Flag Third Sample Volume 1.281L - OK	Not A Valid Test
Perez-Ortiz	Issac	10/23/1982	DUI	80-001235	West Palm Beach PD	West Palm Beach PD	4/8/2006	0:03:15	First Sample Volume 0.593L - No Flag Second Sample Volume 0.816L - No Flag Third Sample Volume 1.281L - OK	Not A Valid Test
Rucker	Curtis	3/12/1945	DUI	80-001248	Broward County SO	Broward County SO	5/7/2006	0:02:11	First Sample Volume 0.347L - No Flag Second Sample Volume 0.816L - No Flag Third Sample Volume 1.755L - OK	Not A Valid Test
Menner	Patricia	1/13/1968	DUI	80-001253	Brevard County SO	Brevard County SO	6/26/2006	22:40:39	First Sample Volume 0.468L - No Flag Second Sample Volume 1.785L - OK Third Sample Volume 1.785L - OK	Not A Valid Test
Brown	Joshua B	8/10/1976	DUI	80-001253	Brevard County SO	Brevard County SO	8/26/2006	23:45:57	First Sample Volume 0.644L - No Flag Second Sample Volume 1.242L - OK Third Sample Volume 1.242L - OK	Not A Valid Test
Doane	Richard G	10/26/1963	DUI	80-001255	Orange County SO	Orange County SO	4/7/2006	3:46:24	First Sample Volume 1.628L - OK Second Sample Volume 0.480L - No Flag Third Sample Volume 3.035L - OK	Valid Test - First and Third Samples
Apointe	Edgardo	10/9/1979	DUI	80-001255	Orange County SO	Orlando PD	9/9/2006	1:43:01	First Sample Volume 0.679L - No Flag Second Sample Volume 0.160L - Flagged Volume Not Met Third Sample - Subject Refused	Not A Valid Test Refusal
Quiros	Ovalle	12/15/1978	DUI	80-001257	Orange County SO	Orlando PD	3/30/2006	4:45:18	First Sample Volume 0.746L - No Flag Second Sample Volume 1.128L - OK Third Sample Volume 1.351L - OK	Valid Test - Second and Third Samples
Meinken	Bradford G	11/3/1976	DUI	80-001257	Orange County SO	Orlando PD	8/5/2006	4:53:30	First Sample Volume 0.339L - No Flag Second Sample Volume 1.312L - OK Third Sample Volume 1.312L - OK	Not A Valid Test
Thomas	Jean C	7/21/1953	DUI	80-001258	Orange County SO	Florida Highway Patrol	5/6/2006	21:11:50	First Sample Volume 0.468L - No Flag Second Sample Volume 1.406L - OK Third Sample Volume 1.406L - OK	Not A Valid Test
McDaniel	Joseph N	10/6/1937	DUI	80-001260	Cocoa PD	Brevard County SO	7/15/2006	1:37:50	First Sample Volume 1.164L - OK Second Sample Volume 1.039L - No Flag Third Sample Volume 0.296L - Flagged Volume Not Met	Not A Valid Test
Gause	Julia F	1/21/1969	DUI	80-001262	Melbourne PD	Melbourne PD	5/29/2006	22:12:15	First Sample Volume 0.519L - No Flag Second Sample - Subject Refused Third Sample - Subject Refused	Not A Valid Test Refusal
Young	Pamela	8/5/1958	DUI	80-001267	Palm Beach County So	Riviera Beach PD	5/30/2006	1:40:07	First Sample Volume 0.097L - Flagged Volume Not Met Second Sample Volume 1.363L - OK Third Sample Volume 0.367L - No Flag	Not A Valid Test
Keenan	Maureen E	10/9/1970	DUI	80-001267	Palm Beach County So	Palm Beach County SO	7/19/2006	1:31:08	First Sample Volume 0.972L - No Flag Second Sample Volume 0.714L - No Flag Third Sample Volume 1.226L - OK	Valid Test - First and Third Samples
Villatoro	Yener R	8/24/1984	DUI	80-001267	Palm Beach County So	Palm Beach County SO	8/6/2006	4:53:20	First Sample Volume 1.445L - OK Second Sample Volume 0.609L - No Flag Third Sample Volume 1.437L - OK	Valid Test - First and Third Samples
Suarez	Jean	3/31/1969	DUI	80-001269	Palm Beach County So	Florida Highway Patrol	5/5/2006	6:55:45	First Sample Volume 1.445L - OK Second Sample Volume 0.609L - No Flag Third Sample Volume 1.437L - OK	Valid Test - First and Third Samples
Goldberg	Paul E	3/18/1969	DUI	80-001269	Palm Beach County So	Florida Highway Patrol	8/24/2006	3:44:46	First Sample Volume 0.410L - No Flag Second Sample Volume 1.519L - OK Third Sample Volume 1.750L - OK	Valid Test - Second and Third Samples

Toft	Carole A	8/6/1964	DUI	80-001269	Palm Beach County So	Palm Beach County SO	9/21/2006	4 12 02	First Sample Volume 0 253L - No Flag Second Sample Volume 1 207L - OK	Not A Valid Test
Cornell	Daved E	2/2/1965	DUI	80-001270	Palm Beach County So	Delray Beach PD	7/23/2006	22 32 26	No Second Sample - Purge Fail	Not A Valid Test
Harri	Kareen R	6/3/1968	DUI	80-001272	Seminole County SO	Oviedo PD	4/2/2006	5 47 26	First Sample Volume 0 867L - No Flag Second Sample Volume 1 703L - OK	Not A Valid Test
Wells	Barry N	9/7/1968	DUI	80-001272	Seminole County SO	Lake Mary PD	4/10/2006	19 11 26	First Sample Volume 0 710L - No Flag Second Sample Volume 0 332L - Flagged Volume Not Met Third Sample Volume 0 785L - Flagged Volume Not Met	Not A Valid Test
Wilson	Glen M	9/15/1964	DUI	80-001272	Seminole County SO	Florida Highway Patrol	6/25/2006	2 12 47	Second Sample Volume 0 257L - Flagged Volume Not Met Third Sample Volume 0 679L - No Flag Second Control Test Outside Tolerance	Not A Valid Test
White	Leslie J	6/6/1960	DUI	80-001272	Seminole County SO	Oviedo PD	7/2/2006	2 46 00	First Sample Volume 0 457L - No Flag Second Control Test Outside Tolerance	Not A Valid Test
Newton	Justin B	10/21/1978	DUI	80-001272	Seminole County SO	Seminole County SO	7/21/2006	22 17 44	First Sample Volume 0 457L - No Flag Second Sample Volume 1 390L - OK	Not A Valid Test
Lopez Hernandez	David	4/4/1978	DUI	80-001273	Pinellas County SO	Pinellas County SO	7/15/2006	3 03 14	First Sample Volume 0 050L - No Flag Second Sample Volume 1 414L - OK Third Sample Volume 1 613L - OK	Valid Test - Second and Third Samples
Portuondo	Enik J	12/11/1975	DUI	80-001276	Miami Dade PD	Miami Dade PD	5/12/2006	0 34 06	First Sample Volume 0 109L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Fandino	Marta L	8/18/1967	DUI	80-001276	Miami Dade PD	Miami Dade PD	5/12/2006	0 34 06	First Sample Volume 0 453L - No Flag Second Sample Volume 1 839L - OK Third Sample Volume 1 292L - OK	Valid Test - Second and Third Samples
Fuhr	Alison R	10/31/1978	DUI	80-001282	University of Florida PD	University of FL PD	4/13/2006	3 34 42	First Sample Volume 0 867L - No Flag Second Sample Volume 0 152L - Flagged Volume Not Met Third Sample Volume 0 082L - Flagged Volume Not Met	Not A Valid Test
Johns	Cynthia J	5/18/1966	DUI	80-001284	Gilchrist County SO	Cedar Keys PD	6/15/2006	1 38 28	First Sample Volume 0 421L - No Flag Second Sample Volume 0 816L - No Flag Third Sample Volume 0 378L - Flagged Volume Not Met	Not A Valid Test
Campbell	Steven M	7/12/1974	DUI	80-001303	Santa Rosa County SO	Santa Rosa County SO	8/6/2006	20 30 09	First Sample Volume 0 289L - Flagged Volume Not Met Second Sample Volume 0 789L - No Flag Third Sample Volume 1 687L - OK	Not A Valid Test
Phillips	Darius K	4/9/1984	OTHER	80-001309	Bay County SO	Tyndall AFB	4/9/2006	13 22 01	First Sample Volume 0 628L - No Flag Second Sample Volume 0 578L - Flagged Volume Not Met No Third Sample - Subject Refused	Not A Valid Test Refusal
Cardwell	Curtis B	11/24/1981	DUI	80-001309	Bay County SO	Bay County SO	7/23/2006	23 35 34	First Sample Volume 0 511L - No Flag Second Sample Volume 0 000L - Flagged Volume Not Met Third Sample Volume 0 437L - Flagged Volume Not Met	Not A Valid Test
Miller	Chadrick L	4/24/1978	DUI	80-001309	Bay County SO	Parker PD	9/16/2006	22 34 35	First Sample Volume 0 437L - No Flag No Second Sample - Subject Refused	Not A Valid Test Refusal
Mings	Lonne L	6/29/1977	DUI	80-001309	Bay County SO	Florida Highway Patrol	9/23/2006	3 41 36	First Sample Volume 0 906L - No Flag Second Sample Volume 0 097L - No Flag Third Sample Volume 1 925L - OK	Valid Test - First and Third Samples
Castillo-Vazquez	Jose	12/6/1960	DUI	80-0001309	Bay County SO	Florida Highway Patrol	9/24/2006	1 50 55	First Sample Volume 0 906L - No Flag Second Sample Volume 2 613L - OK	Not A Valid Test
Clay	Susan S	12/9/1953	DUI	80-001310	Niceville PD	Crestview PD	6/12/2006	0 08 55	First Sample Volume 1 246L - OK Second Sample Volume 0 588L - No Flag Third Sample Volume 1 394L - OK	Valid Test - First and Third Samples
Diamond	Eirrede F	11/1/1936	DUI	80-001328	Indian River County SO	Sebastian PD	8/30/2006	22 47 27	First Sample Volume 0 554L - No Flag Second Sample Volume 1 003L - No Flag	Not A Valid Test
Angus	Anthony	6/11/1959	DUI	80-001328	Indian River County SO	Florida Highway Patrol	9/11/2006	8 34 49	First Sample Volume 1 972L - OK Second Sample Volume 0 156L - Flagged Volume Not Met Third Sample Volume 0 882L - No Flag	Not A Valid Test
Kellowan	Shevron S	3/20/1979	DUI	80-001344	Sarasota County SO	Florida Highway Patrol	4/1/2006	4 55 54	First Sample Volume 0 093L - Flagged Volume Not Met Second Sample Volume 3 066L - OK Third Sample Volume 0 714L - No Flag	Not A Valid Test
Ryhl	Frank L	10/27/1951	DUI	80-001344	Sarasota County SO	Sarasota County SO	4/25/2006	22 40 18	First Sample Volume 0 937L - No Flag Second Sample Volume 1 972L - OK	Not A Valid Test
Kerr	Julia P	2/1/1987	DUI	80-001344	Sarasota County SO	Sarasota County SO	5/3/2006	4 22 57	First Sample Volume 0 468L - No Flag Second Sample Volume 1 871L - OK Third Sample Volume 1 691L - OK	Not A Valid Test

Shapiro	Wayne G	3/25/1959	DUI	80-001344	Sarasota County SO	Sarasota PD	5/28/2006	3 58 01	First Sample Volume 0 656L - No Flag Second Sample Volume 1 683L - OK Third Sample Volume 1 800L - OK	Valid Test - Second and Third Samples
Quattrone	Nancy E	4/9/1957	DUI	80-001344	Sarasota County SO	Sarasota County SO	6/7/2006	0 27 30	First Sample Volume 0 324L - No Flag Second Sample Volume 1 921L - OK	Not A Valid Test
Velle	Victoreno	10/27/1979	DUI	80-001344	Sarasota County SO	Sarasota County SO	6/18/2006	3 26 40	First Sample Volume 1 625L - OK Second Sample Volume 0 636L - No Flag Third Sample Volume 2 160L - OK	Valid Test - First and Third Samples
Palmer	Scott D	7/13/1964	DUI	80-001344	Sarasota County SO	Sarasota PD	6/26/2006	1 53 50	First Sample Volume 0 492L - No Flag Second Sample Volume 0 804L - No Flag No Third Sample - Subject Refused	Not A Valid Test Refusal
Delahanty	Thomas L	5/15/1948	DUI	80-001344	Sarasota County SO	Sarasota County SO	7/8/2006	0 30 50	First Sample Volume 0 130L - No Flag Second Sample Volume 0 191L - Flagged Volume Not Met Third Sample Volume 0 378L - No Flag	Not A Valid Test
Germain	David M	4/8/1957	DUI	80-001367	Pinellas County SO	Florida Highway Patrol	7/16/2006	18 23 45	First Sample Volume 0 906L - No Flag Second Sample Volume 1 125L - OK Third Sample Volume 1 371L - OK	Valid Test - Second and Third Samples
Desmas	John R	2/1/1984	DUI	80-001415	Pinellas County SO	Pinellas Park PD	8/5/2006	3 02 07	First Sample Volume 0 130L - No Flag Second Sample Volume 0 593L - Flagged Volume Not Met Third Sample Volume 0 144L - Flagged Volume Not Met	Not A Valid Test
Dilmome	Linda J	3/13/1962	DUI	80-001416	Orange County SO	Orange County SO	6/6/2006	12 48 14	First Sample Volume 0 835L - No Flag Second Sample Volume 0 437L - Flagged Volume Not Met Third Sample Volume 0 070L - Flagged Volume Not Met	Not A Valid Test
Marinez	Noel I	9/22/1968	DUI	80-001416	Orange County SO	Orange County SO	6/10/2006	4 17 28	First Sample Volume 0 937L - No Flag Second Sample Volume 1 718L - OK	Not A Valid Test
Kars	Chad D	1/4/1967	DUI	80-001416	Orange County SO	Orlando PD	7/10/2006	1 34 39	First Sample Volume 0 042L - Flagged Volume Not Met Second Sample Volume 0 792L - No Flag	Not A Valid Test
Reyes	Arturo	9/11/1965	DUI	80-001418	Orange County SO	Florida Highway Patrol	8/30/2006	1 34 24	First Sample Volume 0 703L - No Flag Second Sample Volume 0 804L - No Flag Third Sample Volume 2 421L - OK	Not A Valid Test
Poole	Nikki K	7/17/1975	DUI	80-001418	Orange County SO	Manland PD	8/30/2006	3 38 20	First Sample Volume 0 261L - Flagged Volume Not Met Second Sample Volume 1 531L - OK	Not A Valid Test
Stephens	Patrick J	11/20/1984	DUI	80-001420	Orange County SO	Orange County SO	6/11/2006	4 28 54	First Sample Volume 2 972L - OK Second Sample Volume 0 394L - No Flag	Not A Valid Test
Buckley	Linda G	5/10/1967	DUI	80-001443	Clay County SO	Clay County SO	5/6/2006	1 46 18	First Sample Volume 1 789L - OK Second Sample Volume 1 464L - OK Third Sample Volume 0 648L - No Flag Second Control Test Outside Tolerance	Not A Valid Test

\*\* Notes are concerning the listed breath tests Additional breath tests of the referenced subject may have been conducted and are not a part of this list



Florida Department of  
Law Enforcement

Gerald M. Bailey  
Interim Commissioner

Criminal Justice  
Professionalism Program  
Alcohol Testing Program

P.O. Box 1489  
Tallahassee, Florida 32302  
(850) 410-8600  
(850) 410-7816 Fax  
<http://www.fdle.state.fl.us>

October 12, 2006

Driver's License Hearing Officer  
Florida Department of Highway Safety and Motor Vehicles  
Division of Driver's Licenses  
Bureau of Administrative Hearings

SUBJECT: Intoxilyzer 8000

Dear Driver's License Hearing Officer:

This letter is to inform you of an issue regarding the Intoxilyzer 8000 that the Florida Department of Law Enforcement Alcohol Testing Program has discovered. We have been in contact with the manufacturer of the Intoxilyzer 8000, CMI, Inc., and have determined there are missing instructions in the Intoxilyzer 8000 software version 8100.26.

The issue involves the instrument's ability to correctly identify the volume of a breath sample delivered past three minutes. The programming in software version 8100.26 was not written to address this very unique and specific scenario. This issue is affecting less than 1% of all breath tests conducted on the Intoxilyzer 8000.

A software revision to add these missing instructions is in the process of being completed and will be installed in your Intoxilyzer 8000 instruments in the near future. **I would like to assure you that software version 8100.26 is operating as written and programmed. The analytical functions of the instrument are not affected by this issue and the instrument continues to correctly analyze all breath samples.** It is only with this unique and specific scenario that the breath sample has not been determined to be reliable.

Enclosed is a summary of the affected breath tests determined as of October 2, 2006. The Alcohol Testing Program is continuing to closely monitor every breath test result obtained in the State of Florida and will contact you regarding any additional affected breath tests conducted by your agency. If you have any questions, please feel free to contact me at (850) 410-7810.

Sincerely,

Gerald M. Bailey  
Interim Commissioner

Laura D. Barfield  
Alcohol Testing Program Manager

LDB/lb

Enclosure

Committed to  
Service - Integrity - Respect - Quality

Subject Last Name	Subject First Name	Subject's Date of Birth	Offense	Instrument Serial Number	Instrument Registered To	Arresting Officer Agency	Date of Test	Time of Test	Breath Sample Volume Results	Notes**
Corredor	Miguel A.	1/7/1986	DUI	80-000223	FDLE (Loaned to Indian River County Sheriff's Office)	Florida Highway Patrol	7/6/2006	22:58:02	First Sample Volume 1.933 L - OK. Second Sample Volume 0.972L - No Flag	Not A Valid Test
Sedovic	Michael T.	11/4/1976	DUI	80-000738	Fort Walton Beach PD	Okaloosa County SO	8/5/2006	4:54:39	First Sample Volume 0.226L - Flagged Volume Not Met No Second Sample - Purge Fail	Not A Valid Test
Pelucola	Temilola A.	12/24/1981	DUI	80-000739	Okaloosa County SO	Florida Highway Patrol	8/21/2006	0:10:12	First Sample Volume 1.066L - No Flag Second Sample Volume 1.164L - OK	Not A Valid Test
Fulford	Kenneth M.	7/7/1951	DUI	80-000744	Escambia County SO	Pensacola PD	5/17/2006	2:04:10	First Sample Volume 1.550L - OK Second Sample Volume 0.605L - No Flag	Not A Valid Test
Burgos	Abel	7/11/1978	DUI	80-000744	Escambia County SO	Florida Highway Patrol	7/22/2006	3:58:14	First Sample Volume 0.684L - No Flag Second Sample Volume 2.105L - OK	Not A Valid Test
East	Easter K.	3/21/1945	DUI	80-000744	Escambia County SO	Florida Highway Patrol	8/12/2004	2:04:13	First Sample Volume 1.082L - No Flag Second Sample - Subject Refused	Not A Valid Test
Rives	Lisa C.	6/8/1959	DUI	80-000744	Escambia County SO	Florida Highway Patrol	9/28/2006	21:07:00	First Sample Volume 0.601L - No Flag Second Sample Volume 1.515L - OK	Not A Valid Test
Swanson	Amanda N.	9/9/1984	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	5/18/2006	5:39:00	First Sample Volume 0.082L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Hidle	Brandi C.	11/9/1972	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	5/25/2006	1:05:16	First Sample Volume 0.129L - Flagged Volume Not Met Second Sample Volume 0.347L - No Flag Third Sample Volume 0.127L - Flagged Volume Not Met	Not A Valid Test
Toney	Patrick S.	3/10/1987	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	5/30/2006	1:29:54	First Sample Volume 1.000L - No Flag Second Sample Volume 1.550L - OK	Not A Valid Test
Boone	Michael J.	11/9/1983	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	6/25/2006	2:30:07	First Sample Volume 0.265L - No Flag Second Sample Volume 0.054L - Flagged Volume Not Met Third Sample Volume 1.398L - OK	Not A Valid Test
Alicandri	Denielle M.	8/11/1971	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	7/25/2006	2:36:03	First Sample Volume 0.742L - No Flag Second Sample Volume 0.855L - Flag Volume Not Met Third Sample Volume 0.214L - Flag Volume Not Met	Not A Valid Test
Jones	Elaine J.	9/18/1969	DUI	80-000772	Santa Rosa County SO	Santa Rosa County SO	5/13/2006	2:38:47	First Sample Volume 0.476L - No Flag Second Sample Volume 0.652L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Bohannon	Lynette D.	12/18/1972	DUI	80-000786	Broward County SO	Broward County SO	5/12/2006	0:10:42	First Sample Volume 0.324L - Flag Volume Not Met Second Sample Volume 0.632L - No Flag Third Sample Volume 0.761L - Flag Volume Not Met	Not A Valid Test
Shelton	Linda M.	8/27/1940	DUI	80-000790	Vero Beach PD	Vero Beach PD	7/13/2006	3:05:22	First Sample Volume 0.851L - Flag Volume Not Met Second Sample Volume 1.292L - OK Third Sample Volume 0.613L - No Flag	Not A Valid Test
Vaught	Michael A.	4/20/1981	DUI	80-000794	St. Lucie County SO	St. Lucie County SO	4/9/2006	2:48:38	First Sample Volume 1.527L - OK Second Sample Volume 0.261L - No Flag Third Sample Volume 1.371L - OK	Not A Valid Test
Lesko	Darryl R.	5/10/1958	DUI	80-000794	St. Lucie County SO	St. Lucie County SO	6/5/2006	1:49:07	First Sample Volume 0.210L - No Flag Second Sample Volume 1.222L - OK Second Control Test Outside Tolerance	Not A Valid Test
Cortez	Javier O.	1/14/1964	DUI	80-000794	St. Lucie County SO	Fort Pierce PD	6/10/2006	0:36:52	First Sample Volume 1.570L - OK Second Sample Volume 0.300L - No Flag Third Sample Volume 2.429L - OK	Not A Valid Test
Schneider	Robert P.	4/23/1962	DUI	80-000794	St. Lucie County SO	Fort Pierce PD	6/19/2006	20:12:44	First Sample Volume 1.113L - OK Second Sample Volume 0.234L - No Flag Second Control Test Outside Tolerance	Not A Valid Test
Slater	Keith	6/24/1972	DUI	80-000794	St. Lucie County SO	Fort Pierce PD	6/21/2006	9:25:13	First Sample Volume 0.515L - No Flag Second Sample Volume 1.343L - OK Third Sample Volume 1.207L - OK	Not A Valid Test
Maliszewski	Dagmar E.	8/5/1987	DUI	80-000794	St. Lucie County SO	Port St. Lucie PD	6/25/2006	23:16:05	First Sample Volume 0.812L - No Flag Second Sample - Subject Refused Third Sample Volume 0.816L - No Flag	Not A Valid Test
Diaz	Alejandro	10/29/1984	DUI	80-000794	St. Lucie County SO	St. Lucie County SO	8/19/2006	3:20:08	First Sample Volume 0.812L - No Flag No Second Sample - Subject Refused	Not A Valid Test
Jimenez	David Y.	1/24/1977	DUI	80-000808	Hillsborough County SO	Hillsborough County SO	3/30/2006	23:07:03	First Sample Volume 0.816L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Pitisa	Petar J.	11/25/1982	DUI	80-000808	Hillsborough County SO	Tampa PD	7/2/2006	3:58:14	First Sample Volume 1.054L - No Flag Second Sample Volume 1.261L - OK	Not A Valid Test

Roberts	Pamela R.	7/23/1984	DUI	80-000808	Hillsborough County SO	Tampa PD	8/24/2006	0:30:17	First Sample Volume 1.070L - No Flag Second Sample Volume 1.679L - OK	Not A Valid Test
Reyes	Tony	6/11/1984	DUI	80-000808	Hillsborough County SO	Plant City PD	9/3/2006	0:04:38	First Sample Volume 0.363L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Clark	Michael C.	4/9/1983	DUI	80-000808	Hillsborough County SO	Florida Highway Patrol	9/9/2006	5:52:04	First Sample Volume 0.246L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Hargrave	Evelyn M.	7/20/1979	DUI	80-000808	Hillsborough County SO	Tampa PD	9/10/2006	3:08:59	First Sample Volume 0.925L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Diaz-Gonzalez	Raul N.	11/24/1957	DUI	80-000813	Marion County SO	Marion County SO	6/21/2006	14:35:11	First Sample Volume 2.738L - OK Second Sample Volume 0.476L - No Flag Purge Fail After Second Sample	Not A Valid Test
Mernott	Brady J.	9/9/1984	DUI	80-000817	Citrus County SO	Citrus County SO	9/14/2006	4:23:44	First Sample Volume 0.398L - No Flag Second Sample Volume 0.179L - No Flag	Not A Valid Test
Welch	Andrei C.	3/20/1966	DUI	80-000824	Leesburg PD	Leesburg PD	4/15/2006	0:19:50	First Sample Volume 0.324L - No Flag Second Sample Volume 1.117L - OK	Not A Valid Test
Garcia	Alfonso R.	8/17/1980	DUI	80-000829	Hillsborough County SO	Tampa PD	4/12/2006	2:16:24	First Sample Volume 0.339L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Nilsen	Michael R.	10/9/1961	DUI	80-000832	Martin County SO	Florida Highway Patrol	5/29/2006	3:47:51	First Sample Volume 1.660L - OK Second Sample Volume 0.972L - No Flag Third Sample Volume 1.539L - OK	Valid Test - First and Third Samples
Camacho	Rafael	4/3/1961	DUI	80-000832	Martin County SO	Stuart PD	6/22/2006	19:54:57	First Sample Volume 0.597L - No Flag Second Sample Volume 0.204L - Flagged Volume Not Met Third Sample Volume 0.219L - Flagged Volume Not Met	Not A Valid Test
Mendoza	Cecilio	11/22/1962	DUI	80-000832	Martin County SO	Martin County SO	7/9/2006	2:12:46	First Sample Volume 0.000L - Flagged Volume Not Met Second Sample Volume 2.226L - OK Third Sample Volume 0.183 - No Flag	Not A Valid Test
Schroeder	Brandon H.	7/28/1987	DUI	80-000832	Martin County SO	Martin County SO	8/13/2006	2:45:47	First Sample Volume 0.468L - Flagged Volume Not Met Second Sample Volume 0.769L - No Flag Third Sample Volume 1.480L - OK	Not A Valid Test
Daher	Jabib	2/3/1948	DUI	80-000832	Martin County SO	FWC West Palm	8/26/2006	23:30:17	First Sample Volume 1.726L - OK Second Sample Volume 0.449L - No Flag	Not A Valid Test
Vandermark	Jeffrey W.	6/18/1945	DUI	80-000832	Martin County SO	Sewall's Point PD	9/4/2006	2:04:32	First Sample Volume 0.808L - No Flag Second Sample Volume 3.394L - OK	Not A Valid Test
Michale	James M.	2/27/1967	DUI	80-000832	Martin County SO	Stuart PD	9/4/2006	21:48:28	First Sample Volume 0.203L - No Flag Second Sample Volume 1.691L - OK Third Sample Volume 1.187L - OK Second Control Test Outside Tolerance	Not A Valid Test
Moretti	Derek P.	6/2/1980	DUI	80-000833	Hillsborough County SO	Tampa PD	5/15/2006	1:09:00	First Sample Volume 0.742L - No Flag Second Sample Volume 0.186L - Flag Volume Not Met No Third Sample - Subject Refused	Not A Valid Test Refusal
Perez	Palermo	11/20/1972	DUI	80-000833	Hillsborough County SO	Hillsborough County SO	6/10/2006	19:51:30	First Sample Volume 0.105L - No Flag Second Sample Volume 0.212L - Flagged Volume Not Met Third Sample Volume 0.219L - Flagged Volume Not Met	Not A Valid Test
Pospischil	Ashley N.	5/10/1983	DUI	80-000833	Hillsborough County SO	Tampa PD	7/3/2006	4:13:57	First Sample Volume 0.957L - No Flag Second Sample Volume 1.648L - OK	Not A Valid Test
Gandanilla	Petronilo R.	9/6/1972	DUI	80-000833	Hillsborough County SO	Plant City PD	7/21/2006	8:12:20	First Sample Volume 0.371L - No Flag Second Sample - Subject Refused	Not A Valid Test Refusal
Rojas	Oscar	8/27/1949	DUI	80-000833	Hillsborough County SO	Tampa PD	8/11/2006	2:46:47	First Sample Volume 0.625L - No Flag Second Sample - Subject Refused	Not A Valid Test Refusal
Rolon	Jose A.	3/8/1961	DUI	80-000833	Hillsborough County SO	Florida Highway Patrol	8/13/2006	4:32:45	First Sample Volume 0.574L - No Flag Second Sample Volume 1.289L - OK	Not A Valid Test
Seymour	Patrick H.	8/26/1967	DUI	80-000859	Cape Coral PD	Cape Coral PD	6/21/2006	20:35:16	First Sample Volume 0.429L - No Flag Second Sample Volume 0.194L - Flagged Volume Not Met Third Sample Volume 0.625L - No Flag	Not A Valid Test
Lisa	Denise A.	12/9/1957	DUI	80-000859	Cape Coral PD	Cape Coral PD	7/10/2006	0:54:05	First Sample Volume 0.464L - No Flag Second Sample Volume 0.168L - Flagged Volume Not Met Third Sample Volume 0.562L - No Flag	Not A Valid Test
Hyland	Cara L.	8/23/1986	DUI	80-000859	Cape Coral PD	Cape Coral PD	8/2/2006	2:37:55	First Sample Volume 1.035L - Flagged Volume Not Met Second Sample Volume 0.980L - No Flag Third Sample Volume 0.015L - Flagged Volume Not Met	Not A Valid Test
Velez	Ernesto	7/22/1966	DUI	80-000859	Cape Coral PD	Cape Coral PD	8/6/2006	1:33:11	First Sample Volume 0.097L - Flagged Volume Not Met Second Sample Volume 0.457L - No Flag Third Sample Volume 0.687L - Flagged Volume Not Met Second Control Test Outside Tolerance	Not A Valid Test Refusal
Jarrett	Donald	12/28/1955	DUI	80-000861	Sarasota County SO	Sarasota County SO	9/12/2006	21:54:31	First Sample Volume 0.035L - No Flag No Second Sample - Subject Refused	Not A Valid Test Refusal



Brown	Nathan	5/1/1979	DUI	80-000868	Monroe County SO	NAS Key West	7/30/2006	2:11:35	First Sample Volume 0.250L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Moniere	Laurent L.	6/24/1947	BUI	80-000869	Monroe County SO	Other State	4/15/2006	20:19:14	First Sample Volume 0.480L - No Flag Second Sample Volume 0.179L - Flagged Volume Not Met Third Sample Volume 1.804L - OK Second Control Test Outside Tolerance	Not A Valid Test
Clarke	Brian D.	12/18/1961	BUI	80-000869	Monroe County SO	FWC	5/20/2006	21:23:30	First Sample Volume 0.687L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Comejo	Jose	4/17/1980	DUI	80-000869	Monroe County SO	Monroe County SO	7/22/2006	4:48:56	First Sample Volume 1.078L - No Flag Second Sample Volume 1.246L - OK	Not A Valid Test
Kelly	Shane L.	2/6/1981	DUI	80-000869	Monroe County SO	Monroe County SO	8/7/2006	5:16:12	First Sample Volume 0.449L - Flagged Volume Not Met Second Sample Volume 0.535L - No Flag Third Sample Volume 1.585L - OK	Not A Valid Test
La Paz	Gilda	11/20/1943	DUI	80-000871	Coral Gables PD	Coral Gables PD	6/4/2006	2:28:46	Second Sample Volume 0.149L - Flagged Volume Not Met Third Sample Volume 1.394L - OK Second Control Test Outside Tolerance	Not A Valid Test
Costa Moran	Enrique A.	2/27/1975	DUI	80-000872	Miami PD	Miami PD	8/25/2006	3:03:40	First Sample Volume 1.578L - OK Second Sample Volume 0.328L - No Flag	Not A Valid Test
Mena	Hollman E.	12/5/1984	DUI	80-000873	Miami PD	Miami PD	9/11/2006	9:26:14	First Sample Volume 0.808L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Mena	Hollman E.	12/5/1984	DUI	80-000872	Miami PD	Miami PD	9/11/2006	9:52:59	First Sample Volume 1.078L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Diaz	Gerardo	7/30/1929	DUI	80-000873	Miami PD	Miami PD	4/1/2006	22:22:46	First Sample Volume 0.535L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Knowles	Patrick J.	5/28/1966	DUI	80-000873	Miami PD	South Miami PD	9/18/2006	1:55:54	First Sample Volume 1.101L - OK Second Sample Volume 1.027L - No Flag	Not A Valid Test
Alonso	Ruben	11/22/1959	DUI	80-000876	Hialeah PD	Hialeah PD	5/11/2006	20:40:29	First Sample Volume 0.312L - No Flag Second Sample Volume 1.105L - OK	Not A Valid Test
Franco	Luis A.	9/30/1967	DUI	80-000876	Hialeah PD	Hialeah PD	6/26/2006	21:54:09	First Sample Volume 0.718L - No Flag Second Sample Volume 1.246L - OK Third Sample Volume 1.148L - OK Second Control Test Outside Tolerance	Not A Valid Test
Bello	Farah T.	10/14/1978	DUI	80-000879	Miami Beach PD	Miami Beach PD	9/16/2006	5:27:26	First Sample Volume 0.687L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Lozano	Cesar U.	10/7/1975	DUI	80-000881	Miami Dade PD	Miami Dade PD	9/20/2006	2:09:52	First Sample Volume 0.304L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Martinez	Gliver O.	9/2/1984	DUI	80-000884	Miami Dade PD	Miami Dade PD	4/8/2006	4:38:12	First Sample Volume 2.617L - OK Second Sample Volume 0.460L - No Flag	Not A Valid Test
Burgos-Donates	Myrna	8/9/1957	DUI	80-000885	Miami Dade PD	Miami Dade PD	5/6/2006	2:43:50	First Sample Volume 0.273L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Velasquez	Valerie A.	3/1/1981	DUI	80-000885	Miami Dade PD	Miami Dade PD	6/10/2006	5:27:19	First Sample Volume 0.433L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Mena	Marianne	10/28/1983	DUI	80-000885	Miami Dade PD	Miami Dade PD	8/20/2006	6:53:05	First Sample Volume 0.339L - No Flag Second Sample Volume 0.132L - Flagged Volume Not Met Third Sample Volume 0.976L - Flagged Volume Not Met	Not A Valid Test
Mena	Marianne	10/28/1983	DUI	80-000885	Miami Dade PD	Miami Dade PD	8/20/2006	7:13:49	First Sample Volume 0.117L - Flagged Volume Not Met Second Sample Volume 0.500L - No Flag Third Sample Volume 0.871L - Flagged Volume Not Met	Not A Valid Test
Winston	Alan	4/17/1942	DUI	80-000886	North Miami Beach PD	North Miami Beach PD	4/22/2006	5:40:59	First Sample Volume 0.632L - No Flag Second Sample Volume 0.597L - No Flag	Not A Valid Test
Ghazanan	Sevak	9/9/1984	DUI	80-000888	Pinellas County SO	Pinellas County SO	6/2/2006	3:28:14	First Sample Volume 0.226L - No Flag Second Sample Volume 0.304L - Flagged Volume Not Met Third Sample Volume 0.101L - Flagged Volume Not Met	Not A Valid Test
Harvey	Raybeth	8/22/1959	DUI	80-000889	Pinellas County SO	Florida Highway Patrol	9/1/2006	21:20:50	First Sample Volume 1.203L - OK Second Sample Volume 0.089L - Flagged Volume Not Met Third Sample Volume 0.507L - No Flag	Not A Valid Test
Lee	Scott W.	3/10/1973	DUI	80-000890	Pinellas County SO	Pinellas County SO	5/20/2006	4:31:38	First Sample Volume 0.777L - No Flag Second Sample Volume 1.074L - Flagged Volume Not Met Third Sample Volume 1.167L - OK	Not A Valid Test
Davis	Kevin M.	5/20/1969	BUI	80-000892	Fish & Wildlife CC	FWC	5/20/2006	14:37:25	First Sample Volume 1.058L - No Flag No Second Sample - Improper Sample	Not A Valid Test
Lopez	Toribio M.	10/5/1981	DUI	80-000907	Manatee County SO	Florida Highway Patrol	8/12/2006	0:51:18	First Sample Volume 1.441L - OK Second Sample Volume 0.554L - No Flag Third Sample Volume 2.554L - OK	Valid Test - First and Third Samples

Tamm	Trinity T.	6/28/1977	DUI	80-000912	Manatee County SO	Manatee County SO	Manatee County SO	4/19/2006	5:10:09	First Sample Volume 0.015L - Flagged Volume Not Met Second Sample Volume 0.082L - No Flag Third Sample Volume 0.286L - Flagged Volume Not Met	Not A Valid Test
Skoliro	Richard D.	6/29/1962	DUI	80-000912	Manatee County SO	Manatee County SO	Florida Highway Patrol	8/28/2006	3:01:18	First Sample Volume 0.699L - No Flag Second Sample Volume 1.339L - OK	Not A Valid Test
Hurtado	Pablo E.	3/22/1967	DUI	80-000939	Lee County SO	Lee County SO	Lee County SO	4/27/2006	2:24:56	First Sample Volume 2.125L - OK Second Sample Volume 0.550L - No Flag	Not A Valid Test
Ferro	Lillian	4/28/1962	DUI	80-000939	Lee County SO	Lee County SO	Lee County SO	6/12/2006	15:27:43	First Sample Volume 0.289L - Flagged Volume Not Met Second Sample Volume 0.835L - No Flag Third Sample Volume 1.496L - OK	Not A Valid Test
Moore	Shannon C.	10/25/1982	DUI	80-000939	Lee County SO	Lee County SO	Lee County SO	7/21/2006	0:18:42	First Sample Volume 0.347L - No Flag Second Sample Volume 1.871L - OK	Not A Valid Test
Conaway	Doreen M.	11/15/1965	DUI	80-000939	Lee County SO	Lee County SO	Lee County SO	7/23/2006	2:57:31	Second Sample Volume 0.000L - Flagged Volume Not Met No Third Sample - Subject Refused	Not A Valid Test
Daughtry	Amanda S.	2/11/1972	DUI	80-000940	Brevard County SO	Brevard County SO	Brevard County SO	8/30/2006	3:25:19	First Sample Volume 1.605L - OK Second Sample Volume 0.230L - No Flag Third Sample Volume 0.757L - No Flag	Not A Valid Test
Cohen	Nikol A.	6/6/1975	DUI	80-000943	Collier County SO	Collier County SO	Naples PD	5/28/2006	3:00:50	First Sample Volume 0.613L - No Flag Second Sample Volume 1.816L - OK	Not A Valid Test
Lyach	Vladimir	2/27/1958	DUI	80-000943	Collier County SO	Collier County SO	Collier County SO	8/7/2006	21:20:07	First Sample Volume 2.144L - OK Second Sample Volume 0.218L - No Flag Third Sample Volume 1.875L - OK	Not A Valid Test
Hall	Deborah L.	11/25/1956	DUI	80-000949	Tallahassee PD	Tallahassee PD	Tallahassee PD	7/26/2006	14:18:55	Second Control Test Outside Tolerance	Not A Valid Test
Bravo-Neiro	Francisco	5/8/1978	DUI	80-000951	Hendry County SO	Hendry County SO	Hendry County SO	7/30/2006	2:44:40	First Sample Volume 0.941L - No Flag Second Sample Volume - 1.417L - OK	Not A Valid Test
Adams	Rodney B.	6/2/1981	DUI	80-000957	Leon County SO	Leon County SO	Leon County SO	3/30/2006	19:14:15	First Sample Volume 0.164L - No Flag Second Sample Volume 2.507L - OK	Not A Valid Test
Tucker	Leroy	10/2/1981	DUI	80-000958	Leon County SO	Leon County SO	Tallahassee PD	4/15/2006	22:23:00	First Sample Volume 1.316L - OK Second Sample Volume 0.410L - No Flag Third Sample Volume 1.570L - OK	Valid Test - First and Third Samples
Furman	Shawn A.	10/21/1975	DUI	80-000958	Leon County SO	Leon County SO	Tallahassee PD	9/9/2006	1:18:33	First Sample Volume 0.429L - No Flag Second Sample Volume 2.250L - OK	Not A Valid Test
Childs	Kendra P.	9/8/1969	DUI	80-000961	Wakulla County SO	Wakulla County SO	Wakulla County SO	4/14/2006	23:27:48	First Sample Volume 0.675L - No Flag Second Sample Volume 0.015L - Flagged Volume Not Met No Third Sample - Purge Fail	Not A Valid Test
Unostegui	Marcos	2/28/1982	DUI	80-000963	Orange County SO	Orange County SO	Orlando PD	5/17/2006	4:51:48	First Sample Volume 0.191L - Flagged Volume Not Met Second Sample Volume 0.183L - No Flag Third Sample Volume 1.453L - OK	Not A Valid Test
Altrabi	Fouad	4/2/1977	DUI	80-000963	Orange County SO	Orange County SO	Orange County SO	5/29/2006	18:30:03	First Sample Volume 0.546L - No Flag Second Sample Volume 0.402L - Flagged Volume Not Met Third Sample Volume 3.941L - OK	Not A Valid Test
Miranda	Arnoldo	5/6/1983	DUI	80-000964	Orange County SO	Orange County SO	Ocoee PD	9/24/2006	4:58:41	First Sample Volume 0.765L - No Flag Second Sample Volume 1.933L - OK	Not A Valid Test
Rodriguez	Candido J.	8/8/1980	DUI	80-000966	Osceola County Corrections	Osceola County Corrections	Osceola County SO	4/22/2006	3:00:40	First Sample Volume 1.453L - OK Second Sample Volume 0.457L - No Flag Third Sample Volume 2.027L - OK	Not A Valid Test
Cannon	Heather M.	3/3/1987	DUI	80-000999	Seminole County SO	Seminole County SO	Seminole County SO	9/22/2006	5:13:02	Second Control Test Outside Tolerance	Not A Valid Test
Harazin	Linda S.	8/24/1947	DUI	80-001003	Pinellas County SO	Pinellas County SO	Pinellas Park PD	5/4/2006	4:47:35	First Sample Volume 0.230L - No Flag Second Sample Volume 0.031L - Flagged Volume Not Met Third Sample Volume 0.554L - Flagged Volume Not Met	Not A Valid Test
Chapman	Douglas R.	12/25/1924	DUI	80-001004	Brevard County SO	Brevard County SO	Brevard County SO	6/25/2006	1:36:28	First Sample Volume 0.394L - No Flag Second Sample Volume 0.429L - Flagged Volume Not Met Third Sample Volume 0.359L - Flagged Volume Not Met	Not A Valid Test
Itby	Howard J.	10/7/1962	DUI	80-001005	Pinellas County SO	Pinellas County SO	Pinellas County SO	4/11/2006	0:17:44	First Sample Volume 1.636L - OK Second Sample Volume 1.062L - No Flag	Not A Valid Test
Richmond	Aubrey L.	2/9/1950	DUI	80-001005	Pinellas County SO	Pinellas County SO	Pinellas County SO	4/15/2006	1:26:05	First Sample Volume 1.844L - OK Second Sample Volume 0.710L - No Flag	Not A Valid Test

Maggs	Robert F.	9/29/1938	DUI	80-001005	Pinellas County SO	Pinellas County SO	Pinellas County SO	7/26/2006	20:38:58	First Sample Volume 0.574L - No Flag Second Sample Volume 0.308L - Flagged Volume Not Met Third Sample Volume 0.281L - Flagged Volume Not Met	Not A Valid Test
Roundtree	Aaron C.	8/22/1971	DUI	80-001005	Pinellas County SO	Pinellas County SO	Pinellas County SO	8/20/2006	3:17:46	First Sample Volume 0.527L - No Flag Second Sample Volume 0.441L - Flagged Volume Not Met Third Sample Volume 1.152L - OK	Not A Valid Test
Smith	Heather L.	10/10/1987	DUI	80-001009	Palm Bay PD	Palm Bay PD	Palm Bay PD	8/28/2006	4:23:50	First Sample Volume 0.429L - No Flag Second Sample Volume 1.648L - OK	Not A Valid Test
Thompson	Catherine S.	2/18/1958	DUI	80-001009	Palm Bay PD	Palm Bay PD	Palm Bay PD	9/11/2006	0:12:25	First Sample Volume 0.714L - Flagged Volume Not Met Second Sample Volume 0.281L - No Flag Third Sample Volume 0.785L - Flagged Volume Not Met	Not A Valid Test
Zarchoz	Roberto	1/30/1981	DUI	80-001043	Highlands County SO	Highlands County SO	Florida Highway Patrol	5/2/2006	0:31:19	No Second Sample - RFI	Not A Valid Test
Cavinder	Brenda A.	9/23/1960	DUI	80-001043	Highlands County SO	Highlands County SO	Sebring PD	6/1/2006	1:56:15	First Sample Volume 1.878L - OK Second Sample Volume 0.687L - No Flag Third Sample Volume 1.402L - OK	Valid Test - First and Third Samples
Marley	Lynne C.	1/10/1959	DUI	80-001043	Highlands County SO	Highlands County SO	Florida Highway Patrol	6/10/2006	0:10:08	First Sample Volume 0.382L - No Flag Second Sample Volume 1.640L - OK	Not A Valid Test
Honorat	Hermane	1/9/1956	DUI	80-001044	Coconut Creek PD	Coconut Creek PD	Coconut Creek PD	6/26/2006	23:34:39	First Sample Volume 1.753L - OK Second Sample Volume 0.304L - No Flag	Not A Valid Test
Blair	Jason L.	7/14/1975	DUI	80-001051	St. Petersburg PD	St. Petersburg PD	St. Petersburg PD	7/4/2006	19:14:37	First Sample Volume 0.351L - No Flag Second Sample Volume 1.863L - OK No Third Sample - Subject Refused	Not A Valid Test
Pruett	Dennis R.	12/4/1951	DUI	80-001051	St. Petersburg PD	St. Petersburg PD	St. Petersburg PD	8/18/2006	19:07:01	First Sample Volume 0.402L - No Flag Second Sample Volume 1.550L - OK Third Sample Volume 1.363L - OK	Valid Test - Second and Third Samples
Zuniga	Sherry C.	7/11/1955	DUI	80-001055	Polk County SO	Polk County SO	Florida Highway Patrol	7/29/2006	2:07:52	First Sample Volume 0.667L - No Flag Second Sample Volume 1.503L - OK	Not A Valid Test
Adelman	Holly L.	10/6/1977	DUI	80-001057	Lakeland PD	Lakeland PD	Lakeland PD	6/18/2006	3:23:44	First Sample Volume 0.300L - No Flag Second Sample Volume 1.125L - OK	Not A Valid Test
Rivera	Victor M.	4/5/1950	DUI	80-001057	Lakeland PD	Lakeland PD	Lakeland PD	6/20/2006	3:12:02	First Sample Volume 0.688L - No Flag Second Sample Volume 1.656L - OK	Not A Valid Test
Godinez Perez	Alberto	1/11/1975	DUI	80-001063	Hollywood PD	Hollywood PD	Hollywood PD	7/1/2006	2:21:25	First Sample Volume 0.109L - No Flag Second Sample Volume 1.488L - OK Second Control Test Outside Tolerance	Not A Valid Test
Buchanan	Bonnie T.	12/11/1940	DUI	80-001067	Broward County SO	Broward County SO	Broward County SO	5/27/2006	19:03:55	First Sample Volume 0.546L - Flagged Volume Not Met Second Sample Volume 0.738L - No Flag Third Sample Volume 1.105L - OK	Not A Valid Test
Lanot	Jesufranc	11/17/1971	DUI	80-001067	Broward County SO	Broward County SO	Broward County SO	7/1/2006	22:47:55	First Sample Volume 0.718L - Flagged Volume Not Met Second Sample Volume 1.574L - OK Third Sample Volume 0.218L - No Flag	Not A Valid Test
Nelson	Stacy A.	7/22/1963	DUI	80-001069	Jacksonville SO	Jacksonville SO	Jacksonville SO	8/21/2006	4:30:22	First Sample Volume 1.093L - No Flag Second Sample Volume 1.957L - OK	Not A Valid Test
Doyle	Stephen V.	7/24/1952	DUI	80-001069	Jacksonville SO	Jacksonville SO	Jacksonville SO	9/5/2006	1:54:10	First Sample Volume 1.382L - OK Second Sample Volume 0.703L - No Flag	Not A Valid Test
Buckley	Linda G.	5/10/1967	DUI	80-001071	Clay County SO	Clay County SO	Clay County SO	5/6/2006	2:11:12	First Sample Volume 0.671L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Davis	Kim P.	11/20/1957	DUI	80-001082	Plantation PD	Plantation PD	Sunrise PD	5/9/2006	16:54:56	First Sample Volume 0.550L - No Flag Second Sample Volume 0.574L - Flagged Volume Not Met Third Sample Volume 1.457L - OK	Not A Valid Test
Nguyen	Thai H.	8/2/1963	DUI	80-001083	Jacksonville SO	Jacksonville SO	Jacksonville SO	5/14/2006	22:56:41	First Sample Volume 0.671L - No Flag Second Sample Volume 0.714L - Flagged Volume Not Met Third Sample Volume 0.179L - Flagged Volume Not Met	Not A Valid Test
Gagne	Audra J.	12/29/1966	DUI	80-001117	Pasco County SO	Pasco County SO	Pasco County SO	4/25/2006	2:54:49	First Sample Volume 0.949L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Drake	Dawn D.	6/9/1970	DUI	80-001117	Pasco County SO	Pasco County SO	Pasco County SO	4/27/2006	1:28:39	First Sample Volume 2.578L - OK Second Sample Volume 0.718L - No Flag	Not A Valid Test
Brooks	Geoffrey D.	4/20/1972	DUI	80-001131	Volusia County SO	Volusia County SO	Volusia County SO	7/4/2006	2:15:13	First Sample Volume 1.015L - No Flag Second Sample Volume 1.101L - OK	Not A Valid Test
Brock	Lyn	7/20/1942	DUI	80-001137	Flagler County SO	Flagler County SO	Flagler County SO	7/22/2006	18:30:02	First Sample Volume 0.097L - Flagged Volume Not Met Second Sample Volume 0.019L - No Flag Third Sample Volume 0.511L - Flagged Volume Not Met	Not A Valid Test
Kenyon	John R.	11/20/1979	DUI	80-001139	Broward County SO	Broward County SO	Broward County SO	6/14/2006	1:48:17	First Sample Volume 1.261L - OK Second Sample Volume 0.972L - No Flag	Not A Valid Test

Jean-Jacques	Patolo	5/10/1979	DUI	80-001139	Broward County SO	Broward County SO	7/8/2006	5:35:25	First Sample Volume 0.410L - No Flag Second Sample Volume 1.519L - OK Third Sample Volume 1.750L - OK	Valid Test - Second and Third Samples
Palacios	Mirian L.	1/21/1978	DUI	80-001139	Broward County SO	Broward County SO	7/22/2006	3:46:52	First Sample Volume 0.519L - No Flag Second Sample Volume 0.097L - Flagged Volume Not Met Third Sample Volume 0.414L - Flagged Volume Not Met	Not A Valid Test
Hendry	Mark J.	5/31/1960	DUI	80-001139	Broward County SO	Broward County SO	9/22/2006	1:11:06	First Sample Volume 1.406L - OK Second Sample Volume 0.531L - No Flag	Not A Valid Test
Barrera	Anistides E.	8/21/1976	DUI	80-001153	Leon County SO	Leon County SO	4/13/2006	21:55:59	First Sample Volume 0.421L - No Flag Second Sample Volume 0.191L - No Flag	Not A Valid Test
Gibson	Amelia C.	7/30/1986	DUI	80-001153	Leon County SO	Leon County SO	8/23/2006	5:26:19	First Sample Volume 1.046L - No Flag Second Sample Volume 0.335L - Flagged Volume Not Met Third Sample Volume 1.199L - OK	Not A Valid Test
Waggoner	Nicholas C.	5/4/1981	DUI	80-001158	Ponce Inlet PD	Daytona Beach PD	4/5/2006	23:12:41	First Sample Volume 0.367L - No Flag Second Breath Sample 0.125L - Flagged Volume Not Met No Third Sample - Sequence Aborted	Not A Valid Test
Currie II	Robert A.	5/11/1955	DUI	80-001166	Brevard County SO	Brevard County SO	4/13/2006	20:11:17	First Sample Volume 2.066L - OK Second Sample Volume 1.058L - No Flag	Not A Valid Test
Zhang	Dong	6/24/1967	DUI	80-001190	Boynton Beach PD	Boynton Beach PD	6/17/2006	8:35:40	First Sample Volume 0.828L - No Flag Second Sample Volume 2.519L - OK	Not A Valid Test
Ferranti	Edward J.	5/26/1970	DUI	80-001190	Boynton Beach PD	Boynton Beach PD	7/22/2006	3:11:52	First Sample Volume 0.015L - Flagged Volume Not Met Second Sample Volume 2.068L - OK Third Sample Volume 0.726L - No Flag	Not A Valid Test
Kaszak	Kenneth	4/16/1958	DUI	80-001190	Boynton Beach PD	Boynton Beach PD	8/13/2006	0:46:51	First Sample Volume 1.406L - OK Second Sample Volume 0.441L - No Flag	Not A Valid Test
Perez-Ortiz	Isaac	10/23/1982	DUI	80-001235	West Palm Beach PD	West Palm Beach PD	4/8/2006	0:03:15	First Sample Volume 1.285L - OK Second Sample Volume 0.488L - No Flag Third Sample Volume 1.281L - OK	Valid Test - First and Third Samples
Rucker	Curtis	3/12/1945	DUI	80-001248	Broward County SO	Broward County SO	5/7/2006	0:02:11	First Sample Volume 0.593L - No Flag Second Sample Volume 0.816L - No Flag	Not A Valid Test
Menner	Patricia	11/3/1968	DUI	80-001253	Brevard County SO	Brevard County SO	6/26/2006	22:40:39	First Sample Volume 0.347L - No Flag Second Sample Volume 1.765L - OK	Not A Valid Test
Brown	Joshua B.	8/10/1976	DUI	80-001253	Brevard County SO	Brevard County SO	8/26/2006	23:45:57	First Sample Volume 0.468L - No Flag Second Sample Volume 1.785L - OK	Not A Valid Test
Doane	Richard G.	10/26/1963	DUI	80-001255	Orange County SO	Orange County SO	4/7/2006	3:46:24	First Sample Volume 0.644L - No Flag Second Sample Volume 1.242L - OK	Not A Valid Test
Aponite	Edgardo	10/9/1979	DUI	80-001255	Orange County SO	Orlando PD	9/9/2006	1:43:01	First Sample Volume 0.679L - No Flag Second Sample Volume 0.160L - Flagged Volume Not Met No Third Sample - Subject Refused	Not A Valid Test Refusal
Quiros	Ovalle	12/15/1978	DUI	80-001257	Orange County SO	Orlando PD	3/30/2006	4:45:18	First Sample Volume 1.628L - OK Second Sample Volume 0.480L - No Flag Third Sample Volume 3.035L - OK	Valid Test - First and Third Samples
Meinken	Bradford G.	11/3/1976	DUI	80-001257	Orange County SO	Orlando PD	8/5/2006	4:53:30	First Sample Volume 0.746L - No Flag Second Sample Volume 1.128L - OK Third Sample Volume 1.351L - OK	Valid Test - Second and Third Samples
Thomas	Jean C.	7/21/1953	DUI	80-001258	Orange County SO	Florida Highway Patrol	5/6/2006	21:11:50	First Sample Volume 0.339L - No Flag Second Sample Volume 1.312L - OK	Not A Valid Test
McDaniel	Joseph N.	10/6/1937	DUI	80-001260	Cocoa PD	Brevard County SO	7/15/2006	1:37:50	First Sample Volume 0.468L - No Flag Second Sample Volume 1.408L - OK	Not A Valid Test
Gause	Julia F.	1/21/1969	DUI	80-001262	Meibourne PD	Meibourne PD	5/29/2006	22:12:15	First Sample Volume 1.164L - OK Second Sample Volume 1.039L - No Flag	Not A Valid Test
Young	Pamela	8/5/1958	DUI	80-001267	Palm Beach County So	Riviera Beach PD	5/30/2006	1:40:07	First Sample Volume 0.296L - Flagged Volume Not Met Second Sample Volume 0.519L - No Flag No Third Sample - Subject Refused	Not A Valid Test Refusal
Keenan	Maureen E.	10/9/1970	DUI	80-001267	Palm Beach County So	Palm Beach County So	7/19/2006	1:31:08	First Sample Volume 0.087L - Flagged Volume Not Met Second Sample Volume 1.363L - OK Third Sample Volume 0.367L - No Flag	Not A Valid Test Refusal
Villatoro	Yener R.	8/24/1984	DUI	80-001267	Palm Beach County So	Palm Beach County So	8/6/2006	4:53:20	First Sample Volume 0.972L - No Flag No Second Sample - Subject Refused	Not A Valid Test Refusal
Suarez	Jean	3/31/1969	DUI	80-001269	Palm Beach County So	Florida Highway Patrol	5/5/2006	6:55:45	First Sample Volume 1.226L - OK Second Sample Volume 0.714L - No Flag Third Sample Volume 1.769L - OK	Valid Test - First and Third Samples
Goldberg	Paul E.	3/18/1969	DUI	80-001269	Palm Beach County So	Florida Highway Patrol	8/24/2006	3:44:46	First Sample Volume 1.445L - OK Second Sample Volume 0.609L - No Flag Third Sample Volume 1.437L - OK	Valid Test - First and Third Samples

Toft	Carole A.	8/6/1964	DUI	80-001269	Palm Beach County So	Palm Beach County So	9/21/2006	4:12:02	First Sample Volume 0.253L - No Flag Second Sample Volume 1.207L - OK	Not A Valid Test
Cornell	Davied E.	2/2/1965	DUI	80-001270	Palm Beach County So	Palm Beach County So	7/23/2006	22:32:26	First Sample Volume 0.613L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Harri	Kareen R.	8/3/1968	DUI	80-001272	Seminole County SO	Oviedo PD	4/2/2006	5:47:26	First Sample Volume 0.867L - No Flag Second Sample Volume 1.703L - OK	Not A Valid Test
Wells	Bary N.	9/7/1968	DUI	80-001272	Seminole County SO	Lake Mary PD	4/10/2006	19:11:26	First Sample Volume 0.710L - No Flag Second Sample Volume 0.332L - Flagged Volume Not Met Third Sample Volume 0.785L - Flagged Volume Not Met	Not A Valid Test
Wilson	Glen M.	9/15/1964	DUI	80-001272	Seminole County SO	Florida Highway Patrol	6/25/2006	2:12:47	First Sample Volume 0.675L - No Flag Second Sample Volume 0.796L - OK Second Control Test Outside Tolerance	Not A Valid Test
White	Leslie J.	6/8/1960	DUI	80-001272	Seminole County SO	Oviedo PD	7/2/2006	2:46:00	Second Sample Volume 0.257L - Flagged Volume Not Met Third Sample Volume 0.679L - No Flag Second Control Test Outside Tolerance	Not A Valid Test
Newton	Justin B.	10/21/1978	DUI	80-001272	Seminole County SO	Seminole County SO	7/21/2006	22:17:44	First Sample Volume 0.457L - No Flag Second Sample Volume 1.390L - OK	Not A Valid Test
Lopez-Hernandez	David	4/4/1978	DUI	80-001273	Pinellas County SO	Pinellas County SO	7/15/2006	3:03:14	First Sample Volume 0.050L - No Flag Second Sample Volume 1.414L - OK Third Sample Volume 1.613L - OK	Valid Test - Second and Third Samples
Portuondo	Erik J.	12/11/1975	DUI	80-001276	Miami Dade PD	Miami Dade PD	5/12/2006	0:34:06	First Sample Volume 0.108L - No Flag No Second Sample - Purge Fail	Not A Valid Test
Fandino	Marta L.	8/18/1967	DUI	80-001276	Miami Dade PD	Miami Dade PD	5/27/2006	4:12:02	First Sample Volume 0.453L - No Flag Second Sample Volume 1.839L - OK Third Sample Volume - 1.292L - OK	Valid Test - Second and Third Samples
Fuhr	Alison R.	10/31/1978	DUI	80-001282	University of Florida PD	University of FL PD	4/13/2006	3:34:42	First Sample Volume 0.867L - No Flag Second Sample Volume 0.152L - Flagged Volume Not Met Third Sample Volume 0.082L - Flagged Volume Not Met	Not A Valid Test
Johns	Cynthia J.	5/18/1966	DUI	80-001284	Gilchrist County SO	Cedar Keys PD	6/15/2006	1:38:28	First Sample Volume 0.421L - No Flag Second Sample Volume 0.378L - Flagged Volume Not Met Third Sample Volume 0.816L - No Flag	Not A Valid Test
Campbell	Steven M.	7/12/1974	DUI	80-001303	Santa Rosa County SO	Santa Rosa County SO	8/6/2006	20:30:09	First Sample Volume 0.289L - Flagged Volume Not Met Second Sample Volume 0.789L - No Flag Third Sample Volume 1.687L - OK	Not A Valid Test
Phillips	Darius K.	4/9/1984	OTHER	80-001309	Bay County SO	Tyndall AFB	4/9/2006	13:22:01	First Sample Volume 0.628L - No Flag Second Sample Volume 0.578L - Flagged Volume Not Met No Third Sample - Subject Refused	Not A Valid Test
Cardwell	Curtis B.	1/24/1981	DUI	80-001309	Bay County SO	Bay County SO	7/23/2006	23:35:34	First Sample Volume 0.511L - No Flag Second Sample Volume 0.000L - Flagged Volume Not Met Third Sample Volume 0.437L - Flagged Volume Not Met	Not A Valid Test
Miller	Chadrick L.	4/24/1978	DUI	80-001309	Bay County SO	Parker PD	9/16/2006	22:34:35	First Sample Volume 0.437L - No Flag No Second Sample - Subject Refused	Not A Valid Test
Mings	Lonnie L.	6/29/1977	DUI	80-001309	Bay County SO	Florida Highway Patrol	9/23/2006	3:41:36	First Sample Volume 2.710L - OK Second Sample Volume 0.097L - No Flag Third Sample Volume 1.925L - OK	Valid Test - First and Third Samples
Castillo-Vazquez	Jose	12/6/1960	DUI	80-0001309	Bay County SO	Florida Highway Patrol	9/24/2006	1:50:55	First Sample Volume 0.906L - No Flag Second Sample Volume 2.613L - OK	Not A Valid Test
Clay	Susan S.	12/9/1953	DUI	80-001310	Niceville PD	Crestview PD	6/12/2006	0:08:55	First Sample Volume 1.246L - OK Second Sample Volume 0.558L - No Flag Third Sample Volume 1.394L - OK	Valid Test - First and Third Samples
Diamond	Elfriede F.	11/1/1936	DUI	80-001328	Indian River County SO	Sebastian PD	8/30/2006	22:47:27	First Sample Volume 0.554L - No Flag Second Sample Volume 1.003L - No Flag	Not A Valid Test
Angus	Anthony	6/11/1958	DUI	80-001328	Indian River County SO	Florida Highway Patrol	9/11/2006	8:34:49	First Sample Volume 1.972L - OK Second Sample Volume 0.156L - Flagged Volume Not Met Third Sample Volume 0.882L - No Flag	Not A Valid Test
Kellowan	Shevron S.	3/20/1979	DUI	80-001344	Sarasota County SO	Florida Highway Patrol	4/1/2006	4:55:54	First Sample Volume 0.093L - Flagged Volume Not Met Second Sample Volume 3.066L - OK Third Sample Volume 0.714L - No Flag	Not A Valid Test
Ryll	Frank L.	10/27/1951	DUI	80-001344	Sarasota County SO	Sarasota County SO	4/25/2006	22:40:18	First Sample Volume 0.937L - No Flag Second Sample Volume 1.972L - OK	Not A Valid Test
Kerr	Julia P.	2/11/1987	DUI	80-001344	Sarasota County SO	Sarasota County SO	5/3/2006	4:22:57	First Sample Volume 0.468L - No Flag Second Sample Volume 1.871L - OK Third Sample Volume 1.691L - OK	Not A Valid Test

Shapiro	Wayne G.	3/25/1959	DUI		80-001344	Sarasota County SO	Sarasota PD	5/28/2006	3:58:01	First Sample Volume 0.656L - No Flag Second Sample Volume 1.683L - OK Third Sample Volume 1.800L - OK	Valid Test - Second and Third Samples
Quattrone	Nancy E.	4/9/1957	DUI		80-001344	Sarasota County SO	Sarasota County SO	6/7/2006	0:27:30	First Sample Volume 0.324L - No Flag Second Sample Volume 1.921L - OK	Not A Valid Test Valid Test - First and Third Samples
Velle	Victoreno	10/27/1979	DUI		80-001344	Sarasota County SO	Sarasota County SO	6/18/2006	3:26:40	First Sample Volume 1.625L - OK Second Sample Volume 0.636L - No Flag Third Sample Volume 2.160L - OK	Not A Valid Test Valid Test - First and Third Samples
Palmer	Scott D.	7/13/1964	DUI		80-001344	Sarasota County SO	Sarasota PD	6/26/2006	1:53:50	First Sample Volume 1.101L - OK Second Sample Volume 0.804L - No Flag No Third Sample - Subject Refused	Not A Valid Test Refusal
Delahanty	Thomas L.	5/15/1948	DUI		80-001344	Sarasota County SO	Sarasota County SO	7/8/2006	0:30:50	First Sample Volume 0.492L - No Flag Second Sample Volume 0.191L - Flagged Volume Not Met Third Sample Volume 0.378L - No Flag	Not A Valid Test
Germain	David M.	4/8/1957	DUI		80-001367	Pinellas County SO	Florida Highway Patrol	7/16/2006	18:23:45	First Sample Volume 0.906L - No Flag Second Sample Volume 1.125L - OK Third Sample Volume 1.371L - OK	Valid Test - Second and Third Samples
Desimas	John R.	2/1/1984	DUI		80-001415	Pinellas County SO	Pinellas Park PD	8/5/2006	3:02:07	First Sample Volume 0.130L - No Flag Second Sample Volume 0.593L - Flagged Volume Not Met Third Sample Volume 0.144L - Flagged Volume Not Met	Not A Valid Test
Dilimone	Linda J.	3/13/1962	DUI		80-001416	Orange County SO	Orange County SO	6/6/2006	12:48:14	First Sample Volume 0.835L - No Flag Second Sample Volume 0.437L - Flagged Volume Not Met Third Sample Volume 0.070L - Flagged Volume Not Met	Not A Valid Test
Martinez	Noel I.	9/22/1968	DUI		80-001416	Orange County SO	Orange County SO	6/10/2006	4:17:28	First Sample Volume 0.937L - No Flag Second Sample Volume 1.718L - OK	Not A Valid Test
Kars	Chad D.	1/4/1967	DUI		80-001416	Orange County SO	Orlando PD	7/10/2006	1:34:38	First Sample Volume 1.308L - OK Second Sample Volume 0.792L - No Flag	Not A Valid Test
Reyes	Arturo	9/11/1965	DUI		80-001418	Orange County SO	Florida Highway Patrol	8/30/2006	1:34:24	First Sample Volume 0.042L - Flagged Volume Not Met Second Sample Volume 0.804L - No Flag Third Sample Volume 2.421L - OK	Not A Valid Test
Poole	Nikki K.	7/17/1975	DUI		80-001418	Orange County SO	Maitland PD	8/30/2006	3:38:20	First Sample Volume 0.703L - No Flag Second Sample Volume 0.261L - Flagged Volume Not Met Third Sample Volume 1.531L - OK	Not A Valid Test
Stephens	Patrick J.	11/20/1984	DUI		80-001420	Orange County SO	Orange County SO	6/11/2006	4:28:54	First Sample Volume 2.972L - OK Second Sample Volume 0.394L - No Flag	Not A Valid Test
Buckley	Linda G.	5/10/1967	DUI		80-001443	Clay County SO	Clay County SO	5/8/2006	1:46:18	First Sample Volume 1.789L - OK Second Sample Volume 1.464L - OK Third Sample Volume 0.648L - No Flag Second Control Test Outside Tolerance	Not A Valid Test

\*\* Notes are concerning the listed breath tests. Additional breath tests of the referenced subject may have been conducted and are not a part of this list.



Florida Department of  
Law Enforcement

Gerald M. Bailey  
*Commissioner*

**Criminal Justice  
Professionalism Program  
Alcohol Testing Program**

P.O. Box 1489  
Tallahassee, Florida 32302  
(850) 410-8600  
(850) 410-7816 Fax  
<http://www.fdle.state.fl.us>

December 7, 2006

NAME  
ADDRESS  
ADDRESS  
ADDRESS

SUBJECT: Intoxilyzer 8000

Dear NAME:

In a letter dated October 5, 2006, you were made aware of an issue regarding the Intoxilyzer 8000 with software version 8100.26. The issue involved the instrument's ability to correctly identify the volume of a breath sample delivered past three minutes due to missing instructions in the software programming.

Since that time, all Intoxilyzer 8000 evidentiary breath test instruments have been upgraded to version 8100.27. This software revision was evaluated by the Florida Department of Law Enforcement on October 9 and 10, 2006 in accordance with Chapter 11D-8, Florida Administrative Code. The software revision upgrade to all Intoxilyzer 8000 evidentiary breath test instruments began on October 11, 2006 and was completed on November 20, 2006.

I have attached a list of all affected breath tests. This list includes tests that you were previously notified of as well as any additional tests since the October 5, 2006 notice.

If you have any questions, please feel free to contact me at (850) 410-7810.

Sincerely,

Gerald M. Bailey  
Commissioner

Laura D. Barfield  
Alcohol Testing Program Manager

LDB/lb

Enclosure

**Final List**

Subject Last Name	Subject First Name	Subject's Date of Birth	Offense	Instrument Serial Number	Circuit	Instrument Registered To	Arresting Officer Agency	Date of Test	Time of Test	Breath Sample Volume Results	Notes**
Corredor	Miguel A.	1/7/1986	DUI	80-000223	Old	FDLE (Loaned to Indian River County Sheriff's Office)	Florida Highway Patrol	7/6/2006	22:58:02	First Sample Volume 1.933 L - OK; Second Sample Volume 0.972L - No Flag	Not A Valid Test - Second Sample Not Flagged
Cruz	Santiago G.	12/29/1965	DUI	80-000224	New 1st	FDLE (Loaned to Okaloosa County Sheriff's Office)	Okaloosa County SO	10/1/2006	4:40:04	First Sample Volume 1.808L - OK Second Sample Volume 0.050L - Flagged Volume Not Met Third Sample Volume - 0.582L - No Flag	Not A Valid Test Third Sample Not Flagged
Sedovic	Michael T.	11/4/1976	DUI	80-000738	Old 1st	Fort Walton Beach PD	Okaloosa County SO	8/5/2006	4:54:39	First Sample Volume 0.226L - Flagged Volume Not Met No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Peiuola	Temilola A.	12/24/1961	DUI	80-000739	Old 1st	Okaloosa County SO	Florida Highway Patrol	8/21/2006	0:10:12	First Sample Volume 1.066L - No Flag Second Sample Volume 1.164L - OK	Not A Valid Test First Sample Not Flagged
Fulford	Kenneth M.	7/7/1951	DUI	80-000744	Old 1st	Escambia County SO	Pensacola PD	5/17/2006	2:04:10	First Sample Volume 1.550L - OK Second Sample Volume 0.605L - No Flag	Not A Valid Test Sample Not Flagged
Burgos	Abel	7/11/1978	DUI	80-000744	Old 1st	Escambia County SO	Florida Highway Patrol	7/22/2006	3:58:14	First Sample Volume 0.664L - No Flag Second Sample Volume 2.105L - OK	Not A Valid Test First Sample Not Flagged
East	Easter K.	3/21/1945	DUI	80-000744	Old 1st	Escambia County SO	Florida Highway Patrol	8/12/2004	2:04:13	First Sample Volume 1.082L - No Flag Second Sample - Subject Refused	Not A Valid Test Refusal
Rives	Lisa C.	6/8/1959	DUI	80-000744	Old 1st	Escambia County SO	Florida Highway Patrol	9/28/2006	21:07:00	First Sample Volume 0.601L - No Flag Second Sample Volume 1.515L - OK	Not A Valid Test First Sample Not Flagged
Meshell	Robert J.	7/12/1981	DUI	80-000752	New 5th	Marion County SO	Ocala PD	10/16/2006	19:51:26	First Sample Volume 0.074L - Flagged Volume Not Met Second Sample Volume 0.566L - No Flag	Not A Valid Test - Need At Least Two Samples
Swanson	Amanda N.	9/9/1984	DUI	80-000763	Old 1st	Okaloosa County SO	Okaloosa County SO	5/18/2006	5:39:00	First Sample Volume 0.062L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Hidle	Brandi C.	11/9/1972	DUI	80-000763	Old 1st	Okaloosa County SO	Okaloosa County SO	5/25/2006	1:05:16	First Sample Volume 0.129L - Flagged Volume Not Met Second Sample Volume 0.347L - No Flag Third Sample Volume 0.127L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Toney	Patrick S.	3/10/1987	DUI	80-000763	Old 1st	Okaloosa County SO	Okaloosa County SO	5/30/2006	1:29:54	First Sample Volume 1.000L - No Flag Second Sample Volume 1.550L - OK	Not A Valid Test First Sample Not Flagged
Boone	Michael J.	11/9/1983	DUI	80-000763	Old 1st	Okaloosa County SO	Okaloosa County SO	6/25/2006	2:30:07	First Sample Volume 0.265L - No Flag Second Sample Volume 0.054L - Flagged Volume Not Met Third Sample Volume 1.398L - OK	Not A Valid Test First Sample Not Flagged
Alicandri	Danielle M.	8/11/1971	DUI	80-000763	Old 1st	Okaloosa County SO	Okaloosa County SO	7/25/2006	2:36:03	First Sample Volume 0.742L - No Flag Second Sample Volume 0.855L - Flag Volume Not Met Third Sample Volume 0.214L - Flag Volume Not Met	Not A Valid Test - Need At Least Two Samples
Jones	Elaine J.	9/18/1969	DUI	80-000772	Old 1st	Santa Rosa County SO	Santa Rosa County SO	5/13/2006	2:38:47	First Sample Volume 0.476L - No Flag Second Sample Volume 1.738L - OK	Not A Valid Test First Sample Not Flagged
Woods	Christopher L.	5/6/1946	DUI	80-000772	New 1st	Santa Rosa County SO	Florida Highway Patrol	9/27/2006	20:29:23	First Sample Volume 0.730L - No Flag Second Sample Volume 1.964L - OK Third Sample Volume 2.000L - OK	Valid Test - Second and Third Samples
Bohannon	Lynette D.	12/18/1972	DUI	80-000786	Old 17th	Broward County SO	Broward County SO	5/12/2006	0:10:42	First Sample Volume 0.652L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Figueredo	Kristina M.	8/2/1985	DUI	80-000786	New 17th	Broward County SO	Other State	10/15/2006	6:03:02	First Sample Volume 0.140L - No Flag Second Sample Volume 1.460L - OK	Not A Valid Test - First Sample Not Flagged
Manella	Thomas	3/28/1951	DUI	80-000786	New 17th	Broward County SO	Broward County SO	10/17/2006	0:58:36	First Sample Volume 0.750L - No Flag Second Sample Volume 1.167L - OK Third Sample Volume - 1.226L - OK	Valid Test - Second and Third Samples
Johnson	Patrick H.	5/2/1951	DUI	80-000786	New 17th	Broward County SO	Broward County SO	10/18/2006	0:06:45	First Sample Volume 0.480L - No Flag Second Sample Volume 1.769L - OK	Not A Valid Test - First Sample Not Flagged
Shelton	Linda M.	8/27/1940	DUI	80-000790	Old 19th	Vero Beach PD	Vero Beach PD	7/13/2006	3:05:22	First Sample Volume 0.324L - Flag Volume Not Met Second Sample Volume 0.632L - No Flag Third Sample Volume 0.761L - Flag Volume Not Met	Not A Valid Test - Need At Least Two Samples
Vaught	Michael A.	4/20/1981	DUI	80-000794	Old 19th	St. Lucie County SO	St. Lucie County SO	4/9/2006	2:49:38	First Sample Volume 0.851L - Flag Volume Not Met Second Sample Volume 1.292L - OK Third Sample Volume 0.613L - No Flag	Not A Valid Test - Third Sample Not Flagged
Lesko	Darryl R.	5/10/1958	DUI	80-000794	Old 19th	St. Lucie County SO	St. Lucie County SO	6/5/2006	1:49:07	First Sample Volume 1.527L - OK Second Sample Volume 0.261L - No Flag Third Sample Volume 1.371L - OK	Not A Valid Test - Second Control Test Outside Tolerance
Cortez	Javier O.	1/14/1964	DUI	80-000794	Old 19th	St. Lucie County SO	Fort Pierce PD	6/10/2006	0:36:52	First Sample Volume 0.210L - No Flag Second Sample Volume 1.222L - OK Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance
Scheider	Robert P.	4/23/1962	DUI	80-000794	Old 19th	St. Lucie County SO	Fort Pierce PD	6/19/2006	20:12:44	First Sample Volume 1.570L - OK Second Sample Volume 0.300L - No Flag Third Sample Volume 2.429L - OK Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance



Slater	Keith	6/24/1972	DUI	80-000794	Old	19th	St. Lucie County SO	Fort Pierce PD	6/21/2006	9:25:13	First Sample Volume 1.113L - OK Second Sample Volume 0.234L - No Flag Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance
Maliszewski	Dagmar E.	8/5/1967	DUI	80-000794	Old	19th	St. Lucie County SO	Port St. Lucie PD	6/25/2006	23:16:05	First Sample Volume 0.515L - No Flag Second Sample Volume 1.343L - OK Third Sample Volume 1.207L - OK	Not A Valid Test - Second Control Test Outside Tolerance
Diaz	Alejandro	10/29/1984	DUI	80-000794	Old	19th	St. Lucie County SO	St. Lucie County SO	8/19/2006	3:20:08	First Sample Volume 0.812L - No Flag No Second Sample - Subject Refused	Not A Valid Test
Jimenez	David Y.	1/24/1977	DUI	80-000808	Old	13th	Hillsborough County SO	Hillsborough County SO	3/30/2006	23:07:03	First Sample Volume 0.816L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Pitisa	Petar J.	11/25/1982	DUI	80-000808	Old	13th	Hillsborough County SO	Tampa PD	7/2/2006	3:58:14	First Sample Volume 1.054L - No Flag Second Sample Volume 1.261L - OK	Not A Valid Test
Roberts	Pamela R.	7/23/1984	DUI	80-000808	Old	13th	Hillsborough County SO	Tampa PD	8/24/2006	0:30:17	First Sample Volume 1.070L - No Flag Second Sample Volume 1.679L - OK	First Sample Not Flagged
Reyes	Tony	6/11/1984	DUI	80-000808	Old	13th	Hillsborough County SO	Plant City PD	9/3/2006	0:04:38	First Sample Volume 0.363L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Clark	Michael C.	4/9/1983	DUI	80-000808	Old	13th	Hillsborough County SO	Florida Highway Patrol	9/9/2006	5:52:04	First Sample Volume 0.246L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Hargrave	Evelyn M.	7/20/1979	DUI	80-000808	Old	13th	Hillsborough County SO	Tampa PD	9/10/2006	3:08:59	First Sample Volume 0.925L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Ballestas	Nicholas E.	12/17/1950	DUI	80-000808	New	13th	Hillsborough County SO	Hillsborough County SO	10/8/2006	21:58:24	First Sample Volume 0.332L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Diaz-Gonzalez	Raul N.	11/24/1957	DUI	80-000813	Old	5th	Marion County SO	Marion County SO	6/21/2006	14:35:11	First Sample Volume 2.738L - OK Second Sample Volume 0.476L - No Flag Purge Fail After Second Sample	Not A Valid Test - Need At Least Two Samples
Merriott	Brady J.	9/9/1984	DUI	80-000817	Old	5th	Citrus County SO	Citrus County SO	9/14/2006	4:23:44	First Sample Volume 0.398L - No Flag Second Sample Volume 0.179L - No Flag	Not A Valid Test - First and Second Sample Not Flagged
Welch	Andrei C.	3/20/1966	DUI	80-000824	Old	5th	Leesburg PD	Leesburg PD	4/15/2006	0:19:50	First Sample Volume 0.324L - No Flag Second Sample Volume 1.117L - OK	Not A Valid Test
Garcia	Alfonso R.	8/17/1980	DUI	80-000829	Old	13th	Hillsborough County SO	Tampa PD	4/12/2006	2:16:24	First Sample Volume 0.339L - No Flag No Second Sample - Purge Fail	First Sample Not Flagged
Nilsen	Michael R.	10/9/1961	DUI	80-000832	Old	19th	Martin County SO	Florida Highway Patrol	5/29/2006	3:47:51	First Sample Volume 1.660L - OK Second Sample Volume 0.972L - No Flag Third Sample Volume 1.539L - OK	Valid Test - First and Third Samples
Camacho	Rafael	4/3/1961	DUI	80-000832	Old	19th	Martin County SO	Stuart PD	6/22/2006	19:54:57	First Sample Volume 0.597L - No Flag Second Sample Volume 0.204L - Flagged Volume Not Met Third Sample Volume 0.219L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Mendoza	Cecilio	11/22/1962	DUI	80-000832	Old	19th	Martin County SO	Martin County SO	7/9/2006	2:12:46	First Sample Volume 0.000L - Flagged Volume Not Met Second Sample Volume 2.226L - OK	Not A Valid Test - Third Sample Not Flagged
Schroeder	Brandon H.	7/28/1987	DUI	80-000832	Old	19th	Martin County SO	Martin County SO	8/13/2006	2:45:47	First Sample Volume 0.468L - Flagged Volume Not Met Second Sample Volume 0.769L - No Flag Third Sample Volume 1.480L - OK	Not A Valid Test - Second Sample Not Flagged
McCelland	Patricia P.	8/25/1950	DUI	80-000832	New	19th	Martin County SO	Martin County SO	8/25/2006	21:47:45	First Sample Volume 0.152L - No Flag No Second Sample - Sequence Aborted	Not A Valid Test - Need At Least Two Samples
Daher	Jabib	2/3/1948	DUI	80-000832	Old	19th	Martin County SO	FWC West Palm	8/26/2006	23:30:17	First Sample Volume 1.726L - OK	Not A Valid Test - Second Sample Not Flagged
Vandermark	Jeffrey W.	6/16/1945	DUI	80-000832	Old	19th	Martin County SO	Sewall's Point PD	9/4/2006	2:04:32	Second Sample Volume 0.449L - No Flag Second Sample Volume 3.394L - OK	Not A Valid Test - First Sample Not Flagged
Michale	James M.	2/27/1967	DUI	80-000832	Old	19th	Martin County SO	Stuart PD	9/4/2006	21:48:28	First Sample Volume 0.203L - No Flag Second Sample Volume 1.691L - OK Third Sample Volume 1.187L - OK	Not A Valid Test - Second Control Test Outside Tolerance
Moretti	Derek P.	6/2/1980	DUI	80-000833	Old	13th	Hillsborough County SO	Tampa PD	5/15/2006	1:09:00	First Sample Volume 0.742L - No Flag Second Sample Volume 0.196L - Flag Volume Not Met No Third Sample - Subject Refused	Not A Valid Test
Perez	Palermo	11/20/1972	DUI	80-000833	Old	13th	Hillsborough County SO	Hillsborough County SO	6/10/2006	19:51:30	First Sample Volume 0.105L - No Flag Second Sample Volume 0.212L - Flagged Volume Not Met Third Sample Volume 0.219L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Pospischil	Ashley N.	5/10/1983	DUI	80-000833	Old	13th	Hillsborough County SO	Tampa PD	7/3/2006	4:13:57	First Sample Volume 0.957L - No Flag Second Sample Volume 1.648L - OK	Not A Valid Test
Gandarilla	Petronilo R.	9/6/1972	DUI	80-000833	Old	13th	Hillsborough County SO	Plant City PD	7/21/2006	9:12:20	First Sample Volume 0.371L - No Flag Second Sample - Subject Refused	First Sample Not Flagged
Rojas	Oscar	8/27/1949	DUI	80-000833	Old	13th	Hillsborough County SO	Tampa PD	8/11/2006	2:46:47	First Sample Volume 0.625L - No Flag Second Sample - Subject Refused	Not A Valid Test
Rolon	Jose A.	3/8/1961	DUI	80-000833	Old	13th	Hillsborough County SO	Florida Highway Patrol	8/13/2006	4:32:45	First Sample Volume 0.574L - No Flag Second Sample Volume 1.289L - OK	Not A Valid Test

Fussell	Christopher N	7/19/1980	DUI	80-000833	New	13th	Hillsborough County SO	Hillsborough County SO	10/7/2006	3:46:57	First Sample Volume 0.289L - No Flag Second Sample Volume 2.910L - OK Third Sample Volume - 3.050L - OK	Valid Test - Second and Third Samples
Seymour	Patrick H.	8/26/1967	DUI	80-000859	Old	20th	Cape Coral PD	Cape Coral PD	6/21/2006	20:35:16	First Sample Volume 0.429L - No Flag Second Sample Volume 0.194L - Flagged Volume Not Met Third Sample Volume 0.625L - No Flag	Not A Valid Test - First and Third Samples Not Flagged
Lisa	Denise A.	12/9/1957	DUI	80-000859	Old	20th	Cape Coral PD	Cape Coral PD	7/10/2006	0:54:05	First Sample Volume 0.464L - No Flag Second Sample Volume 0.168L - Flagged Volume Not Met Third Sample Volume 0.582L - No Flag	Not A Valid Test - First and Third Samples Not Flagged
Hyland	Cara L.	8/23/1966	DUI	80-000859	Old	20th	Cape Coral PD	Cape Coral PD	8/2/2006	2:37:55	First Sample Volume 1.035L - Flagged Volume Not Met Second Sample Volume 0.980L - No Flag Third Sample Volume 0.015L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Velez	Ernesto	7/22/1966	DUI	80-000859	Old	20th	Cape Coral PD	Cape Coral PD	8/6/2006	1:33:11	First Sample Volume 0.097L - Flagged Volume Not Met Second Sample Volume 0.457L - No Flag Third Sample Volume 0.687L - Flagged Volume Not Met Second Control Test Outside Tolerance	Not A Valid Test - Need At Least Two Samples and Second Control Test Outside Tolerance
Jarrett	Donald	12/28/1955	DUI	80-000861	Old	12th	Sarasota County SO	Sarasota County SO	9/12/2006	21:54:31	First Sample Volume 0.035L - No Flag No Second Sample - Subject Refused	Not A Valid Test - Need At Least Two Samples
Brown	Nathan	5/1/1979	DUI	80-000868	Old	16th	Monroe County SO	NAS Key West	7/30/2006	2:11:35	First Sample Volume 0.250L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Moriniere	Laurent L.	6/24/1947	BUI	80-000869	Old	16th	Monroe County SO	Other State	4/15/2006	20:19:14	First Sample Volume 0.480L - No Flag Second Sample Volume 0.179L - Flagged Volume Not Met Third Sample Volume 1.804L - OK	Not A Valid Test - Second Control Test Outside Tolerance
Clarke	Brian D.	12/18/1961	BUI	80-000869	Old	16th	Monroe County SO	FWC	5/20/2006	21:23:30	First Sample Volume 0.687L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Cornejo	Jose	4/17/1980	DUI	80-000869	Old	16th	Monroe County SO	Monroe County SO	7/22/2006	4:48:56	First Sample Volume 1.078L - No Flag Second Sample Volume 1.246L - OK	Not A Valid Test
Kelly	Shane L.	2/6/1981	DUI	80-000869	Old	16th	Monroe County SO	Monroe County SO	8/7/2006	5:16:12	First Sample Volume 0.449L - Flagged Volume Not Met Second Sample Volume 0.535L - No Flag Third Sample Volume 1.585L - OK	Not A Valid Test - Second Sample Not Flagged
La Paz	Gilda	11/20/1943	DUI	80-000871	Old	11th	Coral Gables PD	Coral Gables PD	6/4/2006	2:28:46	Second Sample Volume 0.149L - Flagged Volume Not Met Third Sample Volume 1.394L - OK Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance
Montenegro	Maria H.	5/6/1945	DUI	80-000871	New	11th	Coral Gables PD	Coral Gables PD	10/12/2006	10:30:25	First Sample Volume 0.375L - No Flag Second Sample Volume 1.929L - OK	Not A Valid Test - First Sample Not Flagged
Costa Moran	Enrique A.	2/27/1975	DUI	80-000872	Old	11th	Miami PD	Miami PD	8/25/2006	3:03:40	First Sample Volume 1.578L - OK Second Sample Volume 0.328L - No Flag	Not A Valid Test - Second Sample Not Flagged
Mena	Hollman E.	12/5/1984	DUI	80-000872	Old	11th	Miami PD	Miami PD	9/11/2006	9:52:59	First Sample Volume 1.078L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Nunez	Manuel J.	1/24/1892	DUI	80-000872	New	11th	Miami PD	Miami PD	10/7/2006	8:03:03	First Sample Volume 1.179L - OK Second Sample Volume 0.945L - No Flag	Not A Valid Test - Second Sample Not Flagged
Mena	Hollman E.	12/5/1984	DUI	80-000873	Old	11th	Miami PD	Miami PD	9/11/2006	9:26:14	First Sample Volume 0.808L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Diaz	Gerardo	7/30/1929	DUI	80-000873	Old	11th	Miami PD	Miami PD	4/1/2006	22:22:46	First Sample Volume 0.535L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Knowles	Patrick J.	5/28/1966	DUI	80-000873	Old	11th	Miami PD	South Miami PD	9/16/2006	1:55:54	First Sample Volume 1.101L - OK Second Sample Volume 1.027L - No Flag	Not A Valid Test - Second Sample Not Flagged
Alonso	Ruben	11/22/1959	DUI	80-000876	Old	11th	Hialeah PD	Hialeah PD	5/1/2006	20:40:29	First Sample Volume 0.312L - No Flag Second Sample Volume 1.105L - OK	Not A Valid Test - First Sample Not Flagged
Franco	Luis A.	9/30/1967	DUI	80-000876	Old	11th	Hialeah PD	Hialeah PD	6/26/2006	21:54:09	First Sample Volume 0.718L - No Flag Second Sample Volume 1.246L - OK Third Sample Volume 1.148L - OK	Not A Valid Test - Second Control Test Outside Tolerance
Bello	Farah T.	10/14/1978	DUI	80-000879	Old	11th	Miami Beach PD	Miami Beach PD	9/16/2006	5:27:26	First Sample Volume 0.687L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Lozano	Cesar U.	10/7/1975	DUI	80-000881	Old	11th	Miami Dade PD	Miami Dade PD	9/20/2006	2:09:52	First Sample Volume 0.304L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Martinez	Gilver O.	9/2/1984	DUI	80-000884	Old	11th	Miami Dade PD	Miami Dade PD	4/9/2006	4:38:12	First Sample Volume 2.617L - OK Second Sample Volume 0.460L - No Flag	Not A Valid Test - Second Sample Not Flagged
Burgos-Donates	Myrna	8/9/1957	DUI	80-000885	Old	11th	Miami Dade PD	Miami Dade PD	5/6/2006	2:43:50	First Sample Volume 0.273L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Velasquez	Valerie A.	3/1/1981	DUI	80-000885	Old	11th	Miami Dade PD	Miami Dade PD	6/10/2006	5:27:19	First Sample Volume 0.433L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Mena	Marianne	10/28/1983	DUI	80-000885	Old	11th	Miami Dade PD	Miami Dade PD	8/20/2006	6:53:05	First Sample Volume 0.339L - No Flag Second Sample Volume 0.132L - Flagged Volume Not Met Third Sample Volume 0.976L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Mena	Marianne	10/28/1983	DUI	80-000885	Old	11th	Miami Dade PD	Miami Dade PD	8/20/2006	7:13:49	First Sample Volume 0.117L - Flagged Volume Not Met Second Sample Volume 0.500L - No Flag Third Sample Volume 0.871L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples

Winston	Alan	4/17/1942	DUI	80-000886	Old	11th	North Miami Beach PD	North Miami Beach PD	4/22/2006	5:40:59	First Sample Volume 0.632L - No Flag Second Sample Volume 0.597L - No Flag	Not A Valid Test - First and Second Sample Not Flagged
Ghazarian	Sevak	9/9/1984	DUI	80-000888	Old	6th	Pinellas County SO	Pinellas County SO	6/2/2006	3:28:14	First Sample Volume 0.226L - No Flag Second Sample Volume 0.304L - Flagged Volume Not Met Third Sample Volume 0.101L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Harvey	Raybeth	8/22/1959	DUI	80-000889	Old	6th	Pinellas County SO	Florida Highway Patrol	9/1/2006	21:20:50	First Sample Volume 1.203L - OK Second Sample Volume 0.089L - Flagged Volume Not Met Third Sample Volume 0.507L - No Flag	Not A Valid Test - Third Sample Not Flagged
Lee	Scott W.	3/10/1973	DUI	80-000890	Old	6th	Pinellas County SO	Pinellas County SO	5/20/2006	4:31:38	First Sample Volume 0.777L - No Flag Second Sample Volume 1.074L - Flagged Volume Not Met Third Sample Volume 1.167L - OK	Not A Valid Test - First Sample Not Flagged
Rainbow	Rosa S.	3/16/1964	DUI	80-000890	New	6th	Pinellas County SO	Florida Highway Patrol	10/15/2006	23:59:55	First Sample Volume 0.714L - No Flag Second Sample Volume 1.125L - OK	Not A Valid Test - First Sample Not Flagged
Davis	Kevin M.	5/20/1969	BUI	80-000892	Old	14th	Fish & Wildlife CC	FWC	5/20/2006	14:37:25	First Sample Volume 1.058L - No Flag No Second Sample - Improper Sample	Not A Valid Test - Need At Least Two Samples
Lopez	Toribio M.	10/5/1981	DUI	80-000907	Old	12th	Manatee County SO	Florida Highway Patrol	8/12/2006	0:51:18	First Sample Volume 1.441L - OK Second Sample Volume 0.554L - No Flag Third Sample Volume 2.554L - OK	Valid Test - First and Third Samples
Tamm	Trinity T.	6/28/1977	DUI	80-000912	Old	12th	Manatee County SO	Manatee County SO	4/19/2006	5:10:09	First Sample Volume 0.015L - Flagged Volume Not Met Second Sample Volume 0.082L - No Flag Third Sample Volume 0.296L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Skolroo	Richard D.	6/29/1962	DUI	80-000912	Old	12th	Manatee County SO	Florida Highway Patrol	8/28/2006	3:01:18	First Sample Volume 0.699L - No Flag Second Sample Volume 1.339L - OK	Not A Valid Test - First Sample Not Flagged
Deleon	Maria C.	5/22/1984	DUI	80-000912	New	12th	Manatee County SO	Bradenton PD	11/2/2006	2:01:08	First Sample Volume 0.320L - Flagged Volume Not Met Second Sample Volume 0.332L - No Flag Third Sample Volume 0.015L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Deleon	Maria C.	5/22/1984	DUI	80-000912	New	12th	Manatee County SO	Bradenton PD	11/2/2006	2:22:58	First Sample Volume 0.121L - No Flag Second Sample Volume 0.101L - Flagged Volume Not Met No Third Sample - Subject Refused	Not A Valid Test Refusal
Hurtado	Pablo E.	3/22/1967	DUI	80-000939	Old	20th	Lee County SO	Lee County SO	4/27/2006	2:24:56	First Sample Volume 2.125L - OK Second Sample Volume 0.550L - No Flag	Not A Valid Test - Second Sample Not Flagged
Fenno	Lillian	4/26/1962	DUI	80-000939	Old	20th	Lee County SO	Lee County SO	6/12/2006	15:27:43	First Sample Volume 0.289L - Flagged Volume Not Met Second Sample Volume 0.835L - No Flag Third Sample Volume 1.496L - OK	Not A Valid Test - Second Sample Not Flagged
Moore	Shannon C.	10/25/1982	DUI	80-000939	Old	20th	Lee County SO	Lee County SO	7/21/2006	0:18:42	First Sample Volume 0.347L - No Flag Second Sample Volume 1.871L - OK	Not A Valid Test - First Sample Not Flagged
Conaway	Doreen M.	11/15/1965	DUI	80-000939	Old	20th	Lee County SO	Lee County SO	7/23/2006	2:57:31	Second Sample Volume 0.375L - No Flag No Third Sample - Subject Refused	Not A Valid Test Refusal
Daughtry	Amanda S.	2/11/1972	DUI	80-000940	Old	18th	Brevard County SO	Brevard County SO	8/30/2006	3:25:19	First Sample Volume 1.605L - OK Second Sample Volume 0.230L - No Flag Third Sample Volume 0.757L - No Flag	Not A Valid Test - Second and Third Samples Not Flagged
Cohen	Nikol A.	6/6/1975	DUI	80-000943	Old	20th	Collier County SO	Naples PD	5/28/2006	3:00:50	First Sample Volume 0.613L - No Flag Second Sample Volume 1.816L - OK	Not A Valid Test - First Sample Not Flagged
Lyach	Vladimir	2/27/1958	DUI	80-000943	Old	20th	Collier County SO	Collier County SO	8/7/2006	21:20:07	First Sample Volume 2.144L - OK Second Sample Volume 0.218L - No Flag Third Sample Volume 1.875L - OK	Not A Valid Test - Second Control Test Outside Tolerance
Hall	Deborah L.	11/25/1956	DUI	80-000949	Old	2nd	Tallahassee PD	Tallahassee PD	7/26/2006	14:18:55	Second Control Test Outside Tolerance First Sample Volume 0.947L - No Flag Second Sample Volume - 1.417L - OK	Not A Valid Test First Sample Not Flagged
Bravo-Neiro	Francisco	5/8/1978	DUI	80-000951	Old	20th	Hendry County SO	Hendry County SO	7/30/2006	2:44:40	First Sample Volume 0.164L - No Flag Second Sample Volume 2.507L - OK	Not A Valid Test - First Sample Not Flagged
Castillo	Ruben	4/15/1974	DUI	80-000951	New	20th	Hendry County SO	Hendry County SO	10/8/2006	0:08:39	First Sample Volume 0.266L - No Flag No Second Sample - Subject Refused	Not A Valid Test Refusal
Adams	Rodney B.	6/2/1961	DUI	80-000957	Old	2nd	Leon County SO	Leon County SO	3/30/2006	19:14:15	First Sample Volume 1.316L - OK Second Sample Volume 0.410L - No Flag Third Sample Volume 1.570L - OK	Valid Test - First and Third Samples
Tucker	Leroy	10/2/1981	DUI	80-000958	Old	2nd	Leon County SO	Tallahassee PD	4/15/2006	22:23:00	First Sample Volume 0.429L - No Flag Second Sample Volume 2.250L - OK	Not A Valid Test First Sample Not Flagged
Furman	Shawn A.	10/21/1975	DUI	80-000958	Old	2nd	Leon County SO	Tallahassee PD	9/9/2006	1:18:33	First Sample Volume 0.347L - No Flag Second Sample Volume 0.000L - Flagged Volume Not Met Third Sample Volume 2.382L - OK	Not A Valid Test First Sample Not Flagged
Childs	Kendra P.	9/8/1969	DUI	80-000961	Old	2nd	Wakulla County SO	Wakulla County SO	4/14/2006	23:27:48	First Sample Volume 0.675L - No Flag Second Sample Volume 0.015L - Flagged Volume Not Met No Third Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Uriostegui	Marcos	2/28/1982	DUI	80-000963	Old	9th	Orange County SO	Orlando PD	5/17/2006	4:51:48	First Sample Volume 0.191L - Flagged Volume Not Met Second Sample Volume 0.183L - No Flag Third Sample Volume 1.453L - OK	Not A Valid Test - Second Sample Not Flagged

Altrabi	Fouad	4/2/1977	DUI	80-000963	Old	9th	Orange County SO	Orange County SO	5/29/2006	18:30:03	First Sample Volume 0.546L - No Flag Second Sample Volume 0.402L - Flagged Volume Not Met Third Sample Volume 3.941L - OK	Not A Valid Test First Sample Not Flagged
Miranda	Arnoldo	5/6/1983	DUI	80-000964	Old	9th	Orange County SO	Orange County SO	9/24/2006	4:58:41	First Sample Volume 0.765L - No Flag Second Sample Volume 1.933L - OK	Not A Valid Test First Sample Not Flagged
Rodriguez	Candido J.	8/8/1980	DUI	80-000966	Old	9th	Osceola County Corrections	Osceola County SO	4/22/2006	3:00:40	First Sample Volume 1.453L - OK Second Sample Volume 0.457L - No Flag Third Sample Volume 2.027L - OK Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance
Cannon	Heather M.	3/3/1987	DUI	80-000999	Old	18th	Seminole County SO	Seminole County SO	9/22/2006	5:13:02	First Sample Volume 0.230L - No Flag Second Sample Volume 0.031L - Flagged Volume Not Met Third Sample Volume 0.554L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Harazin	Linda S.	8/24/1947	DUI	80-001003	Old	6th	Pinellas County SO	Pinellas Park PD	5/4/2006	4:47:35	First Sample Volume 0.394L - No Flag Second Sample Volume 0.429L - Flagged Volume Not Met Third Sample Volume 0.359L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Chapman	Douglas R.	12/25/1924	DUI	80-001004	Old	18th	Brevard County SO	Brevard County SO	6/25/2006	1:36:28	First Sample Volume 0.421L - No Flag Second Sample Volume 0.144L - Flagged Volume Not Met Third Sample Volume 1.250L - OK	Not A Valid Test
Irby	Howard J.	10/7/1962	DUI	80-001005	Old	6th	Pinellas County SO	Pinellas County SO	4/11/2006	0:17:44	First Sample Volume 1.636L - OK Second Sample Volume 1.062L - No Flag	Not A Valid Test - Second Sample Not Flagged
Richmond	Aubrey L.	2/9/1950	DUI	80-001005	Old	6th	Pinellas County SO	Pinellas County SO	4/15/2006	1:26:05	First Sample Volume 1.644L - OK Second Sample Volume 0.710L - No Flag	Not A Valid Test - Second Sample Not Flagged
Maggs	Robert F.	9/29/1938	DUI	80-001005	Old	6th	Pinellas County SO	Pinellas County SO	7/26/2006	20:38:58	First Sample Volume 0.574L - No Flag Second Sample Volume 0.308L - Flagged Volume Not Met Third Sample Volume 0.261L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Roundtree	Aaron C.	8/22/1971	DUI	80-001005	Old	6th	Pinellas County SO	Pinellas County SO	8/20/2006	3:17:46	First Sample Volume 0.527L - No Flag Second Sample Volume 0.441L - Flagged Volume Not Met Third Sample Volume 1.152L - OK	Not A Valid Test
Dolch	Matthew W.	2/1/1981	DUI	80-001005	New	6th	Pinellas County SO	Florida Highway Patrol	9/22/2006	4:08:27	First Sample Volume 0.804L - No Flag Second Sample Volume 1.542L - OK	Not A Valid Test - First Sample Not Flagged
Segura-Guadarrama	Ivan	3/24/1984	DUI	80-001007	New	10th	Haines City PD	Haines City PD	9/24/2006	4:22:36	First Sample Volume 0.347L - Flagged Volume Not Met Second Sample Volume 0.199L - No Flag Third Sample Volume 0.078L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Crews	Earl J.	12/16/1957	DUI	80-001007	New	10th	Haines City PD	Haines City PD	10/7/2006	21:06:58	Second Sample Volume 0.441L - No Flag Third Sample Volume 0.394L - No Flag Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance
Smith	Heather L.	10/10/1987	DUI	80-001009	Old	18th	Palm Bay PD	Palm Bay PD	8/28/2006	4:23:50	First Sample Volume 0.429L - No Flag Second Sample Volume 1.648L - OK	Not A Valid Test
Thompson	Catherine S.	2/18/1958	DUI	80-001009	Old	18th	Palm Bay PD	Palm Bay PD	9/11/2006	0:12:25	First Sample Volume 0.714L - Flagged Volume Not Met Second Sample Volume 0.281L - No Flag Third Sample Volume 0.785L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Zanchez	Roberto	1/30/1981	DUI	80-001043	Old	10th	Highlands County SO	Florida Highway Patrol	5/21/2006	0:31:19	First Sample Volume 0.109L - No Flag No Second Sample - RFI	Not A Valid Test Test Aborted Due to RFI
Cavinder	Brenda A.	9/23/1960	DUI	80-001043	Old	10th	Highlands County SO	Sebring PD	6/1/2006	1:56:15	First Sample Volume 1.878L - OK Second Sample Volume 0.687L - No Flag Third Sample Volume 1.402L - OK	Valid Test - First and Third Samples
Marley	Lynne C.	1/10/1959	DUI	80-001043	Old	10th	Highlands County SO	Florida Highway Patrol	6/10/2006	0:10:08	First Sample Volume 0.382L - No Flag Second Sample Volume 1.640L - OK	Not A Valid Test - First Sample Not Flagged
Honorat	Hermiane	1/9/1956	DUI	80-001044	Old	17th	Coconut Creek PD	Coconut Creek PD	6/26/2006	23:34:39	First Sample Volume 1.753L - OK Second Sample Volume 0.304L - No Flag	Not A Valid Test - Second Sample Not Flagged
Osborne	Michele J.	11/2/1966	DUI	80-001047	New	17th	Broward County SO	Broward County SO	10/13/2006	2:13:24	First Sample Volume 0.128L - Flagged Volume Not Met Second Sample Volume 1.019L - No Flag Third Sample Volume 0.015L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Blair	Jason L.	7/14/1975	DUI	80-001051	Old	6th	St. Petersburg PD	St. Petersburg PD	7/4/2006	19:14:37	First Sample Volume 0.351L - No Flag Second Sample Volume 1.863L - OK No Third Sample - Subject Refused	Not A Valid Test Refusal
Pruett	Dennis R.	12/4/1951	DUI	80-001051	Old	6th	St. Petersburg PD	St. Petersburg PD	8/18/2006	19:07:01	First Sample Volume 0.402L - No Flag Second Sample Volume 1.550L - OK Third Sample Volume 1.363L - OK	Valid Test - Second and Third Samples
Zuniga	Sherry C.	7/11/1955	DUI	80-001055	Old	10th	Polk County SO	Florida Highway Patrol	7/29/2006	2:07:52	First Sample Volume 0.667L - No Flag Second Sample Volume 1.503L - OK	Not A Valid Test - First Sample Not Flagged
Addleman	Holly L.	10/6/1977	DUI	80-001057	Old	10th	Lakeland PD	Lakeland PD	6/18/2006	3:23:44	First Sample Volume 0.300L - No Flag Second Sample Volume 1.125L - OK	Not A Valid Test - First Sample Not Flagged
Rivera	Victor M.	4/5/1950	DUI	80-001057	Old	10th	Lakeland PD	Lakeland PD	6/20/2006	3:12:02	First Sample Volume 0.699L - No Flag Second Sample Volume 1.656L - OK	Not A Valid Test - First Sample Not Flagged
Godinez Perez	Alberto	1/11/1975	DUI	80-001063	Old	17th	Hollywood PD	Hollywood PD	7/1/2006	2:21:25	First Sample Volume 0.109L - No Flag Second Sample Volume 1.488L - OK Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance

Miradakis	Kimberly A.	5/7/1960	DUI	80-001065	New	6th	Tarpon Springs PD	Tarpon Springs PD	9/23/2006	3:26:04	First Sample Volume 0.035L - Flagged Volume Not Met Second Sample Volume 0.808L - No Flag Third Sample Volume 0.164L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Buchanan	Bonnie T.	12/11/1940	DUI	80-001067	Old	17th	Broward County SO	Broward County SO	5/27/2006	19:03:55	First Sample Volume 0.546L - Flagged Volume Not Met Second Sample Volume 0.738L - No Flag Third Sample Volume 1.105L - OK	Not A Valid Test - Second Sample Not Flagged
Lanot	Jesufranc	11/17/1971	DUI	80-001067	Old	17th	Broward County SO	Broward County SO	7/1/2006	22:47:55	First Sample Volume 0.718L - Flagged Volume Not Met Second Sample Volume 1.574L - OK Third Sample Volume 0.218L - No Flag	Not A Valid Test - Third Sample Not Flagged
Nelson	Stacy A.	7/22/1963	DUI	80-001069	Old	4th	Jacksonville SO	Jacksonville SO	8/21/2006	4:30:22	First Sample Volume 1.093L - No Flag Second Sample Volume 1.957L - OK	Not A Valid Test
Doyle	Stephen V.	7/24/1952	DUI	80-001069	Old	4th	Jacksonville SO	Jacksonville SO	9/5/2006	1:54:10	First Sample Volume 1.382L - OK Second Sample Volume 0.703L - No Flag	First Sample Not Flagged
Buckley	Linda G.	5/10/1967	DUI	80-001071	Old	4th	Clay County SO	Clay County SO	5/6/2006	2:11:12	First Sample Volume 0.671L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Second Sample Not Flagged
Davis	Kim P.	11/20/1957	DUI	80-001082	Old	17th	Plantation PD	Sunrise PD	5/9/2006	16:54:56	First Sample Volume 0.550L - No Flag Second Sample Volume 0.574L - Flagged Volume Not Met Third Sample Volume 1.457L - OK	Not A Valid Test - First Sample Not Flagged
Nguyen	Thai H.	8/2/1963	DUI	80-001083	Old	4th	Jacksonville SO	Jacksonville SO	5/14/2006	22:56:41	First Sample Volume 0.671L - No Flag Second Sample Volume 0.714L - Flagged Volume Not Met Third Sample Volume 0.179L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Gagne	Audra J.	12/29/1966	DUI	80-001117	Old	6th	Pasco County SO	Pasco County SO	4/25/2006	2:54:49	First Sample Volume 0.949L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Drake	Dawn D.	6/9/1970	DUI	80-001117	Old	6th	Pasco County SO	Pasco County SO	4/27/2006	1:28:39	First Sample Volume 2.578L - OK Second Sample Volume 0.718L - No Flag	Not A Valid Test - Second Sample Not Flagged
Brooks	Geoffrey D.	4/20/1972	DUI	80-001131	Old	7th	Volusia County SO	Volusia County SO	7/4/2006	2:15:13	First Sample Volume 1.015L - No Flag Second Sample Volume 1.101L - OK	Not A Valid Test
Brock	Lyn	7/20/1942	DUI	80-001137	Old	7th	Flagler County SO	Flagler County SO	7/22/2006	18:30:02	First Sample Volume 0.097L - Flagged Volume Not Met Second Sample Volume 0.019L - No Flag Third Sample Volume 0.511L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Kenyon	John R.	11/20/1979	DUI	80-001139	Old	17th	Broward County SO	Broward County SO	6/14/2006	1:48:17	First Sample Volume 1.261L - OK Second Sample Volume 0.972L - No Flag	Not A Valid Test - Second Sample Not Flagged
Jean-Jacques	Paholo	5/10/1979	DUI	80-001139	Old	17th	Broward County SO	Broward County SO	7/8/2006	5:35:25	First Sample Volume 0.410L - No Flag Second Sample Volume 1.519L - OK Third Sample Volume 1.750L - OK	Valid Test - Second and Third Samples
Palacios	Mirian L.	1/21/1978	DUI	80-001139	Old	17th	Broward County SO	Broward County SO	7/22/2006	3:46:52	First Sample Volume 0.519L - No Flag Second Sample Volume 0.097L - Flagged Volume Not Met Third Sample Volume 0.414L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Hendry	Mark J.	5/31/1960	DUI	80-001139	Old	17th	Broward County SO	Broward County SO	9/22/2006	1:11:06	First Sample Volume 1.406L - OK Second Sample Volume 0.531L - No Flag	Not A Valid Test - Second Sample Not Flagged
Barrera	Aristides E.	8/21/1976	DUI	80-001153	Old	2nd	Leon County SO	Leon County SO	4/13/2006	21:55:59	First Sample Volume 0.421L - No Flag Second Sample Volume 0.191L - No Flag	Not A Valid Test - First and Second Sample Not Flagged
Gibson	Amelia C.	7/30/1986	DUI	80-001153	Old	2nd	Leon County SO	Leon County SO	8/23/2006	5:26:19	First Sample Volume 1.046L - No Flag Second Sample Volume 0.335L - Flagged Volume Not Met Third Sample Volume 1.199L - OK	Not A Valid Test
Waggoner	Nicholas C.	5/4/1991	DUI	80-001158	Old	7th	Ponce Inlet PD	Daytona Beach PD	4/5/2006	23:12:41	First Sample Volume 0.367L - No Flag Second Breath Sample 0.125L - Flagged Volume Not Met No Third Sample - Sequence Aborted	Not A Valid Test - Need At Least Two Samples
Currie II	Robert A.	5/11/1955	DUI	80-001166	Old	18th	Brevard County SO	Brevard County SO	4/13/2006	20:11:17	First Sample Volume 2.066L - OK Second Sample Volume 1.058L - No Flag	Not A Valid Test - Second Sample Not Flagged
Ramharack	Rajnauth	10/28/1983	DUI	80-001174	New	5th	Clermont PD	Clermont PD	10/15/2006	23:11:23	Second Sample Volume 1.843L - OK Third Sample Volume 2.464L - OK Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance
Zhang	Dong	6/24/1967	DUI	80-001190	Old	15th	Boynton Beach PD	Boynton Beach PD	6/17/2006	8:35:40	First Sample Volume 0.828L - No Flag Second Sample Volume 2.519L - OK	Not A Valid Test - First Sample Not Flagged
Ferranti	Edward J.	5/26/1970	DUI	80-001190	Old	15th	Boynton Beach PD	Boynton Beach PD	7/22/2006	3:11:52	First Sample Volume 0.015L - Flagged Volume Not Met Second Sample Volume 2.066L - OK Third Sample Volume 0.726L - No Flag	Not A Valid Test - Third Sample Not Flagged
Kaszak	Kenneth	4/16/1958	DUI	80-001190	Old	15th	Boynton Beach PD	Boynton Beach PD	8/13/2006	0:46:51	First Sample Volume 1.406L - OK Second Sample Volume 0.441L - No Flag	Not A Valid Test - Second Sample Not Flagged
Perez-Ortiz	Issac	10/23/1982	DUI	80-001235	Old	15th	West Palm Beach PD	West Palm Beach PD	4/8/2006	0:03:15	First Sample Volume 1.285L - OK Second Sample Volume 0.488L - No Flag Third Sample Volume 1.281L - OK	Valid Test - First and Third Samples
Rucker	Curtis	3/12/1945	DUI	80-001248	Old	17th	Broward County SO	Broward County SO	5/7/2006	0:02:11	First Sample Volume 0.593L - No Flag Second Sample Volume 0.816L - No Flag	Not A Valid Test - First and Second Sample Not Flagged



White	Leslie J.	6/6/1960	DUI	80-001272	Old	18th	Seminole County SO	Oviedo PD	7/2/2006	2:46:00	Second Sample Volume 0.257L - Flagged Volume Not Met Third Sample Volume 0.679L - No Flag	Not A Valid Test - Second Control Test Outside Tolerance
Newton	Justin B.	10/21/1978	DUI	80-001272	Old	18th	Seminole County SO	Seminole County SO	7/21/2006	22:17:44	Second Control Test Outside Tolerance First Sample Volume 0.457L - No Flag Second Sample Volume 1.390L - OK	Not A Valid Test - First Sample Not Flagged
Lopez-Hernandez	David	4/4/1978	DUI	80-001273	Old	6th	Pinellas County SO	Pinellas County SO	7/15/2006	3:03:14	First Sample Volume 0.050L - No Flag Second Sample Volume 1.414L - OK Third Sample Volume 1.613L - OK	Valid Test - Second and Third Samples
Portuondo	Erik J.	12/11/1975	DUI	80-001276	Old	11th	Miami Dade PD	Miami Dade PD	5/12/2006	0:34:06	First Sample Volume 0.109L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Fandino	Marta L.	8/18/1967	DUI	80-001276	Old	11th	Miami Dade PD	Miami Dade PD	5/27/2006	4:12:02	First Sample Volume 0.453L - No Flag Second Sample Volume 1.839L - OK Third Sample Volume - 1.292L - OK	Valid Test - Second and Third Samples
Fuhr	Alison R.	10/31/1978	DUI	80-001282	Old	8th	University of Florida PD	University of Florida PD	4/13/2006	3:34:42	First Sample Volume 0.867L - No Flag Second Sample Volume 0.152L - Flagged Volume Not Met Third Sample Volume 0.082L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Johns	Cynthia J.	5/18/1966	DUI	80-001284	Old	8th	Clchrist County SO	Cedar Keys PD	6/15/2006	1:38:28	First Sample Volume 0.421L - No Flag Second Sample Volume 0.378L - Flagged Volume Not Met Third Sample Volume 0.816L - No Flag	Not A Valid Test - First and Third Samples Not Flagged
Campbell	Steven M.	7/12/1974	DUI	80-001303	Old	1st	Santa Rosa County SO	Santa Rosa County SO	8/6/2006	20:30:09	First Sample Volume 0.289L - Flagged Volume Not Met Second Sample Volume 0.789L - No Flag Third Sample Volume 1.687L - OK	Not A Valid Test - Second Sample Not Flagged
Phillips	Darius K.	4/9/1984	OTHER	80-001309	Old	14th	Bay County SO	Tyndall AFB	4/9/2006	13:22:01	First Sample Volume 0.628L - No Flag Second Sample Volume 0.578L - Flagged Volume Not Met No Third Sample - Subject Refused	Not A Valid Test Refusal
Cardwell	Curtis B.	11/24/1981	DUI	80-001309	Old	14th	Bay County SO	Bay County SO	7/23/2006	23:35:34	First Sample Volume 0.511L - No Flag Second Sample Volume 0.000L - Flagged Volume Not Met Third Sample Volume 0.437L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Miller	Chadrick L.	4/24/1978	DUI	80-001309	Old	14th	Bay County SO	Parker PD	9/16/2006	22:34:35	First Sample Volume 0.437L - No Flag No Second Sample - Subject Refused	Not A Valid Test Refusal
Mings	Lonnice L.	6/29/1977	DUI	80-001309	Old	14th	Bay County SO	Florida Highway Patrol	9/23/2006	3:41:36	First Sample Volume 2.710L - OK Second Sample Volume 0.097L - No Flag Third Sample Volume 1.925L - OK	Valid Test - First and Third Samples
Castillo-Vazquez	Jose	12/6/1960	DUI	80-001309	Old	14th	Bay County SO	Florida Highway Patrol	9/24/2006	1:50:55	First Sample Volume 0.906L - No Flag Second Sample Volume 2.613L - OK	Not A Valid Test - First Sample Not Flagged
Clay	Susan S.	12/9/1953	DUI	80-001310	Old	1st	Niceville PD	Crestview PD	6/12/2006	0:08:55	First Sample Volume 1.246L - OK Second Sample Volume 0.558L - No Flag Third Sample Volume 1.394L - OK	Valid Test - First and Third Samples
Wetherington	Deborah S.	8/10/1964	DUI	80-001310	New	1st	Niceville PD	Niceville PD	10/4/2006	3:39:56	First Sample Volume 0.035L - No Flag No Second Sample Provided No Third Sample Provided	Not A Valid Test - Need At Least Two Samples
Brizendine	William E.	9/4/1963	DUI	80-001313	New	14th	Springfield PD	Panama City PD	9/20/2006	1:55:45	First Sample Volume 1.109L - OK Second Sample Volume 1.011L - No Flag	Not A Valid Test - Second Sample Not Flagged
Diamond	Elfriede F.	11/1/1936	DUI	80-001328	Old	19th	Indian River County SO	Sebastian PD	8/30/2006	22:47:27	First Sample Volume 0.554L - No Flag Second Sample Volume 1.003L - No Flag	Not A Valid Test - First and Second Sample Not Flagged
Angus	Anthony	6/11/1959	DUI	80-001328	Old	19th	Indian River County SO	Florida Highway Patrol	9/11/2006	8:34:49	First Sample Volume 1.972L - OK Second Sample Volume 0.156L - Flagged Volume Not Met Third Sample Volume 0.882L - No Flag	Not A Valid Test - Third Sample Not Flagged
Kellowan	Shevron S.	3/20/1979	DUI	80-001344	Old	12th	Sarasota County SO	Florida Highway Patrol	4/1/2006	4:55:54	First Sample Volume 0.093L - Flagged Volume Not Met Second Sample Volume 3.066L - OK Third Sample Volume 0.714L - No Flag	Not A Valid Test - Third Sample Not Flagged
Ryll	Frank L.	10/27/1951	DUI	80-001344	Old	12th	Sarasota County SO	Sarasota County SO	4/25/2006	22:40:18	First Sample Volume 0.937L - No Flag Second Sample Volume 1.972L - OK	Not A Valid Test - First Sample Not Flagged
Kerr	Julia P.	2/11/1987	DUI	80-001344	Old	12th	Sarasota County SO	Sarasota County SO	5/3/2006	4:22:57	First Sample Volume 0.468L - No Flag Second Sample Volume 1.871L - OK Third Sample Volume 1.691L - OK	Not A Valid Test - No 0.020 Agreement Between Second and Third Samples
Shapiro	Wayne G.	3/25/1959	DUI	80-001344	Old	12th	Sarasota County SO	Sarasota PD	5/28/2006	3:58:01	First Sample Volume 0.656L - No Flag Second Sample Volume 1.683L - OK Third Sample Volume 1.800L - OK	Valid Test - Second and Third Samples
Quattrone	Nancy E.	4/9/1957	DUI	80-001344	Old	12th	Sarasota County SO	Sarasota County SO	6/7/2006	0:27:30	First Sample Volume 0.324L - No Flag Second Sample Volume 1.921L - OK	Not A Valid Test - First Sample Not Flagged
Velle	Victoreno	10/27/1979	DUI	80-001344	Old	12th	Sarasota County SO	Sarasota County SO	6/18/2006	3:26:40	First Sample Volume 1.625L - OK Second Sample Volume 0.636L - No Flag Third Sample Volume 2.160L - OK	Valid Test - First and Third Samples
Palmer	Scott D.	7/13/1964	DUI	80-001344	Old	12th	Sarasota County SO	Sarasota PD	6/26/2006	1:53:50	First Sample Volume 1.101L - OK Second Sample Volume 0.804L - No Flag No Third Sample - Subject Refused	Not A Valid Test Refusal

Delahanty	Thomas L.	5/15/1948	DUI	80-001344	Old	12th	Sarasota County SO	Sarasota County SO	7/8/2006	0:30:50	First Sample Volume 0.492L - No Flag Second Sample Volume 0.191L - Flagged Volume Not Met Third Sample Volume 0.378L - No Flag	Not A Valid Test - First and Third Sample Not Flagged
Nch	Randy R.	6/23/1959	DUI	80-001348	New	12th	Sarasota County SO	North Port PD	10/8/2006	19:25:47	First Sample Volume 0.449L - No Flag Second Sample Volume 2.023L - OK Third Sample Volume 1.519L - OK	Not A Valid Test - No 0.020 Agreement Between Second and Third Samples
Germain	David M.	4/8/1957	DUI	80-001367	Old	6th	Pinellas County SO	Florida Highway Patrol	7/16/2006	18:23:45	First Sample Volume 0.906L - No Flag Second Sample Volume 1.125L - OK Third Sample Volume 1.371L - OK	Valid Test - Second and Third Samples
Desimas	John R.	2/1/1984	DUI	80-001415	Old	6th	Pinellas County SO	Pinellas Park PD	8/5/2006	3:02:07	First Sample Volume 0.130L - No Flag Second Sample Volume 0.593L - Flagged Volume Not Met Third Sample Volume 0.144L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Dilimone	Linda J.	3/13/1962	DUI	80-001416	Old	9th	Orange County SO	Orange County SO	6/6/2006	12:48:14	First Sample Volume 0.835L - No Flag Second Sample Volume 0.437L - Flagged Volume Not Met Third Sample Volume 0.070L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Martinez	Noel I.	9/22/1968	DUI	80-001416	Old	9th	Orange County SO	Orange County SO	6/10/2006	4:17:28	First Sample Volume 0.937L - No Flag Second Sample Volume 1.718L - OK	Not A Valid Test
Kars	Chad D.	1/4/1967	DUI	80-001416	Old	9th	Orange County SO	Orlando PD	7/10/2006	1:34:39	First Sample Volume 1.308L - OK Second Sample Volume 0.792L - No Flag	First Sample Not Flagged
Reyes	Arturo	9/11/1965	DUI	80-001418	Old	9th	Orange County SO	Florida Highway Patrol	8/30/2006	1:34:24	First Sample Volume 0.042L - Flagged Volume Not Met Second Sample Volume 0.804L - No Flag Third Sample Volume 2.421L - OK	Not A Valid Test - Second Sample Not Flagged
Pooler	Nikki K.	7/17/1975	DUI	80-001418	Old	9th	Orange County SO	Maitland PD	8/30/2006	3:38:20	First Sample Volume 0.703L - No Flag Second Sample Volume 0.261L - Flagged Volume Not Met Third Sample Volume 1.531L - OK	Not A Valid Test First Sample Not Flagged
Stephens	Patrick J.	11/20/1984	DUI	80-001420	Old	9th	Orange County SO	Orange County SO	6/11/2006	4:28:54	First Sample Volume 2.972L - OK Second Sample Volume 0.394L - No Flag	Not A Valid Test - Second Sample Not Flagged
Buckley	Linda G.	5/10/1967	DUI	80-001443	Old	4th	Clay County SO	Clay County SO	5/6/2006	1:46:18	First Sample Volume 1.789L - OK Second Sample Volume 1.464L - OK Third Sample Volume 0.648L - No Flag Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance
Previllon	Toni L.	4/25/1964	DUI	80-001443	New	4th	Clay County SO	Florida Highway Patrol	9/24/2006	0:06:04	First Sample Volume 0.265L - Flagged Volume Not Met Second Sample Volume 0.953L - No Flag Third Sample Volume 0.707L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Nonte	William	4/3/1960	DUI	80-001731	New	13th	Hillsborough County SO	Hillsborough County SO	10/15/2006	20:57:01	First Sample Volume 1.710L - OK Second Sample Volume 1.085L - No Flag Third Sample Volume 1.503L - OK	Valid Test - First and Third Samples

\*\* Notes are concerning the listed breath tests. Additional breath tests of the referenced subject may have been conducted and are not a part of this list.





Florida Department of  
Law Enforcement

Gerald M. Bailey  
*Commissioner*

**Criminal Justice  
Professionalism Program  
Alcohol Testing Program**

P.O. Box 1489  
Tallahassee, Florida 32302  
(850) 410-8600  
(850) 410-7816 Fax  
<http://www.fdle.state.fl.us>

## MEMORANDUM

**TO:** Department Inspectors, Agency Inspectors

**FROM:** Laura D. Barfield, Alcohol Testing Program Manager **LDB**

**DATE:** December 7, 2006

**SUBJECT:** Intoxilyzer 8000

---

In a letter dated October 5, 2006, you were made aware of an issue regarding the Intoxilyzer 8000 with software version 8100.26. The issue involved the instrument's ability to correctly identify the volume of a breath sample delivered past three minutes due to missing instructions in the software programming.

Since that time, all Intoxilyzer 8000 evidentiary breath test instruments have been upgraded to version 8100.27. This software revision was evaluated by the Florida Department of Law Enforcement on October 9 and 10, 2006 in accordance with Chapter 11D-8, Florida Administrative Code. The software revision upgrade to all Intoxilyzer 8000 evidentiary breath test instruments began on October 11, 2006 and was completed on November 20, 2006.

I have attached a list of all affected breath tests. This list includes tests that you were previously notified of as well as any additional tests since the October 5, 2006 notice.

If you have any questions, please feel free to contact me at (850) 410-7810.

LDB

Attachment

Final List

Subject Last Name	Subject First Name	Subject's Date of Birth	Offense	Instrument Serial Number	Instrument To	Arresting Officer Agency	Date of Test	Time of Test	Breath Sample Volume Results	Notes**
Corredor	Miguel A.	1/7/1986	DUI	80-000223	FDLE (Loaned to Indian River County Sheriff's Office)	Florida Highway Patrol	7/6/2006	22:58:02	First Sample Volume 1.933 L - OK; Second Sample Volume 0.972L - No Flag	Not A Valid Test - Second Sample Not Flagged
Cruz	Santiago G	12/29/1965	DUI	80-000224	FDLE (Loaned to Okaloosa County Sheriff's Office)	Okaloosa County SO	10/1/2006	4:40:04	First Sample Volume 1.808L - OK Second Sample Volume 0.050L - Flagged Volume Not Met Third Sample Volume - 0.582L - No Flag	Not A Valid Test Third Sample Not Flagged
Sedovic	Michael T.	1/4/1976	DUI	80-000738	Fort Walton Beach PD	Okaloosa County SO	8/5/2006	4:54:39	First Sample Volume 0.226L - Flagged Volume Not Met No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Peluola	Terilola A	12/24/1961	DUI	80-000739	Okaloosa County SO	Florida Highway Patrol	8/21/2006	0:10:12	First Sample Volume 1.066L - No Flag Second Sample Volume 1.164L - OK	Not A Valid Test First Sample Not Flagged
Fulford	Kenneth M	7/7/1951	DUI	80-000744	Escambia County SO	Pensacola PD	5/17/2006	2:04:10	First Sample Volume 0.605L - No Flag Second Sample Volume 0.664L - No Flag	Not A Valid Test - Second Sample Not Flagged
Burgos	Abel	7/11/1978	DUI	80-000744	Escambia County SO	Florida Highway Patrol	7/22/2006	3:58:14	First Sample Volume 1.550L - OK Second Sample Volume 2.105L - OK	Not A Valid Test First Sample Not Flagged
East	Easter K.	3/21/1945	DUI	80-000744	Escambia County SO	Florida Highway Patrol	8/12/2004	2:04:13	First Sample Volume 1.082L - No Flag Second Sample - Subject Refused	Not A Valid Test Refusal
Rives	Lisa C.	6/8/1959	DUI	80-000744	Escambia County SO	Florida Highway Patrol	9/28/2006	21:07:00	First Sample Volume 0.601L - No Flag Second Sample Volume 1.515L - OK	Not A Valid Test First Sample Not Flagged
Meshell	Robert J.	7/12/1981	DUI	80-000752	Marion County SO	Ocala PD	10/16/2006	19:51:26	First Sample Volume 0.074L - Flagged Volume Not Met Second Sample Volume 0.566L - No Flag	Not A Valid Test - Need At Least Two Samples
Swanson	Amanda N.	9/9/1984	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	5/18/2006	5:39:00	First Sample Volume 0.062L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Hiddle	Brandi C.	11/9/1972	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	5/25/2006	1:05:16	First Sample Volume 0.129L - Flagged Volume Not Met Second Sample Volume 0.347L - No Flag	Not A Valid Test - Need At Least Two Samples
Toney	Patrick S.	3/10/1987	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	5/30/2006	1:29:54	First Sample Volume 1.000L - No Flag Second Sample Volume 1.550L - OK	Not A Valid Test First Sample Not Flagged
Boone	Michael J.	11/9/1983	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	6/25/2006	2:30:07	First Sample Volume 0.265L - No Flag Second Sample Volume 0.054L - Flagged Volume Not Met Third Sample Volume 1.398L - OK	Not A Valid Test First Sample Not Flagged
Alcandri	Danielle M.	8/1/1971	DUI	80-000763	Okaloosa County SO	Okaloosa County SO	7/25/2006	2:36:03	First Sample Volume 0.742L - No Flag Second Sample Volume 0.855L - Flag Volume Not Met Third Sample Volume 0.214L - Flag Volume Not Met	Not A Valid Test - Need At Least Two Samples
Jones	Elaine J.	9/18/1969	DUI	80-000772	Santa Rosa County SO	Santa Rosa County SO	5/13/2006	2:38:47	First Sample Volume 0.476L - No Flag Second Sample Volume 1.738L - OK	Not A Valid Test First Sample Not Flagged
Woods	Christopher L.	5/6/1946	DUI	80-000772	Santa Rosa County SO	Florida Highway Patrol	9/27/2006	20:29:23	First Sample Volume 0.730L - No Flag Second Sample Volume 1.964L - OK Third Sample Volume 2.000L - OK	Valid Test - Second and Third Samples
Bohannon	Lynette D.	12/18/1972	DUI	80-000786	Broward County SO	Broward County SO	5/12/2006	0:10:42	First Sample Volume 0.652L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Figueredo	Kristina M.	8/2/1985	DUI	80-000786	Broward County SO	Other State	10/15/2006	6:03:02	First Sample Volume 0.140L - No Flag Second Sample Volume 1.460L - OK	Not A Valid Test - First Sample Not Flagged
Manella	Thomas	3/28/1951	DUI	80-000786	Broward County SO	Broward County SO	10/17/2006	0:58:36	First Sample Volume 0.750L - No Flag Second Sample Volume 1.167L - OK Third Sample Volume - 1.226L - OK	Valid Test - Second and Third Samples
Johnson	Patrick H.	5/2/1951	DUI	80-000786	Broward County SO	Broward County SO	10/18/2006	0:06:45	First Sample Volume 0.480L - No Flag Second Sample Volume 1.769L - OK	Not A Valid Test - First Sample Not Flagged
Shelton	Linda M.	8/27/1940	DUI	80-000790	Vero Beach PD	Vero Beach PD	7/13/2006	3:05:22	First Sample Volume 0.324L - Flag Volume Not Met Second Sample Volume 0.632L - No Flag Third Sample Volume 0.761L - Flag Volume Not Met	Not A Valid Test - Need At Least Two Samples
Vaught	Michael A.	4/20/1981	DUI	80-000794	St. Lucie County SO	St. Lucie County SO	4/9/2006	2:49:38	First Sample Volume 0.851L - Flag Volume Not Met Second Sample Volume 1.292L - OK Third Sample Volume 0.613L - No Flag	Not A Valid Test - Third Sample Not Flagged
Lesko	Darryl R.	5/10/1958	DUI	80-000794	St. Lucie County SO	St. Lucie County SO	6/5/2006	1:49:07	First Sample Volume 1.527L - OK Second Sample Volume 0.261L - No Flag Third Sample Volume 1.371L - OK	Not A Valid Test - Second Control Test Outside Tolerance
Cortez	Javier O.	1/14/1964	DUI	80-000794	St. Lucie County SO	Fort Pierce PD	6/10/2006	0:36:52	First Sample Volume 0.210L - No Flag Second Sample Volume 1.222L - OK Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance
Scheider	Robert P.	4/23/1962	DUI	80-000794	St. Lucie County SO	Fort Pierce PD	6/19/2006	20:12:44	First Sample Volume 1.570L - OK Second Sample Volume 0.300L - No Flag Third Sample Volume 2.429L - OK Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance

Slater	Keith	6/24/1972	DUI	80-000794	Old	19th	St Lucie County SO	Fort Pierce PD	6/21/2006	9 25 13	First Sample Volume 1.113L - OK Second Sample Volume 0.234L - No Flag Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance
Maliszewski	Dagmar E	8/5/1967	DUI	80-000794	Old	19th	St Lucie County SO	Port St. Lucie PD	6/25/2006	23:16:05	First Sample Volume 0.515L - No Flag Second Sample Volume 1.343L - OK Third Sample Volume 1.207L - OK	Not A Valid Test - Second Control Test Outside Tolerance
Diaz	Alejandro	10/29/1984	DUI	80-000794	Old	19th	St. Lucie County SO	St. Lucie County SO	8/19/2006	3 20 08	First Sample Volume 0.812L - No Flag No Second Sample - Subject Refused	Not A Valid Test - Need At Least Two Samples
Jimenez	David Y	1/24/1977	DUI	80-000808	Old	13th	Hillsborough County SO	Hillsborough County SO	3/30/2006	23 07 03	First Sample Volume 0.816L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Pitisa	Petar J.	11/25/1982	DUI	80-000808	Old	13th	Hillsborough County SO	Tampa PD	7/2/2006	3 58 14	First Sample Volume 1.054L - No Flag Second Sample Volume 1.261L - OK	Not A Valid Test - First Sample Not Flagged
Roberts	Pamela R.	7/23/1984	DUI	80-000808	Old	13th	Hillsborough County SO	Tampa PD	8/24/2006	0 30 17	First Sample Volume 1.070L - No Flag Second Sample Volume 1.679L - OK	Not A Valid Test - First Sample Not Flagged
Reyes	Tony	6/11/1984	DUI	80-000808	Old	13th	Hillsborough County SO	Plant City PD	9/3/2006	0 04 38	First Sample Volume 0.363L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Clark	Michael C.	4/9/1983	DUI	80-000808	Old	13th	Hillsborough County SO	Florida Highway Patrol	9/9/2006	5 52 04	First Sample Volume 0.246L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Hargrave	Evelyn M.	7/20/1979	DUI	80-000808	Old	13th	Hillsborough County SO	Tampa PD	9/10/2006	3 08 59	First Sample Volume 0.925L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Ballestas	Nicholas E.	12/17/1950	DUI	80-000808	New	13th	Hillsborough County SO	Hillsborough County SO	10/8/2006	21 58 24	First Sample Volume 0.332L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Diaz-Gonzalez	Raul N	11/24/1957	DUI	80-000813	Old	5th	Marion County SO	Marion County SO	6/21/2006	14 35 11	First Sample Volume 2.738L - OK Second Sample Volume 0.476L - No Flag Purge Fail After Second Sample	Not A Valid Test - Need At Least Two Samples
Merriott	Brady J.	9/9/1984	DUI	80-000817	Old	5th	Citrus County SO	Citrus County SO	9/14/2006	4 23 44	First Sample Volume 0.398L - No Flag Second Sample Volume 0.179L - No Flag	Not A Valid Test - First and Second Sample Not Flagged
Weich	Andrei C.	3/20/1966	DUI	80-000824	Old	5th	Leesburg PD	Leesburg PD	4/15/2006	0 19 50	First Sample Volume 0.324L - No Flag Second Sample Volume 1.117L - OK	Not A Valid Test - First Sample Not Flagged
Garcia	Alfonso R	8/17/1980	DUI	80-000829	Old	13th	Hillsborough County SO	Tampa PD	4/12/2006	2 16 24	First Sample Volume 0.339L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Nilsen	Michael R	10/9/1961	DUI	80-000832	Old	19th	Martin County SO	Florida Highway Patrol	5/29/2006	3 47 51	First Sample Volume 1.660L - OK Second Sample Volume 0.972L - No Flag Third Sample Volume 1.539L - OK	Valid Test - First and Third Samples
Camacho	Rafael	4/3/1961	DUI	80-000832	Old	19th	Martin County SO	Stuart PD	6/22/2006	19 54 57	First Sample Volume 0.597L - No Flag Second Sample Volume 0.204L - Flagged Volume Not Met Third Sample Volume 0.219L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Mendoza	Cecilio	11/22/1962	DUI	80-000832	Old	19th	Martin County SO	Martin County SO	7/9/2006	2 12 46	First Sample Volume 0.000L - Flagged Volume Not Met Second Sample Volume 2.226L - OK Third Sample Volume 0.183 - No Flag	Not A Valid Test - Third Sample Not Flagged
Schroeder	Brandon H.	7/28/1987	DUI	80-000832	Old	19th	Martin County SO	Martin County SO	8/13/2006	2 45 47	First Sample Volume 0.468L - Flagged Volume Not Met Second Sample Volume 0.769L - No Flag Third Sample Volume 1.480L - OK	Not A Valid Test - Second Sample Not Flagged
McClelland	Patricia P.	8/25/1950	DUI	80-000832	New	19th	Martin County SO	Martin County SO	8/25/2006	21 47 45	First Sample Volume 0.152L - No Flag No Second Sample - Sequence Aborted	Not A Valid Test - Need At Least Two Samples
Daher	Jabib	2/3/1948	DUI	80-000832	Old	19th	Martin County SO	FWC West Palm	8/26/2006	23 30 17	First Sample Volume 1.726L - OK Second Sample Volume 0.449L - No Flag	Not A Valid Test - Second Sample Not Flagged
Vandermark	Jeffrey W	6/16/1945	DUI	80-000832	Old	19th	Martin County SO	Sewall's Point PD	9/4/2006	2 04 32	First Sample Volume 0.808L - No Flag Second Sample Volume 3.394L - OK	Not A Valid Test - First Sample Not Flagged
Michale	James M	2/27/1967	DUI	80-000832	Old	19th	Martin County SO	Stuart PD	9/4/2006	21 48 28	First Sample Volume 0.203L - No Flag Second Sample Volume 1.691L - OK Third Sample Volume 1.187L - OK	Not A Valid Test - Second Control Test Outside Tolerance
Moretti	Derek P.	6/2/1980	DUI	80-000833	Old	13th	Hillsborough County SO	Tampa PD	5/15/2006	1 09 00	First Sample Volume 0.742L - No Flag Second Sample Volume 0.196L - Flag Volume Not Met No Third Sample - Subject Refused	Not A Valid Test - Refusal
Perez	Palermo	11/20/1972	DUI	80-000833	Old	13th	Hillsborough County SO	Hillsborough County SO	6/10/2006	19 51 30	First Sample Volume 0.105L - No Flag Second Sample Volume 0.212L - Flagged Volume Not Met Third Sample Volume 0.219L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Pospisil	Ashley N.	5/10/1993	DUI	80-000833	Old	13th	Hillsborough County SO	Tampa PD	7/3/2006	4 13 57	First Sample Volume 0.957L - No Flag Second Sample Volume 1.648L - OK	Not A Valid Test - First Sample Not Flagged
Gandanilla	Petronio R.	9/6/1972	DUI	80-000833	Old	13th	Hillsborough County SO	Plant City PD	7/21/2006	9 12 20	First Sample Volume 0.371L - No Flag Second Sample - Subject Refused	Not A Valid Test - Refusal
Rojas	Oscar	8/27/1949	DUI	80-000833	Old	13th	Hillsborough County SO	Tampa PD	8/11/2006	2 46 47	First Sample Volume 0.625L - No Flag Second Sample - Subject Refused	Not A Valid Test - Refusal
Rolon	Jose A.	3/8/1961	DUI	80-000833	Old	13th	Hillsborough County SO	Florida Highway Patrol	8/13/2006	4 32 45	First Sample Volume 0.574L - No Flag Second Sample Volume 1.289L - OK	Not A Valid Test - First Sample Not Flagged

Fussell	Christopher N	7/19/1980	DUI	80-000833	New	13th	Hillsborough County SO	Hillsborough County SO	10/7/2006	3:46:57	First Sample Volume 0.289L - No Flag Second Sample Volume 2.910L - OK Third Sample Volume - 3.050L - OK	Valid Test - Second and Third Samples
Seymour	Patrick H.	8/26/1967	DUI	80-000859	Old	20th	Cape Coral PD	Cape Coral PD	6/21/2006	20:35:16	First Sample Volume 0.429L - No Flag Second Sample Volume 0.194L - Flagged Volume Not Met Third Sample Volume 0.625L - No Flag	Not A Valid Test - First and Third Samples Not Flagged
Lisa	Denise A.	12/9/1957	DUI	80-000859	Old	20th	Cape Coral PD	Cape Coral PD	7/10/2006	0:54:05	First Sample Volume 0.464L - No Flag Second Sample Volume 0.168L - Flagged Volume Not Met Third Sample Volume 0.582L - No Flag	Not A Valid Test - First and Third Samples Not Flagged
Hyland	Cara L.	8/23/1986	DUI	80-000859	Old	20th	Cape Coral PD	Cape Coral PD	8/2/2006	2:37:55	First Sample Volume 1.035L - Flagged Volume Not Met Second Sample Volume 0.980L - No Flag Third Sample Volume 0.015L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Velez	Emesio	7/22/1966	DUI	80-000859	Old	20th	Cape Coral PD	Cape Coral PD	8/6/2006	1:33:11	First Sample Volume 0.097L - Flagged Volume Not Met Second Sample Volume 0.457L - No Flag Third Sample Volume 0.687L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples and Second Control Test Outside Tolerance
Jarrett	Donald	12/28/1955	DUI	80-000861	Old	12th	Sarasota County SO	Sarasota County SO	9/12/2006	21:54:31	First Sample Volume 0.035L - No Flag No Second Sample - Subject Refused	Not A Valid Test Refusal
Brown	Nathan	5/11/1979	DUI	80-000868	Old	16th	Monroe County SO	NAS Key West	7/30/2006	2:11:35	First Sample Volume 0.250L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Moriniere	Laurent L.	6/24/1947	BUI	80-000869	Old	16th	Monroe County SO	Other State	4/15/2006	20:19:14	First Sample Volume 0.480L - No Flag Second Sample Volume 0.179L - Flagged Volume Not Met Third Sample Volume 1.804L - OK	Not A Valid Test - Second Control Test Outside Tolerance
Clarke	Brian D.	12/18/1961	BUI	80-000869	Old	16th	Monroe County SO	FWC	5/20/2006	21:23:30	First Sample Volume 0.687L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Cornejo	Jose	4/17/1980	DUI	80-000869	Old	16th	Monroe County SO	Monroe County SO	7/22/2006	4:48:56	First Sample Volume 1.246L - OK Second Sample Volume 1.246L - OK	Not A Valid Test First Sample Not Flagged
Kelly	Shane L.	2/6/1981	DUI	80-000869	Old	16th	Monroe County SO	Monroe County SO	8/7/2006	5:16:12	First Sample Volume 0.449L - Flagged Volume Not Met Second Sample Volume 0.535L - No Flag Third Sample Volume 1.585L - OK	Not A Valid Test - Second Sample Not Flagged
La Paz	Gilda	11/20/1943	DUI	80-000871	Old	11th	Coral Gables PD	Coral Gables PD	6/4/2006	2:28:46	Second Control Test Outside Tolerance	Not A Valid Test Second Control Test Outside Tolerance
Montenegro	Maria H.	5/6/1945	DUI	80-000871	New	11th	Coral Gables PD	Coral Gables PD	10/12/2006	10:30:25	First Sample Volume 0.375L - No Flag Second Sample Volume 1.929L - OK	Not A Valid Test - First Sample Not Flagged
Costa Moran	Enrique A.	2/27/1975	DUI	80-000872	Old	11th	Miami PD	Miami PD	8/25/2006	3:03:40	First Sample Volume 1.578L - OK Second Sample Volume 0.328L - No Flag	Not A Valid Test - Second Sample Not Flagged
Mena	Hollman E.	12/5/1984	DUI	80-000872	Old	11th	Miami PD	Miami PD	9/11/2006	9:52:59	First Sample Volume 1.078L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Nunez	Manuel J.	1/24/1892	DUI	80-000872	New	11th	Miami PD	Miami PD	10/7/2006	8:03:03	First Sample Volume 1.179L - OK Second Sample Volume 0.945L - No Flag	Not A Valid Test - Second Sample Not Flagged
Mena	Hollman E.	12/5/1984	DUI	80-000873	Old	11th	Miami PD	Miami PD	9/11/2006	9:26:14	First Sample Volume 0.808L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Diaz	Gerardo	7/30/1929	DUI	80-000873	Old	11th	Miami PD	Miami PD	4/1/2006	22:22:46	First Sample Volume 0.535L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Knowles	Patrick J.	5/28/1966	DUI	80-000873	Old	11th	Miami PD	South Miami PD	9/16/2006	1:55:54	First Sample Volume 1.101L - OK Second Sample Volume 1.027L - No Flag	Not A Valid Test - Second Sample Not Flagged
Alonso	Ruben	11/22/1959	DUI	80-000876	Old	11th	Hialeah PD	Hialeah PD	5/1/2006	20:40:29	First Sample Volume 0.312L - No Flag Second Sample Volume 1.105L - OK	Not A Valid Test - First Sample Not Flagged
Franco	Luis A.	9/30/1967	DUI	80-000876	Old	11th	Hialeah PD	Hialeah PD	6/26/2006	21:54:09	First Sample Volume 0.718L - No Flag Second Sample Volume 1.246L - OK Third Sample Volume 1.148L - OK	Not A Valid Test - Second Control Test Outside Tolerance
Bello	Farah T.	10/14/1978	DUI	80-000879	Old	11th	Miam Beach PD	Miam Beach PD	9/15/2006	5:27:26	First Sample Volume 0.687L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Lozano	Cesar U.	10/7/1975		80-000881	Old	11th	Miami Dade PD	Miami Dade PD	9/20/2006	2:09:52	First Sample Volume 0.304L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Martinez	Gilver O.	9/2/1984	DUI	80-000884	Old	11th	Miami Dade PD	Miami Dade PD	4/9/2006	4:38:12	First Sample Volume 2.617L - OK Second Sample Volume 0.460L - No Flag	Not A Valid Test - Second Sample Not Flagged
Burgos-Donates	Myrna	8/9/1957	DUI	80-000885	Old	11th	Miami Dade PD	Miami Dade PD	5/6/2006	2:43:50	First Sample Volume 0.273L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Velasquez	Valerie A.	3/1/1981	DUI	80-000885	Old	11th	Miami Dade PD	Miami Dade PD	6/10/2006	5:27:19	First Sample Volume 0.433L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Mena	Marianne	10/28/1983	DUI	80-000885	Old	11th	Miami Dade PD	Miami Dade PD	8/20/2006	6:53:05	First Sample Volume 0.339L - No Flag Second Sample Volume 0.132L - Flagged Volume Not Met Third Sample Volume 0.976L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Mena	Marianne	10/28/1983	DUI	80-000885	Old	11th	Miami Dade PD	Miami Dade PD	8/20/2006	7:13:49	First Sample Volume 0.117L - Flagged Volume Not Met Second Sample Volume 0.500L - No Flag Third Sample Volume 0.871L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples

Winston	Alan	4/17/1942	DUI	80-000886	Old	11th	North Miami Beach PD	North Miami Beach PD	4/22/2006	5:40:59	First Sample Volume 0.632L - No Flag Second Sample Volume 0.597L - No Flag	Not A Valid Test - First and Second Sample Not Flagged
Ghazarian	Sevak	9/9/1984	DUI	80-000888	Old	6th	Pinellas County SO	Pinellas County SO	6/2/2006	3:28:14	First Sample Volume 0.226L - No Flag Second Sample Volume 0.304L - Flagged Volume Not Met Third Sample Volume 0.101L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Harvey	Raybeth	8/22/1959	DUI	80-000889	Old	6th	Pinellas County SO	Florida Highway Patrol	9/1/2006	21:20:50	First Sample Volume 1.203L - OK Second Sample Volume 0.089L - Flagged Volume Not Met Third Sample Volume 0.507L - No Flag	Not A Valid Test - Third Sample Not Flagged
Lee	Scott W	3/10/1973	DUI	80-000890	Old	6th	Pinellas County SO	Pinellas County SO	5/20/2006	4:31:38	First Sample Volume 0.777L - No Flag Second Sample Volume 1.074L - Flagged Volume Not Met Third Sample Volume 1.167L - OK	Not A Valid Test - First Sample Not Flagged
Rainbow	Rosa S	3/16/1964	DUI	80-000890	New	6th	Pinellas County SO	Florida Highway Patrol	10/15/2006	23:59:55	First Sample Volume 0.714L - No Flag Second Sample Volume 1.125L - OK	Not A Valid Test - First Sample Not Flagged
Davis	Kevin M	5/20/1969	BUJ	80-000892	Old	14th	Fish & Wildlife CC	FWC	5/20/2006	14:37:25	First Sample Volume 1.058L - No Flag No Second Sample - Improper Sample	Not A Valid Test - Need At Least Two Samples
Lopez	Torbio M	10/5/1981	DUI	80-000907	Old	12th	Manatee County SO	Florida Highway Patrol	8/12/2006	0:51:18	First Sample Volume 1.441L - OK Second Sample Volume 0.554L - No Flag Third Sample Volume 2.554L - OK	Valid Test - First and Third Samples
Tamr	Trinity T.	6/28/1977	DUI	80-000912	Old	12th	Manatee County SO	Manatee County SO	4/19/2006	5:10:09	First Sample Volume 0.015L - Flagged Volume Not Met Second Sample Volume 0.082L - No Flag Third Sample Volume 0.296L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Skolroo	Richard D	6/29/1962	DUI	80-000912	Old	12th	Manatee County SO	Florida Highway Patrol	8/28/2006	3:01:18	First Sample Volume 0.699L - No Flag Second Sample Volume 1.339L - OK	Not A Valid Test - First Sample Not Flagged
Deleon	Maria C	5/22/1984	DUI	80-000912	New	12th	Manatee County SO	Manatee County SO	11/2/2006	2:01:08	First Sample Volume 0.320L - Flagged Volume Not Met Second Sample Volume 0.332L - No Flag Third Sample Volume 0.015L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Deleon	Maria C	5/22/1984	DUI	80-000912	New	12th	Manatee County SO	Manatee County SO	11/2/2006	2:22:58	First Sample Volume 0.121L - No Flag Second Sample Volume 0.101L - Flagged Volume Not Met No Third Sample - Subject Refused	Not A Valid Test - Refusal
Hurtado	Pablo E	3/22/1967	DUI	80-000939	Old	20th	Lee County SO	Lee County SO	4/27/2006	2:24:56	First Sample Volume 2.125L - OK Second Sample Volume 0.550L - No Flag	Not A Valid Test - Second Sample Not Flagged
Fenno	Lillian	4/26/1962	DUI	80-000939	Old	20th	Lee County SO	Lee County SO	6/12/2006	15:27:43	First Sample Volume 0.289L - Flagged Volume Not Met Second Sample Volume 0.835L - No Flag Third Sample Volume 1.496L - OK	Not A Valid Test - Second Sample Not Flagged
Moore	Shannon C.	10/25/1982	DUI	80-000939	Old	20th	Lee County SO	Lee County SO	7/21/2006	0:18:42	First Sample Volume 0.347L - No Flag Second Sample Volume 1.871L - OK	Not A Valid Test - First Sample Not Flagged
Conaway	Doreen M.	11/15/1965	DUI	80-000939	Old	20th	Lee County SO	Lee County SO	7/23/2006	2:57:31	First Sample Volume 0.375L - No Flag Second Sample Volume 0.000L - Flagged Volume Not Met No Third Sample - Subject Refused	Not A Valid Test - Refusal
Daughtry	Amanda S	2/11/1972	DUI	80-000940	Old	18th	Brevard County SO	Brevard County SO	8/30/2006	3:25:19	First Sample Volume 1.605L - OK Second Sample Volume 0.230L - No Flag Third Sample Volume 0.757L - No Flag	Not A Valid Test - Second and Third Samples Not Flagged
Cohen	Nikol A	6/6/1975	DUI	80-000943	Old	20th	Collier County SO	Naples PD	5/28/2006	3:00:50	First Sample Volume 0.613L - No Flag Second Sample Volume 1.816L - OK	Not A Valid Test - First Sample Not Flagged
Lyach	Vladimir	2/27/1958	DUI	80-000943	Old	20th	Collier County SO	Collier County SO	8/7/2006	21:20:07	First Sample Volume 2.144L - OK Second Sample Volume 0.218L - No Flag Third Sample Volume 1.875L - OK Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance
Hall	Deborah L	11/25/1956	DUI	80-000949	Old	2nd	Tallahassee PD	Tallahassee PD	7/26/2006	14:18:55	First Sample Volume 0.941L - No Flag Second Sample Volume - 1.417L - OK	Not A Valid Test
Bravo-Neitro	Francisco	5/8/1978	DUI	80-000951	Old	20th	Hendry County SO	Hendry County SO	7/30/2006	2:44:40	First Sample Volume 0.164L - No Flag Second Sample Volume 2.507L - OK	First Sample Not Flagged
Castillo	Ruben	4/15/1974	DUI	80-000951	New	20th	Hendry County SO	Hendry County SO	10/8/2006	0:08:39	First Sample Volume 0.265L - No Flag No Second Sample - Subject Refused	Not A Valid Test - First Sample Not Flagged
Adams	Rodney B	6/2/1961	DUI	80-000957	Old	2nd	Leon County SO	Leon County SO	3/30/2006	19:14:15	First Sample Volume 1.316L - OK Second Sample Volume 0.410L - No Flag Third Sample Volume 1.570L - OK	Valid Test - First and Third Samples
Tucker	Leroy	10/2/1981	DUI	80-000958	Old	2nd	Leon County SO	Tallahassee PD	4/15/2006	22:23:00	First Sample Volume 0.429L - No Flag Second Sample Volume 2.250L - OK	Not A Valid Test
Furman	Shawn A	10/21/1975	DUI	80-000958	Old	2nd	Leon County SO	Tallahassee PD	9/9/2006	1:18:33	First Sample Volume 0.347L - No Flag Second Sample Volume 0.000L - Flagged Volume Not Met Third Sample Volume 2.382L - OK	Not A Valid Test - First Sample Not Flagged
Childs	Kendra P	9/8/1969	DUI	80-000961	Old	2nd	Wakulla County SO	Wakulla County SO	4/14/2006	23:27:48	First Sample Volume 0.675L - No Flag Second Sample Volume 0.015L - Flagged Volume Not Met No Third Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Uriostegui	Marcos	2/28/1982	DUI	80-000963	Old	9th	Orange County SO	Orlando PD	5/17/2006	4:51:48	First Sample Volume 0.191L - Flagged Volume Not Met Second Sample Volume 0.183L - No Flag Third Sample Volume 1.453L - OK	Not A Valid Test - Second Sample Not Flagged

Altrabi	Fouad	4/2/1977	DUI	80-000963	Old	9th	Orange County SO	Orange County SO	5/29/2006	18 30:03	First Sample Volume 0.546L - No Flag Second Sample Volume 0.402L - Flagged Volume Not Met Third Sample Volume 3.941L - OK	Not A Valid Test First Sample Not Flagged
Miranda	Arnoldo	5/6/1983	DUI	80-000964	Old	9th	Orange County SO	Orange County SO	9/24/2006	4 58:41	First Sample Volume 0.765L - No Flag Second Sample Volume 1.933L - OK Third Sample Volume 1.453L - OK	Not A Valid Test First Sample Not Flagged
Rodriguez	Candido J	8/8/1980	DUI	80-000966	Old	9th	Osceola County Corrections	Osceola County SO	4/22/2006	3 00:40	First Sample Volume 0.457L - No Flag Second Sample Volume 2.027L - OK Third Sample Volume 2.027L - OK Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance
Cannon	Heather M	3/3/1987	DUI	80-000999	Old	18th	Seminole County SO	Seminole County SO	9/22/2006	5 13:02	First Sample Volume 0.230L - No Flag Second Sample Volume 0.031L - Flagged Volume Not Met Third Sample Volume 0.554L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Hatazin	Linda S	8/24/1947	DUI	80-001003	Old	6th	Pinellas County SO	Pinellas County SO	5/4/2006	4 47:35	First Sample Volume 0.394L - No Flag Second Sample Volume 0.429L - Flagged Volume Not Met Third Sample Volume 0.359L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Chapman	Douglas R.	12/25/1924	DUI	80-001004	Old	18th	Brevard County SO	Brevard County SO	6/25/2006	1 36:28	First Sample Volume 0.421L - No Flag Second Sample Volume 0.144L - Flagged Volume Not Met Third Sample Volume 1.250L - OK	Not A Valid Test First Sample Not Flagged
Irby	Howard J	10/7/1962	DUI	80-001005	Old	6th	Pinellas County SO	Pinellas County SO	4/11/2006	0 17:44	First Sample Volume 1.636L - OK Second Sample Volume 1.062L - No Flag Third Sample Volume 1.644L - OK	Not A Valid Test - Second Sample Not Flagged
Richmond	Aubrey L.	2/9/1950	DUI	80-001005	Old	6th	Pinellas County SO	Pinellas County SO	4/15/2006	1 26:05	Second Sample Volume 0.710L - No Flag First Sample Volume 0.574L - No Flag	Not A Valid Test - Second Sample Not Flagged
Maggs	Robert F	9/29/1938	DUI	80-001005	Old	6th	Pinellas County SO	Pinellas County SO	7/26/2006	20:38 58	First Sample Volume 0.308L - Flagged Volume Not Met Second Sample Volume 0.261L - Flagged Volume Not Met Third Sample Volume 1.152L - OK	Not A Valid Test - Need At Least Two Samples
Roundtree	Aaron C	8/22/1971	DUI	80-001005	Old	6th	Pinellas County SO	Pinellas County SO	8/20/2006	3 17:46	First Sample Volume 0.441L - Flagged Volume Not Met Second Sample Volume 0.441L - Flagged Volume Not Met Third Sample Volume 1.52L - OK	Not A Valid Test First Sample Not Flagged
Dolch	Matthew W	2/1/1981	DUI	80-001005	New	6th	Pinellas County SO	Pinellas County SO	9/22/2006	4 08:27	First Sample Volume 0.804L - No Flag Second Sample Volume 1.542L - OK	Not A Valid Test - First Sample Not Flagged
Segura-Guadarrama	Ivan	3/24/1984	DUI	80-001007	New	10th	Haines City PD	Haines City PD	9/24/2006	4 22:36	First Sample Volume 0.347L - Flagged Volume Not Met Second Sample Volume 0.199L - No Flag Third Sample Volume 0.078L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Crews	Earl J	12/16/1957	DUI	80-001007	New	10th	Haines City PD	Haines City PD	10/7/2006	21 06:58	Second Sample Volume 0.441L - No Flag Third Sample Volume 0.394L - No Flag Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance
Smith	Heather L	10/10/1987	DUI	80-001009	Old	18th	Palm Bay PD	Palm Bay PD	8/28/2006	4 23:50	First Sample Volume 0.429L - No Flag Second Sample Volume 1.648L - OK	Not A Valid Test First Sample Not Flagged
Thompson	Catherine S.	2/18/1958	DUI	80-001009	Old	18th	Palm Bay PD	Palm Bay PD	9/11/2006	0 12:25	First Sample Volume 0.714L - Flagged Volume Not Met Second Sample Volume 0.281L - No Flag Third Sample Volume 0.785L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Zanchez	Roberto	1/30/1981	DUI	80-001043	Old	10th	Highlands County SO	Highlands County SO	5/21/2006	0 31:19	First Sample Volume 0.109L - No Flag No Second Sample - RFI	Not A Valid Test At Least Two Samples Test Aborted Due to RFI
Cavinder	Brenda A	9/23/1960	DUI	80-001043	Old	10th	Highlands County SO	Sebring PD	6/1/2006	1 56:15	First Sample Volume 1.878L - OK Second Sample Volume 0.687L - No Flag Third Sample Volume 1.402L - OK	Valid Test - First and Third Samples
Marley	Lynne C	1/10/1959	DUI	80-001043	Old	10th	Highlands County SO	Florida Highway Patrol	6/10/2006	0 10:08	First Sample Volume 0.382L - No Flag Second Sample Volume 1.640L - OK	Not A Valid Test - First Sample Not Flagged
Honorat	Hermene	1/9/1956	DUI	80-001044	Old	17th	Coconut Creek PD	Coconut Creek PD	6/26/2006	23 34:39	First Sample Volume 1.753L - OK Second Sample Volume 0.304L - No Flag	Not A Valid Test - Second Sample Not Flagged
Osborne	Michele J	11/2/1966	DUI	80-001047	New	17th	Broward County SO	Broward County SO	10/13/2006	2 13:24	First Sample Volume 0.128L - Flagged Volume Not Met Second Sample Volume 1.019L - No Flag Third Sample Volume 0.015L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Blair	Jason L.	7/14/1975	DUI	80-001051	Old	6th	St. Petersburg PD	St. Petersburg PD	7/4/2006	19 14:37	First Sample Volume 0.351L - No Flag Second Sample Volume 1.863L - OK No Third Sample - Subject Refused	Not A Valid Test Refusal
Pruett	Dennis R.	12/4/1951	DUI	80-001051	Old	6th	St. Petersburg PD	St. Petersburg PD	8/18/2006	19 07:01	First Sample Volume 0.402L - No Flag Second Sample Volume 1.550L - OK Third Sample Volume 1.363L - OK	Valid Test - Second and Third Samples
Zuniga	Sherry C	7/11/1955	DUI	80-001055	Old	10th	Polk County SO	Florida Highway Patrol	7/29/2006	2 07:52	First Sample Volume 0.667L - No Flag Second Sample Volume 1.503L - OK	Not A Valid Test - First Sample Not Flagged
Addleman	Holly L.	10/6/1977	DUI	80-001057	Old	10th	Lakeland PD	Lakeland PD	6/18/2006	3 23:44	First Sample Volume 0.300L - No Flag Second Sample Volume 1.125L - OK	Not A Valid Test - First Sample Not Flagged
Rivera	Victor M	4/5/1950	DUI	80-001057	Old	10th	Lakeland PD	Lakeland PD	6/20/2006	3 12:02	First Sample Volume 0.699L - No Flag Second Sample Volume 1.656L - OK	Not A Valid Test - First Sample Not Flagged
Godinez Perez	Alberto	1/11/1975	DUI	80-001063	Old	17th	Hollywood PD	Hollywood PD	7/1/2006	2 21:25	First Sample Volume 0.109L - No Flag Second Sample Volume 1.488L - OK Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance

Miradakis	Kimberly A	5/7/1960	DUI	80-001065	New	6th	Tarpon Springs PD	Tarpon Springs PD	9/23/2006	3:26:04	First Sample Volume 0.035L - Flagged Volume Not Met Second Sample Volume 0.808L - No Flag Third Sample Volume 0.164L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Buchanan	Bonnie T	12/11/1940	DUI	80-001067	Old	17th	Broward County SO	Broward County SO	5/27/2006	19:03:55	First Sample Volume 0.546L - Flagged Volume Not Met Second Sample Volume 0.738L - No Flag Third Sample Volume 1.105L - OK	Not A Valid Test - Second Sample Not Flagged
Lanot	Jesufranc	11/17/1971	DUI	80-001067	Old	17th	Broward County SO	Broward County SO	7/11/2006	22:47:55	First Sample Volume 0.718L - Flagged Volume Not Met Second Sample Volume 1.574L - OK Third Sample Volume 0.218L - No Flag	Not A Valid Test - Third Sample Not Flagged
Nelson	Stacy A	7/22/1963	DUI	80-001069	Old	4th	Jacksonville SO	Jacksonville SO	8/21/2006	4:30:22	First Sample Volume 1.093L - No Flag Second Sample Volume 1.957L - OK	Not A Valid Test
Doyle	Stephen V	7/24/1952	DUI	80-001069	Old	4th	Jacksonville SO	Jacksonville SO	9/5/2006	1:54:10	First Sample Volume 1.382L - OK Second Sample Volume 0.703L - No Flag	Not A Valid Test - Second Sample Not Flagged
Buckley	Linda G	5/10/1967	DUI	80-001071	Old	4th	Clay County SO	Clay County SO	5/6/2006	2:11:12	First Sample Volume 0.671L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Davis	Kim P	11/20/1957	DUI	80-001082	Old	17th	Plantation PD	Sunrise PD	5/9/2006	16:54:56	First Sample Volume 0.550L - No Flag Second Sample Volume 0.574L - Flagged Volume Not Met Third Sample Volume 1.457L - OK	Not A Valid Test - First Sample Not Flagged
Nguyen	Thai H	8/2/1963	DUI	80-001083	Old	4th	Jacksonville SO	Jacksonville SO	5/14/2006	22:56:41	First Sample Volume 0.671L - No Flag Second Sample Volume 0.714L - Flagged Volume Not Met Third Sample Volume 0.179L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Gagne	Audra J	12/29/1966	DUI	80-001117	Old	6th	Pasco County SO	Pasco County SO	4/25/2006	2:54:49	First Sample Volume 0.949L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Drake	Dawn D	6/9/1970	DUI	80-001117	Old	6th	Pasco County SO	Pasco County SO	4/27/2006	1:28:39	First Sample Volume 2.578L - OK Second Sample Volume 0.718L - No Flag	Not A Valid Test - Second Sample Not Flagged
Brooks	Geoffrey D	4/20/1972	DUI	80-001131	Old	7th	Volusia County SO	Volusia County SO	7/4/2006	2:15:13	First Sample Volume 1.015L - No Flag Second Sample Volume 1.101L - OK	Not A Valid Test
Brock	Lyn	7/20/1942	DUI	80-001137	Old	7th	Flagler County SO	Flagler County SO	7/22/2006	18:30:02	First Sample Volume 0.097L - Flagged Volume Not Met Second Sample Volume 0.019L - No Flag Third Sample Volume 0.511L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Kenyon	John R	11/20/1979	DUI	80-001139	Old	17th	Broward County SO	Broward County SO	6/14/2006	1:48:17	First Sample Volume 1.261L - OK Second Sample Volume 0.972L - No Flag	Not A Valid Test - Second Sample Not Flagged
Jean-Jacques	Paholo	5/10/1979	DUI	80-001139	Old	17th	Broward County SO	Broward County SO	7/8/2006	5:35:25	First Sample Volume 0.410L - No Flag Second Sample Volume 1.519L - OK Third Sample Volume 1.750L - OK	Valid Test - Second and Third Samples
Palacios	Mirian L	1/21/1978	DUI	80-001139	Old	17th	Broward County SO	Broward County SO	7/22/2006	3:46:52	First Sample Volume 0.519L - No Flag Second Sample Volume 0.097L - Flagged Volume Not Met Third Sample Volume 0.414L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Hendry	Mark J	5/31/1960	DUI	80-001139	Old	17th	Broward County SO	Broward County SO	9/22/2006	1:11:06	First Sample Volume 1.406L - OK Second Sample Volume 0.531L - No Flag	Not A Valid Test - Second Sample Not Flagged
Barrera	Aristides E	8/21/1976	DUI	80-001153	Old	2nd	Leon County SO	Leon County SO	4/13/2006	21:55:59	First Sample Volume 0.421L - No Flag Second Sample Volume 0.191L - No Flag	Not A Valid Test - First and Second Sample Not Flagged
Gibson	Amelia C	7/30/1986	DUI	80-001153	Old	2nd	Leon County SO	Leon County SO	8/23/2006	5:26:19	First Sample Volume 1.046L - No Flag Second Sample Volume 0.335L - Flagged Volume Not Met Third Sample Volume 1.199L - OK	Not A Valid Test
Waggoner	Nicholas C	5/4/1991	DUI	80-001158	Old	7th	Ponce Inlet PD	Daytona Beach PD	4/5/2006	23:12:41	First Sample Volume 0.367L - No Flag Second Breath Sample 0.125L - Flagged Volume Not Met No Third Sample - Sequence Aborted	Not A Valid Test - Need At Least Two Samples
Currie II	Robert A	5/11/1955	DUI	80-001166	Old	18th	Brevard County SO	Brevard County SO	4/13/2006	20:11:17	First Sample Volume 2.066L - OK Second Sample Volume 1.058L - No Flag	Not A Valid Test - Second Sample Not Flagged
Ramharack	Rajnauth	10/28/1983	DUI	80-001174	New	5th	Clermont PD	Clermont PD	10/15/2006	23:11:23	Second Sample Volume 1.843L - OK Third Sample Volume 2.464L - OK Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance
Zhang	Dong	6/24/1967	DUI	80-001190	Old	15th	Boynton Beach PD	Boynton Beach PD	6/17/2006	8:35:40	First Sample Volume 0.828L - No Flag Second Sample Volume 2.519L - OK	Not A Valid Test - First Sample Not Flagged
Ferranti	Edward J	5/26/1970	DUI	80-001190	Old	15th	Boynton Beach PD	Boynton Beach PD	7/22/2006	3:11:52	First Sample Volume 0.015L - Flagged Volume Not Met Second Sample Volume 2.066L - OK Third Sample Volume 0.726L - No Flag	Not A Valid Test - Third Sample Not Flagged
Kaszak	Kenneth	4/16/1958	DUI	80-001190	Old	15th	Boynton Beach PD	Boynton Beach PD	8/13/2006	0:46:51	First Sample Volume 1.406L - OK Second Sample Volume 0.441L - No Flag	Not A Valid Test - Second Sample Not Flagged
Perez-Oniz	Issac	10/23/1982	DUI	80-001235	Old	15th	West Palm Beach PD	West Palm Beach PD	4/8/2006	0:03:15	First Sample Volume 1.285L - OK Second Sample Volume 0.488L - No Flag Third Sample Volume 1.281L - OK	Valid Test - First and Third Samples
Rucker	Curtis	3/12/1945	DUI	80-001248	Old	17th	Broward County SO	Broward County SO	5/7/2006	0:02:11	First Sample Volume 0.593L - No Flag Second Sample Volume 0.816L - No Flag	Not A Valid Test - First and Second Sample Not Flagged

Silva	Natalia R.	7/2/1986	DUI	80-001248	New	17th	Broward County SO	Other State	9/30/2006	9.22.00	First Sample Volume 0.933L - No Flag Second Sample Volume 0.015L - Flagged Volume Not Met Third Sample Volume 1.449L - OK	Not A Valid Test - First Sample Not Flagged
Menner	Patricia	11/3/1968	DUI	80-001253	Old	18th	Brevard County SO	Brevard County SO	6/26/2006	22.40.39	First Sample Volume 0.347L - No Flag Second Sample Volume 1.765L - OK	Not A Valid Test - First Sample Not Flagged
Brown	Joshua B.	8/10/1976	DUI	80-001253	Old	18th	Brevard County SO	Brevard County SO	8/26/2006	23.45.57	First Sample Volume 0.468L - No Flag Second Sample Volume 1.785L - OK	Not A Valid Test - First Sample Not Flagged
Doane	Richard G.	10/26/1963	DUI	80-001255	Old	9th	Orange County SO	Orange County SO	4/7/2006	3.46.24	First Sample Volume 0.644L - No Flag Second Sample Volume 1.242L - OK	Not A Valid Test First Sample Not Flagged
Aponte	Egardo	10/9/1979	DUI	80-001255	Old	9th	Orange County SO	Orlando PD	9/9/2006	1.43.01	First Sample Volume 0.679L - No Flag Second Sample Volume 0.160L - Flagged Volume Not Met No Third Sample - Subject Refused	Not A Valid Test Refusal
Colligan	James J.	1/25/1973	DUI	80-001255	New	9th	Orange County SO	Edgewood PD	10/5/2006	4.07.05	First Sample Volume 0.593L - Flagged Volume Not Met Second Sample Volume 0.335L - No Flag Third Sample Volume 0.937L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Quiros	Ovalle	12/15/1978	DUI	80-001257	Old	9th	Orange County SO	Orlando PD	3/30/2006	4.45.18	First Sample Volume 1.628L - OK Second Sample Volume 0.480L - No Flag Third Sample Volume 3.035L - OK	Valid Test - First and Third Samples
Meinken	Bradford G.	11/3/1976	DUI	80-001257	Old	9th	Orange County SO	Orlando PD	8/5/2006	4.53.30	First Sample Volume 0.746L - No Flag Second Sample Volume 1.128L - OK Third Sample Volume 1.351L - OK	Valid Test - Second and Third Samples
Tripp	John L.	2/23/1964	DUI	80-001257	New	9th	Orange County SO	Florida Highway Patrol	10/6/2006	7.45.02	First Sample Volume 1.109L - OK Second Sample Volume 0.078L - No Flag	Not A Valid Test - Second Sample Not Flagged
Thomas	Jean C.	7/21/1953	DUI	80-001258	Old	9th	Orange County SO	Florida Highway Patrol	5/6/2006	21.11.50	First Sample Volume 0.339L - No Flag Second Sample Volume 1.312L - OK	Not A Valid Test First Sample Not Flagged
Mills	Brian R.	11/9/1977	DUI	80-001258	New	9th	Orange County SO	Florida Highway Patrol	10/12/2006	3.34.59	First Sample Volume 1.066L - No Flag No Second Sample - Improper Sample	Not A Valid Test - Need At Least Two Samples
McDaniel	Joseph N.	10/6/1937	DUI	80-001260	Old	18th	Cocoa PD	Brevard County SO	7/15/2006	1.37.50	First Sample Volume 0.468L - No Flag Second Sample Volume 1.406L - OK	Not A Valid Test - First Sample Not Flagged
Gause	Julia F.	1/21/1969	DUI	80-001262	Old	18th	Melbourne PD	Melbourne PD	5/29/2006	22.12.15	First Sample Volume 1.164L - OK Second Sample Volume 1.039L - No Flag	Not A Valid Test - Second Sample Not Flagged
Young	Pamela	8/5/1958	DUI	80-001267	Old	15th	Palm Beach County SO	Riveria Beach PD	5/30/2006	1.40.07	First Sample Volume 0.296L - Flagged Volume Not Met Second Sample Volume 0.519L - No Flag No Third Sample - Subject Refused	Not A Valid Test Refusal
Keenan	Maureen E.	10/9/1970	DUI	80-001267	Old	15th	Palm Beach County SO	Palm Beach County SO	7/19/2006	1.31.08	First Sample Volume 0.097L - Flagged Volume Not Met Second Sample Volume 1.363L - OK Third Sample Volume 0.367L - No Flag	Not A Valid Test - Third Sample Not Flagged
Villaloro	Yener R.	8/24/1984	DUI	80-001267	Old	15th	Palm Beach County SO	Palm Beach County SO	8/6/2006	4.53.20	First Sample Volume 0.972L - No Flag No Second Sample - Subject Refused	Not A Valid Test Refusal
Suarez	Jean	3/31/1969	DUI	80-001269	Old	15th	Palm Beach County SO	Florida Highway Patrol	5/5/2006	6.55.45	First Sample Volume 1.226L - OK Second Sample Volume 0.714L - No Flag Third Sample Volume 1.769L - OK	Valid Test - First and Third Samples
Goldberg	Paul E.	3/18/1969	DUI	80-001269	Old	15th	Palm Beach County SO	Florida Highway Patrol	8/24/2006	3.44.46	First Sample Volume 1.445L - OK Second Sample Volume 0.609L - No Flag Third Sample Volume 1.437L - OK	Valid Test - First and Third Samples
Toft	Carole A.	8/6/1964	DUI	80-001269	Old	15th	Palm Beach County SO	Palm Beach County SO	9/21/2006	4.12.02	First Sample Volume 0.253L - No Flag Second Sample Volume 1.207L - OK	Not A Valid Test - First Sample Not Flagged
Cornell	Davied E.	2/21/1965	DUI	80-001270	Old	15th	Palm Beach County SO	Delray Beach PD	7/23/2006	22.32.26	First Sample Volume 0.613L - No Flag No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Blair	Robert E.	2/14/1946	DUI	80-001270	New	15th	Palm Beach County SO	Palm Beach County SO	10/7/2006	0.18.56	First Sample Volume 1.398L - OK Second Sample Volume 0.921L - No Flag	Not A Valid Test - Second Sample Not Flagged
Reyes	Iris Y.	8/8/1980	DUI	80-001270	New	15th	Palm Beach County SO	Florida Highway Patrol	10/7/2006	4.02.42	First Sample Volume 0.464L - No Flag Second Sample Volume 1.074L - Flagged Volume Not Met Third Sample Volume 0.171L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Spain	Robert H.	8/21/1951	DUI	80-001270	New	15th	Palm Beach County SO	Riveria Beach PD	10/9/2006	17.11.53	First Sample Volume 2.000L - OK Second Sample Volume 0.710L - No Flag	Not A Valid Test - Second Sample Not Flagged
Rondon	Brian S.	9/8/1980	DUI	80-001270	New	15th	Palm Beach County SO	Palm Beach County SO	10/15/2006	5.27.30	First Sample Volume 0.714L - No Flag Second Sample Volume 1.597L - OK	Not A Valid Test - First Sample Not Flagged
Harri	Kareen R.	6/3/1968	DUI	80-001272	Old	18th	Seminole County SO	Oviedo PD	4/2/2006	5.47.26	First Sample Volume 0.867L - No Flag Second Sample Volume 1.703L - OK	Not A Valid Test - First Sample Not Flagged
Wells	Barry N.	9/7/1968	DUI	80-001272	Old	18th	Seminole County SO	Lake Mary PD	4/10/2006	19.11.26	First Sample Volume 0.710L - No Flag Second Sample Volume 0.332L - Flagged Volume Not Met Third Sample Volume 0.785L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Wilson	Glen M.	9/15/1964	DUI	80-001272	Old	18th	Seminole County SO	Florida Highway Patrol	6/25/2006	2.12.47	First Sample Volume 0.675L - No Flag Second Sample Volume 0.796L - OK Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance



White	Leslie J.	6/6/1960	DUI	80-001272	Old	18th	Seminole County SO	Oviedo PD	7/2/2006	2:46:00	Second Sample Volume 0.257L - Flagged Volume Not Met Third Sample Volume 0.679L - No Flag	Not A Valid Test - Second Control Test Outside Tolerance
Newton	Justin B.	10/21/1978	DUI	80-001272	Old	18th	Seminole County SO	Seminole County SO	7/21/2006	22:17:44	First Sample Volume 0.457L - No Flag Second Sample Volume 1.390L - OK	Not A Valid Test - First Sample Not Flagged
Lopez-Hernandez	David	4/4/1978	DUI	80-001273	Old	6th	Pinellas County SO	Pinellas County SO	7/15/2006	3:03:14	First Sample Volume 0.050L - No Flag Second Sample Volume 1.414L - OK Third Sample Volume 1.613L - OK	Valid Test - Second and Third Samples
Portuondo	Erik J.	12/11/1975	DUI	80-001276	Old	11th	Miami Dade PD	Miami Dade PD	5/12/2006	0:34:06	No Second Sample - Purge Fail	Not A Valid Test - Need At Least Two Samples
Fandino	Marta L.	8/18/1967	DUI	80-001276	Old	11th	Miami Dade PD	Miami Dade PD	5/27/2006	4:12:02	First Sample Volume 0.453L - No Flag Second Sample Volume 1.839L - OK Third Sample Volume - 1.292L - OK	Valid Test - Second and Third Samples
Fuhr	Alison R.	10/31/1978	DUI	80-001282	Old	8th	University of Florida PD	University of FL PD	4/13/2006	3:34:42	First Sample Volume 0.867L - No Flag Second Sample Volume 0.152L - Flagged Volume Not Met Third Sample Volume 0.082L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Johns	Cynthia J.	5/18/1966	DUI	80-001284	Old	8th	Gluchrist County SO	Cedar Keys PD	6/15/2006	1:38:28	First Sample Volume 0.421L - No Flag Second Sample Volume 0.378L - Flagged Volume Not Met Third Sample Volume 0.816L - No Flag	Not A Valid Test - First and Third Samples Not Flagged
Campbell	Steven M.	7/12/1974	DUI	80-001303	Old	1st	Santa Rosa County SO	Santa Rosa County SO	8/6/2006	20:30:09	First Sample Volume 0.289L - Flagged Volume Not Met Second Sample Volume 0.789L - No Flag Third Sample Volume 1.687L - OK	Not A Valid Test - Second Sample Not Flagged
Phillips	Darius K.	4/9/1984	OTHER	80-001309	Old	14th	Bay County SO	Tyndall AFB	4/9/2006	13:22:01	First Sample Volume 0.628L - No Flag Second Sample Volume 0.578L - Flagged Volume Not Met No Third Sample - Subject Refused	Not A Valid Test Refusal
Cardwell	Curtis B.	11/24/1981	DUI	80-001309	Old	14th	Bay County SO	Bay County SO	7/23/2006	23:35:34	First Sample Volume 0.511L - No Flag Second Sample Volume 0.000L - Flagged Volume Not Met Third Sample Volume 0.437L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Miller	Chadrick L.	4/24/1978	DUI	80-001309	Old	14th	Bay County SO	Parker PD	9/16/2006	22:34:35	First Sample Volume 0.437L - No Flag No Second Sample - Subject Refused	Not A Valid Test Refusal
Mings	Lonnie L.	6/29/1977	DUI	80-001309	Old	14th	Bay County SO	Florida Highway Patrol	9/23/2006	3:41:36	First Sample Volume 2.710L - OK Second Sample Volume 0.097L - No Flag Third Sample Volume 1.925L - OK	Valid Test - First and Third Samples
Castillo-Vazquez	Jose	12/6/1960	DUI	80-001309	Old	14th	Bay County SO	Florida Highway Patrol	9/24/2006	1:50:55	First Sample Volume 0.906L - No Flag Second Sample Volume 2.613L - OK	Not A Valid Test - First Sample Not Flagged
Clay	Susan S.	12/9/1953	DUI	80-001310	Old	1st	Niceville PD	Crestview PD	6/12/2006	0:08:55	First Sample Volume 1.246L - OK Second Sample Volume 0.558L - No Flag Third Sample Volume 1.394L - OK	Valid Test - First and Third Samples
Wetherington	Deborah S.	8/10/1964	DUI	80-001310	New	1st	Niceville PD	Niceville PD	10/4/2006	3:39:56	First Sample Volume 0.035L - No Flag No Second Sample Provided No Third Sample Provided	Not A Valid Test - Need At Least Two Samples
Brizendine	William E.	9/4/1963	DUI	80-001313	New	14th	Springfield PD	Panama City PD	9/20/2006	1:55:45	First Sample Volume 1.109L - OK Second Sample Volume 1.011L - No Flag	Not A Valid Test - Second Sample Not Flagged
Diamond	Elfriede F.	11/11/1936	DUI	80-001328	Old	19th	Indian River County SO	Sebastian PD	8/30/2006	22:47:27	First Sample Volume 0.554L - No Flag Second Sample Volume 1.003L - No Flag First Sample Volume 1.972L - OK	Not A Valid Test - First and Second Sample Not Flagged
Angus	Anthony	6/11/1959	DUI	80-001328	Old	19th	Indian River County SO	Florida Highway Patrol	9/11/2006	8:34:49	Second Sample Volume 0.156L - Flagged Volume Not Met Third Sample Volume 0.882L - No Flag	Not A Valid Test - Third Sample Not Flagged
Kellowan	Shevron S.	3/20/1979	DUI	80-001344	Old	12th	Sarasota County SO	Florida Highway Patrol	4/1/2006	4:55:54	First Sample Volume 0.093L - Flagged Volume Not Met Second Sample Volume 3.066L - OK Third Sample Volume 0.714L - No Flag	Not A Valid Test - Third Sample Not Flagged
Ryil	Frank L.	10/27/1951	DUI	80-001344	Old	12th	Sarasota County SO	Sarasota County SO	4/25/2006	22:40:18	First Sample Volume 0.937L - No Flag Second Sample Volume 1.972L - OK	Not A Valid Test - First Sample Not Flagged
Kerr	Julia P.	2/11/1987	DUI	80-001344	Old	12th	Sarasota County SO	Sarasota County SO	5/3/2006	4:22:57	First Sample Volume 0.468L - No Flag Second Sample Volume 1.871L - OK Third Sample Volume 1.691L - OK	Not A Valid Test - No 0 020 Agreement Between Second and Third Samples
Shapiro	Wayne G.	3/25/1959	DUI	80-001344	Old	12th	Sarasota County SO	Sarasota PD	5/28/2006	3:58:01	First Sample Volume 0.656L - No Flag Second Sample Volume 1.683L - OK Third Sample Volume 1.800L - OK	Valid Test - Second and Third Samples
Quattrone	Nancy E.	4/9/1957	DUI	80-001344	Old	12th	Sarasota County SO	Sarasota County SO	6/7/2006	0:27:30	First Sample Volume 0.324L - No Flag Second Sample Volume 1.921L - OK	Not A Valid Test - First Sample Not Flagged
Velle	Victoreno	10/27/1979	DUI	80-001344	Old	12th	Sarasota County SO	Sarasota County SO	6/18/2006	3:26:40	First Sample Volume 1.625L - OK Second Sample Volume 0.636L - No Flag Third Sample Volume 2.160L - OK	Valid Test - First and Third Samples
Palmer	Scott D.	7/13/1964	DUI	80-001344	Old	12th	Sarasota County SO	Sarasota PD	6/26/2006	1:53:50	First Sample Volume 1.101L - OK Second Sample Volume 0.804L - No Flag No Third Sample - Subject Refused	Not A Valid Test Refusal

Delahanty	Thomas L.	5/15/1948	DUI	80-001344	Old	12th	Sarasota County SO	Sarasota County SO	7/8/2006	0:30:50	First Sample Volume 0.492L - No Flag Second Sample Volume 0.191L - Flagged Volume Not Met Third Sample Volume 0.378L - No Flag	Not A Valid Test - First and Third Sample Not Flagged
Nch	Randy R.	6/23/1959	DUI	80-001348	New	12th	Sarasota County SO	North Port PD	10/8/2006	19:25:47	First Sample Volume 0.449L - No Flag Second Sample Volume 2.023L - OK Third Sample Volume 1.519L - OK	Not A Valid Test - No 0 020 Agreement Between Second and Third Samples
Germain	David M.	4/8/1957	DUI	80-001367	Old	6th	Pinellas County SO	Florida Highway Patrol	7/16/2006	18:23:45	First Sample Volume 0.906L - No Flag Second Sample Volume 1.125L - OK Third Sample Volume 1.371L - OK	Valid Test - Second and Third Samples
Desimas	John R.	2/1/1984	DUI	80-001415	Old	6th	Pinellas County SO	Pinellas Park PD	8/5/2006	3:02:07	First Sample Volume 0.130L - No Flag Second Sample Volume 0.593L - Flagged Volume Not Met Third Sample Volume 0.144L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Dilimone	Linda J.	3/13/1962	DUI	80-001416	Old	9th	Orange County SO	Orange County SO	6/6/2006	12:48:14	First Sample Volume 0.835L - No Flag Second Sample Volume 0.437L - Flagged Volume Not Met Third Sample Volume 0.070L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Martinez	Noel I.	9/22/1968	DUI	80-001416	Old	9th	Orange County SO	Orange County SO	6/10/2006	4:17:28	First Sample Volume 0.937L - No Flag Second Sample Volume 1.718L - OK	Not A Valid Test - First Sample Not Flagged
Kars	Chad D.	1/4/1967	DUI	80-001416	Old	9th	Orange County SO	Orlando PD	7/10/2006	1:34:39	First Sample Volume 1.308L - OK Second Sample Volume 0.792L - No Flag	Not A Valid Test - Second Sample Not Flagged
Reyes	Arturo	9/11/1965	DUI	80-001418	Old	9th	Orange County SO	Florida Highway Patrol	8/30/2006	1:34:24	First Sample Volume 0.042L - Flagged Volume Not Met Second Sample Volume 0.804L - No Flag Third Sample Volume 2.421L - OK	Not A Valid Test - Second Sample Not Flagged
Poole	Nikki K.	7/17/1975	DUI	80-001418	Old	9th	Orange County SO	Maitland PD	8/30/2006	3:38:20	First Sample Volume 0.703L - No Flag Second Sample Volume 0.261L - Flagged Volume Not Met Third Sample Volume 1.531L - OK	Not A Valid Test - First Sample Not Flagged
Stephens	Patrick J.	11/20/1984	DUI	80-001420	Old	9th	Orange County SO	Orange County SO	6/11/2006	4:28:54	First Sample Volume 2.972L - OK Second Sample Volume 0.394L - No Flag	Not A Valid Test - Second Sample Not Flagged
Buckley	Linda G.	5/10/1967	DUI	80-001443	Old	4th	Clay County SO	Clay County SO	5/6/2006	1:46:18	First Sample Volume 1.789L - OK Second Sample Volume 1.464L - OK Third Sample Volume 0.648L - No Flag Second Control Test Outside Tolerance	Not A Valid Test - Second Control Test Outside Tolerance
Previllon	Toni L.	4/25/1964	DUI	80-001443	New	4th	Clay County SO	Florida Highway Patrol	9/24/2006	0:06:04	First Sample Volume 0.265L - Flagged Volume Not Met Second Sample Volume 0.953L - No Flag Third Sample Volume 0.707L - Flagged Volume Not Met	Not A Valid Test - Need At Least Two Samples
Nonite	William	4/3/1960	DUI	80-001731	New	13th	Hillsborough County SO	Hillsborough County SO	10/15/2006	20:57:01	First Sample Volume 1.710L - OK Second Sample Volume 1.085L - No Flag Third Sample Volume 1.503L - OK	Valid Test - First and Third Samples

\*\* Notes are concerning the listed breath tests. Additional breath tests of the referenced subject may have been conducted and are not a part of this list.

**FLORIDA DEPARTMENT OF LAW ENFORCEMENT  
ALCOHOL TESTING PROGRAM  
BREATH ALCOHOL TEST AFFIDAVIT**

Instrument Type: Intoxilyzer 8000  
Instrument Registered To: FDLE/OROC/SKIPPER  
Instrument Serial Number: 80-000202 Software: 8100.26  
Date of Test: 09/27/2006

Date of Last Agency Inspection: 08/03/2006  
Observation Period Began: 13:00  
Subject's Name: SAMPLE D LICENSE

DOB: 02/14/1965 Sex: F

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time
	Diagnostics Check	OK	14:24
	Air Blank	0.000	14:24
	Control Test	0.079	14:24
	Air Blank	0.000	14:25
	Subject Sample #1	NSP*	14:28
	Breath Volume	0.000 LITERS	
	Air Blank	0.000	14:29
	Air Blank	0.000	14:31
	Subject Sample #2	0.050	14:34
	Breath Volume	0.750 LITERS	
	Air Blank	0.000	14:34
	Air Blank	0.000	14:36
	Subject Sample #3	0.051	14:39
	Breath Volume	0.578 LITERS	
	Air Blank	0.000	14:40
	Control Test	0.080	14:40
	Air Blank	0.000	14:41
	Diagnostics Check	OK	14:41

*RG/SK  
9-27-06*

\*No Sample Provided

*THIS TEST WAS TO SEE REPORTING WHEN SAMPLE PROVIDED DURING LAST 5 SECONDS OF 3 minute sampling period. AT EXACTLY 3 minutes of sample period instrument stopped analysis and reported sample value and breath volume, even though sample was being introduced through 3 minutes and for 5 seconds afterwards.*

State of Florida, County of \_\_\_\_\_,

Personally appeared before me the undersigned authority, who ( ) is personally known to me or ( ) produced \_\_\_\_\_ as identification, and who after being placed under oath, states:

I ROGER G SKIPPER, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: \_\_\_\_\_ Date: \_\_\_\_\_  
Signature

Sworn to (or affirmed) before me this \_\_\_\_\_ day of \_\_\_\_\_, \_\_\_\_\_

\_\_\_\_\_  
Signature of Notary Public-State of Florida      Printed Name of Notary Public-State of Florida

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.

FLORIDA DEPARTMENT OF LAW ENFORCEMENT  
ALCOHOL TESTING PROGRAM  
BREATH ALCOHOL TEST AFFIDAVIT

Instrument Type: Intoxilyzer 8000  
Instrument Registered To: FDLE/OROC/SKIPPER  
Instrument Serial Number: 80-000202 Software: 8100.26  
Date of Test: 09/27/2006

Date of Last Agency Inspection: 08/03/2006  
Observation Period Began: 13:00  
Subject's Name: SAMPLE D LICENSE

DOB: 02/14/1965 Sex: F

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time
	Diagnostics Check	OK	14:54
	Air Blank	0.000	14:55
	Control Test	0.078	14:55
	Air Blank	0.000	14:56
	Subject Sample #1	0.051	14:59
	Breath Volume	0.589 LITERS	
	Air Blank	0.000	15:00
	Air Blank	0.000	15:01
	Subject Sample #2	VNM*	15:04
	Breath Volume	0.625 LITERS	
	Air Blank	0.000	15:05
	Air Blank	0.000	15:07
	Subject Sample #3	VNM**	15:10
	Breath Volume	0.617 LITERS	
	Air Blank	0.000	15:11
	Control Test	0.077	15:11
	Air Blank	0.000	15:12
	Diagnostics Check	OK	15:12

\*Volume Not Met (0.051 - Breath Sample Not Reliable to Determine Breath Alcohol Level)  
\*\*Volume Not Met (0.048 - Breath Sample Not Reliable to Determine Breath Alcohol Level)

*Rest 9-27-06*

State of Florida, County of \_\_\_\_\_,

Personally appeared before me the undersigned authority, who (\_\_\_) is personally known to me or (\_\_\_) produced \_\_\_\_\_ as identification, and who after being placed under oath, states:

I ROGER G SKIPPER, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: \_\_\_\_\_ Date: \_\_\_\_\_  
Signature

Sworn to (or affirmed) before me this \_\_\_\_\_ day of \_\_\_\_\_, \_\_\_\_\_

Signature of Notary Public-State of Florida \_\_\_\_\_ Printed Name of Notary Public-State of Florida \_\_\_\_\_

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.

**FLORIDA DEPARTMENT OF LAW ENFORCEMENT  
ALCOHOL TESTING PROGRAM  
BREATH ALCOHOL TEST AFFIDAVIT**

Instrument Type: Intoxilyzer 8000  
Instrument Registered To: FDLE/OROC/SKIPPER  
Instrument Serial Number: 80-000202 Software: 8100.26  
Date of Test: 09/27/2006

Date of Last Agency Inspection: 08/03/2006

Observation Period Began: 13:00

Subject's Name: SAMPLE D LICENSE

DOB: 02/14/1965 Sex: F

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:

Test	g/210L	Time
Diagnostics Check	OK	14:11
Air Blank	0.000	14:11
Control Test	0.079	14:12
Air Blank	0.000	14:12
Subject Sample #1	VNM*	14:15
Breath Volume	0.632 LITERS	
Air Blank	0.000	14:16
Air Blank	0.000	14:18
Subject Sample #2	VNM**	14:21
Breath Volume	0.199 LITERS	
Air Blank	0.000	14:22
Control Test	0.074***	14:22
Air Blank	0.000	14:23

*Last Delivery*

*Last Delivery*

*LEFT breath tube  
connected to AS sim  
during purge before*

\*Volume Not Met (0.051 - Breath Sample Not Reliable to Determine Breath Alcohol Level)  
\*\*Volume Not Met (0.043 - Breath Sample Not Reliable to Determine Breath Alcohol Level)  
\*\*\*Control Outside Tolerance

*Fast  
9-27-06*

*THIS TEST WAS TO ESTABLISH PROPER FLAG FOR VNM AND THAT IT REPORTS LAST VOLUME AFTER SEVERAL ATTEMPTS. WORKED PROPERLY.*

State of Florida, County of \_\_\_\_\_,

Personally appeared before me the undersigned authority, who ( ) is personally known to me or ( ) produced \_\_\_\_\_ as identification, and who after being placed under oath, states:

I ROGER G SKIPPER, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: \_\_\_\_\_ Date: \_\_\_\_\_  
Signature

Sworn to (or affirmed) before me this \_\_\_\_\_ day of \_\_\_\_\_, \_\_\_\_\_

Signature of Notary Public-State of Florida \_\_\_\_\_ Printed Name of Notary Public-State of Florida \_\_\_\_\_

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.

FLORIDA DEPARTMENT OF LAW ENFORCEMENT  
ALCOHOL TESTING PROGRAM  
BREATH ALCOHOL TEST AFFIDAVIT

Instrument Type: Intoxilyzer 8000  
Instrument Registered To: FDLE/OROC/SKIPPER  
Instrument Serial Number: 80-000202 Software: 8100.26 (8100.26.01)  
Date of Test: 10/02/2006 BETA

Date of Last Agency Inspection: 10/02/2006  
Observation Period Began: 13:00  
Subject's Name: SAMPLE D LICENSE  
DOB: 02/14/1965 Sex: F

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time
	Diagnostics Check	OK	14:53
	Air Blank	0.000	14:53
	Control Test	0.079	14:54
	Air Blank	0.000	14:54
	Subject Sample #1	0.000	14:55
	Breath Volume	1.273 LITERS	
	Air Blank	0.000	14:55
	Air Blank	0.000	14:57
	Subject Sample #2	VNM*	15:00
	Breath Volume	0.652 LITERS	
	Air Blank	0.000	15:01
	Air Blank	0.000	15:03
	Subject Sample #3	VNM**	15:06
	Breath Volume	0.621 LITERS	
	Air Blank	0.000	15:07
	Control Test	0.079	15:07
	Air Blank	0.000	15:08
	Diagnostics Check	OK	15:08

*RGST*  
*10-02-06*

\*Volume Not Met (0.081 - Breath Sample Not Reliable to Determine Breath Alcohol Level)  
\*\*Volume Not Met (0.081 - Breath Sample Not Reliable to Determine Breath Alcohol Level)

Cylinder Lot: 415601I  
Exp: 06/10/2006

State of Florida, County of \_\_\_\_\_,

Personally appeared before me the undersigned authority, who ( ) is personally known to me or ( ) produced \_\_\_\_\_ as identification, and who after being placed under oath, states:

I ROGER G. SKIPPER, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: \_\_\_\_\_ Date: \_\_\_\_\_  
Signature

Sworn to (or affirmed) before me this \_\_\_\_\_ day of \_\_\_\_\_,

Signature of Notary Public-State of Florida \_\_\_\_\_ Printed Name of Notary Public-State of Florida \_\_\_\_\_

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.

**FLORIDA DEPARTMENT OF LAW ENFORCEMENT  
ALCOHOL TESTING PROGRAM  
BREATH ALCOHOL TEST AFFIDAVIT**

#2

Instrument Type: Intoxilyzer 8000  
Instrument Registered To: FDLE/OROC/SKIPPER  
Instrument Serial Number: 80-000202 Software: 8100.26 (8100.26.01)  
Date of Test: 10/03/2006 BETA

Date of Last Agency Inspection: 10/02/2006  
Observation Period Began: 13:00  
Subject's Name: SAMPLE D LICENSE DOB: 02/14/1965 Sex: F

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time
	Diagnostics Check	OK	23:59
	Air Blank	0.000	23:59
	Control Test	0.079	23:59
	Air Blank	0.000	00:00
	Subject Sample #1	0.081	00:01
	Breath Volume	1.171 LITERS	
	Air Blank	0.000	00:02
	Air Blank	0.000	00:03
	Subject Sample #2	0.081	00:04
	Breath Volume	1.175 LITERS	
	Air Blank	0.000	00:05
	Control Test	0.078	00:05
	Air Blank	0.000	00:05
	Diagnostics Check	OK	00:06

*ICG Skipp  
10-02-06*

Cylinder Lot: 415601I  
Exp: 06/10/2006

State of Florida, County of \_\_\_\_\_,

Personally appeared before me the undersigned authority, who ( ) is personally known to me or ( ) produced \_\_\_\_\_ as identification, and who after being placed under oath, states:

I ROGER G. SKIPPER, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: \_\_\_\_\_ Date: \_\_\_\_\_  
Signature

Sworn to (or affirmed) before me this \_\_\_\_\_ day of \_\_\_\_\_,

\_\_\_\_\_  
Signature of Notary Public-State of Florida      Printed Name of Notary Public-State of Florida

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.

3

FLORIDA DEPARTMENT OF LAW ENFORCEMENT  
ALCOHOL TESTING PROGRAM  
BREATH ALCOHOL TEST AFFIDAVIT

Instrument Type: Intoxilyzer 8000  
Instrument Registered To: FDLE/OROC/SKIPPER  
Instrument Serial Number: 80-000202 Software: 8100.26 (8100.26.01)  
Date of Test: 10/02/2006 Beta

Date of Last Agency Inspection: 10/02/2006  
Observation Period Began: 13:00  
Subject's Name: SAMPLE D LICENSE DOB: 02/14/1965 Sex: F

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time
	Diagnostics Check	OK	15:26
	Air Blank	0.000	15:26
	Control Test	0.078	15:26
	Air Blank	0.000	15:27
	Subject Sample #1	VNM*	15:30
	Breath Volume	0.734 LITERS	
	Air Blank	0.000	15:31
	Air Blank	0.000	15:33
	Subject Sample #2	0.081	15:34
	Breath Volume	1.156 LITERS	
	Air Blank	0.000	15:35
	Air Blank	0.000	15:36
	Subject Sample #3	VNM**	15:40
	Breath Volume	0.507 LITERS	
	Air Blank	0.000	15:40
	Control Test	0.079	15:41
	Air Blank	0.000	15:41
	Diagnostics Check	OK	15:41

*Handwritten signature and date: RGS 10-22-04*

\*Volume Not Met (0.080 - Breath Sample Not Reliable to Determine Breath Alcohol Level)  
\*\*Volume Not Met (0.080 - Breath Sample Not Reliable to Determine Breath Alcohol Level)

Cylinder Lot: 415601I  
Exp: 06/10/2006

State of Florida, County of \_\_\_\_\_,

Personally appeared before me the undersigned authority, who ( ) is personally known to me or ( ) produced \_\_\_\_\_ as identification, and who after being placed under oath, states:

I ROGER G SKIPPER, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: \_\_\_\_\_ Date: \_\_\_\_\_  
Signature

Sworn to (or affirmed) before me this \_\_\_\_\_ day of \_\_\_\_\_,

Signature of Notary Public-State of Florida \_\_\_\_\_ Printed Name of Notary Public-State of Florida \_\_\_\_\_

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.



**FLORIDA DEPARTMENT OF LAW ENFORCEMENT  
ALCOHOL TESTING PROGRAM  
BREATH ALCOHOL TEST AFFIDAVIT**

#24

Instrument Type: Intoxilyzer 8000  
Instrument Registered To: FDLE/OROC/SKIPPER  
Instrument Serial Number: 80-000202 Software: 8100.26 (8100.26.01)  
Date of Test: 10/02/2006 PETA

Date of Last Agency Inspection: 10/02/2006  
Observation Period Began: 13:00  
Subject's Name: SAMPLE D LICENSE DOB: 02/14/1965 Sex: F

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time
	Diagnostics Check	OK	15:44
	Air Blank	0.000	15:44
	Control Test	0.079	15:44
	Air Blank	0.000	15:45
	Subject Sample #1	VNM*	15:48
	Breath Volume	0.878 LITERS	
	Air Blank	0.000	15:49
	Air Blank	0.000	15:51
	Subject Sample #2	0.080	15:54
	Breath Volume	1.109 LITERS	
	Air Blank	0.000	15:55
	Air Blank	0.000	15:56
	Subject Sample #3	0.079	15:57
	Breath Volume	1.179 LITERS	
	Air Blank	0.000	15:57
	Control Test	0.079	15:58
	Air Blank	0.000	15:58
	Diagnostics Check	OK	15:58

*RGST*  
*10-02-04*

\*Volume Not Met (0.080 - Breath Sample Not Reliable to Determine Breath Alcohol Level)

Cylinder Lot: 4156011  
Exp: 06/10/2006

State of Florida, County of \_\_\_\_\_,

Personally appeared before me the undersigned authority, who (\_\_\_) is personally known to me or (\_\_\_) produced \_\_\_\_\_ as identification, and who after being placed under oath, states:

I ROGER G SKIPPER, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: \_\_\_\_\_ Date: \_\_\_\_\_  
Signature

Sworn to (or affirmed) before me this \_\_\_\_\_ day of \_\_\_\_\_,

Signature of Notary Public-State of Florida \_\_\_\_\_ Printed Name of Notary Public-State of Florida \_\_\_\_\_

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.

#5

FLORIDA DEPARTMENT OF LAW ENFORCEMENT  
ALCOHOL TESTING PROGRAM  
BREATH ALCOHOL TEST AFFIDAVIT

Instrument Type: Intoxilyzer 8000  
Instrument Registered To: FDLE/OROC/SKIPPER  
Instrument Serial Number: 80-000202 Software: 8100.26 (8100.26.01)  
Date of Test: 10/02/2006 Bx Ka

Date of Last Agency Inspection: 10/02/2006  
Observation Period Began: 13:00  
Subject's Name: SAMPLE D LICENSE DOB: 02/14/1965 Sex: F

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:

Test	g/210L	Time
Diagnostics Check	OK	16:01
Air Blank	0.000	16:02
Control Test	0.079	16:02
Air Blank	0.000	16:03
Subject Sample #1	VNM*	16:06
Breath Volume	0.476 LITERS	
Air Blank	0.000	16:07
Air Blank	0.000	16:08
Subject Sample #2	VNM**	16:12
Breath Volume	0.750 LITERS	
Air Blank	0.000	16:12
Control Test	0.079	16:12
Air Blank	0.000	16:13
Diagnostics Check	OK	16:13

*[Handwritten Signature]*  
10-02-06

\*Volume Not Met (0.080 - Breath Sample Not Reliable to Determine Breath Alcohol Level)  
\*\*Volume Not Met (0.080 - Breath Sample Not Reliable to Determine Breath Alcohol Level)

Cylinder Lot: 415601I  
Exp: 06/10/2006

State of Florida, County of \_\_\_\_\_,

Personally appeared before me the undersigned authority, who (\_\_\_) is personally known to me or (\_\_\_) produced \_\_\_\_\_ as identification, and who after being placed under oath, states:

I ROGER G SKIPPER, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: \_\_\_\_\_ Date: \_\_\_\_\_  
Signature

Sworn to (or affirmed) before me this \_\_\_\_\_ day of \_\_\_\_\_, \_\_\_\_\_

Signature of Notary Public-State of Florida \_\_\_\_\_ Printed Name of Notary Public-State of Florida \_\_\_\_\_

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.

10-02-04 RSK/AT

#1 3 minute time out test, UNM on 2<sup>nd</sup> + 3<sup>rd</sup> Sample

1<sup>st</sup> Sample

↑ 1.1 Liters introduced with 0.000 Breath Sample @ 0:05 took sample.

2<sup>nd</sup> Sample

↓ 1.1 Liters introduced @ 2:55 minutes. Ended Sample @ 3:00 Requested 3<sup>rd</sup> Sample on Displayed NO 0.020 Agreement. ASKED FOR 3<sup>rd</sup> sample.

3<sup>rd</sup> Sample

↓ 1.1 Liters introduced @ 2:55 minutes. Ended Sample @ 3:00 + Displayed UNM.

Functioned Properly

#2

Time of TEST DATE change from ~~23:58~~ time START TEST Button Pushed to END OF TEST TIME. (TEST STARTED BEFORE MIDNIGHT + ended AFTER MIDNIGHT)

- START TEST Button pushed @ 23:58 - OK, Reported Date of TEST Properly

#3

3 minute time out TEST, UNM 1<sup>st</sup> + 3<sup>rd</sup> Sample

1<sup>st</sup> Sample  
2<sup>nd</sup> Sample

↓ 1.1 Liters introduced @ 2:55. Ended Sample @ 3:00 + Displayed UNM

↑ 1.1 Liters introduced @ 1:00. Took sample + Displayed NO 0.020 Agreement. asked for 3<sup>rd</sup> sample.

↓ 1.1 Liters introduced @ 2:55. Took sample + displayed UNM @ 3:00

Functioned Properly

10-27-06  
R. S. K. W. P. C.

#4 3 minute timeout TEST, 1<sup>st</sup> sample VNM, 2<sup>nd</sup> sample a 2:55 but  
1<sup>st</sup> volume met, 3<sup>rd</sup> sample  $\uparrow$  1.1 liters.

1<sup>st</sup> sample  $\downarrow$  1.1 liters introduced @ 2:55. END OF sample @ 3:00 + displayed VNM

2<sup>nd</sup> sample  $\uparrow$  1.1 liters introduced @ 2:55, +  $\uparrow$  1.1 liters before 3:00. Displayed

NO 0.020 Agreement + added for 3<sup>rd</sup> sample.

3<sup>rd</sup> sample  $\uparrow$  1.1 liters introduced @ 0:05.

FUNCTIONAL PROPERLY

#5 3 minute timeout TEST, 1<sup>st</sup> + 2<sup>nd</sup> sample VNM.

1<sup>st</sup> sample  $\downarrow$  1.1 liters introduced @ 2:55. END OF sample @ 3:00 + displayed VNM

2<sup>nd</sup> sample  $\downarrow$  1.1 liters introduced @ 2:55. END OF sample @ 3:00 + displayed VNM

#1

FLORIDA DEPARTMENT OF LAW ENFORCEMENT  
ALCOHOL TESTING PROGRAM  
BREATH ALCOHOL TEST AFFIDAVIT

Instrument Type: Intoxilyzer 8000  
Instrument Registered To: FDLE OROC  
Instrument Serial Number: 80-000227 Software: 8100.26 (8100.26.01)  
Date of Test: 10/03/2006 BETA

Date of Last Agency Inspection: 10/03/2006  
Observation Period Began: 13:00  
Subject's Name: SAMPLE D LICENSE DOB: 02/14/1965 Sex: F

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time
	Diagnostics Check	OK	13:53
	Air Blank	0.000	13:54
	Control Test	0.078	13:54
	Air Blank	0.000	13:55
	Subject Sample #1	VNM*	13:58
	Breath Volume	1.023 LITERS	
	Air Blank	0.000	13:58
	Air Blank	0.000	14:00
	Subject Sample #2	VNM**	14:04
	Breath Volume	0.539 LITERS	
	Air Blank	0.000	14:04
	Control Test	0.079	14:04
	Air Blank	0.000	14:05
	Diagnostics Check	OK	14:05

10-03-06  
REG

\*Volume Not Met (0.081 - Breath Sample Not Reliable to Determine Breath Alcohol Level)  
\*\*Volume Not Met (0.080 - Breath Sample Not Reliable to Determine Breath Alcohol Level)

Cylinder Lot: 518702I  
Exp: 07/08/2007

State of Florida, County of \_\_\_\_\_,

Personally appeared before me the undersigned authority, who ( ) is personally known to me or ( ) produced \_\_\_\_\_ as identification, and who after being placed under oath, states:

I ROGER G SKIPPER, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: \_\_\_\_\_ Date: \_\_\_\_\_  
Signature

Sworn to (or affirmed) before me this \_\_\_\_\_ day of \_\_\_\_\_, \_\_\_\_\_

Signature of Notary Public-State of Florida \_\_\_\_\_ Printed Name of Notary Public-State of Florida \_\_\_\_\_

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.

**FLORIDA DEPARTMENT OF LAW ENFORCEMENT  
ALCOHOL TESTING PROGRAM  
BREATH ALCOHOL TEST AFFIDAVIT**

Instrument Type: Intoxilyzer 8000  
Instrument Registered To: FDLE OROC  
Instrument Serial Number: 80-000227 Software: 8100.26 (8100.26.01)  
Date of Test: 10/03/2006 Beta

Date of Last Agency Inspection: 10/03/2006  
Observation Period Began: 13:00  
Subject's Name: SAMPLE D LICENSE

DOB: 02/14/1965 Sex: F

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time
	Diagnostics Check	OK	14:07
	Air Blank	0.000	14:07
	Control Test	0.078	14:08
	Air Blank	0.000	14:08
	Subject Sample #1	VNM*	14:11
	Breath Volume	0.968 LITERS	
	Air Blank	0.000	14:12
	Air Blank	0.000	14:14
	Subject Sample #2	0.080	14:14
	Breath Volume	1.222 LITERS	
	Air Blank	0.000	14:15
	Air Blank	0.000	14:17
	Subject Sample #3	VNM**	14:20
	Breath Volume	0.738 LITERS	
	Air Blank	0.000	14:20
	Control Test	0.079	14:21
	Air Blank	0.000	14:21
	Diagnostics Check	OK	14:21

*10-03-06*  
*[Signature]*

\*Volume Not Met (0.079 - Breath Sample Not Reliable to Determine Breath Alcohol Level)  
\*\*Volume Not Met (0.079 - Breath Sample Not Reliable to Determine Breath Alcohol Level)

Cylinder Lot: 518702I  
Exp: 07/08/2007

State of Florida, County of \_\_\_\_\_,

Personally appeared before me the undersigned authority, who ( ) is personally known to me or ( ) produced \_\_\_\_\_ as identification, and who after being placed under oath, states:

I ROGER G. SKIPPER, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: \_\_\_\_\_ Date: \_\_\_\_\_  
Signature

Sworn to (or affirmed) before me this \_\_\_\_\_ day of \_\_\_\_\_, \_\_\_\_\_

Signature of Notary Public-State of Florida \_\_\_\_\_ Printed Name of Notary Public-State of Florida \_\_\_\_\_

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.

3  
X  
FLORIDA DEPARTMENT OF LAW ENFORCEMENT  
ALCOHOL TESTING PROGRAM  
BREATH ALCOHOL TEST AFFIDAVIT

Instrument Type: Intoxilyzer 8000  
Instrument Registered To: FDLE OROC  
Instrument Serial Number: 80-000227 Software: 8100.26 (8100.26.01)  
Date of Test: 10/03/2006 1387a

Date of Last Agency Inspection: 10/03/2006  
Observation Period Began: 13:00  
Subject's Name: SAMPLE D LICENSE DOB: 02/14/1965 Sex: F

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:

Test	g/210L	Time
Diagnostics Check	OK	14:40
Air Blank	0.000	14:40
Control Test	0.079	14:40
Air Blank	0.000	14:41
Subject Sample #1	0.080	14:44
Breath Volume	1.214 LITERS	
Air Blank	0.000	14:44
Air Blank	0.000	14:46
Subject Sample #2	VNM*	14:49
Breath Volume	0.917 LITERS	
Air Blank	0.000	14:50
Air Blank	0.000	14:52
Subject Sample #3	VNM**	14:55
Breath Volume	0.328 LITERS	
Air Blank	0.000	14:56
Control Test	0.078	14:56
Air Blank	0.000	14:56
Diagnostics Check	OK	14:56

10-3-06  
R. G. SKIPPER

\*Volume Not Met (0.079 - Breath Sample Not Reliable to Determine Breath Alcohol Level)  
\*\*Volume Not Met (0.076 - Breath Sample Not Reliable to Determine Breath Alcohol Level)

Cylinder Lot: 5187021  
Exp: 07/08/2007

State of Florida, County of \_\_\_\_\_,

Personally appeared before me the undersigned authority, who ( ) is personally known to me or ( ) produced \_\_\_\_\_ as identification, and who after being placed under oath, states:

I ROGER G SKIPPER, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: \_\_\_\_\_ Date: \_\_\_\_\_  
Signature

Sworn to (or affirmed) before me this \_\_\_\_\_ day of \_\_\_\_\_,

Signature of Notary Public-State of Florida Printed Name of Notary Public-State of Florida

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.

#4  
FLORIDA DEPARTMENT OF LAW ENFORCEMENT  
ALCOHOL TESTING PROGRAM  
BREATH ALCOHOL TEST AFFIDAVIT

Instrument Type: Intoxilyzer 8000  
Instrument Registered To: FDLE OROC  
Instrument Serial Number: 80-000227 Software: 8100.26 (8100.26.01)  
Date of Test: 10/03/2006 BETA

Date of Last Agency Inspection: 10/03/2006

Observation Period Began: 13:00

Subject's Name: SAMPLE D LICENSE

DOB: 02/14/1965 Sex: F

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:

Test	g/210L	Time
Diagnostics Check	OK	14:59
Air Blank	0.000	14:59
Control Test	0.078	15:00
Air Blank	0.000	15:00
Subject Sample #1	0.079	15:00
Breath Volume	1.171 LITERS	
Air Blank	0.000	15:01
Air Blank	0.000	15:03
Subject Sample #2	0.079	15:04
Breath Volume	1.167 LITERS	
Air Blank	0.000	15:04
Control Test	0.079	15:04
Air Blank	0.000	15:05
Diagnostics Check	OK	15:05

10-3-06  
FCS

Cylinder Lot: 518702I  
Exp: 07/08/2007

State of Florida, County of \_\_\_\_\_,

Personally appeared before me the undersigned authority, who (\_\_\_) is personally known to me or (\_\_\_) produced \_\_\_\_\_ as identification, and who after being placed under oath, states:

I ROGER G SKIPPER, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: \_\_\_\_\_ Date: \_\_\_\_\_  
Signature

Sworn to (or affirmed) before me this \_\_\_\_\_ day of \_\_\_\_\_, \_\_\_\_\_

Signature of Notary Public-State of Florida \_\_\_\_\_ Printed Name of Notary Public-State of Florida \_\_\_\_\_

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.



#5  
FLORIDA DEPARTMENT OF LAW ENFORCEMENT  
ALCOHOL TESTING PROGRAM  
BREATH ALCOHOL TEST AFFIDAVIT

Instrument Type: Intoxilyzer 8000  
Instrument Registered To: FDLE OROC  
Instrument Serial Number: 80-000227 Software: 8100.26 (8100.26.01)  
Date of Test: 10/04/2006 *BETA*

Date of Last Agency Inspection: 10/03/2006  
Observation Period Began: 13:00  
Subject's Name: SAMPLE D LICENSE

DOB: 02/14/1965 Sex: F

The subject was observed for at least twenty-minutes prior to the administration of the breath test to ensure that the subject did not take anything orally and did not regurgitate.

Results:	Test	g/210L	Time
	Diagnostics Check	OK	23:59
	Air Blank	0.000	23:59
	Control Test	0.079	23:59
	Air Blank	0.000	00:00
	Subject Sample #1	0.080	00:00
	Breath Volume	1.183 LITERS	
	Air Blank	0.000	00:01
	Air Blank	0.000	00:03
	Subject Sample #2	0.079	00:03
	Breath Volume	1.171 LITERS	
	Air Blank	0.000	00:04
	Control Test	0.078	00:04
	Air Blank	0.000	00:05
	Diagnostics Check	OK	00:05

*10-03-06*  
*RGS*

Cylinder Lot: 518702I  
Exp: 07/08/2007

State of Florida, County of \_\_\_\_\_,

Personally appeared before me the undersigned authority, who (\_\_\_) is personally known to me or (\_\_\_) produced \_\_\_\_\_ as identification, and who after being placed under oath, states:

I ROGER G SKIPPER, hold a valid Breath Test Operator permit issued by the Florida Department of Law Enforcement, I administered the above breath test to the subject named above in accordance with Chapter 11D-8, Florida Administrative Code, and this form is a true and accurate report of that breath test.

Breath Test Operator: \_\_\_\_\_ Date: \_\_\_\_\_  
Signature

Sworn to (or affirmed) before me this \_\_\_\_\_ day of \_\_\_\_\_, \_\_\_\_\_

Signature of Notary Public-State of Florida \_\_\_\_\_ Printed Name of Notary Public-State of Florida \_\_\_\_\_

Note: Pursuant to section 117.10, Florida Statutes, law enforcement officers, correctional officers, traffic accident investigation officers and traffic infraction enforcement officers are notaries public when engaged in the performance of official duties. In accordance with section 316.1934(5), F.S., this completed form is admissible without further authentication and is presumptive proof of the results herein. To be used in accordance with Section 316.1934(5), F.S., and in administrative proceedings pursuant to 322.2615, F.S.

3 minute TIMEOUT TESTS,

#1

sample #1 VNM ON 1<sup>st</sup> + 2<sup>nd</sup> Samples  
 ↓ 1.1 liters introduced @ 2:55. END OF sample @ 3:00 + Displayed VNM  
 sample #2 ↓ 1.1 liters introduced @ 2:55. END OF sample @ 3:00 + Displayed VNM  
 OK

#2

sample #1 VNM ON 1<sup>st</sup> + 3<sup>rd</sup> Samples  
 ↓ 1.1 liters introduced @ 2:55. END OF sample @ 3:00 + Displayed VNM  
 sample #2 ↑ 1.1 liters introduced @ 0:05. NO 0.020 AGREEMENT.  
 sample #3 ↓ 1.1 liters introduced @ 2:55. END OF sample @ 3:00 + Displayed VNM  
 OK

#3

sample #1 VNM 2<sup>nd</sup> + 3<sup>rd</sup> Samples with 3<sup>rd</sup> sample VNM @ 2:00  
 ↑ 1.1 liters introduced @ 0:05. OK  
 sample #2 ↓ 1.1 liters introduced @ 2:55. Displayed VNM + NO 0.020 AGREEMENT  
 sample #3 ↓ 1.1 liters introduced @ 2:00. Timed out @ 3:00, Displayed VNM.  
 OK

#4

Base volume met TEST  
 sample #1 ↑ 1.1 liters  
 sample #2 ↑ 1.1 liters OK

#5

TIME OF TEST DATE CHANGE FROM PUSH START TEST BUTTON TO END OF TEST TIME. (TEST STARTED) BEFORE MIDNIGHT + ended AFTER MIDNIGHT.