

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

Agency: ESCAMBIA COUNTY SO

Date/Time: 10/31/2016 06:07:58

Data Download Date/Time: 10/31/2016 07:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: ESCAMBIA COUNTY SO

Date/Time: 12/05/2016 04:45:20

Data Download Date/Time: 12/05/2016 06:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000758

Agency: FPSI TRAINING

Date/Time: 06/07/2016 09:19:59

Data Download Date/Time: 10/13/2016 13:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

Agency: FL HIGHWAY PATROL

Date/Time: 01/04/2016 09:19:51

Data Download Date/Time: 01/28/2016 15:36

Voltage/Current Test	Fail
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: FL HIGHWAY PATROL

Date/Time: 01/04/2016 09:21:44

Data Download Date/Time: 01/28/2016 15:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

Agency: FL HIGHWAY PATROL

Date/Time: 01/04/2016 09:22:55

Data Download Date/Time: 01/28/2016 15:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: FL HIGHWAY PATROL

Date/Time: 05/16/2016 15:11:31

Data Download Date/Time: 05/16/2016 16:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

Agency: FL HIGHWAY PATROL

Date/Time: 07/25/2016 10:05:25

Data Download Date/Time: 08/14/2016 15:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: FL HIGHWAY PATROL

Date/Time: 07/25/2016 10:25:33

Data Download Date/Time: 08/14/2016 15:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000774

Agency: MIAMI PD TRAINING

Date/Time: 12/30/2016 15:04:50

Data Download Date/Time: 12/30/2016 15:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

Agency: ST LUCIE COUNTY SO

Date/Time: 09/01/2016 09:38:45

Data Download Date/Time: 09/27/2016 14:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

Agency: HERNANDO COUNTY S.O.

Date/Time: 01/07/2016 11:19:38

Data Download Date/Time: 01/25/2016 15:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: HERNANDO COUNTY S.O.

Date/Time: 04/20/2016 14:56:55

Data Download Date/Time: 04/20/2016 17:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

Agency: HERNANDO COUNTY S.O.

Date/Time: 06/22/2016 13:34:03

Data Download Date/Time: 06/23/2016 10:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: HERNANDO COUNTY S.O.

Date/Time: 10/21/2016 10:46:43

Data Download Date/Time: 11/23/2016 16:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

Agency: HERNANDO COUNTY S.O.

Date/Time: 10/27/2016 14:22:38

Data Download Date/Time: 11/23/2016 16:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: HERNANDO COUNTY S.O.

Date/Time: 10/27/2016 15:04:56

Data Download Date/Time: 11/23/2016 16:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

Agency: HERNANDO COUNTY S.O.

Date/Time: 10/28/2016 13:03:05

Data Download Date/Time: 11/23/2016 16:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: HERNANDO COUNTY S.O.

Date/Time: 10/31/2016 12:38:14

Data Download Date/Time: 11/23/2016 16:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000812

Agency: HERNANDO COUNTY S.O.

Date/Time: 10/31/2016 15:56:32

Data Download Date/Time: 11/23/2016 16:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

Agency: HERNANDO COUNTY S.O.

Date/Time: 12/27/2016 22:16:37

Data Download Date/Time: 01/19/2017 08:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

Agency: SUMTER COUNTY SO

Date/Time: 05/22/2016 04:59:59

Data Download Date/Time: 05/30/2016 12:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

Agency: CITRUS COUNTY SO

Date/Time: 05/23/2016 08:10:37

Data Download Date/Time: 06/29/2016 11:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: CITRUS COUNTY SO

Date/Time: 05/23/2016 08:28:11

Data Download Date/Time: 06/29/2016 11:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

Agency: MARION COUNTY SO

Date/Time: 05/25/2016 20:19:37

Data Download Date/Time: 05/27/2016 12:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

Agency: HILLSBOROUGH CO SO

Date/Time: 09/34/2016 22:54:02

Data Download Date/Time: 10/05/2016 17:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	Fail
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

Agency: LYNN HAVEN PD

Date/Time: 06/25/2016 09:20:57

Data Download Date/Time: 06/25/2016 10:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

Agency: CAPE CORAL PD

Date/Time: 03/09/2016 08:34:24

Data Download Date/Time: 03/09/2016 09:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency: MIAMI PD

Date/Time: 11/01/2016 14:48:56

Data Download Date/Time: 11/29/2016 16:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

Agency: MIAMI PD

Date/Time: 05/25/2016 17:14:25

Data Download Date/Time: 05/25/2016 19:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

Agency: MIAMI PD

Date/Time: 11/29/2016 15:39:41

Data Download Date/Time: 11/29/2016 16:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

Agency: CALHOUN COUNTY SO

Date/Time: 02/03/2016 14:55:16

Data Download Date/Time: 02/18/2016 09:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CALHOUN COUNTY SO

Date/Time: 02/08/2016 15:13:45

Data Download Date/Time: 02/18/2016 09:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

Agency: CALHOUN COUNTY SO

Date/Time: 02/08/2016 15:14:15

Data Download Date/Time: 02/18/2016 09:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: CALHOUN COUNTY SO

Date/Time: 02/08/2016 15:14:50

Data Download Date/Time: 02/18/2016 09:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency: HIALEAH PD

Date/Time: 02/11/2016 15:18:24

Data Download Date/Time: 03/31/2016 10:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: HIALEAH PD

Date/Time: 04/19/2016 13:50:07

Data Download Date/Time: 05/18/2016 11:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency: HIALEAH PD

Date/Time: 04/19/2016 13:52:19

Data Download Date/Time: 05/18/2016 11:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: HIALEAH PD

Date/Time: 04/19/2016 13:53:49

Data Download Date/Time: 05/18/2016 11:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency: HIALEAH PD

Date/Time: 04/19/2016 13:55:03

Data Download Date/Time: 05/18/2016 11:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: HIALEAH PD

Date/Time: 04/19/2016 13:56:32

Data Download Date/Time: 05/18/2016 11:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency: HIALEAH PD

Date/Time: 04/20/2016 09:26:59

Data Download Date/Time: 05/18/2016 11:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: HIALEAH PD

Date/Time: 05/18/2016 13:05:35

Data Download Date/Time: 06/22/2016 12:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

Agency: HIALEAH PD

Date/Time: 07/14/2016 14:00:31

Data Download Date/Time: 08/22/2016 12:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

Agency: MIAMI-DADE PD

Date/Time: 11/10/2016 13:05:37

Data Download Date/Time: 11/14/2016 12:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

Agency: NORTH MIAMI BEACH PD

Date/Time: 12/01/2016 08:31:12

Data Download Date/Time: 12/01/2016 09:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

Agency: PINELLAS COUNTY SO

Date/Time: 04/29/2016 13:33:06

Data Download Date/Time: 05/03/2016 11:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PINELLAS COUNTY SO

Date/Time: 05/02/2016 15:32:06

Data Download Date/Time: 05/03/2016 11:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

Agency: PINELLAS COUNTY SO

Date/Time: 04/12/2016 08:41:29

Data Download Date/Time: 04/12/2016 15:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PINELLAS COUNTY SO

Date/Time: 06/21/2016 15:39:21

Data Download Date/Time: 06/22/2016 16:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

Agency: PINELLAS COUNTY SO

Date/Time: 06/21/2016 16:35:55

Data Download Date/Time: 06/22/2016 16:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

Agency: FWC

Date/Time: 11/15/2016 11:17:37

Data Download Date/Time: 11/15/2016 16:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

Agency: FFWCC TAMPA

Date/Time: 02/25/2016 17:59:02

Data Download Date/Time: 02/25/2016 19:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

Agency: FISH & WILDLIFE CC

Date/Time: 05/10/2016 11:44:02

Data Download Date/Time: 05/10/2016 12:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

Agency: PINECREST PD

Date/Time: 02/25/2016 07:33:10

Data Download Date/Time: 02/25/2016 15:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PINECREST PD

Date/Time: 08/18/2016 14:39:50

Data Download Date/Time: 08/23/2016 13:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

Agency: SANIBEL ISLAND PD

Date/Time: 09/02/2016 23:16:19

Data Download Date/Time: 10/10/2016 17:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

Agency: LEE COUNTY SO

Date/Time: 08/17/2016 15:36:23

Data Download Date/Time: 08/23/2016 12:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

Agency: LEE COUNTY SO

Date/Time: 08/08/2016 07:47:43

Data Download Date/Time: 08/12/2016 14:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

Agency: COLLIER COUNTY SO

Date/Time: 11/14/2016 15:34:01

Data Download Date/Time: 11/15/2016 16:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

Agency: COLLIER COUNTY SO

Date/Time: 06/22/2016 15:47:32

Data Download Date/Time: 06/23/2016 16:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

Agency: GLADES COUNTY S.O.

Date/Time: 03/29/2016 07:13:35

Data Download Date/Time: 03/29/2016 08:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

Agency: CLEWISTON P.D.

Date/Time: 10/25/2016 14:20:36

Data Download Date/Time: 11/21/2016 16:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CLEWISTON P.D.

Date/Time: 10/25/2016 14:21:49

Data Download Date/Time: 11/21/2016 16:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

Agency: APOPKA P.D.

Date/Time: 07/15/2016 21:05:47

Data Download Date/Time: 07/15/2016 22:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: APOPKA P.D.

Date/Time: 11/25/2016 23:39:23

Data Download Date/Time: 11/28/2016 22:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000994

Agency: WINTER PARK P.D.

Date/Time: 02/01/2016 12:42:38

Data Download Date/Time: 02/01/2016 13:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: WINTER PARK P.D.

Date/Time: 02/26/2016 14:29:15

Data Download Date/Time: 02/26/2016 15:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000994

Agency: WINTER PARK P.D.

Date/Time: 03/24/2016 14:30:35

Data Download Date/Time: 03/24/2016 15:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: WINTER PARK P.D.

Date/Time: 04/19/2016 14:45:02

Data Download Date/Time: 04/19/2016 15:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000994

Agency: WINTER PARK P.D.

Date/Time: 05/20/2016 09:11:11

Data Download Date/Time: 05/20/2016 10:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: WINTER PARK P.D.

Date/Time: 06/13/2016 13:13:50

Data Download Date/Time: 06/13/2016 14:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000994

Agency: WINTER PARK P.D.

Date/Time: 07/19/2016 13:51:33

Data Download Date/Time: 07/19/2016 14:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: WINTER PARK P.D.

Date/Time: 08/17/2016 13:54:26

Data Download Date/Time: 08/17/2016 14:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000994

Agency: WINTER PARK P.D.

Date/Time: 09/16/2016 18:23:06

Data Download Date/Time: 09/16/2016 19:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: WINTER PARK P.D.

Date/Time: 10/04/2016 10:00:58

Data Download Date/Time: 10/04/2016 11:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000994

Agency: WINTER PARK P.D.

Date/Time: 11/08/2016 13:06:56

Data Download Date/Time: 11/08/2016 14:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: WINTER PARK P.D.

Date/Time: 12/15/2016 09:47:57

Data Download Date/Time: 12/15/2016 13:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000997

Agency: ALTAMONTE SPRINGS PD

Date/Time: 01/06/2016 03:46:16

Data Download Date/Time: 01/06/2016 04:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: ALTAMONTE SPRINGS PD

Date/Time: 04/04/2016 05:09:57

Data Download Date/Time: 04/06/2016 13:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000997

Agency: ALTAMONTE SPRINGS PD

Date/Time: 05/16/2016 01:19:16

Data Download Date/Time: 05/16/2016 02:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

Agency: HARDEE COUNTY SO

Date/Time: 08/29/2016 13:33:45

Data Download Date/Time: 08/29/2016 14:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

Agency: PINELLAS COUNTY SO

Date/Time: 11/03/2016 23:16:22

Data Download Date/Time: 11/17/2016 23:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001004

Agency: BREVARD COUNTY S.O.

Date/Time: 09/30/2016 15:14:03

Data Download Date/Time: 10/03/2016 16:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

Agency: PINELLAS PARK PD

Date/Time: 04/12/2016 15:18:23

Data Download Date/Time: 04/13/2016 10:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PINELLAS PARK PD

Date/Time: 04/12/2016 15:26:08

Data Download Date/Time: 04/13/2016 10:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/15/2016 13:25:50

Data Download Date/Time: 01/15/2016 13:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 01/21/2016 19:54:45

Data Download Date/Time: 01/21/2016 21:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CLAY COUNTY SO

Date/Time: 02/07/2016 15:10:38

Data Download Date/Time: 02/21/2016 18:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 02/07/2016 15:23:37

Data Download Date/Time: 02/21/2016 18:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CLAY COUNTY SO

Date/Time: 03/21/2016 15:12:07

Data Download Date/Time: 03/21/2016 16:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 04/27/2016 13:15:27

Data Download Date/Time: 04/27/2016 14:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CLAY COUNTY SO

Date/Time: 05/21/2016 13:59:25

Data Download Date/Time: 05/21/2016 14:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 07/03/2016 13:11:19

Data Download Date/Time: 07/03/2016 13:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CLAY COUNTY SO

Date/Time: 09/25/2016 18:03:23

Data Download Date/Time: 09/25/2016 18:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 10/09/2016 09:00:21

Data Download Date/Time: 10/09/2016 09:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CLAY COUNTY SO

Date/Time: 11/06/2016 11:26:40

Data Download Date/Time: 11/06/2016 12:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

Agency: PASCO COUNTY SO

Date/Time: 02/25/2016 19:11:38

Data Download Date/Time: 02/26/2016 13:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 03/16/2016 09:14:54

Data Download Date/Time: 03/16/2016 10:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

Agency: PASCO COUNTY SO

Date/Time: 04/08/2016 11:14:58

Data Download Date/Time: 04/08/2016 12:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 05/05/2016 08:35:00

Data Download Date/Time: 05/05/2016 09:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

Agency: PASCO COUNTY SO

Date/Time: 06/03/2016 08:46:29

Data Download Date/Time: 06/03/2016 09:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 07/05/2016 00:22:22

Data Download Date/Time: 07/07/2016 10:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

Agency: PASCO COUNTY SO

Date/Time: 08/19/2016 10:06:35

Data Download Date/Time: 08/19/2016 10:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 09/15/2016 09:33:00

Data Download Date/Time: 09/15/2016 10:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

Agency: PASCO COUNTY SO

Date/Time: 10/14/2016 07:46:08

Data Download Date/Time: 10/14/2016 08:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 11/03/2016 08:20:45

Data Download Date/Time: 11/03/2016 09:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

Agency: PASCO COUNTY SO

Date/Time: 12/01/2016 11:03:04

Data Download Date/Time: 12/01/2016 11:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency: NASSAU CO SO

Date/Time: 06/27/2016 04:26:35

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: NASSAU CO SO

Date/Time: 06/27/2016 05:30:01

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency: NASSAU CO SO

Date/Time: 06/27/2016 05:31:55

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: NASSAU CO SO

Date/Time: 06/27/2016 05:33:42

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	Fail
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency: NASSAU CO SO

Date/Time: 06/27/2016 05:38:15

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: NASSAU CO SO

Date/Time: 06/27/2016 05:45:15

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency: NASSAU CO SO

Date/Time: 06/27/2016 05:45:53

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: NASSAU CO SO

Date/Time: 06/29/2016 13:13:14

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency: NASSAU CO SO

Date/Time: 06/29/2016 13:14:12

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: NASSAU CO SO

Date/Time: 06/29/2016 13:16:20

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency: NASSAU CO SO

Date/Time: 06/29/2016 13:17:21

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: NASSAU CO SO

Date/Time: 06/29/2016 13:19:03

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency: NASSAU CO SO

Date/Time: 06/29/2016 13:19:40

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	Fail
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: NASSAU CO SO

Date/Time: 06/29/2016 13:21:59

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency: NASSAU CO SO

Date/Time: 06/30/2016 08:43:47

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: NASSAU CO SO

Date/Time: 07/18/2016 10:32:41

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	Fail
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency: NASSAU CO SO

Date/Time: 07/20/2016 12:43:48

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: NASSAU CO SO

Date/Time: 07/20/2016 12:44:27

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	Fail
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency: NASSAU CO SO

Date/Time: 07/20/2016 12:45:11

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: NASSAU CO SO

Date/Time: 07/20/2016 12:45:51

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	Fail
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

Agency: NASSAU CO SO

Date/Time: 07/20/2016 12:46:26

Data Download Date/Time: 08/12/2016 14:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/15/2016 13:29:16

Data Download Date/Time: 01/15/2016 14:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: HIGHLANDS COUNTY SO

Date/Time: 06/17/2016 13:41:00

Data Download Date/Time: 06/17/2016 14:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

Agency: POLK COUNTY SO

Date/Time: 11/10/2016 08:35:27

Data Download Date/Time: 11/10/2016 15:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

Agency: DAVIE PD

Date/Time: 03/08/2016 15:57:35

Data Download Date/Time: 05/04/2016 14:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

Agency: WINTER HAVEN PD

Date/Time: 02/21/2016 03:03:29

Data Download Date/Time: 02/26/2016 00:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

Agency: TREASURE ISLAND PD

Date/Time: 03/28/2016 10:12:09

Data Download Date/Time: 04/14/2016 14:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

Agency: NEW PORT RICHEY PD

Date/Time: 10/05/2016 20:33:04

Data Download Date/Time: 10/06/2016 19:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: NEW PORT RICHEY PD

Date/Time: 10/21/2016 10:27:39

Data Download Date/Time: 11/23/2016 16:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

Agency: NEW PORT RICHEY PD

Date/Time: 10/27/2016 14:21:25

Data Download Date/Time: 11/23/2016 16:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: NEW PORT RICHEY PD

Date/Time: 10/27/2016 15:03:33

Data Download Date/Time: 11/23/2016 16:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

Agency: NEW PORT RICHEY PD

Date/Time: 10/28/2016 13:01:41

Data Download Date/Time: 11/23/2016 16:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: NEW PORT RICHEY PD

Date/Time: 10/31/2016 12:37:07

Data Download Date/Time: 11/23/2016 16:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

Agency: NEW PORT RICHEY PD

Date/Time: 10/31/2016 15:55:16

Data Download Date/Time: 11/23/2016 16:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

Agency: TARPON SPRINGS PD

Date/Time: 05/09/2016 11:26:45

Data Download Date/Time: 05/10/2016 10:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: TARPON SPRINGS PD

Date/Time: 07/08/2016 10:57:58

Data Download Date/Time: 07/21/2016 14:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

Agency: DUNNELLON PD

Date/Time: 02/01/2016 14:34:35

Data Download Date/Time: 02/01/2016 15:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

Agency: CLAY COUNTY SO

Date/Time: 01/21/2016 19:50:16

Data Download Date/Time: 01/21/2016 20:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CLAY COUNTY SO

Date/Time: 03/21/2016 14:46:56

Data Download Date/Time: 03/21/2016 15:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

Agency: CLAY COUNTY SO

Date/Time: 03/21/2016 15:06:10

Data Download Date/Time: 03/21/2016 15:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CLAY COUNTY SO

Date/Time: 10/09/2016 09:46:19

Data Download Date/Time: 10/09/2016 10:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency: PASCO COUNTY SO

Date/Time: 02/25/2016 19:10:03

Data Download Date/Time: 02/26/2016 13:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 03/16/2016 09:14:13

Data Download Date/Time: 03/16/2016 10:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency: PASCO COUNTY SO

Date/Time: 04/08/2016 11:14:30

Data Download Date/Time: 04/08/2016 11:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 05/05/2016 08:34:57

Data Download Date/Time: 05/05/2016 09:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency: PASCO COUNTY SO

Date/Time: 06/03/2016 08:45:57

Data Download Date/Time: 06/03/2016 09:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 07/07/2016 09:01:31

Data Download Date/Time: 07/07/2016 09:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency: PASCO COUNTY SO

Date/Time: 08/19/2016 10:07:25

Data Download Date/Time: 08/19/2016 11:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 09/15/2016 09:32:09

Data Download Date/Time: 09/15/2016 10:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency: PASCO COUNTY SO

Date/Time: 10/14/2016 07:48:24

Data Download Date/Time: 10/14/2016 08:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 11/03/2016 08:25:06

Data Download Date/Time: 11/03/2016 09:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency: PASCO COUNTY SO

Date/Time: 12/01/2016 11:05:48

Data Download Date/Time: 12/01/2016 11:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency: ZEPHYRHILLS PD

Date/Time: 01/28/2016 18:32:08

Data Download Date/Time: 01/28/2016 19:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: ZEPHYRHILLS PD

Date/Time: 03/26/2016 21:23:19

Data Download Date/Time: 03/27/2016 00:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

Agency: ZEPHYRHILLS PD

Date/Time: 03/26/2016 21:23:56

Data Download Date/Time: 03/27/2016 00:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: ZEPHYRHILLS PD

Date/Time: 03/26/2016 21:31:15

Data Download Date/Time: 03/27/2016 00:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency: LARGO PD

Date/Time: 01/20/2016 10:55:10

Data Download Date/Time: 01/20/2016 12:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: LARGO PD

Date/Time: 01/20/2016 12:36:51

Data Download Date/Time: 01/20/2016 13:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency: LARGO PD

Date/Time: 02/22/2016 20:52:49

Data Download Date/Time: 02/22/2016 21:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: LARGO PD

Date/Time: 04/14/2016 10:29:04

Data Download Date/Time: 04/14/2016 11:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency: LARGO PD

Date/Time: 05/18/2016 08:13:47

Data Download Date/Time: 05/18/2016 08:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: LARGO PD

Date/Time: 05/26/2016 06:58:32

Data Download Date/Time: 05/26/2016 07:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency: LARGO PD

Date/Time: 09/19/2016 04:50:54

Data Download Date/Time: 09/19/2016 06:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: LARGO PD

Date/Time: 10/17/2016 16:38:42

Data Download Date/Time: 11/15/2016 14:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

Agency: PASCO COUNTY SO

Date/Time: 02/25/2016 19:12:57

Data Download Date/Time: 02/26/2016 13:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 03/16/2016 09:14:16

Data Download Date/Time: 03/16/2016 09:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

Agency: PASCO COUNTY SO

Date/Time: 04/08/2016 11:14:34

Data Download Date/Time: 04/08/2016 11:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 05/05/2016 08:35:39

Data Download Date/Time: 05/05/2016 09:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

Agency: PASCO COUNTY SO

Date/Time: 06/03/2016 08:46:12

Data Download Date/Time: 06/03/2016 09:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 07/07/2016 09:01:33

Data Download Date/Time: 07/07/2016 09:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

Agency: PASCO COUNTY SO

Date/Time: 08/19/2016 10:05:55

Data Download Date/Time: 08/19/2016 10:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 09/15/2016 09:31:27

Data Download Date/Time: 09/15/2016 10:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

Agency: PASCO COUNTY SO

Date/Time: 10/14/2016 07:46:05

Data Download Date/Time: 10/14/2016 08:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 11/03/2016 08:20:54

Data Download Date/Time: 11/03/2016 09:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

Agency: PASCO COUNTY SO

Date/Time: 12/01/2016 11:03:43

Data Download Date/Time: 12/01/2016 11:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

Agency: FL HIGHWAY PATROL

Date/Time: 01/15/2016 20:14:40

Data Download Date/Time: 01/15/2016 22:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: FL HIGHWAY PATROL

Date/Time: 01/15/2016 20:33:56

Data Download Date/Time: 01/15/2016 22:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

Agency: FL HIGHWAY PATROL

Date/Time: 01/15/2016 20:15:04

Data Download Date/Time: 03/09/2016 09:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: FL HIGHWAY PATROL

Date/Time: 01/15/2016 20:34:11

Data Download Date/Time: 03/09/2016 09:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001117

Agency: PASCO COUNTY SO

Date/Time: 02/25/2016 19:13:58

Data Download Date/Time: 02/26/2016 13:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency: FL HIGHWAY PATROL

Date/Time: 02/01/2016 14:33:32

Data Download Date/Time: 02/01/2016 15:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: FL HIGHWAY PATROL

Date/Time: 12/06/2016 09:04:37

Data Download Date/Time: 01/10/2017 12:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency: FL HIGHWAY PATROL

Date/Time: 12/06/2016 09:10:36

Data Download Date/Time: 01/10/2017 12:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

Agency: FL HIGHWAY PATROL

Date/Time: 12/29/2016 13:46:58

Data Download Date/Time: 12/29/2016 15:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001120

Agency: FHP TROOP K

Date/Time: 04/19/2016 07:56:41

Data Download Date/Time: 04/21/2016 09:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency: FHP TROOP D ORLANDO

Date/Time: 02/25/2016 16:29:41

Data Download Date/Time: 02/26/2016 13:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 07/20/2016 09:24:14

Data Download Date/Time: 07/20/2016 14:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

Agency: DAYTONA BEACH P.D.

Date/Time: 07/25/2016 11:07:19

Data Download Date/Time: 08/04/2016 10:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

Agency: VOLUSIA COUNTY SO

Date/Time: 06/29/2016 11:39:41

Data Download Date/Time: 07/11/2016 07:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

Agency: BOCA RATON PD

Date/Time: 10/25/2016 10:43:28

Data Download Date/Time: 10/25/2016 11:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

Agency: COLLIER COUNTY SO

Date/Time: 09/30/2016 13:50:59

Data Download Date/Time: 10/03/2016 11:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

Agency: COLLIER COUNTY SO

Date/Time: 07/22/2016 13:59:34

Data Download Date/Time: 07/25/2016 16:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

Agency: PANAMA CITY PD

Date/Time: 03/22/2016 01:46:47

Data Download Date/Time: 04/28/2016 16:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PANAMA CITY PD

Date/Time: 03/23/2016 21:44:20

Data Download Date/Time: 04/28/2016 16:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

Agency: LEE COUNTY SO

Date/Time: 08/08/2016 08:02:44

Data Download Date/Time: 08/12/2016 09:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

Agency: TYNDALL AFB

Date/Time: 01/12/2016 08:54:10

Data Download Date/Time: 01/21/2016 20:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: TYNDALL AFB

Date/Time: 01/12/2016 09:00:08

Data Download Date/Time: 01/21/2016 20:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 01/17/2016 20:20:28

Data Download Date/Time: 01/17/2016 20:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: AUBURNDALE PD

Date/Time: 07/27/2016 01:35:29

Data Download Date/Time: 08/29/2016 04:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

Agency: AUBURNDALE PD

Date/Time: 10/23/2016 21:14:52

Data Download Date/Time: 10/23/2016 21:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: AUBURNDALE PD

Date/Time: 11/25/2016 02:50:10

Data Download Date/Time: 11/25/2016 03:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency: EDGEWATER P.D.

Date/Time: 07/28/2016 06:29:53

Data Download Date/Time: 08/20/2016 07:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

Agency: BROWARD COUNTY SO

Date/Time: 04/20/2016 10:48:49

Data Download Date/Time: 04/25/2016 18:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

Agency: BROWARD COUNTY SO

Date/Time: 12/23/2016 12:12:39

Data Download Date/Time: 12/23/2016 12:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

Agency: ORANGE COUNTY S.O.

Date/Time: 04/21/2016 09:19:10

Data Download Date/Time: 04/21/2016 10:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001259

Agency: PINELLAS COUNTY SO

Date/Time: 04/20/2016 13:59:45

Data Download Date/Time: 04/29/2016 17:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PINELLAS COUNTY SO

Date/Time: 04/20/2016 14:29:12

Data Download Date/Time: 04/29/2016 17:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001259

Agency: PINELLAS COUNTY SO

Date/Time: 04/20/2016 16:18:49

Data Download Date/Time: 04/29/2016 17:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PINELLAS COUNTY SO

Date/Time: 04/20/2016 16:19:44

Data Download Date/Time: 04/29/2016 17:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001259

Agency: PINELLAS COUNTY SO

Date/Time: 04/20/2016 16:24:13

Data Download Date/Time: 04/29/2016 17:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PINELLAS COUNTY SO

Date/Time: 04/20/2016 16:27:27

Data Download Date/Time: 04/29/2016 17:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001259

Agency: PINELLAS COUNTY SO

Date/Time: 04/26/2016 10:13:02

Data Download Date/Time: 04/29/2016 17:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PINELLAS COUNTY SO

Date/Time: 04/26/2016 10:15:06

Data Download Date/Time: 04/29/2016 17:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001259

Agency: PINELLAS COUNTY SO

Date/Time: 04/29/2016 11:26:21

Data Download Date/Time: 04/29/2016 17:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

Agency: HIALEAH GARDENS PD

Date/Time: 05/09/2016 09:07:41

Data Download Date/Time: 05/09/2016 11:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

Agency: NASSAU COUNTY SO

Date/Time: 07/18/2016 10:36:26

Data Download Date/Time: 07/21/2016 05:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

Agency: HIGH SPRINGS PD

Date/Time: 04/19/2016 05:57:32

Data Download Date/Time: 04/19/2016 07:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

Agency: UNION COUNTY SO

Date/Time: 01/29/2016 13:29:17

Data Download Date/Time: 01/29/2016 15:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: UNION COUNTY SO

Date/Time: 01/29/2016 13:31:12

Data Download Date/Time: 01/29/2016 15:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

Agency: UNION COUNTY SO

Date/Time: 04/01/2016 09:33:46

Data Download Date/Time: 04/29/2016 11:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

Agency: UNION COUNTY SO

Date/Time: 03/01/2016 15:39:36

Data Download Date/Time: 03/01/2016 17:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: UNION COUNTY SO

Date/Time: 04/01/2016 09:29:26

Data Download Date/Time: 04/29/2016 11:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

Agency: BRADFORD COUNTY SO

Date/Time: 02/24/2016 23:52:43

Data Download Date/Time: 02/25/2016 10:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: BRADFORD COUNTY SO

Date/Time: 10/18/2016 09:48:18

Data Download Date/Time: 10/18/2016 10:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

Agency: SANTA ROSA COUNTY SO

Date/Time: 01/08/2016 10:28:49

Data Download Date/Time: 01/08/2016 12:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: SANTA ROSA COUNTY SO

Date/Time: 01/08/2016 11:11:42

Data Download Date/Time: 02/26/2016 16:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

Agency: SANTA ROSA COUNTY SO

Date/Time: 01/08/2016 11:12:40

Data Download Date/Time: 02/26/2016 16:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: SANTA ROSA COUNTY SO

Date/Time: 01/08/2016 11:13:42

Data Download Date/Time: 02/26/2016 16:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

Agency: SANTA ROSA COUNTY SO

Date/Time: 01/08/2016 11:14:51

Data Download Date/Time: 02/26/2016 16:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

Agency: SANTA ROSA COUNTY SO

Date/Time: 01/08/2016 10:27:19

Data Download Date/Time: 01/08/2016 11:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: SANTA ROSA COUNTY SO

Date/Time: 01/08/2016 11:01:38

Data Download Date/Time: 02/26/2016 16:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

Agency: SANTA ROSA COUNTY SO

Date/Time: 01/08/2016 11:02:42

Data Download Date/Time: 02/26/2016 16:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: SANTA ROSA COUNTY SO

Date/Time: 03/02/2016 03:40:03

Data Download Date/Time: 03/31/2016 12:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

Agency: SANTA ROSA COUNTY SO

Date/Time: 03/02/2016 03:40:48

Data Download Date/Time: 03/31/2016 12:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

Agency: WALTON COUNTY SO

Date/Time: 06/13/2016 08:25:45

Data Download Date/Time: 06/13/2016 12:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: WALTON COUNTY SO

Date/Time: 06/13/2016 08:26:55

Data Download Date/Time: 06/13/2016 12:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

Agency: WALTON COUNTY SO

Date/Time: 06/14/2016 14:18:25

Data Download Date/Time: 06/21/2016 09:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: WALTON COUNTY SO

Date/Time: 06/14/2016 14:19:50

Data Download Date/Time: 06/21/2016 09:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

Agency: WALTON COUNTY SO

Date/Time: 06/14/2016 14:30:30

Data Download Date/Time: 06/21/2016 09:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: WALTON COUNTY SO

Date/Time: 06/21/2016 13:39:33

Data Download Date/Time: 06/27/2016 08:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

Agency: WALTON COUNTY SO

Date/Time: 07/19/2016 10:47:06

Data Download Date/Time: 08/01/2016 15:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: WALTON COUNTY SO

Date/Time: 07/21/2016 15:21:05

Data Download Date/Time: 08/01/2016 15:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

Agency: WALTON COUNTY SO

Date/Time: 08/10/2016 10:32:40

Data Download Date/Time: 08/10/2016 15:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001313

Agency: SPRINGFIELD PD

Date/Time: 03/29/2016 13:05:14

Data Download Date/Time: 03/29/2016 15:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

Agency: INDIAN RIVER CO. SO

Date/Time: 02/01/2016 09:03:21

Data Download Date/Time: 02/01/2016 11:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

Agency: DESOTO COUNTY SO

Date/Time: 02/29/2016 02:46:39

Data Download Date/Time: 02/29/2016 03:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: DESOTO COUNTY SO

Date/Time: 05/04/2016 10:39:10

Data Download Date/Time: 05/31/2016 06:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

Agency: DESOTO COUNTY SO

Date/Time: 05/31/2016 05:59:02

Data Download Date/Time: 05/31/2016 06:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

Agency: DESOTO COUNTY SO

Date/Time: 02/01/2016 14:55:36

Data Download Date/Time: 06/25/2016 02:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

Agency: EUSTIS P.D.

Date/Time: 01/18/2016 22:04:18

Data Download Date/Time: 01/18/2016 22:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: EUSTIS P.D.

Date/Time: 08/23/2016 09:35:59

Data Download Date/Time: 08/23/2016 10:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

Agency: EUSTIS P.D.

Date/Time: 09/06/2016 14:46:32

Data Download Date/Time: 09/06/2016 15:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: EUSTIS P.D.

Date/Time: 09/07/2016 00:55:14

Data Download Date/Time: 10/24/2016 17:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

Agency: EUSTIS P.D.

Date/Time: 12/02/2016 09:33:13

Data Download Date/Time: 12/02/2016 10:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001360

Agency: EVERGLADES NAT PARK

Date/Time: 08/05/2016 15:23:27

Data Download Date/Time: 11/23/2016 10:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

Agency: CLAY COUNTY SO

Date/Time: 01/21/2016 19:50:41

Data Download Date/Time: 01/21/2016 20:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CLAY COUNTY SO

Date/Time: 03/21/2016 14:49:41

Data Download Date/Time: 03/21/2016 16:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

Agency: CLAY COUNTY SO

Date/Time: 03/21/2016 16:05:51

Data Download Date/Time: 03/21/2016 16:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CLAY COUNTY SO

Date/Time: 03/21/2016 16:07:53

Data Download Date/Time: 03/21/2016 16:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

Agency: CLAY COUNTY SO

Date/Time: 03/21/2016 16:17:44

Data Download Date/Time: 03/21/2016 16:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CLAY COUNTY SO

Date/Time: 04/27/2016 13:45:43

Data Download Date/Time: 04/27/2016 14:21

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

Agency: CLAY COUNTY SO

Date/Time: 07/03/2016 14:01:27

Data Download Date/Time: 07/03/2016 15:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CLAY COUNTY SO

Date/Time: 09/25/2016 18:18:37

Data Download Date/Time: 09/25/2016 18:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

Agency: CLAY COUNTY SO

Date/Time: 10/09/2016 09:45:20

Data Download Date/Time: 10/09/2016 10:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CLAY COUNTY SO

Date/Time: 11/06/2016 12:25:19

Data Download Date/Time: 11/06/2016 16:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency: PATRICK AFB SFS

Date/Time: 04/20/2016 09:34:53

Data Download Date/Time: 04/20/2016 12:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

Agency: TYNDALL AFB

Date/Time: 01/12/2016 08:45:55

Data Download Date/Time: 01/20/2016 21:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: TYNDALL AFB

Date/Time: 01/12/2016 08:50:46

Data Download Date/Time: 01/20/2016 21:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001650

Agency: KEY BISCAYNE PD

Date/Time: 11/09/2016 12:48:00

Data Download Date/Time: 11/18/2016 14:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency: ST PETERSBURG PD

Date/Time: 02/05/2016 01:24:56

Data Download Date/Time: 02/24/2016 23:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: ST PETERSBURG PD

Date/Time: 07/25/2016 14:38:58

Data Download Date/Time: 07/26/2016 20:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

Agency: MIAMI PD

Date/Time: 11/29/2016 15:34:15

Data Download Date/Time: 11/29/2016 16:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency: PALM BAY P.D.

Date/Time: 11/23/2016 05:03:31

Data Download Date/Time: 12/28/2016 12:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

Agency: LAKELAND PD

Date/Time: 04/25/2016 10:08:27

Data Download Date/Time: 04/29/2016 14:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: LAKELAND PD

Date/Time: 04/25/2016 10:46:06

Data Download Date/Time: 04/29/2016 14:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

Agency: LAKELAND PD

Date/Time: 04/29/2016 09:48:54

Data Download Date/Time: 04/29/2016 14:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: LAKELAND PD

Date/Time: 11/14/2016 12:31:00

Data Download Date/Time: 11/15/2016 12:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002096

Agency: MIAMI-DADE COLLEGE

Date/Time: 03/29/2016 16:38:38

Data Download Date/Time: 03/29/2016 16:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: MIAMI-DADE COLLEGE

Date/Time: 03/29/2016 16:56:21

Data Download Date/Time: 03/29/2016 16:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002096

Agency: MIAMI-DADE COLLEGE

Date/Time: 11/08/2016 10:33:24

Data Download Date/Time: 12/30/2016 15:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

Agency: MIAMI BEACH PD

Date/Time: 02/13/2016 02:57:36

Data Download Date/Time: 02/23/2016 10:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: MIAMI BEACH PD

Date/Time: 09/14/2016 11:02:47

Data Download Date/Time: 09/26/2016 12:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

Agency: HIALEAH PD

Date/Time: 02/11/2016 15:59:32

Data Download Date/Time: 03/23/2016 13:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: HIALEAH PD

Date/Time: 04/25/2016 13:16:09

Data Download Date/Time: 05/18/2016 11:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

Agency: HIALEAH PD

Date/Time: 05/18/2016 13:28:07

Data Download Date/Time: 06/22/2016 13:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency: HIALEAH PD

Date/Time: 02/11/2016 11:51:11

Data Download Date/Time: 02/11/2016 12:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: HIALEAH PD

Date/Time: 05/18/2016 12:18:47

Data Download Date/Time: 06/22/2016 12:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency: HIALEAH PD

Date/Time: 06/22/2016 13:14:01

Data Download Date/Time: 07/14/2016 13:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: HIALEAH PD

Date/Time: 07/14/2016 13:18:58

Data Download Date/Time: 08/22/2016 11:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency: HIALEAH PD

Date/Time: 08/22/2016 12:19:17

Data Download Date/Time: 09/21/2016 11:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

Agency: HIALEAH PD

Date/Time: 01/20/2016 12:35:44

Data Download Date/Time: 01/21/2016 11:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: HIALEAH PD

Date/Time: 02/11/2016 14:56:11

Data Download Date/Time: 03/31/2016 12:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

Agency: HIALEAH PD

Date/Time: 02/11/2016 14:57:44

Data Download Date/Time: 03/31/2016 12:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: HIALEAH PD

Date/Time: 02/11/2016 15:01:36

Data Download Date/Time: 03/31/2016 12:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

Agency: HIALEAH PD

Date/Time: 04/19/2016 14:22:44

Data Download Date/Time: 05/18/2016 12:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: HIALEAH PD

Date/Time: 05/18/2016 12:43:52

Data Download Date/Time: 06/22/2016 13:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

Agency: HIALEAH PD

Date/Time: 06/22/2016 13:51:05

Data Download Date/Time: 07/14/2016 13:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: HIALEAH PD

Date/Time: 06/22/2016 13:52:23

Data Download Date/Time: 07/14/2016 13:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

Agency: HIALEAH PD

Date/Time: 07/14/2016 13:43:23

Data Download Date/Time: 08/22/2016 12:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

Agency: FWC

Date/Time: 02/24/2016 16:39:51

Data Download Date/Time: 02/25/2016 10:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency: PASCO COUNTY SO

Date/Time: 02/25/2016 19:09:25

Data Download Date/Time: 02/26/2016 14:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 03/16/2016 09:15:03

Data Download Date/Time: 03/16/2016 10:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency: PASCO COUNTY SO

Date/Time: 04/08/2016 11:14:57

Data Download Date/Time: 04/08/2016 12:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 05/05/2016 08:34:56

Data Download Date/Time: 05/05/2016 09:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency: PASCO COUNTY SO

Date/Time: 06/03/2016 08:46:46

Data Download Date/Time: 06/03/2016 09:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 07/07/2016 09:08:27

Data Download Date/Time: 07/07/2016 10:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency: PASCO COUNTY SO

Date/Time: 08/19/2016 10:06:49

Data Download Date/Time: 08/19/2016 11:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 09/15/2016 09:31:58

Data Download Date/Time: 09/15/2016 10:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency: PASCO COUNTY SO

Date/Time: 10/14/2016 07:49:57

Data Download Date/Time: 10/14/2016 08:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PASCO COUNTY SO

Date/Time: 11/03/2016 08:25:49

Data Download Date/Time: 11/03/2016 09:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency: PASCO COUNTY SO

Date/Time: 12/01/2016 11:05:31

Data Download Date/Time: 12/01/2016 12:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency: FL HIGHWAY PATROL

Date/Time: 02/01/2016 14:34:35

Data Download Date/Time: 02/01/2016 15:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency: FHP TROOP E MIAMI

Date/Time: 04/19/2016 07:55:35

Data Download Date/Time: 04/21/2016 10:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency: FL HIGHWAY PATROL

Date/Time: 02/01/2016 14:33:35

Data Download Date/Time: 02/01/2016 15:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: FL HIGHWAY PATROL

Date/Time: 03/03/2016 15:44:28

Data Download Date/Time: 04/06/2016 10:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency: FL HIGHWAY PATROL

Date/Time: 03/03/2016 15:45:01

Data Download Date/Time: 04/06/2016 10:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: FL HIGHWAY PATROL

Date/Time: 03/03/2016 15:45:47

Data Download Date/Time: 04/06/2016 10:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency: FL HIGHWAY PATROL

Date/Time: 03/03/2016 17:06:40

Data Download Date/Time: 04/06/2016 10:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency: FL HIGHWAY PATROL

Date/Time: 05/27/2016 10:19:25

Data Download Date/Time: 05/31/2016 16:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: FL HIGHWAY PATROL

Date/Time: 08/02/2016 15:49:34

Data Download Date/Time: 08/02/2016 17:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency: HURLBURT FIELD PD

Date/Time: 06/02/2016 10:08:35

Data Download Date/Time: 06/02/2016 11:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: HURLBURT FIELD PD

Date/Time: 06/02/2016 10:09:08

Data Download Date/Time: 06/02/2016 11:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency: HURLBURT FIELD PD

Date/Time: 06/02/2016 10:09:49

Data Download Date/Time: 06/02/2016 11:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: HURLBURT FIELD PD

Date/Time: 06/02/2016 10:11:56

Data Download Date/Time: 06/02/2016 11:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency: HURLBURT FIELD PD

Date/Time: 06/02/2016 10:13:20

Data Download Date/Time: 06/02/2016 11:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: HURLBURT FIELD PD

Date/Time: 06/02/2016 10:14:05

Data Download Date/Time: 06/02/2016 11:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency: HURLBURT FIELD PD

Date/Time: 06/02/2016 11:06:40

Data Download Date/Time: 06/02/2016 12:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 01/26/2016 16:36:35

Data Download Date/Time: 01/28/2016 16:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CRESTVIEW PD

Date/Time: 02/29/2016 15:08:48

Data Download Date/Time: 02/29/2016 16:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 04/04/2016 16:05:19

Data Download Date/Time: 04/04/2016 17:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CRESTVIEW PD

Date/Time: 05/31/2016 20:13:30

Data Download Date/Time: 06/01/2016 16:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 06/30/2016 15:35:40

Data Download Date/Time: 06/30/2016 17:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CRESTVIEW PD

Date/Time: 07/28/2016 16:23:53

Data Download Date/Time: 08/02/2016 10:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 08/29/2016 15:56:04

Data Download Date/Time: 09/29/2016 16:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CRESTVIEW PD

Date/Time: 09/29/2016 14:54:14

Data Download Date/Time: 09/29/2016 16:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 10/24/2016 15:46:51

Data Download Date/Time: 10/24/2016 17:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CRESTVIEW PD

Date/Time: 11/17/2016 14:12:37

Data Download Date/Time: 11/21/2016 09:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 12/12/2016 15:55:09

Data Download Date/Time: 12/12/2016 17:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

Agency: PATRICK AFB

Date/Time: 05/14/2016 13:00:09

Data Download Date/Time: 05/14/2016 13:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PATRICK AFB

Date/Time: 06/23/2016 13:22:08

Data Download Date/Time: 07/21/2016 17:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

Agency: PATRICK AFB

Date/Time: 06/23/2016 13:25:47

Data Download Date/Time: 07/21/2016 17:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: PATRICK AFB

Date/Time: 06/23/2016 13:27:45

Data Download Date/Time: 07/21/2016 17:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005146

Agency: PATRICK AFB

Date/Time: 06/23/2016 13:29:03

Data Download Date/Time: 07/21/2016 17:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency: HERNANDO COUNTY SO

Date/Time: 12/28/2016 00:39:44

Data Download Date/Time: 01/19/2017 18:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency: FORT MYERS PD

Date/Time: 04/13/2016 09:27:23

Data Download Date/Time: 04/13/2016 10:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency: MIAMI DADE PD

Date/Time: 11/10/2016 16:13:41

Data Download Date/Time: 11/14/2016 16:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency: FORT MEYERS PD

Date/Time: 08/18/2016 13:03:24

Data Download Date/Time: 08/23/2016 14:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

Agency: LAKELAND PD

Date/Time: 11/14/2016 11:13:49

Data Download Date/Time: 11/15/2016 10:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006221

Agency: LAKE CITY PD

Date/Time: 08/09/2016 01:18:43

Data Download Date/Time: 09/06/2016 00:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

Agency: ASTATULA PD

Date/Time: 07/03/2016 01:28:47

Data Download Date/Time: 07/03/2016 02:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: ASTATULA PD

Date/Time: 07/22/2016 02:36:25

Data Download Date/Time: 07/22/2016 03:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

Agency: BREVARD COUNTY S.O.

Date/Time: 08/24/2016 17:58:20

Data Download Date/Time: 08/24/2016 19:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006235

Agency: ASTATULA PD

Date/Time: 07/22/2016 02:38:11

Data Download Date/Time: 07/22/2016 03:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 01/21/2016 19:47:13

Data Download Date/Time: 01/21/2016 20:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CLAY COUNTY SO

Date/Time: 03/21/2016 14:44:43

Data Download Date/Time: 03/21/2016 15:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 03/21/2016 15:04:56

Data Download Date/Time: 03/21/2016 15:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CLAY COUNTY SO

Date/Time: 04/27/2016 13:13:55

Data Download Date/Time: 04/27/2016 13:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 05/21/2016 13:57:25

Data Download Date/Time: 05/21/2016 14:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CLAY COUNTY SO

Date/Time: 07/03/2016 13:08:39

Data Download Date/Time: 07/03/2016 13:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 09/25/2016 18:00:13

Data Download Date/Time: 09/25/2016 18:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: CLAY COUNTY SO

Date/Time: 10/09/2016 08:56:54

Data Download Date/Time: 10/09/2016 09:41

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 11/06/2016 11:23:43

Data Download Date/Time: 11/06/2016 12:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

Agency: FL HIGHWAY PATROL

Date/Time: 04/21/2016 13:11:17

Data Download Date/Time: 04/25/2016 13:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

Agency: HILLSBOROUGH CO SO

Date/Time: 09/30/2016 17:14:31

Data Download Date/Time: 10/03/2016 14:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

Agency: ESCAMBIA COUNTY SO

Date/Time: 03/10/2016 23:14:46

Data Download Date/Time: 03/27/2016 06:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: ESCAMBIA COUNTY SO

Date/Time: 03/27/2016 04:35:21

Data Download Date/Time: 03/27/2016 06:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

Agency: ESCAMBIA COUNTY SO

Date/Time: 03/10/2016 23:01:32

Data Download Date/Time: 05/03/2016 15:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

---

Agency: ESCAMBIA COUNTY SO

Date/Time: 03/10/2016 23:09:42

Data Download Date/Time: 05/03/2016 15:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

Agency: ESCAMBIA COUNTY SO

Date/Time: 03/12/2016 01:49:12

Data Download Date/Time: 05/03/2016 15:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

Agency: ESCAMBIA COUNTY SO

Date/Time: 06/29/2016 22:47:01

Data Download Date/Time: 07/12/2016 08:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

Agency: ESCAMBIA COUNTY SO

Date/Time: 07/01/2016 15:56:32

Data Download Date/Time: 07/12/2016 08:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

Agency: FL HIGHWAY PATROL

Date/Time: 08/02/2016 17:04:00

Data Download Date/Time: 08/02/2016 18:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Diagnostic Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

Agency: FL HIGHWAY PATROL

Date/Time: 11/30/2016 12:21:33

Data Download Date/Time: 11/30/2016 15:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

---