

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER COUNTY SO      **Date** 01/29/2018      **Type** Agency Inspection  
**Inspector** JACOBSON,JAMES,W      **Time** 01:55:54      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A3  
**Expiration** 03/08/2018      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/29/2018 01:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

|                                  |                        |                                   |
|----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> SUMTER COUNTY SO   | <b>Date</b> 02/08/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> BELL,DANIELLE,M | <b>Time</b> 16:44:23   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG708807 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 03/29/2019   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.082  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.081  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0009 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 02/08/2018 16:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | SUMTER COUNTY SO | <b>Date</b>       | 02/25/2018 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | FORINASH,GARY,B  | <b>Time</b>       | 19:44:25   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201603D          | <b>0.08 Lot #</b> | 201611B    | <b>0.20 Lot #</b> | 201702B           | <b>0.08 Gas Lot #</b> | 20016080A3 |
| <b>Expiration</b>           | 03/08/2018       | <b>Expiration</b> | 11/15/2018 | <b>Expiration</b> | 02/23/2019        | <b>Expiration</b>     | 09/05/2018 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/25/2018 19:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER COUNTY SO      **Date** 03/31/2018      **Type** Agency Inspection  
**Inspector** JACOBSON,JAMES,W      **Time** 01:44:33      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A3  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/31/2018 01:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | SUMTER COUNTY SO | <b>Date</b>       | 04/26/2018 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | FORINASH,GARY,B  | <b>Time</b>       | 10:30:13   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201703A          | <b>0.08 Lot #</b> | 201611B    | <b>0.20 Lot #</b> | 201702B           | <b>0.08 Gas Lot #</b> | 20016080A3 |
| <b>Expiration</b>           | 03/17/2019       | <b>Expiration</b> | 11/15/2018 | <b>Expiration</b> | 02/23/2019        | <b>Expiration</b>     | 09/05/2018 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 04/26/2018 10:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER COUNTY SO      **Date** 05/30/2018      **Type** Agency Inspection  
**Inspector** JACOBSON,JAMES,W      **Time** 13:26:05      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A3  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/30/2018 13:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

|                  |                  |             |            |                 |                   |
|------------------|------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | SUMTER COUNTY SO | <b>Date</b> | 06/28/2018 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | FORINASH,GARY,B  | <b>Time</b> | 08:06:23   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |   |                   |     |
|-----------------------------|---|-------------------|-----|
| <b>Number of Simulators</b> | 5 | <b>Compliance</b> | Yes |
|-----------------------------|---|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A:WAITED TOO LONG.

**Data Download Date/Time** 06/28/2018 09:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | SUMTER COUNTY SO | <b>Date</b>       | 06/28/2018 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | FORINASH,GARY,B  | <b>Time</b>       | 09:04:03   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201703A          | <b>0.08 Lot #</b> | 201611B    | <b>0.20 Lot #</b> | 201702B           | <b>0.08 Gas Lot #</b> | 20016080A3 |
| <b>Expiration</b>           | 03/07/2019       | <b>Expiration</b> | 11/15/2018 | <b>Expiration</b> | 02/23/2019        | <b>Expiration</b>     | 09/05/2018 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 06/28/2018 09:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER COUNTY SO      **Date** 07/20/2018      **Type** Agency Inspection  
**Inspector** JACOBSON,JAMES,W      **Time** 13:58:05      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 13518080A5  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 08/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/20/2018 14:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER COUNTY SO      **Date** 08/28/2018      **Type** Agency Inspection  
**Inspector** FORINASH,GARY,B      **Time** 19:44:27      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 13518080A5  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 08/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/28/2018 21:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER COUNTY SO      **Date** 09/27/2018      **Type** Agency Inspection  
**Inspector** JACOBSON,JAMES,W      **Time** 09:27:21      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 13518080A5  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 08/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/27/2018 09:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER COUNTY SO      **Date** 10/28/2018      **Type** Agency Inspection  
**Inspector** FORINASH,GARY,B      **Time** 09:54:17      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 13518080A5  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 08/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/28/2018 10:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

**Agency** SUMTER COUNTY SO      **Date** 11/16/2018      **Type** Agency Inspection  
**Inspector** JACOBSON,JAMES,W      **Time** 12:08:36      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 13518080A5  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 08/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/16/2018 12:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | SUMTER COUNTY SO | <b>Date</b>       | 12/23/2018 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | FORINASH,GARY,B  | <b>Time</b>       | 09:21:08   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201703A          | <b>0.08 Lot #</b> | 201709A    | <b>0.20 Lot #</b> | 201702B           | <b>0.08 Gas Lot #</b> | 13518080A5 |
| <b>Expiration</b>           | 03/07/2019       | <b>Expiration</b> | 09/12/2019 | <b>Expiration</b> | 02/23/2019        | <b>Expiration</b>     | 08/05/2020 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/23/2018 09:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

**Agency** CITRUS COUNTY SO      **Date** 01/24/2018      **Type** Agency Inspection  
**Inspector** ROBERSON, WILLIAM, E      **Time** 11:26:55      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 26316080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 11/05/2018

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/24/2018 11:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

**Agency** CITRUS COUNTY SO      **Date** 02/27/2018      **Type** Agency Inspection  
**Inspector** ROBERSON, WILLIAM, E      **Time** 11:46:17      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 26316080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 11/05/2018

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/27/2018 11:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

**Agency** CITRUS COUNTY SO      **Date** 03/21/2018      **Type** Agency Inspection  
**Inspector** ROBERSON, WILLIAM, E      **Time** 10:04:05      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 26316080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 11/05/2018

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/21/2018 10:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

**Agency** CITRUS COUNTY SO      **Date** 04/30/2018      **Type** Agency Inspection  
**Inspector** BOWERMASTER,MARY,E      **Time** 13:56:31      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 26316080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 11/05/2018

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/30/2018 14:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

**Agency** CITRUS COUNTY SO      **Date** 05/30/2018      **Type** Agency Inspection  
**Inspector** ROBERSON, WILLIAM, E      **Time** 07:30:54      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 26316080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 11/05/2018

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.052  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.083  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/30/2018 07:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

|                                       |                              |                               |                                  |
|---------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> CITRUS COUNTY SO        | <b>Date</b> 06/29/2018       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> ROBERSON, WILLIAM, E | <b>Time</b> 10:05:43         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201703A             | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 26316080A2 |
| <b>Expiration</b> 03/07/2019          | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 11/05/2018     |
| <b>Number of Simulators</b> 3         | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/29/2018 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> CITRUS COUNTY SO     | <b>Date</b> 07/09/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SHANAHAN, JAKE, L | <b>Time</b> 11:43:26   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805702 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.082  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.082  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.082  |
| 0.20 g/210L Test # 7                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.082  |
| 0.20 g/210L Test # 8                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.082  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.082  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0008 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 07/09/2018 11:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

**Agency** CITRUS COUNTY SO      **Date** 07/31/2018      **Type** Agency Inspection  
**Inspector** ROBERSON, WILLIAM, E      **Time** 11:25:34      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 13518080A5  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 08/05/2020

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/31/2018 11:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

**Agency** CITRUS COUNTY SO      **Date** 08/31/2018      **Type** Agency Inspection  
**Inspector** ROBERSON, WILLIAM, E      **Time** 09:00:16      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 13518080A5  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 08/05/2020

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/31/2018 09:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

|                                      |                              |                               |                                  |
|--------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> CITRUS COUNTY SO       | <b>Date</b> 09/24/2018       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> ROBERSON, WILIAM, E | <b>Time</b> 14:25:35         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201703A            | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 13518080A5 |
| <b>Expiration</b> 03/07/2019         | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 08/05/2020     |
| <b>Number of Simulators</b> 3        | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/24/2018 14:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

**Agency** CITRUS COUNTY SO      **Date** 10/31/2018      **Type** Agency Inspection  
**Inspector** ROBERSON, WILLIAM, E      **Time** 09:12:26      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 13518080A5  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 08/05/2020

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/05/2018 10:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

**Agency** CITRUS COUNTY SO      **Date** 11/15/2018      **Type** Agency Inspection  
**Inspector** ROBERSON, WILLIAM, E      **Time** 17:53:03      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 13518080A5  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 08/05/2020

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/15/2018 17:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000817

**Agency** CITRUS COUNTY SO      **Date** 12/31/2018      **Type** Agency Inspection  
**Inspector** ROBERSON, WILLIAM, E      **Time** 09:59:48      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201808E      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 13518080A5  
**Expiration** 03/07/2019      **Expiration** 08/22/2020      **Expiration** 02/23/2019      **Expiration** 08/05/2020

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/31/2018 10:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

**Agency** CITRUS COUNTY SO      **Date** 01/24/2018      **Type** Agency Inspection  
**Inspector** ROBERSON, WILLIAM, E      **Time** 11:31:06      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 26316080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 11/05/2018

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/24/2018 11:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

**Agency** CITRUS COUNTY SO      **Date** 02/27/2018      **Type** Agency Inspection  
**Inspector** ROBERSON, WILLIAM, E      **Time** 11:50:36      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 26316080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 11/05/2018

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.204  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/27/2018 11:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

**Agency** CITRUS COUNTY SO      **Date** 03/21/2018      **Type** Agency Inspection  
**Inspector** ROBERSON, WILLIAM, E      **Time** 10:08:11      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 26316080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 11/05/2018

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.204  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.205  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

ACETONE BOTTLE TO CLOSE.

**Data Download Date/Time** 03/21/2018 10:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

**Agency** CITRUS COUNTY SO      **Date** 04/30/2018      **Type** Agency Inspection  
**Inspector** ROBERSON, WILLIAM, E      **Time** 14:00:27      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 26316080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 11/05/2018

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.206  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.207  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.207  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/30/2018 14:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

**Agency** CITRUS COUNTY SO      **Date** 05/30/2018      **Type** Agency Inspection  
**Inspector** ROBERSON, WILLIAM, E      **Time** 07:56:31      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 26316080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 11/05/2018

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/30/2018 07:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

|                  |                  |             |            |                 |                   |
|------------------|------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | CITRUS COUNTY SO | <b>Date</b> | 11/14/2018 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | SHANAHAN,JAKE,L  | <b>Time</b> | 07:10:06   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASS TO OPERATE

**Data Download Date/Time** 11/14/2018 09:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> CITRUS COUNTY SO     | <b>Date</b> 11/14/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SHANAHAN, JAKE, L | <b>Time</b> 09:31:22   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG806701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1                      | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1                      | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.050  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.081  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.200  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.000  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.050  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.081  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.079  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3               | INT    |
|                                             |        | Alcohol Free Test # 4                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 4                      | 0.050  |
|                                             |        | 0.08 g/210L Test # 4                      | 0.081  |
|                                             |        | 0.20 g/210L Test # 4                      | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4          | 0.079  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4               | INT    |
|                                             |        | Alcohol Free Test # 5                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 5                      | 0.050  |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

|                                            |       |
|--------------------------------------------|-------|
| 0.08 g/210L Test # 5                       | 0.081 |
| 0.20 g/210L Test # 5                       | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5           | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 5  | 0.079 |
| Interferent Detect Test # 5                | INT   |
| Alcohol Free Test # 6                      | 0.000 |
| 0.05 g/210L Test # 6                       | 0.050 |
| 0.08 g/210L Test # 6                       | 0.081 |
| 0.20 g/210L Test # 6                       | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6           | 0.080 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 6  | 0.079 |
| Interferent Detect Test # 6                | INT   |
| Alcohol Free Test # 7                      | 0.000 |
| 0.05 g/210L Test # 7                       | 0.050 |
| 0.08 g/210L Test # 7                       | 0.081 |
| 0.20 g/210L Test # 7                       | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7           | 0.080 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 7  | 0.079 |
| Interferent Detect Test # 7                | INT   |
| Alcohol Free Test # 8                      | 0.000 |
| 0.05 g/210L Test # 8                       | 0.050 |
| 0.08 g/210L Test # 8                       | 0.081 |
| 0.20 g/210L Test # 8                       | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8           | 0.080 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 8  | 0.079 |
| Interferent Detect Test # 8                | INT   |
| Alcohol Free Test # 9                      | 0.000 |
| 0.05 g/210L Test # 9                       | 0.050 |
| 0.08 g/210L Test # 9                       | 0.081 |
| 0.20 g/210L Test # 9                       | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9           | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 9  | 0.080 |
| Interferent Detect Test # 9                | INT   |
| Alcohol Free Test # 10                     | 0.000 |
| 0.05 g/210L Test # 10                      | 0.050 |
| 0.08 g/210L Test # 10                      | 0.081 |
| 0.20 g/210L Test # 10                      | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10          | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 10 | 0.079 |
| Interferent Detect Test # 10               | INT   |

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

|                                                    |        |
|----------------------------------------------------|--------|
| 0.05 g/210L Test (Std Dev)                         | 0.0003 |
| 0.08 g/210L Test (Std Dev)                         | 0.0000 |
| 0.20 g/210L Test (Std Dev)                         | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev)             | 0.0334 |
| 0.08 g/210L Dry Gas Std Test (Std Dev)<br>(Repeat) | 0.0005 |

**Remarks:**

08: Control Outside Tolerance DGS NOT ATTACHED RETEST.

---

**Data Download Date/Time** 11/14/2018 09:34

**Printed: 14 January 2019**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

**Agency** CITRUS COUNTY SO      **Date** 12/31/2018      **Type** Agency Inspection  
**Inspector** ROBERSON, WILLIAM, E      **Time** 10:03:36      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201808E      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 13518080A5  
**Expiration** 03/07/2019      **Expiration** 08/22/2020      **Expiration** 02/23/2019      **Expiration** 08/05/2020

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/31/2018 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

**Agency** LAKE COUNTY S.O. **Date** 01/04/2018 **Type** Agency Inspection  
**Inspector** PITCHER,CHESSHER,/ **Time** 17:55:03 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 34416080A2  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 02/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/04/2018 18:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

**Agency** LAKE COUNTY S.O. **Date** 02/01/2018 **Type** Agency Inspection  
**Inspector** PITCHER,CHESSHER,/ **Time** 18:44:56 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 34416080A2  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 02/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/01/2018 19:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

**Agency** LAKE COUNTY S.O. **Date** 03/09/2018 **Type** Agency Inspection  
**Inspector** CHESSHER,PITCHER,/ **Time** 18:12:59 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 34416080A2  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 02/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/09/2018 18:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

**Agency** LAKE COUNTY S.O. **Date** 04/07/2018 **Type** Agency Inspection  
**Inspector** PITCHER,CHESSHER,/ **Time** 22:29:46 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 34416080A2  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 02/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/07/2018 22:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

|                                  |                        |                                   |
|----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> LAKE COUNTY S.O.   | <b>Date</b> 04/11/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> BELL,DANIELLE,M | <b>Time</b> 15:39:06   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0006      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0008 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0008 |

#### Remarks:

Data Download Date/Time 04/11/2018 15:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

**Agency** LAKE COUNTY S.O. **Date** 04/26/2018 **Type** Agency Inspection  
**Inspector** CHESSHER,BINDER,/ **Time** 22:01:28 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34416080A2  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A:ALCOHOL STILL IN MOUTH.

**Data Download Date/Time** 04/26/2018 22:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

**Agency** LAKE COUNTY S.O.  
**Inspector** PITCHER,BINDER,

**Date** 05/03/2018  
**Time** 18:01:08

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34416080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2019

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/03/2018 18:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

**Agency** LAKE COUNTY S.O. **Date** 06/07/2018 **Type** Agency Inspection  
**Inspector** BINDER,PITCHER,/ **Time** 21:56:34 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 814181  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/24/2020

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/07/2018 22:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

**Agency** LAKE COUNTY S.O. **Date** 07/06/2018 **Type** Agency Inspection  
**Inspector** BINDER,PITCHER,/ **Time** 23:21:40 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34416080A2  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/06/2018 23:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> LAKE COUNTY S.O.     | <b>Date</b> 08/15/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> BINDER,PITCHER, / | <b>Time</b> 23:21:22   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 814181 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 07/24/2020 |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.187  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.199  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.192  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.199  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

20: Control Outside Tolerance RECONNECTED SIM HOSE .

**Data Download Date/Time** 08/15/2018 23:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

**Agency** LAKE COUNTY S.O. **Date** 09/04/2018 **Type** Agency Inspection  
**Inspector** BINDER,PITCHER,/ **Time** 12:48:30 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 814181  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 07/24/2020

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/04/2018 13:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> LAKE COUNTY S.O.     | <b>Date</b> 10/18/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> PITCHER,BINDER, / | <b>Time</b> 20:41:44   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201802B    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 814181 |
| <b>Expiration</b> 02/06/2020 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 07/24/2020 |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

A F / M A: No Sample Provide/STARTED WITH ALCOHOL .

**Data Download Date/Time** 10/18/2018 20:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

**Agency** LAKE COUNTY S.O.  
**Inspector** PITCHER,BINDER,/

**Date** 11/15/2018  
**Time** 21:04:02

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 814181  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 07/24/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/15/2018 21:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

|                                    |                              |                               |                              |
|------------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> LAKE COUNTY S.O.     | <b>Date</b> 12/06/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> BINDER,PITCHER, / | <b>Time</b> 00:34:06         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201802B          | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201802G     | <b>0.08 Gas Lot #</b> 814181 |
| <b>Expiration</b> 02/06/2020       | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/22/2020  | <b>Expiration</b> 07/24/2020 |
| <b>Number of Simulators</b> 5      | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/06/2018 00:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

**Agency** LAKE COUNTY S.O. **Date** 01/01/2019 **Type** Agency Inspection  
**Inspector** BINDER,PITCHER,/ **Time** 20:52:14 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 814181  
**Expiration** 02/06/2020 **Expiration** 09/12/2919 **Expiration** 02/22/2020 **Expiration** 07/24/2020

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/01/2019 21:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

**Agency** LAKE COUNTY SO **Date** 01/04/2018 **Type** Agency Inspection  
**Inspector** PITCHER,CHESSHER,/ **Time** 18:02:57 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 12716080A7  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/05/2018

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/04/2018 18:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

**Agency** LAKE COUNTY SO      **Date** 02/01/2018      **Type** Agency Inspection  
**Inspector** PITCHER,CHESSHER,/      **Time** 18:56:04      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 12716080A7  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 06/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/01/2018 19:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

**Agency** LAKE COUNTY SO **Date** 03/09/2018 **Type** Agency Inspection  
**Inspector** CHESSHER,PITCHER,/ **Time** 18:19:21 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 12716080A7  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/05/2018

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/09/2018 18:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

|                                     |                              |                               |                                  |
|-------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> LAKE COUNTY SO        | <b>Date</b> 04/07/2018       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> PITCHER,CHESSHER,/ | <b>Time</b> 22:43:30         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201703A           | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 12716080A7 |
| <b>Expiration</b> 03/07/2019        | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 06/05/2018     |
| <b>Number of Simulators</b> 5       | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/07/2018 22:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

**Agency** LAKE COUNTY SO  
**Inspector** PITCHER,BINDER, /

**Date** 05/03/2018  
**Time** 17:49:35

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 12716080A7  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 06/05/2018

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/03/2018 17:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

|                                  |                        |                                   |
|----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> LAKE COUNTY SO     | <b>Date</b> 05/15/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> BELL,DANIELLE,M | <b>Time</b> 15:51:34   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805702 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0007      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.082  |
|                                             |        | 0.20 g/210L Test # 5             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.077  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.082  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.076  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.082  |
| 0.20 g/210L Test # 7                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.076  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.083  |
| 0.20 g/210L Test # 8                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.077  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.083  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.076  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.083  |
| 0.20 g/210L Test # 10                  | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.076  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0008 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0012 |

#### Remarks:

Data Download Date/Time 05/15/2018 15:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

**Agency** LAKE COUNTY SO  
**Inspector** BINDER,PITCHER,/

**Date** 06/07/2018  
**Time** 22:23:49

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 814181  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 07/24/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/07/2018 22:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

**Agency** LAKE COUNTY SO  
**Inspector** BINDER,PITCHER,/

**Date** 07/06/2018  
**Time** 23:29:33

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 814181  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 07/24/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/06/2018 23:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

**Agency** LAKE COUNTY SO  
**Inspector** BINDER,PITCHER,/

**Date** 08/15/2018  
**Time** 23:31:48

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 970670  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 06/07/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 08/15/2018 23:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

**Agency** LAKE COUNTY SO  
**Inspector** PITCHER,BINDER,/

**Date** 09/04/2018  
**Time** 12:41:49

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 970670  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 06/07/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 09/04/2018 13:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> LAKE COUNTY SO       | <b>Date</b> 10/18/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> PITCHER,BINDER, / | <b>Time</b> 20:44:53   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201802B    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 970670 |
| <b>Expiration</b> 02/06/2018 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 06/07/2021 |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

A F / M A: No Sample Provided/ STARTED WITH MOUTH ALCOHOL .

**Data Download Date/Time** 10/18/2018 20:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

**Agency** LAKE COUNTY SO  
**Inspector** PITCHER,BINDER,/

**Date** 11/15/2018  
**Time** 21:01:00

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 970670  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 06/07/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 11/15/2018 21:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

**Agency** LAKE COUNTY SO  
**Inspector** BINDER,PITCHER,/

**Date** 12/06/2018  
**Time** 00:51:27

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 970670  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 06/07/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 12/06/2018 01:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

**Agency** LAKE COUNTY SO  
**Inspector** BINDER,PITCHER,/

**Date** 01/01/2019  
**Time** 21:02:31

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 970670  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 06/07/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/01/2019 21:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

|                                     |                        |                               |  |
|-------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> LEESBURG P.D.         | <b>Date</b> 01/22/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> SOMMERSDORF,JOHN,G | <b>Time</b> 02:22:40   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 01316080A2 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 02/05/2018     |

  

**Number of Simulators** 3 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.036  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.048  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.077  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.192  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.036  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.048  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.037  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

05: Control Outside Tolerance BAD ORING REPLACED.

**Data Download Date/Time** 01/22/2018 02:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 02/27/2018 **Type** Agency Inspection  
**Inspector** SOMMERSDORF,JOHN,G **Time** 02:38:22 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A2  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 02/05/2020

**Number of Simulators** 3 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/27/2018 03:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 03/09/2018 **Type** Agency Inspection  
**Inspector** SOMMERSDORF,JOHN,G **Time** 02:38:52 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** **Expiration**

**Number of Simulators** 3 **Compliance** No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1          | 0.031  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1 | 0.041  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 1   | INT    |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Alcohol Free Test # 2         | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # 2          | 0.049  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.08 g/210L Test # 2          | 0.057  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # (Repeat) 2 | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2   | INT    |
| 0.08 g/210L Test: OK                        | No     | Alcohol Free Test # 3         | 0.000  |
| 0.20 g/210L Test: OK                        | No     | 0.05 g/210L Test # 3          | 0.049  |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | 0.08 g/210L Test # 3          | 0.059  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3   | INT    |

#### Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:BAD SEAL ON SIMULATOR JAR.

**Data Download Date/Time** 03/09/2018 03:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

|                                     |                        |                               |  |
|-------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> LEESBURG P.D.         | <b>Date</b> 03/09/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> SOMMERSDORF,JOHN,G | <b>Time</b> 03:35:03   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 34917080A2 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 02/05/2020     |

  

**Number of Simulators** 3 **Compliance** Yes

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                 | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                 | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                 | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1     | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1          | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # (Repeat) 1 | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                 | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                 | 0.079  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                 | 0.198  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2     | 0.082  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2          | 0.000  |
|                                             |        | Interferent Detect Test # (Repeat) 2 | INT    |
|                                             |        | Alcohol Free Test # 3                | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                 | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                 | 0.079  |
|                                             |        | 0.20 g/210L Test # 3                 | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3     | 0.081  |
|                                             |        | Interferent Detect Test # 3          | 0.000  |
|                                             |        | Interferent Detect Test # (Repeat) 3 | INT    |

**Remarks:**

Int Det:LOOSE JAR.

**Data Download Date/Time** 03/09/2018 03:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 04/15/2018 **Type** Agency Inspection  
**Inspector** SOMMERSDORF,JOHN,G **Time** 11:08:29 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A2  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

**Number of Simulators** 3 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/15/2018 11:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

|                                     |                        |                               |  |
|-------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> LEESBURG P.D.         | <b>Date</b> 05/22/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> SOMMERSDORF,JOHN,G | <b>Time</b> 09:27:24   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 34917080A2 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 02/05/2020     |

  

**Number of Simulators** 3 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

WAITED TO LONG AFTER MOUTH ALCOHOL SOLUTIONRFI Detect.

**Data Download Date/Time** 05/22/2018 09:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

|                                  |                        |                                   |
|----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> LEESBURG P.D.      | <b>Date</b> 06/05/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> BELL,DANIELLE,M | <b>Time</b> 11:22:43   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805702 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 06/05/2018 11:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 06/28/2018 **Type** Agency Inspection  
**Inspector** SOMMERSDORF,JOHN,G **Time** 00:05:31 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A2  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

**Number of Simulators** 3 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/28/2018 00:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 07/13/2018 **Type** Agency Inspection  
**Inspector** SOMMERSDORF,JOHN,G **Time** 03:55:57 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A2  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

**Number of Simulators** 3 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/13/2018 03:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 08/15/2018 **Type** Agency Inspection  
**Inspector** SOMMERSDORF,JOHN,G **Time** 03:36:42 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A2  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

**Number of Simulators** 3 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/15/2018 03:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 09/25/2018 **Type** Agency Inspection  
**Inspector** SOMMERSDORF,JOHN,G **Time** 08:13:23 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A2  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

**Number of Simulators** 3 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/25/2018 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 10/18/2018 **Type** Agency Inspection  
**Inspector** SOMMERSDORF,JOHN,G **Time** 14:41:06 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A2  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

**Number of Simulators** 3 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/18/2018 14:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 11/25/2018 **Type** Agency Inspection  
**Inspector** SOMMERSDORF,JOHN,G **Time** 21:34:12 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34917080A2  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2020

**Number of Simulators** 3 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/25/2018 21:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

**Agency** LEESBURG P.D. **Date** 12/12/2018 **Type** Agency Inspection  
**Inspector** SOMMERSDORF,JOHN,G **Time** 22:43:33 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 34917080A2  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 02/05/2020

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.190  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.192  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.192  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/12/2018 23:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

|                                   |                        |                               |  |
|-----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MOUNT DORA P.D.     | <b>Date</b> 01/31/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> DANCEL, KAREN, J | <b>Time</b> 04:13:21   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 583457 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 10/20/2019 |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                 | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                 | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                 | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1     | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1          | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # (Repeat) 1 | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                 | 0.050  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                 | 0.079  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                 | 0.200  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2     | 0.080  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2          | 0.000  |
|                                             |        | Interferent Detect Test # (Repeat) 2 | INT    |
|                                             |        | Alcohol Free Test # 3                | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                 | 0.050  |
|                                             |        | 0.08 g/210L Test # 3                 | 0.080  |
|                                             |        | 0.20 g/210L Test # 3                 | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3     | 0.080  |
|                                             |        | Interferent Detect Test # 3          | 0.000  |
|                                             |        | Interferent Detect Test # (Repeat) 3 | INT    |

**Remarks:**

Int Det: ADDED ACETONE TOO EARLY.

**Data Download Date/Time** 01/31/2018 04:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

|                                  |                        |                                   |
|----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MOUNT DORA P.D.    | <b>Date</b> 02/08/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> BELL,DANIELLE,M | <b>Time</b> 11:35:53   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG708807 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 03/29/2019   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 02/08/2018 11:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

**Agency** MOUNT DORA P.D. **Date** 02/23/2018 **Type** Agency Inspection  
**Inspector** DANCEL,KAREN,J **Time** 04:28:11 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 583457  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/20/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/23/2018 04:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

**Agency** MOUNT DORA P.D. **Date** 03/29/2018 **Type** Agency Inspection  
**Inspector** TAYLOR,KEITH,A **Time** 03:30:31 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 583457  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/20/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/29/2018 03:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

**Agency** MOUNT DORA P.D. **Date** 04/30/2018 **Type** Agency Inspection  
**Inspector** DANCEL,KAREN,J **Time** 21:11:21 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 583457  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/20/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/30/2018 21:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

**Agency** MOUNT DORA P.D. **Date** 05/31/2018 **Type** Agency Inspection  
**Inspector** DANCEL,KAREN,J **Time** 22:41:51 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 583457  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/20/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/31/2018 22:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

**Agency** MOUNT DORA P.D. **Date** 06/30/2018 **Type** Agency Inspection  
**Inspector** DANCEL,KAREN,J **Time** 19:52:14 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 583457  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/20/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/30/2018 19:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

**Agency** MOUNT DORA P.D. **Date** 07/31/2018 **Type** Agency Inspection  
**Inspector** DANCEL,KAREN,J **Time** 21:30:46 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 583457  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/20/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/31/2018 21:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

**Agency** MOUNT DORA P.D. **Date** 08/31/2018 **Type** Agency Inspection  
**Inspector** DANCEL,KAREN,J **Time** 22:37:28 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 583457  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/20/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 08/31/2018 22:43

**Printed: 14 January 2019**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

**Agency** MOUNT DORA P.D. **Date** 09/30/2018 **Type** Agency Inspection  
**Inspector** DANCEL,KAREN,J **Time** 06:33:16 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 583457  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/20/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/30/2018 06:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

**Agency** MOUNT DORA P.D. **Date** 10/30/2018 **Type** Agency Inspection  
**Inspector** DANCEL,KAREN,J **Time** 20:52:50 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 583457  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/20/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/02/2018 11:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

**Agency** MOUNT DORA P.D. **Date** 11/30/2018 **Type** Agency Inspection  
**Inspector** DANCEL,KAREN,J **Time** 23:48:28 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 583457  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/20/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Time-Date changed.

**Data Download Date/Time** 12/14/2018 08:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

**Agency** MOUNT DORA P.D. **Date** 12/31/2018 **Type** Agency Inspection  
**Inspector** DANCEL,KAREN,J **Time** 20:15:57 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 583457  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/20/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/04/2019 00:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

**Agency** MOUNT DORA P.D.  
**Inspector** DANCEL, KAREN, J

**Date** 01/04/2019  
**Time** 00:29:12

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 583457  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 10/20/2019

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/04/2019 00:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

**Agency** MARION COUNTY SO      **Date** 01/16/2018      **Type** Agency Inspection  
**Inspector** RAKER,JEFFERY,A      **Time** 18:18:35      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 770784  
**Expiration** 03/08/2018      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 04/27/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

WITHIN COMPLIANCE

**Data Download Date/Time** 01/16/2018 18:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

**Agency** MARION COUNTY SO      **Date** 02/21/2018      **Type** Agency Inspection  
**Inspector** RAKER,JEFFERY,A      **Time** 14:20:31      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 770784  
**Expiration** 03/08/2018      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 04/27/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

WITHIN COMPLIANCE

**Data Download Date/Time** 02/21/2018 14:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

**Agency** MARION COUNTY SO      **Date** 03/22/2018      **Type** Agency Inspection  
**Inspector** RAKER,JEFFERY,A      **Time** 13:33:24      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 770784  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 04/27/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.052  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

WITHIN COMPLIANCE

**Data Download Date/Time** 03/22/2018 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

**Agency** MARION COUNTY SO      **Date** 04/27/2018      **Type** Agency Inspection  
**Inspector** RAKER,JEFFERY,A      **Time** 02:48:51      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 770784  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 04/27/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

WITHIN COMPLIANCE

**Data Download Date/Time** 04/27/2018 02:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

**Agency** MARION COUNTY SO      **Date** 05/30/2018      **Type** Agency Inspection  
**Inspector** RAKER,JEFFERY,A      **Time** 16:21:01      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 770784  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 04/27/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

WITHIN COMPLIANCE

**Data Download Date/Time** 05/30/2018 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

**Agency** MARION COUNTY SO      **Date** 06/25/2018      **Type** Agency Inspection  
**Inspector** RAKER,JEFFERY,A      **Time** 13:27:54      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 644476  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 12/01/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

WITHIN COMPLIANCE

**Data Download Date/Time** 06/25/2018 13:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

|                                  |                        |                               |  |
|----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MARION COUNTY SO   | <b>Date</b> 07/21/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> RAKER,JEFFERY,A | <b>Time</b> 15:41:39   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 644476 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 12/01/2019 |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

WITHIN COMPLIANCE

**Data Download Date/Time** 07/21/2018 15:44

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

**Agency** MARION COUNTY SO      **Date** 08/27/2018      **Type** Agency Inspection  
**Inspector** RAKER,JEFFERY,A      **Time** 13:15:24      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 770784  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 04/27/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

WITHIN COMPLIANCE

**Data Download Date/Time** 08/27/2018 13:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

**Agency** MARION COUNTY SO      **Date** 09/29/2018      **Type** Agency Inspection  
**Inspector** RAKER,JEFFERY,A      **Time** 12:43:15      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 770784  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 04/27/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

WITHIN COMPLIANCE

**Data Download Date/Time** 09/29/2018 12:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

**Agency** MARION COUNTY SO      **Date** 10/29/2018      **Type** Agency Inspection  
**Inspector** RAKER,JEFFERY,A      **Time** 17:28:22      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 770784  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 04/27/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

WITHIN COMPLIANCE

**Data Download Date/Time** 10/30/2018 16:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MARION COUNTY SO     | <b>Date</b> 11/05/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SHANAHAN, JAKE, L | <b>Time</b> 13:24:11   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0007      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.082  |
|                                             |        | 0.20 g/210L Test # 5             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.082  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.082  |
| 0.20 g/210L Test # 7                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.082  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.082  |
| 0.20 g/210L Test # 9                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.082  |
| 0.20 g/210L Test # 10                  | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0007 |
| 0.08 g/210L Test (Std Dev)             | 0.0006 |
| 0.20 g/210L Test (Std Dev)             | 0.0011 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 11/05/2018 13:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

**Agency** MARION COUNTY SO      **Date** 11/16/2018      **Type** Agency Inspection  
**Inspector** RAKER,JEFFERY,A      **Time** 14:08:14      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 770784  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 04/27/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

WITHIN COMPLIANCE.

**Data Download Date/Time** 11/16/2018 15:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | MARION COUNTY SO | <b>Date</b>       | 12/21/2018 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | RAKER,JEFFERY,A  | <b>Time</b>       | 12:22:38   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201802B          | <b>0.08 Lot #</b> | 201709A    | <b>0.20 Lot #</b> | 201802G           | <b>0.08 Gas Lot #</b> | 770784     |
| <b>Expiration</b>           | 02/06/2020       | <b>Expiration</b> | 09/12/2019 | <b>Expiration</b> | 02/22/2020        | <b>Expiration</b>     | 04/27/2020 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

WITHIN COMPLIANCE

**Data Download Date/Time** 12/21/2018 13:15

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

|                                  |                        |                               |  |
|----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MARION COUNTY SO   | <b>Date</b> 01/14/2019 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> RAKER,JEFFERY,A | <b>Time</b> 09:50:13   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201802B    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201802G    | <b>0.08 Gas Lot #</b> 875088 |
| <b>Expiration</b> 02/06/2020 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/22/2020 | <b>Expiration</b> 12/12/2020 |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

00: Ambient Fail.

**Data Download Date/Time** 01/14/2019 09:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

**Agency** HILLSBOROUGH CO SO      **Date** 01/03/2018      **Type** Agency Inspection  
**Inspector** SKIPPER, ROGER, G      **Time** 19:44:00      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 03417080A1  
**Expiration** 03/08/2018      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 03/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 01/03/2018 19:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

**Agency** HILLSBOROUGH CO SO  
**Inspector** SKIPPER, ROGER, G

**Date** 02/13/2018  
**Time** 19:04:41

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 03417080A1  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 03/05/2019

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 02/14/2018 10:13

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> HILLSBOROUGH CO SO   | <b>Date</b> 03/05/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> SKIPPER, ROGER, G | <b>Time</b> 18:17:01   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 03417080A1 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 03/05/2019     |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 03/07/2018 07:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

**Agency** HILLSBOROUGH CO SO  
**Inspector** SKIPPER, ROGER, G

**Date** 04/05/2018  
**Time** 18:53:43

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 03417080A1  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 03/05/2019

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 04/05/2018 19:28

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

|                                  |                        |                                   |
|----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> HILLSBOROUGH CO SO | <b>Date</b> 04/11/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> BELL,DANIELLE,M | <b>Time</b> 16:26:30   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.051  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

|  |
|--|
|  |
|--|

Data Download Date/Time 04/11/2018 16:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

**Agency** HILLSBOROUGH CO SO      **Date** 05/03/2018      **Type** Agency Inspection  
**Inspector** SKIPPER,ROGER,G      **Time** 17:07:45      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 03417080A1  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 03/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 05/03/2018 17:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> HILLSBOROUGH CO SO   | <b>Date</b> 06/19/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> SKIPPER, ROGER, G | <b>Time</b> 20:14:28   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 31517080A1 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 01/05/2020     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 06/19/2018 20:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

**Agency** HILLSBOROUGH CO SO      **Date** 07/12/2018      **Type** Agency Inspection  
**Inspector** SKIPPER,ROGER,G      **Time** 17:20:28      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 31517080A1  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 01/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 07/12/2018 17:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> HILLSBOROUGH CO SO   | <b>Date</b> 08/07/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> SKIPPER, ROGER, G | <b>Time</b> 14:58:19   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201802B    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201802G    | <b>0.08 Gas Lot #</b> 31517080A1 |
| <b>Expiration</b> 02/06/2020 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/22/2020 | <b>Expiration</b> 01/05/2020     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 4 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

AI MSGT T. BELL #4152 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 08/23/2018 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

**Agency** HILLSBOROUGH CO SO      **Date** 09/06/2018      **Type** Agency Inspection  
**Inspector** SKIPPER, ROGER, G      **Time** 12:44:52      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 31517080A1  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 01/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI MSGT T. BELL #4152 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 09/06/2018 14:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

**Agency** HILLSBOROUGH CO SO      **Date** 10/04/2018      **Type** Agency Inspection  
**Inspector** SKIPPER,ROGER,G      **Time** 11:49:09      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 31517080A1  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 01/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI MSGT T. BELL PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 10/04/2018 11:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> HILLSBOROUGH CO SO   | <b>Date</b> 11/01/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> SKIPPER, ROGER, G | <b>Time</b> 10:37:02   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201802B    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201802G    | <b>0.08 Gas Lot #</b> 31517080A1 |
| <b>Expiration</b> 02/06/2020 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/22/2020 | <b>Expiration</b> 01/05/2020     |

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

AI MSGT T. BELL #4152 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 11/01/2018 11:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

|                                    |                              |                               |                                  |
|------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> HILLSBOROUGH CO SO   | <b>Date</b> 12/18/2018       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> SKIPPER, ROGER, G | <b>Time</b> 19:22:52         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201802B          | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201802G     | <b>0.08 Gas Lot #</b> 31517080A1 |
| <b>Expiration</b> 02/06/2020       | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/22/2020  | <b>Expiration</b> 01/05/2020     |
| <b>Number of Simulators</b> 5      | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: IMPROPER SAMPLE GIVEN. AI J DAIL #254236 PRESENT.

**Data Download Date/Time** 12/18/2018 19:31

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

**Agency** HILLSBOROUGH CO SO      **Date** 01/10/2019      **Type** Agency Inspection  
**Inspector** SKIPPER,ROGER,G      **Time** 19:31:27      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 31517080A1  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 01/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI J DAIL #254236 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 01/10/2019 19:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

|                                       |                        |                               |  |
|---------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MARTIN COUNTY SO        | <b>Date</b> 01/23/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> THOMAS,CHRISTOPHER,L | <b>Time</b> 10:51:33   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201603D    | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 06616080A2 |
| <b>Expiration</b> 03/08/2018 | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 04/05/2018     |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.000  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.047  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.076  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.000  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.047  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.000  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

05: Control Outside Tolerance/BAD SIMULATOR, REPLACED .

**Data Download Date/Time** 01/23/2018 10:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

**Agency** MARTIN COUNTY SO      **Date** 02/02/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 16:51:30      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06616080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 04/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/02/2018 16:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

**Agency** MARTIN COUNTY SO      **Date** 03/15/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 14:32:28      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06616080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 04/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/15/2018 14:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

|                                 |                        |                                   |
|---------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MARTIN COUNTY SO  | <b>Date</b> 03/21/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> PLATT,SHAYLA,D | <b>Time</b> 11:54:35   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805702 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 03/21/2018 11:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

**Agency** MARTIN COUNTY SO      **Date** 04/27/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 09:37:13      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A7  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/27/2018 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

**Agency** MARTIN COUNTY SO      **Date** 05/27/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 16:56:19      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A7  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/27/2018 17:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

**Agency** MARTIN COUNTY SO      **Date** 06/26/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 14:17:37      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A7  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/26/2018 14:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

**Agency** MARTIN COUNTY SO **Date** 07/21/2018 **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L **Time** 01:21:19 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 06717080A7  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 05/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/21/2018 01:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

**Agency** MARTIN COUNTY SO      **Date** 08/12/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 22:04:12      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A7  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/12/2018 22:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

**Agency** MARTIN COUNTY SO      **Date** 09/17/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 15:08:53      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A7  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/17/2018 15:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

**Agency** MARTIN COUNTY SO      **Date** 10/24/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 11:50:22      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A7  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/24/2018 11:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

**Agency** MARTIN COUNTY SO      **Date** 11/29/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 15:32:28      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A7  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/29/2018 15:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

**Agency** MARTIN COUNTY SO      **Date** 12/19/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 22:07:27      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A7  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/19/2018 22:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

**Agency** MARTIN COUNTY SO      **Date** 01/13/2019      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 22:53:19      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 06717080A7  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/13/2019 23:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

**Agency** MARTIN COUNTY SO      **Date** 01/23/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 10:58:51      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A7  
**Expiration** 03/08/2018      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.000  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.049  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.078  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.000  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.048  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.000  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

05: Control Outside Tolerance/BAD SIMULATOR, REPLACED.

**Data Download Date/Time** 01/23/2018 11:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

**Agency** MARTIN COUNTY SO      **Date** 02/02/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 16:47:40      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A7  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/02/2018 16:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

**Agency** MARTIN COUNTY SO      **Date** 03/15/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 14:27:26      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A7  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/15/2018 14:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

**Agency** MARTIN COUNTY SO      **Date** 04/27/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 10:46:03      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A7  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/27/2018 10:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

|                                 |                        |                                   |
|---------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MARTIN COUNTY SO  | <b>Date</b> 05/15/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> PLATT,SHAYLA,D | <b>Time</b> 11:29:27   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805702 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.048  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Test (Std Dev)             | 0.0006 |
| 0.20 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 05/15/2018 11:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

**Agency** MARTIN COUNTY SO      **Date** 05/27/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 18:12:08      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 05218080A4  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 05/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/27/2018 18:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

**Agency** MARTIN COUNTY SO      **Date** 06/26/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 14:29:02      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 05218080A4  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 05/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/26/2018 14:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

**Agency** MARTIN COUNTY SO      **Date** 07/21/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 02:48:54      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 05218080A4  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 05/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/21/2018 02:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

**Agency** MARTIN COUNTY SO      **Date** 08/12/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 22:01:26      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 05218080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/12/2018 22:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

**Agency** MARTIN COUNTY SO      **Date** 09/17/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 15:20:11      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 05218080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/17/2018 15:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

**Agency** MARTIN COUNTY SO      **Date** 10/24/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 11:57:23      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 05218080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/24/2018 12:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

**Agency** MARTIN COUNTY SO      **Date** 11/29/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 16:56:00      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 05218080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.052  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.052  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/29/2018 17:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

**Agency** MARTIN COUNTY SO      **Date** 12/20/2018      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 00:08:38      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 05218080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.052  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/20/2018 00:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000832

**Agency** MARTIN COUNTY SO      **Date** 01/14/2019      **Type** Agency Inspection  
**Inspector** THOMAS,CHRISTOPHER,L      **Time** 00:19:42      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 05218080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 05/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.052  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/14/2019 00:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> HILLSBOROUGH CO SO   | <b>Date</b> 01/03/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> SKIPPER, ROGER, G | <b>Time</b> 18:08:48   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201603D    | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 03417080A1 |
| <b>Expiration</b> 03/08/2018 | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 03/05/2019     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 01/03/2018 18:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> HILLSBOROUGH CO SO   | <b>Date</b> 02/03/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> SKIPPER, ROGER, G | <b>Time</b> 18:59:01   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 03417080A1 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 03/05/2019     |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 02/13/2018 18:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

|                                  |                        |                                   |
|----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> HILLSBOROUGH CO SO | <b>Date</b> 03/05/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> BELL,DANIELLE,M | <b>Time</b> 16:20:37   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG715202 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 06/01/2019   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0006      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.082  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.082  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.082  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.082  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0009 |

#### Remarks:

Data Download Date/Time 03/05/2018 16:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

**Agency** HILLSBOROUGH CO SO      **Date** 04/05/2018      **Type** Agency Inspection  
**Inspector** SKIPPER,ROGER,G      **Time** 17:18:47      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 03417080A1  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 03/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 04/05/2018 17:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> HILLSBOROUGH CO SO   | <b>Date</b> 05/03/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> SKIPPER, ROGER, G | <b>Time</b> 18:27:30   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 03417080A1 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 03/05/2019     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 05/03/2018 18:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

|                                     |                              |                               |                                  |
|-------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> HILLSBOROUGH CO SO    | <b>Date</b> 06/19/2018       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> GRANNIS, RONALD, R | <b>Time</b> 20:24:27         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201703A           | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 03417080A1 |
| <b>Expiration</b> 03/07/2019        | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 03/05/2019     |
| <b>Number of Simulators</b> 5       | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R. SKIPPER #246109 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 06/19/2018 20:37

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

**Agency** HILLSBOROUGH CO SO  
**Inspector** SKIPPER, ROGER, G

**Date** 07/12/2018  
**Time** 17:31:50

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 03417080A1  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 03/05/2019

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R. GRANNIS #256794 PRESENT ENTIRE INSPECTION.

**Data Download Date/Time** 07/12/2018 17:42

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

**Agency** UMATILLA PD(LAKE SO) **Date** 01/04/2018 **Type** Agency Inspection  
**Inspector** PITCHER,CHESSHER,/ **Time** 18:05:42 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 34416080A2  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 02/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/04/2018 18:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

**Agency** UMATILLA PD(LAKE SO) **Date** 02/01/2018 **Type** Agency Inspection  
**Inspector** PITCHER,CHESSHER,/ **Time** 18:54:01 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 34416080A2  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 02/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/01/2018 19:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

|                                     |                        |                               |  |
|-------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> UMATILLA PD(LAKE SO)  | <b>Date</b> 03/09/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> CHESSHER,PITCHER,/ | <b>Time</b> 18:29:40   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 34416080A2 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 02/05/2019     |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.000  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.050  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.080  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.000  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.050  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.000  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

05:SIM NOT CONNECTED. CONNECTED AND REPEATED .

**Data Download Date/Time** 03/09/2018 18:53

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

**Agency** UMATILLA PD(LAKE SO) **Date** 04/07/2018 **Type** Agency Inspection  
**Inspector** PITCHER,CHESSHER,/ **Time** 22:10:58 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | Yes    |      |        |
| Mouth Alcohol Test: Slope Not Met           | Yes    |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

CLEANED BREATH TEST TUBE . Non-compliance: AMB FAIL. BREATH TEST TUBE HAD WATER DROPS .

**Data Download Date/Time** 04/07/2018 23:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

**Agency** UMATILLA PD(LAKE SO) **Date** 04/07/2018 **Type** Agency Inspection  
**Inspector** CHESSHER,PITCHER,/ **Time** 22:54:52 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 34416080A2  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 02/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: Ambient Fail TUBE HAS WATER DROPLETS.

**Data Download Date/Time** 04/07/2018 23:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

**Agency** UMATILLA PD(LAKE SO) **Date** 05/03/2018 **Type** Agency Inspection  
**Inspector** PITCHER,BINDER,/ **Time** 17:43:54 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34416080A2  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/03/2018 17:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

**Agency** UMATILLA PD(LAKE SO) **Date** 06/07/2018 **Type** Agency Inspection  
**Inspector** BINDER,PITCHER,/ **Time** 22:18:15 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34416080A2  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/07/2018 22:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> UMATILLA PD(LAKE SO) | <b>Date</b> 07/06/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> BINDER,PITCHER, / | <b>Time</b> 23:43:11   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 34416080A2 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 02/05/2019     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.050  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.080  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.200  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.000  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.051  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.080  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.080  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3               | INT    |

**Remarks:**

RECONNECTED DRY GAS HOSE.

**Data Download Date/Time** 07/06/2018 23:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

**Agency** UMATILLA PD(LAKE SO) **Date** 08/15/2018 **Type** Agency Inspection  
**Inspector** BINDER,PITCHER,/ **Time** 23:27:09 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34416080A2  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 08/15/2018 23:48

**Printed: 14 January 2019**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> UMATILLA PD(LAKE SO) | <b>Date</b> 09/04/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> BINDER,PITCHER, / | <b>Time</b> 12:52:09   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 34416080A2 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 02/05/2019     |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.049  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.079  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.049  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.000  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

05: Control Outside Tolerance DISCONNECTED ACCIDENTALLY . ACCIDENTALLY DISCONNECTED

**Data Download Date/Time** 09/04/2018 13:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

**Agency** UMATILLA PD(LAKE SO) **Date** 10/18/2018 **Type** Agency Inspection  
**Inspector** PITCHER,BINDER,/ **Time** 20:50:09 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 34416080A2  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/18/2018 20:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

**Agency** UMATILLA PD(LAKE SO) **Date** 11/15/2018 **Type** Agency Inspection  
**Inspector** PITCHER,BINDER,/ **Time** 20:19:34 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 34416080A2  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 02/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/15/2018 20:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

**Agency** UMATILLA PD(LAKE SO) **Date** 12/06/2018 **Type** Agency Inspection  
**Inspector** BINDER,PITCHER,/ **Time** 00:36:54 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 34416080A2  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 02/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/06/2018 00:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000836

**Agency** UMATILLA PD(LAKE SO)      **Date** 01/01/2019      **Type** Agency Inspection  
**Inspector** BINDER,PITCHER,/      **Time** 20:50:32      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 34416080A2  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 02/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/01/2019 20:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

|                                     |                        |                                   |  |
|-------------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> PANAMA CITY BEACH PD  | <b>Date</b> 03/15/2018 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> MURPHY, PATRICK, J | <b>Time</b> 13:27:57   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> 17817080A2 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 08/05/2019     |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0005 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.082  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0007 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 03/15/2018 14:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

**Agency** PANAMA CITY BEACH PD      **Date** 03/28/2018      **Type** Agency Inspection  
**Inspector** MOORE,JACOB,A      **Time** 12:31:44      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 1861309  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/19/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE NOTED

**Data Download Date/Time** 03/28/2018 14:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

**Agency** PANAMA CITY BEACH PD      **Date** 04/30/2018      **Type** Agency Inspection  
**Inspector** MOORE,JACOB,A      **Time** 22:42:01      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration** 03/07/2019      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.000  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1 | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | Interferent Detect Test # 1   | INT    |
| Alcohol Free Subject Test: 0.000            | Yes    | Alcohol Free Test # 2         | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.05 g/210L Test # 2          | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # (Repeat) 2 | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 2   | INT    |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # 3         | 0.000  |
| 0.05 g/210L Test: OK                        | No     | 0.05 g/210L Test # 3          | 0.000  |
| 0.08 g/210L Test: OK                        | No     | 0.05 g/210L Test # (Repeat) 3 | 0.000  |
| 0.20 g/210L Test: OK                        | No     | Interferent Detect Test # 3   | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                               |        |

#### Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 05/01/2018 15:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> PANAMA CITY BEACH PD | <b>Date</b> 05/01/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> MOORE,JACOB,A     | <b>Time</b> 14:46:55   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                               |
|------------------------------|------------------------------|------------------------------|-------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 1861309 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 05/19/2018  |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.195  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.000  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.196  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.000  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time** 05/01/2018 15:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

**Agency** PANAMA CITY BEACH PD      **Date** 06/22/2018      **Type** Agency Inspection  
**Inspector** BARGE,SUSAN,M      **Time** 17:11:45      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 906364  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/07/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NA

**Data Download Date/Time** 06/22/2018 18:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

**Agency** PANAMA CITY BEACH PD      **Date** 06/22/2018      **Type** Agency Inspection  
**Inspector** BARGE,SUSAN,M      **Time** 17:18:13      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators**      **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: Sequence Aborted.

**Data Download Date/Time** 06/22/2018 18:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

**Agency** PANAMA CITY BEACH PD      **Date** 07/06/2018      **Type** Agency Inspection  
**Inspector** BARGE,SUSAN,M      **Time** 17:17:04      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 906364  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/07/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NA

**Data Download Date/Time** 07/06/2018 18:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

**Agency** PANAMA CITY BEACH PD      **Date** 08/03/2018      **Type** Agency Inspection  
**Inspector** BARGE,SUSAN,M      **Time** 11:21:08      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 906364  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/07/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 08/03/2018 12:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

**Agency** PANAMA CITY BEACH PD      **Date** 09/10/2018      **Type** Agency Inspection  
**Inspector** CHERI,DIANE,M      **Time** 13:41:38      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 906364  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 02/07/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 09/10/2018 14:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

**Agency** PANAMA CITY BEACH PD      **Date** 10/02/2018      **Type** Agency Inspection  
**Inspector** CHERI,DIANE,M      **Time** 15:37:41      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 906364  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/07/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 10/02/2018 17:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

**Agency** PANAMA CITY BEACH PD      **Date** 11/01/2018      **Type** Agency Inspection  
**Inspector** CHERI,DIANE,M      **Time** 18:23:48      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 906364  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 02/07/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 11/01/2018 20:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

**Agency** PANAMA CITY BEACH PD      **Date** 12/02/2018      **Type** Agency Inspection  
**Inspector** CHERI,DIANE,M      **Time** 16:50:46      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 906364  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 02/07/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 12/02/2018 18:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

**Agency** PANAMA CITY BEACH PD      **Date** 01/04/2019      **Type** Agency Inspection  
**Inspector** CHERI,DIANE,M      **Time** 08:43:21      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 906364  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 02/07/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A:TOO MUCH MA.

**Data Download Date/Time** 01/04/2019 10:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

|                  |                  |             |            |                 |                   |
|------------------|------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | LYNN HAVEN PD    | <b>Date</b> | 01/04/2018 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | MURPHY,PATRICK,J | <b>Time</b> | 09:54:38   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

SKIPPED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 01/04/2018 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

|                                   |                        |                                   |
|-----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> LYNN HAVEN PD       | <b>Date</b> 01/04/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> MURPHY,PATRICK,J | <b>Time</b> 13:25:34   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG715202 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 06/01/2019   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.083  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.084  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.084  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.051  |
|                                             |        | 0.08 g/210L Test # 4             | 0.084  |
|                                             |        | 0.20 g/210L Test # 4             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.082  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.084  |
|                                             |        | 0.20 g/210L Test # 5             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.084  |
| 0.20 g/210L Test # 6                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.052  |
| 0.08 g/210L Test # 7                   | 0.084  |
| 0.20 g/210L Test # 7                   | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.052  |
| 0.08 g/210L Test # 8                   | 0.084  |
| 0.20 g/210L Test # 8                   | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.052  |
| 0.08 g/210L Test # 9                   | 0.084  |
| 0.20 g/210L Test # 9                   | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.052  |
| 0.08 g/210L Test # 10                  | 0.084  |
| 0.20 g/210L Test # 10                  | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0011 |

#### Remarks:

Data Download Date/Time 01/04/2018 13:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

**Agency** LYNN HAVEN PD      **Date** 01/19/2018      **Type** Agency Inspection  
**Inspector** HAMILTON,PATRICK,T      **Time** 08:48:05      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 09216080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 05/05/2018

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.053  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.052  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.053  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/19/2018 09:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

|                                       |                        |                               |  |
|---------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> LYNN HAVEN PD           | <b>Date</b> 02/26/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> HAMILTON, PATRICK, T | <b>Time</b> 15:57:51   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 09216080A2 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 05/05/2018     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 3 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.192  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.189  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.192  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.191  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

A F / M A: Range Exceeded. 20: Control Outside Tolerance.

**Data Download Date/Time** 03/03/2018 08:15

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

**Agency** LYNN HAVEN PD      **Date** 03/31/2018      **Type** Agency Inspection  
**Inspector** HAMILTON,PATRICK,T      **Time** 08:36:01      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 09216080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 05/05/2018

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.052  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.052  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/31/2018 10:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

**Agency** LYNN HAVEN PD      **Date** 04/29/2018      **Type** Agency Inspection  
**Inspector** HAMILTON,PATRICK,T      **Time** 03:36:27      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 05218080A3  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 05/05/2020

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Time-Date changed.

**Data Download Date/Time** 04/29/2018 22:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

**Agency** LYNN HAVEN PD      **Date** 05/31/2018      **Type** Agency Inspection  
**Inspector** HAMILTON,PATRICK,T      **Time** 03:30:09      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 05218080A3  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 05/05/2020

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.052  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.052  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/31/2018 21:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

|                                       |                        |                               |  |
|---------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> LYNN HAVEN PD           | <b>Date</b> 06/26/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> HAMILTON, PATRICK, T | <b>Time</b> 11:19:46   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 05218080A3 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 05/05/2020     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 3 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.189  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.203  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.202  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time** 06/27/2018 08:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

|                                       |                        |                               |  |
|---------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> LYNN HAVEN PD           | <b>Date</b> 08/09/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> HAMILTON, PATRICK, T | <b>Time</b> 00:23:13   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 05218080A3 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 05/05/2020     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 3 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 08/09/2018 01:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

|                                       |                        |                               |  |
|---------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> LYNN HAVEN PD           | <b>Date</b> 09/12/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> HAMILTON, PATRICK, T | <b>Time</b> 02:08:56   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 05218080A3 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 05/05/2020     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 3 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.086  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.082  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.192  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.085  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.192  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.083  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.191  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

A F / M A: Improper Sample. 08: Control Outside Tolerance.

**Data Download Date/Time** 09/12/2018 03:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

**Agency** CAPE CORAL PD      **Date** 01/10/2018      **Type** Agency Inspection  
**Inspector** CLAFLIN, DAVID, S      **Time** 10:48:13      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201611B      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 29017080A2  
**Expiration** 03/08/2018      **Expiration** 11/15/2018      **Expiration** 04/05/2018      **Expiration** 12/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/10/2018 10:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

**Agency** CAPE CORAL PD      **Date** 02/14/2018      **Type** Agency Inspection  
**Inspector** CLAFLIN, DAVID, S      **Time** 08:59:50      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 29017080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 12/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/14/2018 08:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

**Agency** CAPE CORAL PD  
**Inspector** CLAFLIN, DAVID, S

**Date** 03/07/2018  
**Time** 09:59:20

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 29017080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 12/05/2019

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/07/2018 09:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

|                                 |                        |                                   |  |
|---------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> CAPE CORAL PD     | <b>Date</b> 04/25/2018 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> PLATT,SHAYLA,D | <b>Time</b> 11:46:49   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0006 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1                      | 0.082  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1                      | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.087  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.082  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.081  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.197  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.085  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.081  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.080  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.084  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3               | INT    |
|                                             |        | Alcohol Free Test # 4                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 4                      | 0.048  |
|                                             |        | 0.08 g/210L Test # 4                      | 0.081  |
|                                             |        | 0.20 g/210L Test # 4                      | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4          | 0.082  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 4 | 0.081  |
|                                             |        | Interferent Detect Test # 4               | INT    |
|                                             |        | Alcohol Free Test # 5                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 5                      | 0.048  |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

|                                            |       |
|--------------------------------------------|-------|
| 0.08 g/210L Test # 5                       | 0.080 |
| 0.20 g/210L Test # 5                       | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 5           | 0.083 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 5  | 0.081 |
| Interferent Detect Test # 5                | INT   |
| Alcohol Free Test # 6                      | 0.000 |
| 0.05 g/210L Test # 6                       | 0.048 |
| 0.08 g/210L Test # 6                       | 0.080 |
| 0.20 g/210L Test # 6                       | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6           | 0.082 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 6  | 0.081 |
| Interferent Detect Test # 6                | INT   |
| Alcohol Free Test # 7                      | 0.000 |
| 0.05 g/210L Test # 7                       | 0.049 |
| 0.08 g/210L Test # 7                       | 0.081 |
| 0.20 g/210L Test # 7                       | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7           | 0.082 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 7  | 0.081 |
| Interferent Detect Test # 7                | INT   |
| Alcohol Free Test # 8                      | 0.000 |
| 0.05 g/210L Test # 8                       | 0.049 |
| 0.08 g/210L Test # 8                       | 0.080 |
| 0.20 g/210L Test # 8                       | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8           | 0.082 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 8  | 0.081 |
| Interferent Detect Test # 8                | INT   |
| Alcohol Free Test # 9                      | 0.000 |
| 0.05 g/210L Test # 9                       | 0.049 |
| 0.08 g/210L Test # 9                       | 0.081 |
| 0.20 g/210L Test # 9                       | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9           | 0.081 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 9  | 0.081 |
| Interferent Detect Test # 9                | INT   |
| Alcohol Free Test # 10                     | 0.000 |
| 0.05 g/210L Test # 10                      | 0.049 |
| 0.08 g/210L Test # 10                      | 0.081 |
| 0.20 g/210L Test # 10                      | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10          | 0.081 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 10 | 0.080 |
| Interferent Detect Test # 10               | INT   |

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

|                                                    |        |
|----------------------------------------------------|--------|
| 0.05 g/210L Test (Std Dev)                         | 0.0005 |
| 0.08 g/210L Test (Std Dev)                         | 0.0006 |
| 0.20 g/210L Test (Std Dev)                         | 0.0011 |
| 0.08 g/210L Dry Gas Std Test (Std Dev)             | 0.0019 |
| 0.08 g/210L Dry Gas Std Test (Std Dev)<br>(Repeat) | 0.0005 |

**Remarks:**

08: Control Outside Tolerance.

---

**Data Download Date/Time** 04/25/2018 11:48

**Printed: 14 January 2019**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

**Agency** CAPE CORAL PD  
**Inspector** CLAFLIN, DAVID, S

**Date** 04/27/2018  
**Time** 09:44:01

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 29017080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 12/05/2019

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | No     | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: TOO MUCH M/A.

**Data Download Date/Time** 04/27/2018 09:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

|                                    |                              |                               |                                  |
|------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> CAPE CORAL PD        | <b>Date</b> 05/09/2018       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> CLAFLIN, DAVID, S | <b>Time</b> 11:16:08         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201703A          | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 23016080A3 |
| <b>Expiration</b> 03/07/2019       | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 09/05/2018     |
| <b>Number of Simulators</b> 5      | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/09/2018 11:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

**Agency** CAPE CORAL PD      **Date** 06/06/2018      **Type** Agency Inspection  
**Inspector** CLAFLIN, DAVID, S      **Time** 10:37:34      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 26316080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 11/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/06/2018 10:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

**Agency** CAPE CORAL PD      **Date** 07/18/2018      **Type** Agency Inspection  
**Inspector** CLAFLIN, DAVID, S      **Time** 08:59:59      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 26316080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 11/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/18/2018 09:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

|                                    |                              |                               |                                  |
|------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> CAPE CORAL PD        | <b>Date</b> 08/01/2018       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> CLAFLIN, DAVID, S | <b>Time</b> 09:07:07         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201802B          | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201802G     | <b>0.08 Gas Lot #</b> 26316080A2 |
| <b>Expiration</b> 02/06/2020       | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/20/2020  | <b>Expiration</b> 11/05/2018     |
| <b>Number of Simulators</b> 5      | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 08/01/2018 09:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

**Agency** SARASOTA COUNTY SO      **Date** 01/31/2018      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 08:39:31      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 09216080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 05/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 01/31/2018 08:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

**Agency** SARASOTA COUNTY SO      **Date** 02/23/2018      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 08:49:27      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 09216080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 05/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 02/23/2018 08:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

**Agency** SARASOTA COUNTY SO      **Date** 03/29/2018      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 08:36:33      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 09216080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 05/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 03/29/2018 08:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

**Agency** SARASOTA COUNTY SO      **Date** 04/28/2018      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 07:41:15      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 09216080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 05/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 04/28/2018 07:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | SARASOTA COUNTY SO | <b>Date</b>       | 05/31/2018 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | FELTOVIC,DEP,M     | <b>Time</b>       | 07:39:11   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201703A            | <b>0.08 Lot #</b> | 201611B    | <b>0.20 Lot #</b> | 201702B           | <b>0.08 Gas Lot #</b> | 19817080A2 |
| <b>Expiration</b>           | 03/07/2019         | <b>Expiration</b> | 11/15/2018 | <b>Expiration</b> | 02/23/2019        | <b>Expiration</b>     | 09/05/2019 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 05/31/2018 07:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

**Agency** SARASOTA COUNTY SO      **Date** 06/29/2018      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 08:01:34      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 19817080A2  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 09/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 06/29/2018 08:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

|                                  |                        |                                   |
|----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> SARASOTA COUNTY SO | <b>Date</b> 07/05/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> PLATT,SHAYLA,D  | <b>Time</b> 10:40:11   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0006 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 07/05/2018 10:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

**Agency** SARASOTA COUNTY SO      **Date** 07/30/2018      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 15:50:20      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 10818080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 06/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 07/30/2018 15:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

|                                  |                        |                               |
|----------------------------------|------------------------|-------------------------------|
| <b>Agency</b> SARASOTA COUNTY SO | <b>Date</b> 08/21/2018 | <b>Type</b> Agency Inspection |
| <b>Inspector</b> FELTOVIC,DEP,M  | <b>Time</b> 07:43:25   | <b>Software</b> 8100.27       |

|                              |                              |                              |                       |
|------------------------------|------------------------------|------------------------------|-----------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b>     |

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1          | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1          | 0.187  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1 | 0.188  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1   | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2         | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.05 g/210L Test # 2          | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2          | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2          | 0.188  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2 | 0.188  |
| 0.20 g/210L Test: OK                        | No     | Interferent Detect Test # 2   | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | Alcohol Free Test # 3         | 0.000  |
|                                             |        | 0.05 g/210L Test # 3          | 0.049  |
|                                             |        | 0.08 g/210L Test # 3          | 0.078  |
|                                             |        | 0.20 g/210L Test # 3          | 0.188  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3 | 0.188  |
|                                             |        | Interferent Detect Test # 3   | INT    |

**Remarks:**

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: CONTROL OUTSIDE TOL.

**Data Download Date/Time** 08/21/2018 07:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | SARASOTA COUNTY SO | <b>Date</b>       | 08/21/2018 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | FELTOVIC,DEP,M     | <b>Time</b>       | 08:37:32   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201703A            | <b>0.08 Lot #</b> | 201709A    | <b>0.20 Lot #</b> | 201702B           | <b>0.08 Gas Lot #</b> | 10818080A4 |
| <b>Expiration</b>           | 03/07/2019         | <b>Expiration</b> | 09/12/2019 | <b>Expiration</b> | 02/23/2019        | <b>Expiration</b>     | 06/05/2020 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 08/21/2018 08:44

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

**Agency** SARASOTA COUNTY SO      **Date** 09/30/2018      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 08:33:38      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 10818080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 06/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 11/01/2018 11:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

|                  |                    |             |            |                 |                   |
|------------------|--------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | SARASOTA COUNTY SO | <b>Date</b> | 12/04/2018 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | PLATT,SHAYLA,D     | <b>Time</b> | 10:22:16   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 12/04/2018 13:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

|                                  |                        |                                   |
|----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> SARASOTA COUNTY SO | <b>Date</b> 12/04/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> PLATT,SHAYLA,D  | <b>Time</b> 12:37:49   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.048  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.048  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0009 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 12/04/2018 13:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

|                                  |                        |                               |  |
|----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> SARASOTA COUNTY SO | <b>Date</b> 12/28/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> FELTOVIC,DEP,M  | <b>Time</b> 09:47:56   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201802B    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201802G    | <b>0.08 Gas Lot #</b> 19817080A2 |
| <b>Expiration</b> 02/06/2020 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/22/2020 | <b>Expiration</b> 09/05/2019     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

INSPECTED AT N JAIL

**Data Download Date/Time** 12/28/2018 09:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

|                                   |                        |                                   |
|-----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> BLOUNTSTOWN PD      | <b>Date</b> 03/09/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> MURPHY,PATRICK,J | <b>Time</b> 14:20:43   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG715202 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 06/01/2019   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.082  |
| 0.20 g/210L Test # 8                   | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0007 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 03/09/2018 14:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

**Agency** BLOUNTSTOWN PD      **Date** 03/20/2018      **Type** Agency Inspection  
**Inspector** BARGE,SUSAN,M      **Time** 02:07:34      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 770784  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 07/23/2019      **Expiration** 04/27/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NA

**Data Download Date/Time** 03/20/2018 03:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

**Agency** BLOUNTSTOWN PD      **Date** 04/02/2018      **Type** Agency Inspection  
**Inspector** BOWMAN,JOANNA,N      **Time** 21:21:37      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 770784  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 04/27/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 04/02/2018 23:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

**Agency** BLOUNTSTOWN PD      **Date** 05/08/2018      **Type** Agency Inspection  
**Inspector** BARGE,SUSAN,M      **Time** 13:48:20      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 770784  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 04/27/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NA

**Data Download Date/Time** 05/08/2018 15:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

**Agency** BLOUNTSTOWN PD      **Date** 06/17/2018      **Type** Agency Inspection  
**Inspector** BOWMAN,JOANNA,N      **Time** 23:30:20      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 770784  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 04/27/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 06/18/2018 00:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

**Agency** BLOUNTSTOWN PD      **Date** 07/06/2018      **Type** Agency Inspection  
**Inspector** BARGE,SUSAN,M      **Time** 17:27:00      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 770784  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 04/27/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NA

**Data Download Date/Time** 07/06/2018 18:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

**Agency** BLOUNTSTOWN PD      **Date** 08/22/2018      **Type** Agency Inspection  
**Inspector** BARGE,SUSAN,M      **Time** 11:46:56      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 770784  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 04/27/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NA

**Data Download Date/Time** 08/22/2018 14:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

**Agency** BLOUNTSTOWN PD      **Date** 09/14/2018      **Type** Agency Inspection  
**Inspector** BARGE,SUSAN,M      **Time** 11:01:16      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 770784  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 04/27/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NA

**Data Download Date/Time** 09/14/2018 12:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000862

**Agency** BLOUNTSTOWN PD      **Date** 10/08/2018      **Type** Agency Inspection  
**Inspector** HOLLIDAY,DAISY,J      **Time** 15:46:48      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 770784  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 04/27/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/08/2018 18:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

**Agency** MONROE COUNTY SO      **Date** 01/10/2018      **Type** Agency Inspection  
**Inspector** KEITH,JASON,T      **Time** 12:46:49      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 686331  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 01/31/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

KW INTOX

**Data Download Date/Time** 01/10/2018 13:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

Agency MONROE COUNTY SO Date 02/01/2018 Type Agency Inspection  
Inspector KEITH,JASON,T Time 11:15:50 Software 8100.27

0.05 Lot # 201703A 0.08 Lot # 201611B 0.20 Lot # 201702B 0.08 Gas Lot # 686331  
Expiration 03/07/2019 Expiration 11/15/2018 Expiration 02/23/2019 Expiration 01/31/2020

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

KW INTOX

Data Download Date/Time 02/01/2018 11:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

**Agency** MONROE COUNTY SO      **Date** 03/01/2018      **Type** Agency Inspection  
**Inspector** KEITH,JASON,T      **Time** 11:49:43      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 686331  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 01/31/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

KW INTOX. PRE FDLE DEPT INSPECTION DONE

**Data Download Date/Time** 03/01/2018 11:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MONROE COUNTY SO          | <b>Date</b> 03/07/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 09:57:56   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG626605 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 09/22/2018   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.082  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.082  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 03/07/2018 10:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 04/04/2018  
**Time** 10:33:44

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 686331  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 01/31/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE

**Data Download Date/Time** 04/06/2018 11:22

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

**Agency** MONROE COUNTY SO      **Date** 05/06/2018      **Type** Agency Inspection  
**Inspector** KEITH,JASON,T      **Time** 15:03:21      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 686331  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 01/31/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE

**Data Download Date/Time** 05/06/2018 15:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 06/06/2018  
**Time** 13:45:54

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 823252  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 10/05/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX

**Data Download Date/Time** 06/06/2018 14:01

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 07/12/2018  
**Time** 14:00:03

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 686331  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 01/31/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX

**Data Download Date/Time** 07/12/2018 14:15

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 08/02/2018  
**Time** 16:06:01

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 686331  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 01/31/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX

**Data Download Date/Time** 08/02/2018 16:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

**Agency** MONROE COUNTY SO      **Date** 09/06/2018      **Type** Agency Inspection  
**Inspector** KEITH,JASON,T      **Time** 14:31:40      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 686331  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 01/31/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX

**Data Download Date/Time** 09/06/2018 14:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

**Agency** MONROE COUNTY SO      **Date** 10/04/2018      **Type** Agency Inspection  
**Inspector** KEITH,JASON,T      **Time** 08:41:58      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 686331  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 01/31/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX

**Data Download Date/Time** 10/04/2018 08:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 11/01/2018  
**Time** 08:43:39

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 823252  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 10/05/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX

**Data Download Date/Time** 11/01/2018 13:53

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 12/13/2018  
**Time** 14:04:06

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 823252  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 10/05/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX

**Data Download Date/Time** 12/13/2018 14:20

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 01/02/2019  
**Time** 11:00:55

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 823252  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 10/05/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE. PRE-EVIDENTIARY AL COMPLETED.

**Data Download Date/Time** 01/02/2019 11:08

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

|                                |                        |                               |  |
|--------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MONROE COUNTY SO | <b>Date</b> 01/10/2019 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> KEITH,JASON,T | <b>Time</b> 13:28:35   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201802B    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201802G    | <b>0.08 Gas Lot #</b> 915310 |
| <b>Expiration</b> 02/06/2020 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/22/2020 | <b>Expiration</b> 03/07/2021 |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.058  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.053  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.083  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.203  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.056  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.053  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.055  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

05: Control Outside Tolerance.

**Data Download Date/Time** 01/10/2019 13:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MONROE COUNTY SO          | <b>Date</b> 02/22/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 12:43:05   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG626605 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 09/22/2018   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0007 |
| 0.08 g/210L Test (Std Dev)             | 0.0006 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 02/22/2018 12:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 04/04/2018  
**Time** 10:55:12

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 686331  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 01/31/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE

**Data Download Date/Time** 04/06/2018 11:20

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 05/06/2018  
**Time** 15:13:05

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 686331  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 01/31/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE

**Data Download Date/Time** 05/06/2018 15:23

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 06/06/2018  
**Time** 13:51:20

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 823252  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 10/05/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX

**Data Download Date/Time** 06/06/2018 13:57

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 07/12/2018  
**Time** 14:04:15

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 686331  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 01/31/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX

**Data Download Date/Time** 07/12/2018 14:20

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 08/02/2018  
**Time** 16:10:21

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 686331  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 01/31/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX

**Data Download Date/Time** 08/02/2018 16:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

Agency MONROE COUNTY SO Date 09/06/2018 Type Agency Inspection  
Inspector KEITH,JASON,T Time 14:42:44 Software 8100.27

0.05 Lot # 201703A 0.08 Lot # 201709A 0.20 Lot # 201702B 0.08 Gas Lot # 686331  
Expiration 03/07/2019 Expiration 09/12/2019 Expiration 02/23/2019 Expiration 01/31/2020

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX

Data Download Date/Time 09/06/2018 14:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 09/27/2018  
**Time** 20:18:22

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 823252  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 10/05/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX - PUT IN SERVICE IN KW

**Data Download Date/Time** 09/27/2018 20:21

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 10/04/2018  
**Time** 13:42:02

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 915310  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 03/07/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

KW INTOX

**Data Download Date/Time** 10/04/2018 13:47

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

**Agency** MONROE COUNTY SO **Date** 10/12/2018 **Type** Agency Inspection  
**Inspector** KEITH,JASON,T **Time** 16:14:55 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 823252  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 10/05/2020

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NOW SPARE. POST-EVIDENTIARY USE AI DONE.

**Data Download Date/Time** 10/12/2018 16:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 11/01/2018  
**Time** 08:37:34

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 823252  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 10/05/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX

**Data Download Date/Time** 11/01/2018 13:50

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

**Agency** MONROE COUNTY SO **Date** 12/13/2018 **Type** Agency Inspection  
**Inspector** KEITH,JASON,T **Time** 14:07:33 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 823252  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 10/05/2020

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX

**Data Download Date/Time** 12/13/2018 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

|                                |                        |                               |  |
|--------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MONROE COUNTY SO | <b>Date</b> 01/02/2019 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> KEITH,JASON,T | <b>Time</b> 14:10:59   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201802B    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201802G    | <b>0.08 Gas Lot #</b> 823252 |
| <b>Expiration</b> 02/06/2020 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/22/2020 | <b>Expiration</b> 10/05/2020 |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

SPARE. PRE-EVIDENTIARY AI COMPLETED. KV INTOX.

**Data Download Date/Time** 01/02/2019 14:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

**Agency** MONROE COUNTY SO      **Date** 01/09/2019      **Type** Agency Inspection  
**Inspector** KEITH,JASON,T      **Time** 10:50:40      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 823252  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 10/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

KV. SPARE. POST-EVIDENTIARY AI COMPLETED.

**Data Download Date/Time** 01/09/2019 11:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

|                                |                        |                               |  |
|--------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MONROE COUNTY SO | <b>Date</b> 01/18/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> KEITH,JASON,T | <b>Time</b> 13:01:46   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 686331 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 01/31/2020 |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.205  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.051  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.082  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.204  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.000  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.051  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.082  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.000  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3               | INT    |

**Remarks:**

08: Control Outside Tolerance. FORGOT TO ATTACH GAS TUBE

**Data Download Date/Time** 01/18/2018 13:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MONROE COUNTY SO          | <b>Date</b> 01/24/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 10:24:41   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG626604 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 09/22/2018   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.082  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.082  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.082  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.082  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.082  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.082  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0010 |

#### Remarks:

Data Download Date/Time 01/24/2018 10:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

**Agency** MONROE COUNTY SO      **Date** 02/01/2018      **Type** Agency Inspection  
**Inspector** KEITH,JASON,T      **Time** 16:22:15      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 686331  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 01/31/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX

**Data Download Date/Time** 02/01/2018 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

|                                |                              |                               |                              |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> MONROE COUNTY SO | <b>Date</b> 03/01/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> KEITH,JASON,T | <b>Time</b> 16:49:34         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201703A      | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 686331 |
| <b>Expiration</b> 03/07/2019   | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 01/31/2020 |
| <b>Number of Simulators</b> 5  | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

PUT INTO SERVICE IN KV DUE TO 80-001254 FDLE DEPT INSP

**Data Download Date/Time** 03/01/2018 16:52

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 04/04/2018  
**Time** 10:19:18

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 686331  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 01/31/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

KV INTOX

**Data Download Date/Time** 04/06/2018 11:12

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

**Agency** MONROE COUNTY SO      **Date** 05/06/2018      **Type** Agency Inspection  
**Inspector** KEITH,JASON,T      **Time** 14:52:29      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 686331  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 01/31/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

KV INTOX

**Data Download Date/Time** 05/06/2018 15:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

**Agency** MONROE COUNTY SO      **Date** 06/06/2018      **Type** Agency Inspection  
**Inspector** KEITH,JASON,T      **Time** 13:37:09      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 823252  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 10/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

KV INTOX

**Data Download Date/Time** 06/06/2018 14:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

**Agency** MONROE COUNTY SO **Date** 07/12/2018 **Type** Agency Inspection  
**Inspector** KEITH,JASON,T **Time** 13:51:01 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 823252  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/05/2020

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

KV INTOX

**Data Download Date/Time** 07/12/2018 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

**Agency** MONROE COUNTY SO      **Date** 08/02/2018      **Type** Agency Inspection  
**Inspector** KEITH,JASON,T      **Time** 15:59:07      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 823252  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 10/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

WAITED TOO LONG TO PROVIDE MA SAMPLE

**Data Download Date/Time** 08/02/2018 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

**Agency** MONROE COUNTY SO **Date** 09/06/2018 **Type** Agency Inspection  
**Inspector** KEITH,JASON,T **Time** 14:25:20 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 823252  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/05/2020

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

KV INTOX

**Data Download Date/Time** 09/06/2018 14:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

Agency MONROE COUNTY SO Date 10/04/2018 Type Agency Inspection  
Inspector KEITH,JASON,T Time 08:34:37 Software 8100.27

0.05 Lot # 201703A 0.08 Lot # 201709A 0.20 Lot # 201702B 0.08 Gas Lot # 823252  
Expiration 03/07/2019 Expiration 09/12/2019 Expiration 02/23/2019 Expiration 10/05/2020

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

KV INTOX

Data Download Date/Time 10/04/2018 08:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 11/01/2018  
**Time** 08:30:28

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 823252  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 10/05/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

KV INTOX

**Data Download Date/Time** 11/01/2018 13:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

Agency MONROE COUNTY SO Date 12/13/2018 Type Agency Inspection  
Inspector KEITH,JASON,T Time 13:55:43 Software 8100.27

0.05 Lot # 201802B 0.08 Lot # 201709A 0.20 Lot # 201802G 0.08 Gas Lot # 823252  
Expiration 02/06/2020 Expiration 09/12/2019 Expiration 02/22/2020 Expiration 10/05/2020

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

KV INTOX

Data Download Date/Time 12/13/2018 14:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

|                                |                        |                               |  |
|--------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MONROE COUNTY SO | <b>Date</b> 01/02/2019 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> KEITH,JASON,T | <b>Time</b> 14:03:00   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201802B    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201802G    | <b>0.08 Gas Lot #</b> 823252 |
| <b>Expiration</b> 02/06/2020 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/22/2020 | <b>Expiration</b> 10/05/2020 |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

KV. POST-EVIDENTIARY AI COMPLETED.

**Data Download Date/Time** 01/02/2019 15:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MONROE COUNTY SO          | <b>Date</b> 01/03/2019 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 09:12:40   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.083  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.082  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.083  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.083  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.083  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.082  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.082  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0008 |

#### Remarks:

Data Download Date/Time 01/03/2019 09:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

**Agency** MONROE COUNTY SO **Date** 01/09/2019 **Type** Agency Inspection  
**Inspector** KEITH,JASON,T **Time** 11:16:14 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 823252  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 10/05/2020

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

KV INTOX. POST-FDLE DI AI COMPLETED.

**Data Download Date/Time** 01/09/2019 11:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

**Agency** MONROE COUNTY SO      **Date** 01/18/2018      **Type** Agency Inspection  
**Inspector** KEITH,JASON,T      **Time** 13:07:24      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 686331  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 01/31/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.083  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.207  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.052  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.083  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.206  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.083  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX

**Data Download Date/Time** 01/18/2018 13:08

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MONROE COUNTY SO          | <b>Date</b> 01/24/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 10:35:35   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG626604 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 09/22/2018   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.204  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.205  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.205  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.082  |
|                                             |        | 0.20 g/210L Test # 4             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.082  |
| 0.20 g/210L Test # 6                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.082  |
| 0.20 g/210L Test # 7                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.083  |
| 0.20 g/210L Test # 8                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.082  |
| 0.20 g/210L Test # 9                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.083  |
| 0.20 g/210L Test # 10                  | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0008 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 01/24/2018 10:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 02/01/2018  
**Time** 16:17:15

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 686331  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 01/31/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.083  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.209  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.208  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.209  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX

**Data Download Date/Time** 02/01/2018 16:18

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

|                                |                        |                               |  |
|--------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MONROE COUNTY SO | <b>Date</b> 03/01/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> KEITH,JASON,T | <b>Time</b> 16:42:22   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 686331 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 01/31/2020 |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.052  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.206  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.052  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.083  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.206  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.205  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

PUT INTO SERVICE IN PK DUE TO 80-006693 GOING FOR DEPT INSP

**Data Download Date/Time** 03/01/2018 16:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

**Agency** MONROE COUNTY SO      **Date** 04/04/2018      **Type** Agency Inspection  
**Inspector** KEITH,JASON,T      **Time** 14:49:58      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 686331  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 01/31/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.204  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.207  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.208  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Time-Date changed. TIME BEHIND BY 3 MINS. PK INTOX

**Data Download Date/Time** 04/04/2018 14:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 05/04/2018  
**Time** 14:04:01

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 823252  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 10/05/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.206  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.207  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.083  |
|                                             |        | 0.20 g/210L Test # 3             | 0.206  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

PK INTOX

**Data Download Date/Time** 05/04/2018 14:12

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

**Agency** MONROE COUNTY SO      **Date** 06/08/2018      **Type** Agency Inspection  
**Inspector** KEITH,JASON,T      **Time** 13:55:14      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 823252  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 10/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.206  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.205  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.207  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

PK INTOX

**Data Download Date/Time** 06/08/2018 14:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 07/13/2018  
**Time** 13:08:51

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 823252  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 10/05/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.206  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

PK INTOX

**Data Download Date/Time** 07/13/2018 13:25

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

**Agency** MONROE COUNTY SO      **Date** 08/01/2018      **Type** Agency Inspection  
**Inspector** KEITH,JASON,T      **Time** 14:32:55      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 823252  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 10/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.205  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.206  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.207  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

PK INTOX

**Data Download Date/Time** 08/01/2018 14:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

**Agency** MONROE COUNTY SO **Date** 09/05/2018 **Type** Agency Inspection  
**Inspector** KEITH,JASON,T **Time** 15:23:33 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 823252  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 10/05/2020

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.204  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

PK INTOX

**Data Download Date/Time** 09/05/2018 15:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

**Agency** MONROE COUNTY SO      **Date** 10/03/2018      **Type** Agency Inspection  
**Inspector** KEITH,JASON,T      **Time** 14:28:21      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 823252  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 10/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.205  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

WAS PK INTOX - NOW SPARE INTOX

**Data Download Date/Time** 10/03/2018 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 11/07/2018  
**Time** 13:58:56

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 823252  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 10/05/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.206  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.205  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX

**Data Download Date/Time** 11/07/2018 14:04

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

Agency MONROE COUNTY SO Date 12/19/2018 Type Agency Inspection  
Inspector KEITH,JASON,T Time 12:30:25 Software 8100.27

0.05 Lot # 201802B 0.08 Lot # 201808E 0.20 Lot # 201802G 0.08 Gas Lot # 823252  
Expiration 02/06/2020 Expiration 08/22/2020 Expiration 02/22/2020 Expiration 10/05/2020

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.206  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE INTOX

Data Download Date/Time 12/19/2018 12:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

**Agency** MONROE COUNTY SO  
**Inspector** KEITH,JASON,T

**Date** 01/02/2019  
**Time** 17:05:46

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201808E      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 1012059  
**Expiration** 02/06/2020      **Expiration** 08/22/2020      **Expiration** 02/22/2020      **Expiration** 09/05/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.205  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.205  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.206  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

SPARE. PRE-FDLE DI AI COMPLETED.

**Data Download Date/Time** 01/02/2019 17:13

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

|                             |                 |                   |            |                   |                   |                       |            |
|-----------------------------|-----------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | CORAL GABLES PD | <b>Date</b>       | 01/29/2018 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SMITH,DANIEL,A  | <b>Time</b>       | 13:02:07   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201703A         | <b>0.08 Lot #</b> | 201709A    | <b>0.20 Lot #</b> | 201702B           | <b>0.08 Gas Lot #</b> | 26316080AI |
| <b>Expiration</b>           | 03/07/2019      | <b>Expiration</b> | 09/12/2019 | <b>Expiration</b> | 02/23/2019        | <b>Expiration</b>     | 11/05/2018 |
| <b>Number of Simulators</b> | 4               | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/30/2018 14:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

|                             |                 |                   |            |                   |                   |                       |            |
|-----------------------------|-----------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | CORAL GABLES PD | <b>Date</b>       | 02/23/2018 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | SMITH,DANIEL,A  | <b>Time</b>       | 12:26:32   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201703A         | <b>0.08 Lot #</b> | 201709A    | <b>0.20 Lot #</b> | 201702B           | <b>0.08 Gas Lot #</b> | 26316080AI |
| <b>Expiration</b>           | 03/07/2019      | <b>Expiration</b> | 09/12/2019 | <b>Expiration</b> | 02/23/2019        | <b>Expiration</b>     | 11/05/2018 |
| <b>Number of Simulators</b> | 4               | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: Ambient Fail.

**Data Download Date/Time** 02/23/2018 12:30

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

**Agency** CORAL GABLES PD  
**Inspector** SMITH,DANIEL,A

**Date** 03/23/2018  
**Time** 12:58:37

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 26316080AI  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 11/05/2018

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: Ambient Fail.

**Data Download Date/Time** 03/23/2018 13:02

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> CORAL GABLES PD           | <b>Date</b> 09/04/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 10:47:21   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.048  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 09/04/2018 11:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

|                                 |                        |                               |
|---------------------------------|------------------------|-------------------------------|
| <b>Agency</b> CORAL GABLES PD   | <b>Date</b> 09/07/2018 | <b>Type</b> Agency Inspection |
| <b>Inspector</b> SMITH,DANIEL,A | <b>Time</b> 10:20:56   | <b>Software</b> 8100.27       |

|                              |                              |                   |                       |
|------------------------------|------------------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> | <b>Expiration</b>     |

**Number of Simulators** 4 **Compliance** No

| Check or Test                               | Result | Test                           | Result |
|---------------------------------------------|--------|--------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1 | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1           | 0.044  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1  | 0.045  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1           | 0.073  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Test # (Repeat) 1  | 0.051  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1    | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2 | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.05 g/210L Test # 2           | 0.045  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # (Repeat) 2  | 0.045  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2           | 0.073  |
| 0.08 g/210L Test: OK                        | No     | 0.08 g/210L Test # (Repeat) 2  | 0.048  |
| 0.20 g/210L Test: OK                        | No     | Interferent Detect Test # 2    | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | Alcohol Free Test # (Repeat) 3 | 0.000  |
|                                             |        | 0.05 g/210L Test # 3           | 0.045  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3  | 0.046  |
|                                             |        | 0.08 g/210L Test # 3           | 0.073  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3  | 0.052  |
|                                             |        | Interferent Detect Test # 3    | INT    |

**Remarks:**

00: Ambient Fail. 05: Control Outside Tolerance. 08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 09/07/2018 10:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> CORAL GABLES PD           | <b>Date</b> 09/12/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 12:56:37   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.048  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0010 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 09/12/2018 13:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

**Agency** CORAL GABLES PD      **Date** 09/14/2018      **Type** Agency Inspection  
**Inspector** SMITH,DANIEL,A      **Time** 11:18:18      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration** 03/07/2019      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 4      **Compliance** No

| Check or Test                               | Result | Test                           | Result |
|---------------------------------------------|--------|--------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1 | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1           | 0.038  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1  | 0.035  |
| Barometric Pressure Sensor Check: OK        |        | Interferent Detect Test # 1    | INT    |
| Alcohol Free Subject Test: 0.000            | Yes    | Alcohol Free Test # (Repeat) 2 | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.05 g/210L Test # 2           | 0.040  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # (Repeat) 2  | 0.035  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 2    | INT    |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # (Repeat) 3 | 0.000  |
| 0.05 g/210L Test: OK                        | No     | 0.05 g/210L Test # 3           | 0.035  |
| 0.08 g/210L Test: OK                        | No     | 0.05 g/210L Test # (Repeat) 3  | 0.035  |
| 0.20 g/210L Test: OK                        | No     | Interferent Detect Test # 3    | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                                |        |

#### Remarks:

00: Ambient Fail. 05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 09/14/2018 11:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

**Agency** CORAL GABLES PD      **Date** 12/27/2018      **Type** Agency Inspection  
**Inspector** PASTOR,JOSE,F      **Time** 14:39:15      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 620051  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 11/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/27/2018 15:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MIAMI PD             | <b>Date</b> 01/30/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> REYES, BARBARA, T | <b>Time</b> 13:12:38   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 640976 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 11/21/2019 |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.205  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

A F / M A: LACK OF MOUTH ALCOHOL SOLUTION.

**Data Download Date/Time** 01/30/2018 14:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency MIAMI PD Date 02/27/2018 Type Agency Inspection  
Inspector PASTOR,JOSE,F Time 11:58:38 Software 8100.27

0.05 Lot # 201703A 0.08 Lot # 201611B 0.20 Lot # 201702B 0.08 Gas Lot # 640976  
Expiration 03/07/2019 Expiration 11/15/2018 Expiration 02/23/2019 Expiration 11/21/2019

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.205  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.206  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.207  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Data Download Date/Time 02/27/2018 13:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

**Agency** MIAMI PD **Date** 03/27/2018 **Type** Agency Inspection  
**Inspector** PASTOR,JOSE,F **Time** 13:23:35 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 640976  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 11/21/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.204  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.205  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/27/2018 14:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

|                                       |                        |                                   |
|---------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MIAMI PD                | <b>Date</b> 04/04/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA,DAVID,E | <b>Time</b> 10:01:10   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG715202 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 06/01/2019   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.082  |
|                                             |        | 0.20 g/210L Test # 4             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.076  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.082  |
|                                             |        | 0.20 g/210L Test # 5             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.076  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.076  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.082  |
| 0.20 g/210L Test # 7                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.076  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.082  |
| 0.20 g/210L Test # 8                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.076  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.082  |
| 0.20 g/210L Test # 9                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.076  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.082  |
| 0.20 g/210L Test # 10                  | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.076  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0006 |
| 0.20 g/210L Test (Std Dev)             | 0.0008 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 04/04/2018 10:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

**Agency** MIAMI PD **Date** 04/26/2018 **Type** Agency Inspection  
**Inspector** PASTOR,JOSE,F **Time** 20:06:14 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 640976  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 11/21/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/26/2018 20:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

**Agency** MIAMI PD **Date** 05/14/2018 **Type** Agency Inspection  
**Inspector** FERNANDEZ,ANGEL, **Time** 15:27:11 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 640976 **0.08 Gas Lot #** 640976  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 11/21/2019 **Expiration** 11/21/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/14/2018 15:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MIAMI PD             | <b>Date</b> 06/28/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> FERNANDEZ, ANGEL, | <b>Time</b> 10:28:45   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 640976 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 11/21/2019 |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.184  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.203  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.202  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

20:CONNETORS WERE LOOSE READJUSTED & REPEAT TEST..

**Data Download Date/Time** 06/28/2018 11:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

|                                |                              |                               |                              |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> MIAMI PD         | <b>Date</b> 07/30/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> PASTOR,JOSE,F | <b>Time</b> 18:08:32         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201703A      | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 640976 |
| <b>Expiration</b> 03/07/2019   | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 11/21/2019 |
| <b>Number of Simulators</b> 5  | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.084  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/30/2018 18:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

**Agency** MIAMI PD **Date** 08/29/2018 **Type** Agency Inspection  
**Inspector** PASTOR,JOSE,F **Time** 20:17:52 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 640976  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 11/21/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/29/2018 20:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

**Agency** MIAMI PD **Date** 09/30/2018 **Type** Agency Inspection  
**Inspector** PASTOR,JOSE,F **Time** 21:16:10 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 640976  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 11/21/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/30/2018 21:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency MIAMI PD Date 10/30/2018 Type Agency Inspection  
Inspector PASTOR,JOSE,F Time 16:55:02 Software 8100.27

0.05 Lot # 201802B 0.08 Lot # 201611B 0.20 Lot # 201702B 0.08 Gas Lot # 640976  
Expiration 02/06/2020 Expiration 11/15/2018 Expiration 02/23/2019 Expiration 11/21/2019

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.204  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Data Download Date/Time 10/30/2018 17:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

Agency MIAMI PD Date 11/27/2018 Type Agency Inspection  
Inspector PASTOR,JOSE,F Time 12:59:41 Software 8100.27

0.05 Lot # 201802B 0.08 Lot # 201709A 0.20 Lot # 201702B 0.08 Gas Lot # 640976  
Expiration 02/06/2020 Expiration 09/12/2019 Expiration 02/23/2019 Expiration 11/21/2019

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Data Download Date/Time 11/27/2018 13:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

**Agency** MIAMI PD **Date** 12/27/2018 **Type** Agency Inspection  
**Inspector** PASTOR,JOSE,F **Time** 14:12:24 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 640976  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 11/21/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: Range Exceeded.

**Data Download Date/Time** 12/27/2018 15:20

**Printed: 14 January 2019**  
COBRA ©1997-2014 CMI, Inc.



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

|                                |                        |                               |
|--------------------------------|------------------------|-------------------------------|
| <b>Agency</b> MIAMI PD         | <b>Date</b> 02/27/2018 | <b>Type</b> Agency Inspection |
| <b>Inspector</b> PASTOR,JOSE,F | <b>Time</b> 10:42:57   | <b>Software</b> 8100.27       |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

FOROT ALCOHOL.

**Data Download Date/Time** 02/27/2018 12:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

|                                |                              |                               |                              |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> MIAMI PD         | <b>Date</b> 02/27/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> PASTOR,JOSE,F | <b>Time</b> 12:07:25         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201703A      | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 203498 |
| <b>Expiration</b> 03/07/2019   | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 12/10/2018 |
| <b>Number of Simulators</b> 5  | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 02/27/2018 12:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

|                                |                              |                               |                              |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> MIAMI PD         | <b>Date</b> 03/27/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> PASTOR,JOSE,F | <b>Time</b> 13:04:28         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201703A      | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 203498 |
| <b>Expiration</b> 03/07/2019   | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 12/10/2018 |
| <b>Number of Simulators</b> 5  | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/27/2018 14:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

|                                |                              |                               |                              |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> MIAMI PD         | <b>Date</b> 04/26/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> PASTOR,JOSE,F | <b>Time</b> 19:33:47         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201703A      | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 203498 |
| <b>Expiration</b> 03/07/2019   | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 12/10/2018 |
| <b>Number of Simulators</b> 5  | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 04/26/2018 19:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

**Agency** MIAMI PD      **Date** 05/14/2018      **Type** Agency Inspection  
**Inspector** FERNANDEZ,ANGEL,5      **Time** 14:59:22      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 203498      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 203498  
**Expiration** 03/07/2019      **Expiration** 12/10/2018      **Expiration** 02/23/2019      **Expiration** 12/10/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/14/2018 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

|                                   |                              |                               |                              |
|-----------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> MIAMI PD            | <b>Date</b> 06/28/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> FERNANDEZ,ANGEL, | <b>Time</b> 10:50:35         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201703A         | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 203498 |
| <b>Expiration</b> 03/07/2019      | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 12/10/2018 |
| <b>Number of Simulators</b> 5     | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/28/2018 13:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

|                                |                              |                               |                              |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> MIAMI PD         | <b>Date</b> 07/30/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> PASTOR,JOSE,F | <b>Time</b> 17:48:27         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201703A      | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 203498 |
| <b>Expiration</b> 03/07/2019   | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 12/10/2018 |
| <b>Number of Simulators</b> 5  | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/30/2018 18:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

|                                |                              |                               |                              |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> MIAMI PD         | <b>Date</b> 08/29/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> PASTOR,JOSE,F | <b>Time</b> 20:08:36         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201703A      | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 203498 |
| <b>Expiration</b> 03/07/2019   | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 12/10/2018 |
| <b>Number of Simulators</b> 5  | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/29/2018 20:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

|                                |                              |                               |                              |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> MIAMI PD         | <b>Date</b> 09/30/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> PASTOR,JOSE,F | <b>Time</b> 22:26:20         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201703A      | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 203498 |
| <b>Expiration</b> 03/07/2019   | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 12/10/2018 |
| <b>Number of Simulators</b> 5  | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 09/30/2018 22:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

|                                |                              |                               |                              |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> MIAMI PD         | <b>Date</b> 10/30/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> PASTOR,JOSE,F | <b>Time</b> 16:58:29         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201802B      | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 203498 |
| <b>Expiration</b> 02/06/2020   | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 12/10/2018 |
| <b>Number of Simulators</b> 5  | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/30/2018 17:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

**Agency** MIAMI PD **Date** 11/27/2018 **Type** Agency Inspection  
**Inspector** PASTOR,JOSE,F **Time** 13:33:14 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 620051  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 11/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/27/2018 14:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

|                                       |                        |                                   |
|---------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MIAMI PD                | <b>Date</b> 12/10/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA,DAVID,E | <b>Time</b> 12:00:02   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0006      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0006 |
| 0.20 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0008 |

#### Remarks:

Data Download Date/Time 12/10/2018 12:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

**Agency** MIAMI PD **Date** 12/27/2018 **Type** Agency Inspection  
**Inspector** PASTOR,JOSE,F **Time** 14:44:07 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 620051  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 11/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/27/2018 15:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

|                                    |                              |                               |                              |
|------------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> MIAMI PD             | <b>Date</b> 01/30/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> REYES, BARBARA, T | <b>Time</b> 13:06:29         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201703A          | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 203498 |
| <b>Expiration</b> 03/07/2019       | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 12/10/2018 |
| <b>Number of Simulators</b> 5      | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/30/2018 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

|                                |                              |                               |                              |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> MIAMI PD         | <b>Date</b> 02/27/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> PASTOR,JOSE,F | <b>Time</b> 11:49:24         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201703A      | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 203498 |
| <b>Expiration</b> 03/07/2019   | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 12/10/2018 |
| <b>Number of Simulators</b> 5  | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 02/27/2018 13:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

|                                |                              |                               |                              |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> MIAMI PD         | <b>Date</b> 03/27/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> PASTOR,JOSE,F | <b>Time</b> 12:51:09         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201703A      | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 203498 |
| <b>Expiration</b> 03/07/2019   | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 12/10/2018 |
| <b>Number of Simulators</b> 5  | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/27/2018 13:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

|                                |                              |                               |                              |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> MIAMI PD         | <b>Date</b> 04/26/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> PASTOR,JOSE,F | <b>Time</b> 19:56:08         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201703A      | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 203498 |
| <b>Expiration</b> 03/07/2019   | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 12/10/2018 |
| <b>Number of Simulators</b> 5  | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/26/2018 20:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

**Agency** MIAMI PD      **Date** 05/14/2018      **Type** Agency Inspection  
**Inspector** FERNANDEZ, ANGEL,      **Time** 14:30:41      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201711B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 203498  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 12/10/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

OFC PASTOR WAS PRESENT

**Data Download Date/Time** 05/14/2018 15:23

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

## Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

**Agency** MIAMI PD  
**Inspector** FERNANDEZ, ANGEL,

**Date** 06/28/2018  
**Time** 10:40:36

|          |                   |
|----------|-------------------|
| Type     | Agency Inspection |
| Software | 8100.27           |

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 203498 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 12/10/2018 |

Number of Simulators 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

|  |
|--|
|  |
|--|

**Data Download Date/Time** 06/28/2018 12:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

|                                |                              |                               |                              |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> MIAMI PD         | <b>Date</b> 07/30/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> PASTOR,JOSE,F | <b>Time</b> 17:43:20         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201703A      | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 203498 |
| <b>Expiration</b> 03/07/2019   | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 12/10/2018 |
| <b>Number of Simulators</b> 5  | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/30/2018 18:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

|                                |                              |                               |                              |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> MIAMI PD         | <b>Date</b> 08/29/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> PASTOR,JOSE,F | <b>Time</b> 19:52:09         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201703A      | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 203498 |
| <b>Expiration</b> 03/07/2019   | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 12/10/2018 |
| <b>Number of Simulators</b> 5  | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/29/2018 19:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MIAMI PD                  | <b>Date</b> 09/04/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 14:23:28   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.077  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.077  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.077  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 09/04/2018 14:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

|                                |                              |                               |                              |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> MIAMI PD         | <b>Date</b> 09/30/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> PASTOR,JOSE,F | <b>Time</b> 22:02:52         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201703A      | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 203498 |
| <b>Expiration</b> 03/07/2019   | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 12/10/2018 |
| <b>Number of Simulators</b> 5  | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/30/2018 22:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

|                                |                              |                               |                              |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> MIAMI PD         | <b>Date</b> 10/30/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> PASTOR,JOSE,F | <b>Time</b> 16:39:08         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201802B      | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 203498 |
| <b>Expiration</b> 02/06/2020   | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 12/10/2018 |
| <b>Number of Simulators</b> 5  | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/30/2018 16:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

**Agency** MIAMI PD                      **Date** 11/27/2018                      **Type** Agency Inspection  
**Inspector** PASTOR,JOSE,F                      **Time** 13:21:43                      **Software** 8100.27

**0.05 Lot #** 201802B                      **0.08 Lot #** 201709A                      **0.20 Lot #** 201702B                      **0.08 Gas Lot #** 620051  
**Expiration** 02/06/2020                      **Expiration** 09/12/2019                      **Expiration** 02/23/2019                      **Expiration** 11/05/2019

**Number of Simulators** 5                      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 11/27/2018 13:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

**Agency** MIAMI PD **Date** 12/27/2018 **Type** Agency Inspection  
**Inspector** PASTOR,JOSE,F **Time** 14:24:30 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 620051  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 11/05/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/27/2018 15:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

|                                   |                              |                               |                                  |
|-----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> CALHOUN COUNTY SO   | <b>Date</b> 01/18/2018       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> WILSON,STANLEY,T | <b>Time</b> 08:42:49         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201603D         | <b>0.08 Lot #</b> 201601F    | <b>0.20 Lot #</b> 201604C     | <b>0.08 Gas Lot #</b> 06717080A6 |
| <b>Expiration</b> 03/08/2018      | <b>Expiration</b> 01/26/2018 | <b>Expiration</b> 04/05/2018  | <b>Expiration</b> 05/05/2019     |
| <b>Number of Simulators</b> 5     | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NONE

**Data Download Date/Time** 01/18/2018 09:52

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

**Agency** CALHOUN COUNTY SO      **Date** 02/05/2018      **Type** Agency Inspection  
**Inspector** WILSON,STANLEY,T      **Time** 09:02:53      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201709A      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 06717080A6  
**Expiration** 03/08/2018      **Expiration** 09/12/2019      **Expiration** 04/05/2018      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 02/05/2018 10:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

**Agency** CALHOUN COUNTY SO      **Date** 03/05/2018      **Type** Agency Inspection  
**Inspector** WILSON,STANLEY,T      **Time** 09:43:49      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A6  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 03/05/2018 11:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

**Agency** CALHOUN COUNTY SO      **Date** 04/02/2018      **Type** Agency Inspection  
**Inspector** WILSON,STANLEY,T      **Time** 08:43:24      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A6  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 04/02/2018 09:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

**Agency** CALHOUN COUNTY SO      **Date** 05/16/2018      **Type** Agency Inspection  
**Inspector** WILSON,STANLEY,T      **Time** 08:34:53      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A6  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.207  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.206  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 05/16/2018 09:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

**Agency** CALHOUN COUNTY SO      **Date** 06/05/2018      **Type** Agency Inspection  
**Inspector** WILSON,STANLEY,T      **Time** 08:44:42      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A6  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 06/05/2018 10:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

**Agency** CALHOUN COUNTY SO      **Date** 07/19/2018      **Type** Agency Inspection  
**Inspector** WILSON,STANLEY,T      **Time** 10:47:15      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A6  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 07/19/2018 12:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

**Agency** CALHOUN COUNTY SO      **Date** 08/09/2018      **Type** Agency Inspection  
**Inspector** WILSON,STANLEY,T      **Time** 08:34:20      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A6  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 08/09/2018 09:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

|                                   |                              |                               |                                  |
|-----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> CALHOUN COUNTY SO   | <b>Date</b> 09/12/2018       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> WILSON,STANLEY,T | <b>Time</b> 09:00:50         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201703A         | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 06717080A6 |
| <b>Expiration</b> 03/07/2019      | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 05/05/2019     |
| <b>Number of Simulators</b> 5     | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NONE

**Data Download Date/Time** 09/12/2018 10:15

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

|                                   |                        |                                   |
|-----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> CALHOUN COUNTY SO   | <b>Date</b> 09/27/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> MURPHY,PATRICK,J | <b>Time</b> 10:54:48   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 09/27/2018 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

**Agency** CALHOUN COUNTY SO      **Date** 10/25/2018      **Type** Agency Inspection  
**Inspector** WILSON,STANLEY,T      **Time** 08:48:54      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A6  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 10/25/2018 09:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

**Agency** CALHOUN COUNTY SO      **Date** 11/13/2018      **Type** Agency Inspection  
**Inspector** WILSON,STANLEY,T      **Time** 12:26:14      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A6  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 11/13/2018 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

**Agency** CALHOUN COUNTY SO      **Date** 12/03/2018      **Type** Agency Inspection  
**Inspector** WILSON,STANLEY,T      **Time** 08:55:36      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A6  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 12/03/2018 10:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

**Agency** CALHOUN COUNTY SO      **Date** 01/03/2019      **Type** Agency Inspection  
**Inspector** WILSON,STANLEY,T      **Time** 09:09:18      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 06717080A6  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 05/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 01/03/2019 11:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

**Agency** HIALEAH PD      **Date** 01/24/2018      **Type** Agency Inspection  
**Inspector** MONTROYA, GENARO, N      **Time** 12:14:27      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201611B      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 712031  
**Expiration** 03/08/2018      **Expiration** 11/15/2018      **Expiration** 04/05/2018      **Expiration** 02/21/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/24/2018 12:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

**Agency** HIALEAH PD      **Date** 02/14/2018      **Type** Agency Inspection  
**Inspector** MONTOYA, GENARO, N      **Time** 13:11:41      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 712031  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/21/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/14/2018 13:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

**Agency** HIALEAH PD      **Date** 03/31/2018      **Type** Agency Inspection  
**Inspector** MONTROYA, GENARO, N      **Time** 09:58:58      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 712031  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/21/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/31/2018 10:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> HIALEAH PD                | <b>Date</b> 04/04/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 10:15:30   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG715202 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 06/01/2019   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 04/04/2018 10:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

**Agency** HIALEAH PD **Date** 04/27/2018 **Type** Agency Inspection  
**Inspector** MONTROYA, GENARO, N **Time** 12:37:17 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 712031  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 02/21/2020

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/27/2018 12:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> HIALEAH PD                | <b>Date</b> 11/26/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 09:16:37   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 11/26/2018 09:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

|                                      |                              |                               |                              |
|--------------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> HIALEAH PD             | <b>Date</b> 12/19/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> MONTROYA, GENARO, N | <b>Time</b> 10:41:03         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201703A            | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 712031 |
| <b>Expiration</b> 03/07/2019         | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 02/21/2020 |
| <b>Number of Simulators</b> 5        | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/19/2018 10:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

**Agency** HOMESTEAD PD  
**Inspector** PEREZ,RAUL,

**Date** 01/29/2018  
**Time** 15:31:54

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 1870563  
**Expiration** 03/08/2018      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 06/17/2018

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 01/29/2018 15:40

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

**Agency** HOMESTEAD PD  
**Inspector** PEREZ,RAUL,

**Date** 02/27/2018  
**Time** 14:53:42

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 1870563  
**Expiration** 03/08/2018      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 06/17/2018

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NONE

**Data Download Date/Time** 02/27/2018 14:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

**Agency** HOMESTEAD PD  
**Inspector** PEREZ,RAUL,

**Date** 03/29/2018  
**Time** 09:11:51

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 1870563  
**Expiration** 03/07/2019      **Expiration** 09/12/2018      **Expiration** 02/23/2019      **Expiration** 06/17/2018

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.204  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NONE

**Data Download Date/Time** 03/29/2018 09:16

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> HOMESTEAD PD              | <b>Date</b> 04/04/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 10:30:24   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG715202 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 06/01/2019   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 04/04/2018 10:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

**Agency** HOMESTEAD PD  
**Inspector** PEREZ,RAUL,

**Date** 04/25/2018  
**Time** 09:15:12

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 1870563  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 06/17/2018

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NONE

**Data Download Date/Time** 04/25/2018 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

**Agency** HOMESTEAD PD  
**Inspector** PEREZ,RAUL,

**Date** 05/29/2018  
**Time** 10:30:39

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 937240  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 04/07/2021

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.083  |
|                                             |        | 0.20 g/210L Test # 3             | 0.205  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NONE

**Data Download Date/Time** 05/29/2018 10:37

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

**Agency** HOMESTEAD PD  
**Inspector** PEREZ,RAUL,

**Date** 06/20/2018  
**Time** 10:35:51

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 937240  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 04/07/2021

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.204  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.205  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 06/20/2018 10:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

**Agency** HOMESTEAD PD  
**Inspector** PEREZ,RAUL,

**Date** 07/10/2018  
**Time** 09:40:23

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 937240  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 04/07/2021

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NONE

**Data Download Date/Time** 07/10/2018 09:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

**Agency** HOMESTEAD PD  
**Inspector** PEREZ,RAUL,

**Date** 08/25/2018  
**Time** 22:26:45

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 937240  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 04/07/2021

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NONE

**Data Download Date/Time** 08/25/2018 22:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

**Agency** HOMESTEAD PD  
**Inspector** PEREZ,RAUL,

**Date** 09/25/2018  
**Time** 09:12:52

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 937240  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 04/07/2021

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.083  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.204  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.083  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.083  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Time-Date changed. INCORRECT TIME

**Data Download Date/Time** 09/25/2018 09:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

|                               |                        |                               |  |
|-------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> HOMESTEAD PD    | <b>Date</b> 10/23/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> PEREZ, RAUL, | <b>Time</b> 13:16:11   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201802B    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201802G    | <b>0.08 Gas Lot #</b> 937240 |
| <b>Expiration</b> 02/06/2020 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/22/2020 | <b>Expiration</b> 04/07/2021 |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 4 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.082  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.051  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.082  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.201  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.000  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.082  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.051  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.082  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.000  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3               | INT    |

**Remarks:**

08: Control Outside Tolerance DRY GAS TUBE DISCONNECTED.

**Data Download Date/Time** 10/23/2018 13:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

**Agency** HOMESTEAD PD  
**Inspector** PEREZ,RAUL,

**Date** 11/27/2018  
**Time** 15:04:25

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 937240  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 04/07/2021

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 11/27/2018 15:12

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

**Agency** HOMESTEAD PD  
**Inspector** PEREZ,RAUL,

**Date** 12/19/2018  
**Time** 09:45:51

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 937240  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 04/07/2021

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NONE

**Data Download Date/Time** 12/19/2018 09:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

**Agency** MIAMI BEACH PD  
**Inspector** MILLAN,KEVIN,

**Date** 01/31/2018  
**Time** 10:43:32

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 390314  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 04/19/2019

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/31/2018 10:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

**Agency** MIAMI BEACH PD  
**Inspector** MILLAN,KEVIN,

**Date** 02/26/2018  
**Time** 14:43:51

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201611B      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 390314  
**Expiration** 03/07/2019      **Expiration** 11/15/2018      **Expiration** 02/23/2019      **Expiration** 04/19/2019

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/26/2018 15:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

**Agency** MIAMI BEACH PD      **Date** 03/21/2018      **Type** Agency Inspection  
**Inspector** MILLAN,KEVIN,      **Time** 11:20:48      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 814181  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 07/24/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/21/2018 11:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

**Agency** MIAMI BEACH PD  
**Inspector** MILLAN,KEVIN,

**Date** 04/26/2018  
**Time** 11:52:35

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 814181  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 07/24/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/30/2018 11:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

**Agency** MIAMI BEACH PD  
**Inspector** MILLAN,KEVIN,

**Date** 05/31/2018  
**Time** 10:58:51

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 814181  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 07/24/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 05/31/2018 11:05

**Printed: 14 January 2019**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MIAMI BEACH PD            | <b>Date</b> 06/19/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 09:47:06   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG715202 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 06/01/2019   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0008      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.085  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.085  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.085  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.085  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.085  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.048  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.085  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.085  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0024 |

#### Remarks:

Data Download Date/Time 06/19/2018 09:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

|                                |                        |                               |  |
|--------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MIAMI BEACH PD   | <b>Date</b> 06/25/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> MILLAN,KEVIN, | <b>Time</b> 11:55:59   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 814181 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 07/24/2020 |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.086  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.079  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.200  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.085  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.081  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.048  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.079  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.085  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3               | INT    |

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 06/25/2018 12:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

**Agency** MIAMI BEACH PD  
**Inspector** MILLAN,KEVIN,

**Date** 07/23/2018  
**Time** 12:06:22

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 814181  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 07/24/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/23/2018 12:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

**Agency** MIAMI BEACH PD  
**Inspector** MILLAN,KEVIN,

**Date** 08/22/2018  
**Time** 10:45:19

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 814181  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 07/24/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 08/22/2018 11:11

**Printed: 14 January 2019**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

**Agency** MIAMI BEACH PD      **Date** 09/25/2018      **Type** Agency Inspection  
**Inspector** MILLAN,KEVIN,      **Time** 10:06:53      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 814181  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 07/24/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/25/2018 10:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

**Agency** MIAMI BEACH PD  
**Inspector** MILLAN,KEVIN,

**Date** 10/29/2018  
**Time** 14:19:51

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 814181  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 07/24/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/29/2018 14:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

|                                |                              |                               |
|--------------------------------|------------------------------|-------------------------------|
| <b>Agency</b> MIAMI BEACH PD   | <b>Date</b> 11/30/2018       | <b>Type</b> Agency Inspection |
| <b>Inspector</b> MILLAN,KEVIN, | <b>Time</b> 14:25:23         | <b>Software</b> 8100.27       |
| <b>0.05 Lot #</b> 201802B      | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201802G     |
| <b>0.08 Gas Lot #</b> 814181   |                              |                               |
| <b>Expiration</b> 02/06/2020   | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/22/2020  |
| <b>Expiration</b> 07/24/2020   |                              |                               |
| <b>Number of Simulators</b> 5  |                              | <b>Compliance</b> Yes         |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.195  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.076  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

08: Control Outside Tolerance WRONG SIMULATOR.

**Data Download Date/Time** 11/30/2018 14:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

**Agency** MIAMI BEACH PD  
**Inspector** MILLAN,KEVIN,

**Date** 12/27/2018  
**Time** 14:33:55

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 814181  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 07/24/2020

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/27/2018 14:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

**Agency** MIAMI-DADE PD      **Date** 01/01/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:36:54      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201601F      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/08/2018      **Expiration** 01/26/2018      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.052  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/01/2018 11:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MIAMI-DADE PD             | <b>Date</b> 01/09/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 11:47:50   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG626604 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 09/22/2018   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0008      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.084  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.083  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.083  |
|                                             |        | 0.20 g/210L Test # 4             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.082  |
|                                             |        | 0.20 g/210L Test # 5             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.082  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.083  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.077  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.083  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.082  |
| 0.20 g/210L Test # 9                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0009 |
| 0.08 g/210L Test (Std Dev)             | 0.0008 |
| 0.20 g/210L Test (Std Dev)             | 0.0011 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0006 |

#### Remarks:

Data Download Date/Time 01/09/2018 11:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

**Agency** MIAMI-DADE PD      **Date** 01/11/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 09:02:17      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201601F      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/08/2018      **Expiration** 01/26/2018      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/11/2018 09:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

**Agency** MIAMI-DADE PD      **Date** 02/24/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 22:56:05      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201709A      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/08/2018      **Expiration** 09/12/2019      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/24/2018 22:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

**Agency** MIAMI-DADE PD      **Date** 03/01/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 09:47:38      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/01/2018 10:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

**Agency** MIAMI-DADE PD      **Date** 04/20/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 09:41:34      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/20/2018 09:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

**Agency** MIAMI-DADE PD      **Date** 05/09/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:03:18      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/09/2018 13:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

**Agency** MIAMI-DADE PD      **Date** 06/29/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:47:06      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/29/2018 11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> MIAMI-DADE PD      | <b>Date</b> 07/04/2018       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> ESPINOZA,PABLO, | <b>Time</b> 09:57:27         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201703A        | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 20016080A2 |
| <b>Expiration</b> 03/07/2019     | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 09/05/2018     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/04/2018 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

|                                  |                        |                               |
|----------------------------------|------------------------|-------------------------------|
| <b>Agency</b> MIAMI-DADE PD      | <b>Date</b> 08/17/2018 | <b>Type</b> Agency Inspection |
| <b>Inspector</b> ESPINOZA,PABLO, | <b>Time</b> 11:48:56   | <b>Software</b> 8100.27       |

|                              |                              |                   |                       |
|------------------------------|------------------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> 201802B    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> 02/06/2020 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> | <b>Expiration</b>     |

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1          | 0.029  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1 | 0.031  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 1   | INT    |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Alcohol Free Test # 2         | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # 2          | 0.048  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.08 g/210L Test # 2          | 0.029  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # (Repeat) 2 | 0.032  |
| 0.05 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2   | INT    |
| 0.08 g/210L Test: OK                        | No     | Alcohol Free Test # 3         | 0.000  |
| 0.20 g/210L Test: OK                        | No     | 0.05 g/210L Test # 3          | 0.050  |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | 0.08 g/210L Test # 3          | 0.027  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3 | 0.029  |
|                                             |        | Interferent Detect Test # 3   | INT    |

**Remarks:**

A F / M A: Interferent Detect/CHANGED MOUTHPIECE. 08: Control Outside Tolerance/TIGHTENED SIM, Control Outsid. Non-compliance:FAILED .08 TEST.

**Data Download Date/Time** 08/31/2018 12:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> MIAMI-DADE PD      | <b>Date</b> 08/31/2018       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> ESPINOZA,PABLO, | <b>Time</b> 12:24:26         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201802B        | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 20016080A2 |
| <b>Expiration</b> 02/06/2020     | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 09/05/2018     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.053  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.051  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.082  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.058  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.052  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.053  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

05: Control Outside Tolerance/CHANGED SIM.

**Data Download Date/Time** 08/31/2018 12:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

**Agency** MIAMI-DADE PD      **Date** 09/03/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 13:02:25      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.053  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.053  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.054  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/03/2018 13:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

**Agency** MIAMI-DADE PD      **Date** 10/05/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:25:35      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.052  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.075  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/05/2018 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

|                                  |                        |                               |  |
|----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MIAMI-DADE PD      | <b>Date</b> 11/07/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> ESPINOZA,PABLO, | <b>Time</b> 09:45:05   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 34917080A4 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 02/05/2020     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.053  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.051  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.077  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.056  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.050  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

05: Control Outside Tolerance/CHANGED SIM.

**Data Download Date/Time** 11/07/2018 10:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> MIAMI-DADE PD      | <b>Date</b> 12/06/2018       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> ESPINOZA,PABLO, | <b>Time</b> 09:49:53         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201703A        | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 34917080A4 |
| <b>Expiration</b> 03/07/2019     | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 02/05/2020     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.083  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/06/2018 12:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

**Agency** MIAMI DADE PD      **Date** 01/12/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 09:27:47      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201601F      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/08/2018      **Expiration** 01/26/2018      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/12/2018 09:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MIAMI DADE PD             | <b>Date</b> 01/17/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 08:53:14   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG626604 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 09/22/2018   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.051  |
|                                             |        | 0.08 g/210L Test # 4             | 0.082  |
|                                             |        | 0.20 g/210L Test # 4             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.082  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.082  |
|                                             |        | 0.20 g/210L Test # 5             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.082  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.083  |
| 0.20 g/210L Test # 6                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.082  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.083  |
| 0.20 g/210L Test # 7                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.082  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.082  |
| 0.20 g/210L Test # 8                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.082  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.082  |
| 0.20 g/210L Test # 9                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.082  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.083  |
| 0.20 g/210L Test # 10                  | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.082  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0006 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 01/17/2018 08:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

**Agency** MIAMI DADE PD      **Date** 01/23/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:54:56      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201601F      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/08/2018      **Expiration** 01/26/2018      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/23/2018 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

**Agency** MIAMI DADE PD      **Date** 02/24/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 22:50:00      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201709A      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/08/2018      **Expiration** 09/12/2019      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/24/2018 23:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

**Agency** MIAMI DADE PD      **Date** 03/01/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 09:42:17      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/01/2018 10:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

**Agency** MIAMI DADE PD      **Date** 04/20/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 09:35:21      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/20/2018 09:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

|                                  |                        |                               |  |
|----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MIAMI DADE PD      | <b>Date</b> 05/09/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> ESPINOZA,PABLO, | <b>Time</b> 10:16:59   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 20016080A2 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 09/05/2018     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.086  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.080  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.202  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.084  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.080  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.084  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3               | INT    |

**Remarks:**

08: Control Outside Tolerance/MONITOR AND CONTINUE PER FDLE.

**Data Download Date/Time** 05/09/2018 13:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

**Agency** MIAMI DADE PD      **Date** 06/29/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:50:21      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/29/2018 11:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

**Agency** MIAMI DADE PD  
**Inspector** ESPINOZA,PABLO,

**Date** 07/04/2018  
**Time** 09:59:34

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 07/04/2018 10:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

**Agency** MIAMI DADE PD      **Date** 08/17/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 11:54:53      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.075  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 08/20/2018 14:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

**Agency** MIAMI DADE PD      **Date** 09/03/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 11:12:55      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/03/2018 11:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> MIAMI DADE PD      | <b>Date</b> 10/23/2018       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> ESPINOZA,PABLO, | <b>Time</b> 10:26:41         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201703A        | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 34917080A4 |
| <b>Expiration</b> 03/07/2019     | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 02/05/2020     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/25/2018 08:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

**Agency** MIAMI DADE PD      **Date** 11/08/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 09:01:15      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/08/2018 09:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

**Agency** MIAMI DADE PD      **Date** 12/14/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 11:47:33      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/14/2018 13:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

**Agency** MIAMI DADE PD      **Date** 01/10/2019      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 09:56:46      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: Ambient Fail/CLEARED AREA.

**Data Download Date/Time** 01/10/2019 10:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

**Agency** MIAMI-DADE PD      **Date** 01/25/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:21:54      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201601F      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/08/2018      **Expiration** 01/26/2018      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/25/2018 11:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MIAMI-DADE PD             | <b>Date</b> 01/29/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 13:05:56   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG626604 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 09/22/2018   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.082  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.082  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 01/29/2018 13:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

**Agency** MIAMI-DADE PD      **Date** 02/01/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:22:24      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201709A      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/08/2018      **Expiration** 09/12/2019      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/01/2018 13:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

**Agency** MIAMI-DADE PD      **Date** 03/07/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 09:23:09      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/07/2018 09:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

**Agency** MIAMI-DADE PD      **Date** 04/17/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:18:50      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/17/2018 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

**Agency** MIAMI-DADE PD      **Date** 05/15/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:10:43      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/15/2018 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

**Agency** MIAMI-DADE PD      **Date** 06/07/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 11:23:13      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/07/2018 14:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

**Agency** MIAMI-DADE PD      **Date** 07/10/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 11:09:43      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/10/2018 12:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

|                                  |                        |                               |  |
|----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MIAMI-DADE PD      | <b>Date</b> 08/07/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> ESPINOZA,PABLO, | <b>Time</b> 11:13:58   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201802B    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 20016080A2 |
| <b>Expiration</b> 02/06/2020 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 09/05/2018     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.080  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.203  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.080  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.080  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.079  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3               | INT    |

**Remarks:**

08: Control Outside Tolerance/FORGOT TO ATTACH HOSE.

**Data Download Date/Time** 08/07/2018 11:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

**Agency** MIAMI-DADE PD      **Date** 09/03/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 11:10:01      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.053  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/03/2018 11:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

**Agency** MIAMI-DADE PD      **Date** 10/23/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:25:24      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/25/2018 08:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

**Agency** MIAMI-DADE PD      **Date** 11/08/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 08:55:33      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/08/2018 09:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

**Agency** MIAMI-DADE PD      **Date** 12/14/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 11:44:58      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 12/14/2018 13:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

**Agency** MIAMI-DADE PD      **Date** 01/04/2019      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 11:56:01      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/08/2019 13:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

**Agency** MIAMI-DADE PD      **Date** 01/15/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:15:59      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201601F      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/08/2018      **Expiration** 01/26/2018      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/15/2018 10:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

**Agency** MIAMI-DADE PD      **Date** 02/08/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 09:18:02      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201709A      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/08/2018      **Expiration** 09/12/2019      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/08/2018 09:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MIAMI-DADE PD             | <b>Date</b> 02/13/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 09:43:45   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG626605 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 09/22/2018   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.051  |
|                                             |        | 0.08 g/210L Test # 4             | 0.082  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.081  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.052  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.082  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.082  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.052  |
| 0.08 g/210L Test # 10                  | 0.082  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 02/13/2018 10:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

**Agency** MIAMI-DADE PD      **Date** 02/19/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:35:15      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201709A      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/08/2018      **Expiration** 09/12/2019      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/19/2018 11:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

**Agency** MIAMI-DADE PD      **Date** 03/09/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:43:13      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/09/2018 10:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

**Agency** MIAMI-DADE PD      **Date** 04/03/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 11:11:27      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/03/2018 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

**Agency** MIAMI-DADE PD      **Date** 05/02/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 09:58:40      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.075  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/02/2018 12:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

**Agency** MIAMI-DADE PD      **Date** 06/13/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:20:36      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.053  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/13/2018 11:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

**Agency** MIAMI-DADE PD      **Date** 07/04/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 11:18:51      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.083  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.205  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.084  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/04/2018 11:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

**Agency** MIAMI-DADE PD      **Date** 08/07/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 11:05:23      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A:NOT ENOUGH MA.

**Data Download Date/Time** 08/07/2018 11:15

**Printed: 14 January 2019**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

|                                  |                        |                               |  |
|----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MIAMI-DADE PD      | <b>Date</b> 10/31/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> ESPINOZA,PABLO, | <b>Time</b> 09:51:53   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 34917080A4 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 02/05/2020     |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | INT    |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.077  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.197  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.048  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.077  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3               | INT    |

**Remarks:**

08: Interferent Detect/CHECKED AREA.

**Data Download Date/Time** 10/31/2018 13:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

**Agency** MIAMI-DADE PD      **Date** 11/02/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:23:21      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/06/2018 07:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

**Agency** MIAMI-DADE PD      **Date** 12/06/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 09:46:35      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/06/2018 12:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

**Agency** MIAMI-DADE PD      **Date** 01/25/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:11:55      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201601F      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/08/2018      **Expiration** 01/26/2018      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/25/2018 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

|                                       |                        |                                   |
|---------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MIAMI-DADE PD           | <b>Date</b> 01/31/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA,DAVID,E | <b>Time</b> 08:53:22   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG626604 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 09/22/2018   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.082  |
|                                             |        | 0.20 g/210L Test # 4             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.082  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.082  |
| 0.20 g/210L Test # 8                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 01/31/2018 08:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

**Agency** MIAMI-DADE PD      **Date** 02/02/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 09:22:54      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201709A      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/08/2018      **Expiration** 09/12/2019      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/02/2018 10:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

**Agency** MIAMI-DADE PD      **Date** 03/22/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 11:27:33      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/22/2018 11:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> MIAMI-DADE PD      | <b>Date</b> 04/03/2018       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> ESPINOZA,PABLO, | <b>Time</b> 09:21:14         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201703A        | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 20016080A2 |
| <b>Expiration</b> 03/07/2019     | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 09/05/2018     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/03/2018 09:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

**Agency** MIAMI-DADE PD      **Date** 05/18/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:15:02      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/21/2018 13:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

**Agency** MIAMI-DADE PD      **Date** 06/15/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:29:59      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/18/2018 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> MIAMI-DADE PD      | <b>Date</b> 07/12/2018       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> ESPINOZA,PABLO, | <b>Time</b> 09:29:00         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201703A        | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 20016080A2 |
| <b>Expiration</b> 03/07/2019     | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 09/05/2018     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/12/2018 09:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

**Agency** MIAMI-DADE PD      **Date** 08/16/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 09:26:25      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/16/2018 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

**Agency** MIAMI-DADE PD      **Date** 09/04/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 13:32:18      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.075  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/05/2018 09:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

**Agency** MIAMI-DADE PD      **Date** 10/26/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:02:24      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.075  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/26/2018 10:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

|                                  |                        |                               |  |
|----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MIAMI-DADE PD      | <b>Date</b> 11/01/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> ESPINOZA,PABLO, | <b>Time</b> 10:17:35   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 34917080A4 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 02/05/2020     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.074  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.075  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.195  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.075  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.075  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.075  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.075  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

08: Control Outside Tolerance/TIGHTENED SIM.

**Data Download Date/Time** 11/01/2018 12:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

**Agency** MIAMI-DADE PD      **Date** 12/28/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:03:00      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.052  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.052  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/28/2018 10:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

**Agency** MIAMI-DADE PD      **Date** 01/04/2019      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 12:01:26      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/08/2019 13:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

**Agency** MIAMI-DADE PD      **Date** 01/01/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:32:50      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201601F      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/08/2018      **Expiration** 01/26/2018      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/01/2018 10:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

|                                       |                        |                                   |
|---------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MIAMI-DADE PD           | <b>Date</b> 01/09/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA,DAVID,E | <b>Time</b> 12:04:22   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG626604 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 09/22/2018   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.077  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.077  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 01/09/2018 12:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

**Agency** MIAMI-DADE PD      **Date** 01/11/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 09:01:35      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201601F      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/08/2018      **Expiration** 01/26/2018      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/11/2018 09:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

|                                  |                        |                               |  |
|----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MIAMI-DADE PD      | <b>Date</b> 02/08/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> ESPINOZA,PABLO, | <b>Time</b> 11:41:11   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201603D    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201604C    | <b>0.08 Gas Lot #</b> 20016080A2 |
| <b>Expiration</b> 03/08/2018 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 04/05/2018 | <b>Expiration</b> 09/05/2018     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                 | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                 | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                 | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1     | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1          | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # (Repeat) 1 | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                 | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                 | 0.078  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                 | 0.197  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2     | 0.079  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2          | 0.000  |
|                                             |        | Interferent Detect Test # (Repeat) 2 | INT    |
|                                             |        | Alcohol Free Test # 3                | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                 | 0.048  |
|                                             |        | 0.08 g/210L Test # 3                 | 0.078  |
|                                             |        | 0.20 g/210L Test # 3                 | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3     | 0.079  |
|                                             |        | Interferent Detect Test # 3          | 0.000  |
|                                             |        | Interferent Detect Test # (Repeat) 3 | INT    |

**Remarks:**

A F / M A:NOT ENOUGH MA. Int Det:SIM NOT CLOSED PROPERLY.

**Data Download Date/Time** 02/08/2018 11:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

**Agency** MIAMI-DADE PD      **Date** 03/22/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 11:30:54      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 04/05/2018      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/22/2018 11:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

**Agency** MIAMI-DADE PD      **Date** 04/06/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:22:42      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 04/06/2018 10:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

**Agency** MIAMI-DADE PD      **Date** 05/18/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:11:27      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/21/2018 13:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

**Agency** MIAMI-DADE PD      **Date** 06/15/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:36:46      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/18/2018 11:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> MIAMI-DADE PD      | <b>Date</b> 07/12/2018       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> ESPINOZA,PABLO, | <b>Time</b> 11:14:14         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201703A        | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 20016080A2 |
| <b>Expiration</b> 03/07/2019     | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 09/05/2018     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 07/12/2018 11:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

|                                  |                        |                               |  |
|----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MIAMI-DADE PD      | <b>Date</b> 08/16/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> ESPINOZA,PABLO, | <b>Time</b> 09:33:02   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201802B    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 20016080A2 |
| <b>Expiration</b> 02/06/2020 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 09/05/2018     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.078  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.199  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.000  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.048  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.078  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.000  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3               | INT    |

**Remarks:**

08: Control Outside Tolerance/FORGOT TO CONNECT HOSE.

**Data Download Date/Time** 08/16/2018 09:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

**Agency** MIAMI-DADE PD      **Date** 09/04/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 13:38:47      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 20016080A2  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 09/05/2018

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/06/2018 10:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

**Agency** MIAMI-DADE PD      **Date** 10/26/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 10:11:03      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/26/2018 10:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

**Agency** MIAMI-DADE PD      **Date** 11/01/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 09:26:51      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/01/2018 12:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

**Agency** MIAMI-DADE PD      **Date** 12/21/2018      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 09:15:25      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/24/2018 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

**Agency** MIAMI-DADE PD      **Date** 01/10/2019      **Type** Agency Inspection  
**Inspector** ESPINOZA,PABLO,      **Time** 09:51:17      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 34917080A4  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 02/05/2020

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/10/2019 10:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

**Agency** NORTH MIAMI BEACH PD      **Date** 01/30/2018      **Type** Agency Inspection  
**Inspector** MANN,CORA,L      **Time** 21:31:28      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201709A      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 12317080A3  
**Expiration** 03/08/2018      **Expiration** 09/12/2019      **Expiration** 04/05/2018      **Expiration** 07/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/30/2018 21:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

**Agency** NORTH MIAMI BEACH PD      **Date** 02/27/2018      **Type** Agency Inspection  
**Inspector** MANN,CORA,L      **Time** 19:12:34      **Software** 8100.27

**0.05 Lot #** 201603D      **0.08 Lot #** 201709A      **0.20 Lot #** 201604C      **0.08 Gas Lot #** 12317080A3  
**Expiration** 03/08/2018      **Expiration** 09/12/2019      **Expiration** 04/05/2018      **Expiration** 07/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/27/2018 19:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> NORTH MIAMI BEACH PD | <b>Date</b> 03/30/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> MANN,CORA,L       | <b>Time</b> 00:25:12   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201604C    | <b>0.08 Gas Lot #</b> 12317080A3 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 04/05/2018 | <b>Expiration</b> 07/05/2019     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.193  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.000  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.194  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.172  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

20: Control Outside Tolerance/IMPROPER SEAL.

**Data Download Date/Time** 03/30/2018 00:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> NORTH MIAMI BEACH PD | <b>Date</b> 04/30/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> MANN,CORA,L       | <b>Time</b> 20:45:02   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201703A    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201702B    | <b>0.08 Gas Lot #</b> 12317080A3 |
| <b>Expiration</b> 03/07/2019 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/23/2019 | <b>Expiration</b> 07/05/2019     |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.000  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.048  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.079  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.169  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.192  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Interferent Detect Test # 1      | INT    |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # 2            | 0.000  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # 2             | 0.000  |
| 0.08 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.049  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.20 g/210L Test # 2             | 0.176  |
|                                             |        | 0.20 g/210L Test # (Repeat) 2    | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.000  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.181  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

05: Control Outside Toleranc/IMPROPER SEAL. 20: Control Outside Tolerance/IMPROPER SEAL.

**Data Download Date/Time** 04/30/2018 20:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

**Agency** NORTH MIAMI BEACH PD      **Date** 05/26/2018      **Type** Agency Inspection  
**Inspector** MANN,CORA,L      **Time** 07:44:31      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 12317080A3  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 07/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/26/2018 07:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

**Agency** NORTH MIAMI BEACH PD      **Date** 06/26/2018      **Type** Agency Inspection  
**Inspector** MANN,CORA,L      **Time** 23:34:45      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 12317080A3  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 07/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/26/2018 23:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

**Agency** NORTH MIAMI BEACH PD      **Date** 07/31/2018      **Type** Agency Inspection  
**Inspector** MANN,CORA,L      **Time** 00:19:00      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 12317080A3  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 07/05/2029

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 07/31/2018 00:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

**Agency** NORTH MIAMI BEACH PD      **Date** 08/29/2018      **Type** Agency Inspection  
**Inspector** BAKER,KENNETH,      **Time** 01:02:44      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 12317080A3  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 07/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/29/2018 01:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

**Agency** NORTH MIAMI BEACH PD      **Date** 09/30/2018      **Type** Agency Inspection  
**Inspector** MANN,CORA,L      **Time** 18:10:47      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 12317080A3  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 07/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 09/30/2018 18:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

|                             |                      |                   |            |                   |                   |                       |            |
|-----------------------------|----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | NORTH MIAMI BEACH PD | <b>Date</b>       | 10/26/2018 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | MANN,CORA,L          | <b>Time</b>       | 00:27:26   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201703A              | <b>0.08 Lot #</b> | 201709A    | <b>0.20 Lot #</b> | 201702B           | <b>0.08 Gas Lot #</b> | 12317080A3 |
| <b>Expiration</b>           | 03/07/2019           | <b>Expiration</b> | 09/12/2019 | <b>Expiration</b> | 02/23/2019        | <b>Expiration</b>     | 07/05/2019 |
| <b>Number of Simulators</b> | 5                    | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/26/2018 00:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> NORTH MIAMI BEACH PD      | <b>Date</b> 10/29/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 14:15:40   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0006      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.083  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.051  |
|                                             |        | 0.08 g/210L Test # 4             | 0.082  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.082  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.082  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.082  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.082  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.082  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.082  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0010 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 10/29/2018 14:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

**Agency** NORTH MIAMI BEACH PD      **Date** 11/01/2018      **Type** Agency Inspection  
**Inspector** MANN,CORA,L      **Time** 21:06:58      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 12317080A3  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 07/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 11/01/2018 21:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

**Agency** NORTH MIAMI BEACH PD      **Date** 12/27/2018      **Type** Agency Inspection  
**Inspector** MANN,CORA,L      **Time** 22:38:52      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 12317080A3  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 07/05/2019

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/27/2018 22:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

**Agency** CHATTAHOOCHEE PD      **Date** 08/14/2018      **Type** Agency Inspection  
**Inspector** MURPHY,PATRICK,J      **Time** 14:32:50      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators**      **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

SKIPPED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 08/15/2018 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

|                                   |                        |                                   |
|-----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> CHATTAHOOCHEE PD    | <b>Date</b> 08/15/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> MURPHY,PATRICK,J | <b>Time</b> 12:20:17   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.077  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.077  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.077  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.077  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.077  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0007 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 08/15/2018 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

**Agency** CHATTAHOOCHEE PD      **Date** 08/22/2018      **Type** Agency Inspection  
**Inspector** BARGE,SUSAN,M      **Time** 13:21:44      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 970670  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 06/07/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NA

**Data Download Date/Time** 08/22/2018 16:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

**Agency** CHATTAHOOCHEE PD      **Date** 09/11/2018      **Type** Agency Inspection  
**Inspector** OBRIAN,FREDERICK,C      **Time** 12:57:56      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | Interferent Detect Test # 1          | 0.000  |
| Minimum Sample Volume Check: OK             |        | Interferent Detect Test # (Repeat) 1 | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 2          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # (Repeat) 2 | 0.000  |
| Interferent Detect Test: Interferent Detect | No     | Alcohol Free Test # 3                | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 3          | 0.000  |
| Alcohol Free Test: OK                       | Yes    | Interferent Detect Test # (Repeat) 3 | 0.000  |
| 0.05 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Test: OK                        | No     |                                      |        |
| 0.20 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                                      |        |

#### Remarks:

NA, .

**Data Download Date/Time** 09/14/2018 17:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

**Agency** CHATTAHOOCHEE PD      **Date** 09/11/2018      **Type** Agency Inspection  
**Inspector** OBRIAN,FREDERICK,C      **Time** 13:12:49      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                        | Result |
|---------------------------------------------|--------|-----------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1       | 0.000  |
| Date and/or Time Adjusted                   | No     | Interferent Detect Test # 1 | 0.000  |
| Minimum Sample Volume Check: OK             |        | Alcohol Free Test # 2       | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | Interferent Detect Test # 2 | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | Alcohol Free Test # 3       | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 3 | 0.000  |
| Interferent Detect Test: Interferent Detect | No     |                             |        |
| Diagnostic Check (Post-Inspection): OK      | No     |                             |        |
| Alcohol Free Test: OK                       | Yes    |                             |        |
| 0.05 g/210L Test: OK                        | No     |                             |        |
| 0.08 g/210L Test: OK                        | No     |                             |        |
| 0.20 g/210L Test: OK                        | No     |                             |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                             |        |

#### Remarks:

Int Det: .

**Data Download Date/Time** 09/14/2018 17:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

|                  |                    |             |            |                 |                   |
|------------------|--------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | CHATTAHOOCHEE PD   | <b>Date</b> | 09/11/2018 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | OBRIAN,FREDERICK,C | <b>Time</b> | 13:38:06   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |   |                   |    |
|-----------------------------|---|-------------------|----|
| <b>Number of Simulators</b> | 5 | <b>Compliance</b> | No |
|-----------------------------|---|-------------------|----|

| Check or Test                               | Result | Test                        | Result |
|---------------------------------------------|--------|-----------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1       | 0.000  |
| Date and/or Time Adjusted                   | No     | Interferent Detect Test # 1 | 0.000  |
| Minimum Sample Volume Check: OK             |        | Alcohol Free Test # 2       | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | Interferent Detect Test # 2 | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | Alcohol Free Test # 3       | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 3 | 0.000  |
| Interferent Detect Test: Interferent Detect | No     |                             |        |
| Diagnostic Check (Post-Inspection): OK      | No     |                             |        |
| Alcohol Free Test: OK                       | Yes    |                             |        |
| 0.05 g/210L Test: OK                        | No     |                             |        |
| 0.08 g/210L Test: OK                        | No     |                             |        |
| 0.20 g/210L Test: OK                        | No     |                             |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                             |        |

#### Remarks:

NA. Non-compliance: .

**Data Download Date/Time** 09/14/2018 17:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

|                  |                    |             |            |                 |                   |
|------------------|--------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | CHATTAHOOCHEE PD   | <b>Date</b> | 09/11/2018 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | OBRIAN,FREDERICK,C | <b>Time</b> | 13:51:02   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |   |                   |     |
|-----------------------------|---|-------------------|-----|
| <b>Number of Simulators</b> | 5 | <b>Compliance</b> | Yes |
|-----------------------------|---|-------------------|-----|

| Check or Test                               | Result | Test                        | Result |
|---------------------------------------------|--------|-----------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1       | 0.000  |
| Date and/or Time Adjusted                   | No     | Interferent Detect Test # 1 | 0.000  |
| Minimum Sample Volume Check: OK             |        | Alcohol Free Test # 2       | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | Interferent Detect Test # 2 | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | Alcohol Free Test # 3       | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 3 | 0.000  |
| Interferent Detect Test: Interferent Detect | No     |                             |        |
| Diagnostic Check (Post-Inspection): OK      | No     |                             |        |
| Alcohol Free Test: OK                       | Yes    |                             |        |
| 0.05 g/210L Test: OK                        | No     |                             |        |
| 0.08 g/210L Test: OK                        | No     |                             |        |
| 0.20 g/210L Test: OK                        | No     |                             |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                             |        |

#### Remarks:

Int Det: .

**Data Download Date/Time** 09/14/2018 17:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

**Agency** CHATTAHOOCHEE PD      **Date** 09/14/2018      **Type** Agency Inspection  
**Inspector** OBRIAN,FREDERICK,C      **Time** 14:44:35      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** No

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | Interferent Detect Test # 1          | 0.000  |
| Minimum Sample Volume Check: OK             |        | Interferent Detect Test # (Repeat) 1 | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 2          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # (Repeat) 2 | 0.000  |
| Interferent Detect Test: Interferent Detect | No     | Alcohol Free Test # 3                | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 3          | 0.000  |
| Alcohol Free Test: OK                       | Yes    | Interferent Detect Test # (Repeat) 3 | 0.000  |
| 0.05 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Test: OK                        | No     |                                      |        |
| 0.20 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                                      |        |

#### Remarks:

ACETONE OUT OF DATE, . Non-compliance: .

**Data Download Date/Time** 09/14/2018 17:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

**Agency** CHATTAHOOCHEE PD      **Date** 09/14/2018      **Type** Agency Inspection  
**Inspector** OBRIAN,FREDERICK,C      **Time** 17:00:16      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 970670  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 06/07/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                 | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                 | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                 | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1     | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1          | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # (Repeat) 1 | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                 | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                 | 0.079  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                 | 0.203  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2     | 0.080  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2          | INT    |
|                                             |        | Interferent Detect Test # (Repeat) 2 | INT    |
|                                             |        | Alcohol Free Test # 3                | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                 | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                 | 0.080  |
|                                             |        | 0.20 g/210L Test # 3                 | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3     | 0.079  |
|                                             |        | Interferent Detect Test # 3          | INT    |
|                                             |        | Interferent Detect Test # (Repeat) 3 | INT    |

#### Remarks:

SIM LEAK TIGHTEN SIM.

**Data Download Date/Time** 09/14/2018 17:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

|                                       |                        |                               |  |
|---------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> CHATTAHOOCHEE PD        | <b>Date</b> 11/08/2018 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> OBRIAN, FREDERICK, C | <b>Time</b> 13:59:36   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201802B    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201802G    | <b>0.08 Gas Lot #</b> 970670 |
| <b>Expiration</b> 02/06/2020 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/22/2020 | <b>Expiration</b> 06/07/2021 |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                 | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                 | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                 | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1     | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1          | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # (Repeat) 1 | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                 | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                 | 0.082  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                 | 0.195  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2     | 0.079  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2          | 0.000  |
|                                             |        | Interferent Detect Test # (Repeat) 2 | INT    |
|                                             |        | Alcohol Free Test # 3                | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                 | 0.048  |
|                                             |        | 0.08 g/210L Test # 3                 | 0.082  |
|                                             |        | 0.20 g/210L Test # 3                 | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3     | 0.079  |
|                                             |        | Interferent Detect Test # 3          | 0.000  |
|                                             |        | Interferent Detect Test # (Repeat) 3 | INT    |

**Remarks:**

TIGHTEN SIM CAP.

**Data Download Date/Time** 11/08/2018 15:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

**Agency** CHATTAHOOCHEE PD      **Date** 12/17/2018      **Type** Agency Inspection  
**Inspector** OBRIAN,FREDERICK,C      **Time** 15:11:35      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 970670  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 06/07/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 12/17/2018 15:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

|                                       |                        |                               |  |
|---------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> CHATTAHOOCHEE PD        | <b>Date</b> 01/10/2019 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> OBRIAN, FREDERICK, C | <b>Time</b> 12:28:10   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201802B    | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201802G    | <b>0.08 Gas Lot #</b> 970670 |
| <b>Expiration</b> 02/06/2020 | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/22/2020 | <b>Expiration</b> 06/07/2021 |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                 | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                 | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                 | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1     | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1          | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # (Repeat) 1 | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                 | 0.051  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                 | 0.081  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                 | 0.199  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2     | 0.080  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2          | 0.000  |
|                                             |        | Interferent Detect Test # (Repeat) 2 | INT    |
|                                             |        | Alcohol Free Test # 3                | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                 | 0.050  |
|                                             |        | 0.08 g/210L Test # 3                 | 0.081  |
|                                             |        | 0.20 g/210L Test # 3                 | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3     | 0.079  |
|                                             |        | Interferent Detect Test # 3          | 0.000  |
|                                             |        | Interferent Detect Test # (Repeat) 3 | INT    |

**Remarks:**

Int Det: TIGHTEN SIM

**Data Download Date/Time** 01/10/2019 12:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

|                  |                    |             |            |                 |                   |
|------------------|--------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | PINELLAS COUNTY SO | <b>Date</b> | 01/22/2018 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | BELL,DANIELLE,M    | <b>Time</b> | 15:07:47   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

**Data Download Date/Time** 01/24/2018 17:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

|                                  |                        |                                   |
|----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PINELLAS COUNTY SO | <b>Date</b> 01/24/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> BELL,DANIELLE,M | <b>Time</b> 17:11:57   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG708807 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 03/29/2019   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.082  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.082  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.077  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.077  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 01/24/2018 17:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

|                  |                    |             |            |                 |                   |
|------------------|--------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | PINELLAS COUNTY SO | <b>Date</b> | 01/08/2019 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | GRAHAM,THOMAS,J    | <b>Time</b> | 12:37:44   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

**Data Download Date/Time** 01/08/2019 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PINELLAS COUNTY SO   | <b>Date</b> 01/08/2019 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GRAHAM, THOMAS, J | <b>Time</b> 15:17:53   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000888

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Test (Std Dev)             | 0.0006 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 01/08/2019 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

|                  |                    |             |            |                 |                   |
|------------------|--------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | PINELLAS COUNTY SO | <b>Date</b> | 01/22/2018 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | BELL,DANIELLE,M    | <b>Time</b> | 15:45:36   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

**Data Download Date/Time** 01/24/2018 15:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

|                                  |                        |                                   |
|----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PINELLAS COUNTY SO | <b>Date</b> 01/24/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> BELL,DANIELLE,M | <b>Time</b> 15:33:32   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG708807 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 03/29/2019   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0008      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.083  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.206  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.082  |
|                                             |        | 0.20 g/210L Test # 4             | 0.207  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.082  |
|                                             |        | 0.20 g/210L Test # 5             | 0.206  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000889

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.083  |
| 0.20 g/210L Test # 6                   | 0.205  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.082  |
| 0.20 g/210L Test # 7                   | 0.207  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.083  |
| 0.20 g/210L Test # 8                   | 0.205  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.052  |
| 0.08 g/210L Test # 9                   | 0.083  |
| 0.20 g/210L Test # 9                   | 0.205  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.082  |
| 0.20 g/210L Test # 10                  | 0.206  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0008 |
| 0.08 g/210L Test (Std Dev)             | 0.0007 |
| 0.20 g/210L Test (Std Dev)             | 0.0014 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0006 |

#### Remarks:

Data Download Date/Time 01/24/2018 15:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

|                  |                    |             |            |                 |                   |
|------------------|--------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | PINELLAS COUNTY SO | <b>Date</b> | 01/22/2018 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | BELL,DANIELLE,M    | <b>Time</b> | 15:13:34   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

**Data Download Date/Time** 01/24/2018 14:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

|                                  |                        |                                   |
|----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PINELLAS COUNTY SO | <b>Date</b> 01/24/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> BELL,DANIELLE,M | <b>Time</b> 14:51:23   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG708807 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 03/29/2019   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.082  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 01/24/2018 14:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

|                  |                    |             |            |                 |                   |
|------------------|--------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | PINELLAS COUNTY SO | <b>Date</b> | 01/08/2019 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | GRAHAM,THOMAS,J    | <b>Time</b> | 12:36:41   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. NOT A COMPLIANCE CHECK.

**Data Download Date/Time** 01/08/2019 15:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PINELLAS COUNTY SO   | <b>Date</b> 01/08/2019 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GRAHAM, THOMAS, J | <b>Time</b> 15:06:11   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000890

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 01/08/2019 15:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

**Agency** FWC **Date** 01/29/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 14:26:58 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 507129  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 08/04/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/29/2018 15:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

**Agency** FWC **Date** 02/03/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 13:55:59 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 507129  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 08/04/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/03/2018 14:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

**Agency** FWC **Date** 03/03/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 13:57:01 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 507129  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/04/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/03/2018 14:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

**Agency** FWC **Date** 04/25/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 13:14:42 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 507129  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/04/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/25/2018 14:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

**Agency** FWC **Date** 05/09/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 12:34:50 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 507129  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/04/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/09/2018 13:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

**Agency** FWC **Date** 06/13/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 17:21:33 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 507129  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/04/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/13/2018 18:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

**Agency** FWC **Date** 07/23/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 14:18:06 **Software** 8100.27

**0.05 Lot #** 201702B **0.08 Lot #** 201909A **0.20 Lot #** 201702B **0.08 Gas Lot #** 507129  
**Expiration** 02/23/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/04/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/23/2018 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

|                                   |                        |                                   |
|-----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> FWC                 | <b>Date</b> 07/25/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> MURPHY,PATRICK,J | <b>Time</b> 11:13:20   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 07/25/2018 12:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

**Agency** FWC **Date** 07/27/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 15:54:34 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 507129  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/04/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/27/2018 16:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

**Agency** FWC **Date** 08/28/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 13:48:05 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 507129  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/04/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/28/2018 14:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

**Agency** FWC **Date** 09/13/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 17:19:23 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 507129  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 08/04/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/13/2018 18:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

**Agency** FWC **Date** 10/27/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 09:47:01 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 507129  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 08/04/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/27/2018 10:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

**Agency** FWC **Date** 11/26/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 15:42:08 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 507129  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 08/04/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 11/26/2018 16:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

**Agency** FWC **Date** 12/17/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 11:50:48 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 962587  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/04/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/17/2018 12:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

|                                  |                              |                               |                               |
|----------------------------------|------------------------------|-------------------------------|-------------------------------|
| <b>Agency</b> FWC                | <b>Date</b> 01/29/2018       | <b>Type</b> Agency Inspection |                               |
| <b>Inspector</b> MOLNAR,JARROD,T | <b>Time</b> 14:20:21         | <b>Software</b> 8100.27       |                               |
| <b>0.05 Lot #</b> 201703A        | <b>0.08 Lot #</b> 201611B    | <b>0.20 Lot #</b> 201702B     | <b>0.08 Gas Lot #</b> 201611B |
| <b>Expiration</b> 03/07/2019     | <b>Expiration</b> 11/15/2018 | <b>Expiration</b> 02/23/2019  | <b>Expiration</b> 11/15/2018  |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> No         |                               |                               |

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | No     | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.050  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.082  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.205  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.000  |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.000  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.050  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.083  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.205  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.000  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3               | INT    |

#### Remarks:

08: .08 SIM CONNECTED. .08 DRY GAS HUMAN ERROR.

**Data Download Date/Time** 01/29/2018 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

**Agency** FWC **Date** 01/29/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 15:00:55 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 477627  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/29/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.205  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.205  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.206  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/29/2018 16:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

**Agency** FWC **Date** 02/03/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 13:48:49 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201611B **0.20 Lot #** 201702B **0.08 Gas Lot #** 477627  
**Expiration** 03/07/2019 **Expiration** 11/15/2018 **Expiration** 02/23/2019 **Expiration** 06/29/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/03/2018 14:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

**Agency** FWC **Date** 03/03/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 13:51:56 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 477627  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/29/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/03/2018 14:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

**Agency** FWC **Date** 04/25/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 13:07:20 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 477627  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/29/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/25/2018 14:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

**Agency** FWC **Date** 05/09/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 12:21:09 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 477627  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/29/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.206  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/09/2018 13:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

**Agency** FWC **Date** 06/13/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 17:09:20 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 477627  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/29/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/13/2018 18:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

|                                  |                              |                               |                              |
|----------------------------------|------------------------------|-------------------------------|------------------------------|
| <b>Agency</b> FWC                | <b>Date</b> 07/23/2018       | <b>Type</b> Agency Inspection |                              |
| <b>Inspector</b> MOLNAR,JARROD,T | <b>Time</b> 14:07:05         | <b>Software</b> 8100.27       |                              |
| <b>0.05 Lot #</b> 201703A        | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201703A     | <b>0.08 Gas Lot #</b> 477627 |
| <b>Expiration</b> 03/07/2019     | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 03/07/2019  | <b>Expiration</b> 06/29/2019 |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                              |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/23/2018 15:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

|                                   |                        |                                   |  |
|-----------------------------------|------------------------|-----------------------------------|--|
| <b>Agency</b> FWC                 | <b>Date</b> 07/25/2018 | <b>Type</b> Department Inspection |  |
| <b>Inspector</b> MURPHY,PATRICK,J | <b>Time</b> 11:52:28   | <b>Software</b> 8100.27           |  |

  

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805701 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

  

|                               |                                          |                       |
|-------------------------------|------------------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Average Standard Deviation</b> 0.0004 | <b>Compliance</b> Yes |
|-------------------------------|------------------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0009 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 07/25/2018 12:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

**Agency** FWC **Date** 07/27/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 15:41:57 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 477627  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/29/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/27/2018 16:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

**Agency** FWC **Date** 08/28/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 13:36:52 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 477627  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/29/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/28/2018 14:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

**Agency** FWC **Date** 09/13/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 17:09:11 **Software** 8100.27

**0.05 Lot #** 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 477627  
**Expiration** 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/29/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/13/2018 18:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

**Agency** FWC **Date** 10/27/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 09:34:39 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 477627  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/29/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/27/2018 10:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

**Agency** FWC **Date** 11/26/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 15:33:40 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 477627  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/29/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/26/2018 16:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

**Agency** FWC **Date** 12/17/2018 **Type** Agency Inspection  
**Inspector** MOLNAR,JARROD,T **Time** 11:52:22 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 477627  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/29/2019

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/17/2018 13:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

|                                  |                        |                               |
|----------------------------------|------------------------|-------------------------------|
| <b>Agency</b> FWC                | <b>Date</b> 03/21/2018 | <b>Type</b> Agency Inspection |
| <b>Inspector</b> SHANAHAN,JAKE,L | <b>Time</b> 11:05:52   | <b>Software</b> 8100.27       |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                       |
|-----------------------------|-----------------------|
| <b>Number of Simulators</b> | <b>Compliance</b> Yes |
|-----------------------------|-----------------------|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASS TO OPERATE

**Data Download Date/Time** 03/21/2018 13:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> FWC                  | <b>Date</b> 03/21/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SHANAHAN, JAKE, L | <b>Time</b> 13:51:28   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805702 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 03/21/2018 13:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

**Agency** FWC **Date** 05/25/2018 **Type** Agency Inspection  
**Inspector** JOHNSTON,ROBERT,B **Time** 19:26:30 **Software** 8100.27

**0.05 Lot #** 201802C **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 09117080A2  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/05/2019

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSTRUMENT BACK IN SERVICE

**Data Download Date/Time** 05/25/2018 20:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

**Agency** FWC **Date** 06/27/2018 **Type** Agency Inspection  
**Inspector** JOHNSTON,ROBERT,B **Time** 20:38:27 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 09117080A2  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/05/2019

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 06/29/2018 13:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

**Agency** FWC **Date** 07/16/2018 **Type** Agency Inspection  
**Inspector** JOHNSTON,ROBERT,B **Time** 20:30:50 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 09117080A2  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/05/2019

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 07/19/2018 10:39

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

**Agency** FWC **Date** 08/30/2018 **Type** Agency Inspection  
**Inspector** JOHNSTON,ROBERT,B **Time** 10:01:57 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 09117080A2  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/05/2019

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 09/04/2018 11:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000893

|                                    |                              |                               |                                  |
|------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> FWC                  | <b>Date</b> 09/30/2018       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> JOHNSTON,ROBERT,B | <b>Time</b> 22:48:26         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201802B          | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201802G     | <b>0.08 Gas Lot #</b> 09117080A2 |
| <b>Expiration</b> 02/06/2020       | <b>Expiration</b> 09/12/2018 | <b>Expiration</b> 02/22/2020  | <b>Expiration</b> 06/05/2019     |
| <b>Number of Simulators</b> 4      | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 10/05/2018 11:50

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

|                                  |                        |                               |
|----------------------------------|------------------------|-------------------------------|
| <b>Agency</b> FWC                | <b>Date</b> 03/21/2018 | <b>Type</b> Agency Inspection |
| <b>Inspector</b> SHANAHAN,JAKE,L | <b>Time</b> 11:12:23   | <b>Software</b> 8100.27       |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                       |
|-----------------------------|-----------------------|
| <b>Number of Simulators</b> | <b>Compliance</b> Yes |
|-----------------------------|-----------------------|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASS TO OPERATE

**Data Download Date/Time** 03/21/2018 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> FWC                  | <b>Date</b> 03/21/2018 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SHANAHAN, JAKE, L | <b>Time</b> 13:36:09   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201707D    | <b>0.08 Lot #</b> 201707E    | <b>0.20 Lot #</b> 201707C    | <b>0.08 Gas Lot #</b> AG805702 |
| <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/25/2019 | <b>Expiration</b> 07/24/2019 | <b>Expiration</b> 02/26/2020   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.081  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0008 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 03/21/2018 13:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

**Agency** FWC **Date** 05/25/2018 **Type** Agency Inspection  
**Inspector** JOHNSTON,ROBERT,B **Time** 19:19:21 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 09117080A2  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/05/2019

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSTRUMENT BACK IN SERVICE

**Data Download Date/Time** 05/25/2018 20:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

**Agency** FWC **Date** 06/27/2018 **Type** Agency Inspection  
**Inspector** JOHNSTON,ROBERT,B **Time** 20:32:34 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 09117080A2  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/05/2019

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 06/29/2018 13:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

|                                    |                              |                               |                                  |
|------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> FWC                  | <b>Date</b> 07/16/2018       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> JOHNSTON,ROBERT,B | <b>Time</b> 20:32:21         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201802B          | <b>0.08 Lot #</b> 201709A    | <b>0.20 Lot #</b> 201802G     | <b>0.08 Gas Lot #</b> 09117080A2 |
| <b>Expiration</b> 02/06/2020       | <b>Expiration</b> 09/12/2019 | <b>Expiration</b> 02/22/2020  | <b>Expiration</b> 06/05/2019     |
| <b>Number of Simulators</b> 4      | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 07/19/2018 10:36

**Printed: 14 January 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

**Agency** FWC **Date** 08/30/2018 **Type** Agency Inspection  
**Inspector** JOHNSTON,ROBERT,B **Time** 10:04:43 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 09117080A2  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/05/2019

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 09/04/2018 11:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

**Agency** FWC **Date** 09/30/2018 **Type** Agency Inspection  
**Inspector** JOHNSTON,ROBERT,B **Time** 22:38:47 **Software** 8100.27

**0.05 Lot #** 201802B **0.08 Lot #** 201709A **0.20 Lot #** 201802G **0.08 Gas Lot #** 09117080A2  
**Expiration** 02/06/2020 **Expiration** 09/12/2019 **Expiration** 02/22/2020 **Expiration** 06/05/2019

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSTRUMENT IN COMPLIANCE

**Data Download Date/Time** 10/05/2018 11:34