

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency BISCAYNE NAT PARK | Date 01/09/2019 | Type Agency Inspection |
| Inspector ESPINOZA,PABLO, | Time 09:43:05 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 34917080A4 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 02/05/2020 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/09/2019 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

Agency BISCAYNE NAT PARK **Date** 01/15/2019 **Type** Department Inspection
Inspector REYES-RIVERA, DAVID, E **Time** 19:43:43 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0008 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.085 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.050 |
| | | 0.08 g/210L Test # 4 | 0.082 |
| | | 0.20 g/210L Test # 4 | 0.205 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.083 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.050 |
| | | 0.08 g/210L Test # 5 | 0.082 |
| | | 0.20 g/210L Test # 5 | 0.204 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.082 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.082 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.082 |
| 0.20 g/210L Test # 7 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.082 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.082 |
| 0.20 g/210L Test # 8 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.082 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.082 |
| 0.20 g/210L Test # 9 | 0.203 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.081 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.082 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.082 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0014 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0010 |

Remarks:

Data Download Date/Time 01/15/2019 19:47

Printed: 31 May 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency BISCAYNE NAT PARK | Date 01/18/2019 | Type Agency Inspection |
| Inspector ESPINOZA,PABLO, | Time 11:31:12 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 34917080A4 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 02/05/2020 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/18/2019 12:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency BISCAYNE NAT PARK | Date 02/22/2019 | Type Agency Inspection |
| Inspector ESPINOZA,PABLO, | Time 13:49:04 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 34917080A4 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 02/05/2020 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.046 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

A F / M A:NOT ENOUGH MA.

Data Download Date/Time 02/27/2019 10:30

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency BISCAYNE NAT PARK | Date 03/01/2019 | Type Agency Inspection |
| Inspector ESPINOZA,PABLO, | Time 10:57:53 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201811C | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 34917080A4 |
| Expiration 11/13/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 02/05/2020 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/01/2019 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency BISCAYNE NAT PARK | Date 04/17/2019 | Type Agency Inspection |
| Inspector ESPINOZA,PABLO, | Time 09:45:45 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201811C | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 34917080A4 |
| Expiration 11/13/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 02/05/2020 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/17/2019 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency CLAY COUNTY SO | Date 01/11/2019 | Type Agency Inspection |
| Inspector HARWOOD, CHARLES, M | Time 15:47:31 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 17618080A1 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 08/05/2020 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/11/2019 16:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency CLAY COUNTY SO | Date 02/22/2019 | Type Agency Inspection |
| Inspector PATRONE,RICHARD,L | Time 11:31:19 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 17618080A1 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 08/05/2020 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

MONTHLY INSPECTION

Data Download Date/Time 02/22/2019 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency CLAY COUNTY SO | Date 03/25/2019 | Type Agency Inspection |
| Inspector HARWOOD, CHARLES, M | Time 12:27:51 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 17618080A1 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 08/05/2020 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

MONTHLY INSPECTION

Data Download Date/Time 03/25/2019 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency CLAY COUNTY SO | Date 04/21/2019 | Type Agency Inspection |
| Inspector PATRONE,RICHARD,L | Time 08:32:28 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 17618080A1 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 08/05/2020 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

MONTHLY INSPECTION

Data Download Date/Time 04/21/2019 08:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency PATRICK AFB SFS **Date** 01/18/2019 **Type** Agency Inspection
Inspector OUVERSON,JESSE,T **Time** 16:03:13 **Software** 8100.27

0.05 Lot # 201802B **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration 02/06/2020 **Expiration** **Expiration** **Expiration**

Number of Simulators 4

Compliance No

| Check or Test | Result | Test | Result |
|---|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.043 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.044 |
| Barometric Pressure Sensor Check: OK | | Interferent Detect Test # 1 | INT |
| Alcohol Free Subject Test: 0.000 | Yes | Alcohol Free Test # 2 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.05 g/210L Test # 2 | 0.043 |
| Interferent Detect Test: Interferent Detect | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.045 |
| Diagnostic Check (Post-Inspection): OK | No | Interferent Detect Test # 2 | INT |
| Alcohol Free Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| 0.05 g/210L Test: OK | No | 0.05 g/210L Test # 3 | 0.045 |
| 0.08 g/210L Test: OK | No | 0.05 g/210L Test # (Repeat) 3 | 0.045 |
| 0.20 g/210L Test: OK | No | Interferent Detect Test # 3 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Data Download Date/Time 01/28/2019 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency PATRICK AFB SFS | Date 01/22/2019 | Type Agency Inspection |
| Inspector OUVERSON,JESSE,T | Time 17:12:44 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 894572 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 08/20/2019 |

Number of Simulators 4 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.191 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.193 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/28/2019 13:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency PATRICK AFB SFS | Date 02/28/2019 | Type Agency Inspection |
| Inspector OUVERSON,JESSE,T | Time 12:33:43 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 894572 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 07/21/2021 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 4 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/28/2019 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

| | | | |
|-----------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency PATRICK AFB SFS | Date 04/22/2019 | Type Agency Inspection | |
| Inspector OUVERSON,JESSE,T | Time 11:38:45 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 894572 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 08/12/2019 |
| Number of Simulators 4 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/22/2019 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency TYNDALL AFB | Date 01/03/2019 | Type Agency Inspection |
| Inspector MURPHY,PATRICK,J | Time 09:34:17 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

TESTING FORMS, NOT AN INSPECTION

Data Download Date/Time 01/04/2019 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.081 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0007 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 01/04/2019 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

| | | | |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency TYNDALL AFB | Date 01/11/2019 | Type Agency Inspection | |
| Inspector BARGE,SUSAN,M | Time 21:20:48 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 970670 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 06/07/2021 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.083 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

NA

Data Download Date/Time 01/11/2019 22:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

| | | | |
|-------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency TYNDALL AFB | Date 02/12/2019 | Type Agency Inspection | |
| Inspector LORD,LANNY,A | Time 09:39:32 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 970670 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 06/07/2021 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|---|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.082 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

08: Control Outside Tolerance RECONNECT.

Data Download Date/Time 02/12/2019 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

| | | |
|-------------------------------|------------------------|-------------------------------|
| Agency TYNDALL AFB | Date 03/13/2019 | Type Agency Inspection |
| Inspector LORD,LANNY,A | Time 10:36:18 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 970670 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 06/07/2021 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/13/2019 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

| | | | |
|-------------------------------|------------------------------|------------------------------|------------------------------|
| Agency TYNDALL AFB | Date 04/16/2019 | Type Agency Inspection | |
| Inspector LORD,LANNY,A | Time 09:07:54 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 970670 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 06/07/2021 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.083 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.083 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/16/2019 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency CAPE CORAL PD | Date 01/09/2019 | Type Agency Inspection |
| Inspector CLAFLIN, DAVID, S | Time 11:12:41 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 05218080A3 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 05/05/2020 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/09/2019 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency CAPE CORAL PD | Date 02/15/2019 | Type Agency Inspection |
| Inspector CLAFLIN, DAVID, S | Time 09:44:07 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 05218080A3 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 05/05/2020 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/15/2019 09:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency CAPE CORAL PD | Date 03/06/2019 | Type Agency Inspection |
| Inspector CLAFLIN, DAVID, S | Time 12:10:00 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 05218080A3 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 05/05/2020 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/06/2019 12:14

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency CAPE CORAL PD | Date 04/17/2019 | Type Agency Inspection |
| Inspector CLAFLIN, DAVID, S | Time 10:37:08 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 28818080A1 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 12/05/2020 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/17/2019 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency KENNETH CITY PD | Date 01/29/2019 | Type Agency Inspection |
| Inspector VIENO,LT. MIKE,J | Time 18:56:29 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 06717080A7 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 05/05/2019 |

Number of Simulators 3

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.189 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.192 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

MOISTURE IN CLEAR SIM. TUBING SWITCHED TUBING.RETEST

Data Download Date/Time 01/29/2019 20:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency KENNETH CITY PD | Date 02/27/2019 | Type Agency Inspection |
| Inspector VIENO,LT. MICHAEL,J | Time 18:23:06 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|-------------------|-----------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration | Expiration |

Number of Simulators 3

Compliance No

| Check or Test | Result | Test | Result |
|---|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.072 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # (Repeat) 1 | 0.073 |
| Alcohol Free Subject Test: 0.000 | Yes | Interferent Detect Test # 1 | INT |
| Mouth Alcohol Test: Slope Not Met | Yes | Alcohol Free Test # 2 | 0.000 |
| Interferent Detect Test: Interferent Detect | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Diagnostic Check (Post-Inspection): OK | No | 0.08 g/210L Test # 2 | 0.074 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # (Repeat) 2 | 0.073 |
| 0.05 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Test: OK | No | Alcohol Free Test # 3 | 0.000 |
| 0.20 g/210L Test: OK | No | 0.05 g/210L Test # 3 | 0.049 |
| 0.08 g/210L Dry Gas Std Test: OK | No | 0.08 g/210L Test # 3 | 0.074 |
| | | 0.08 g/210L Test # (Repeat) 3 | 0.074 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

08: MOISTURE IN CLR SIM TUB CHANGED TUBING . Non-compliance: .

Data Download Date/Time 03/07/2019 08:59

Printed: 31 May 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency KENNETH CITY PD | Date 03/07/2019 | Type Agency Inspection |
| Inspector GRAHAM, THOMAS, J | Time 06:42:47 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

NOT A COMPLIANCE CHECK. BYPASSED AI TO OPERATE INSTRUMENT.

Data Download Date/Time 03/07/2019 08:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

Agency KENNETH CITY PD **Date** 03/07/2019 **Type** Department Inspection
Inspector GRAHAM, THOMAS, J **Time** 08:51:22 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.052 |
| | | 0.08 g/210L Test # 4 | 0.081 |
| | | 0.20 g/210L Test # 4 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.052 |
| | | 0.08 g/210L Test # 5 | 0.081 |
| | | 0.20 g/210L Test # 5 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.052 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.082 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 03/07/2019 08:59

Printed: 31 May 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

| | | |
|---------------------------------------|------------------------|-------------------------------|
| Agency COLLIER COUNTY SO | Date 01/02/2019 | Type Agency Inspection |
| Inspector NEIDERHISER, JOHN, E | Time 10:15:30 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 19817080A1 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 09/05/2019 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|---|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.042 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.050 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.064 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.200 |
| Interferent Detect Test: Interferent Detect | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.076 |
| Alcohol Free Test: OK | Yes | Interferent Detect Test # 1 | INT |
| 0.05 g/210L Test: OK | Yes | Alcohol Free Test # 2 | 0.000 |
| 0.08 g/210L Test: OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| 0.20 g/210L Test: OK | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.050 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| | | 0.20 g/210L Test # 2 | 0.105 |
| | | 0.20 g/210L Test # (Repeat) 2 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.077 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.05 g/210L Test # (Repeat) 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.000 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

05: Control Outside Tolerance LOOSE CONNECTION. 20: Control Outside Tolerance LOOSE CONNECTION. 08: Control Outside Tolerance.

**Florida Department of Law Enforcement
Alcohol Testing Program**

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Data Download Date/Time 01/02/2019 11:14

Printed: 31 May 2019
COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

| | | |
|---------------------------------------|------------------------|-------------------------------|
| Agency COLLIER COUNTY SO | Date 01/02/2019 | Type Agency Inspection |
| Inspector NEIDERHISER, JOHN, E | Time 11:05:34 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 19817080A1 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 09/05/2019 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

TESTED AT SPEC OPS

Data Download Date/Time 01/02/2019 11:14

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO **Date** 02/04/2019 **Type** Agency Inspection
Inspector NEIDERHISER,JOHN,E **Time** 09:55:27 **Software** 8100.27

0.05 Lot # 201703A **0.08 Lot #** 201709A **0.20 Lot #** 201702B **0.08 Gas Lot #** 10818080A6
Expiration 03/07/2019 **Expiration** 09/12/2019 **Expiration** 02/23/2019 **Expiration** 06/05/2020

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/04/2019 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO **Date** 02/05/2019 **Type** Department Inspection
Inspector REYES-RIVERA, DAVID, E **Time** 08:42:27 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.080 |
| | | 0.20 g/210L Test # 4 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.049 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 02/05/2019 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency COLLIER COUNTY SO | Date 02/05/2019 | Type Agency Inspection |
| Inspector NEIDERHISER,JOHN,E | Time 15:13:27 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 19817080A1 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 09/05/2019 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|---|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.072 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.076 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.073 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.076 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

08: Control Outside Tolerance LOW PSI.

Data Download Date/Time 02/05/2019 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency COLLIER COUNTY SO | Date 03/04/2019 | Type Agency Inspection |
| Inspector NEIDERHISER,JOHN,E | Time 13:20:32 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 31517080A1 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 01/05/2020 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

TESTED AT SPEC OPS

Data Download Date/Time 03/04/2019 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

| | | |
|---------------------------------------|------------------------|-------------------------------|
| Agency COLLIER COUNTY SO | Date 04/02/2019 | Type Agency Inspection |
| Inspector NEIDERHISER, JOHN, E | Time 10:45:40 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201811C | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 31517080A1 |
| Expiration 11/13/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 01/05/2020 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

TESTED AT SPEC OPS

Data Download Date/Time 04/02/2019 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

| | | |
|---------------------------------------|------------------------|-------------------------------|
| Agency ST PETERSBURG PD | Date 01/15/2019 | Type Agency Inspection |
| Inspector WEISKOPF, MICHAEL, D | Time 08:25:23 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 13518080A5 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 08/05/2020 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.051 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/15/2019 08:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

| | | | |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency MIAMI PD | Date 01/29/2019 | Type Agency Inspection | |
| Inspector PASTOR,JOSE,F | Time 11:32:53 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 620051 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 11/05/2019 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/29/2019 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

| | | | |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency MIAMI PD | Date 02/27/2019 | Type Agency Inspection | |
| Inspector PASTOR,JOSE,F | Time 14:50:35 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 620051 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 11/05/2019 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/27/2019 15:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

| | | |
|--------------------------------|------------------------------|-------------------------------|
| Agency MIAMI PD | Date 03/27/2019 | Type Agency Inspection |
| Inspector PASTOR,JOSE,F | Time 10:07:15 | Software 8100.27 |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 |
| | | 0.08 Gas Lot # 620051 |
| | | Expiration 11/05/2019 |
| Number of Simulators 5 | | Compliance Yes |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/27/2019 11:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

| | | |
|-------------------------------|------------------------|-------------------------------|
| Agency MIAMI PD | Date 04/29/2019 | Type Agency Inspection |
| Inspector PASTOR,JOSE, | Time 13:26:19 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 620051 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 11/05/2019 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

OFC FERNANDEZ IBM 28394 PRESENT

Data Download Date/Time 05/02/2019 09:30

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

| | | | |
|------------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency ST JOHNS COUNTY SO | Date 01/10/2019 | Type Agency Inspection | |
| Inspector MCGINNIS,JOSEPH,D | Time 19:36:13 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 366849 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 03/30/2019 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/10/2019 19:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

| | | | |
|------------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency ST JOHNS COUNTY SO | Date 02/26/2019 | Type Agency Inspection | |
| Inspector MCGINNIS,JOSEPH,D | Time 14:26:48 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 640976 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 11/21/2019 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/26/2019 14:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

| | | | |
|-------------------------------|------------------------------|------------------------------|------------------------------|
| Agency ST JOHNS COUNTY SO | Date 03/26/2019 | Type Agency Inspection | |
| Inspector MCGINNIS,JOSEPH,D | Time 11:41:41 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 640976 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 11/21/2019 |
| Number of Simulators 5 | Compliance Yes | | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/26/2019 11:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

| | | | |
|------------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency ST JOHNS COUNTY SO | Date 04/18/2019 | Type Agency Inspection | |
| Inspector MCGINNIS,JOSEPH,D | Time 14:43:27 | Software 8100.27 | |
| 0.05 Lot # 201811C | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 640976 |
| Expiration 11/13/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 11/21/2019 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.204 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/18/2019 14:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency NORTH MIAMI PD | Date 01/02/2019 | Type Agency Inspection |
| Inspector DE JESUS, RAY, | Time 09:52:54 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 32316080A3 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 01/05/2019 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/02/2019 10:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD **Date** 01/07/2019 **Type** Department Inspection
Inspector REYES-RIVERA,DAVID,E **Time** 10:05:35 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.051 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.051 |
| | | 0.08 g/210L Test # 4 | 0.083 |
| | | 0.20 g/210L Test # 4 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.051 |
| | | 0.08 g/210L Test # 5 | 0.082 |
| | | 0.20 g/210L Test # 5 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.083 |
| 0.20 g/210L Test # 6 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.083 |
| 0.20 g/210L Test # 7 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.081 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.052 |
| 0.08 g/210L Test # 8 | 0.083 |
| 0.20 g/210L Test # 8 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.084 |
| 0.20 g/210L Test # 9 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.083 |
| 0.20 g/210L Test # 10 | 0.202 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0006 |
| 0.20 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 01/07/2019 10:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

| | | | |
|---------------------------------|------------------------------|-------------------------------|----------------------------------|
| Agency NORTH MIAMI PD | Date 01/09/2019 | Type Agency Inspection | |
| Inspector DE JESUS, RAY, | Time 15:43:12 | Software 8100.27 | |
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 34917080A4 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 02/05/2020 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/09/2019 15:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency NORTH MIAMI PD | Date 02/07/2019 | Type Agency Inspection |
| Inspector DE JESUS, RAY, | Time 12:13:28 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 34418080A3 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 02/05/2021 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/07/2019 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency NORTH MIAMI PD | Date 03/07/2019 | Type Agency Inspection |
| Inspector DE JESUS, RAY, | Time 12:46:28 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 34418080A3 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 02/05/2021 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/07/2019 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency NORTH MIAMI PD | Date 04/23/2019 | Type Agency Inspection |
| Inspector DE JESUS, RAY, | Time 15:44:39 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 34418080A3 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 02/05/2021 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/23/2019 15:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency MICCOSUKEE PD | Date 01/10/2019 | Type Agency Inspection |
| Inspector ESPINOZA,PABLO, | Time 12:40:33 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 34917080A2 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 02/05/2020 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.082 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

A F / M A:NOT ENOUGH MA.

Data Download Date/Time 01/10/2019 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

| | | |
|----------------------------------|-----------------|----------------------------|
| Agency MICCOSUKEE PD | Date 01/15/2019 | Type Department Inspection |
| Inspector REYES-RIVERA, DAVID, E | Time 15:52:27 | Software 8100.27 |

| | | | |
|-----------------------|-----------------------|-----------------------|-------------------------|
| 0.05 Lot # 201707D | 0.08 Lot # 201707E | 0.20 Lot # 201707C | 0.08 Gas Lot # AG805701 |
| Expiration 07/25/2019 | Expiration 07/25/2019 | Expiration 07/24/2019 | Expiration 02/26/2020 |

Number of Simulators 5 Average Standard Deviation 0.0003 Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.050 |
| | | 0.08 g/210L Test # 4 | 0.082 |
| | | 0.20 g/210L Test # 4 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.051 |
| | | 0.08 g/210L Test # 5 | 0.081 |
| | | 0.20 g/210L Test # 5 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.082 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.082 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.082 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.082 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 01/15/2019 16:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency MICCOSUKEE PD | Date 01/17/2019 | Type Agency Inspection |
| Inspector ESPINOZA,PABLO, | Time 12:33:10 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 34917080A2 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 02/05/2020 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/17/2019 12:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency MICCOSUKEE PD | Date 02/01/2019 | Type Agency Inspection |
| Inspector ESPINOZA,PABLO, | Time 09:41:49 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 34917080A2 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 02/05/2020 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.194 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.187 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.187 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

20: Control Outside Tolerance/TIGHTENED SIM.

Data Download Date/Time 02/01/2019 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency MICCOSUKEE PD | Date 03/27/2019 | Type Agency Inspection |
| Inspector ESPINOZA,PABLO, | Time 09:37:02 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201811C | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 34917080A2 |
| Expiration 11/13/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 02/05/2020 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/27/2019 09:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001657

| | | | |
|-------------------------------|------------------------------|------------------------------|----------------------------------|
| Agency MICCOSUKEE PD | Date 04/30/2019 | Type Agency Inspection | |
| Inspector ESPINOZA,PABLO, | Time 08:57:02 | Software 8100.27 | |
| 0.05 Lot # 201811C | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 34917080A2 |
| Expiration 11/13/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 02/05/2020 |
| Number of Simulators 5 | Compliance Yes | | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 05/01/2019 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001714

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency UNIVERSITY OF FL PD | Date 01/31/2019 | Type Agency Inspection |
| Inspector PETERSON, KYLE, A | Time 11:52:10 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201702B | 0.08 Gas Lot # 22817080A4 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/23/2019 | Expiration 10/05/2019 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/31/2019 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001714

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency UNIVERSITY OF FL PD | Date 02/28/2019 | Type Agency Inspection |
| Inspector PETERSON, KYLE, A | Time 11:21:35 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 22817080A4 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 10/05/2019 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/28/2019 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001714

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency UNIVERSITY OF FL PD | Date 03/29/2019 | Type Agency Inspection |
| Inspector PETERSON, KYLE, A | Time 13:58:14 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 22817080A4 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 10/05/2019 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/29/2019 14:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001714

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency UNIVERSITY OF FL PD | Date 04/30/2019 | Type Agency Inspection |
| Inspector PETERSON, KYLE, A | Time 10:26:00 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 22817080A4 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 10/05/2019 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/30/2019 10:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

| | | | |
|-----------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency OSCEOLA COUNTY S.O. | Date 01/07/2019 | Type Agency Inspection | |
| Inspector LYONS,DANIEL,E | Time 00:38:14 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 646645 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 12/05/2019 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.045 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.046 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/07/2019 00:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency OSCEOLA COUNTY S.O. | Date 01/24/2019 | Type Agency Inspection |
| Inspector LYONS,DANIEL,E | Time 22:03:51 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 646645 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 12/05/2019 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.192 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/24/2019 22:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency OSCEOLA COUNTY S.O. **Date** 02/01/2019 **Type** Department Inspection
Inspector GRAHAM,THOMAS,J **Time** 10:59:51 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.050 |
| | | 0.08 g/210L Test # 4 | 0.080 |
| | | 0.20 g/210L Test # 4 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.050 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0006 |
| 0.20 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

Remarks:

Data Download Date/Time 02/01/2019 11:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

| | | | |
|-----------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency OSCEOLA COUNTY S.O. | Date 02/07/2019 | Type Agency Inspection | |
| Inspector LYONS,DANIEL,E | Time 19:19:53 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 646645 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 12/05/2019 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.193 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/07/2019 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency OSCEOLA COUNTY S.O. | Date 03/18/2019 | Type Agency Inspection |
| Inspector LYONS,DANIEL,E | Time 22:33:44 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 646645 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 12/05/2019 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.041 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.047 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.192 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.043 |
| 0.05 g/210L Test: OK | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.047 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.05 g/210L Test # (Repeat) 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

05: Control Outside Tolerance.

Data Download Date/Time 03/18/2019 22:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

| | | | |
|-----------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency OSCEOLA COUNTY S.O. | Date 04/03/2019 | Type Agency Inspection | |
| Inspector LYONS,DANIEL,E | Time 20:41:08 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 646645 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 12/05/2019 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/03/2019 19:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

| | | | |
|---------------------------------|------------------------------|-------------------------------|--------------------------------|
| Agency ALACHUA PD | Date 01/14/2019 | Type Agency Inspection | |
| Inspector PARKS,DONALD,M | Time 00:49:53 | Software 8100.27 | |
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # AG805701 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 02/26/2020 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/14/2019 00:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

| | | | |
|---------------------------------|------------------------------|-------------------------------|--------------------------------|
| Agency ALACHUA PD | Date 02/24/2019 | Type Agency Inspection | |
| Inspector PARKS,DONALD,M | Time 23:31:27 | Software 8100.27 | |
| 0.05 Lot # 201703A | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # AG805701 |
| Expiration 03/07/2019 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 02/26/2020 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/24/2019 23:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

| | | | |
|---------------------------------|------------------------------|-------------------------------|--------------------------------|
| Agency ALACHUA PD | Date 03/28/2019 | Type Agency Inspection | |
| Inspector PARKS,DONALD,M | Time 03:56:09 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # AG805701 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 02/26/2020 |
| Number of Simulators 5 | Compliance Yes | | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/28/2019 04:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency ALACHUA PD | Date 04/30/2019 | Type Agency Inspection |
| Inspector PARKS,DONALD,M | Time 01:55:25 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # AG805701 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 02/26/2020 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/30/2019 06:01

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

| | | | |
|---------------------------------|------------------------|-------------------------------|--|
| Agency PALM BAY P.D. | Date 01/27/2019 | Type Agency Inspection | |
| Inspector LUTZ, EDWIN, J | Time 22:26:31 | Software 8100.27 | |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 22817080A4 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 10/05/2019 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.191 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/27/2019 22:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency PALM BAY P.D. | Date 02/26/2019 | Type Agency Inspection |
| Inspector LUTZ, EDWIN, J | Time 10:54:39 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 22817080A4 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 10/05/2019 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/26/2019 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

| | | | |
|---------------------------------|------------------------|-------------------------------|--|
| Agency PALM BAY P.D. | Date 03/28/2019 | Type Agency Inspection | |
| Inspector LUTZ, EDWIN, J | Time 18:14:37 | Software 8100.27 | |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 22817080A4 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 10/05/2019 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.193 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/28/2019 18:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

| | | | |
|--------------------------|-----------------|------------------------|--|
| Agency PALM BAY P.D. | Date 04/28/2019 | Type Agency Inspection | |
| Inspector LUTZ, EDWIN, J | Time 15:24:06 | Software 8100.27 | |

| | | | |
|-----------------------|-----------------------|-----------------------|---------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 24818080A2 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 10/05/2020 |

| | |
|------------------------|----------------|
| Number of Simulators 5 | Compliance Yes |
|------------------------|----------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/28/2019 15:30

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

Agency PINELLAS PARK P.D. Date 01/04/2019 Type Agency Inspection
 Inspector KOLBICKA,LAWRENCE,R Time 10:53:03 Software 8100.27

0.05 Lot # 201703A 0.08 Lot # 201709A 0.20 Lot # 201702B 0.08 Gas Lot # 05218080A3
 Expiration 03/07/2019 Expiration 09/12/2019 Expiration 02/23/2019 Expiration 05/05/2020

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/04/2019 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency PINELLAS PARK P.D. | Date 03/13/2019 | Type Agency Inspection |
| Inspector KOLBICKA,LAWRENCE,R | Time 15:30:18 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 05218080A3 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 05/05/2020 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/13/2019 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

| | | |
|--|------------------------|-------------------------------|
| Agency PINELLAS PARK P.D. | Date 04/04/2019 | Type Agency Inspection |
| Inspector KOLBICKA, LAWRENCE, R | Time 15:03:15 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 05218080A3 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 05/05/2020 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|---|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

08: Ambient Fail HOSE NOT CONNECTED.

Data Download Date/Time 04/04/2019 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

| | | | |
|----------------------------------|------------------------------|-------------------------------|--------------------------------|
| Agency SOUTH MIAMI PD | Date 01/11/2019 | Type Agency Inspection | |
| Inspector ESPINOZA,PABLO, | Time 10:50:51 | Software 8100.27 | |
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # AG827002 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 09/27/2020 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/11/2019 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

Agency SOUTH MIAMI PD **Date** 01/15/2019 **Type** Department Inspection
Inspector REYES-RIVERA, DAVID, E **Time** 14:49:17 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.081 |
| | | 0.20 g/210L Test # 4 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.081 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.049 |
| | | 0.08 g/210L Test # 5 | 0.081 |
| | | 0.20 g/210L Test # 5 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.082 |
| | | Interferent Detect Test # 5 | INT |

**Florida Department of Law Enforcement
Alcohol Testing Program**

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.082 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.082 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.081 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.082 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.082 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

Remarks:

Data Download Date/Time 01/15/2019 14:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

| | | | |
|-------------------------------|------------------------------|------------------------------|--------------------------------|
| Agency SOUTH MIAMI PD | Date 01/17/2019 | Type Agency Inspection | |
| Inspector ESPINOZA,PABLO, | Time 10:53:34 | Software 8100.27 | |
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # AG827002 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 09/27/2020 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/17/2019 11:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency SOUTH MIAMI PD | Date 02/15/2019 | Type Agency Inspection |
| Inspector ESPINOZA,PABLO, | Time 09:35:05 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # AG827002 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 09/27/2020 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.046 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/15/2019 09:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

| | | | |
|-------------------------------|------------------------------|------------------------------|--------------------------------|
| Agency SOUTH MIAMI PD | Date 03/26/2019 | Type Agency Inspection | |
| Inspector ESPINOZA,PABLO, | Time 09:40:28 | Software 8100.27 | |
| 0.05 Lot # 201811C | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # AG827002 |
| Expiration 11/13/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 09/27/2020 |
| Number of Simulators 5 | Compliance Yes | | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.203 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/26/2019 10:14

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

| | | | |
|----------------------------------|------------------------------|-------------------------------|--------------------------------|
| Agency SOUTH MIAMI PD | Date 04/26/2019 | Type Agency Inspection | |
| Inspector ESPINOZA,PABLO, | Time 12:43:52 | Software 8100.27 | |
| 0.05 Lot # 201811C | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # AG827002 |
| Expiration 11/13/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 09/27/2020 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

AI RESTARTED DUE TO POWER LOSS

Data Download Date/Time 04/26/2019 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency CHARLOTTE COUNTY SO | Date 02/07/2019 | Type Agency Inspection |
| Inspector PLATT,SHAYLA,D | Time 08:47:05 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Data Download Date/Time 02/07/2019 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency CHARLOTTE COUNTY SO **Date** 02/07/2019 **Type** Department Inspection
Inspector PLATT,SHAYLA,D **Time** 10:48:32 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0008 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.049 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.078 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0009 |
| 0.08 g/210L Test (Std Dev) | 0.0007 |
| 0.20 g/210L Test (Std Dev) | 0.0013 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

Remarks:

Data Download Date/Time 02/07/2019 10:52

Printed: 31 May 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency CHARLOTTE COUNTY SO | Date 02/27/2019 | Type Agency Inspection |
| Inspector KERN, MICHAEL, P | Time 07:14:00 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

A F / M A: .

Data Download Date/Time 04/29/2019 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency CHARLOTTE COUNTY SO | Date 02/27/2019 | Type Agency Inspection |
| Inspector KERN, MICHAEL, P | Time 07:52:34 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 10818080A5 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 06/05/2020 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 04/29/2019 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency CHARLOTTE COUNTY SO | Date 03/28/2019 | Type Agency Inspection |
| Inspector KERN, MICHAEL, P | Time 08:29:00 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 10818080A5 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 06/05/2020 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 04/29/2019 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency CHARLOTTE COUNTY SO | Date 04/29/2019 | Type Agency Inspection |
| Inspector KERN, MICHAEL, P | Time 09:16:58 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201709A | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 10818080A5 |
| Expiration 09/12/2019 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 06/05/2020 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/29/2019 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency CLEARWATER PD | Date 01/22/2019 | Type Agency Inspection |
| Inspector NUGENT, DAVID, A | Time 19:18:56 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 02618080A2 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 03/05/2020 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

MOUTH ALCOHOL 2017-B ACETONE 2018-A

Data Download Date/Time 01/22/2019 19:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

| | | |
|--------------------------------------|------------------------|-------------------------------|
| Agency CLEARWATER PD | Date 02/20/2019 | Type Agency Inspection |
| Inspector NEGERSMITH,DANIEL,M | Time 20:27:57 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201811C | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 02618080A2 |
| Expiration 11/13/2020 | Expiration 09/12/2019 | Expiration 02/20/2020 | Expiration 03/05/2020 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

ACETONE LOT2018-A MOUTH ALC SOL LOT2017-B

Data Download Date/Time 02/20/2019 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency CLEARWATER PD **Date** 02/27/2019 **Type** Department Inspection
Inspector GRAHAM,THOMAS,J **Time** 09:16:46 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.049 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.079 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

Remarks:

Data Download Date/Time 02/27/2019 09:21

Printed: 31 May 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

| | | | |
|-------------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency PALM BEACH CO SO | Date 01/18/2019 | Type Agency Inspection | |
| Inspector KARLECKE, JASON, C | Time 13:05:25 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 646645 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 12/05/2019 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/18/2019 13:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

| | | | |
|-------------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency PALM BEACH CO SO | Date 02/08/2019 | Type Agency Inspection | |
| Inspector KARLECKE, JASON, C | Time 13:08:54 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 646645 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 12/05/2019 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.052 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.052 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/08/2019 13:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency PALM BEACH CO SO | Date 03/15/2019 | Type Agency Inspection |
| Inspector KARLECKE, JASON, C | Time 13:38:08 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 646645 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 12/05/2019 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.051 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/15/2019 13:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency PALM BEACH CO SO | Date 04/19/2019 | Type Agency Inspection |
| Inspector KARLECKE, JASON, C | Time 13:49:03 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 646645 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 12/05/2019 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/19/2019 14:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

| | | |
|---------------------------------|------------------------|-----------------------------------|
| Agency PALM BEACH CO SO | Date 04/30/2019 | Type Department Inspection |
| Inspector PLATT,SHAYLA,D | Time 11:16:50 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| 0.05 Lot # 201707D | 0.08 Lot # 201707E | 0.20 Lot # 201707C | 0.08 Gas Lot # AG831804 |
| Expiration 07/25/2019 | Expiration 07/25/2019 | Expiration 07/24/2019 | Expiration 11/14/2020 |

| | | |
|-------------------------------|--|-----------------------|
| Number of Simulators 5 | Average Standard Deviation 0.0004 | Compliance Yes |
|-------------------------------|--|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.050 |
| | | 0.08 g/210L Test # 4 | 0.080 |
| | | 0.20 g/210L Test # 4 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.049 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 04/30/2019 11:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

| | | | |
|---------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency MANATEE COUNTY SO | Date 01/21/2019 | Type Agency Inspection | |
| Inspector CLINE, GLENN, | Time 12:36:25 | Software 8100.27 | |
| 0.05 Lot # 201703A | 0.08 Lot # 201808E | 0.20 Lot # 201702B | 0.08 Gas Lot # 867139 |
| Expiration 03/07/2019 | Expiration 08/22/2020 | Expiration 02/23/2019 | Expiration 11/15/2020 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/21/2019 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

| | | | |
|---------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency MANATEE COUNTY SO | Date 02/26/2019 | Type Agency Inspection | |
| Inspector CLINE, GLENN, | Time 11:38:44 | Software 8100.27 | |
| 0.05 Lot # 201703A | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 867139 |
| Expiration 03/07/2019 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 11/15/2020 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/26/2019 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

| | | | |
|---------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency MANATEE COUNTY SO | Date 03/25/2019 | Type Agency Inspection | |
| Inspector CLINE,GLENN, | Time 08:08:52 | Software 8100.27 | |
| 0.05 Lot # 201811C | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 840875 |
| Expiration 11/13/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 09/19/2020 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/25/2019 08:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

| | | | |
|---------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency MANATEE COUNTY SO | Date 04/23/2019 | Type Agency Inspection | |
| Inspector CLINE,GLENN, | Time 16:07:58 | Software 8100.27 | |
| 0.05 Lot # 201811C | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 840875 |
| Expiration 11/13/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 09/19/2020 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/23/2019 16:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency LAKELAND PD | Date 01/14/2019 | Type Agency Inspection |
| Inspector FRANZ,GORDON,L | Time 10:13:12 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 22817080A5 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 10/05/2019 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.074 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # (Repeat) 1 | 0.077 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.193 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.075 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # (Repeat) 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.08 g/210L Test # (Repeat) 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 01/14/2019 10:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency LAKELAND PD | Date 02/18/2019 | Type Agency Inspection |
| Inspector FRANZ,GORDON,L | Time 10:08:03 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 22817080A5 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 10/05/2019 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/18/2019 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.051 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.051 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.051 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.078 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.051 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.051 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

Remarks:

Data Download Date/Time 02/25/2019 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency LAKELAND PD | Date 03/25/2019 | Type Agency Inspection |
| Inspector FRANZ,GORDON,L | Time 11:19:17 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 22817080A5 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 10/05/2019 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/25/2019 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency LAKELAND PD | Date 04/24/2019 | Type Agency Inspection |
| Inspector FRANZ,GORDON,L | Time 10:45:48 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 2018O2B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 22817080A5 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 10/05/2019 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/24/2019 10:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency PORT ST LUCIE PD | Date 01/31/2019 | Type Agency Inspection |
| Inspector HARRIS, IAN, | Time 19:20:25 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 19817080A1 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 09/05/2019 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 4 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.075 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/31/2019 19:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

| | | | |
|--------------------------------|------------------------|-------------------------------|--|
| Agency PORT ST LUCIE PD | Date 02/25/2019 | Type Agency Inspection | |
| Inspector HARRIS, IAN, | Time 13:47:01 | Software 8100.27 | |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201811C | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 19817080A1 |
| Expiration 11/13/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 09/05/2019 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 4 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.045 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.075 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.075 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/25/2019 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

| | | | |
|--------------------------------|------------------------|-------------------------------|--|
| Agency PORT ST LUCIE PD | Date 03/28/2019 | Type Agency Inspection | |
| Inspector HARRIS, IAN, | Time 18:32:57 | Software 8100.27 | |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201811C | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 19817080A1 |
| Expiration 11/13/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 09/05/2019 |

Number of Simulators 4

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/28/2019 18:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency PORT ST LUCIE PD | Date 04/29/2019 | Type Agency Inspection |
| Inspector HARRIS, IAN, | Time 18:18:27 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201811C | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 03519080A4 |
| Expiration 11/13/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 04/05/2021 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 4 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/29/2019 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency NAS JACKSONVILLE PD | Date 01/14/2019 | Type Agency Inspection |
| Inspector JUSINO,LUIS,H | Time 20:23:00 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 03417080A1 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 03/05/2019 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/14/2019 20:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency NAS JACKSONVILLE PD | Date 02/24/2019 | Type Agency Inspection |
| Inspector GARDZALLA,ROBERT,S | Time 01:45:10 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 03417080A1 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 03/05/2019 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.045 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.045 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/24/2019 01:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency NAS JACKSONVILLE PD | Date 03/23/2019 | Type Agency Inspection |
| Inspector WRIGHT,ANDREW,E | Time 05:59:21 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 00919080A3 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 03/05/2021 |

Number of Simulators 4

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.045 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.045 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.046 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/23/2019 06:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency NAS JACKSONVILLE PD | Date 04/21/2019 | Type Agency Inspection |
| Inspector GARDZALLA,ROBERT,S | Time 07:41:35 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 00919080A3 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 03/05/2021 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/27/2019 00:30

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

| | | |
|---|------------------------|-------------------------------|
| Agency NAS KEY WEST | Date 01/02/2019 | Type Agency Inspection |
| Inspector REYES-RIVERA, DAVID, E | Time 05:50:24 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED

Data Download Date/Time 01/02/2019 06:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

Agency NAS KEY WEST **Date** 01/02/2019 **Type** Department Inspection
Inspector REYES-RIVERA, DAVID, E **Time** 10:09:58 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG805701
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 02/26/2020

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.050 |
| | | 0.08 g/210L Test # 4 | 0.081 |
| | | 0.20 g/210L Test # 4 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.050 |
| | | 0.08 g/210L Test # 5 | 0.081 |
| | | 0.20 g/210L Test # 5 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.078 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.082 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.078 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0007 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

Remarks:

Data Download Date/Time 01/02/2019 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency NAS KEY WEST | Date 03/07/2019 | Type Agency Inspection |
| Inspector ANDERSON,LEE,B | Time 14:45:17 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 792976 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 06/06/2020 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|---|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

08: Control Outside Tolerance.TUBE NOT CONNECTED, PASSED 3

Data Download Date/Time 03/07/2019 14:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency NEW SMYRNA BEACH PD | Date 01/27/2019 | Type Agency Inspection |
| Inspector REVE,JASON,R | Time 08:31:24 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 10818080A5 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 06/05/2020 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/27/2019 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency NEW SMYRNA BEACH PD | Date 02/28/2019 | Type Agency Inspection |
| Inspector REVE,JASON,R | Time 19:27:25 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 10818080A5 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/05/2020 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/28/2019 19:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency NEW SMYRNA BEACH PD | Date 03/31/2019 | Type Agency Inspection |
| Inspector REVE,JASON,R | Time 00:00:45 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201811C | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 10818080A5 |
| Expiration 11/13/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/05/2020 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/31/2019 00:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency NEW SMYRNA BEACH PD | Date 04/29/2019 | Type Agency Inspection |
| Inspector REVE,JASON,R | Time 16:56:10 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 10818080A5 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/05/2020 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/29/2019 17:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

| | | | |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency MIAMI BEACH PD | Date 01/28/2019 | Type Agency Inspection | |
| Inspector MILLAN,KEVIN, | Time 11:31:04 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 390314 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 04/19/2019 |
| Number of Simulators 5 | Compliance Yes | | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/28/2019 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

| | | | |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency MIAMI BEACH PD | Date 02/27/2019 | Type Agency Inspection | |
| Inspector MILLAN,KEVIN, | Time 12:47:02 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 390314 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 04/19/2019 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/27/2019 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency MIAMI BEACH PD | Date 03/27/2019 | Type Agency Inspection |
| Inspector MILLAN,KEVIN, | Time 09:09:10 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201811C | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 390314 |
| Expiration 11/13/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 04/19/2019 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/27/2019 09:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

Agency MIAMI BEACH PD **Date** 04/03/2019 **Type** Department Inspection
Inspector REYES-RIVERA,DAVID,E **Time** 11:54:24 **Software** 8100.27

0.05 Lot # 201707D **0.08 Lot #** 201707E **0.20 Lot #** 201707C **0.08 Gas Lot #** AG831804
Expiration 07/25/2019 **Expiration** 07/25/2019 **Expiration** 07/24/2019 **Expiration** 11/14/2020

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.050 |
| | | 0.08 g/210L Test # 4 | 0.081 |
| | | 0.20 g/210L Test # 4 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.050 |
| | | 0.08 g/210L Test # 5 | 0.082 |
| | | 0.20 g/210L Test # 5 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.082 |
| 0.20 g/210L Test # 6 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.082 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.082 |
| 0.20 g/210L Test # 8 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.082 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.082 |
| 0.20 g/210L Test # 10 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0007 |
| 0.20 g/210L Test (Std Dev) | 0.0009 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

Remarks:

Data Download Date/Time 04/03/2019 12:01

Printed: 31 May 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

| | | | |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency MIAMI BEACH PD | Date 04/26/2019 | Type Agency Inspection | |
| Inspector MILLAN,KEVIN, | Time 11:22:36 | Software 8100.27 | |
| 0.05 Lot # 201811C | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 980558 |
| Expiration 11/13/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 07/09/2021 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.083 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/26/2019 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

| | | | |
|-------------------------------------|------------------------|-------------------------------|--|
| Agency HIALEAH PD | Date 01/16/2019 | Type Agency Inspection | |
| Inspector MONTOYA, GENARO, N | Time 11:46:04 | Software 8100.27 | |

| | | | |
|------------------------------|------------------------------|------------------------------|-----------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration |

Number of Simulators 5

Compliance No

| Check or Test | Result | Test | Result |
|---|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.076 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.172 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.198 |
| 0.20 g/210L Test: OK | No | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | No | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.181 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.198 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

LOOSE HOSE / CONTACTED THE DEPT. . Non-compliance:.20 G/210 L FAILED. RETEST PER DEPT..

Data Download Date/Time 01/16/2019 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency HIALEAH PD | Date 01/16/2019 | Type Agency Inspection |
| Inspector MONTOYA, GENARO, N | Time 12:31:01 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201702B | 0.08 Gas Lot # 712031 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/23/2019 | Expiration 02/21/2020 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.023 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.047 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.195 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.075 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.027 |
| 0.05 g/210L Test: OK | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.048 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.05 g/210L Test # (Repeat) 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

05: Control Outside Tolerance / SIM PADDLE STOPPED .

Data Download Date/Time 01/16/2019 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency HIALEAH PD | Date 02/13/2019 | Type Agency Inspection |
| Inspector MONTOYA, GENARO, N | Time 11:15:58 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 712031 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 02/21/2020 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.043 |
| Minimum Sample Volume Check: OK | | 0.05 g/210L Test # (Repeat) 1 | 0.050 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.196 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.040 |
| 0.05 g/210L Test: OK | Yes | 0.05 g/210L Test # (Repeat) 2 | 0.050 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.038 |
| | | 0.05 g/210L Test # (Repeat) 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

05: Control Outside Tolerance.

Data Download Date/Time 02/13/2019 11:32

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 03/04/2019 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

| | | | |
|-------------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency HIALEAH PD | Date 03/26/2019 | Type Agency Inspection | |
| Inspector MONTOYA, GENARO, N | Time 10:48:48 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 712031 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 02/21/2020 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # (Repeat) 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # (Repeat) 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # (Repeat) 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.052 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

00: Ambient Fail.

Data Download Date/Time 03/26/2019 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

| | | | |
|-------------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency HIALEAH PD | Date 04/16/2019 | Type Agency Inspection | |
| Inspector MONTOYA, GENARO, N | Time 11:06:31 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 712031 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 02/21/2020 |
| Number of Simulators 5 | Compliance Yes | | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/16/2019 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

| | | | |
|-------------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency HIALEAH PD | Date 01/16/2019 | Type Agency Inspection | |
| Inspector MONTOYA, GENARO, N | Time 11:19:44 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 712031 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 02/21/2020 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.076 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/16/2019 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

| | | | |
|-------------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency HIALEAH PD | Date 02/13/2019 | Type Agency Inspection | |
| Inspector MONTOYA, GENARO, N | Time 11:00:35 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 712031 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 02/21/2020 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.193 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/13/2019 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

| | | | |
|-------------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency HIALEAH PD | Date 03/26/2019 | Type Agency Inspection | |
| Inspector MONTOYA, GENARO, N | Time 10:43:43 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 712031 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 02/21/2020 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/26/2019 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency HIALEAH PD | Date 04/16/2019 | Type Agency Inspection |
| Inspector MONTOYA, GENARO, N | Time 11:20:34 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 712031 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 02/21/2020 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/16/2019 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency HIALEAH PD | Date 01/16/2019 | Type Agency Inspection |
| Inspector MONTOYA, GENARO, N | Time 11:43:39 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 712031 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 02/21/2020 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | 0.000 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # (Repeat) 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.076 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Interferent Detect Test # 2 | 0.000 |
| | | Interferent Detect Test # (Repeat) 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | 0.000 |
| | | Interferent Detect Test # (Repeat) 3 | INT |

Remarks:

Int Det: HOSE DISCONNECTED.

Data Download Date/Time 01/16/2019 11:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency HIALEAH PD | Date 02/13/2019 | Type Agency Inspection |
| Inspector MONTOYA, GENARO, N | Time 11:14:08 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 712031 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 02/21/2020 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.186 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.193 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.076 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.188 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.194 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.193 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

20: Control Outside Tolerance.

Data Download Date/Time 02/13/2019 11:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

| | | |
|---|------------------------|-----------------------------------|
| Agency HIALEAH PD | Date 03/04/2019 | Type Department Inspection |
| Inspector REYES-RIVERA, DAVID, E | Time 09:59:29 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| 0.05 Lot # 201707D | 0.08 Lot # 201707E | 0.20 Lot # 201707C | 0.08 Gas Lot # AG805702 |
| Expiration 07/25/2019 | Expiration 07/25/2019 | Expiration 07/24/2019 | Expiration 02/26/2020 |

| | | |
|-------------------------------|--|-----------------------|
| Number of Simulators 5 | Average Standard Deviation 0.0004 | Compliance Yes |
|-------------------------------|--|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.079 |
| | | 0.20 g/210L Test # 4 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.078 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.048 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| | | Interferent Detect Test # 5 | INT |

**Florida Department of Law Enforcement
Alcohol Testing Program**

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.079 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.079 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.079 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.079 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.048 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

Remarks:

Data Download Date/Time 03/04/2019 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

| | | | |
|-------------------------------|-----------------|------------------------|--|
| Agency HIALEAH PD | Date 03/26/2019 | Type Agency Inspection | |
| Inspector MONTROYA, GENARO, N | Time 10:52:10 | Software 8100.27 | |

| | | | |
|-----------------------|-----------------------|-----------------------|----------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration |

Number of Simulators 5 Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.079 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # (Repeat) 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.079 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # (Repeat) 2 | 0.079 |
| 0.20 g/210L Test: OK | No | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | No | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # (Repeat) 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

20: Control Outside Tolerance, WRONG SIM. CONECTED .

Data Download Date/Time 03/26/2019 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

| | | |
|-------------------------------------|------------------------|-------------------------------|
| Agency HIALEAH PD | Date 03/26/2019 | Type Agency Inspection |
| Inspector MONTOYA, GENARO, N | Time 11:26:55 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 712031 |
| Expiration 02/06/2020 | Expiration 08/22/2018 | Expiration 02/22/2020 | Expiration 02/21/2020 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/26/2019 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

| | | | |
|-------------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency HIALEAH PD | Date 04/16/2019 | Type Agency Inspection | |
| Inspector MONTOYA, GENARO, N | Time 10:59:43 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 712031 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 02/21/2020 |
| Number of Simulators 5 | Compliance Yes | | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.193 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/16/2019 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002566

| | | | |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency SURFSIDE PD | Date 01/28/2019 | Type Agency Inspection | |
| Inspector MILLAN,KEVIN, | Time 11:36:58 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 390314 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 04/19/2019 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/28/2019 11:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002566

| | | | |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency SURFSIDE PD | Date 02/27/2019 | Type Agency Inspection | |
| Inspector MILLAN,KEVIN, | Time 13:04:38 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 390314 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 04/19/2019 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.077 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/27/2019 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002566

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.082 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.079 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.082 |
| 0.20 g/210L Test # 7 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.082 |
| 0.20 g/210L Test # 8 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.082 |
| 0.20 g/210L Test # 9 | 0.201 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.079 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.083 |
| 0.20 g/210L Test # 10 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0003 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

Remarks:

Data Download Date/Time 03/07/2019 12:40

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002566

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency SURFSIDE PD | Date 03/27/2019 | Type Agency Inspection |
| Inspector MILLAN,KEVIN, | Time 09:53:02 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201811C | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 814181 |
| Expiration 11/13/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 07/24/2020 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.076 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.201 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.077 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/27/2019 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency MIAMI GARDENS PD | Date 01/31/2019 | Type Agency Inspection |
| Inspector OMALLEY,BRIAN,P | Time 15:14:10 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 09117080A2 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/05/2019 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | No | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

A F / M A: Range Exceeded Ambient Fail.

Data Download Date/Time 01/31/2019 15:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency MIAMI GARDENS PD | Date 02/28/2019 | Type Agency Inspection |
| Inspector OMALLEY,BRIAN,P | Time 15:06:31 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 09117080A2 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/05/2019 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.205 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.204 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/28/2019 15:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency MIAMI GARDENS PD | Date 03/28/2019 | Type Agency Inspection |
| Inspector OMALLEY,BRIAN,P | Time 11:08:22 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 09117080A2 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/05/2019 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMPLIANCE

Data Download Date/Time 03/28/2019 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency MIAMI GARDENS PD | Date 04/30/2019 | Type Agency Inspection |
| Inspector OMALLEY,BRIAN,P | Time 15:21:46 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 09117080A2 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/05/2019 |

Number of Simulators 5

Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMPLIANCE

Data Download Date/Time 04/30/2019 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency MIAMI GARDENS PD | Date 01/31/2019 | Type Agency Inspection |
| Inspector OMALLEY,BRIAN,P | Time 15:58:34 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 09117080A2 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/05/2019 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.046 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.192 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.046 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.194 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.046 |
| | | 0.08 g/210L Test # 3 | 0.076 |
| | | 0.20 g/210L Test # 3 | 0.194 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/31/2019 16:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency MIAMI GARDENS PD | Date 02/28/2019 | Type Agency Inspection |
| Inspector OMALLEY,BRIAN,P | Time 15:11:06 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 09117080A2 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/05/2019 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/28/2019 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency MIAMI GARDENS PD | Date 03/28/2019 | Type Agency Inspection |
| Inspector OMALLEY,BRIAN,P | Time 11:07:51 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 09117080A2 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/05/2019 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.194 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.195 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMPLIANCE

Data Download Date/Time 03/28/2019 11:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

| | | |
|----------------------------------|------------------------|-------------------------------|
| Agency MIAMI GARDENS PD | Date 04/30/2019 | Type Agency Inspection |
| Inspector OMALLEY,BRIAN,P | Time 15:21:32 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 09117080A2 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/05/2019 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

IN COMPLIANCE

Data Download Date/Time 04/30/2019 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency TALLAHASSEE PD | Date 01/25/2019 | Type Agency Inspection |
| Inspector MAHAN, TERRY, | Time 12:24:08 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 34416080A2 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 02/05/2019 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/25/2019 12:29

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency TALLAHASSEE PD | Date 02/22/2019 | Type Agency Inspection |
| Inspector MAHAN, TERRY, | Time 11:14:46 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201703A | 0.08 Lot # 201709A | 0.20 Lot # 201702B | 0.08 Gas Lot # 32518080A1 |
| Expiration 03/07/2019 | Expiration 09/12/2019 | Expiration 02/23/2019 | Expiration 02/05/2021 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.047 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.077 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.195 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.077 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/22/2019 11:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

| | | | |
|--------------------------------|------------------------------|-------------------------------|----------------------------------|
| Agency TALLAHASSEE PD | Date 03/27/2019 | Type Agency Inspection | |
| Inspector MAHAN, TERRY, | Time 14:46:53 | Software 8100.27 | |
| 0.05 Lot # 201811C | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 32518080A1 |
| Expiration 11/13/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 02/05/2021 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/28/2019 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency TALLAHASSEE PD | Date 04/30/2019 | Type Agency Inspection |
| Inspector MAHAN, TERRY, | Time 15:56:33 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| 0.05 Lot # 201811C | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 32518080A1 |
| Expiration 11/13/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 02/05/2021 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 05/02/2019 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

| | | |
|------------------------------------|------------------------|-------------------------------|
| Agency FWC | Date 04/25/2019 | Type Agency Inspection |
| Inspector SHANAHAN, JAKE, L | Time 07:11:55 | Software 8100.27 |

| | | | |
|-------------------|-------------------|-------------------|-----------------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK | No | | |
| Date and/or Time Adjusted | No | | |
| Minimum Sample Volume Check: OK | | | |
| Barometric Pressure Sensor Check: OK | | | |
| Alcohol Free Subject Test: 0.000 | No | | |
| Mouth Alcohol Test: Slope Not Met | No | | |
| Interferent Detect Test: Interferent Detect | No | | |
| Diagnostic Check (Post-Inspection): OK | No | | |
| Alcohol Free Test: OK | No | | |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

BYPASS TO OPERATE

Data Download Date/Time 04/25/2018 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

Agency FWC Date 04/25/2019 Type Department Inspection
 Inspector SHANAHAN, JAKE, L Time 08:47:56 Software 8100.27

0.05 Lot # 201707D 0.08 Lot # 201707E 0.20 Lot # 201707C 0.08 Gas Lot # AG805702
 Expiration 07/25/2019 Expiration 07/25/2019 Expiration 07/24/2019 Expiration 02/26/2020

Number of Simulators 5 Average Standard Deviation 0.0005 Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.082 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.050 |
| | | 0.08 g/210L Test # 4 | 0.081 |
| | | 0.20 g/210L Test # 4 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.049 |
| | | 0.08 g/210L Test # 5 | 0.080 |
| | | 0.20 g/210L Test # 5 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.080 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0004 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0006 |

Remarks:

Data Download Date/Time 04/25/2018 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

| | | | |
|---------------------------------|------------------------------|-------------------------------|--------------------------------|
| Agency FWC | Date 01/07/2019 | Type Agency Inspection | |
| Inspector RICE, MICHAEL, | Time 20:43:55 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # AG816701 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/16/2020 |
| Number of Simulators 4 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/08/2019 08:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

| | | | |
|-------------------------------|------------------------------|------------------------------|--------------------------------|
| Agency FWC | Date 02/28/2019 | Type Agency Inspection | |
| Inspector RICE, MICHAEL, | Time 19:59:04 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # AG816701 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/16/2020 |
| Number of Simulators 4 | Compliance Yes | | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/04/2019 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

| | | | |
|-----------------------------------|--|-----------------------------------|--------------------------------|
| Agency FWC | Date 03/20/2019 | Type Department Inspection | |
| Inspector KIRKLAND,BRETT,H | Time 12:29:08 | Software 8100.27 | |
| 0.05 Lot # 201707D | 0.08 Lot # 201707E | 0.20 Lot # 201707C | 0.08 Gas Lot # AG805701 |
| Expiration 07/25/2019 | Expiration 07/25/2019 | Expiration 07/24/2019 | Expiration 02/26/2020 |
| Number of Simulators 5 | Average Standard Deviation 0.0005 | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|---|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.084 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.08 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.20 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.000 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.084 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.000 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.084 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.081 |
| | | 0.20 g/210L Test # 4 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.087 |
| | | 0.08 g/210L Dry Gas Std Test # (Repeat) 4 | 0.084 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.050 |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

| | |
|--|-------|
| 0.08 g/210L Test # 5 | 0.080 |
| 0.20 g/210L Test # 5 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 5 | 0.086 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 5 | 0.084 |
| Interferent Detect Test # 5 | INT |
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.086 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 6 | 0.084 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.086 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 7 | 0.084 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.085 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 8 | 0.084 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.085 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 9 | 0.084 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.084 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 10 | 0.083 |
| Interferent Detect Test # 10 | INT |

Printed: 31 May 2019

COBRA ©1997-2014 CMI, Inc.

**Florida Department of Law Enforcement
Alcohol Testing Program**

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

| | |
|--|--------|
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0009 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0413 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) (Repeat) | 0.0003 |

Remarks:

08: Control Outside FORGOT TO PLUG IN HOSE

Data Download Date/Time 03/20/2019 12:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

Agency FWC Date 03/27/2019 Type Department Inspection
 Inspector KIRKLAND,BRETT,H Time 10:58:35 Software 8100.27

0.05 Lot # 201707D 0.08 Lot # 201707E 0.20 Lot # 201707C 0.08 Gas Lot # AG805701
 Expiration 07/25/2019 Expiration 07/25/2019 Expiration 07/24/2019 Expiration 02/26/2020

Number of Simulators 5 Average Standard Deviation 0.0005 Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.199 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.048 |
| | | 0.08 g/210L Test # 4 | 0.081 |
| | | 0.20 g/210L Test # 4 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.049 |
| | | 0.08 g/210L Test # 5 | 0.081 |
| | | 0.20 g/210L Test # 5 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.048 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.048 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Test (Std Dev) | 0.0005 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

Remarks:

Data Download Date/Time 03/27/2019 11:12

Printed: 31 May 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency FWC | Date 04/01/2019 | Type Agency Inspection |
| Inspector RICE, MICHAEL, | Time 10:06:50 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # AG816701 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/16/2020 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 4 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.077 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.047 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/01/2019 20:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

| | | | |
|---------------------------------|------------------------------|-------------------------------|--------------------------------|
| Agency FWC | Date 01/07/2019 | Type Agency Inspection | |
| Inspector RICE, MICHAEL, | Time 20:40:48 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # AG816701 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/16/2020 |
| Number of Simulators 4 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.077 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/08/2019 08:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

| | | | |
|---------------------------------|------------------------------|-------------------------------|--------------------------------|
| Agency FWC | Date 02/28/2019 | Type Agency Inspection | |
| Inspector RICE, MICHAEL, | Time 19:54:36 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # AG816701 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/16/2020 |
| Number of Simulators 4 | Compliance Yes | | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.078 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.078 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/04/2019 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

| | | |
|-----------------------------------|------------------------|-----------------------------------|
| Agency FWC | Date 03/20/2019 | Type Department Inspection |
| Inspector KIRKLAND,BRETT,H | Time 11:54:36 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| 0.05 Lot # 201707D | 0.08 Lot # 201707E | 0.20 Lot # 201707C | 0.08 Gas Lot # AG805701 |
| Expiration 07/25/2019 | Expiration 07/25/2019 | Expiration 07/24/2019 | Expiration 02/26/2020 |

| | | |
|-------------------------------|--|----------------------|
| Number of Simulators 5 | Average Standard Deviation 0.0005 | Compliance No |
|-------------------------------|--|----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.078 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.049 |
| | | 0.08 g/210L Test # 4 | 0.080 |
| | | 0.20 g/210L Test # 4 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.079 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.049 |
| | | 0.08 g/210L Test # 5 | 0.081 |
| | | 0.20 g/210L Test # 5 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.079 |
| | | Interferent Detect Test # 5 | INT |

**Florida Department of Law Enforcement
Alcohol Testing Program**

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.078 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.078 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.080 |
| 0.20 g/210L Test # 8 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.079 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.049 |
| 0.08 g/210L Test # 9 | 0.080 |
| 0.20 g/210L Test # 9 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.078 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.049 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.196 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0004 |
| 0.20 g/210L Test (Std Dev) | 0.0011 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

Remarks:

BUMPED POWER CORD DURING DIAGNOSTIC CHECK WILL REPEAT INSPENon-compliance:BUMPED POWER CORD DURING DIAGNOSTIC CHECK .

Data Download Date/Time 03/20/2019 12:03

Printed: 31 May 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.049 |
| 0.08 g/210L Test # 6 | 0.080 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.082 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.049 |
| 0.08 g/210L Test # 7 | 0.081 |
| 0.20 g/210L Test # 7 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.082 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.049 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.082 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.079 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.082 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.080 |
| 0.20 g/210L Test # 10 | 0.197 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.082 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0006 |
| 0.08 g/210L Test (Std Dev) | 0.0007 |
| 0.20 g/210L Test (Std Dev) | 0.0010 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

Remarks:

Data Download Date/Time 03/20/2019 14:11

Printed: 31 May 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

| | | |
|---------------------------------|------------------------|-------------------------------|
| Agency FWC | Date 04/01/2019 | Type Agency Inspection |
| Inspector RICE, MICHAEL, | Time 10:13:15 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # AG816701 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/16/2020 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 4 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.047 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.195 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.079 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.196 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.079 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.078 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.079 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 04/01/2019 10:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

| | | |
|----------------------------------|------------------------------|-------------------------------|
| Agency FWC | Date 01/22/2019 | Type Agency Inspection |
| Inspector MOLNAR,JARROD,T | Time 13:18:52 | Software 8100.27 |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 |
| | | 0.08 Gas Lot # 477627 |
| | | Expiration 06/29/2019 |
| Number of Simulators 5 | | Compliance Yes |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.200 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.200 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.200 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/22/2019 14:28

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

| | | | |
|----------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency FWC | Date 02/25/2019 | Type Agency Inspection | |
| Inspector MOLNAR,JARROD,T | Time 13:52:49 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 477627 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/29/2019 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.082 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.201 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.082 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.082 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/25/2019 15:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

| | | |
|-----------------------------------|------------------------|-------------------------------|
| Agency FWC | Date 03/24/2019 | Type Agency Inspection |
| Inspector BASFORD,NICOLE,M | Time 14:34:37 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 477627 |
| Expiration 02/06/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 06/29/2019 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.201 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/24/2019 15:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

| | | | |
|----------------------------|-----------------|------------------------|--|
| Agency FWC | Date 04/16/2019 | Type Agency Inspection | |
| Inspector BASFORD,NICOLE,M | Time 11:50:11 | Software 8100.27 | |

| | | | |
|------------|------------|------------|----------------|
| 0.05 Lot # | 0.08 Lot # | 0.20 Lot # | 0.08 Gas Lot # |
| Expiration | Expiration | Expiration | Expiration |

Number of Simulators 5 Compliance No

| Check or Test | Result | Test | Result |
|---|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | Interferent Detect Test # 1 | 0.000 |
| Minimum Sample Volume Check: OK | | Interferent Detect Test # (Repeat) 1 | 0.000 |
| Barometric Pressure Sensor Check: OK | | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Subject Test: 0.000 | Yes | Interferent Detect Test # 2 | 0.000 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # (Repeat) 2 | 0.000 |
| Interferent Detect Test: Interferent Detect | No | Alcohol Free Test # 3 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | No | Interferent Detect Test # 3 | 0.000 |
| Alcohol Free Test: OK | Yes | Interferent Detect Test # (Repeat) 3 | 0.000 |
| 0.05 g/210L Test: OK | No | | |
| 0.08 g/210L Test: OK | No | | |
| 0.20 g/210L Test: OK | No | | |
| 0.08 g/210L Dry Gas Std Test: OK | No | | |

Remarks:

DID NOT READ ACETONE, . SEAL ON JAR DEFFECTIVE.

Data Download Date/Time 04/16/2019 13:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

| | | | |
|-----------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency FWC | Date 04/16/2019 | Type Agency Inspection | |
| Inspector BASFORD,NICOLE,M | Time 12:32:52 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 477627 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 06/29/2019 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.080 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.202 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.202 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.202 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

NA

Data Download Date/Time 04/16/2019 13:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency MIAMI BEACH PD | Date 01/28/2019 | Type Agency Inspection |
| Inspector MILLAN,KEVIN, | Time 11:23:59 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 366849 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 03/30/2019 |

Number of Simulators 5 **Compliance** Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.196 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.079 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.197 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 01/28/2019 12:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

| | | | |
|--------------------------------|------------------------------|-------------------------------|------------------------------|
| Agency MIAMI BEACH PD | Date 02/27/2019 | Type Agency Inspection | |
| Inspector MILLAN,KEVIN, | Time 12:54:22 | Software 8100.27 | |
| 0.05 Lot # 201802B | 0.08 Lot # 201709A | 0.20 Lot # 201802G | 0.08 Gas Lot # 390314 |
| Expiration 02/06/2020 | Expiration 09/12/2019 | Expiration 02/22/2020 | Expiration 04/19/2019 |
| Number of Simulators 5 | | Compliance Yes | |

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.078 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.197 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.078 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.197 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.196 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.081 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 02/27/2019 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

Agency MIAMI BEACH PD Date 03/07/2019 Type Department Inspection
 Inspector REYES-RIVERA,DAVID,E Time 12:46:19 Software 8100.27

0.05 Lot # 201707D 0.08 Lot # 201707E 0.20 Lot # 201707C 0.08 Gas Lot # AG805702
 Expiration 07/25/2019 Expiration 07/25/2019 Expiration 07/24/2019 Expiration 02/26/2020

Number of Simulators 5 Average Standard Deviation 0.0002 Compliance Yes

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.050 |
| Minimum Sample Volume Check: OK | Yes | 0.08 g/210L Test # 1 | 0.081 |
| Barometric Pressure Sensor Check: OK | Yes | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.050 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.081 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.199 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.050 |
| | | 0.08 g/210L Test # 3 | 0.081 |
| | | 0.20 g/210L Test # 3 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |
| | | Alcohol Free Test # 4 | 0.000 |
| | | 0.05 g/210L Test # 4 | 0.050 |
| | | 0.08 g/210L Test # 4 | 0.081 |
| | | 0.20 g/210L Test # 4 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 4 | 0.080 |
| | | Interferent Detect Test # 4 | INT |
| | | Alcohol Free Test # 5 | 0.000 |
| | | 0.05 g/210L Test # 5 | 0.050 |
| | | 0.08 g/210L Test # 5 | 0.081 |
| | | 0.20 g/210L Test # 5 | 0.199 |
| | | 0.08 g/210L Dry Gas Std Test # 5 | 0.080 |
| | | Interferent Detect Test # 5 | INT |

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

| | |
|--|--------|
| Alcohol Free Test # 6 | 0.000 |
| 0.05 g/210L Test # 6 | 0.050 |
| 0.08 g/210L Test # 6 | 0.081 |
| 0.20 g/210L Test # 6 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6 | 0.080 |
| Interferent Detect Test # 6 | INT |
| Alcohol Free Test # 7 | 0.000 |
| 0.05 g/210L Test # 7 | 0.050 |
| 0.08 g/210L Test # 7 | 0.080 |
| 0.20 g/210L Test # 7 | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7 | 0.080 |
| Interferent Detect Test # 7 | INT |
| Alcohol Free Test # 8 | 0.000 |
| 0.05 g/210L Test # 8 | 0.050 |
| 0.08 g/210L Test # 8 | 0.081 |
| 0.20 g/210L Test # 8 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 8 | 0.080 |
| Interferent Detect Test # 8 | INT |
| Alcohol Free Test # 9 | 0.000 |
| 0.05 g/210L Test # 9 | 0.050 |
| 0.08 g/210L Test # 9 | 0.081 |
| 0.20 g/210L Test # 9 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9 | 0.080 |
| Interferent Detect Test # 9 | INT |
| Alcohol Free Test # 10 | 0.000 |
| 0.05 g/210L Test # 10 | 0.050 |
| 0.08 g/210L Test # 10 | 0.081 |
| 0.20 g/210L Test # 10 | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.081 |
| Interferent Detect Test # 10 | INT |
| 0.05 g/210L Test (Std Dev) | 0.0000 |
| 0.08 g/210L Test (Std Dev) | 0.0003 |
| 0.20 g/210L Test (Std Dev) | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

Remarks:

Data Download Date/Time 03/07/2019 12:54

Printed: 31 May 2019

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency MIAMI BEACH PD | Date 03/27/2019 | Type Agency Inspection |
| Inspector MILLAN,KEVIN, | Time 09:16:52 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201811C | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 390314 |
| Expiration 11/13/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 04/19/2019 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.048 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.079 |
| Barometric Pressure Sensor Check: OK | | 0.20 g/210L Test # 1 | 0.198 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.080 |
| Mouth Alcohol Test: Slope Not Met | Yes | Interferent Detect Test # 1 | INT |
| Interferent Detect Test: Interferent Detect | Yes | Alcohol Free Test # 2 | 0.000 |
| Diagnostic Check (Post-Inspection): OK | Yes | 0.05 g/210L Test # 2 | 0.048 |
| Alcohol Free Test: OK | Yes | 0.08 g/210L Test # 2 | 0.080 |
| 0.05 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.080 |
| 0.20 g/210L Test: OK | Yes | Interferent Detect Test # 2 | INT |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.048 |
| | | 0.08 g/210L Test # 3 | 0.080 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.080 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

Data Download Date/Time 03/27/2019 11:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

| | | |
|--------------------------------|------------------------|-------------------------------|
| Agency MIAMI BEACH PD | Date 04/26/2019 | Type Agency Inspection |
| Inspector MILLAN,KEVIN, | Time 11:47:36 | Software 8100.27 |

| | | | |
|------------------------------|------------------------------|------------------------------|------------------------------|
| 0.05 Lot # 201811C | 0.08 Lot # 201808E | 0.20 Lot # 201802G | 0.08 Gas Lot # 814181 |
| Expiration 11/13/2020 | Expiration 08/22/2020 | Expiration 02/22/2020 | Expiration 07/24/2020 |

| | |
|-------------------------------|-----------------------|
| Number of Simulators 5 | Compliance Yes |
|-------------------------------|-----------------------|

| Check or Test | Result | Test | Result |
|---|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK | Yes | Alcohol Free Test # 1 | 0.000 |
| Date and/or Time Adjusted | No | 0.05 g/210L Test # 1 | 0.049 |
| Minimum Sample Volume Check: OK | | 0.08 g/210L Test # 1 | 0.091 |
| Barometric Pressure Sensor Check: OK | | 0.08 g/210L Test # (Repeat) 1 | 0.082 |
| Alcohol Free Subject Test: 0.000 | Yes | 0.20 g/210L Test # 1 | 0.197 |
| Mouth Alcohol Test: Slope Not Met | Yes | 0.08 g/210L Dry Gas Std Test # 1 | 0.081 |
| Interferent Detect Test: Interferent Detect | Yes | Interferent Detect Test # 1 | INT |
| Diagnostic Check (Post-Inspection): OK | Yes | Alcohol Free Test # 2 | 0.000 |
| Alcohol Free Test: OK | Yes | 0.05 g/210L Test # 2 | 0.049 |
| 0.05 g/210L Test: OK | Yes | 0.08 g/210L Test # 2 | 0.087 |
| 0.08 g/210L Test: OK | Yes | 0.08 g/210L Test # (Repeat) 2 | 0.081 |
| 0.20 g/210L Test: OK | Yes | 0.20 g/210L Test # 2 | 0.198 |
| 0.08 g/210L Dry Gas Std Test: OK | Yes | 0.08 g/210L Dry Gas Std Test # 2 | 0.081 |
| | | Interferent Detect Test # 2 | INT |
| | | Alcohol Free Test # 3 | 0.000 |
| | | 0.05 g/210L Test # 3 | 0.049 |
| | | 0.08 g/210L Test # 3 | 0.085 |
| | | 0.08 g/210L Test # (Repeat) 3 | 0.079 |
| | | 0.20 g/210L Test # 3 | 0.198 |
| | | 0.08 g/210L Dry Gas Std Test # 3 | 0.082 |
| | | Interferent Detect Test # 3 | INT |

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 04/26/2019 11:54