

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

<b>Agency</b> FFWCC TAMPA	<b>Date</b> 01/02/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> PULASKI,DAMON,J	<b>Time</b> 12:47:32	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 13518080A5
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 08/05/2020

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 01/02/2019 13:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

<b>Agency</b> FFWCC TAMPA	<b>Date</b> 02/13/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> PULASKI,DAMON,J	<b>Time</b> 20:59:38	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 13518080A5
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 08/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 02/13/2019 22:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

<b>Agency</b> FFWCC TAMPA	<b>Date</b> 03/18/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> PULASKI,DAMON,J	<b>Time</b> 10:34:53	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 13518080A5
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 08/05/2020

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 03/18/2019 10:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000901

<b>Agency</b> FFWCC TAMPA	<b>Date</b> 04/24/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> PULASKI,DAMON,J	<b>Time</b> 19:07:13	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 13518080A5
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 08/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 04/24/2019 19:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

<b>Agency</b> FFWCC TAMPA	<b>Date</b> 01/02/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> PULASKI,DAMON,J	<b>Time</b> 12:10:10	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 13518080A5
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 08/05/2020

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 01/02/2019 12:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

<b>Agency</b> FFWCC TAMPA	<b>Date</b> 02/13/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> PULASKI,DAMON,J	<b>Time</b> 21:43:09	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 13518080A5
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 08/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 02/13/2019 21:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000902

<b>Agency</b> FFWCC TAMPA	<b>Date</b> 03/18/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> PULASKI,DAMON,J	<b>Time</b> 11:57:32	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 13518080A5
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 08/05/2020

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 03/18/2019 12:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

<b>Agency</b> FFWCC SOUTH REGION	<b>Date</b> 01/23/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> BLANCO, JUAN, S	<b>Time</b> 20:52:50	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 666491
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/30/2019
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # (Repeat) 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

REPEATED .08G/210L TEST DUE TO RFI. AI

**Data Download Date/Time** 01/23/2019 21:11



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

<b>Agency</b> FFWCC SOUTH REGION	<b>Date</b> 02/27/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> PASTOR,JOSE,F	<b>Time</b> 13:25:35	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 666491
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/30/2019

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/27/2019 15:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

<b>Agency</b> FFWCC SOUTH REGION	<b>Date</b> 03/26/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> BLANCO, JUAN, S	<b>Time</b> 23:04:47	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 666491
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/30/2019
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

AI

**Data Download Date/Time** 03/26/2019 23:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

<b>Agency</b> FFWCC SOUTH REGION	<b>Date</b> 04/30/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> BLANCO, JUAN, S	<b>Time</b> 23:16:33	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 666491
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/30/2019
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

AI

**Data Download Date/Time** 04/30/2019 23:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000905

<b>Agency</b> FWC	<b>Date</b> 01/07/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> RICE, MICHAEL,	<b>Time</b> 20:49:12	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> AG728303
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 10/10/2019
<b>Number of Simulators</b> 4		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/08/2019 08:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000905

<b>Agency</b> FWC	<b>Date</b> 02/28/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> RICE, MICHAEL,	<b>Time</b> 19:48:56	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> AG728303
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 10/10/2019

  

<b>Number of Simulators</b> 4	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/04/2019 14:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000905

<b>Agency</b> FWC	<b>Date</b> 03/20/2019	<b>Type</b> Department Inspection	
<b>Inspector</b> KIRKLAND,BRETT,H	<b>Time</b> 12:08:04	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201707D	<b>0.08 Lot #</b> 201707E	<b>0.20 Lot #</b> 201707C	<b>0.08 Gas Lot #</b> AG805701
<b>Expiration</b> 07/25/2019	<b>Expiration</b> 07/25/2019	<b>Expiration</b> 07/24/2019	<b>Expiration</b> 02/26/2020
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0006	<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000905

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

---

**Data Download Date/Time** 03/20/2019 12:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000905

<b>Agency</b> FWC	<b>Date</b> 04/01/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> RICE, MICHAEL,	<b>Time</b> 10:17:19	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> AG728303
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 10/10/2019

  

<b>Number of Simulators</b> 4	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

00: Ambient Fail.

**Data Download Date/Time** 04/01/2019 11:03



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

Agency FISH & WILDLIFE CC	Date 01/02/2019	Type Agency Inspection	
Inspector NEIDERHISER,JOHN,E	Time 10:21:27	Software 8100.27	
<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 666491
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 12/30/2019
<b>Number of Simulators</b> 5	<b>Compliance</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

TESTED AT SPEC OPS

Data Download Date/Time 01/02/2019 10:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

Agency FISH & WILDLIFE CC      Date 02/04/2019      Type Agency Inspection  
Inspector NEIDERHISER,JOHN,E      Time 11:27:59      Software 8100.27

0.05 Lot # 201703A      0.08 Lot # 201709A      0.20 Lot # 201702B      0.08 Gas Lot # 666491  
Expiration 03/07/2019      Expiration 09/12/2019      Expiration 02/23/2019      Expiration 12/30/2019

Number of Simulators 5      Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.053
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

#### Remarks:

Data Download Date/Time 02/04/2019 11:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

**Agency** FISH & WILDLIFE CC      **Date** 02/05/2019      **Type** Department Inspection  
**Inspector** REYES-RIVERA, DAVID, E      **Time** 08:28:29      **Software** 8100.27

**0.05 Lot #** 201707D      **0.08 Lot #** 201707E      **0.20 Lot #** 201707C      **0.08 Gas Lot #** AG805701  
**Expiration** 07/25/2019      **Expiration** 07/25/2019      **Expiration** 07/24/2019      **Expiration** 02/26/2020

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

**Data Download Date/Time** 02/05/2019 08:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

**Agency** FISH & WILDLIFE CC      **Date** 02/05/2019      **Type** Agency Inspection  
**Inspector** NEIDERHISER,JOHN,E      **Time** 15:19:41      **Software** 8100.27

**0.05 Lot #** 201703A      **0.08 Lot #** 201709A      **0.20 Lot #** 201702B      **0.08 Gas Lot #** 666491  
**Expiration** 03/07/2019      **Expiration** 09/12/2019      **Expiration** 02/23/2019      **Expiration** 12/30/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

TESTED AT SPEC OPS

**Data Download Date/Time** 02/06/2019 13:57



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

<b>Agency</b> FISH & WILDLIFE CC	<b>Date</b> 04/01/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER, JOHN, E	<b>Time</b> 11:37:27	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 666491
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 09/12/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/30/2019

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.052
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # (Repeat) 3	INT

**Remarks:**

Int Det: RFI Detect CELL PHONE.

**Data Download Date/Time** 04/01/2019 11:44

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

<b>Agency</b> PORT ST JOE PD	<b>Date</b> 03/04/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> MURPHY, PATRICK, J	<b>Time</b> 14:39:00	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

<b>Number of Simulators</b>	<b>Compliance</b> Yes
-----------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

SKIPPED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 03/05/2019 11:25



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

**Agency** PORT ST JOE PD                      **Date** 03/05/2019                      **Type** Department Inspection  
**Inspector** MURPHY,PATRICK,J              **Time** 11:13:45                      **Software** 8100.27

**0.05 Lot #** 201707D              **0.08 Lot #** 201707E              **0.20 Lot #** 201707C              **0.08 Gas Lot #** AG831804  
**Expiration** 07/25/2019              **Expiration** 07/25/2019              **Expiration** 07/24/2019              **Expiration** 11/14/2020

**Number of Simulators** 5                      **Average Standard Deviation** 0.0003                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

Data Download Date/Time 03/05/2019 11:25

Printed: 31 May 2019

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

<b>Agency</b> PORT ST JOE PD	<b>Date</b> 03/14/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> BURCH,RUSSELL,L	<b>Time</b> 19:17:41	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 17618080A1
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 08/05/2020

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/14/2019 20:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

<b>Agency</b> PORT ST JOE PD	<b>Date</b> 04/21/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> BURCH,RUSSELL,L	<b>Time</b> 19:06:22	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 17618080A1
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 08/05/2020
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/21/2019 19:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

<b>Agency</b> PINECREST PD	<b>Date</b> 01/31/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> ESPINOZA,PABLO,	<b>Time</b> 09:58:48	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 22817080A4
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 10/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/31/2019 13:28



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0008
0.08 g/210L Test (Std Dev)	0.0009
0.20 g/210L Test (Std Dev)	0.0012
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

---

**Data Download Date/Time** 02/05/2019 11:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

Agency PINECREST PD	Date 02/12/2019	Type Agency Inspection	
Inspector ESPINOZA,PABLO,	Time 10:56:04	Software 8100.27	
<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 22817080A4
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 10/05/2019

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

Data Download Date/Time 02/12/2019 13:09



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

<b>Agency</b> FORT MYERS PD	<b>Date</b> 01/09/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN, DAVID, S	<b>Time</b> 13:44:57	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 29017080A2
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/09/2019 13:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

<b>Agency</b> FORT MYERS PD	<b>Date</b> 02/15/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN, DAVID, S	<b>Time</b> 12:38:58	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 29017080A2
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/15/2019 12:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

<b>Agency</b> FORT MYERS PD	<b>Date</b> 03/06/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN, DAVID, S	<b>Time</b> 13:23:49	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 29017080A2
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/06/2019 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

<b>Agency</b> FORT MYERS PD	<b>Date</b> 04/17/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN, DAVID, S	<b>Time</b> 13:59:07	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 29017080A2
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/17/2019 14:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

<b>Agency</b> SANIBEL ISLAND PD	<b>Date</b> 01/09/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN, DAVID, S	<b>Time</b> 14:49:20	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 10818080A4
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 06/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/09/2019 15:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

<b>Agency</b> SANIBEL ISLAND PD	<b>Date</b> 02/15/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN, DAVID, S	<b>Time</b> 11:04:54	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 10818080A4
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 06/05/2020

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/15/2019 11:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

<b>Agency</b> SANIBEL ISLAND PD	<b>Date</b> 03/06/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN, DAVID, S	<b>Time</b> 12:08:42	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 10818080A4
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 06/05/2020

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/06/2019 12:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

<b>Agency</b> SANIBEL ISLAND PD	<b>Date</b> 04/17/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN, DAVID, S	<b>Time</b> 12:03:21	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 10818080A4
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 06/05/2020

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/17/2019 12:27



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 01/09/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN, DAVID, S	<b>Time</b> 11:00:00	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 10818080A4
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 06/05/2020

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/09/2019 11:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 02/15/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN, DAVID, S	<b>Time</b> 09:35:50	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 10818080A4
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 06/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/15/2019 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 03/06/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN, DAVID, S	<b>Time</b> 10:46:42	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 10818080A4
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 06/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

00: Ambient Fail CLEARED ROOM AIR

**Data Download Date/Time** 03/06/2019 10:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 04/17/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN, DAVID, S	<b>Time</b> 11:58:10	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 10818080A4
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 06/05/2020

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/17/2019 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 01/09/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN, DAVID, S	<b>Time</b> 13:42:51	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 10818080A4
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 06/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/09/2019 13:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 02/15/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> CLAFLIN, DAVID, S	<b>Time</b> 11:06:59	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 28818080A1
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2020
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/15/2019 11:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 03/06/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN, DAVID, S	<b>Time</b> 10:56:20	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 28818080A1
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2020

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/06/2019 11:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

<b>Agency</b> LEE COUNTY SO	<b>Date</b> 04/17/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> CLAFLIN, DAVID, S	<b>Time</b> 10:10:33	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 28818080A1
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/17/2019 10:55



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

<b>Agency</b> BREVARD COUNTY S.O.	<b>Date</b> 03/27/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> PLATT,SHAYLA,D	<b>Time</b> 09:25:50	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 03/27/2019 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

Agency BREVARD COUNTY S.O.                      Date 03/27/2019                      Type Department Inspection  
 Inspector PLATT,SHAYLA,D                      Time 11:29:44                      Software 8100.27

0.05 Lot # 201707D                      0.08 Lot # 201707E                      0.20 Lot # 201707C                      0.08 Gas Lot # AG831804  
 Expiration 07/25/2019                      Expiration 07/25/2019                      Expiration 07/24/2019                      Expiration 11/14/2020

Number of Simulators 5                      Average Standard Deviation 0.0005                      Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

**Remarks:**

**Data Download Date/Time** 03/27/2019 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

<b>Agency</b> BREVARD COUNTY S.O.	<b>Date</b> 04/24/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> GARRITY, THOMAS, A	<b>Time</b> 17:33:43	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 19817080A3
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 09/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/24/2019 17:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 01/02/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER,JOHN,E	<b>Time</b> 15:44:47	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 19817080A1
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 09/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

TESTED AT SPEC OPS

**Data Download Date/Time** 01/02/2019 17:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

**Agency** COLLIER COUNTY SO                      **Date** 01/03/2019                      **Type** Department Inspection  
**Inspector** REYES-RIVERA, DAVID, E              **Time** 12:47:51                      **Software** 8100.27

**0.05 Lot #** 201707D                      **0.08 Lot #** 201707E                      **0.20 Lot #** 201707C                      **0.08 Gas Lot #** AG805701  
**Expiration** 07/25/2019                      **Expiration** 07/25/2019                      **Expiration** 07/24/2019                      **Expiration** 02/26/2020

**Number of Simulators** 5                      **Average Standard Deviation** 0.0005                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0010
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

**Data Download Date/Time** 01/03/2019 12:51

**Printed: 31 May 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 01/07/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER, JOHN, E	<b>Time</b> 12:46:35	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 19817080A1
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 09/05/2019

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

TESTED AT SPEC OPS

**Data Download Date/Time** 01/07/2019 13:43



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 02/04/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER, JOHN, E	<b>Time</b> 09:37:44	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 19817080A1
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 09/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.051
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	0.000
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	0.000
		Interferent Detect Test # (Repeat) 3	INT

**Remarks:**

Int Det: POOR JAR SEAL.

**Data Download Date/Time** 02/04/2019 10:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 03/04/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER,JOHN,E	<b>Time</b> 15:22:26	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 03417080A2
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 03/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

TESTED AT SPEC OPS

**Data Download Date/Time** 03/04/2019 15:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 04/01/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER, JOHN, E	<b>Time</b> 09:49:44	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 31517080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 01/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

TESTED AT SPEC OPS

**Data Download Date/Time** 04/01/2019 10:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 01/02/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER, JOHN, E	<b>Time</b> 15:38:17	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 19817080A1
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 09/05/2019

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

00: Ambient Fail MA NEAR.

**Data Download Date/Time** 01/02/2019 17:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

**Agency** COLLIER COUNTY SO                      **Date** 01/03/2019                      **Type** Department Inspection  
**Inspector** REYES-RIVERA, DAVID, E              **Time** 14:34:57                      **Software** 8100.27

**0.05 Lot #** 201707D                      **0.08 Lot #** 201707E                      **0.20 Lot #** 201707C                      **0.08 Gas Lot #** AG805701  
**Expiration** 07/25/2019                      **Expiration** 07/25/2019                      **Expiration** 07/24/2019                      **Expiration** 02/26/2020

**Number of Simulators** 5                      **Average Standard Deviation** 0.0005                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

---

**Data Download Date/Time** 01/03/2019 14:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 01/07/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER, JOHN, E	<b>Time</b> 12:40:05	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 19817080A1
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 09/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

**Remarks:**

TESTED AT SPEC OPS

**Data Download Date/Time** 01/07/2019 13:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 02/04/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER, JOHN, E	<b>Time</b> 09:56:30	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 19817080A1
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 09/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.052
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.077
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.076
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

08: Control Outside Tolerance NO GAS LINE.

**Data Download Date/Time** 02/04/2019 10:15



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 03/04/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER,JOHN,E	<b>Time</b> 15:11:02	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 31517080A1
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 01/05/2020

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

TESTED AT SPEC OPS

**Data Download Date/Time** 03/05/2019 12:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 04/01/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER,JOHN,E	<b>Time</b> 10:09:35	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 31517080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 01/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

TESTED AT SPEC OPS

**Data Download Date/Time** 04/01/2019 10:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

**Agency** COLLIER COUNTY SO                      **Date** 01/02/2019                      **Type** Department Inspection  
**Inspector** REYES-RIVERA, DAVID, E              **Time** 13:49:41                      **Software** 8100.27

**0.05 Lot #** 201707D                      **0.08 Lot #** 201707E                      **0.20 Lot #** 201707C                      **0.08 Gas Lot #** AG805701  
**Expiration** 07/25/2019                      **Expiration** 07/25/2019                      **Expiration** 07/24/2019                      **Expiration** 02/26/2020

**Number of Simulators** 5                      **Average Standard Deviation** 0.0006                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

**Data Download Date/Time** 01/02/2019 14:03

**Printed: 31 May 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 01/03/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER, JOHN, E	<b>Time</b> 11:42:43	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 19817080A1
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 09/05/2019

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

TESTED AT SPEC OPS

**Data Download Date/Time** 01/03/2019 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 02/05/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER,JOHN,E	<b>Time</b> 14:11:27	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 19817080A1
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 09/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

TESTED AT SPEC OPS

**Data Download Date/Time** 02/05/2019 14:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 03/04/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER, JOHN, E	<b>Time</b> 10:34:27	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 10818080A6
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 06/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

TESTED AT SPEC OPS

**Data Download Date/Time** 03/04/2019 10:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

<b>Agency</b> COLLIER COUNTY SO	<b>Date</b> 04/01/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> NEIDERHISER, JOHN, E	<b>Time</b> 14:25:22	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 31517080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 01/05/2020

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

TESTED AT SPEC OPS

**Data Download Date/Time** 04/02/2019 09:23



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 01/30/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> KERN, MICHAEL, P	<b>Time</b> 07:27:21	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 10818080A5
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 06/05/2020

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

CLEANED AND INSPECTED

**Data Download Date/Time** 01/31/2019 06:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 02/27/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> KERN, MICHAEL, P	<b>Time</b> 07:05:28	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 10818080A5
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 06/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

CLEANED AND INSPECTED

**Data Download Date/Time** 03/02/2019 11:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 03/28/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> KERN, MICHAEL, P	<b>Time</b> 07:20:46	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 10818080A5
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 06/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.075
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

CLEANED AND INSPECTED

**Data Download Date/Time** 04/01/2019 06:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

<b>Agency</b> CHARLOTTE COUNTY SO	<b>Date</b> 04/29/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> KERN, MICHAEL, P	<b>Time</b> 07:27:12	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 10818080A5
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 06/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

00: Ambient Fail.

**Data Download Date/Time** 04/29/2019 10:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

<b>Agency</b> GLADES COUNTY S.O.	<b>Date</b> 01/31/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> BONILLA, JOSUE, D	<b>Time</b> 11:09:40	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 31517080A1
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 01/05/2020

**Number of Simulators** 3

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.188
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	0.000
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test # (Repeat) 1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # 2	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.191
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.193
		0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	0.000
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.192
		0.20 g/210L Test # (Repeat) 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	0.000
		Interferent Detect Test # (Repeat) 3	INT

**Remarks:**

Int Det: BAD ACETONE. 20: Control Outside RADIO IN THE ROOM.

**Data Download Date/Time** 01/31/2019 11:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

<b>Agency</b> GLADES COUNTY S.O.	<b>Date</b> 02/28/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> MADRUGO, FELICIA,	<b>Time</b> 19:40:52	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 31517080A1
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 01/05/2020

**Number of Simulators** 2

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.188
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test # (Repeat) 1	0.192
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test # 1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # 2	0.046
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.20 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.075
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test # 2	0.191
		0.20 g/210L Test # (Repeat) 2	0.194
		0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.08 g/210L Test # (Repeat) 3	0.076
		0.20 g/210L Test # 3	0.192
		0.20 g/210L Test # (Repeat) 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

LOOSE HOSE. SEAL NOT TIGHT.

**Data Download Date/Time** 02/28/2019 19:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

<b>Agency</b> GLADES COUNTY S.O.	<b>Date</b> 03/28/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> MADRUGO, FELICIA,	<b>Time</b> 17:24:51	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 31517080A1
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 01/05/2020

**Number of Simulators** 2

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # (Repeat) 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # (Repeat) 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # (Repeat) 3	INT

**Remarks:**

Int Det: Ambient Fail.

**Data Download Date/Time** 03/28/2019 17:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

**Agency** GLADES COUNTY S.O.      **Date** 02/25/2019      **Type** Department Inspection  
**Inspector** GRAHAM, THOMAS, J      **Time** 09:34:39      **Software** 8100.27

**0.05 Lot #** 201707D      **0.08 Lot #** 201707E      **0.20 Lot #** 201707C      **0.08 Gas Lot #** AG805701  
**Expiration** 07/25/2019      **Expiration** 07/25/2019      **Expiration** 07/24/2019      **Expiration** 02/26/2020

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.195
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.196
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.195
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.196
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0009
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

---

**Data Download Date/Time** 02/25/2019 09:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

<b>Agency</b> TALLAHASSEE PD	<b>Date</b> 01/25/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> MAHAN, TERRY,	<b>Time</b> 12:18:32	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 34416080A2
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 02/05/2019

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/25/2019 12:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

<b>Agency</b> TALLAHASSEE PD	<b>Date</b> 02/22/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> MAHAN, TERRY,	<b>Time</b> 11:07:54	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 32518080A1
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 02/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/22/2019 11:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

<b>Agency</b> TALLAHASSEE PD	<b>Date</b> 03/27/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> MAHAN, TERRY,	<b>Time</b> 14:53:11	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 32518080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 02/05/2021

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/28/2019 11:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

<b>Agency</b> TALLAHASSEE PD	<b>Date</b> 04/30/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> MAHAN, TERRY,	<b>Time</b> 12:58:02	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 32518080A1
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 02/05/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 05/02/2019 11:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

<b>Agency</b> HAVANA PD	<b>Date</b> 01/07/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 10:12:31	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 12/05/2019

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.205
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 01/07/2019 10:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

<b>Agency</b> HAVANA PD	<b>Date</b> 02/12/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 10:38:26	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 12/05/2019

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 02/12/2019 10:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

<b>Agency</b> HAVANA PD	<b>Date</b> 03/13/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 09:51:11	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 03/13/2019 10:01



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

<b>Agency</b> HAVANA PD	<b>Date</b> 04/15/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 12:30:16	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> N
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019
<b>Number of Simulators</b> 5	<b>Compliance</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 04/15/2019 12:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

**Agency** HENDRY COUNTY S.O.      **Date** 04/30/2019      **Type** Agency Inspection  
**Inspector** NORWOOD,CHRISTOPHER,J      **Time** 02:01:47      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201808E      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 437738  
**Expiration** 02/06/2020      **Expiration** 08/22/2020      **Expiration** 02/22/2020      **Expiration** 06/08/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/30/2019 02:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

Agency FRANKLIN COUNTY SO      Date 04/16/2019      Type Department Inspection  
 Inspector MURPHY,PATRICK,J      Time 11:18:41      Software 8100.27

0.05 Lot # 201707D      0.08 Lot # 201707E      0.20 Lot # 201707C      0.08 Gas Lot # AG831804  
 Expiration 07/25/2019      Expiration 07/25/2019      Expiration 07/24/2019      Expiration 11/14/2020

Number of Simulators 5      Average Standard Deviation 0.0006      Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.076
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0013

**Remarks:**

---

**Data Download Date/Time** 04/16/2019 12:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

Agency FRANKLIN COUNTY SO      Date 01/15/2019      Type Agency Inspection  
 Inspector WILSON,STANLEY,T      Time 10:53:35      Software 8100.27

0.05 Lot # 201703A      0.08 Lot # 201709A      0.20 Lot # 201702B      0.08 Gas Lot # 686331  
 Expiration 03/07/2019      Expiration 09/12/2019      Expiration 02/23/2019      Expiration 01/31/2020

Number of Simulators 5      Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

Data Download Date/Time 01/15/2019 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

Agency FRANKLIN COUNTY SO      Date 02/19/2019      Type Agency Inspection  
 Inspector WILSON,STANLEY,T      Time 11:04:23      Software 8100.27

0.05 Lot # 201703A      0.08 Lot # 201709A      0.20 Lot # 201702B      0.08 Gas Lot # 686331  
 Expiration 03/07/2019      Expiration 09/12/2019      Expiration 02/23/2019      Expiration 01/31/2020

Number of Simulators 5      Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

Data Download Date/Time 02/19/2019 11:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

<b>Agency</b> FRANKLIN COUNTY SO	<b>Date</b> 03/19/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 11:40:42	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 686331
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 01/31/2020

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 03/19/2019 11:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

<b>Agency</b> FRANKLIN COUNTY SO	<b>Date</b> 04/23/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 12:26:59	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 686331
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 01/31/2020
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 04/23/2019 12:40



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

<b>Agency</b> JEFFERSON COUNTY SO	<b>Date</b> 01/09/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 10:31:00	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 12/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 01/09/2019 10:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

<b>Agency</b> JEFFERSON COUNTY SO	<b>Date</b> 02/05/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 10:59:41	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 12/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 02/05/2019 11:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

<b>Agency</b> JEFFERSON COUNTY SO	<b>Date</b> 03/20/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 13:56:38	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 03/20/2019 14:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

<b>Agency</b> JEFFERSON COUNTY SO	<b>Date</b> 04/08/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 11:10:17	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 04/08/2019 11:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

<b>Agency</b> JEFFERSON COUNTY SO	<b>Date</b> 01/09/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 10:35:51	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 12/05/2019

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 01/09/2019 10:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

<b>Agency</b> JEFFERSON COUNTY SO	<b>Date</b> 02/05/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 11:01:43	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 12/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 02/05/2019 11:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

<b>Agency</b> JEFFERSON COUNTY SO	<b>Date</b> 03/20/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 13:58:26	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 03/20/2019 14:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

<b>Agency</b> JEFFERSON COUNTY SO	<b>Date</b> 04/08/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 11:17:53	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 04/08/2019 11:24



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

<b>Agency</b> LIBERTY COUNTY SO	<b>Date</b> 01/03/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 10:20:12	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 12/05/2019

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 01/03/2019 11:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

<b>Agency</b> LIBERTY COUNTY SO	<b>Date</b> 02/06/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 09:53:35	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 12/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 02/06/2019 10:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

<b>Agency</b> LIBERTY COUNTY SO	<b>Date</b> 03/18/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 11:16:45	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 03/18/2019 11:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

<b>Agency</b> LIBERTY COUNTY SO	<b>Date</b> 04/16/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 09:35:50	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.045
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 04/16/2019 09:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

**Agency** LIBERTY COUNTY SO                      **Date** 04/24/2019                      **Type** Department Inspection  
**Inspector** GRAHAM, THOMAS, J                      **Time** 14:18:54                      **Software** 8100.27

**0.05 Lot #** 201707D                      **0.08 Lot #** 201707E                      **0.20 Lot #** 201707C                      **0.08 Gas Lot #** AG831804  
**Expiration** 07/25/2019                      **Expiration** 07/25/2019                      **Expiration** 07/24/2019                      **Expiration** 11/14/2020

**Number of Simulators** 5                      **Average Standard Deviation** 0.0004                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.077
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.077
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.047
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

---

**Data Download Date/Time** 04/24/2019 14:27

**Printed: 31 May 2019**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

<b>Agency</b> WAKULLA COUNTY SO	<b>Date</b> 01/09/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> MURPHY,PATRICK,J	<b>Time</b> 13:31:54	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

SKIPPED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 01/10/2019 11:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

Agency WAKULLA COUNTY SO      Date 01/10/2019      Type Department Inspection  
 Inspector MURPHY,PATRICK,J      Time 11:35:33      Software 8100.27

0.05 Lot # 201707D      0.08 Lot # 201707E      0.20 Lot # 201707C      0.08 Gas Lot # AG805701  
 Expiration 07/25/2019      Expiration 07/25/2019      Expiration 07/24/2019      Expiration 02/26/2020

Number of Simulators 5      Average Standard Deviation 0.0004      Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT



**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

**Data Download Date/Time** 01/10/2019 11:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

<b>Agency</b> WAKULLA COUNTY SO	<b>Date</b> 01/28/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 11:17:33	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 06717080A6
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 05/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 01/28/2019 11:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

<b>Agency</b> WAKULLA COUNTY SO	<b>Date</b> 02/11/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 10:50:14	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 06717080A6
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 05/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 02/11/2019 10:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

<b>Agency</b> WAKULLA COUNTY SO	<b>Date</b> 03/27/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 11:57:23	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 06717080A6
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 05/05/2019

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 03/27/2019 12:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

<b>Agency</b> WAKULLA COUNTY SO	<b>Date</b> 04/09/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 11:22:11	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 06717080A6
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 05/05/2019

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 04/09/2019 11:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

<b>Agency</b> WAKULLA COUNTY SO	<b>Date</b> 01/28/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 11:15:11	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 06717080A6
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 05/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 01/28/2019 11:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

<b>Agency</b> WAKULLA COUNTY SO	<b>Date</b> 02/11/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 10:45:42	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 06717080A6
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 05/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.205
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 02/11/2019 10:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

<b>Agency</b> WAKULLA COUNTY SO	<b>Date</b> 03/27/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 11:33:04	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test # (Repeat) 1	0.000
Minimum Sample Volume Check: OK		Interferent Detect Test # 1	0.000
Barometric Pressure Sensor Check: OK		Interferent Detect Test # (Repeat) 1	0.000
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test # (Repeat) 2	0.000
Interferent Detect Test: Interferent Detect	No	Interferent Detect Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # (Repeat) 2	0.000
Alcohol Free Test: OK	Yes	Alcohol Free Test # 3	0.000
0.05 g/210L Test: OK	No	Alcohol Free Test # (Repeat) 3	0.000
0.08 g/210L Test: OK	No	Interferent Detect Test # 3	0.000
0.20 g/210L Test: OK	No	Interferent Detect Test # (Repeat) 3	0.000
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

00: RFI Detect: CO TALKED ON RADIO.. Int Det:NO AIR COMING OUT OF SIM RETURN INTO THE SIM., . Non-compliance:SAME ISSUE..

**Data Download Date/Time** 03/27/2019 11:37



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

<b>Agency</b> ORANGE COUNTY S.O.	<b>Date</b> 01/16/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> MELVILLE,KELLY,M	<b>Time</b> 10:19:01	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 13917080A3
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 02/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 07/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/16/2019 10:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

<b>Agency</b> ORANGE COUNTY S.O.	<b>Date</b> 02/13/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> MELVILLE,KELLY,M	<b>Time</b> 10:10:55	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 13917080A3
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 07/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/13/2019 10:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

**Agency** ORANGE COUNTY S.O.                      **Date** 03/04/2019                      **Type** Department Inspection  
**Inspector** GRAHAM, THOMAS, J                      **Time** 12:21:09                      **Software** 8100.27

**0.05 Lot #** 201707D                      **0.08 Lot #** 201707E                      **0.20 Lot #** 201707C                      **0.08 Gas Lot #** AG805701  
**Expiration** 07/25/2019                      **Expiration** 07/25/2019                      **Expiration** 07/24/2019                      **Expiration** 02/26/2020

**Number of Simulators** 5                      **Average Standard Deviation** 0.0003                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

---

**Data Download Date/Time** 03/04/2019 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

<b>Agency</b> ORANGE COUNTY S.O.	<b>Date</b> 03/26/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> MILLS, ARMAND, G	<b>Time</b> 14:09:32	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 13917080A3
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 07/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

SIMULATORS PREPARED BY AI MELVILLE

**Data Download Date/Time** 03/26/2019 14:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

<b>Agency</b> ORANGE COUNTY S.O.	<b>Date</b> 04/17/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> MELVILLE,KELLY,M	<b>Time</b> 10:06:37	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 13518080A5
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 08/05/2020

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/18/2019 10:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

<b>Agency</b> ORANGE COUNTY SO	<b>Date</b> 01/16/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> MELVILLE,KELLY,M	<b>Time</b> 11:29:43	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 13917080A3
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 07/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/16/2019 11:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

<b>Agency</b> ORANGE COUNTY SO	<b>Date</b> 02/13/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> MELVILLE,KELLY,M	<b>Time</b> 11:29:22	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 13917080A3
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 07/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/13/2019 11:38



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

<b>Agency</b> ORANGE COUNTY SO	<b>Date</b> 03/26/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> MELVILLE,KELLY,M	<b>Time</b> 11:25:41	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 13917080A3
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 07/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/26/2019 11:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

<b>Agency</b> ORANGE COUNTY SO	<b>Date</b> 04/17/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> MELVILLE,KELLY,M	<b>Time</b> 11:10:00	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 13917080A3
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 07/05/2019

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/18/2019 11:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

<b>Agency</b> QUINCY PD	<b>Date</b> 01/14/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 09:34:29	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 12/05/2019

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 01/14/2019 09:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

<b>Agency</b> QUINCY PD	<b>Date</b> 02/13/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 12:01:31	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201703A	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 03/07/2019	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 12/05/2019

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 02/13/2019 12:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

<b>Agency</b> QUINCY PD	<b>Date</b> 03/25/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 11:44:51	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 03/25/2019 12:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

<b>Agency</b> QUINCY PD	<b>Date</b> 04/02/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> WILSON,STANLEY,T	<b>Time</b> 11:21:08	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 29017080A3
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

NONE

**Data Download Date/Time** 04/02/2019 11:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 01/06/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> LYONS,DANIEL,E	<b>Time</b> 22:29:54	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 646645
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	No	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/06/2019 22:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 01/06/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> LYONS,DANIEL,E	<b>Time</b> 23:51:06	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 646645
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/06/2019 23:55



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 01/24/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> LYONS,DANIEL,E	<b>Time</b> 23:23:55	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 646645
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/24/2019 23:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 02/07/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> LYONS,DANIEL,E	<b>Time</b> 18:07:27	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 646645
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/07/2019 18:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 01/06/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> LYONS,DANIEL,E	<b>Time</b> 22:19:22	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 646645
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

AF/MA;RISD MOUTH ALCOHOL SNM READING, RETESTED

**Data Download Date/Time** 01/06/2019 22:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

**Agency** OSCEOLA CORRECTIONS      **Date** 01/14/2019      **Type** Department Inspection  
**Inspector** GRAHAM, THOMAS, J      **Time** 11:26:06      **Software** 8100.27

**0.05 Lot #** 201707D      **0.08 Lot #** 201707E      **0.20 Lot #** 201707C      **0.08 Gas Lot #** AG805701  
**Expiration** 07/25/2019      **Expiration** 07/25/2019      **Expiration** 07/24/2019      **Expiration** 02/26/2020

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.195
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.196
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.196
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.196
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

**Data Download Date/Time** 01/14/2019 11:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 01/24/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> LYONS,DANIEL,E	<b>Time</b> 23:29:17	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 1060911
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/11/2021
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/24/2019 23:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 02/07/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> LYONS,DANIEL,E	<b>Time</b> 19:24:34	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 1060911
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/11/2021
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.075
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/07/2019 19:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 03/18/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> LYONS,DANIEL,E	<b>Time</b> 23:43:01	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 1060911
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/11/2021
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/18/2019 23:49



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

<b>Agency</b> OSCEOLA CORRECTIONS	<b>Date</b> 04/03/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> LYONS,DANIEL,E	<b>Time</b> 18:29:27	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 1060911
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/11/2021
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/03/2019 18:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

<b>Agency</b> COCOA BEACH P.D.	<b>Date</b> 01/21/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> SCHEINER,JAMES,R	<b>Time</b> 13:31:22	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 17618080A3
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 08/05/2020

  

<b>Number of Simulators</b> 4	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/21/2019 13:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

<b>Agency</b> COCOA BEACH P.D.	<b>Date</b> 02/21/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> BETTS, RONALD, E	<b>Time</b> 09:56:56	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 17618080A3
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 08/05/2020

**Number of Simulators** 2

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.183
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test # (Repeat) 1	0.195
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test # 1	0.000
Alcohol Free Test: OK	Yes	Interferent Detect Test # (Repeat) 1	INT
0.05 g/210L Test: OK	Yes	Alcohol Free Test # 2	0.000
0.08 g/210L Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.20 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.078
		0.20 g/210L Test # 2	0.191
		0.20 g/210L Test # (Repeat) 2	0.195
		0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	0.000
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.08 g/210L Test # (Repeat) 3	0.079
		0.20 g/210L Test # 3	0.193
		0.20 g/210L Test # (Repeat) 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	0.000
		Interferent Detect Test # (Repeat) 3	INT

**Remarks:**

Int Det: . 08: Control Outside Tolerance. 20: Control Outside SEE NOTES

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

---

**Data Download Date/Time** 03/06/2019 08:17

**Printed: 31 May 2019**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

**Agency** COCOA BEACH P.D.                      **Date** 03/06/2019                      **Type** Department Inspection  
**Inspector** PLATT,SHAYLA,D                      **Time** 11:42:35                      **Software** 8100.27

**0.05 Lot #** 201707D                      **0.08 Lot #** 201707E                      **0.20 Lot #** 201707C                      **0.08 Gas Lot #** AG831804  
**Expiration** 07/25/2019                      **Expiration** 07/25/2019                      **Expiration** 07/24/2019                      **Expiration** 11/14/2020

**Number of Simulators** 5                      **Average Standard Deviation** 0.0007                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0020
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

---

**Data Download Date/Time** 03/06/2019 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

<b>Agency</b> COCOA BEACH P.D.	<b>Date</b> 03/22/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> BETTS, RONALD, E	<b>Time</b> 12:53:11	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 17618080A3
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 08/05/2020

**Number of Simulators** 2

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.043
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.05 g/210L Test # (Repeat) 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.192
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

05: Control Outside Tolerance.

**Data Download Date/Time** 04/12/2019 13:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency TITUSVILLE P.D.	Date 01/24/2019	Type Agency Inspection	
Inspector SANCHEZ,JASON,M	Time 12:44:51	Software 8100.27	

0.05 Lot # 201802B	0.08 Lot # 201709A	0.20 Lot # 201702B	0.08 Gas Lot #
Expiration 02/06/2020	Expiration 09/12/2019	Expiration 02/23/2019	Expiration

Number of Simulators 4

Compliance No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.186
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.187
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.189
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.190
0.20 g/210L Test: OK	No	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.190
		0.20 g/210L Test # (Repeat) 3	0.189
		Interferent Detect Test # 3	INT

**Remarks:**

20: Control Outside Tolerance, Control Outside Tolerance. OUT OF TOL.

Data Download Date/Time 01/25/2019 19:02



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

<b>Agency</b> TITUSVILLE P.D.	<b>Date</b> 01/25/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> SANCHEZ,JASON,M	<b>Time</b> 17:35:59	<b>Software</b> 8100.27	

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** 4 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

A F / M A: . Non-compliance: .

**Data Download Date/Time** 01/25/2019 19:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

<b>Agency</b> TITUSVILLE P.D.	<b>Date</b> 01/25/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> SANCHEZ,JASON,M	<b>Time</b> 18:48:01	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 29017080A2
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 12/05/2019

**Number of Simulators** 4

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/25/2019 19:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

**Agency** TITUSVILLE P.D.                      **Date** 02/26/2019                      **Type** Agency Inspection  
**Inspector** SANCHEZ,JASON,M                      **Time** 11:34:47                      **Software** 8100.27

**0.05 Lot #** 201802B                      **0.08 Lot #** 201709A                      **0.20 Lot #** 201702B                      **0.08 Gas Lot #**  
**Expiration** 02/06/2020                      **Expiration** 09/12/2019                      **Expiration** 02/23/2019                      **Expiration**

**Number of Simulators** 4

**Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.186
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.189
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test # 2	0.045
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.190
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.190
0.20 g/210L Test: OK	No	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	No	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.075
		0.20 g/210L Test # 3	0.191
		0.20 g/210L Test # (Repeat) 3	0.191
		Interferent Detect Test # 3	INT

**Remarks:**

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 02/26/2019 11:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

<b>Agency</b> TITUSVILLE P.D.	<b>Date</b> 02/27/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> SANCHEZ,JASON,M	<b>Time</b> 21:17:19	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 29017080A2
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 12/05/2019

**Number of Simulators** 4

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/27/2019 21:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

<b>Agency</b> TITUSVILLE P.D.	<b>Date</b> 03/27/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> HOLDEN, ROBERT, A	<b>Time</b> 16:17:10	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 29017080A2
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 12/05/2019

**Number of Simulators** 4

**Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.191
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.08 g/210L Test # (Repeat) 3	0.076
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

08: Control Outside Tolerance. Non-compliance: DID NOT MEET STANDARD.

**Data Download Date/Time** 03/27/2019 16:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

<b>Agency</b> TITUSVILLE P.D.	<b>Date</b> 04/09/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> HOLDEN, ROBERT, A	<b>Time</b> 10:56:42	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201811C	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 29017080A2
<b>Expiration</b> 11/13/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019

**Number of Simulators** 4

**Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.190
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.08 g/210L Test # (Repeat) 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

08: Control Outside Tolerance. OUTSIDE TOLERANCE.

**Data Download Date/Time** 04/09/2019 11:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

Agency SEMINOLE COUNTY S.O.      Date 01/27/2019      Type Agency Inspection  
 Inspector BETHAM,KEITH,D      Time 01:44:55      Software 8100.27

0.05 Lot # 201802B      0.08 Lot # 201709A      0.20 Lot # 201702B      0.08 Gas Lot # 282638  
 Expiration 02/06/2020      Expiration 09/12/2019      Expiration 02/23/2019      Expiration 02/15/2019

Number of Simulators 5      Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

Data Download Date/Time 01/28/2019 20:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

<b>Agency</b> SEMINOLE COUNTY S.O.	<b>Date</b> 02/15/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> BETHAM,KEITH,D	<b>Time</b> 23:14:33	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 915310
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 03/07/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.077
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

08: HOSE DETACHED RERUN.

**Data Download Date/Time** 02/17/2019 20:59



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

Agency SEMINOLE COUNTY S.O.      Date 02/16/2019      Type Agency Inspection  
 Inspector BETHAM,KEITH,D      Time 00:26:56      Software 8100.27

0.05 Lot # 201802B      0.08 Lot # 201709A      0.20 Lot # 201702B      0.08 Gas Lot # 915310  
 Expiration 02/06/2020      Expiration 09/12/2019      Expiration 02/23/2019      Expiration 03/07/2021

Number of Simulators 5      Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	0.000
		Interferent Detect Test # (Repeat) 3	INT

**Remarks:**

Int Det:REPLACED DEFECTIVE INLET TUBE.

Data Download Date/Time 02/17/2019 20:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

**Agency** SEMINOLE COUNTY S.O.      **Date** 03/31/2019      **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D      **Time** 01:25:44      **Software** 8100.27

**0.05 Lot #** 201802B      **0.08 Lot #** 201709A      **0.20 Lot #** 201802G      **0.08 Gas Lot #** 915310  
**Expiration** 02/06/2020      **Expiration** 09/12/2019      **Expiration** 02/22/2020      **Expiration** 03/07/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	0.000
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	0.000
		Interferent Detect Test # (Repeat) 3	INT

**Remarks:**

Int Det:NO ACETONE WAS ADDED TO SOLUTION.

**Data Download Date/Time** 03/31/2019 20:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

<b>Agency</b> SEMINOLE COUNTY S.O.	<b>Date</b> 04/26/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> BETHAM,KEITH,D	<b>Time</b> 00:48:44	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 915310
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 03/07/2021

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	INT
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

08: Interferent Detect.

**Data Download Date/Time** 04/27/2019 20:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

<b>Agency</b> SEMINOLE COUNTY S.O.	<b>Date</b> 01/27/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> BETHAM,KEITH,D	<b>Time</b> 01:45:38	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 282638
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 02/15/2019

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.051
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

MOUTH ALCOHOL INSTEAD OF ALCOHOL FREE. 08: Control Outside Tolerance GAS NOT ATTACHED

**Data Download Date/Time** 01/28/2019 20:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

<b>Agency</b> SEMINOLE COUNTY S.O.	<b>Date</b> 02/22/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> BETHAM,KEITH,D	<b>Time</b> 21:57:26	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201702B	<b>0.08 Gas Lot #</b> 915310
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/23/2019	<b>Expiration</b> 03/07/2021

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/04/2019 00:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

<b>Agency</b> SEMINOLE COUNTY S.O.	<b>Date</b> 03/31/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> BETHAM,KEITH,D	<b>Time</b> 01:30:46	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802B	<b>0.08 Gas Lot #</b> 915310
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/06/2020	<b>Expiration</b> 03/07/2021

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

CORRECT.20 LOT#201802G EXP.03/07/20

**Data Download Date/Time** 03/31/2019 20:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

<b>Agency</b> SEMINOLE COUNTY S.O.	<b>Date</b> 04/25/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> BETHAM,KEITH,D	<b>Time</b> 23:44:34	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 915310
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 03/07/2021

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/27/2019 20:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

<b>Agency</b> COCOA P.D.	<b>Date</b> 01/30/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> FRAZIER,JAMES,H	<b>Time</b> 20:16:54	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 02618080A3
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 03/05/2020

**Number of Simulators** 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/30/2019 20:23



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

<b>Agency</b> COCOA P.D.	<b>Date</b> 02/20/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> FRAZIER,JAMES,H	<b>Time</b> 21:58:55	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 02618080A3
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 03/05/2020

  

<b>Number of Simulators</b> 4	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/20/2019 22:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

<b>Agency</b> COCOA P.D.	<b>Date</b> 03/19/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> FRAZIER,JAMES,H	<b>Time</b> 21:50:53	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 02618080A3
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 03/05/2020

  

<b>Number of Simulators</b> 4	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/19/2019 22:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001000

<b>Agency</b> COCOA P.D.	<b>Date</b> 04/23/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> FRAZIER,JAMES,H	<b>Time</b> 20:22:42	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201709A	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 02618080A3
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 09/12/2019	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 03/05/2020

**Number of Simulators** 4

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/23/2019 20:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

<b>Agency</b> SATELLITE BEACH P.D.	<b>Date</b> 01/23/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> TRIEBELL,DON,R	<b>Time</b> 05:28:00	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 646645
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 12/05/2019
<b>Number of Simulators</b> 4		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/23/2019 05:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

<b>Agency</b> SATELLITE BEACH P.D.	<b>Date</b> 02/11/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> TRIEBELL,DON,R	<b>Time</b> 05:21:14	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 1076482
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 01/15/2022
<b>Number of Simulators</b> 4		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/11/2019 05:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

<b>Agency</b> SATELLITE BEACH P.D.	<b>Date</b> 03/30/2019	<b>Type</b> Agency Inspection	
<b>Inspector</b> TRIEBELL,DON,R	<b>Time</b> 06:35:32	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 1076482
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 01/15/2022
<b>Number of Simulators</b> 4		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/01/2019 07:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

<b>Agency</b> SATELLITE BEACH P.D.	<b>Date</b> 04/26/2019	<b>Type</b> Agency Inspection
<b>Inspector</b> TRIEBELL,DON,R	<b>Time</b> 18:44:46	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 201802B	<b>0.08 Lot #</b> 201808E	<b>0.20 Lot #</b> 201802G	<b>0.08 Gas Lot #</b> 1076482
<b>Expiration</b> 02/06/2020	<b>Expiration</b> 08/22/2020	<b>Expiration</b> 02/22/2020	<b>Expiration</b> 01/15/2022

  

<b>Number of Simulators</b> 4	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/26/2019 18:52