

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

**Agency** MELBOURNE P.D. **Date** 01/28/2021 **Type** Agency Inspection  
**Inspector** WHITRIGHT, MARK, D **Time** 11:44:08 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1298240  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 03/19/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/01/2021 08:36

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

**Agency** MELBOURNE P.D. **Date** 02/23/2021 **Type** Agency Inspection  
**Inspector** LIVINGSTON, CHRISTOPHER, S **Time** 12:05:51 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1298240  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 03/19/2023

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N

**Data Download Date/Time** 02/25/2021 15:53

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

**Agency** MELBOURNE P.D. **Date** 03/30/2021 **Type** Agency Inspection  
**Inspector** LINEHAN,DANIEL,J **Time** 09:07:41 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 4 **Compliance** No

| Check or Test                               | Result | Test                  | Result |
|---------------------------------------------|--------|-----------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1 | 0.000  |
| Date and/or Time Adjusted                   | No     | Alcohol Free Test # 2 | 0.000  |
| Minimum Sample Volume Check: OK             |        | Alcohol Free Test # 3 | 0.000  |
| Barometric Pressure Sensor Check: OK        |        |                       |        |
| Alcohol Free Subject Test: 0.000            | Yes    |                       |        |
| Mouth Alcohol Test: Slope Not Met           | Yes    |                       |        |
| Interferent Detect Test: Interferent Detect | No     |                       |        |
| Diagnostic Check (Post-Inspection): OK      | No     |                       |        |
| Alcohol Free Test: OK                       | Yes    |                       |        |
| 0.05 g/210L Test: OK                        | No     |                       |        |
| 0.08 g/210L Test: OK                        | No     |                       |        |
| 0.20 g/210L Test: OK                        | No     |                       |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                       |        |

#### Remarks:

Int Det: Ambient Fail, Ambient Fail. Non-compliance:AMB FAIL.

**Data Download Date/Time** 03/31/2021 06:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency MELBOURNE P.D. Date 03/30/2021 Type Agency Inspection  
Inspector LINEHAN,DANIEL,J Time 10:04:21 Software 8100.27

0.05 Lot # 201910D 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # 1298240  
Expiration 10/22/2021 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 03/19/2023

Number of Simulators 4 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 03/31/2021 06:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

|                                   |                        |                               |
|-----------------------------------|------------------------|-------------------------------|
| <b>Agency</b> MELBOURNE P.D.      | <b>Date</b> 04/20/2021 | <b>Type</b> Agency Inspection |
| <b>Inspector</b> LINEHAN,DANIEL,J | <b>Time</b> 07:58:26   | <b>Software</b> 8100.27       |

|                              |                              |                   |                       |
|------------------------------|------------------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> | <b>Expiration</b>     |

**Number of Simulators** 4

**Compliance** No

| Check or Test                               | Result | Test                        | Result |
|---------------------------------------------|--------|-----------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1       | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1        | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1        | 0.074  |
| Barometric Pressure Sensor Check: OK        |        | Interferent Detect Test # 1 | INT    |
| Alcohol Free Subject Test: 0.000            | Yes    | Alcohol Free Test # 2       | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.05 g/210L Test # 2        | 0.046  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.08 g/210L Test # 2        | 0.074  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 2 | INT    |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # 3       | 0.000  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # 3        | 0.047  |
| 0.08 g/210L Test: OK                        | No     | 0.08 g/210L Test # 3        | 0.075  |
| 0.20 g/210L Test: OK                        | No     | Interferent Detect Test # 3 | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                             |        |

#### Remarks:

08: Control Outside Tolerance, Ambient Fail. Non-compliance:.08 BELOW LVL.

**Data Download Date/Time** 04/21/2021 07:11

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency MELBOURNE P.D. Date 04/20/2021 Type Agency Inspection  
Inspector LINEHAN,DANIEL,J Time 08:54:38 Software 8100.27

0.05 Lot # 201910D 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # 1408275  
Expiration 10/22/2021 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 11/19/2023

Number of Simulators 4 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # (Repeat) 1    | 0.190  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

20: Ambient Fail.

Data Download Date/Time 04/21/2021 07:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

**Agency** MELBOURNE P.D. **Date** 05/27/2021 **Type** Agency Inspection  
**Inspector** LIVINGSTON, CHRISTOPHER, S **Time** 11:37:30 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1408275  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 11/19/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/01/2021 06:35

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

**Agency** MELBOURNE P.D. **Date** 06/28/2021 **Type** Agency Inspection  
**Inspector** LIVINGSTON, CHRISTOPHER, S **Time** 11:26:09 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1408275  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 11/19/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/29/2021 09:20

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency MELBOURNE P.D. Date 07/28/2021 Type Agency Inspection  
Inspector LIVINGSTON, CHRISTOPHER, S Time 15:12:37 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # 1408275  
Expiration 01/12/2023 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 11/19/2023

Number of Simulators 4 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 07/29/2021 06:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

**Agency** MELBOURNE P.D. **Date** 08/31/2021 **Type** Agency Inspection  
**Inspector** LIVINGSTON, CHRISTOPHER, S **Time** 09:12:28 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1408275  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 11/19/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 09/03/2021 07:49

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency MELBOURNE P.D. Date 09/29/2021 Type Agency Inspection  
Inspector WHITRIGHT, MARK, D Time 15:43:39 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # 1482048  
Expiration 01/12/2023 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 04/21/2024

Number of Simulators 4 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.190  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 10/11/2021 14:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency MELBOURNE P.D. Date 10/25/2021 Type Agency Inspection  
Inspector WHITRIGHT, MARK, D Time 12:55:23 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # 1482048  
Expiration 01/12/2023 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 04/21/2024

Number of Simulators 4 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 10/25/2021 15:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency MELBOURNE P.D. Date 11/30/2021 Type Agency Inspection  
Inspector WHITRIGHT, MARK, D Time 11:18:01 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # 1482048  
Expiration 01/12/2023 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 04/21/2024

Number of Simulators 4 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 12/08/2021 11:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

**Agency** MELBOURNE P.D. **Date** 12/07/2021 **Type** Agency Inspection  
**Inspector** REYES-RIVERA,DAVID,E **Time** 12:37:06 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED.

**Data Download Date/Time** 12/08/2021 11:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MELBOURNE P.D.            | <b>Date</b> 12/08/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 11:10:45   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG115904 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 06/08/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 12/08/2021 11:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency MELBOURNE P.D. Date 12/14/2021 Type Agency Inspection  
Inspector WHITRIGHT, MARK, D Time 11:59:33 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # 1482048  
Expiration 01/12/2023 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 04/21/2024

Number of Simulators 4 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.075  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.075  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 12/14/2021 12:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

**Agency** INDIAN HARBOUR PD      **Date** 02/14/2021      **Type** Agency Inspection  
**Inspector** WIMMER,MATTHEW,S      **Time** 13:02:46      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** No

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | Interferent Detect Test # 1          | 0.007  |
| Minimum Sample Volume Check: OK             |        | Interferent Detect Test # (Repeat) 1 | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 2          | 0.007  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # (Repeat) 2 | 0.000  |
| Interferent Detect Test: Interferent Detect | No     | Alcohol Free Test # 3                | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 3          | 0.000  |
| Alcohol Free Test: OK                       | Yes    | Interferent Detect Test # (Repeat) 3 | 0.000  |
| 0.05 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Test: OK                        | No     |                                      |        |
| 0.20 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                                      |        |

#### Remarks:

Int Det: VALVE NEEDED TO BE TIGHTENED, . Non-compliance: INTERFERENT UNDETECTED.

**Data Download Date/Time** 02/18/2021 05:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

**Agency** INDIAN HARBOUR PD      **Date** 02/14/2021      **Type** Agency Inspection  
**Inspector** WIMMER,MATTHEW,S      **Time** 13:35:21      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** No

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | Interferent Detect Test # 1          | 0.000  |
| Minimum Sample Volume Check: OK             |        | Interferent Detect Test # (Repeat) 1 | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 2          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # (Repeat) 2 | 0.000  |
| Interferent Detect Test: Interferent Detect | No     | Alcohol Free Test # 3                | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 3          | 0.000  |
| Alcohol Free Test: OK                       | Yes    | Interferent Detect Test # (Repeat) 3 | 0.000  |
| 0.05 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Test: OK                        | No     |                                      |        |
| 0.20 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                                      |        |

#### Remarks:

Int Det: NOT DETECTED, . Non-compliance: INTEFERENT UNDETECTED.

**Data Download Date/Time** 02/18/2021 05:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

|                                   |                        |                               |
|-----------------------------------|------------------------|-------------------------------|
| <b>Agency</b> INDIAN HARBOUR PD   | <b>Date</b> 02/14/2021 | <b>Type</b> Agency Inspection |
| <b>Inspector</b> WIMMER,MATTHEW,S | <b>Time</b> 14:39:50   | <b>Software</b> 8100.27       |

|                              |                              |                   |                       |
|------------------------------|------------------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> | <b>Expiration</b>     |

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.057  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1 | 0.055  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1          | 0.087  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Test # (Repeat) 1 | 0.086  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1   | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2         | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.05 g/210L Test # 2          | 0.056  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # (Repeat) 2 | 0.055  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2          | 0.087  |
| 0.08 g/210L Test: OK                        | No     | 0.08 g/210L Test # (Repeat) 2 | 0.085  |
| 0.20 g/210L Test: OK                        | No     | Interferent Detect Test # 2   | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | Alcohol Free Test # 3         | 0.000  |
|                                             |        | 0.05 g/210L Test # 3          | 0.055  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3 | 0.054  |
|                                             |        | 0.08 g/210L Test # 3          | 0.086  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3 | 0.085  |
|                                             |        | Interferent Detect Test # 3   | INT    |

**Remarks:**

05: Control Outside Tolerance. 08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: OUTSIDE TOLERANCE BY 2.

**Data Download Date/Time** 02/18/2021 05:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

**Agency** INDIAN HARBOUR PD      **Date** 02/14/2021      **Type** Agency Inspection  
**Inspector** WIMMER,MATTHEW,S      **Time** 15:25:37      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 28719080A1  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 12/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.055  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.085  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.208  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.055  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.085  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.208  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.054  |
|                                             |        | 0.08 g/210L Test # 3             | 0.085  |
|                                             |        | 0.20 g/210L Test # 3             | 0.208  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NO REMARKS

**Data Download Date/Time** 02/18/2021 05:55

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

**Agency** INDIAN HARBOUR PD      **Date** 03/28/2021      **Type** Agency Inspection  
**Inspector** WIMMER,MATTHEW,S      **Time** 13:56:44      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 28719080A1  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 12/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.052  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.083  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.052  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.083  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/05/2021 14:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

**Agency** INDIAN HARBOUR PD      **Date** 04/30/2021      **Type** Agency Inspection  
**Inspector** MCGRORY,DIARMUID,J      **Time** 13:42:09      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

PASSED.

**Data Download Date/Time** 05/17/2021 18:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

**Agency** INDIAN HARBOUR PD      **Date** 04/30/2021      **Type** Agency Inspection  
**Inspector** MCGRORY, DIARMUID, J      **Time** 14:42:43      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 28719080A1  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 12/05/2021

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: Subject Test Refused.

**Data Download Date/Time** 05/17/2021 18:40

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

**Agency** INDIAN HARBOUR PD      **Date** 05/28/2021      **Type** Agency Inspection  
**Inspector** MCGRORY,DIARMUID,J      **Time** 07:36:52      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 28719080A1  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 12/05/2021

**Number of Simulators** 3      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/05/2021 02:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

**Agency** INDIAN HARBOUR PD      **Date** 07/20/2021      **Type** Agency Inspection  
**Inspector** WIMMER,MATTHEW,S      **Time** 16:52:23      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 28719080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 12/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NO REMARKS

**Data Download Date/Time** 07/21/2021 11:26

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

**Agency** INDIAN HARBOUR PD      **Date** 08/26/2021      **Type** Agency Inspection  
**Inspector** WIMMER,MATTHEW,S      **Time** 12:01:12      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 28719080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 12/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NO REMARKS

**Data Download Date/Time** 09/01/2021 11:38

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

|                                   |                        |                               |
|-----------------------------------|------------------------|-------------------------------|
| <b>Agency</b> INDIAN HARBOUR PD   | <b>Date</b> 09/30/2021 | <b>Type</b> Agency Inspection |
| <b>Inspector</b> WIMMER,MATTHEW,S | <b>Time</b> 10:42:23   | <b>Software</b> 8100.27       |

|                              |                              |                              |                       |
|------------------------------|------------------------------|------------------------------|-----------------------|
| <b>0.05 Lot #</b> 202101C    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> 01/12/2023 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b>     |

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1          | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1          | 0.172  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1 | 0.185  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1   | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2         | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.05 g/210L Test # 2          | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2          | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2          | 0.149  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2 | 0.191  |
| 0.20 g/210L Test: OK                        | No     | Interferent Detect Test # 2   | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | Alcohol Free Test # 3         | 0.000  |
|                                             |        | 0.05 g/210L Test # 3          | 0.050  |
|                                             |        | 0.08 g/210L Test # 3          | 0.080  |
|                                             |        | 0.20 g/210L Test # 3          | 0.142  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3 | 0.193  |
|                                             |        | Interferent Detect Test # 3   | INT    |

**Remarks:**

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 10/01/2021 13:53

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

**Agency** INDIAN HARBOUR PD      **Date** 09/30/2021      **Type** Agency Inspection  
**Inspector** WIMMER,MATTHEW,S      **Time** 11:39:12      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** No

| Check or Test                               | Result | Test                           | Result |
|---------------------------------------------|--------|--------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1          | INT    |
| Date and/or Time Adjusted                   | No     | Alcohol Free Test # (Repeat) 1 | INT    |
| Minimum Sample Volume Check: OK             |        |                                |        |
| Barometric Pressure Sensor Check: OK        |        |                                |        |
| Alcohol Free Subject Test: 0.000            | Yes    |                                |        |
| Mouth Alcohol Test: Slope Not Met           | Yes    |                                |        |
| Interferent Detect Test: Interferent Detect | No     |                                |        |
| Diagnostic Check (Post-Inspection): OK      | No     |                                |        |
| Alcohol Free Test: OK                       | No     |                                |        |
| 0.05 g/210L Test: OK                        | No     |                                |        |
| 0.08 g/210L Test: OK                        | No     |                                |        |
| 0.20 g/210L Test: OK                        | No     |                                |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                                |        |

#### Remarks:

00: Interferent Detect, Interferent Detect. Non-compliance: INTERFERENT DETECT.

**Data Download Date/Time** 10/01/2021 13:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

**Agency** INDIAN HARBOUR PD      **Date** 09/30/2021      **Type** Agency Inspection  
**Inspector** WIMMER,MATTHEW,S      **Time** 12:39:59      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 28719080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 12/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 10/01/2021 13:53

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

**Agency** INDIAN HARBOUR PD      **Date** 10/20/2021      **Type** Agency Inspection  
**Inspector** WIMMER,MATTHEW,S      **Time** 14:05:53      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 28719080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 12/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NO REMARKS

**Data Download Date/Time** 10/20/2021 15:15

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> INDIAN HARBOUR PD         | <b>Date</b> 11/10/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 10:24:16   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG026705 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 09/23/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 11/10/2021 10:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

**Agency** INDIAN HARBOUR PD      **Date** 11/24/2021      **Type** Agency Inspection  
**Inspector** WIMMER,MATTHEW,S      **Time** 13:17:34      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 28719080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 12/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 11/24/2021 14:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

**Agency** BREVARD COUNTY S.O.      **Date** 01/21/2021      **Type** Agency Inspection  
**Inspector** ROBIN,GASTON,A      **Time** 12:08:14      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 22419080A2  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 10/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/24/2021 10:36

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

**Agency** BREVARD COUNTY S.O. **Date** 02/24/2021 **Type** Agency Inspection  
**Inspector** ROBIN,GASTON,A **Time** 14:49:27 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 22419080A2  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 10/05/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/24/2021 15:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

**Agency** BREVARD COUNTY S.O. **Date** 03/25/2021 **Type** Agency Inspection  
**Inspector** BIVINS, HEATHER, C **Time** 18:02:25 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 22419080A2  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 10/05/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 04/02/2021 12:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

**Agency** BREVARD COUNTY S.O. **Date** 04/02/2021 **Type** Agency Inspection  
**Inspector** ROBIN,GASTON,A **Time** 14:01:51 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 22419080A2  
**Expiration** 10/22/2021 **Expiration** 07/12/2022 **Expiration** 05/12/2022 **Expiration** 10/05/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 04/02/2021 14:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

**Agency** BREVARD COUNTY S.O. **Date** 05/27/2021 **Type** Agency Inspection  
**Inspector** ROBIN,GASTON,A **Time** 20:06:11 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 22419080A2  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 10/05/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/27/2021 20:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

**Agency** BREVARD COUNTY S.O. **Date** 06/30/2021 **Type** Agency Inspection  
**Inspector** ROBIN,GASTON,A **Time** 19:44:13 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 22419080A2  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 10/05/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 07/23/2021 20:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

**Agency** BREVARD COUNTY S.O. **Date** 07/23/2021 **Type** Agency Inspection  
**Inspector** ROBIN,GASTON,A **Time** 23:36:59 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 22419080A2  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 10/05/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 08/19/2021 11:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

**Agency** BREVARD COUNTY S.O. **Date** 08/19/2021 **Type** Agency Inspection  
**Inspector** ROBIN,GASTON,A **Time** 13:33:28 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 22419080A2  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 10/05/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/19/2021 15:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

**Agency** BREVARD COUNTY S.O. **Date** 09/21/2021 **Type** Agency Inspection  
**Inspector** ROBIN,GASTON,A **Time** 10:41:27 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 22419080A2  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 10/05/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/21/2021 10:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

**Agency** BREVARD COUNTY S.O.      **Date** 10/28/2021      **Type** Agency Inspection  
**Inspector** ROBIN,GASTON,A      **Time** 16:15:31      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 24620080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 10/28/2021 17:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

**Agency** BREVARD COUNTY S.O.      **Date** 11/17/2021      **Type** Agency Inspection  
**Inspector** WAGNER,ROBERT,F      **Time** 12:42:09      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 25721080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 11/29/2021 14:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

|                                   |                        |                                   |
|-----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> BREVARD COUNTY S.O. | <b>Date</b> 11/30/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,     | <b>Time</b> 09:52:25   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0007      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.081  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.082  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.081  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0009 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0007 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0010 |

#### Remarks:

Data Download Date/Time 11/30/2021 09:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

**Agency** BREVARD COUNTY S.O. **Date** 12/29/2021 **Type** Agency Inspection  
**Inspector** WAGNER,ROBERT,F **Time** 11:28:54 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202005A **0.08 Gas Lot #** 25721080A2  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 05/12/2022 **Expiration** 11/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Time-Date changed.

**Data Download Date/Time** 12/29/2021 11:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

|                               |                              |                               |                                  |
|-------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> PALM BAY P.D.   | <b>Date</b> 01/29/2021       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LUTZ,EDWIN,J | <b>Time</b> 19:15:26         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201910D     | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> 24620080A2 |
| <b>Expiration</b> 10/22/2021  | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 11/05/2022     |
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/29/2021 19:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

**Agency** PALM BAY P.D. **Date** 02/27/2021 **Type** Agency Inspection  
**Inspector** LUTZ,EDWIN,J **Time** 17:04:03 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 24620080A2  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 11/05/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/01/2021 12:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

**Agency** PALM BAY P.D. **Date** 03/21/2021 **Type** Agency Inspection  
**Inspector** LUTZ,EDWIN,J **Time** 17:12:13 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 24620080A2  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 11/05/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/26/2021 15:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

|                               |                              |                               |                                  |
|-------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> PALM BAY P.D.   | <b>Date</b> 04/27/2021       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LUTZ,EDWIN,J | <b>Time</b> 18:34:49         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201910D     | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> 24620080A2 |
| <b>Expiration</b> 10/22/2021  | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 11/05/2022     |
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/27/2021 18:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

|                               |                              |                               |                                  |
|-------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> PALM BAY P.D.   | <b>Date</b> 05/28/2021       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LUTZ,EDWIN,J | <b>Time</b> 16:41:12         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201910D     | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> 24620080A2 |
| <b>Expiration</b> 10/22/2021  | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 11/05/2022     |
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/28/2021 10:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

|                               |                              |                               |                                  |
|-------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> PALM BAY P.D.   | <b>Date</b> 06/28/2021       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LUTZ,EDWIN,J | <b>Time</b> 15:25:35         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201910D     | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> 24620080A2 |
| <b>Expiration</b> 10/22/2021  | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 11/05/2022     |
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 06/28/2021 15:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

**Agency** PALM BAY P.D. **Date** 07/29/2021 **Type** Agency Inspection  
**Inspector** LUTZ,EDWIN,J **Time** 13:51:04 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 24620080A2  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 11/05/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/29/2021 14:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

**Agency** PALM BAY P.D. **Date** 08/26/2021 **Type** Agency Inspection  
**Inspector** LUTZ,EDWIN,J **Time** 12:20:34 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 24620080A2  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 11/05/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/26/2021 12:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

**Agency** PALM BAY P.D. **Date** 09/28/2021 **Type** Agency Inspection  
**Inspector** LUTZ,EDWIN,J **Time** 16:59:58 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 24620080A2  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 11/05/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/28/2021 17:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

|                               |                              |                               |                                  |
|-------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> PALM BAY P.D.   | <b>Date</b> 10/27/2021       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LUTZ,EDWIN,J | <b>Time</b> 16:43:19         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202101C     | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> 24620080A2 |
| <b>Expiration</b> 01/12/2023  | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 11/05/2022     |
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/27/2021 16:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

|                                |                        |                                   |
|--------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PALM BAY P.D.    | <b>Date</b> 11/04/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO, ISRAEL, | <b>Time</b> 12:52:30   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 11/04/2021 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

|                               |                              |                               |                                  |
|-------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> PALM BAY P.D.   | <b>Date</b> 01/29/2021       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LUTZ,EDWIN,J | <b>Time</b> 19:22:26         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201910D     | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> 22620080A2 |
| <b>Expiration</b> 10/22/2021  | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 10/05/2022     |
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/06/2021 00:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

|                               |                              |                               |                                  |
|-------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> PALM BAY P.D.   | <b>Date</b> 02/27/2021       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LUTZ,EDWIN,J | <b>Time</b> 16:59:24         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201910D     | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> 24620080A2 |
| <b>Expiration</b> 10/22/2021  | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 11/05/2022     |
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/01/2021 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

**Agency** PALM BAY P.D. **Date** 03/21/2021 **Type** Agency Inspection  
**Inspector** LUTZ,EDWIN,J **Time** 17:03:04 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 24620080A2  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 11/05/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 04/09/2021 14:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

**Agency** PALM BAY P.D. **Date** 04/27/2021 **Type** Agency Inspection  
**Inspector** LUTZ,EDWIN,J **Time** 18:30:25 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 24620080A2  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 11/05/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/27/2021 18:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

|                                 |                        |                               |  |
|---------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> PALM BAY P.D.     | <b>Date</b> 05/28/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> LUTZ, EDWIN, J | <b>Time</b> 16:48:08   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 24620080A2 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 11/05/2022     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.077  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.199  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.000  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.048  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.077  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.080  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3               | INT    |

**Remarks:**

08: Control Outside Tolerance AI ERR; FORGOT CONNECT DGST .

**Data Download Date/Time** 05/28/2021 17:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

**Agency** PALM BAY P.D. **Date** 06/28/2021 **Type** Agency Inspection  
**Inspector** LUTZ,EDWIN,J **Time** 15:14:20 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 24620080A2  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 11/05/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/28/2021 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

**Agency** PALM BAY P.D. **Date** 07/29/2021 **Type** Agency Inspection  
**Inspector** LUTZ,EDWIN,J **Time** 13:55:26 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 24620080A2  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 11/05/2022

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 07/29/2021 14:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

|                               |                              |                               |                                  |
|-------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> PALM BAY P.D.   | <b>Date</b> 08/26/2021       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LUTZ,EDWIN,J | <b>Time</b> 12:25:38         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201910D     | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> 24620080A2 |
| <b>Expiration</b> 10/22/2021  | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 11/05/2022     |
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/26/2021 12:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

|                               |                              |                               |                                  |
|-------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> PALM BAY P.D.   | <b>Date</b> 09/28/2021       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LUTZ,EDWIN,J | <b>Time</b> 17:03:55         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201910D     | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> 24620080A2 |
| <b>Expiration</b> 10/22/2021  | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 11/05/2022     |
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/04/2021 18:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

|                               |                              |                               |                                  |
|-------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> PALM BAY P.D.   | <b>Date</b> 10/27/2021       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> LUTZ,EDWIN,J | <b>Time</b> 16:36:25         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202101C     | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> 24620080A2 |
| <b>Expiration</b> 01/12/2023  | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 11/05/2022     |
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/27/2021 17:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

|                               |                        |                                   |
|-------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PALM BAY P.D.   | <b>Date</b> 11/04/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL, | <b>Time</b> 12:33:36   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.048  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 11/04/2021 12:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

**Agency** PALM BAY P.D.  
**Inspector** LUTZ, EDWIN, J

**Date** 11/29/2021  
**Time** 14:36:30

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 24620080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2022

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 11/30/2021 14:04

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency PALM BAY P.D. Date 12/27/2021 Type Agency Inspection  
Inspector LUTZ, EDWIN, J Time 17:08:20 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202007A 0.20 Lot # 202106B 0.08 Gas Lot # 24620080A2  
Expiration 01/12/2023 Expiration 07/21/2022 Expiration 06/22/2023 Expiration 11/05/2022

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 12/27/2021 17:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

**Agency** SEMINOLE COUNTY S.O. **Date** 01/29/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 01:01:10 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 201902E **0.08 Gas Lot #** 915310  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 02/21/2021 **Expiration** 03/07/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.205  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.206  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.206  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/30/2021 00:09

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

**Agency** SEMINOLE COUNTY S.O. **Date** 01/29/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 01:56:45 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 201902E **0.08 Gas Lot #** 915310  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 02/20/2021 **Expiration** 03/07/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.205  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.206  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.206  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/30/2021 00:09

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

**Agency** SEMINOLE COUNTY S.O. **Date** 02/26/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 23:16:39 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 915310  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 03/07/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.204  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 03/02/2021 20:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

**Agency** SEMINOLE COUNTY S.O. **Date** 03/30/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 02:20:46 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/30/2021 23:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

**Agency** SEMINOLE COUNTY S.O. **Date** 04/29/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 23:20:50 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.204  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 04/30/2021 23:26

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

**Agency** SEMINOLE COUNTY S.O. **Date** 05/29/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 01:18:34 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 05/31/2021 00:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

**Agency** SEMINOLE COUNTY S.O. **Date** 06/12/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 00:35:58 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/13/2021 22:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

**Agency** SEMINOLE COUNTY S.O. **Date** 07/31/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 01:33:20 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 08/02/2021 21:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

**Agency** SEMINOLE COUNTY S.O. **Date** 08/29/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 00:06:43 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 2012007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 09/13/2021 10:48

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> SEMINOLE COUNTY S.O. | <b>Date</b> 09/13/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 13:19:51   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG026705 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 09/23/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0001      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 09/13/2021 13:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency SEMINOLE COUNTY S.O. Date 10/29/2021 Type Agency Inspection  
Inspector BETHAM,KEITH,D Time 23:47:17 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # 1433248  
Expiration 01/12/2023 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 01/06/2024

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.050  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.079  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.049  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | INT    |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

05: Interferent Detect/CLEAR RERUN.

Data Download Date/Time 11/02/2021 23:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

**Agency** SEMINOLE COUNTY S.O. **Date** 11/27/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 23:15:30 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 11/28/2021 22:03

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

**Agency** SEMINOLE COUNTY S.O. **Date** 12/09/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 00:25:23 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/09/2021 02:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

**Agency** SEMINOLE COUNTY S.O. **Date** 03/01/2021 **Type** Agency Inspection  
**Inspector** SOTO,ISRAEL, **Time** 12:54:13 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 03/01/2021 15:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> SEMINOLE COUNTY S.O. | <b>Date</b> 03/01/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,      | <b>Time</b> 15:21:31   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.077  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.077  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0006 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0008 |

#### Remarks:

Data Download Date/Time 03/01/2021 15:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

**Agency** SEMINOLE COUNTY S.O. **Date** 03/30/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 00:44:19 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/30/2021 23:34

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

**Agency** SEMINOLE COUNTY S.O. **Date** 04/30/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 00:19:15 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/04/2021 02:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

**Agency** SEMINOLE COUNTY S.O. **Date** 05/29/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 00:07:42 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/31/2021 00:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

**Agency** SEMINOLE COUNTY S.O. **Date** 06/11/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 23:33:05 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/14/2021 21:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

**Agency** SEMINOLE COUNTY S.O. **Date** 07/31/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 00:40:47 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 08/02/2021 21:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

**Agency** SEMINOLE COUNTY S.O. **Date** 08/29/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 01:14:31 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 09/01/2021 23:12

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> SEMINOLE COUNTY S.O. | <b>Date</b> 09/30/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> BETHAM,KEITH,D    | <b>Time</b> 19:57:42   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                               |
|------------------------------|------------------------------|------------------------------|-------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 1433248 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 01/06/2024  |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.069  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.075  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.193  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.070  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.075  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.070  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.075  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

08: BAD SEAL ON SIMULATOR CHANGE OUT SIMULATOR .

**Data Download Date/Time** 10/03/2021 22:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

**Agency** SEMINOLE COUNTY S.O. **Date** 10/30/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 01:09:17 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 10/31/2021 04:48

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

**Agency** SEMINOLE COUNTY S.O. **Date** 11/28/2021 **Type** Agency Inspection  
**Inspector** BETHAM,KEITH,D **Time** 00:15:41 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 11/28/2021 22:21

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

Agency SEMINOLE COUNTY S.O. Date 12/08/2021 Type Agency Inspection  
Inspector BETHAM,KEITH,D Time 23:11:25 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # 1433248  
Expiration 01/12/2023 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 01/06/2024

Number of Simulators 5 Compliance No

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Non-compliance: YES IN COMPLIANCE.

Data Download Date/Time 12/09/2021 02:35

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 01/21/2021  
**Time** 01:40:04

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG916501  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 06/14/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/23/2021 08:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 02/04/2021  
**Time** 21:27:26

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG916501  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 06/14/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/05/2021 19:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PINELLAS COUNTY SO   | <b>Date</b> 02/25/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW,TAYLOR,D | <b>Time</b> 13:54:26   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG003005 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 01/30/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 02/25/2021 14:00

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 03/19/2021  
**Time** 21:41:43

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG916501  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 06/14/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/19/2021 21:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 04/14/2021  
**Time** 20:04:39

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG916501  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 06/14/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/14/2021 20:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

|                                  |                              |                               |                                |
|----------------------------------|------------------------------|-------------------------------|--------------------------------|
| <b>Agency</b> PINELLAS COUNTY SO | <b>Date</b> 05/12/2021       | <b>Type</b> Agency Inspection |                                |
| <b>Inspector</b> DEANE,M,        | <b>Time</b> 20:50:23         | <b>Software</b> 8100.27       |                                |
| <b>0.05 Lot #</b> 201910D        | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> AG012501 |
| <b>Expiration</b> 10/22/2021     | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 05/04/2022   |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/13/2021 19:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

|                                  |                              |                               |                                |
|----------------------------------|------------------------------|-------------------------------|--------------------------------|
| <b>Agency</b> PINELLAS COUNTY SO | <b>Date</b> 06/16/2021       | <b>Type</b> Agency Inspection |                                |
| <b>Inspector</b> DEANE,M,        | <b>Time</b> 19:50:13         | <b>Software</b> 8100.27       |                                |
| <b>0.05 Lot #</b> 201910D        | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> AG012501 |
| <b>Expiration</b> 10/22/2021     | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 05/04/2022   |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/16/2021 21:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 07/08/2021  
**Time** 20:22:57

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG012501  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 05/04/2022

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 07/09/2021 03:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 08/19/2021  
**Time** 20:19:35

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG012501  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 05/04/2022

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 08/19/2021 20:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

|                                  |                              |                               |                                |
|----------------------------------|------------------------------|-------------------------------|--------------------------------|
| <b>Agency</b> PINELLAS COUNTY SO | <b>Date</b> 09/23/2021       | <b>Type</b> Agency Inspection |                                |
| <b>Inspector</b> DEANE,M,        | <b>Time</b> 22:29:19         | <b>Software</b> 8100.27       |                                |
| <b>0.05 Lot #</b> 202101C        | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> AG012501 |
| <b>Expiration</b> 01/12/2023     | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 05/04/2022   |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/23/2021 23:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 10/28/2021  
**Time** 19:23:47

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202105B      **0.08 Gas Lot #** AG106803  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/22/2023      **Expiration** 03/09/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 10/29/2021 19:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 11/24/2021  
**Time** 19:30:28

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202105B      **0.08 Gas Lot #** AG106803  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/22/2023      **Expiration** 03/09/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 11/24/2021 20:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 12/09/2021  
**Time** 20:57:39

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202105B      **0.08 Gas Lot #** AG106803  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/22/2023      **Expiration** 03/09/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 12/13/2021 10:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 01/21/2021  
**Time** 01:07:06

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG916501  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 06/14/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/23/2021 08:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 02/04/2021  
**Time** 20:04:10

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG916501  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 06/14/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/05/2021 19:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

|                                      |                        |                                   |
|--------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PINELLAS COUNTY SO     | <b>Date</b> 02/25/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW, TAYLOR, D | <b>Time</b> 13:41:58   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG003005 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 01/30/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 02/25/2021 14:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 03/19/2021  
**Time** 22:19:10

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG012501  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 05/04/2022

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/22/2021 10:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

|                                  |                              |                               |                                |
|----------------------------------|------------------------------|-------------------------------|--------------------------------|
| <b>Agency</b> PINELLAS COUNTY SO | <b>Date</b> 04/14/2021       | <b>Type</b> Agency Inspection |                                |
| <b>Inspector</b> DEANE,M,        | <b>Time</b> 21:44:53         | <b>Software</b> 8100.27       |                                |
| <b>0.05 Lot #</b> 201910D        | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> AG012501 |
| <b>Expiration</b> 10/22/2021     | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 05/04/2022   |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/14/2021 21:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 05/12/2021  
**Time** 23:15:51

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG012501  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 05/04/2022

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 05/13/2021 19:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

|                                  |                              |                               |                                |
|----------------------------------|------------------------------|-------------------------------|--------------------------------|
| <b>Agency</b> PINELLAS COUNTY SO | <b>Date</b> 06/16/2021       | <b>Type</b> Agency Inspection |                                |
| <b>Inspector</b> DEANE,M,        | <b>Time</b> 21:23:07         | <b>Software</b> 8100.27       |                                |
| <b>0.05 Lot #</b> 201910D        | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> AG012501 |
| <b>Expiration</b> 10/22/2021     | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 05/04/2022   |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/16/2021 22:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 07/08/2021  
**Time** 21:41:48

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG012501  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 05/04/2022

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 07/09/2021 04:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 08/19/2021  
**Time** 21:30:19

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG012501  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 05/04/2022

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/19/2021 21:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 09/23/2021  
**Time** 23:12:33

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG021605  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 08/03/2022

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 09/24/2021 04:22

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

|                                  |                        |                               |  |
|----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> PINELLAS COUNTY SO | <b>Date</b> 10/28/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> DEANE,M,        | <b>Time</b> 20:42:30   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202101C    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202105B    | <b>0.08 Gas Lot #</b> AG021605 |
| <b>Expiration</b> 01/12/2023 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/22/2023 | <b>Expiration</b> 08/03/2022   |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                 | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                 | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                 | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1     | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # (Repeat) 1 | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2                | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2                 | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2                 | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                 | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2     | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # (Repeat) 2 | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3                | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                 | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                 | 0.079  |
|                                             |        | 0.20 g/210L Test # 3                 | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3     | 0.079  |
|                                             |        | Interferent Detect Test # (Repeat) 3 | INT    |

**Remarks:**

Int Det: Ambient Fail UNKNOWN PROBLEM REPEAT OK.

**Data Download Date/Time** 10/30/2021 02:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

**Agency** PINELLAS COUNTY SO  
**Inspector** DEANE,M,

**Date** 11/24/2021  
**Time** 21:12:42

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202105B      **0.08 Gas Lot #** AG021605  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/22/2023      **Expiration** 08/03/2022

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 11/26/2021 20:41

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

|                                  |                              |                               |                                |
|----------------------------------|------------------------------|-------------------------------|--------------------------------|
| <b>Agency</b> PINELLAS COUNTY SO | <b>Date</b> 12/09/2021       | <b>Type</b> Agency Inspection |                                |
| <b>Inspector</b> DEANE,M,        | <b>Time</b> 22:23:06         | <b>Software</b> 8100.27       |                                |
| <b>0.05 Lot #</b> 202101C        | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202105B     | <b>0.08 Gas Lot #</b> AG021605 |
| <b>Expiration</b> 01/12/2023     | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/22/2023  | <b>Expiration</b> 08/03/2022   |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/13/2021 10:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

**Agency** HIALEAH GARDENS PD      **Date** 03/18/2021      **Type** Agency Inspection  
**Inspector** REYES-RIVERA,DAVID,E      **Time** 08:46:03      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators**      **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

COMPLIANCE NOT DETERMINED, AI NOT CONDUCTED

**Data Download Date/Time** 03/18/2021 12:05

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> HIALEAH GARDENS PD        | <b>Date</b> 03/18/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 12:00:24   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG026705 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 09/23/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 03/18/2021 12:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

**Agency** HIALEAH GARDENS PD      **Date** 04/30/2021      **Type** Agency Inspection  
**Inspector** PASTOR,JOSE,F      **Time** 11:40:32      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/07/2021 13:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

**Agency** HIALEAH GARDENS PD      **Date** 05/30/2021      **Type** Agency Inspection  
**Inspector** HUERTA, EDWIN,      **Time** 00:46:46      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/30/2021 19:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

**Agency** HIALEAH GARDENS PD      **Date** 06/30/2021      **Type** Agency Inspection  
**Inspector** HUERTA, EDWIN,      **Time** 04:50:22      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/30/2021 16:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

**Agency** HIALEAH GARDENS PD      **Date** 07/30/2021      **Type** Agency Inspection  
**Inspector** HUERTA, EDWIN,      **Time** 00:57:16      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 07/30/2021 01:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

**Agency** HIALEAH GARDENS PD      **Date** 08/30/2021      **Type** Agency Inspection  
**Inspector** HUERTA, EDWIN,      **Time** 21:09:02      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 08/30/2021 21:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

**Agency** HIALEAH GARDENS PD      **Date** 09/30/2021      **Type** Agency Inspection  
**Inspector** HUERTA, EDWIN,      **Time** 22:31:19      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 09/30/2021 22:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

**Agency** HIALEAH GARDENS PD      **Date** 10/30/2021      **Type** Agency Inspection  
**Inspector** HUERTA, EDWIN,      **Time** 01:11:13      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 10/30/2021 01:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

**Agency** HIALEAH GARDENS PD      **Date** 11/30/2021      **Type** Agency Inspection  
**Inspector** HUERTA, EDWIN,      **Time** 21:19:03      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 11/30/2021 21:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

**Agency** HIALEAH GARDENS PD      **Date** 12/30/2021      **Type** Agency Inspection  
**Inspector** HUERTA, EDWIN,      **Time** 03:14:09      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/30/2021 03:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

**Agency** MIAMI-DADE PD      **Date** 01/23/2021      **Type** Agency Inspection  
**Inspector** MYRTIL,REGINALD,J      **Time** 02:40:44      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** 05620080A1  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 02/20/2021      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/27/2021 14:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

**Agency** MIAMI-DADE PD      **Date** 02/20/2021      **Type** Agency Inspection  
**Inspector** MYRTIL,REGINALD,J      **Time** 01:38:49      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** 05620080A1  
**Expiration** 10/22/2021      **Expiration** 10/22/2021      **Expiration** 02/20/2021      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/20/2021 01:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

**Agency** MIAMI-DADE PD      **Date** 03/27/2021      **Type** Agency Inspection  
**Inspector** MYRTIL,REGINALD,J      **Time** 03:20:36      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/27/2021 04:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

**Agency** MIAMI-DADE PD      **Date** 04/24/2021      **Type** Agency Inspection  
**Inspector** CLOSIUS,RICHARD,L      **Time** 01:08:36      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI MYTRIL PRESENT

**Data Download Date/Time** 04/26/2021 16:47

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

**Agency** MIAMI-DADE PD      **Date** 05/18/2021      **Type** Agency Inspection  
**Inspector** CLOSIUS,RICHARD,L      **Time** 05:13:41      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R MYRTIL PRESENT

**Data Download Date/Time** 05/18/2021 06:19

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

|                                    |                              |                               |                                  |
|------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> MIAMI-DADE PD        | <b>Date</b> 06/03/2021       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> CLOSIUS,RICHARD,L | <b>Time</b> 04:18:16         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201910D          | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> 05620080A1 |
| <b>Expiration</b> 10/22/2021       | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 04/05/2022     |
| <b>Number of Simulators</b> 5      | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R MYRTIL PRESENT

**Data Download Date/Time** 06/03/2021 10:39

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

|                                    |                              |                               |                                  |
|------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> MIAMI-DADE PD        | <b>Date</b> 07/14/2021       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> CLOSIUS,RICHARD,L | <b>Time</b> 00:30:52         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201910D          | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> 05620080A1 |
| <b>Expiration</b> 10/22/2021       | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 04/05/2022     |
| <b>Number of Simulators</b> 5      | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/14/2021 10:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

|                                    |                              |                               |                                  |
|------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> MIAMI-DADE PD        | <b>Date</b> 08/05/2021       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> CLOSIUS,RICHARD,L | <b>Time</b> 04:59:58         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202101C          | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> 05620080A1 |
| <b>Expiration</b> 01/12/2023       | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 04/05/2022     |
| <b>Number of Simulators</b> 5      | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R MYRTIL PRESENT

**Data Download Date/Time** 08/06/2021 10:16

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

**Agency** MIAMI-DADE PD      **Date** 09/07/2021      **Type** Agency Inspection  
**Inspector** CLOSIUS,RICHARD,L      **Time** 03:47:38      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R MYRTIL PRESENT

**Data Download Date/Time** 09/07/2021 07:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

**Agency** MIAMI-DADE PD      **Date** 10/07/2021      **Type** Agency Inspection  
**Inspector** CLOSIUS,RICHARD,L      **Time** 06:23:59      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R MYRTIL PRESENT

**Data Download Date/Time** 10/07/2021 07:06

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MIAMI-DADE PD             | <b>Date</b> 11/01/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 09:20:10   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG026705 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 09/23/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.077  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.076  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.076  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.076  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.076  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.076  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.076  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0007 |

#### Remarks:

Data Download Date/Time 11/01/2021 09:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

|                                    |                              |                               |                                  |
|------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> MIAMI-DADE PD        | <b>Date</b> 11/11/2021       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> CLOSIUS,RICHARD,L | <b>Time</b> 03:42:40         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202101C          | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> 05620080A1 |
| <b>Expiration</b> 01/12/2023       | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 04/05/2022     |
| <b>Number of Simulators</b> 4      | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R MYRTIL PRESENT

**Data Download Date/Time** 11/11/2021 12:22

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

**Agency** MIAMI-DADE PD      **Date** 12/02/2021      **Type** Agency Inspection  
**Inspector** CLOSIUS,RICHARD,L      **Time** 03:44:23      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R MYRTIL PRESENT

**Data Download Date/Time** 12/02/2021 07:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

**Agency** MIAMI-DADE PD      **Date** 01/23/2021      **Type** Agency Inspection  
**Inspector** MYRTIL,REGINALD,J      **Time** 02:31:22      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** 05620080A1  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 02/20/2021      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/26/2021 12:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

**Agency** MIAMI-DADE PD      **Date** 02/20/2021      **Type** Agency Inspection  
**Inspector** MYRTIL,REGINALD,J      **Time** 01:44:31      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** 05620080A1  
**Expiration** 10/22/2021      **Expiration** 10/22/2021      **Expiration** 02/20/2021      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/20/2021 02:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

**Agency** MIAMI-DADE PD      **Date** 03/27/2021      **Type** Agency Inspection  
**Inspector** MYRTIL,REGINALD,J      **Time** 03:20:04      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/27/2021 03:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

**Agency** MIAMI-DADE PD      **Date** 04/24/2021      **Type** Agency Inspection  
**Inspector** CLOSIUS,RICHARD,      **Time** 01:28:21      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI MYRTIL PRESENT

**Data Download Date/Time** 04/26/2021 17:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

**Agency** MIAMI-DADE PD      **Date** 05/20/2021      **Type** Agency Inspection  
**Inspector** CLOSIUS,RICHARD,L      **Time** 03:53:28      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI MYRTIL PRESENT

**Data Download Date/Time** 05/20/2021 09:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

**Agency** MIAMI-DADE PD      **Date** 06/08/2021      **Type** Agency Inspection  
**Inspector** CLOSIUS,RICHARD,L      **Time** 05:10:58      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R MYRTIL PRESENT

**Data Download Date/Time** 06/08/2021 05:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

|                                       |                        |                                   |
|---------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MIAMI-DADE PD           | <b>Date</b> 06/24/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA,DAVID,E | <b>Time</b> 14:32:05   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG026705 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 09/23/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.051  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.052  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 06/24/2021 14:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

**Agency** MIAMI-DADE PD      **Date** 07/01/2021      **Type** Agency Inspection  
**Inspector** CLOSIUS,RICHARD,L      **Time** 03:03:08      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R MYRTIL PRESENT

**Data Download Date/Time** 07/02/2021 04:13

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

|                                    |                              |                               |                                  |
|------------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> MIAMI-DADE PD        | <b>Date</b> 08/19/2021       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> CLOSIUS,RICHARD,L | <b>Time</b> 03:22:12         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 202101C          | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> 05620080A1 |
| <b>Expiration</b> 01/12/2023       | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 04/05/2022     |
| <b>Number of Simulators</b> 5      | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R MYRTIL PRESENT

**Data Download Date/Time** 08/19/2021 03:53

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

**Agency** MIAMI-DADE PD      **Date** 09/16/2021      **Type** Agency Inspection  
**Inspector** CLOSIUS,RICHARD,L      **Time** 04:02:13      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/16/2021 06:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

**Agency** MIAMI-DADE PD      **Date** 10/14/2021      **Type** Agency Inspection  
**Inspector** CLOSIUS,RICHARD,L      **Time** 03:32:03      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R MYRTIL PRESENT

**Data Download Date/Time** 10/14/2021 03:51

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

**Agency** MIAMI-DADE PD      **Date** 11/11/2021      **Type** Agency Inspection  
**Inspector** CLOSIUS,RICHARD,L      **Time** 03:45:02      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R MYRTIL PRESENT

**Data Download Date/Time** 11/11/2021 12:19

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

**Agency** MIAMI-DADE PD      **Date** 12/09/2021      **Type** Agency Inspection  
**Inspector** CLOSIUS,RICHARD,L      **Time** 03:43:25      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 05620080A1  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 05/12/2022      **Expiration** 04/05/2022

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

AI R MYRTIL PRESENT

**Data Download Date/Time** 12/09/2021 03:57

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

**Agency** JACKSONVILLE SO      **Date** 01/13/2021      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 11:03:30      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** 09820080A2  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 02/20/2021      **Expiration** 06/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/13/2021 11:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

**Agency** JACKSONVILLE SO      **Date** 01/28/2021      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 08:19:13      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** 09820080A2  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 02/20/2021      **Expiration** 06/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/28/2021 08:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

**Agency** JACKSONVILLE SO      **Date** 02/24/2021      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 05:42:34      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 09820080A2  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 06/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/24/2021 15:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

**Agency** JACKSONVILLE SO      **Date** 03/10/2021      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 10:49:55      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 09820080A2  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 06/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/12/2021 09:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

**Agency** JACKSONVILLE SO      **Date** 04/15/2021      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 12:02:19      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 22620080A2  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 10/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

REMOVED FROM EVED. USE AT PTDF 4-15-2021

**Data Download Date/Time** 04/15/2021 12:24

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> JACKSONVILLE SO      | <b>Date</b> 05/12/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> LIGHTSEY,ROBERT,D | <b>Time</b> 12:03:12   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 22620080A2 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 10/05/2022     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

PLACED INTO EVED USE AT PTDF ON 5-12-2021

**Data Download Date/Time** 05/12/2021 12:12

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

**Agency** JACKSONVILLE SO      **Date** 05/27/2021      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 09:33:39      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 22620080A2  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 10/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/27/2021 09:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | JACKSONVILLE SO   | <b>Date</b>       | 06/16/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LIGHTSEY,ROBERT,D | <b>Time</b>       | 10:18:19   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201910D           | <b>0.08 Lot #</b> | 201908B    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 22620080A2 |
| <b>Expiration</b>           | 10/22/2021        | <b>Expiration</b> | 08/07/2021 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 10/05/2022 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 06/17/2021 05:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

**Agency** JACKSONVILLE SO      **Date** 06/30/2021      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 12:38:02      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 22620080A2  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 10/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 07/01/2021 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | JACKSONVILLE SO   | <b>Date</b>       | 07/15/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LIGHTSEY,ROBERT,D | <b>Time</b>       | 09:21:38   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201910D           | <b>0.08 Lot #</b> | 201908B    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 22620080A2 |
| <b>Expiration</b>           | 10/22/2021        | <b>Expiration</b> | 08/07/2021 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 10/05/2022 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 07/15/2021 12:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

**Agency** JACKSONVILLE SO      **Date** 07/29/2021      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 12:30:51      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 22620080A2  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 10/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.206  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/29/2021 12:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

**Agency** JACKSONVILLE SO      **Date** 08/13/2021      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 06:51:01      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 22620080A2  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 10/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/13/2021 07:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | JACKSONVILLE SO   | <b>Date</b>       | 08/26/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LIGHTSEY,ROBERT,D | <b>Time</b>       | 08:38:05   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201910D           | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 22620080A2 |
| <b>Expiration</b>           | 10/22/2021        | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 10/05/2022 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/26/2021 11:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

**Agency** JACKSONVILLE SO      **Date** 09/17/2021      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 11:10:02      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 22620080A2  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 10/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

REMOVED FROM EVED USE FOR DEPT INSPEC AND P/M 9-17-2021

**Data Download Date/Time** 09/17/2021 11:56

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

|                                     |                        |                                   |
|-------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> JACKSONVILLE SO       | <b>Date</b> 11/29/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> WILLIAMS,RICHARD,A | <b>Time</b> 13:36:33   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.077  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.077  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.077  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.077  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.077  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.077  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 11/29/2021 14:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

**Agency** JACKSONVILLE SO      **Date** 12/14/2021      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 11:35:49      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 22620080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 10/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

PLACED INTO EVED. USE AT PTDF ON 12-14-2021

**Data Download Date/Time** 12/14/2021 11:38

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

**Agency** JACKSONVILLE SO      **Date** 12/30/2021      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 09:07:21      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 22620080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 10/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 12/30/2021 09:10

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

|                                      |                        |                               |  |
|--------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> JACKSONVILLE SO        | <b>Date</b> 01/13/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> LIGHTSEY, ROBERT, D | <b>Time</b> 14:42:55   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 201902C    | <b>0.08 Gas Lot #</b> 22419080A2 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 02/20/2021 | <b>Expiration</b> 10/05/2021     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

REMOVED FROM EVED USE AT PTDF ON 1-13-2021

**Data Download Date/Time** 01/14/2021 11:41

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

**Agency** JACKSONVILLE SO      **Date** 02/24/2021      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 01:16:09      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 22419080A2  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 10/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

PLACED INTO EVED. USE AT PTDF ON 2-24-2021

**Data Download Date/Time** 02/24/2021 15:17

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

**Agency** JACKSONVILLE SO      **Date** 03/10/2021      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 09:07:42      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 22419080A2  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 10/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/12/2021 09:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

|                                      |                        |                               |  |
|--------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> JACKSONVILLE SO        | <b>Date</b> 04/15/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> LIGHTSEY, ROBERT, D | <b>Time</b> 12:08:58   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 22620080A2 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 10/05/2022     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

REMOVED FROM EVED. USE AT PTDF 4-15-2021

**Data Download Date/Time** 04/15/2021 12:55

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

|                                      |                        |                               |  |
|--------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> JACKSONVILLE SO        | <b>Date</b> 05/12/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> LIGHTSEY, ROBERT, D | <b>Time</b> 11:59:27   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 22620080A2 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 10/05/2022     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

PLACED INTO EVED USE AT PTDF ON 5-12-2021

**Data Download Date/Time** 05/12/2021 12:10

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

**Agency** JACKSONVILLE SO      **Date** 05/27/2021      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 09:42:39      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 22620080A2  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 10/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/28/2021 15:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

**Agency** JACKSONVILLE SO      **Date** 06/16/2021      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 10:19:29      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 22620080A2  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 10/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/17/2021 07:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | JACKSONVILLE SO   | <b>Date</b>       | 06/30/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LIGHTSEY,ROBERT,D | <b>Time</b>       | 12:46:13   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201910D           | <b>0.08 Lot #</b> | 201908B    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 22620080A2 |
| <b>Expiration</b>           | 10/22/2021        | <b>Expiration</b> | 08/07/2021 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 10/05/2022 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Data Download Date/Time 07/01/2021 10:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

|                                      |                        |                               |  |
|--------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> JACKSONVILLE SO        | <b>Date</b> 07/15/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> LIGHTSEY, ROBERT, D | <b>Time</b> 09:31:28   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 22620080A2 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 10/05/2022     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.082  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             |        |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

08: RFI Detect.RADIO TRAFF FROM SALLY PORT CLEARED RERAN 08

**Data Download Date/Time** 07/16/2021 10:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

|                                      |                        |                               |  |
|--------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> JACKSONVILLE SO        | <b>Date</b> 07/29/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> LIGHTSEY, ROBERT, D | <b>Time</b> 12:49:34   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 22620080A2 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 10/05/2022     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

REMOVED FROM EVED USE AT PTDF 7-29-2021

**Data Download Date/Time** 07/29/2021 13:22

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | JACKSONVILLE SO   | <b>Date</b>       | 08/12/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LIGHTSEY,ROBERT,D | <b>Time</b>       | 10:07:32   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201910D           | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 22620080A2 |
| <b>Expiration</b>           | 10/22/2021        | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 10/05/2022 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.083  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

PLACED INTO EVED USE AT MAYPORT NAS 8-12-2021

**Data Download Date/Time** 08/12/2021 10:53

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

|                                      |                        |                               |  |
|--------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> JACKSONVILLE SO        | <b>Date</b> 09/30/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> LIGHTSEY, ROBERT, D | <b>Time</b> 10:30:34   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 22620080A2 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 10/05/2022     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

R/D FROM NAS M/P EVED AND PLACED IN PTDF EVED ON 9-30-2021

**Data Download Date/Time** 09/30/2021 10:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

|                             |                   |                   |            |                   |                   |                       |            |
|-----------------------------|-------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | JACKSONVILLE SO   | <b>Date</b>       | 10/14/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LIGHTSEY,ROBERT,D | <b>Time</b>       | 10:23:01   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201910D           | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 22620080A2 |
| <b>Expiration</b>           | 10/22/2021        | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 10/05/2022 |
| <b>Number of Simulators</b> | 5                 | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/26/2021 10:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

**Agency** JACKSONVILLE SO      **Date** 11/03/2021      **Type** Agency Inspection  
**Inspector** LIGHTSEY,ROBERT,D      **Time** 11:24:12      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 22620080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 10/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/03/2021 11:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

|                                      |                        |                               |  |
|--------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> JACKSONVILLE SO        | <b>Date</b> 11/23/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> LIGHTSEY, ROBERT, D | <b>Time</b> 10:36:04   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202101C    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 22620080A2 |
| <b>Expiration</b> 01/12/2023 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 10/05/2022     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

REMOVED FROM EVED USE AT PTDF FOR P/M AND D/I ON 11-23-2021

**Data Download Date/Time** 11/23/2021 10:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

|                                     |                        |                                   |
|-------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> JACKSONVILLE SO       | <b>Date</b> 12/03/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> WILLIAMS,RICHARD,A | <b>Time</b> 16:06:03   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.051  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 12/03/2021 16:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

|                               |                        |                                   |
|-------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> JACKSONVILLE SO | <b>Date</b> 01/22/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL, | <b>Time</b> 16:02:58   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0001      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 01/22/2021 16:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

|                               |                        |                                   |
|-------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> JACKSONVILLE SO | <b>Date</b> 04/12/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL, | <b>Time</b> 10:05:19   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.048  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 04/12/2021 10:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

**Agency** NASSAU COUNTY SO **Date** 01/28/2021 **Type** Agency Inspection  
**Inspector** HIGGINBOTHAM,BRIAN,K **Time** 21:47:07 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** 10/22/2021 **Expiration** **Expiration** **Expiration**

**Number of Simulators** 4 **Compliance** No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.042  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1 | 0.042  |
| Barometric Pressure Sensor Check: OK        |        | Interferent Detect Test # 1   | INT    |
| Alcohol Free Subject Test: 0.000            | Yes    | Alcohol Free Test # 2         | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.05 g/210L Test # 2          | 0.042  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # (Repeat) 2 | 0.043  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 2   | INT    |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # 3         | 0.000  |
| 0.05 g/210L Test: OK                        | No     | 0.05 g/210L Test # 3          | 0.042  |
| 0.08 g/210L Test: OK                        | No     | 0.05 g/210L Test # (Repeat) 3 | 0.042  |
| 0.20 g/210L Test: OK                        | No     | Interferent Detect Test # 3   | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                               |        |

#### Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:.05 TOO LOW.

**Data Download Date/Time** 02/08/2021 14:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

|                                |                        |                                   |
|--------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> NASSAU COUNTY SO | <b>Date</b> 02/10/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,  | <b>Time</b> 13:24:39   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.075  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.076  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.077  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.076  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.077  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.077  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.076  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.077  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.077  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.077  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.048  |
| 0.08 g/210L Test # 10                  | 0.077  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0007 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 02/10/2021 14:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

|                                         |                        |                               |  |
|-----------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> NASSAU COUNTY SO          | <b>Date</b> 02/25/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> HIGGINBOTHAM, BRIAN, K | <b>Time</b> 22:24:06   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 14020080A1 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 07/05/2022     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 4 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.044  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.046  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.077  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.195  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.044  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.046  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.045  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.045  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

05: Control Outside Tolerance.

**Data Download Date/Time** 03/29/2021 14:24

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

**Agency** NASSAU COUNTY SO      **Date** 03/29/2021      **Type** Agency Inspection  
**Inspector** HIGGINBOTHAM,BRIAN,K      **Time** 16:37:53      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 14020080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2022

**Number of Simulators** 2      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.045  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.045  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/29/2021 17:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

**Agency** NASSAU COUNTY SO      **Date** 04/29/2021      **Type** Agency Inspection  
**Inspector** HIGGINBOTHAM,BRIAN,K      **Time** 13:57:17      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 14020080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2022

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/01/2021 17:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

|                             |                      |                   |            |                   |                   |                       |            |
|-----------------------------|----------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | NASSAU COUNTY SO     | <b>Date</b>       | 05/26/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | HIGGINBOTHAM,BRIAN,K | <b>Time</b>       | 00:42:03   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201910D              | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 14020080A1 |
| <b>Expiration</b>           | 10/22/2021           | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 07/05/2022 |
| <b>Number of Simulators</b> | 4                    | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/28/2021 18:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

**Agency** NASSAU COUNTY SO      **Date** 06/28/2021      **Type** Agency Inspection  
**Inspector** HIGGINBOTHAM,BRIAN,K      **Time** 05:49:23      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 14020080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2022

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/30/2021 23:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

|                                         |                        |                               |  |
|-----------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> NASSAU COUNTY SO          | <b>Date</b> 07/25/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> HIGGINBOTHAM, BRIAN, K | <b>Time</b> 22:18:58   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202101C    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 14020080A1 |
| <b>Expiration</b> 01/12/2023 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 07/05/2022     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 4 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.061  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.076  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.041  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.041  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

|                                                 |                                                 |
|-------------------------------------------------|-------------------------------------------------|
| 08: GLASS SIMULATOR BROKE DURING TEST TOLERANCE | . SIMULATOR CRACKED DURING TEST CONTROL OUTSIDE |
|-------------------------------------------------|-------------------------------------------------|

**Data Download Date/Time** 07/26/2021 00:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

**Agency** NASSAU COUNTY SO      **Date** 08/22/2021      **Type** Agency Inspection  
**Inspector** HIGGINBOTHAM,BRIAN,K      **Time** 20:02:41      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 14020080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2022

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 08/22/2021 20:19

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

**Agency** NASSAU COUNTY SO      **Date** 09/28/2021      **Type** Agency Inspection  
**Inspector** HIGGINBOTHAM,BRIAN,K      **Time** 00:05:34      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 14020080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2022

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/30/2021 12:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

**Agency** NASSAU COUNTY SO      **Date** 10/27/2021      **Type** Agency Inspection  
**Inspector** HIGGINBOTHAM,BRIAN,K      **Time** 05:37:51      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 14020080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2022

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/30/2021 02:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

**Agency** NASSAU COUNTY SO      **Date** 11/27/2021      **Type** Agency Inspection  
**Inspector** HIGGINBOTHAM,BRIAN,K      **Time** 00:10:04      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 14020080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2022

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1             | 0.045  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Time-Date changed.

**Data Download Date/Time** 12/01/2021 12:49

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

**Agency** NASSAU COUNTY SO      **Date** 12/22/2021      **Type** Agency Inspection  
**Inspector** HIGGINBOTHAM,BRIAN,K      **Time** 03:21:28      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 14020080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2022

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.045  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.045  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/22/2021 03:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

**Agency** UNIVERSITY OF FL PD      **Date** 12/07/2021      **Type** Agency Inspection  
**Inspector** SOTO,ISRAEL,      **Time** 08:06:13      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators**      **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 12/07/2021 11:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

|                                   |                        |                                   |
|-----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> UNIVERSITY OF FL PD | <b>Date</b> 12/07/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,     | <b>Time</b> 11:53:24   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.051  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.077  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.077  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.077  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.051  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.077  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.204  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.077  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 12/07/2021 11:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

|                                   |                        |                               |  |
|-----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> HIGH SPRINGS PD     | <b>Date</b> 02/28/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> YAKUBSIN,JAMES,R | <b>Time</b> 16:34:35   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 22620080A3 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 10/05/2022     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.187  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.194  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.189  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.195  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

20: Control Outside SIMULATOR HOSE AND LID NOT CONN RIGHT .

**Data Download Date/Time** 07/15/2021 13:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

|                  |                 |             |            |                 |                   |
|------------------|-----------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | HIGH SPRINGS PD | <b>Date</b> | 07/15/2021 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | SOTO,ISRAEL,    | <b>Time</b> | 08:48:23   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 07/15/2021 13:14

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

|                                     |                        |                                   |
|-------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> HIGH SPRINGS PD       | <b>Date</b> 07/15/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> WILLIAMS,RICHARD,A | <b>Time</b> 13:08:28   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AGO11102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.081  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.081  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.081  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 07/15/2021 13:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

**Agency** GILCHRIST COUNTY SO      **Date** 01/01/2021      **Type** Agency Inspection  
**Inspector** LANGFORD,BRIAN,A      **Time** 10:30:22      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** N      **0.08 Gas Lot #** 1224982  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 10/31/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | INT    |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.077  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.195  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.078  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

08: Interferent Detect.

**Data Download Date/Time** 01/06/2021 06:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

**Agency** GILCHRIST COUNTY SO      **Date** 02/25/2021      **Type** Agency Inspection  
**Inspector** LANGFORD,BRIAN,A      **Time** 09:57:50      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1224982  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 10/31/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/01/2021 10:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

**Agency** GILCHRIST COUNTY SO      **Date** 03/29/2021      **Type** Agency Inspection  
**Inspector** LANGFORD,BRIAN,A      **Time** 02:24:47      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1224982  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 10/31/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 04/06/2021 01:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

**Agency** GILCHRIST COUNTY SO      **Date** 05/08/2021      **Type** Agency Inspection  
**Inspector** LANGFORD,BRIAN,A      **Time** 08:44:45      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1224982  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 10/31/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/18/2021 08:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

**Agency** GILCHRIST COUNTY SO      **Date** 06/15/2021      **Type** Agency Inspection  
**Inspector** LANGFORD,BRIAN,A      **Time** 17:22:56      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1224982  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 10/31/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/16/2021 10:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

**Agency** GILCHRIST COUNTY SO      **Date** 07/31/2021      **Type** Agency Inspection  
**Inspector** LANGFORD,BRIAN,A      **Time** 08:47:56      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1224982  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 10/31/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 08/06/2021 15:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

**Agency** GILCHRIST COUNTY SO      **Date** 09/21/2021      **Type** Agency Inspection  
**Inspector** LANGFORD,BRIAN,A      **Time** 03:13:43      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202002A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1224982  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 10/31/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 10/04/2021 16:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

**Agency** GILCHRIST COUNTY SO      **Date** 11/02/2021      **Type** Agency Inspection  
**Inspector** LANGFORD,BRIAN,A      **Time** 02:37:37      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1224982  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 10/31/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 11/11/2021 16:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

|                                     |                        |                                   |
|-------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> GILCHRIST COUNTY SO   | <b>Date</b> 12/29/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> WILLIAMS,RICHARD,A | <b>Time</b> 11:46:47   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 12/29/2021 11:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

**Agency** GILCHRIST COUNTY SO      **Date** 01/01/2021      **Type** Agency Inspection  
**Inspector** LANGFORD,BRIAN,N      **Time** 10:33:03      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1048227  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 11/15/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/06/2021 06:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

|                                     |                        |                               |  |
|-------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> GILCHRIST COUNTY SO   | <b>Date</b> 07/31/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> LANGFORD, BRIAN, A | <b>Time</b> 08:42:56   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                               |
|------------------------------|------------------------------|------------------------------|-------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 1048227 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 11/15/2021  |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

A F / M A: RFI Detect.

**Data Download Date/Time** 08/06/2021 15:22

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

**Agency** GILCHRIST COUNTY SO      **Date** 08/30/2021      **Type** Agency Inspection  
**Inspector** LEVY, JAYSON, I      **Time** 08:51:15      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1048227  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/15/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.205  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.205  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.205  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/29/2021 14:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

**Agency** GILCHRIST COUNTY SO      **Date** 09/21/2021      **Type** Agency Inspection  
**Inspector** LANGFORD,BRIAN,A      **Time** 03:12:00      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1048227  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 11/15/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 11/29/2021 14:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

|                  |                     |             |            |                 |                   |
|------------------|---------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | GILCHRIST COUNTY SO | <b>Date</b> | 11/29/2021 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | WILLIAMS,RICHARD,A  | <b>Time</b> | 14:54:08   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT.

**Data Download Date/Time** 11/30/2021 12:44

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

|                                     |                        |                                   |
|-------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> GILCHRIST COUNTY SO   | <b>Date</b> 11/30/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> WILLIAMS,RICHARD,A | <b>Time</b> 12:36:13   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Time-Date changed.

Data Download Date/Time 11/30/2021 12:44

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001286

**Agency** BAKER COUNTY SO **Date** 01/04/2021 **Type** Agency Inspection  
**Inspector** SOTO,ISRAEL, **Time** 14:02:40 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 01/07/2021 12:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001286

|                               |                        |                                   |
|-------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> BAKER COUNTY SO | <b>Date</b> 01/07/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL, | <b>Time</b> 10:53:28   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201905A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 05/14/2021 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001286

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0007 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0010 |

#### Remarks:

Data Download Date/Time 01/07/2021 12:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001286

|                  |                              |             |            |                 |                   |
|------------------|------------------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | BAKER COUNTY SO              | <b>Date</b> | 02/18/2021 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | DEREK WATSON,CALEB COLLINS,/ | <b>Time</b> | 16:41:58   | <b>Software</b> | 8100.27           |

|                   |            |                   |  |                   |  |                       |  |
|-------------------|------------|-------------------|--|-------------------|--|-----------------------|--|
| <b>0.05 Lot #</b> | 201910D    | <b>0.08 Lot #</b> |  | <b>0.20 Lot #</b> |  | <b>0.08 Gas Lot #</b> |  |
| <b>Expiration</b> | 10/22/2021 | <b>Expiration</b> |  | <b>Expiration</b> |  | <b>Expiration</b>     |  |

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                        | Result |
|---------------------------------------------|--------|-----------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1       | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1        | 0.044  |
| Minimum Sample Volume Check: OK             |        | Interferent Detect Test # 1 | INT    |
| Barometric Pressure Sensor Check: OK        |        | Alcohol Free Test # 2       | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.05 g/210L Test # 2        | 0.046  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 2 | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 3       | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.05 g/210L Test # 3        | 0.047  |
| Alcohol Free Test: OK                       | Yes    | Interferent Detect Test # 3 | INT    |
| 0.05 g/210L Test: OK                        | No     |                             |        |
| 0.08 g/210L Test: OK                        | No     |                             |        |
| 0.20 g/210L Test: OK                        | No     |                             |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                             |        |

#### Remarks:

05: Control Outside Tolerance.

**Data Download Date/Time** 10/05/2021 15:27

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001286

|                  |                              |             |            |                 |                   |
|------------------|------------------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | BAKER COUNTY SO              | <b>Date</b> | 02/18/2021 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | DEREK WATSON,CALEB COLLINS,/ | <b>Time</b> | 17:07:11   | <b>Software</b> | 8100.27           |

|                   |            |                   |            |                   |  |                       |  |
|-------------------|------------|-------------------|------------|-------------------|--|-----------------------|--|
| <b>0.05 Lot #</b> | 201910D    | <b>0.08 Lot #</b> | 201908B    | <b>0.20 Lot #</b> |  | <b>0.08 Gas Lot #</b> |  |
| <b>Expiration</b> | 10/22/2021 | <b>Expiration</b> | 08/07/2021 | <b>Expiration</b> |  | <b>Expiration</b>     |  |

**Number of Simulators** 5

**Compliance** No

| Check or Test                               | Result | Test                        | Result |
|---------------------------------------------|--------|-----------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1       | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1        | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1        | 0.072  |
| Barometric Pressure Sensor Check: OK        |        | Interferent Detect Test # 1 | INT    |
| Alcohol Free Subject Test: 0.000            | Yes    | Alcohol Free Test # 2       | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.05 g/210L Test # 2        | 0.046  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.08 g/210L Test # 2        | 0.074  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 2 | INT    |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # 3       | 0.000  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # 3        | 0.047  |
| 0.08 g/210L Test: OK                        | No     | 0.08 g/210L Test # 3        | 0.075  |
| 0.20 g/210L Test: OK                        | No     | Interferent Detect Test # 3 | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                             |        |

#### Remarks:

08: Control Outside Tolerance. Non-compliance:CONTROL OUTSIDE TOLERANCE.

**Data Download Date/Time** 10/05/2021 15:27

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

|                  |                              |             |            |                 |                   |
|------------------|------------------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | BAKER COUNTY SO              | <b>Date</b> | 01/10/2021 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | DEREK WATSON,CALEB COLLINS,/ | <b>Time</b> | 04:37:10   | <b>Software</b> | 8100.27           |

|                   |            |                   |            |                   |            |                       |            |
|-------------------|------------|-------------------|------------|-------------------|------------|-----------------------|------------|
| <b>0.05 Lot #</b> | 201910D    | <b>0.08 Lot #</b> | 201908B    | <b>0.20 Lot #</b> | 202005A    | <b>0.08 Gas Lot #</b> | 24119080A1 |
| <b>Expiration</b> | 10/22/2021 | <b>Expiration</b> | 08/07/2019 | <b>Expiration</b> | 05/12/2022 | <b>Expiration</b>     | 11/05/2021 |

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/26/2021 08:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

**Agency** BAKER COUNTY SO **Date** 03/01/2021 **Type** Agency Inspection  
**Inspector** SOTO,ISRAEL, **Time** 10:37:40 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 03/01/2021 12:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

|                               |                        |                                   |
|-------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> BAKER COUNTY SO | <b>Date</b> 03/01/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL, | <b>Time</b> 12:16:48   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 03/01/2021 12:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

|                  |                                  |             |            |                 |                   |
|------------------|----------------------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | BAKER COUNTY SO                  | <b>Date</b> | 06/06/2021 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | CALEB COLLINS,DEREK<br>WATSON, / | <b>Time</b> | 22:55:36   | <b>Software</b> | 8100.27           |

|                   |            |                   |            |                   |            |                       |            |
|-------------------|------------|-------------------|------------|-------------------|------------|-----------------------|------------|
| <b>0.05 Lot #</b> | 201910D    | <b>0.08 Lot #</b> | 201908B    | <b>0.20 Lot #</b> | 202005A    | <b>0.08 Gas Lot #</b> | 06021080A2 |
| <b>Expiration</b> | 10/22/2021 | <b>Expiration</b> | 08/07/2021 | <b>Expiration</b> | 05/12/2022 | <b>Expiration</b>     | 04/05/2023 |

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/08/2021 01:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

**Agency** BAKER COUNTY SO **Date** 07/29/2021 **Type** Agency Inspection  
**Inspector** WATSON,COLLINS,/ **Time** 03:29:52 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: Range Exceeded. Non-compliance: .

**Data Download Date/Time** 10/05/2021 15:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

**Agency** BAKER COUNTY SO  
**Inspector** COLLINS,WATSON, /

**Date** 07/29/2021  
**Time** 04:04:56

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 06021080A2  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 04/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/05/2021 15:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

**Agency** BAKER COUNTY SO **Date** 11/12/2021 **Type** Agency Inspection  
**Inspector** COLLINS, CALEB, A **Time** 23:38:11 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: Range Exceeded. Non-compliance: .

**Data Download Date/Time** 11/13/2021 00:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

**Agency** BAKER COUNTY SO  
**Inspector** COLLINS, CALEB, A

**Date** 11/13/2021  
**Time** 00:20:18

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 06021080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/04/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 11/13/2021 00:43

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

|                  |                              |             |            |                 |                   |
|------------------|------------------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | BAKER COUNTY SO              | <b>Date</b> | 12/13/2021 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | DEREK WATSON,CALEB COLLINS,/ | <b>Time</b> | 01:40:19   | <b>Software</b> | 8100.27           |

|                   |            |                   |            |                   |            |                       |            |
|-------------------|------------|-------------------|------------|-------------------|------------|-----------------------|------------|
| <b>0.05 Lot #</b> | 202101C    | <b>0.08 Lot #</b> | 202108D    | <b>0.20 Lot #</b> | 202005A    | <b>0.08 Gas Lot #</b> | 06021080A2 |
| <b>Expiration</b> | 01/12/2023 | <b>Expiration</b> | 08/17/2023 | <b>Expiration</b> | 05/12/2022 | <b>Expiration</b>     | 04/05/2023 |

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 12/14/2021 08:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> UNION COUNTY SO      | <b>Date</b> 01/29/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> MCDANIEL,MARCUS,G | <b>Time</b> 09:43:04   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                               |
|------------------------------|------------------------------|------------------------------|-------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 201902C    | <b>0.08 Gas Lot #</b> 1190213 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 02/20/2021 | <b>Expiration</b> 08/28/2022  |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 4 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 01/29/2021 09:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

**Agency** UNION COUNTY SO **Date** 02/26/2021 **Type** Agency Inspection  
**Inspector** MCDANIEL,MARCUS,G **Time** 09:28:00 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1190213  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 08/28/2022

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/26/2021 10:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> UNION COUNTY SO      | <b>Date</b> 03/29/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> MCDANIEL,MARCUS,G | <b>Time</b> 10:05:06   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                               |
|------------------------------|------------------------------|------------------------------|-------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 1190213 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 08/28/2022  |

  

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.184  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.195  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.190  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.196  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time** 03/29/2021 10:32

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

**Agency** UNION COUNTY SO **Date** 04/28/2021 **Type** Agency Inspection  
**Inspector** MCDANIEL,MARCUS,G **Time** 10:02:49 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1190213  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 08/28/2022

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/28/2021 10:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> UNION COUNTY SO      | <b>Date</b> 05/29/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> MCDANIEL,MARCUS,G | <b>Time</b> 09:31:04   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                               |
|------------------------------|------------------------------|------------------------------|-------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 1190213 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 08/28/2022  |

  

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.074  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.078  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.075  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

A F / M A: Improper SampleImproper Sample. 08: Control Outside Tolerance.

**Data Download Date/Time** 05/29/2021 09:41

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

|                                     |                        |                                   |
|-------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> UNION COUNTY SO       | <b>Date</b> 06/30/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> WILLIAMS,RICHARD,A | <b>Time</b> 16:27:00   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010B    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001288

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

A F / M A:RECEIVED VALUE, NOMINAL VALUE OBTAINED UPON RETRY.

Data Download Date/Time 06/30/2021 16:32

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

**Agency** UNION COUNTY SO **Date** 01/29/2021 **Type** Agency Inspection  
**Inspector** MCDANIEL,MARCUS,G **Time** 09:47:56 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 201902C **0.08 Gas Lot #** 1190213  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 02/20/2021 **Expiration** 08/28/2022

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/29/2021 09:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

**Agency** UNION COUNTY SO **Date** 02/26/2021 **Type** Agency Inspection  
**Inspector** MCDANIEL,MARCUS,G **Time** 09:32:55 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1190213  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 08/28/2022

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 02/26/2021 09:55

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

**Agency** UNION COUNTY SO **Date** 03/29/2021 **Type** Agency Inspection  
**Inspector** MCDANIEL,MARCUS,G **Time** 10:09:17 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1190213  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 08/28/2022

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/29/2021 10:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

**Agency** UNION COUNTY SO **Date** 04/28/2021 **Type** Agency Inspection  
**Inspector** MCDANIEL,MARCUS,G **Time** 10:05:55 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1190213  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 08/28/2022

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 04/28/2021 10:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

**Agency** UNION COUNTY SO **Date** 05/29/2021 **Type** Agency Inspection  
**Inspector** MCDANIEL,MARCUS,G **Time** 09:26:40 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** **Expiration**

**Number of Simulators** 4 **Compliance** No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.045  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1 | 0.047  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1          | 0.077  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 1   | INT    |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Alcohol Free Test # 2         | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # 2          | 0.044  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.05 g/210L Test # (Repeat) 2 | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2          | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2   | INT    |
| 0.08 g/210L Test: OK                        | Yes    | Alcohol Free Test # 3         | 0.000  |
| 0.20 g/210L Test: OK                        | No     | 0.05 g/210L Test # 3          | 0.046  |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | 0.05 g/210L Test # (Repeat) 3 | 0.047  |
|                                             |        | 0.08 g/210L Test # 3          | 0.078  |
|                                             |        | Interferent Detect Test # 3   | INT    |

#### Remarks:

A F / M A: Improper Sample. 05: Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 05/29/2021 10:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> UNION COUNTY SO      | <b>Date</b> 05/29/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> MCDANIEL,MARCUS,G | <b>Time</b> 10:26:01   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                               |
|------------------------------|------------------------------|------------------------------|-------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 1190213 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 08/28/2022  |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 4 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.074  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.078  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 05/29/2021 10:49

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

Agency BRADFORD COUNTY SO Date 01/18/2021 Type Agency Inspection  
Inspector LEVY, JAYSON, I Time 17:42:33 Software 8100.27

0.05 Lot # 201905A 0.08 Lot # 201908B 0.20 Lot # 201902C 0.08 Gas Lot # AG000606  
Expiration 05/14/2021 Expiration 08/07/2021 Expiration 02/20/2021 Expiration 01/06/2022

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 01/26/2021 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

**Agency** BRADFORD COUNTY SO      **Date** 02/13/2021      **Type** Agency Inspection  
**Inspector** LEVY, JAYSON, I      **Time** 20:19:55      **Software** 8100.27

**0.05 Lot #** 201905A      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** AG000606  
**Expiration** 05/14/2021      **Expiration** 08/07/2021      **Expiration** 02/20/2021      **Expiration** 01/06/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/29/2021 11:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

Agency BRADFORD COUNTY SO Date 03/20/2021 Type Agency Inspection  
Inspector LEVY, JAYSON, I Time 14:53:07 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # AG000606  
Expiration 01/12/2023 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 01/06/2022

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 03/29/2021 11:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

Agency BRADFORD COUNTY SO Date 04/05/2021 Type Agency Inspection  
Inspector LEVY, JAYSON, I Time 20:51:59 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # AG000606  
Expiration 01/12/2023 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 01/06/2022

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 04/12/2021 08:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

|                                  |                        |                                   |
|----------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> BRADFORD COUNTY SO | <b>Date</b> 04/12/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,    | <b>Time</b> 10:20:22   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 04/12/2021 10:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

Agency BRADFORD COUNTY SO Date 04/26/2021 Type Agency Inspection  
Inspector LEVY, JAYSON, I Time 05:51:50 Software 8100.27

0.05 Lot # 201910D 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # AG106803  
Expiration 10/22/2021 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 03/09/2023

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 04/26/2021 09:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

**Agency** BRADFORD COUNTY SO      **Date** 05/27/2021      **Type** Agency Inspection  
**Inspector** LEVY,JAYSON,I      **Time** 21:35:54      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** No

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: Sequence Aborted.INCORRECT READINGNon-compliance: INCORRECT READING INSP ERROR.

**Data Download Date/Time** 06/18/2021 08:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | BRADFORD COUNTY SO | <b>Date</b>       | 05/27/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LEVY,JAYSON,I      | <b>Time</b>       | 22:19:19   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201910D            | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | AG106803   |
| <b>Expiration</b>           | 10/22/2021         | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 03/09/2023 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: Ambient Fail UNKNOWN.

**Data Download Date/Time** 06/18/2021 08:59

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

|                           |                 |                        |  |
|---------------------------|-----------------|------------------------|--|
| Agency BRADFORD COUNTY SO | Date 06/02/2021 | Type Agency Inspection |  |
| Inspector LEVY, JAYSON, I | Time 17:41:12   | Software 8100.27       |  |

  

|                       |                       |                       |                         |
|-----------------------|-----------------------|-----------------------|-------------------------|
| 0.05 Lot # 201910D    | 0.08 Lot # 202007A    | 0.20 Lot # 202005A    | 0.08 Gas Lot # AG106803 |
| Expiration 10/22/2021 | Expiration 07/21/2022 | Expiration 05/12/2022 | Expiration 03/09/2023   |

  

|                        |                |
|------------------------|----------------|
| Number of Simulators 5 | Compliance Yes |
|------------------------|----------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

A F / M A: TOO MUCH MOUTH ALCOHOL.

Data Download Date/Time 06/18/2021 08:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

**Agency** BRADFORD COUNTY SO      **Date** 07/01/2021      **Type** Agency Inspection  
**Inspector** LEVY, JAYSON, I      **Time** 21:14:04      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG106803  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 03/09/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.208  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.207  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.206  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/30/2021 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

Agency BRADFORD COUNTY SO Date 08/30/2021 Type Agency Inspection  
Inspector LEVY, JAYSON, I Time 08:10:18 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202005A 0.20 Lot # 202005A 0.08 Gas Lot # AG106803  
Expiration 01/12/2023 Expiration 05/12/2022 Expiration 05/12/2022 Expiration 03/09/2023

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Data Download Date/Time 08/30/2021 12:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

**Agency** BRADFORD COUNTY SO      **Date** 10/08/2021      **Type** Agency Inspection  
**Inspector** LEVY, JAYSON, I      **Time** 10:39:15      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG000606  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 01/06/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 10/23/2021 07:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

**Agency** BRADFORD COUNTY SO      **Date** 12/12/2021      **Type** Agency Inspection  
**Inspector** LEVY, JAYSON, I      **Time** 16:44:20      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG000606  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 01/06/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 12/13/2021 11:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

**Agency** BRADFORD COUNTY SO      **Date** 01/30/2021      **Type** Agency Inspection  
**Inspector** LEVY, JAYSON, I      **Time** 20:57:24      **Software** 8100.27

**0.05 Lot #** 201905A      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** AG000606  
**Expiration** 05/14/2021      **Expiration** 08/07/2021      **Expiration** 02/20/2021      **Expiration** 01/06/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/09/2021 07:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

**Agency** BRADFORD COUNTY SO      **Date** 02/02/2021      **Type** Agency Inspection  
**Inspector** LEVY, JAYSON, I      **Time** 21:16:31      **Software** 8100.27

**0.05 Lot #** 201905A      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** AG000606  
**Expiration** 05/14/2021      **Expiration** 08/07/2021      **Expiration** 02/20/2021      **Expiration** 01/06/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.045  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/09/2021 07:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

**Agency** BRADFORD COUNTY SO **Date** 03/31/2021 **Type** Agency Inspection  
**Inspector** LEVY,JAYSON,I **Time** 20:47:06 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: Sequence Aborted BUTTON PUSHEDSequence Aborted. Non-compliance: MA.

**Data Download Date/Time** 04/05/2021 12:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | BRADFORD COUNTY SO | <b>Date</b>       | 03/31/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LEVY,JAYSON,I      | <b>Time</b>       | 22:53:00   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C            | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | AG000606   |
| <b>Expiration</b>           | 01/12/2023         | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 01/06/2022 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Data Download Date/Time 04/05/2021 12:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

Agency BRADFORD COUNTY SO Date 04/03/2021 Type Agency Inspection  
Inspector LEVY, JAYSON, I Time 23:20:52 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # AG000606  
Expiration 01/12/2023 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 01/06/2022

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 04/05/2021 12:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

Agency BRADFORD COUNTY SO Date 04/30/2021 Type Agency Inspection  
Inspector LEVY, JAYSON, I Time 18:27:19 Software 8100.27

0.05 Lot # 201910D 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # AG106803  
Expiration 10/22/2021 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 03/09/2023

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.045  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 05/03/2021 13:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

|                                     |                        |                                   |
|-------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> BRADFORD COUNTY SO    | <b>Date</b> 05/11/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> WILLIAMS,RICHARD,A | <b>Time</b> 13:04:57   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1                      | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1                      | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.080  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.199  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.000  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.080  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.079  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3               | INT    |
|                                             |        | Alcohol Free Test # 4                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 4                      | 0.049  |
|                                             |        | 0.08 g/210L Test # 4                      | 0.080  |
|                                             |        | 0.20 g/210L Test # 4                      | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4          | 0.079  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4               | INT    |
|                                             |        | Alcohol Free Test # 5                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 5                      | 0.049  |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

|                                            |       |
|--------------------------------------------|-------|
| 0.08 g/210L Test # 5                       | 0.080 |
| 0.20 g/210L Test # 5                       | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5           | 0.078 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 5  | 0.079 |
| Interferent Detect Test # 5                | INT   |
| Alcohol Free Test # 6                      | 0.000 |
| 0.05 g/210L Test # 6                       | 0.049 |
| 0.08 g/210L Test # 6                       | 0.080 |
| 0.20 g/210L Test # 6                       | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6           | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 6  | 0.079 |
| Interferent Detect Test # 6                | INT   |
| Alcohol Free Test # 7                      | 0.000 |
| 0.05 g/210L Test # 7                       | 0.049 |
| 0.08 g/210L Test # 7                       | 0.080 |
| 0.20 g/210L Test # 7                       | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 7           | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 7  | 0.078 |
| Interferent Detect Test # 7                | INT   |
| Alcohol Free Test # 8                      | 0.000 |
| 0.05 g/210L Test # 8                       | 0.049 |
| 0.08 g/210L Test # 8                       | 0.080 |
| 0.20 g/210L Test # 8                       | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8           | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 8  | 0.079 |
| Interferent Detect Test # 8                | INT   |
| Alcohol Free Test # 9                      | 0.000 |
| 0.05 g/210L Test # 9                       | 0.049 |
| 0.08 g/210L Test # 9                       | 0.080 |
| 0.20 g/210L Test # 9                       | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 9           | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 9  | 0.079 |
| Interferent Detect Test # 9                | INT   |
| Alcohol Free Test # 10                     | 0.000 |
| 0.05 g/210L Test # 10                      | 0.049 |
| 0.08 g/210L Test # 10                      | 0.080 |
| 0.20 g/210L Test # 10                      | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 10          | 0.079 |
| 0.08 g/210L Dry Gas Std Test # (Repeat) 10 | 0.079 |
| Interferent Detect Test # 10               | INT   |

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

|                                                    |        |
|----------------------------------------------------|--------|
| 0.05 g/210L Test (Std Dev)                         | 0.0000 |
| 0.08 g/210L Test (Std Dev)                         | 0.0000 |
| 0.20 g/210L Test (Std Dev)                         | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev)             | 0.0332 |
| 0.08 g/210L Dry Gas Std Test (Std Dev)<br>(Repeat) | 0.0004 |

**Remarks:**

|                                                                                                                   |
|-------------------------------------------------------------------------------------------------------------------|
| 08: Control Outside Tolerance, FORGOT TO PLUG IN HOSE. . 08: CONTOL OUTSIDE OF TOLERANCE, FORGOT TO PLUG IN HOSE. |
|-------------------------------------------------------------------------------------------------------------------|

---

**Data Download Date/Time** 05/11/2021 13:11

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

**Agency** BRADFORD COUNTY SO      **Date** 05/24/2021      **Type** Agency Inspection  
**Inspector** LEVY,JAYSON,I      **Time** 17:59:50      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** No

| Check or Test                               | Result | Test                        | Result |
|---------------------------------------------|--------|-----------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1       | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1        | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1        | 0.071  |
| Barometric Pressure Sensor Check: OK        |        | Interferent Detect Test # 1 | INT    |
| Alcohol Free Subject Test: 0.000            | Yes    | Alcohol Free Test # 2       | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.05 g/210L Test # 2        | 0.047  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.08 g/210L Test # 2        | 0.072  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 2 | INT    |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # 3       | 0.000  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # 3        | 0.047  |
| 0.08 g/210L Test: OK                        | No     | 0.08 g/210L Test # 3        | 0.072  |
| 0.20 g/210L Test: OK                        | No     | Interferent Detect Test # 3 | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                             |        |

#### Remarks:

08: Control Outside Tolerance SIM ISSUE. Non-compliance: .

**Data Download Date/Time** 05/26/2021 11:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

Agency BRADFORD COUNTY SO Date 05/25/2021 Type Agency Inspection  
Inspector LEVY, JAYSON, I Time 19:13:41 Software 8100.27

0.05 Lot # 201910D 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # AG106803  
Expiration 10/22/2021 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 03/09/2023

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 05/26/2021 11:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

Agency BRADFORD COUNTY SO Date 06/30/2021 Type Agency Inspection  
Inspector LEVY, JAYSON, I Time 19:27:20 Software 8100.27

0.05 Lot # 201910D 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # AG106803  
Expiration 10/22/2021 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 03/09/2023

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 07/01/2021 11:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

Agency BRADFORD COUNTY SO Date 07/01/2021 Type Agency Inspection  
Inspector LEVY, JAYSON, I Time 07:46:12 Software 8100.27

0.05 Lot # 201910D 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # AG106803  
Expiration 10/22/2021 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 03/09/2023

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 07/01/2021 11:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

Agency BRADFORD COUNTY SO Date 08/30/2021 Type Agency Inspection  
Inspector LEVY, JAYSON, I Time 16:26:09 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # AG106803  
Expiration 01/12/2023 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 03/09/2023

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 09/26/2021 08:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

Agency BRADFORD COUNTY SO Date 09/06/2021 Type Agency Inspection  
Inspector LEVY, JAYSON, I Time 11:18:09 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # AG106803  
Expiration 01/12/2023 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 03/09/2023

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.079  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.080  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

08: RFI Detect RADIO WAS KEYED UP.

Data Download Date/Time 09/26/2021 08:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

**Agency** BRADFORD COUNTY SO      **Date** 10/31/2021      **Type** Agency Inspection  
**Inspector** LEVY, JAYSON, I      **Time** 20:51:22      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG000606  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 01/06/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/02/2021 14:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | BRADFORD COUNTY SO | <b>Date</b>       | 11/01/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LEVY,JAYSON,I      | <b>Time</b>       | 09:24:19   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C            | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | AG000606   |
| <b>Expiration</b>           | 01/12/2023         | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 01/06/2022 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/02/2021 14:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | BRADFORD COUNTY SO | <b>Date</b>       | 12/09/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | LEVY,JAYSON,I      | <b>Time</b>       | 10:29:23   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C            | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | AG000606   |
| <b>Expiration</b>           | 01/12/2023         | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 01/06/2022 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Data Download Date/Time 01/03/2022 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

**Agency** LEVY COUNTY SO      **Date** 05/10/2021      **Type** Agency Inspection  
**Inspector** WILLIAMS,RICHARD,A      **Time** 15:29:24      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators**      **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT, NOT A COMPLIANCE CHECK

**Data Download Date/Time** 05/11/2021 12:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

|                                     |                        |                                   |
|-------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> LEVY COUNTY SO        | <b>Date</b> 05/11/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> WILLIAMS,RICHARD,A | <b>Time</b> 12:37:17   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.047  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.047  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.047  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.047  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.047  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.047  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.047  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 05/11/2021 12:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

Agency LEVY COUNTY SO Date 05/24/2021 Type Agency Inspection  
Inspector LEVY, JAYSON, I Time 17:33:43 Software 8100.27

0.05 Lot # 201910D 0.08 Lot # 0.20 Lot # 0.08 Gas Lot #  
Expiration 10/22/2021 Expiration Expiration Expiration

Number of Simulators 5 Compliance No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.041  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1 | 0.044  |
| Barometric Pressure Sensor Check: OK        |        | Interferent Detect Test # 1   | INT    |
| Alcohol Free Subject Test: 0.000            | Yes    | Alcohol Free Test # 2         | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.05 g/210L Test # 2          | 0.042  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # (Repeat) 2 | 0.045  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 2   | INT    |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # 3         | 0.000  |
| 0.05 g/210L Test: OK                        | No     | 0.05 g/210L Test # 3          | 0.042  |
| 0.08 g/210L Test: OK                        | No     | 0.05 g/210L Test # (Repeat) 3 | 0.045  |
| 0.20 g/210L Test: OK                        | No     | Interferent Detect Test # 3   | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                               |        |

#### Remarks:

05: Control Outside Tolerance SIMULATOR ISSUE, Control Outs. Non-compliance: .

Data Download Date/Time 05/26/2021 11:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

**Agency** LEVY COUNTY SO  
**Inspector** LEVY, JAYSON, I

**Date** 05/25/2021  
**Time** 19:30:07

**Type** Agency Inspection  
**Software** 8100.27

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> AG106803 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 03/09/2023   |

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.044  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.050  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.072  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Test # (Repeat) 1    | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.20 g/210L Test # 1             | 0.195  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Interferent Detect Test # 1      | INT    |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # 2            | 0.000  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # 2             | 0.044  |
| 0.08 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.050  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.081  |
|                                             |        | 0.20 g/210L Test # 2             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.044  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

05: Control Outside Tolerance UNKNOWN. 08: Control Outside Tolerance HOSE NOT CONNECTED ALL THE WA.

**Data Download Date/Time** 05/26/2021 11:53

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

Agency LEVY COUNTY SO Date 06/30/2021 Type Agency Inspection  
Inspector LEVY, JAYSON, I Time 17:24:32 Software 8100.27

0.05 Lot # 201910D 0.08 Lot # 0.20 Lot # 0.08 Gas Lot #  
Expiration 10/22/2021 Expiration Expiration Expiration

Number of Simulators 5 Compliance No

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                 | 0.044  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1        | 0.056  |
| Barometric Pressure Sensor Check: OK        |        | Interferent Detect Test # 1          | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # (Repeat) 1 | INT    |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Alcohol Free Test # 2                | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # 2                 | 0.051  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 2          | INT    |
| Alcohol Free Test: OK                       | Yes    | Interferent Detect Test # (Repeat) 2 | INT    |
| 0.05 g/210L Test: OK                        | No     | Alcohol Free Test # 3                | 0.000  |
| 0.08 g/210L Test: OK                        | No     | 0.05 g/210L Test # 3                 | 0.055  |
| 0.20 g/210L Test: OK                        | No     | Interferent Detect Test # 3          | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | Interferent Detect Test # (Repeat) 3 | INT    |

#### Remarks:

Int Det:HOSE NOT CONNECTED. 05: Control Outside Tolerance SIM TOP ISSUE, Seq ABT ERROR . Non-compliance:SIMULATOR ISSUE.

Data Download Date/Time 07/01/2021 11:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

**Agency** LEVY COUNTY SO  
**Inspector** LEVY, JAYSON, I

**Date** 06/30/2021  
**Time** 20:18:57

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG106803  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 03/09/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 07/01/2021 11:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

**Agency** LEVY COUNTY SO  
**Inspector** LEVY, JAYSON, I

**Date** 07/01/2021  
**Time** 07:42:15

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG106803  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 03/09/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 07/01/2021 11:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

**Agency** LEVY COUNTY SO  
**Inspector** LEVY, JAYSON, I

**Date** 08/30/2021  
**Time** 11:27:10

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG106803  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 03/09/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: Ambient Fail UNK WHY.

**Data Download Date/Time** 09/26/2021 08:19

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

**Agency** LEVY COUNTY SO  
**Inspector** LEVY, JAYSON, I

**Date** 09/06/2021  
**Time** 11:22:35

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG106803  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 03/09/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.045  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.045  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 09/26/2021 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

|                                  |                        |                               |  |
|----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> LEVY COUNTY SO     | <b>Date</b> 10/31/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> LEVY, JAYSON, I | <b>Time</b> 20:51:21   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202101C    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> AG000606 |
| <b>Expiration</b> 01/12/2023 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 01/06/2022   |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.000  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.046  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.075  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.047  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

05: Control Outside Tolerance HOSE NOT CONNECTED.

**Data Download Date/Time** 11/02/2021 14:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

**Agency** LEVY COUNTY SO  
**Inspector** LEVY, JAYSON, I

**Date** 11/01/2021  
**Time** 08:27:56

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG000606  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 01/06/2022

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 11/02/2021 14:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

|                           |                 |                        |  |
|---------------------------|-----------------|------------------------|--|
| Agency LEVY COUNTY SO     | Date 12/15/2021 | Type Agency Inspection |  |
| Inspector LEVY, JAYSON, I | Time 21:57:38   | Software 8100.27       |  |

  

|                       |                       |                       |                         |
|-----------------------|-----------------------|-----------------------|-------------------------|
| 0.05 Lot # 202101C    | 0.08 Lot # 202007A    | 0.20 Lot # 202106B    | 0.08 Gas Lot # AG000606 |
| Expiration 01/12/2023 | Expiration 07/21/2022 | Expiration 06/22/2023 | Expiration 01/06/2022   |

  

|                        |                |
|------------------------|----------------|
| Number of Simulators 5 | Compliance Yes |
|------------------------|----------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.000  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.051  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.078  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.052  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.051  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.054  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

05: Control Outside Tolerance HOSE NOT CONNECTED.

Data Download Date/Time 01/03/2022 09:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

|                  |                 |             |            |                 |                   |
|------------------|-----------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | WILLISTON PD    | <b>Date</b> | 01/27/2021 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | JOHNSON,DAVID,E | <b>Time</b> | 17:44:47   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |   |                   |    |
|-----------------------------|---|-------------------|----|
| <b>Number of Simulators</b> | 4 | <b>Compliance</b> | No |
|-----------------------------|---|-------------------|----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | Yes    |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: . Non-compliance: .

**Data Download Date/Time** 01/28/2021 14:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

**Agency** WILLISTON PD      **Date** 01/27/2021      **Type** Agency Inspection  
**Inspector** JOHNSON, DAVID, E      **Time** 18:30:26      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1355293  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 07/29/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/28/2021 14:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

**Agency** WILLISTON PD **Date** 02/25/2021 **Type** Agency Inspection  
**Inspector** JOHNSON,DAVID,E **Time** 20:19:03 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | Yes    |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: .

**Data Download Date/Time** 03/02/2021 10:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

**Agency** WILLISTON PD **Date** 02/25/2021 **Type** Agency Inspection  
**Inspector** JOHNSON,DAVID,E **Time** 20:47:56 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** 10/22/2021 **Expiration** **Expiration** **Expiration**

**Number of Simulators** 4 **Compliance** No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.036  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1 | 0.037  |
| Barometric Pressure Sensor Check: OK        |        | Interferent Detect Test # 1   | INT    |
| Alcohol Free Subject Test: 0.000            | Yes    | Alcohol Free Test # 2         | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.05 g/210L Test # 2          | 0.036  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # (Repeat) 2 | 0.038  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 2   | INT    |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # 3         | 0.000  |
| 0.05 g/210L Test: OK                        | No     | 0.05 g/210L Test # 3          | 0.036  |
| 0.08 g/210L Test: OK                        | No     | 0.05 g/210L Test # (Repeat) 3 | 0.039  |
| 0.20 g/210L Test: OK                        | No     | Interferent Detect Test # 3   | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                               |        |

#### Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 03/02/2021 10:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

**Agency** WILLISTON PD      **Date** 02/27/2021      **Type** Agency Inspection  
**Inspector** JOHNSON,DAVID,E      **Time** 01:12:01      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1355293  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 07/29/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/02/2021 10:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> WILLISTON PD         | <b>Date</b> 03/25/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> JOHNSON, DAVID, E | <b>Time</b> 17:08:03   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 135593 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 07/29/2023 |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 4 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                 | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                 | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                 | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1     | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1          | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # (Repeat) 1 | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                 | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                 | 0.079  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                 | 0.201  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2     | 0.078  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2          | 0.000  |
|                                             |        | Interferent Detect Test # (Repeat) 2 | INT    |
|                                             |        | Alcohol Free Test # 3                | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                 | 0.050  |
|                                             |        | 0.08 g/210L Test # 3                 | 0.080  |
|                                             |        | 0.20 g/210L Test # 3                 | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3     | 0.078  |
|                                             |        | Interferent Detect Test # 3          | 0.000  |
|                                             |        | Interferent Detect Test # (Repeat) 3 | INT    |

**Remarks:**

Int Det: .

**Data Download Date/Time** 03/26/2021 11:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

**Agency** WILLISTON PD **Date** 04/30/2021 **Type** Agency Inspection  
**Inspector** JOHNSOM,DAVID,E **Time** 14:04:08 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 135593  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 07/29/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/30/2021 14:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

**Agency** WILLISTON PD      **Date** 05/26/2021      **Type** Agency Inspection  
**Inspector** JOHNSON,DAVID,E      **Time** 11:53:33      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: .

**Data Download Date/Time** 05/26/2021 12:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

**Agency** WILLISTON PD  
**Inspector** JOHNSON, DAVID, E

**Date** 05/26/2021  
**Time** 12:48:48

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 135593  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/29/2023

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 05/26/2021 12:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

**Agency** WILLISTON PD **Date** 07/14/2021 **Type** Agency Inspection  
**Inspector** WILLIAMS,RICHARD,A **Time** 11:19:19 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

**Remarks:**

**Data Download Date/Time** 07/14/2021 13:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

|                                     |                        |                                   |
|-------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> WILLISTON PD          | <b>Date</b> 07/14/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> WILLIAMS,RICHARD,A | <b>Time</b> 13:24:50   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.076  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.076  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.076  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.076  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.076  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.076  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.076  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.076  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 07/14/2021 13:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

**Agency** WILLISTON PD **Date** 07/26/2021 **Type** Agency Inspection  
**Inspector** JOHNSON,DAVID,E **Time** 13:15:28 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 135593  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 07/29/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/26/2021 14:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

**Agency** WILLISTON PD **Date** 08/26/2021 **Type** Agency Inspection  
**Inspector** JOHNSON, DAVID, E **Time** 16:05:35 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005 **0.08 Gas Lot #** 135593  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 07/29/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 09/30/2021 07:19

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

**Agency** WILLISTON PD      **Date** 08/29/2021      **Type** Agency Inspection  
**Inspector** JOHNSON,DAVID,E      **Time** 15:08:11      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 0      **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

**Remarks:**

**Data Download Date/Time** 09/30/2021 07:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

Agency WILLISTON PD Date 09/26/2021 Type Agency Inspection  
Inspector JOHNSON, DAVID, E Time 16:27:29 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # 135593  
Expiration 01/12/2023 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 07/29/2023

Number of Simulators 4 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 09/30/2021 07:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

**Agency** WILLISTON PD      **Date** 09/29/2021      **Type** Agency Inspection  
**Inspector** JOHNSON,DAVID,E      **Time** 11:06:50      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 4      **Compliance** No

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | Yes    |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: . Non-compliance: .

**Data Download Date/Time** 09/30/2021 07:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

**Agency** WILLISTON PD **Date** 09/29/2021 **Type** Agency Inspection  
**Inspector** JOHNSON, DAVID, E **Time** 11:46:46 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 135593  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 07/29/2023

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | No     | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 09/30/2021 07:19

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

**Agency** WILLISTON PD **Date** 09/29/2021 **Type** Agency Inspection  
**Inspector** JOHNSON,DAVID,E **Time** 15:19:05 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 135593  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 07/29/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 09/30/2021 07:19

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

Agency WILLISTON PD Date 11/22/2021 Type Agency Inspection  
Inspector JOHNSON, DAVID, E Time 16:52:57 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # 135593  
Expiration 01/12/2023 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 07/29/2023

Number of Simulators 4 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 12/16/2021 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

**Agency** WILLISTON PD **Date** 12/16/2021 **Type** Agency Inspection  
**Inspector** JOHNSON, DAVID, E **Time** 12:16:16 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 135593  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 07/29/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 12/16/2021 12:18

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

|                  |                 |             |            |                 |                   |
|------------------|-----------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | WILLISTON PD    | <b>Date</b> | 12/16/2021 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | JOHNSON,DAVID,E | <b>Time</b> | 12:21:06   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

Remarks:

**Data Download Date/Time** 12/16/2021 12:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

**Agency** SANTA ROSA COUNTY SO **Date** 01/30/2021 **Type** Agency Inspection  
**Inspector** NELSON,DURON,A **Time** 17:30:09 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1349789  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 07/15/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.053  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.085  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.208  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.053  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.085  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.207  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.053  |
|                                             |        | 0.08 g/210L Test # 3             | 0.085  |
|                                             |        | 0.20 g/210L Test # 3             | 0.208  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/01/2021 12:33

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

**Agency** SANTA ROSA COUNTY SO      **Date** 02/26/2021      **Type** Agency Inspection  
**Inspector** NELSON,DURON,A      **Time** 16:19:00      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1349789  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/15/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/02/2021 17:50

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

**Agency** SANTA ROSA COUNTY SO      **Date** 03/31/2021      **Type** Agency Inspection  
**Inspector** NELSON,DURON,A      **Time** 18:51:12      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1349789  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/15/2022      **Expiration** 07/15/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 04/05/2021 10:42

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

**Agency** SANTA ROSA COUNTY SO      **Date** 04/30/2021      **Type** Agency Inspection  
**Inspector** NELSON,DURON,A      **Time** 16:34:11      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1349789  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/15/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 05/05/2021 10:17

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

**Agency** SANTA ROSA COUNTY SO **Date** 05/28/2021 **Type** Agency Inspection  
**Inspector** NELSON,DURON,A **Time** 19:01:15 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1349789  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 07/15/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/28/2021 20:11

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

**Agency** SANTA ROSA COUNTY SO **Date** 06/24/2021 **Type** Agency Inspection  
**Inspector** NELSON,DURON,A **Time** 14:09:44 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1349789  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 07/15/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/28/2021 17:44

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

**Agency** SANTA ROSA COUNTY SO **Date** 07/30/2021 **Type** Agency Inspection  
**Inspector** NELSON,DURON,A **Time** 00:26:41 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1349789  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 07/15/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 08/02/2021 17:33

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

**Agency** SANTA ROSA COUNTY SO      **Date** 09/01/2021      **Type** Agency Inspection  
**Inspector** NELSON,DURON,A      **Time** 00:31:04      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1482048  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/21/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 09/01/2021 01:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> SANTA ROSA COUNTY SO | <b>Date</b> 09/08/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,      | <b>Time</b> 09:49:24   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 09/08/2021 10:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

**Agency** SANTA ROSA COUNTY SO **Date** 09/29/2021 **Type** Agency Inspection  
**Inspector** NELSON,DURON,A **Time** 20:12:24 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1482048  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 04/21/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 10/01/2021 15:52

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

**Agency** SANTA ROSA COUNTY SO      **Date** 10/29/2021      **Type** Agency Inspection  
**Inspector** NELSON,DURON,A      **Time** 17:11:47      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1482048  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/21/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 11/03/2021 12:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

**Agency** SANTA ROSA COUNTY SO      **Date** 11/30/2021      **Type** Agency Inspection  
**Inspector** NELSON,DURON,A      **Time** 20:05:33      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1482048  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 04/21/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.082  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.079  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.198  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.081  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.078  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3               | INT    |

#### Remarks:

08: RFI Detect.

**Data Download Date/Time** 11/30/2021 21:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

**Agency** SANTA ROSA COUNTY SO      **Date** 12/31/2021      **Type** Agency Inspection  
**Inspector** NELSON,DURON,A      **Time** 23:10:32      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1482048  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/21/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/04/2022 00:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

**Agency** SANTA ROSA COUNTY SO **Date** 01/30/2021 **Type** Agency Inspection  
**Inspector** NELSON,DURON,A **Time** 19:04:32 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1349789  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 07/15/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/01/2021 12:35

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

**Agency** SANTA ROSA COUNTY SO      **Date** 02/26/2021      **Type** Agency Inspection  
**Inspector** NELSON,DURON,A      **Time** 16:28:11      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1349789  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/15/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 03/02/2021 17:43

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

**Agency** SANTA ROSA COUNTY SO      **Date** 03/31/2021      **Type** Agency Inspection  
**Inspector** NELSON,DURON,A      **Time** 18:43:33      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1349789  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/15/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.083  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.205  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.083  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.205  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.083  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 04/05/2021 10:40

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

**Agency** SANTA ROSA COUNTY SO **Date** 04/30/2021 **Type** Agency Inspection  
**Inspector** NELSON,DURON,A **Time** 17:11:27 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1349789  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 07/15/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.205  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/05/2021 20:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

**Agency** SANTA ROSA COUNTY SO      **Date** 05/28/2021      **Type** Agency Inspection  
**Inspector** NELSON,DURON,A      **Time** 19:14:47      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1349789  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/15/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.083  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 06/24/2021 14:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

**Agency** SANTA ROSA COUNTY SO      **Date** 06/24/2021      **Type** Agency Inspection  
**Inspector** NELSON,DURON,A      **Time** 14:06:39      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1349789  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/15/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/28/2021 17:41

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

**Agency** SANTA ROSA COUNTY SO **Date** 07/30/2021 **Type** Agency Inspection  
**Inspector** NELSON,DURON,A **Time** 00:20:07 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1349789  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 07/15/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 08/02/2021 17:20

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> SANTA ROSA COUNTY SO | <b>Date</b> 08/06/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,      | <b>Time</b> 13:26:35   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.202  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.203  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Test (Std Dev)             | 0.0006 |
| 0.20 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 08/06/2021 14:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

**Agency** SANTA ROSA COUNTY SO      **Date** 09/01/2021      **Type** Agency Inspection  
**Inspector** NELSON,DURON,A      **Time** 00:43:11      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1482048  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/21/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 09/04/2021 23:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

**Agency** SANTA ROSA COUNTY SO      **Date** 09/29/2021      **Type** Agency Inspection  
**Inspector** NELSON,DURON,A      **Time** 20:03:39      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1482048  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/21/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 10/01/2021 15:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

**Agency** SANTA ROSA COUNTY SO **Date** 10/29/2021 **Type** Agency Inspection  
**Inspector** NELSON,DURON,A **Time** 17:14:02 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1 | 0.048  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1          | 0.078  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Test # (Repeat) 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1   | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2         | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.05 g/210L Test # 2          | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # (Repeat) 2 | 0.047  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2          | 0.077  |
| 0.08 g/210L Test: OK                        | No     | Interferent Detect Test # 2   | INT    |
| 0.20 g/210L Test: OK                        | No     | Alcohol Free Test # 3         | 0.000  |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | 0.05 g/210L Test # (Repeat) 3 | 0.047  |
|                                             |        | Interferent Detect Test # 3   | INT    |

#### Remarks:

05: Ambient Fail. 08: Ambient Fail, Ambient Fail. AMB FAIL.

**Data Download Date/Time** 11/08/2021 14:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> SANTA ROSA COUNTY SO | <b>Date</b> 11/09/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,      | <b>Time</b> 10:44:16   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.051  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0006 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 11/09/2021 11:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

**Agency** SANTA ROSA COUNTY SO      **Date** 11/30/2021      **Type** Agency Inspection  
**Inspector** NELSON,DURON,A      **Time** 20:19:25      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1482048  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 04/21/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: RFI DetectRFI Detect.

**Data Download Date/Time** 11/30/2021 21:33

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

**Agency** SANTA ROSA COUNTY SO **Date** 12/31/2021 **Type** Agency Inspection  
**Inspector** NELSON,DURON,A **Time** 23:17:16 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1482048  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 04/21/2024

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/04/2022 00:12

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

|                                 |                        |                               |  |
|---------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MARIANNA PD       | <b>Date</b> 01/17/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> RUBEL, NEIL, F | <b>Time</b> 02:53:38   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                              |
|------------------------------|------------------------------|------------------------------|------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 201902C    | <b>0.08 Gas Lot #</b> 970670 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 02/20/2021 | <b>Expiration</b> 06/07/2021 |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.192  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.190  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

A F / M A: RFI Detect. LAPTOP AIR CARD TURNED ON

**Data Download Date/Time** 01/19/2021 11:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

|                               |                        |                                   |
|-------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MARIANNA PD     | <b>Date</b> 01/25/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL, | <b>Time</b> 09:12:18   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.081  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.081  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 01/25/2021 10:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

|                  |              |             |            |                 |                   |
|------------------|--------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | MARIANNA PD  | <b>Date</b> | 04/05/2021 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | SOTO,ISRAEL, | <b>Time</b> | 08:20:01   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

**Data Download Date/Time** 04/05/2021 12:50

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

|                               |                        |                                   |
|-------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MARIANNA PD     | <b>Date</b> 04/05/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL, | <b>Time</b> 11:45:36   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.081  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.081  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.081  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 04/05/2021 12:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

**Agency** MARIANNA PD **Date** 04/08/2021 **Type** Agency Inspection  
**Inspector** BARGE,SUSAN,M **Time** 15:26:25 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908D **0.20 Lot #** 202005A **0.08 Gas Lot #** 1346146  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 07/07/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NA

**Data Download Date/Time** 04/09/2021 10:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

**Agency** MARIANNA PD      **Date** 05/10/2021      **Type** Agency Inspection  
**Inspector** BARGE,SUSAN,M      **Time** 13:26:55      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1346146  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 07/07/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NA

**Data Download Date/Time** 05/10/2021 14:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

**Agency** MARIANNA PD  
**Inspector** RUBEL, NEIL, F

**Date** 06/09/2021  
**Time** 11:18:20

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1346146  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 07/07/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

N/A

**Data Download Date/Time** 06/17/2021 10:52

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

**Agency** MARIANNA PD **Date** 07/01/2021 **Type** Agency Inspection  
**Inspector** BARGE,SUSAN,M **Time** 13:30:48 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 1346146  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 07/07/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NA

**Data Download Date/Time** 07/07/2021 10:45

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

**Agency** MARIANNA PD **Date** 08/29/2021 **Type** Agency Inspection  
**Inspector** LEE,JERRY,W **Time** 10:04:32 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | Yes    |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: .

**Data Download Date/Time** 09/05/2021 03:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

**Agency** MARIANNA PD  
**Inspector** LEE,JERRY,W

**Date** 08/29/2021  
**Time** 10:40:09

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1346146  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/07/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 09/05/2021 03:27

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

|                  |             |             |            |                 |                   |
|------------------|-------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | MARIANNA PD | <b>Date</b> | 09/30/2021 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | LEE,JERRY,W | <b>Time</b> | 08:18:06   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |   |                   |    |
|-----------------------------|---|-------------------|----|
| <b>Number of Simulators</b> | 5 | <b>Compliance</b> | No |
|-----------------------------|---|-------------------|----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: Range Exceeded. Non-compliance:TESTED IN THE WRONG ORDER.

**Data Download Date/Time** 10/11/2021 10:17

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

|                  |             |             |            |                 |                   |
|------------------|-------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | MARIANNA PD | <b>Date</b> | 09/30/2021 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | LEE,JERRY,W | <b>Time</b> | 08:19:45   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

Remarks:

**Data Download Date/Time** 10/11/2021 10:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

**Agency** MARIANNA PD      **Date** 09/30/2021      **Type** Agency Inspection  
**Inspector** LEE,JERRY,W      **Time** 08:33:30      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                        | Result |
|---------------------------------------------|--------|-----------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1       | 0.000  |
| Date and/or Time Adjusted                   | No     | Interferent Detect Test # 1 | 0.000  |
| Minimum Sample Volume Check: OK             |        | Alcohol Free Test # 2       | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | Interferent Detect Test # 2 | 0.000  |
| Alcohol Free Subject Test: 0.000            | No     | Alcohol Free Test # 3       | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 3 | 0.000  |
| Interferent Detect Test: Interferent Detect | No     |                             |        |
| Diagnostic Check (Post-Inspection): OK      | No     |                             |        |
| Alcohol Free Test: OK                       | Yes    |                             |        |
| 0.05 g/210L Test: OK                        | No     |                             |        |
| 0.08 g/210L Test: OK                        | No     |                             |        |
| 0.20 g/210L Test: OK                        | No     |                             |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                             |        |

#### Remarks:

Int Det: .

**Data Download Date/Time** 10/11/2021 10:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

**Agency** MARIANNA PD **Date** 09/30/2021 **Type** Agency Inspection  
**Inspector** LEE,JERRY,W **Time** 08:42:25 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                  | Result |
|---------------------------------------------|--------|-----------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1 | 0.000  |
| Date and/or Time Adjusted                   | No     | Alcohol Free Test # 2 | 0.000  |
| Minimum Sample Volume Check: OK             |        | Alcohol Free Test # 3 | 0.000  |
| Barometric Pressure Sensor Check: OK        |        |                       |        |
| Alcohol Free Subject Test: 0.000            | Yes    |                       |        |
| Mouth Alcohol Test: Slope Not Met           | Yes    |                       |        |
| Interferent Detect Test: Interferent Detect | No     |                       |        |
| Diagnostic Check (Post-Inspection): OK      | No     |                       |        |
| Alcohol Free Test: OK                       | Yes    |                       |        |
| 0.05 g/210L Test: OK                        | No     |                       |        |
| 0.08 g/210L Test: OK                        | No     |                       |        |
| 0.20 g/210L Test: OK                        | No     |                       |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                       |        |

**Remarks:**

**Data Download Date/Time** 10/11/2021 10:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

|                  |             |             |            |                 |                   |
|------------------|-------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | MARIANNA PD | <b>Date</b> | 09/30/2021 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | LEE,JERRY,W | <b>Time</b> | 08:49:48   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |   |                   |    |
|-----------------------------|---|-------------------|----|
| <b>Number of Simulators</b> | 5 | <b>Compliance</b> | No |
|-----------------------------|---|-------------------|----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | Yes    |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: Range Exceeded. Non-compliance:IMPROPER TEST.

**Data Download Date/Time** 10/11/2021 10:17

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

**Agency** MARIANNA PD **Date** 09/30/2021 **Type** Agency Inspection  
**Inspector** LEE,JERRY,W **Time** 09:56:46 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1346146  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 07/07/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 10/11/2021 10:17

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

**Agency** MARIANNA PD  
**Inspector** LEE,JERRY,W

**Date** 10/11/2021  
**Time** 08:56:50

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1346146  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 07/07/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/11/2021 10:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

**Agency** MARIANNA PD  
**Inspector** LEE,JERRY,W

**Date** 11/30/2021  
**Time** 11:12:08

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1346146  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/07/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 12/13/2021 10:07

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

**Agency** MARIANNA PD  
**Inspector** LEE,JERRY,W

**Date** 12/03/2021  
**Time** 09:28:04

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1346146  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/07/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 12/13/2021 10:07

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

|                                      |                        |                               |  |
|--------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> GULF BREEZE PD         | <b>Date</b> 01/04/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> ROBERSON, THOMAS, C | <b>Time</b> 14:03:37   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                               |
|------------------------------|------------------------------|------------------------------|-------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 1298240 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 03/19/2023  |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 4 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.085  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.082  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.047  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.082  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.198  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.086  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.082  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.048  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.082  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.086  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3               | INT    |

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 01/05/2021 08:33

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

|                                      |                        |                               |  |
|--------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> GULF BREEZE PD         | <b>Date</b> 02/04/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> ROBERSON, THOMAS, C | <b>Time</b> 00:37:18   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 17919080A1 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 08/05/2021     |

  

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                 | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                 | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                 | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1     | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1          | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # (Repeat) 1 | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                 | 0.047  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                 | 0.082  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                 | 0.201  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2     | 0.081  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2          | INT    |
|                                             |        | Interferent Detect Test # (Repeat) 2 | INT    |
|                                             |        | Alcohol Free Test # 3                | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                 | 0.048  |
|                                             |        | 0.08 g/210L Test # 3                 | 0.082  |
|                                             |        | 0.20 g/210L Test # 3                 | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3     | 0.081  |
|                                             |        | Interferent Detect Test # 3          | INT    |
|                                             |        | Interferent Detect Test # (Repeat) 3 | INT    |

**Remarks:**

Int Det: .

**Data Download Date/Time** 02/04/2021 03:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

**Agency** GULF BREEZE PD      **Date** 03/04/2021      **Type** Agency Inspection  
**Inspector** ROBERSON, THOMAS, C      **Time** 04:00:51      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 17919080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 08/05/2021

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.083  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.083  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/05/2021 14:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

|                                      |                        |                               |  |
|--------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> GULF BREEZE PD         | <b>Date</b> 04/11/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> ROBERSON, THOMAS, C | <b>Time</b> 22:09:42   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 17919080A1 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 08/05/2021     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 4 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.201  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.000  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.201  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.000  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time** 04/12/2021 01:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

**Agency** GULF BREEZE PD      **Date** 05/04/2021      **Type** Agency Inspection  
**Inspector** ROBERSON, THOMAS, C      **Time** 03:00:42      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 17919080A  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 08/05/2021

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/04/2021 04:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

**Agency** GULF BREEZE PD      **Date** 06/01/2021      **Type** Agency Inspection  
**Inspector** ROBERSON, THOMAS, C      **Time** 22:29:58      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 17919080A  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 08/05/2021

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/02/2021 03:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

**Agency** GULF BREEZE PD **Date** 07/12/2021 **Type** Agency Inspection  
**Inspector** ROBERSON, THOMAS, C **Time** 11:22:02 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 07/12/2021 12:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

|                               |                        |                                   |
|-------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> GULF BREEZE PD  | <b>Date</b> 07/14/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL, | <b>Time</b> 09:34:00   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 07/14/2021 10:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

**Agency** GULF BREEZE PD **Date** 07/17/2021 **Type** Agency Inspection  
**Inspector** ROBERSON, THOMAS, C **Time** 09:52:37 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 07/19/2021 10:30

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

**Agency** GULF BREEZE PD **Date** 08/05/2021 **Type** Agency Inspection  
**Inspector** ROBERSON, THOMAS, C **Time** 08:56:31 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1433248  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 01/06/2024

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 08/05/2021 11:08

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

|                                      |                        |                               |  |
|--------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> GULF BREEZE PD         | <b>Date</b> 09/02/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> ROBERSON, THOMAS, C | <b>Time</b> 10:18:39   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                               |
|------------------------------|------------------------------|------------------------------|-------------------------------|
| <b>0.05 Lot #</b> 202101C    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 1433248 |
| <b>Expiration</b> 01/12/2023 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 01/06/2024  |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 4 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.081  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.201  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.000  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.081  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.048  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.081  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.000  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3               | INT    |

**Remarks:**

08: Control Outside Tolerance.INTAKE HOSE NOT CONECTED. REC

**Data Download Date/Time** 11/05/2021 09:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

**Agency** GULF BREEZE PD      **Date** 11/05/2021      **Type** Agency Inspection  
**Inspector** ROBERSON, THOMAS, C      **Time** 09:48:45      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1433248  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 01/06/2024

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.082  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 11/05/2021 12:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001306

**Agency** GULF BREEZE PD      **Date** 12/18/2021      **Type** Agency Inspection  
**Inspector** ROBERSON, THOMAS, C      **Time** 10:04:43      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1356262  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 08/03/2023

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 12/18/2021 11:23

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

|                                  |                        |                               |  |
|----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MADISON COUNTY SO  | <b>Date</b> 01/11/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> HENING, FRED, C | <b>Time</b> 11:08:46   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 201902C    | <b>0.08 Gas Lot #</b> 00919080A2 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 02/20/2021 | <b>Expiration</b> 03/05/2021     |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.212  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.201  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.208  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.197  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.206  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

20: Control Outside Tolerance.

**Data Download Date/Time** 01/11/2021 11:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

**Agency** MADISON COUNTY SO      **Date** 02/17/2021      **Type** Agency Inspection  
**Inspector** HENING,FRED,C      **Time** 13:36:10      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** 00919080A2  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 02/20/2021      **Expiration** 03/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.045  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 02/19/2021 10:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

|                                 |                        |                                   |
|---------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> MADISON COUNTY SO | <b>Date</b> 02/19/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,   | <b>Time</b> 13:36:27   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 02/19/2021 13:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

|                                  |                              |                               |                                  |
|----------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> MADISON COUNTY SO  | <b>Date</b> 03/22/2021       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> HENING, FRED, C | <b>Time</b> 09:46:11         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201910D        | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> 34520080A1 |
| <b>Expiration</b> 10/22/2021     | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 02/05/2023     |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

NONE

**Data Download Date/Time** 03/22/2021 10:00

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

**Agency** MADISON COUNTY SO      **Date** 03/24/2021      **Type** Agency Inspection  
**Inspector** HENING,FRED,C      **Time** 14:20:14      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration** 10/22/2021      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** No

| Check or Test                               | Result | Test                           | Result |
|---------------------------------------------|--------|--------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1 | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1           | 0.045  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1  | 0.043  |
| Barometric Pressure Sensor Check: OK        |        | Interferent Detect Test # 1    | INT    |
| Alcohol Free Subject Test: 0.000            | Yes    | Alcohol Free Test # (Repeat) 2 | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.05 g/210L Test # 2           | 0.044  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # (Repeat) 2  | 0.044  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 2    | INT    |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # (Repeat) 3 | 0.000  |
| 0.05 g/210L Test: OK                        | No     | 0.05 g/210L Test # 3           | 0.045  |
| 0.08 g/210L Test: OK                        | No     | 0.05 g/210L Test # (Repeat) 3  | 0.044  |
| 0.20 g/210L Test: OK                        | No     | Interferent Detect Test # 3    | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                                |        |

#### Remarks:

00: Ambient Fail. 05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 03/25/2021 13:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

|                                  |                        |                               |  |
|----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MADISON COUNTY SO  | <b>Date</b> 03/25/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> HENING, FRED, C | <b>Time</b> 11:35:26   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 34520080A1 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 02/05/2023     |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.082  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.080  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.203  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.000  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.083  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.080  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.082  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3               | INT    |

**Remarks:**

08: Control Outside Tolerance INLET TUBE NOT SECURE.

**Data Download Date/Time** 03/25/2021 13:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

**Agency** MADISON COUNTY SO      **Date** 04/26/2021      **Type** Agency Inspection  
**Inspector** HENING,FRED,C      **Time** 13:28:17      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 34520080A1  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 04/26/2021 13:36

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

**Agency** MADISON COUNTY SO      **Date** 05/18/2021      **Type** Agency Inspection  
**Inspector** HENING,FRED,C      **Time** 11:52:31      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 34520080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 05/18/2021 11:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

**Agency** MADISON COUNTY SO      **Date** 06/09/2021      **Type** Agency Inspection  
**Inspector** HENING,FRED,C      **Time** 11:04:48      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 34520080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 06/09/2021 11:09

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

**Agency** MADISON COUNTY SO      **Date** 07/14/2021      **Type** Agency Inspection  
**Inspector** HENING,FRED,C      **Time** 13:02:25      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 34520080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 07/14/2021 13:06

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

**Agency** MADISON COUNTY SO      **Date** 08/11/2021      **Type** Agency Inspection  
**Inspector** HENING,FRED,C      **Time** 14:08:20      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 34520080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 08/11/2021 14:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

**Agency** MADISON COUNTY SO      **Date** 09/22/2021      **Type** Agency Inspection  
**Inspector** HENING,FRED,C      **Time** 12:13:00      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 34520080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 09/22/2021 12:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

**Agency** MADISON COUNTY SO      **Date** 10/20/2021      **Type** Agency Inspection  
**Inspector** HENING,FRED,C      **Time** 14:55:10      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 34520080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 02/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 10/20/2021 15:00

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

|                                  |                        |                               |  |
|----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> MADISON COUNTY SO  | <b>Date</b> 11/08/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> HENING, FRED, C | <b>Time</b> 14:11:56   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202101C    | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B    | <b>0.08 Gas Lot #</b> 34520080A1 |
| <b>Expiration</b> 01/12/2023 | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023 | <b>Expiration</b> 02/05/2023     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.074  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.077  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.075  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

08: Control Outside Tolerance UNKNOWN.

**Data Download Date/Time** 11/08/2021 14:17

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

**Agency** MADISON COUNTY SO      **Date** 12/28/2021      **Type** Agency Inspection  
**Inspector** HENING,FRED,C      **Time** 14:39:56      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 03521080A3  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 05/12/2022      **Expiration** 03/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.075  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.075  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

NONE

**Data Download Date/Time** 12/28/2021 14:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 01/25/2021  
**Time** 16:00:00

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG923908  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 08/27/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/28/2021 10:00

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 02/26/2021  
**Time** 09:25:39

**Type** Agency Inspection  
**Software** 8100.27

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> AG923908 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 08/27/2021   |

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.078  |
| Alcohol Free Subject Test: 0.000            | No     | 0.20 g/210L Test # 1             | 0.199  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.000  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.000  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

SIMULATOR NOT PLUGGED INTO SIDE OF INSTRUMENT.

**Data Download Date/Time** 02/26/2021 12:00

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

|                               |                        |                                   |
|-------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> BAY COUNTY SO   | <b>Date</b> 03/05/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL, | <b>Time</b> 09:43:41   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.081  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.081  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0007 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 03/05/2021 10:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

**Agency** BAY COUNTY SO **Date** 03/05/2021 **Type** Agency Inspection  
**Inspector** SOTO,ISRAEL, **Time** 10:05:07 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

CHECKING IF FORM 40 IS ON INSTRUMENT

**Data Download Date/Time** 03/05/2021 11:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 03/12/2021  
**Time** 14:05:27

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG923908  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 08/27/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 03/12/2021 15:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 03/30/2021  
**Time** 10:37:28

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG923908  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 08/27/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 03/30/2021 13:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 04/20/2021  
**Time** 10:51:45

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG923908  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 08/27/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 04/20/2021 14:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 05/21/2021  
**Time** 14:38:20

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG923908  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 08/27/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 05/21/2021 17:00

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 06/28/2021  
**Time** 12:17:16

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG923908  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 08/27/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 06/28/2021 16:33

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 07/16/2021  
**Time** 15:23:50

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG923908  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 08/27/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 07/16/2021 17:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 08/03/2021  
**Time** 13:21:38

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** AG923908  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 08/27/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 08/04/2021 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 09/17/2021  
**Time** 10:07:56

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1305520  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 04/01/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 09/17/2021 11:36

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 10/13/2021  
**Time** 09:45:29

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1305520  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 04/01/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 10/13/2021 12:27

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 11/17/2021  
**Time** 09:40:08

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1305520  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 04/01/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 11/17/2021 13:57

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 12/21/2021  
**Time** 09:24:58

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1305520  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 04/01/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 12/21/2021 12:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 01/25/2021  
**Time** 15:26:46

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1303107  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 03/23/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/27/2021 11:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

|                               |                        |                                   |
|-------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> BAY COUNTY SO   | <b>Date</b> 01/27/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL, | <b>Time</b> 10:33:44   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.048  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.048  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 01/27/2021 11:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 01/27/2021  
**Time** 14:36:26

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1303107  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 03/28/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/28/2021 09:53

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 02/26/2021  
**Time** 09:20:15

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1303107  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 03/28/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 02/26/2021 11:45

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 03/30/2021  
**Time** 11:20:41

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1303107  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 03/28/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 03/30/2021 12:57

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 04/16/2021  
**Time** 13:30:32

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1303107  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 03/28/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/20/2021 10:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

**Agency** BAY COUNTY SO  
**Inspector** MILLER,BRIAN,D

**Date** 05/21/2021  
**Time** 15:25:55

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1303107  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 03/28/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 05/21/2021 16:58

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

**Agency** BAY COUNTY SO **Date** 07/02/2021 **Type** Agency Inspection  
**Inspector** WILLIAMS,RICHARD,A **Time** 08:42:35 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI TO OPERAE INSTRUMENT NOT A COMPLIANCE CHECK.

**Data Download Date/Time** 07/07/2021 12:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

|                                    |                        |                                   |
|------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> BAY COUNTY SO        | <b>Date</b> 12/22/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GRAHAM, THOMAS, J | <b>Time</b> 12:08:02   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.051  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.077  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.077  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.051  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.051  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.077  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 12/22/2021 13:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

**Agency** NICEVILLE PD      **Date** 01/19/2021      **Type** Agency Inspection  
**Inspector** CARRICO,KELLY,D      **Time** 10:50:09      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** 1283438  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 02/20/2021      **Expiration** 02/25/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/19/2021 11:55

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

|                                  |                              |                               |                               |
|----------------------------------|------------------------------|-------------------------------|-------------------------------|
| <b>Agency</b> NICEVILLE PD       | <b>Date</b> 02/11/2021       | <b>Type</b> Agency Inspection |                               |
| <b>Inspector</b> CARRICO,KELLY,D | <b>Time</b> 10:05:52         | <b>Software</b> 8100.27       |                               |
| <b>0.05 Lot #</b> 201910D        | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 201902C     | <b>0.08 Gas Lot #</b> 1283438 |
| <b>Expiration</b> 10/22/2021     | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 02/20/2021  | <b>Expiration</b> 02/25/2023  |
| <b>Number of Simulators</b> 5    | <b>Compliance</b> Yes        |                               |                               |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/11/2021 11:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

|                               |                        |                                   |
|-------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> NICEVILLE PD    | <b>Date</b> 04/29/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL, | <b>Time</b> 09:56:51   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.048  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0003 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 04/29/2021 10:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

Agency NICEVILLE PD Date 04/30/2021 Type Agency Inspection  
Inspector CARRICO,KELLY,D Time 14:32:12 Software 8100.27

0.05 Lot # 201910D 0.08 Lot # 201908B 0.20 Lot # 202005A 0.08 Gas Lot # 1283438  
Expiration 10/22/2021 Expiration 08/07/2021 Expiration 05/12/2022 Expiration 02/25/2023

Number of Simulators 5 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

Data Download Date/Time 04/30/2021 16:52

Printed: 14 January 2022  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

**Agency** NICEVILLE PD **Date** 05/27/2021 **Type** Agency Inspection  
**Inspector** CARRICO,KELLY,D **Time** 17:00:12 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 1283438  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 02/25/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/30/2021 14:57

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

**Agency** NICEVILLE PD **Date** 06/28/2021 **Type** Agency Inspection  
**Inspector** CARRICO,KELLY,D **Time** 16:36:39 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 1283438  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 02/25/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 07/01/2021 14:03

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

|                                  |                        |                               |  |
|----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> NICEVILLE PD       | <b>Date</b> 07/02/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> EAKINS,SIERRA,M | <b>Time</b> 12:53:06   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                               |
|------------------------------|------------------------------|------------------------------|-------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 1283438 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 02/25/2023  |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1                 | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                 | 0.075  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1        | 0.077  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1                 | 0.195  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1     | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1          | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Interferent Detect Test # (Repeat) 1 | INT    |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # 2                | 0.000  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # 2                 | 0.047  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                 | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2        | 0.077  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.20 g/210L Test # 2                 | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 2     | 0.080  |
|                                             |        | Interferent Detect Test # 2          | 0.000  |
|                                             |        | Interferent Detect Test # (Repeat) 2 | INT    |
|                                             |        | Alcohol Free Test # 3                | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                 | 0.047  |
|                                             |        | 0.08 g/210L Test # 3                 |        |
|                                             |        | 0.08 g/210L Test # (Repeat) 3        | 0.077  |
|                                             |        | 0.20 g/210L Test # 3                 | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3     | 0.080  |
|                                             |        | Interferent Detect Test # 3          | INT    |
|                                             |        | Interferent Detect Test # (Repeat) 3 | INT    |

**Remarks:**

Time-Date changed. Int Det:NO INT DETECTED. 08: RFI Detect.

**Data Download Date/Time** 07/15/2021 13:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

|                  |                 |             |            |                 |                   |
|------------------|-----------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | NICEVILLE PD    | <b>Date</b> | 08/12/2021 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | EAKINS,SIERRA,M | <b>Time</b> | 15:39:28   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |   |                   |    |
|-----------------------------|---|-------------------|----|
| <b>Number of Simulators</b> | 5 | <b>Compliance</b> | No |
|-----------------------------|---|-------------------|----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | Yes    |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A:IMPROPER READING. Non-compliance:INCORRECT READING.

**Data Download Date/Time** 09/03/2021 09:54

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

|                                  |                        |                               |
|----------------------------------|------------------------|-------------------------------|
| <b>Agency</b> NICEVILLE PD       | <b>Date</b> 08/12/2021 | <b>Type</b> Agency Inspection |
| <b>Inspector</b> EAKINS,SIERRA,M | <b>Time</b> 17:13:48   | <b>Software</b> 8100.27       |

|                              |                              |                   |                       |
|------------------------------|------------------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> | <b>Expiration</b>     |

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1          | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1 | 0.074  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 1   | INT    |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Alcohol Free Test # 2         | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # 2          | 0.047  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.08 g/210L Test # 2          | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # (Repeat) 2 | 0.075  |
| 0.05 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2   | INT    |
| 0.08 g/210L Test: OK                        | No     | Alcohol Free Test # 3         | 0.000  |
| 0.20 g/210L Test: OK                        | No     | 0.05 g/210L Test # 3          | 0.047  |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | 0.08 g/210L Test # 3          | 0.000  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3 | 0.076  |
|                                             |        | Interferent Detect Test # 3   | INT    |

**Remarks:**

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:OUTSIDE TOLERANCE.

**Data Download Date/Time** 09/03/2021 09:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

**Agency** NICEVILLE PD      **Date** 08/13/2021      **Type** Agency Inspection  
**Inspector** EAKINS,SIERRA,M      **Time** 14:08:57      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** No

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | Interferent Detect Test # 1          | 0.000  |
| Minimum Sample Volume Check: OK             |        | Interferent Detect Test # (Repeat) 1 | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 2          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # (Repeat) 2 | 0.000  |
| Interferent Detect Test: Interferent Detect | No     | Alcohol Free Test # 3                | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 3          | 0.000  |
| Alcohol Free Test: OK                       | Yes    | Interferent Detect Test # (Repeat) 3 | 0.000  |
| 0.05 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Test: OK                        | No     |                                      |        |
| 0.20 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                                      |        |

#### Remarks:

Int Det:NO INTERFERANT, . Non-compliance:NO INTERFERANT FOR ACETONE.

**Data Download Date/Time** 09/03/2021 09:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

**Agency** NICEVILLE PD **Date** 08/30/2021 **Type** Agency Inspection  
**Inspector** CARRICO,KELLY,D **Time** 14:03:00 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1283438  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 02/25/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                 | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                 | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                 | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1     | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # (Repeat) 1 | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2                | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2                 | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2                 | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                 | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2     | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # (Repeat) 2 | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3                | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                 | 0.048  |
|                                             |        | 0.08 g/210L Test # 3                 | 0.077  |
|                                             |        | 0.20 g/210L Test # 3                 | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3     | 0.080  |
|                                             |        | Interferent Detect Test # (Repeat) 3 | INT    |

#### Remarks:

Int Det: RFI Detect.

**Data Download Date/Time** 09/03/2021 09:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

**Agency** NICEVILLE PD      **Date** 09/07/2021      **Type** Agency Inspection  
**Inspector** MACBETH,MATTHEW,R      **Time** 15:06:48      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1283438  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 02/25/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 09/09/2021 14:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

**Agency** NICEVILLE PD **Date** 10/19/2021 **Type** Agency Inspection  
**Inspector** EAKINS,SIERRA,M **Time** 14:13:25 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1283438  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 02/25/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 10/22/2021 15:45

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

**Agency** NICEVILLE PD **Date** 11/04/2021 **Type** Agency Inspection  
**Inspector** MACBETH, MATTHEW, R **Time** 15:36:44 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1283438  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 02/25/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 11/05/2021 11:28

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

**Agency** NICEVILLE PD **Date** 12/08/2021 **Type** Agency Inspection  
**Inspector** EAKINS,SIERRA,M **Time** 15:01:44 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A:TEST CONDUCTED OUT OF SEQUENCE . Non-compliance:TESTED OUT OF SEQUENCE.

**Data Download Date/Time** 12/10/2021 11:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

**Agency** NICEVILLE PD **Date** 12/08/2021 **Type** Agency Inspection  
**Inspector** EAKINS,SIERRA,M **Time** 15:35:20 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1283438  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 02/25/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 12/10/2021 11:55

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

**Agency** WALTON COUNTY SO **Date** 01/23/2021 **Type** Agency Inspection  
**Inspector** PULLINGS, ALLEN, C **Time** 10:08:17 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** No

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: No Sample Provided. Non-compliance: .

**Data Download Date/Time** 02/01/2021 13:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

**Agency** WALTON COUNTY SO **Date** 01/24/2021 **Type** Agency Inspection  
**Inspector** GAYDOS, JONATHAN, A **Time** 04:12:29 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 980558  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 07/09/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/01/2021 13:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

**Agency** WALTON COUNTY SO **Date** 02/01/2021 **Type** Agency Inspection  
**Inspector** GAYDOS,JONATHAN,A **Time** 12:04:09 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 980558  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 07/09/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/01/2021 13:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

|                                |                        |                                   |
|--------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> WALTON COUNTY SO | <b>Date</b> 02/02/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,  | <b>Time</b> 14:03:40   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.081  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.081  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.081  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.081  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.081  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.081  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 02/02/2021 15:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

**Agency** WALTON COUNTY SO **Date** 02/02/2021 **Type** Agency Inspection  
**Inspector** GAYDOS,JONATHAN,A **Time** 18:48:22 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 980558  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 07/09/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.083  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.052  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.204  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.052  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/02/2021 19:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

**Agency** WALTON COUNTY SO **Date** 03/15/2021 **Type** Agency Inspection  
**Inspector** GAYDOS, JONATHAN, A **Time** 17:15:47 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 980558  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 07/09/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/17/2021 09:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

**Agency** WALTON COUNTY SO **Date** 04/17/2021 **Type** Agency Inspection  
**Inspector** PULLINGS, ALLEN, C **Time** 08:05:12 **Software** 8100.27

**0.05 Lot #** 20191D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 980558  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 07/09/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.083  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 04/17/2021 09:13

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

**Agency** WALTON COUNTY SO **Date** 05/15/2021 **Type** Agency Inspection  
**Inspector** GAYDOS,JONATHAN,A **Time** 00:30:41 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 980558  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 07/09/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/15/2021 01:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> WALTON COUNTY SO     | <b>Date</b> 06/16/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> LEWIS, TABITHA, J | <b>Time</b> 21:59:53   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                               |
|------------------------------|------------------------------|------------------------------|-------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 1346773 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 07/08/2023  |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.087  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.085  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.078  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.198  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.085  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.084  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.078  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.085  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3               | INT    |

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 06/22/2021 10:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

|                                |                        |                                   |
|--------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> WALTON COUNTY SO | <b>Date</b> 06/22/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,  | <b>Time</b> 12:35:00   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG931603 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 11/12/2021   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.080  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 06/22/2021 13:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

**Agency** WALTON COUNTY SO **Date** 07/15/2021 **Type** Agency Inspection  
**Inspector** LEWIS,TABITHA,J **Time** 21:33:01 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1346773  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 07/08/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

IN SERVICE FROM FDLE

**Data Download Date/Time** 07/19/2021 19:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

**Agency** WALTON COUNTY SO **Date** 08/21/2021 **Type** Agency Inspection  
**Inspector** LEWIS,TABITHA,J **Time** 19:59:56 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1346773  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 07/08/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 08/21/2021 21:11

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

**Agency** WALTON COUNTY SO **Date** 09/19/2021 **Type** Agency Inspection  
**Inspector** LEWIS,TABITHA,J **Time** 19:29:39 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: MESSED UP TIMING WITH THE MOUTH ALCOHOL .

**Data Download Date/Time** 09/19/2021 21:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

**Agency** WALTON COUNTY SO **Date** 09/19/2021 **Type** Agency Inspection  
**Inspector** LEWIS,TABITHA,J **Time** 20:03:46 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1346773  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 07/08/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 09/19/2021 21:10

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

**Agency** WALTON COUNTY SO **Date** 10/26/2021 **Type** Agency Inspection  
**Inspector** PULLINGS, ALLEN, C **Time** 09:45:56 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 20207A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1346773  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 07/08/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 10/26/2021 10:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

**Agency** WALTON COUNTY SO **Date** 11/22/2021 **Type** Agency Inspection  
**Inspector** GAYDOS,JONATHAN,A **Time** 13:37:40 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1346773  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/08/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/22/2021 14:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

**Agency** WALTON COUNTY SO      **Date** 12/20/2021      **Type** Agency Inspection  
**Inspector** GAYDOS,JONATHAN,A      **Time** 13:51:26      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** WO32302427  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 09/11/2024

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 12/20/2021 15:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

**Agency** WALTON COUNTY SO **Date** 01/24/2021 **Type** Agency Inspection  
**Inspector** GAYDOS,JONATHAN,A **Time** 04:25:52 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 181692  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 03/05/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 01/29/2021 09:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

**Agency** WALTON COUNTY SO **Date** 02/02/2021 **Type** Agency Inspection  
**Inspector** GAYDOS,JONATHAN,A **Time** 18:55:02 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 181692  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 03/05/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.051  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.082  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.205  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.204  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/02/2021 20:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

|                                      |                        |                               |  |
|--------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> WALTON COUNTY SO       | <b>Date</b> 03/15/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> GAYDOS, JONATHAN, A | <b>Time</b> 17:16:44   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                               |
|------------------------------|------------------------------|------------------------------|-------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 201908B    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 1012059 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 08/07/2021 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 09/05/2021  |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.079  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.200  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.000  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.048  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.078  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.000  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3               | INT    |

**Remarks:**

OUT OF SERVICE, PENDING FDLE ANNUAL INSPECTION

**Data Download Date/Time** 04/16/2021 09:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

|                                |                        |                                   |
|--------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> WALTON COUNTY SO | <b>Date</b> 04/29/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> SOTO,ISRAEL,  | <b>Time</b> 14:09:27   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG011102 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 04/20/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.080  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.080  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.080  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.080  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 04/29/2021 15:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

**Agency** WALTON COUNTY SO **Date** 05/15/2021 **Type** Agency Inspection  
**Inspector** GAYDOS,JONATHAN,A **Time** 00:19:35 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 1408275  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 11/19/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/15/2021 01:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

**Agency** WALTON COUNTY SO **Date** 06/16/2021 **Type** Agency Inspection  
**Inspector** LEWIS,TABITHA,J **Time** 21:32:47 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 1408275  
**Expiration** 10/22/2021 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 11/19/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 06/16/2021 22:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

**Agency** WALTON COUNTY SO  
**Inspector** LEWIS, TABITHA, J

**Date** 07/15/2021  
**Time** 20:01:42

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1408275  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/19/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 07/15/2021 21:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

**Agency** WALTON COUNTY SO **Date** 08/21/2021 **Type** Agency Inspection  
**Inspector** LEWIS,TABITHA,J **Time** 20:14:26 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1408275  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 11/19/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 08/21/2021 21:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

**Agency** WALTON COUNTY SO **Date** 09/19/2021 **Type** Agency Inspection  
**Inspector** LEWIS,TABITHA,J **Time** 20:12:41 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1408275  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 11/19/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 09/19/2021 21:21

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

**Agency** WALTON COUNTY SO **Date** 10/26/2021 **Type** Agency Inspection  
**Inspector** PULLINGS, ALLEN, C **Time** 13:21:43 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1408275  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 11/19/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 10/26/2021 14:34

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

|                                      |                        |                               |  |
|--------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> WALTON COUNTY SO       | <b>Date</b> 11/22/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> GAYDOS, JONATHAN, A | <b>Time</b> 13:19:00   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                               |
|------------------------------|------------------------------|------------------------------|-------------------------------|
| <b>0.05 Lot #</b> 202101C    | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B    | <b>0.08 Gas Lot #</b> 1408275 |
| <b>Expiration</b> 01/12/2023 | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023 | <b>Expiration</b> 11/19/2023  |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.027  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.048  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.078  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.082  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.031  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.048  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

0.05 TEST - LOOSE HOSE, TIGHTENED SEAL ON SIMULATOR .

**Data Download Date/Time** 11/22/2021 14:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

**Agency** WALTON COUNTY SO **Date** 12/20/2021 **Type** Agency Inspection  
**Inspector** GAYDOS,JONATHAN,A **Time** 13:39:27 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1408275  
**Expiration** 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 11/19/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 12/20/2021 15:20

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

**Agency** OKEECHOBEE CO SO  
**Inspector** WRIGHT, MARK, A

**Date** 01/04/2021  
**Time** 16:30:23

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** 1076482  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 02/20/2021      **Expiration** 01/15/2022

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/05/2021 05:14

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

**Agency** OKEECHOBEE CO SO  
**Inspector** WRIGHT, MARK, A

**Date** 02/19/2021  
**Time** 10:03:26

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** 1076482  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 02/20/2021      **Expiration** 01/15/2022

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 02/19/2021 10:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

**Agency** OKEECHOBEE CO SO  
**Inspector** WRIGHT, MARK, A

**Date** 03/09/2021  
**Time** 06:24:41

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1076482  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 01/15/2022

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 03/09/2021 06:32

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | OKEECHOBEE CO SO | <b>Date</b>       | 04/06/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | GARLAND, MARK, W | <b>Time</b>       | 08:41:34   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201910D          | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 1076482    |
| <b>Expiration</b>           | 10/22/2021       | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 01/15/2022 |
| <b>Number of Simulators</b> | 4                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 04/06/2021 08:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

|                                      |                        |                                   |
|--------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> OKEECHOBEE CO SO       | <b>Date</b> 04/16/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> HAUGHEY, MICHAEL, D | <b>Time</b> 11:43:22   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG026705 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 09/23/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 04/16/2021 12:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

**Agency** OKEECHOBEE CO SO  
**Inspector** GARLAND, MARK, W

**Date** 04/27/2021  
**Time** 13:28:07

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1076482  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 01/15/2022

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.203  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 04/27/2021 13:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | OKEECHOBEE CO SO | <b>Date</b>       | 05/23/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | WRIGHT,MARK,A    | <b>Time</b>       | 08:36:12   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201910D          | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 1474747    |
| <b>Expiration</b>           | 10/22/2021       | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 04/12/2024 |
| <b>Number of Simulators</b> | 4                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.078  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.000  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.000  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 05/25/2021 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

**Agency** OKEECHOBEE CO SO      **Date** 05/24/2021      **Type** Agency Inspection  
**Inspector** GARLAND, MARK, W      **Time** 13:10:51      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1474747  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/12/2024

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/25/2021 09:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

**Agency** OKEECHOBEE CO SO  
**Inspector** MAJERE, ARIC,

**Date** 06/18/2021  
**Time** 04:04:21

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1474747  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/12/2024

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            |        |
| Date and/or Time Adjusted                   | No     | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # 1             | 0.047  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.077  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: RFI Detect.

**Data Download Date/Time** 06/21/2021 11:39

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

**Agency** OKEECHOBEE CO SO  
**Inspector** MAJERE, ARIC,

**Date** 07/02/2021  
**Time** 03:34:32

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1474747  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/12/2024

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 07/02/2021 04:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

**Agency** OKEECHOBEE CO SO  
**Inspector** WRIGHT, MARK, A

**Date** 08/19/2021  
**Time** 21:20:23

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1474747  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/12/2024

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 08/23/2021 10:29

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | OKEECHOBEE CO SO | <b>Date</b>       | 09/12/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | WRIGHT,MARK,A    | <b>Time</b>       | 09:51:52   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 202101C          | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 1474747    |
| <b>Expiration</b>           | 01/12/2023       | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 04/12/2024 |
| <b>Number of Simulators</b> | 4                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/13/2021 14:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

|                                |                        |                               |  |
|--------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> OKEECHOBEE CO SO | <b>Date</b> 10/28/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> MAJERE, ARIC, | <b>Time</b> 17:17:49   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                               |
|------------------------------|------------------------------|------------------------------|-------------------------------|
| <b>0.05 Lot #</b> 202101C    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 1474747 |
| <b>Expiration</b> 01/12/2023 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 04/12/2024  |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 4 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.079  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.201  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.000  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.078  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.000  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3               | INT    |

**Remarks:**

A F / M A: Improper SampleAmbient Fail. 08: Control Outside Tolerance.

**Data Download Date/Time** 11/01/2021 15:40

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

**Agency** OKEECHOBEE CO SO **Date** 11/09/2021 **Type** Agency Inspection  
**Inspector** GARLAND, MARK, W **Time** 11:52:24 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1474747  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 04/12/2024

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 11/09/2021 12:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

**Agency** OKEECHOBEE CO SO **Date** 12/07/2021 **Type** Agency Inspection  
**Inspector** GARLAND, MARK, W **Time** 10:37:23 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1474747  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 04/12/2024

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 12/07/2021 10:40

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

**Agency** OKEECHOBEE COUNTY SO      **Date** 01/04/2021      **Type** Agency Inspection  
**Inspector** WRIGHT, MARK, A      **Time** 15:47:55      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** 1076482  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 02/20/2021      **Expiration** 01/15/2022

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/04/2021 15:51

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

**Agency** OKEECHOBEE COUNTY SO      **Date** 02/19/2021      **Type** Agency Inspection  
**Inspector** WRIGHT, MARK, A      **Time** 10:12:12      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** 1076482  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 02/20/2021      **Expiration** 01/15/2022

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 02/19/2021 10:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

**Agency** OKEECHOBEE COUNTY SO      **Date** 03/09/2021      **Type** Agency Inspection  
**Inspector** WRIGHT, MARK, A      **Time** 06:34:49      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201910B      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1076482  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 05/12/2022      **Expiration** 01/15/2022

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 03/09/2021 11:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

|                                      |                        |                                   |
|--------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> OKEECHOBEE COUNTY SO   | <b>Date</b> 03/17/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> HAUGHEY, MICHAEL, D | <b>Time</b> 12:11:08   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG003005 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 01/30/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.195  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0007 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 03/17/2021 12:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

|                                    |                        |                               |  |
|------------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> OKEECHOBEE COUNTY SO | <b>Date</b> 04/06/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> GARLAND, MARK, W  | <b>Time</b> 08:53:56   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                               |
|------------------------------|------------------------------|------------------------------|-------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 1076482 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 01/15/2022  |

  

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # 1             | 0.047  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.077  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.194  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | INT    |
|                                             |        | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

00: Interferent Detect.

**Data Download Date/Time** 04/06/2021 09:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

**Agency** OKEECHOBEE COUNTY SO      **Date** 05/24/2021      **Type** Agency Inspection  
**Inspector** GARLAND, MARK, W      **Time** 13:05:54      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1076482  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 01/15/2022

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 05/24/2021 13:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

**Agency** OKEECHOBEE COUNTY SO      **Date** 06/18/2021      **Type** Agency Inspection  
**Inspector** MAJERE, ARIC,      **Time** 04:02:14      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1076482  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 01/15/2022

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/21/2021 07:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

**Agency** OKEECHOBEE COUNTY SO      **Date** 07/02/2021      **Type** Agency Inspection  
**Inspector** MAJERE, ARIC,      **Time** 03:31:46      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1474747  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/12/2024

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: Range Exceeded.

**Data Download Date/Time** 07/02/2021 04:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

**Agency** OKEECHOBEE COUNTY SO      **Date** 08/19/2021      **Type** Agency Inspection  
**Inspector** WRIGHT, MARK, A      **Time** 21:17:13      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1474747  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/12/2024

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 08/19/2021 21:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

**Agency** OKEECHOBEE COUNTY SO      **Date** 09/12/2021      **Type** Agency Inspection  
**Inspector** WRIGHT, MARK, A      **Time** 09:47:09      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1474747  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/12/2024

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/13/2021 14:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

**Agency** OKEECHOBEE COUNTY SO **Date** 10/28/2021 **Type** Agency Inspection  
**Inspector** MAJERE, ARIC, **Time** 17:09:14 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1474747  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 04/12/2024

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: Range ExceededRange Exceeded.

**Data Download Date/Time** 11/01/2021 15:08

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

**Agency** OKEECHOBEE COUNTY SO      **Date** 11/09/2021      **Type** Agency Inspection  
**Inspector** GARLAND, MARK, W      **Time** 11:46:53      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1474747  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/12/2024

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 11/09/2021 11:55

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

**Agency** OKEECHOBEE COUNTY SO      **Date** 12/07/2021      **Type** Agency Inspection  
**Inspector** GARLAND, MARK, W      **Time** 10:34:30      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1474747  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/12/2024

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 12/07/2021 10:36

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

**Agency** OKEECHOBEE P.D. **Date** 04/08/2021 **Type** Agency Inspection  
**Inspector** HAUGHEY, MICHAEL, D **Time** 09:54:22 **Software** 8100.27

**0.05 Lot #** **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**  
**Expiration** **Expiration** **Expiration** **Expiration**

**Number of Simulators** **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

COMPLIANCE NOT DETERMINED. AI BY-PASS.

**Data Download Date/Time** 04/08/2021 14:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

|                                      |                        |                                   |
|--------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> OKEECHOBEE P.D.        | <b>Date</b> 04/08/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> HAUGHEY, MICHAEL, D | <b>Time</b> 14:43:33   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG026705 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 09/23/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 04/08/2021 14:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

|                  |                 |             |            |                 |                   |
|------------------|-----------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | OKEECHOBEE P.D. | <b>Date</b> | 05/04/2021 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | MARRERO,RAUL,E  | <b>Time</b> | 17:06:31   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |   |                   |    |
|-----------------------------|---|-------------------|----|
| <b>Number of Simulators</b> | 5 | <b>Compliance</b> | No |
|-----------------------------|---|-------------------|----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

A F / M A: Range Exceeded. Non-compliance: .

**Data Download Date/Time** 05/07/2021 16:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

**Agency** OKEECHOBEE P.D. **Date** 05/04/2021 **Type** Agency Inspection  
**Inspector** MARRERO, RAUL, E **Time** 17:41:57 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 201908B **0.20 Lot #** 202005A **0.08 Gas Lot #** 1355293  
**Expiration** 01/12/2023 **Expiration** 08/07/2021 **Expiration** 05/12/2022 **Expiration** 07/29/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/07/2021 16:32

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

**Agency** PORT ST LUCIE PD  
**Inspector** HARRIS, IAN,

**Date** 01/29/2021  
**Time** 14:00:05

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 201908B      **0.20 Lot #** 201902C      **0.08 Gas Lot #** 24620080A3  
**Expiration** 10/22/2021      **Expiration** 08/07/2021      **Expiration** 02/20/2021      **Expiration** 11/05/2022

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/29/2021 14:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

|                                |                        |                               |  |
|--------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> PORT ST LUCIE PD | <b>Date</b> 02/22/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> HARRIS, IAN,  | <b>Time</b> 13:58:27   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 24620080A3 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 11/05/2022     |

  

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                 | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                 | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                 | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1     | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1          | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # (Repeat) 1 | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                 | 0.047  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                 | 0.077  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                 | 0.197  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2     | 0.080  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2          | INT    |
|                                             |        | Interferent Detect Test # (Repeat) 2 | INT    |
|                                             |        | Alcohol Free Test # 3                | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                 | 0.047  |
|                                             |        | 0.08 g/210L Test # 3                 | 0.077  |
|                                             |        | 0.20 g/210L Test # 3                 | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3     | 0.080  |
|                                             |        | Interferent Detect Test # 3          | INT    |
|                                             |        | Interferent Detect Test # (Repeat) 3 | INT    |

**Remarks:**

Int Det: LOOSE HOSE, REFITTED AND CONTINUED.

**Data Download Date/Time** 02/22/2021 14:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

**Agency** PORT ST LUCIE PD      **Date** 03/31/2021      **Type** Agency Inspection  
**Inspector** HARRIS, IAN,      **Time** 18:19:08      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 4      **Compliance** No

| Check or Test                               | Result | Test                                 | Result |
|---------------------------------------------|--------|--------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                | 0.000  |
| Date and/or Time Adjusted                   | No     | Interferent Detect Test # 1          | 0.000  |
| Minimum Sample Volume Check: OK             |        | Interferent Detect Test # (Repeat) 1 | 0.000  |
| Barometric Pressure Sensor Check: OK        |        | Alcohol Free Test # 2                | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 2          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # (Repeat) 2 | 0.000  |
| Interferent Detect Test: Interferent Detect | No     | Alcohol Free Test # 3                | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | Interferent Detect Test # 3          | 0.000  |
| Alcohol Free Test: OK                       | Yes    | Interferent Detect Test # (Repeat) 3 | 0.000  |
| 0.05 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Test: OK                        | No     |                                      |        |
| 0.20 g/210L Test: OK                        | No     |                                      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |                                      |        |

#### Remarks:

Int Det: REMIXED SOLUTION STILL NOT WORKING. Non-compliance: INTERFERENT NOT DETECTED.

**Data Download Date/Time** 04/16/2021 08:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

|                                      |                        |                                   |
|--------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> PORT ST LUCIE PD       | <b>Date</b> 04/16/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> HAUGHEY, MICHAEL, D | <b>Time</b> 12:08:59   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG026705 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 09/23/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0001      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0000 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 04/16/2021 12:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

**Agency** PORT ST LUCIE PD  
**Inspector** HARRIS, IAN,

**Date** 04/30/2021  
**Time** 11:29:38

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 24620080A3  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2022

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 04/30/2021 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

**Agency** PORT ST LUCIE PD  
**Inspector** HARRIS, IAN,

**Date** 05/29/2021  
**Time** 14:02:08

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 24620080A3  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2022

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 05/29/2021 15:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

|                                |                        |                               |  |
|--------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> PORT ST LUCIE PD | <b>Date</b> 06/25/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> HARRIS, IAN,  | <b>Time</b> 21:42:21   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 24620080A3 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 11/05/2022     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 4 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.043  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.045  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.071  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Test # (Repeat) 1    | 0.075  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.20 g/210L Test # 1             | 0.188  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.194  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Alcohol Free Test: OK                       | Yes    | Interferent Detect Test # 1      | INT    |
| 0.05 g/210L Test: OK                        | Yes    | Alcohol Free Test # 2            | 0.000  |
| 0.08 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # 2             | 0.044  |
| 0.20 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.045  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Test # 2             | 0.078  |
|                                             |        | 0.08 g/210L Test # (Repeat) 2    | 0.076  |
|                                             |        | 0.20 g/210L Test # 2             | 0.189  |
|                                             |        | 0.20 g/210L Test # (Repeat) 2    | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.074  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.075  |
|                                             |        | 0.20 g/210L Test # 3             | 0.191  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.192  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

05: Control Outside Tolerance RESET INLET RETEST. 08: Control CONTROL TESTS OUTSIDE TOLERANCE. RECHECKED IN/OUTLETS RETEST Outside Tolerance RECHECKED TUBES AND RETEST.

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

---

**Data Download Date/Time** 06/25/2021 22:55

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

|                                |                        |                               |
|--------------------------------|------------------------|-------------------------------|
| <b>Agency</b> PORT ST LUCIE PD | <b>Date</b> 07/23/2021 | <b>Type</b> Agency Inspection |
| <b>Inspector</b> HARRIS, IAN,  | <b>Time</b> 16:50:34   | <b>Software</b> 8100.27       |

|                              |                              |                              |                       |
|------------------------------|------------------------------|------------------------------|-----------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b>     |

**Number of Simulators** 4 **Compliance** No

| Check or Test                               | Result | Test                          | Result |
|---------------------------------------------|--------|-------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1         | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1          | 0.045  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1          | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1          | 0.189  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1 | 0.187  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1   | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2         | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.05 g/210L Test # 2          | 0.045  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2          | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2          | 0.188  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2 | 0.198  |
| 0.20 g/210L Test: OK                        | No     | Interferent Detect Test # 2   | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | Alcohol Free Test # 3         | 0.000  |
|                                             |        | 0.05 g/210L Test # 3          | 0.046  |
|                                             |        | 0.08 g/210L Test # 3          | 0.078  |
|                                             |        | 0.20 g/210L Test # 3          | 0.191  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3 | 0.198  |
|                                             |        | Interferent Detect Test # 3   | INT    |

**Remarks:**

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 07/23/2021 18:15

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

|                                |                        |                               |  |
|--------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> PORT ST LUCIE PD | <b>Date</b> 07/23/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> HARRIS, IAN,  | <b>Time</b> 17:37:06   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 24620080A3 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 11/05/2022     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 4 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.044  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # (Repeat) 1    | 0.046  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.076  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.046  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

05: Control Outside Tolerance.

**Data Download Date/Time** 07/23/2021 18:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

**Agency** PORT ST LUCIE PD  
**Inspector** HARRIS, IAN,

**Date** 08/31/2021  
**Time** 14:21:36

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 24620080A3  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2022

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 08/31/2021 17:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

**Agency** PORT ST LUCIE PD  
**Inspector** HARRIS, IAN,

**Date** 09/27/2021  
**Time** 11:25:30

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 24620080A3  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2022

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 09/27/2021 11:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

|                                |                              |                               |                                  |
|--------------------------------|------------------------------|-------------------------------|----------------------------------|
| <b>Agency</b> PORT ST LUCIE PD | <b>Date</b> 10/20/2021       | <b>Type</b> Agency Inspection |                                  |
| <b>Inspector</b> HARRIS, IAN,  | <b>Time</b> 13:49:48         | <b>Software</b> 8100.27       |                                  |
| <b>0.05 Lot #</b> 201910D      | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A     | <b>0.08 Gas Lot #</b> 24620080A3 |
| <b>Expiration</b> 10/22/2021   | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022  | <b>Expiration</b> 11/05/2022     |
| <b>Number of Simulators</b> 4  | <b>Compliance</b> Yes        |                               |                                  |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.078  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.193  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.074  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.190  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.075  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.191  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

08: Control Outside Tolerance, RESEATED TUBES.

**Data Download Date/Time** 10/20/2021 14:13

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

|                                |                        |                               |  |
|--------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> PORT ST LUCIE PD | <b>Date</b> 11/30/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> HARRIS, IAN,  | <b>Time</b> 18:42:01   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 202101C    | <b>0.08 Lot #</b> 202108D    | <b>0.20 Lot #</b> 202106B    | <b>0.08 Gas Lot #</b> 24620080A3 |
| <b>Expiration</b> 01/12/2023 | <b>Expiration</b> 08/17/2023 | <b>Expiration</b> 06/22/2023 | <b>Expiration</b> 11/05/2022     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 4 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

A F / M A: FAILED TO GIVE ENOUGH VOLUME.

**Data Download Date/Time** 11/30/2021 19:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

**Agency** PORT ST LUCIE PD  
**Inspector** HARRIS, IAN,

**Date** 12/15/2021  
**Time** 18:21:57

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202108D      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 24620080A3  
**Expiration** 01/12/2023      **Expiration** 08/17/2023      **Expiration** 06/22/2023      **Expiration** 11/05/2022

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 12/16/2021 09:24

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

**Agency** INDIAN RIVER CO. SO **Date** 01/08/2021 **Type** Agency Inspection  
**Inspector** KANTORSKI, MARK, A **Time** 12:46:14 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1298240  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 03/19/2023

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/08/2021 13:54

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

**Agency** INDIAN RIVER CO. SO **Date** 02/21/2021 **Type** Agency Inspection  
**Inspector** KANTORSKI, MARK, A **Time** 17:59:29 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1298240  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 03/19/2023

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.203  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.203  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: RFI DetectSequence Aborted.

**Data Download Date/Time** 02/21/2021 18:16

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

Agency INDIAN RIVER CO. SO Date 03/30/2021 Type Agency Inspection  
Inspector KANTORSKI, MARK, A Time 14:03:32 Software 8100.27

0.05 Lot # 201910D 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # 1298240  
Expiration 10/22/2021 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 03/19/2023

Number of Simulators 4 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Time-Date changed. RESIDUAL M/A.

Data Download Date/Time 03/30/2021 14:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

**Agency** INDIAN RIVER CO. SO **Date** 04/29/2021 **Type** Agency Inspection  
**Inspector** TANGEL,MATTHEW,A **Time** 10:53:35 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1298240  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 03/19/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/03/2021 09:08

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

Agency INDIAN RIVER CO. SO Date 05/24/2021 Type Agency Inspection  
Inspector TANGEL,MATTHEW,A Time 11:08:32 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202007A 0.20 Lot # 202005A 0.08 Gas Lot # 1298240  
Expiration 01/12/2023 Expiration 07/21/2022 Expiration 05/12/2022 Expiration 03/19/2023

Number of Simulators 4 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: Improper Sample.

Data Download Date/Time 05/24/2021 12:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

**Agency** INDIAN RIVER CO. SO **Date** 06/28/2021 **Type** Agency Inspection  
**Inspector** TANGEL,MATTHEW,A **Time** 12:58:27 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1298240  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 03/19/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/28/2021 13:10

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> INDIAN RIVER CO. SO       | <b>Date</b> 07/13/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 08:36:48   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG026705 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 09/23/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.077  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.080  |
|                                             |        | 0.20 g/210L Test # 5             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.077  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.077  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.077  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.049  |
| 0.08 g/210L Test # 8                   | 0.080  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.077  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.080  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.077  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.080  |
| 0.20 g/210L Test # 10                  | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.077  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0007 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 07/13/2021 08:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

**Agency** INDIAN RIVER CO. SO **Date** 07/16/2021 **Type** Agency Inspection  
**Inspector** TANGEL,MATTHEW,A **Time** 17:04:57 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1474747  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 04/12/2024

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 07/16/2021 17:23

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

**Agency** INDIAN RIVER CO. SO      **Date** 08/31/2021      **Type** Agency Inspection  
**Inspector** TANGEL,MATTHEW,A      **Time** 16:33:10      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1474747  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/12/2024

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 09/04/2021 23:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

**Agency** INDIAN RIVER CO. SO **Date** 09/23/2021 **Type** Agency Inspection  
**Inspector** TANGEL,MATTHEW,A **Time** 10:30:30 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1474747  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 04/12/2024

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 09/23/2021 10:42

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

**Agency** INDIAN RIVER CO. SO **Date** 10/28/2021 **Type** Agency Inspection  
**Inspector** TANGEL,MATTHEW,A **Time** 07:50:17 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 10/28/2021 08:18

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

**Agency** INDIAN RIVER CO. SO **Date** 11/26/2021 **Type** Agency Inspection  
**Inspector** TANGEL,MATTHEW,A **Time** 13:31:18 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202106B **0.08 Gas Lot #** 1474747  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 06/22/2023 **Expiration** 04/12/2024

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | Yes    | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Time-Date changed.

**Data Download Date/Time** 11/26/2021 13:52

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

Agency INDIAN RIVER CO. SO Date 12/22/2021 Type Agency Inspection  
Inspector RUDOLPH,BRITTANY,D Time 09:39:37 Software 8100.27

0.05 Lot # 202101C 0.08 Lot # 202007A 0.20 Lot # 202106B 0.08 Gas Lot # 1474747  
Expiration 01/12/2023 Expiration 07/21/2022 Expiration 06/22/2023 Expiration 04/12/2024

Number of Simulators 4 Compliance Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

Data Download Date/Time 12/22/2021 10:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

**Agency** INDIAN RIVER CO. SO **Date** 01/08/2021 **Type** Agency Inspection  
**Inspector** KANTORSKI, MARK, A **Time** 12:51:27 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1298240  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 03/19/2023

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 01/11/2021 12:50

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

**Agency** INDIAN RIVER CO. SO **Date** 02/21/2021 **Type** Agency Inspection  
**Inspector** KANTORSKI, MARK, A **Time** 18:03:37 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1298240  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2020 **Expiration** 03/19/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

A F / M A: Range ExceededSequence Aborted.

**Data Download Date/Time** 02/21/2021 18:58

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

**Agency** INDIAN RIVER CO. SO **Date** 03/30/2021 **Type** Agency Inspection  
**Inspector** KANTORSKI, MARK, A **Time** 12:58:57 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1298240  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 03/19/2023

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/30/2021 13:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

**Agency** INDIAN RIVER CO. SO **Date** 04/29/2021 **Type** Agency Inspection  
**Inspector** TANGEL,MATTHEW,A **Time** 10:51:51 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1298240  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 03/19/2023

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 05/03/2021 11:02

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

**Agency** INDIAN RIVER CO. SO **Date** 05/24/2021 **Type** Agency Inspection  
**Inspector** TANGEL,MATTHEW,A **Time** 11:02:57 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1474747  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 04/12/2024

**Number of Simulators** 4 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 05/24/2021 12:12

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

**Agency** INDIAN RIVER CO. SO **Date** 06/28/2021 **Type** Agency Inspection  
**Inspector** TANGEL,MATTHEW,A **Time** 13:05:36 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1474747  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 04/12/2024

**Number of Simulators** 4

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 06/28/2021 13:27

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

**Agency** INDIAN RIVER CO. SO      **Date** 07/16/2021      **Type** Agency Inspection  
**Inspector** TANGEL,MATTHEW,A      **Time** 17:02:38      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1474747  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 04/12/2024

**Number of Simulators** 4      **Compliance** No

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.074  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.078  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | No     | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

08: Control Outside Tolerance. Non-compliance: .

**Data Download Date/Time** 07/16/2021 17:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

|                                     |                        |                                   |
|-------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> INDIAN RIVER CO. SO   | <b>Date</b> 10/07/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> WILLIAMS,RICHARD,A | <b>Time</b> 13:23:12   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG113403 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 05/14/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.201  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.080  |
| 0.20 g/210L Test # 6                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.080  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.050  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0004 |

#### Remarks:

Data Download Date/Time 10/07/2021 13:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

**Agency** INDIAN RIVER CO. SO      **Date** 10/28/2021      **Type** Agency Inspection  
**Inspector** TANGEL,MATTHEW,A      **Time** 07:58:54      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 1474747  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 04/12/2024

**Number of Simulators** 4      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 10/28/2021 08:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

**Agency** DESOTO COUNTY SO      **Date** 01/27/2021      **Type** Agency Inspection  
**Inspector** PITTS,JOSHUA,E      **Time** 10:00:58      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 24119080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 01/28/2021 07:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

**Agency** DESOTO COUNTY SO      **Date** 02/24/2021      **Type** Agency Inspection  
**Inspector** PITTS,JOSHUA,E      **Time** 09:07:26      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 24119080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/26/2021 10:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

**Agency** DESOTO COUNTY SO      **Date** 03/29/2021      **Type** Agency Inspection  
**Inspector** PITTS,JOSHUA,E      **Time** 10:48:25      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 24119080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 03/30/2021 07:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

|                                 |                        |                               |  |
|---------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> DESOTO COUNTY SO  | <b>Date</b> 04/21/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> PITTS,JOSHUA,E | <b>Time</b> 09:23:52   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 24119080A1 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 11/05/2021     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 04/22/2021 07:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

**Agency** DESOTO COUNTY SO      **Date** 05/26/2021      **Type** Agency Inspection  
**Inspector** PITTS,JOSHUA,E      **Time** 09:39:35      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 24119080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 05/28/2021 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

**Agency** DESOTO COUNTY SO      **Date** 06/21/2021      **Type** Agency Inspection  
**Inspector** PITTS,JOSHUA,E      **Time** 09:41:32      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 24119080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/22/2021 13:49

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

**Agency** DESOTO COUNTY SO      **Date** 07/27/2021      **Type** Agency Inspection  
**Inspector** PITTS,JOSHUA,E      **Time** 04:55:48      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 24119080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 07/30/2021 12:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

**Agency** DESOTO COUNTY SO      **Date** 08/25/2021      **Type** Agency Inspection  
**Inspector** PITTS,JOSHUA,E      **Time** 13:33:58      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 24119080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.193  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.077  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 08/25/2021 16:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

**Agency** DESOTO COUNTY SO      **Date** 09/21/2021      **Type** Agency Inspection  
**Inspector** PITTS,JOSHUA,E      **Time** 10:02:31      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 24119080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.190  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.192  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.193  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/21/2021 10:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

**Agency** DESOTO COUNTY SO      **Date** 10/27/2021      **Type** Agency Inspection  
**Inspector** PITTS,JOSHUA,E      **Time** 06:27:28      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 19021080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 09/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.192  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 10/31/2021 09:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

**Agency** DESOTO COUNTY SO      **Date** 11/22/2021      **Type** Agency Inspection  
**Inspector** PITTS,JOSHUA,E      **Time** 14:27:30      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 19021080A2  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 09/05/2023

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.075  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.193  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.075  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: RFI Detect RADIO PASSING LOCATION.

**Data Download Date/Time** 11/22/2021 15:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

|                                         |                        |                                   |
|-----------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> DESOTO COUNTY SO          | <b>Date</b> 12/07/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> REYES-RIVERA, DAVID, E | <b>Time</b> 12:04:32   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG115904 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 06/08/2023   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.049  |
|                                             |        | 0.08 g/210L Test # 5             | 0.078  |
|                                             |        | 0.20 g/210L Test # 5             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001340

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.078  |
| 0.20 g/210L Test # 6                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.048  |
| 0.08 g/210L Test # 7                   | 0.078  |
| 0.20 g/210L Test # 7                   | 0.196  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.078  |
| 0.20 g/210L Test # 8                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.078  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.078  |
| 0.20 g/210L Test # 9                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.078  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.078  |
| 0.20 g/210L Test # 10                  | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0000 |

#### Remarks:

Data Download Date/Time 12/07/2021 12:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

|                                 |                        |                               |  |
|---------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> DESOTO COUNTY SO  | <b>Date</b> 01/27/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> PITTS,JOSHUA,E | <b>Time</b> 09:52:06   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 24119080A1 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 11/05/2021     |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.081  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.079  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.198  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.081  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.049  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.079  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          |        |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3               | INT    |

**Remarks:**

08: RFI Detect RADIO MOVED FURTHER AWAY

**Data Download Date/Time** 01/28/2021 07:20

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

**Agency** DESOTO COUNTY SO      **Date** 02/24/2021      **Type** Agency Inspection  
**Inspector** PITTS,JOSHUA,E      **Time** 09:09:09      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 24119080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 02/26/2021 10:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

|                             |                  |                   |            |                   |                   |                       |            |
|-----------------------------|------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | DESOTO COUNTY SO | <b>Date</b>       | 03/29/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | PITTS,JOSHUA,E   | <b>Time</b>       | 10:38:58   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201910D          | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 24119080A1 |
| <b>Expiration</b>           | 10/22/2021       | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 11/05/2021 |
| <b>Number of Simulators</b> | 5                | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 03/30/2021 07:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

**Agency** DESOTO COUNTY SO      **Date** 04/21/2021      **Type** Agency Inspection  
**Inspector** PITTS,JOSHUA,E      **Time** 09:11:07      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 24119080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 04/22/2021 08:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

|                                 |                        |                               |  |
|---------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> DESOTO COUNTY SO  | <b>Date</b> 05/26/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> PITTS,JOSHUA,E | <b>Time</b> 09:28:25   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 24119080A1 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 11/05/2021     |

  

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                                      | Result |
|---------------------------------------------|--------|-------------------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1                     | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1                      | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1                      | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1                      | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1          | 0.000  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1               | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2                     | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2                      | 0.048  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2                      | 0.079  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2                      | 0.197  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2          | 0.000  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # (Repeat) 2 | 0.081  |
|                                             |        | Interferent Detect Test # 2               | INT    |
|                                             |        | Alcohol Free Test # 3                     | 0.000  |
|                                             |        | 0.05 g/210L Test # 3                      | 0.048  |
|                                             |        | 0.08 g/210L Test # 3                      | 0.078  |
|                                             |        | 0.20 g/210L Test # 3                      | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3          | 0.000  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # (Repeat) 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3               | INT    |

**Remarks:**

08: Control Outside Tolerance NOT CONNECTED.

**Data Download Date/Time** 05/28/2021 15:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

**Agency** DESOTO COUNTY SO      **Date** 06/21/2021      **Type** Agency Inspection  
**Inspector** PITTS,JOSHUA,E      **Time** 09:32:56      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 24119080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 11/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 06/22/2021 13:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

|                  |                    |             |            |                 |                   |
|------------------|--------------------|-------------|------------|-----------------|-------------------|
| <b>Agency</b>    | SARASOTA COUNTY SO | <b>Date</b> | 02/09/2021 | <b>Type</b>     | Agency Inspection |
| <b>Inspector</b> | GUTSCHOW,TAYLOR,D  | <b>Time</b> | 09:41:07   | <b>Software</b> | 8100.27           |

|                   |                   |                   |                       |
|-------------------|-------------------|-------------------|-----------------------|
| <b>0.05 Lot #</b> | <b>0.08 Lot #</b> | <b>0.20 Lot #</b> | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b> | <b>Expiration</b>     |

|                             |                   |     |
|-----------------------------|-------------------|-----|
| <b>Number of Simulators</b> | <b>Compliance</b> | Yes |
|-----------------------------|-------------------|-----|

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | No     |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | No     |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

BYPASSED AI. COMPLIANCE NOT DETERMINED.

**Data Download Date/Time** 02/09/2021 12:39

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

|                                      |                        |                                   |
|--------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> SARASOTA COUNTY SO     | <b>Date</b> 02/09/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> HAUGHEY, MICHAEL, D | <b>Time</b> 12:24:07   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG003005 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 01/30/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.048  |
|                                             |        | 0.08 g/210L Test # 4             | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.048  |
|                                             |        | 0.08 g/210L Test # 5             | 0.079  |
|                                             |        | 0.20 g/210L Test # 5             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.078  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.049  |
| 0.08 g/210L Test # 6                   | 0.079  |
| 0.20 g/210L Test # 6                   | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.078  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.049  |
| 0.08 g/210L Test # 7                   | 0.079  |
| 0.20 g/210L Test # 7                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.078  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.048  |
| 0.08 g/210L Test # 8                   | 0.079  |
| 0.20 g/210L Test # 8                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.079  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.049  |
| 0.08 g/210L Test # 9                   | 0.079  |
| 0.20 g/210L Test # 9                   | 0.198  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.049  |
| 0.08 g/210L Test # 10                  | 0.079  |
| 0.20 g/210L Test # 10                  | 0.197  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.078  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Test (Std Dev)             | 0.0004 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

Data Download Date/Time 02/09/2021 12:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

**Agency** SARASOTA COUNTY SO      **Date** 03/16/2021      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 13:10:52      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 14020080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.045  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.077  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 03/16/2021 13:13

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

**Agency** SARASOTA COUNTY SO      **Date** 03/27/2021      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 13:19:57      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 14020080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.194  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.194  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.194  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 03/31/2021 14:49

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

|                   |                    |                   |            |                   |                   |                       |            |
|-------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>     | SARASOTA COUNTY SO | <b>Date</b>       | 04/11/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>  | FELTOVIC,DEP,M     | <b>Time</b>       | 13:56:49   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b> | 201910D            | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 14020080A1 |
| <b>Expiration</b> | 10/22/2021         | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 07/05/2022 |

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 04/14/2021 16:23

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

**Agency** SARASOTA COUNTY SO      **Date** 05/27/2021      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 14:57:05      **Software** 8100.27

**0.05 Lot #** 20191OD      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 14020080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTION AT N JAIL

**Data Download Date/Time** 06/01/2021 06:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | SARASOTA COUNTY SO | <b>Date</b>       | 06/24/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | FELTOVIC,DEP,M     | <b>Time</b>       | 15:05:21   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201910D            | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 14020080A1 |
| <b>Expiration</b>           | 10/22/2021         | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 07/05/2022 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.045  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.076  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.045  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.195  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 06/28/2021 14:46

Printed: 14 January 2022

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

|                                  |                        |                               |  |
|----------------------------------|------------------------|-------------------------------|--|
| <b>Agency</b> SARASOTA COUNTY SO | <b>Date</b> 07/30/2021 | <b>Type</b> Agency Inspection |  |
| <b>Inspector</b> FELTOVIC,DEP,M  | <b>Time</b> 12:34:07   | <b>Software</b> 8100.27       |  |

  

|                              |                              |                              |                                  |
|------------------------------|------------------------------|------------------------------|----------------------------------|
| <b>0.05 Lot #</b> 201910D    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> 14020080A1 |
| <b>Expiration</b> 10/22/2021 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b> 07/05/2022     |

  

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.074  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # (Repeat) 1    | 0.076  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.194  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.076  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.076  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.076  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.076  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

08: Control Outside Tolerance.CHECK SIM CONNECT

**Data Download Date/Time** 07/30/2021 12:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

**Agency** SARASOTA COUNTY SO      **Date** 08/28/2021      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 12:21:26      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 14020080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.046  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 09/06/2021 13:30

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | SARASOTA COUNTY SO | <b>Date</b>       | 09/15/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | FELTOVIC,DEP,M     | <b>Time</b>       | 12:10:11   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201910D            | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 14020080A1 |
| <b>Expiration</b>           | 10/22/2021         | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 07/05/2022 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.197  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 09/15/2021 12:16

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

**Agency** SARASOTA COUNTY SO      **Date** 10/23/2021      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 11:22:33      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 14020080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 07/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.196  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.196  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.197  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 10/27/2021 09:50

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

|                                      |                        |                                   |
|--------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> EUSTIS P.D.            | <b>Date</b> 08/19/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> GUTSCHOW, TAYLOR, D | <b>Time</b> 15:42:45   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG026705 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 09/23/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0002      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.081  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |
|                                             |        | 0.08 g/210L Test # 5             | 0.081  |
|                                             |        | 0.20 g/210L Test # 5             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 5 | 0.079  |
|                                             |        | Interferent Detect Test # 5      | INT    |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

|                                        |        |
|----------------------------------------|--------|
| Alcohol Free Test # 6                  | 0.000  |
| 0.05 g/210L Test # 6                   | 0.050  |
| 0.08 g/210L Test # 6                   | 0.081  |
| 0.20 g/210L Test # 6                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6       | 0.079  |
| Interferent Detect Test # 6            | INT    |
| Alcohol Free Test # 7                  | 0.000  |
| 0.05 g/210L Test # 7                   | 0.050  |
| 0.08 g/210L Test # 7                   | 0.081  |
| 0.20 g/210L Test # 7                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 7       | 0.079  |
| Interferent Detect Test # 7            | INT    |
| Alcohol Free Test # 8                  | 0.000  |
| 0.05 g/210L Test # 8                   | 0.050  |
| 0.08 g/210L Test # 8                   | 0.081  |
| 0.20 g/210L Test # 8                   | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8       | 0.080  |
| Interferent Detect Test # 8            | INT    |
| Alcohol Free Test # 9                  | 0.000  |
| 0.05 g/210L Test # 9                   | 0.050  |
| 0.08 g/210L Test # 9                   | 0.081  |
| 0.20 g/210L Test # 9                   | 0.201  |
| 0.08 g/210L Dry Gas Std Test # 9       | 0.079  |
| Interferent Detect Test # 9            | INT    |
| Alcohol Free Test # 10                 | 0.000  |
| 0.05 g/210L Test # 10                  | 0.051  |
| 0.08 g/210L Test # 10                  | 0.081  |
| 0.20 g/210L Test # 10                  | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 10      | 0.079  |
| Interferent Detect Test # 10           | INT    |
| 0.05 g/210L Test (Std Dev)             | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0005 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

#### Remarks:

Data Download Date/Time 08/19/2021 16:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

**Agency** EUSTIS P.D. **Date** 09/14/2021 **Type** Agency Inspection  
**Inspector** WADE,CODY,L **Time** 10:16:20 **Software** 8100.27

**0.05 Lot #** 201910D **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 1031121  
**Expiration** 10/22/2021 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 10/12/2021

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.201  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

**Data Download Date/Time** 09/14/2021 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

**Agency** EUSTIS P.D.  
**Inspector** WADE,CODY,L

**Date** 10/11/2021  
**Time** 10:42:13

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 1031121  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 10/12/2021

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.202  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.202  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.082  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.082  |
|                                             |        | 0.20 g/210L Test # 3             | 0.202  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.082  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 10/11/2021 13:55

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

**Agency** EUSTIS P.D. **Date** 11/30/2021 **Type** Agency Inspection  
**Inspector** WADE,CODY,L **Time** 10:28:11 **Software** 8100.27

**0.05 Lot #** 202101C **0.08 Lot #** 202007A **0.20 Lot #** 202005A **0.08 Gas Lot #** 211143  
**Expiration** 01/12/2023 **Expiration** 07/21/2022 **Expiration** 05/12/2022 **Expiration** 07/05/2023

**Number of Simulators** 5 **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

Remarks:

**Data Download Date/Time** 12/01/2021 12:45

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

|                              |                        |                               |
|------------------------------|------------------------|-------------------------------|
| <b>Agency</b> EUSTIS P.D.    | <b>Date</b> 12/16/2021 | <b>Type</b> Agency Inspection |
| <b>Inspector</b> WADE,CODY,L | <b>Time</b> 11:34:50   | <b>Software</b> 8100.27       |

|                              |                              |                              |                       |
|------------------------------|------------------------------|------------------------------|-----------------------|
| <b>0.05 Lot #</b> 202101C    | <b>0.08 Lot #</b> 202007A    | <b>0.20 Lot #</b> 202005A    | <b>0.08 Gas Lot #</b> |
| <b>Expiration</b> 01/12/2023 | <b>Expiration</b> 07/21/2022 | <b>Expiration</b> 05/12/2022 | <b>Expiration</b>     |

|                               |                       |
|-------------------------------|-----------------------|
| <b>Number of Simulators</b> 5 | <b>Compliance</b> Yes |
|-------------------------------|-----------------------|

| Check or Test                               | Result | Test                        | Result |
|---------------------------------------------|--------|-----------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1       | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1        | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1        | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1        | 0.195  |
| Alcohol Free Subject Test: 0.000            | Yes    | Interferent Detect Test # 1 | INT    |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Alcohol Free Test # 2       | 0.000  |
| Interferent Detect Test: Interferent Detect | Yes    | 0.05 g/210L Test # 2        | 0.049  |
| Diagnostic Check (Post-Inspection): OK      | No     | 0.08 g/210L Test # 2        | 0.081  |
| Alcohol Free Test: OK                       | Yes    | 0.20 g/210L Test # 2        | 0.195  |
| 0.05 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2 | INT    |
| 0.08 g/210L Test: OK                        | Yes    | Alcohol Free Test # 3       | 0.000  |
| 0.20 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # 3        | 0.049  |
| 0.08 g/210L Dry Gas Std Test: OK            | No     | 0.08 g/210L Test # 3        | 0.080  |
|                                             |        | 0.20 g/210L Test # 3        | 0.196  |
|                                             |        | Interferent Detect Test # 3 | INT    |

Remarks:

**Data Download Date/Time** 12/16/2021 15:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

**Agency** EUSTIS P.D.  
**Inspector** WADE,CODY,L

**Date** 12/16/2021  
**Time** 12:14:36

**Type** Agency Inspection  
**Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 211143  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2023

**Number of Simulators** 5

**Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.081  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.081  |
|                                             |        | Interferent Detect Test # 3      | INT    |

**Remarks:**

**Data Download Date/Time** 12/16/2021 15:13

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

|                                      |                        |                                   |
|--------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> SARASOTA COUNTY SO     | <b>Date</b> 08/11/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> HAUGHEY, MICHAEL, D | <b>Time</b> 13:30:46   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG026705 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 09/23/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             | Yes    | 0.05 g/210L Test # (Repeat) 1    | 0.049  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.08 g/210L Test # 1             | 0.080  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| 0.05 g/210L Test: OK                        | Yes    | 0.05 g/210L Test # (Repeat) 2    | 0.050  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.05 g/210L Test # (Repeat) 3    | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.05 g/210L Test # (Repeat) 4    | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.080  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.078  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

|                                   |       |
|-----------------------------------|-------|
| 0.05 g/210L Test # (Repeat) 5     | 0.050 |
| 0.08 g/210L Test # 5              | 0.080 |
| 0.20 g/210L Test # 5              | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 5  | 0.079 |
| Interferent Detect Test # 5       | INT   |
| Alcohol Free Test # 6             | 0.000 |
| 0.05 g/210L Test # 6              | 0.050 |
| 0.05 g/210L Test # (Repeat) 6     | 0.050 |
| 0.08 g/210L Test # 6              | 0.081 |
| 0.20 g/210L Test # 6              | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 6  | 0.079 |
| Interferent Detect Test # 6       | INT   |
| Alcohol Free Test # 7             | 0.000 |
| 0.05 g/210L Test # 7              | 0.050 |
| 0.05 g/210L Test # (Repeat) 7     | 0.050 |
| 0.08 g/210L Test # 7              | 0.080 |
| 0.20 g/210L Test # 7              | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7  | 0.079 |
| Interferent Detect Test # 7       | INT   |
| Alcohol Free Test # 8             | 0.000 |
| 0.05 g/210L Test # 8              | 0.050 |
| 0.05 g/210L Test # (Repeat) 8     | 0.050 |
| 0.08 g/210L Test # 8              | 0.081 |
| 0.20 g/210L Test # 8              | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8  | 0.079 |
| Interferent Detect Test # 8       | INT   |
| Alcohol Free Test # 9             | 0.000 |
| 0.05 g/210L Test # 9              | 0.000 |
| 0.05 g/210L Test # (Repeat) 9     | 0.050 |
| 0.08 g/210L Test # 9              | 0.081 |
| 0.20 g/210L Test # 9              | 0.200 |
| 0.08 g/210L Dry Gas Std Test # 9  | 0.079 |
| Interferent Detect Test # 9       | INT   |
| Alcohol Free Test # 10            | 0.000 |
| 0.05 g/210L Test # 10             | 0.000 |
| 0.05 g/210L Test # (Repeat) 10    | 0.050 |
| 0.08 g/210L Test # 10             | 0.081 |
| 0.20 g/210L Test # 10             | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10      | INT   |

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

|                                        |        |
|----------------------------------------|--------|
| 0.05 g/210L Test (Std Dev)             | 0.0210 |
| 0.05 g/210L Test (Std Dev) (Repeat)    | 0.0003 |
| 0.08 g/210L Test (Std Dev)             | 0.0005 |
| 0.20 g/210L Test (Std Dev)             | 0.0004 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0003 |

**Remarks:**

05: Control Outside Tolerance. DISCONNECTED SIM. REATTACHED.

---

**Data Download Date/Time** 08/11/2021 13:51

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

**Agency** SARASOTA COUNTY SO      **Date** 08/28/2021      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 12:18:06      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 14020080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.050  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.081  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.078  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.078  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 08/30/2021 21:34

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

**Agency** SARASOTA COUNTY SO      **Date** 09/15/2021      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 11:33:29      **Software** 8100.27

**0.05 Lot #** 20191OD      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 14020080AI  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.080  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.200  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.081  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.050  |
|                                             |        | 0.08 g/210L Test # 3             | 0.081  |
|                                             |        | 0.20 g/210L Test # 3             | 0.201  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 09/15/2021 11:39

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

**Agency** SARASOTA COUNTY SO      **Date** 10/23/2021      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 12:03:02      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 14020080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 07/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 10/27/2021 09:45

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

**Agency** SARASOTA COUNTY SO      **Date** 11/30/2021      **Type** Agency Inspection  
**Inspector** BRECKLE, TIMOTHY, M      **Time** 18:47:10      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 14020080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 07/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.080  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.080  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 11/30/2021 18:56

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

**Agency** SARASOTA COUNTY SO      **Date** 12/09/2021      **Type** Agency Inspection  
**Inspector** SANDERS,DEPUTY,J      **Time** 15:14:25      **Software** 8100.27

**0.05 Lot #** 202101C      **0.08 Lot #** 202007A      **0.20 Lot #** 202106B      **0.08 Gas Lot #** 14020080A1  
**Expiration** 01/12/2023      **Expiration** 07/21/2022      **Expiration** 06/22/2023      **Expiration** 07/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.078  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

PASS AT SCSO JAIL

**Data Download Date/Time** 12/09/2021 15:20

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

|                                      |                        |                                   |
|--------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> SARASOTA COUNTY SO     | <b>Date</b> 01/12/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> HAUGHEY, MICHAEL, D | <b>Time</b> 11:48:34   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG003005 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 01/30/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.08 g/210L Test # (Repeat) 1    | 0.077  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.199  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.049  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # (Repeat) 2    | 0.077  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.08 g/210L Test # (Repeat) 3    | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.049  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.08 g/210L Test # (Repeat) 4    | 0.078  |
|                                             |        | 0.20 g/210L Test # 4             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.079  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.050  |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

|                                     |        |
|-------------------------------------|--------|
| 0.08 g/210L Test # 5                | 0.080  |
| 0.08 g/210L Test # (Repeat) 5       | 0.078  |
| 0.20 g/210L Test # 5                | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 5    | 0.079  |
| Interferent Detect Test # 5         | INT    |
| Alcohol Free Test # 6               | 0.000  |
| 0.05 g/210L Test # 6                | 0.050  |
| 0.08 g/210L Test # 6                | 0.079  |
| 0.08 g/210L Test # (Repeat) 6       | 0.078  |
| 0.20 g/210L Test # 6                | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 6    | 0.079  |
| Interferent Detect Test # 6         | INT    |
| Alcohol Free Test # 7               | 0.000  |
| 0.05 g/210L Test # 7                | 0.049  |
| 0.08 g/210L Test # 7                | 0.080  |
| 0.08 g/210L Test # (Repeat) 7       | 0.078  |
| 0.20 g/210L Test # 7                | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 7    | 0.078  |
| Interferent Detect Test # 7         | INT    |
| Alcohol Free Test # 8               | 0.000  |
| 0.05 g/210L Test # 8                | 0.049  |
| 0.08 g/210L Test # 8                | 0.079  |
| 0.08 g/210L Test # (Repeat) 8       | 0.078  |
| 0.20 g/210L Test # 8                | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 8    | 0.079  |
| Interferent Detect Test # 8         | INT    |
| Alcohol Free Test # 9               | 0.000  |
| 0.05 g/210L Test # 9                | 0.049  |
| 0.08 g/210L Test # (Repeat) 9       | 0.078  |
| 0.20 g/210L Test # 9                | 0.200  |
| 0.08 g/210L Dry Gas Std Test # 9    | 0.079  |
| Interferent Detect Test # 9         | INT    |
| Alcohol Free Test # 10              | 0.000  |
| 0.05 g/210L Test # 10               | 0.049  |
| 0.08 g/210L Test # (Repeat) 10      | 0.077  |
| 0.20 g/210L Test # 10               | 0.199  |
| 0.08 g/210L Dry Gas Std Test # 10   | 0.079  |
| Interferent Detect Test # 10        | INT    |
| 0.05 g/210L Test (Std Dev)          | 0.0004 |
| 0.08 g/210L Test (Std Dev) (Repeat) | 0.0005 |

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

0.20 g/210L Test (Std Dev) 0.0005

0.08 g/210L Dry Gas Std Test (Std Dev) 0.0005

**Remarks:**

08: Ambient Fail. UNK. CLEARED ROOM.

---

**Data Download Date/Time** 01/12/2021 12:11

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

**Agency** SARASOTA COUNTY SO      **Date** 01/29/2021      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 11:34:33      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test | Result |
|---------------------------------------------|--------|------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    |      |        |
| Date and/or Time Adjusted                   | No     |      |        |
| Minimum Sample Volume Check: OK             |        |      |        |
| Barometric Pressure Sensor Check: OK        |        |      |        |
| Alcohol Free Subject Test: 0.000            | No     |      |        |
| Mouth Alcohol Test: Slope Not Met           | Yes    |      |        |
| Interferent Detect Test: Interferent Detect | No     |      |        |
| Diagnostic Check (Post-Inspection): OK      | No     |      |        |
| Alcohol Free Test: OK                       | No     |      |        |
| 0.05 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Test: OK                        | No     |      |        |
| 0.20 g/210L Test: OK                        | No     |      |        |
| 0.08 g/210L Dry Gas Std Test: OK            | No     |      |        |

#### Remarks:

TOO MUCH MOUTH ALCOHOL.

**Data Download Date/Time** 01/29/2021 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

**Agency** SARASOTA COUNTY SO      **Date** 01/29/2021      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 12:17:03      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 14020080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.198  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.049  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.198  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

RESIDUAL MOUTH ALCOHOL.

**Data Download Date/Time** 01/29/2021 12:28

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

**Agency** SARASOTA COUNTY SO      **Date** 02/10/2021      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 14:14:16      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 14020080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.047  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.078  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.199  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.200  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 02/10/2021 14:18

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

**Agency** SARASOTA COUNTY SO      **Date** 03/16/2021      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 13:04:09      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 14020080A1  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 07/05/2022

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.046  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.197  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.079  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.046  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.079  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.047  |
|                                             |        | 0.08 g/210L Test # 3             | 0.077  |
|                                             |        | 0.20 g/210L Test # 3             | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 03/16/2021 13:15

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

**Agency** SARASOTA COUNTY SO      **Date** 03/27/2021      **Type** Agency Inspection  
**Inspector** FELTOVIC,DEP,M      **Time** 13:07:39      **Software** 8100.27

**0.05 Lot #** 201910D      **0.08 Lot #** 202007A      **0.20 Lot #** 202005A      **0.08 Gas Lot #** 17919080A2  
**Expiration** 10/22/2021      **Expiration** 07/21/2022      **Expiration** 05/12/2022      **Expiration** 08/05/2021

**Number of Simulators** 5      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.048  |
| Minimum Sample Volume Check: OK             |        | 0.08 g/210L Test # 1             | 0.077  |
| Barometric Pressure Sensor Check: OK        |        | 0.20 g/210L Test # 1             | 0.198  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | Interferent Detect Test # 1      | INT    |
| Interferent Detect Test: Interferent Detect | Yes    | Alcohol Free Test # 2            | 0.000  |
| Diagnostic Check (Post-Inspection): OK      | Yes    | 0.05 g/210L Test # 2             | 0.048  |
| Alcohol Free Test: OK                       | Yes    | 0.08 g/210L Test # 2             | 0.078  |
| 0.05 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.08 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.081  |
| 0.20 g/210L Test: OK                        | Yes    | Interferent Detect Test # 2      | INT    |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.080  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

INSPECTED AT N JAIL

**Data Download Date/Time** 03/27/2021 13:12

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

|                                      |                        |                                   |
|--------------------------------------|------------------------|-----------------------------------|
| <b>Agency</b> SARASOTA COUNTY SO     | <b>Date</b> 04/08/2021 | <b>Type</b> Department Inspection |
| <b>Inspector</b> HAUGHEY, MICHAEL, D | <b>Time</b> 14:35:34   | <b>Software</b> 8100.27           |

|                              |                              |                              |                                |
|------------------------------|------------------------------|------------------------------|--------------------------------|
| <b>0.05 Lot #</b> 202010A    | <b>0.08 Lot #</b> 202010B    | <b>0.20 Lot #</b> 202010D    | <b>0.08 Gas Lot #</b> AG026705 |
| <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/05/2022 | <b>Expiration</b> 10/06/2022 | <b>Expiration</b> 09/23/2022   |

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | 0.05 g/210L Test # 1             | 0.049  |
| Minimum Sample Volume Check: OK             | Yes    | 0.08 g/210L Test # 1             | 0.079  |
| Barometric Pressure Sensor Check: OK        | Yes    | 0.20 g/210L Test # 1             | 0.000  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # (Repeat) 1    | 0.199  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.080  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # 2            | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.050  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.079  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.199  |
| 0.20 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # (Repeat) 2    | 0.199  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.080  |
|                                             |        | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # 3            | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.051  |
|                                             |        | 0.08 g/210L Test # 3             | 0.079  |
|                                             |        | 0.20 g/210L Test # 3             | 0.200  |
|                                             |        | 0.20 g/210L Test # (Repeat) 3    | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.079  |
|                                             |        | Interferent Detect Test # 3      | INT    |
|                                             |        | Alcohol Free Test # 4            | 0.000  |
|                                             |        | 0.05 g/210L Test # 4             | 0.050  |
|                                             |        | 0.08 g/210L Test # 4             | 0.079  |
|                                             |        | 0.20 g/210L Test # 4             | 0.199  |
|                                             |        | 0.20 g/210L Test # (Repeat) 4    | 0.199  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 4 | 0.080  |
|                                             |        | Interferent Detect Test # 4      | INT    |
|                                             |        | Alcohol Free Test # 5            | 0.000  |
|                                             |        | 0.05 g/210L Test # 5             | 0.051  |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

|                                   |       |
|-----------------------------------|-------|
| 0.08 g/210L Test # 5              | 0.079 |
| 0.20 g/210L Test # 5              | 0.200 |
| 0.20 g/210L Test # (Repeat) 5     | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 5  | 0.079 |
| Interferent Detect Test # 5       | INT   |
| Alcohol Free Test # 6             | 0.000 |
| 0.05 g/210L Test # 6              | 0.051 |
| 0.08 g/210L Test # 6              | 0.079 |
| 0.20 g/210L Test # 6              | 0.201 |
| 0.20 g/210L Test # (Repeat) 6     | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 6  | 0.079 |
| Interferent Detect Test # 6       | INT   |
| Alcohol Free Test # 7             | 0.000 |
| 0.05 g/210L Test # 7              | 0.050 |
| 0.08 g/210L Test # 7              | 0.079 |
| 0.20 g/210L Test # 7              | 0.201 |
| 0.20 g/210L Test # (Repeat) 7     | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 7  | 0.080 |
| Interferent Detect Test # 7       | INT   |
| Alcohol Free Test # 8             | 0.000 |
| 0.05 g/210L Test # 8              | 0.051 |
| 0.08 g/210L Test # 8              | 0.079 |
| 0.20 g/210L Test # 8              | 0.201 |
| 0.20 g/210L Test # (Repeat) 8     | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 8  | 0.080 |
| Interferent Detect Test # 8       | INT   |
| Alcohol Free Test # 9             | 0.000 |
| 0.05 g/210L Test # 9              | 0.051 |
| 0.08 g/210L Test # 9              | 0.079 |
| 0.20 g/210L Test # 9              | 0.201 |
| 0.20 g/210L Test # (Repeat) 9     | 0.198 |
| 0.08 g/210L Dry Gas Std Test # 9  | 0.080 |
| Interferent Detect Test # 9       | INT   |
| Alcohol Free Test # 10            | 0.000 |
| 0.05 g/210L Test # 10             | 0.051 |
| 0.08 g/210L Test # 10             | 0.079 |
| 0.20 g/210L Test # 10             | 0.202 |
| 0.20 g/210L Test # (Repeat) 10    | 0.199 |
| 0.08 g/210L Dry Gas Std Test # 10 | 0.079 |
| Interferent Detect Test # 10      | INT   |

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

|                                        |        |
|----------------------------------------|--------|
| 0.05 g/210L Test (Std Dev)             | 0.0007 |
| 0.08 g/210L Test (Std Dev)             | 0.0000 |
| 0.20 g/210L Test (Std Dev)             | 0.0633 |
| 0.20 g/210L Test (Std Dev) (Repeat)    | 0.0003 |
| 0.08 g/210L Dry Gas Std Test (Std Dev) | 0.0005 |

#### Remarks:

20: Control Outside Tolerance SIM NOT CONNECTED.

---

**Data Download Date/Time** 04/08/2021 14:40

**Printed: 14 January 2022**  
COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

|                             |                    |                   |            |                   |                   |                       |            |
|-----------------------------|--------------------|-------------------|------------|-------------------|-------------------|-----------------------|------------|
| <b>Agency</b>               | SARASOTA COUNTY SO | <b>Date</b>       | 05/27/2021 | <b>Type</b>       | Agency Inspection |                       |            |
| <b>Inspector</b>            | FELTOVIC,DEP,M     | <b>Time</b>       | 13:59:04   | <b>Software</b>   | 8100.27           |                       |            |
| <b>0.05 Lot #</b>           | 201910D            | <b>0.08 Lot #</b> | 202007A    | <b>0.20 Lot #</b> | 202005A           | <b>0.08 Gas Lot #</b> | 14020080A1 |
| <b>Expiration</b>           | 10/22/2021         | <b>Expiration</b> | 07/21/2022 | <b>Expiration</b> | 05/12/2022        | <b>Expiration</b>     | 07/05/2022 |
| <b>Number of Simulators</b> | 5                  | <b>Compliance</b> | Yes        |                   |                   |                       |            |

| Check or Test                               | Result | Test                             | Result |
|---------------------------------------------|--------|----------------------------------|--------|
| Diagnostic Check (Pre-Inspection): OK       | Yes    | Alcohol Free Test # 1            | 0.000  |
| Date and/or Time Adjusted                   | No     | Alcohol Free Test # (Repeat) 1   | 0.000  |
| Minimum Sample Volume Check: OK             |        | 0.05 g/210L Test # 1             | 0.046  |
| Barometric Pressure Sensor Check: OK        |        | 0.08 g/210L Test # 1             | 0.077  |
| Alcohol Free Subject Test: 0.000            | Yes    | 0.20 g/210L Test # 1             | 0.195  |
| Mouth Alcohol Test: Slope Not Met           | Yes    | 0.08 g/210L Dry Gas Std Test # 1 | 0.084  |
| Interferent Detect Test: Interferent Detect | Yes    | Interferent Detect Test # 1      | INT    |
| Diagnostic Check (Post-Inspection): OK      | Yes    | Alcohol Free Test # (Repeat) 2   | 0.000  |
| Alcohol Free Test: OK                       | Yes    | 0.05 g/210L Test # 2             | 0.047  |
| 0.05 g/210L Test: OK                        | Yes    | 0.08 g/210L Test # 2             | 0.077  |
| 0.08 g/210L Test: OK                        | Yes    | 0.20 g/210L Test # 2             | 0.195  |
| 0.20 g/210L Test: OK                        | Yes    | 0.08 g/210L Dry Gas Std Test # 2 | 0.084  |
| 0.08 g/210L Dry Gas Std Test: OK            | Yes    | Interferent Detect Test # 2      | INT    |
|                                             |        | Alcohol Free Test # (Repeat) 3   | 0.000  |
|                                             |        | 0.05 g/210L Test # 3             | 0.048  |
|                                             |        | 0.08 g/210L Test # 3             | 0.078  |
|                                             |        | 0.20 g/210L Test # 3             | 0.196  |
|                                             |        | 0.08 g/210L Dry Gas Std Test # 3 | 0.083  |
|                                             |        | Interferent Detect Test # 3      | INT    |

#### Remarks:

00: Ambient Fail CLEAR ROOM.

**Data Download Date/Time** 05/27/2021 14:03

**Printed: 14 January 2022**

COBRA ©1997-2014 CMI, Inc.