Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

Agency FFWCC **Date** 01/12/2023 **Type** Agency Inspection
Inspector MARROQUIN,RICHARD,J **Time** 11:53:02 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** AG128602
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 10/13/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.192
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.193
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/12/2023 12:35

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

Agency FFWCC **Date** 01/18/2023 **Type** Department Inspection
Inspector NICODEMO,PHIL, **Time** 13:39:54 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.077
0.20 g/210L Test # 4	0.195
0.08 g/210L Dry Gas Std Test # 4	0.077
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.077
0.20 g/210L Test # 5	0.195
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.077
0.20 g/210L Test # 6	0.195
0.08 g/210L Dry Gas Std Test # 6	0.076
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.077
0.20 g/210L Test # 7	0.195
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.077
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.077
0.20 g/210L Test # 9	0.195
0.08 g/210L Dry Gas Std Test # 9	0.076
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.077
0.20 g/210L Test # 10	0.195
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Data Download Date/Time 01/18/2023 13:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

Agency FFWCC **Date** 03/06/2023 **Type** Agency Inspection
Inspector MARROQUIN,RICHARD,J **Time** 10:35:52 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** AG128602
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 10/13/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 03/06/2023 10:55

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

Agency FFWCC **Date** 04/28/2023 **Type** Agency Inspection
Inspector MARROQUIN,RICHARD,J **Time** 11:18:22 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** AG128602
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 10/13/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.192
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 04/28/2023 11:47

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

Agency FFWCC **Date** 05/16/2023 **Type** Agency Inspection
Inspector MARROQUIN,RICHARD,J **Time** 12:19:21 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** AG128602
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 10/13/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/16/2023 12:47

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

Agency FFWCC **Date** 06/20/2023 **Type** Agency Inspection
Inspector MARROQUIN,RICHARD,J **Time** 13:06:55 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG128602
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/13/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/20/2023 13:38

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

Agency	FFWCC	**Date**	07/24/2023	**Type**	Agency Inspection
Inspector	MARROQUIN,RICHARD,J	**Time**	15:04:29	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	AG128602
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	10/13/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/24/2023 15:40

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

Agency FFWCC **Date** 08/28/2023 **Type** Agency Inspection
Inspector MARROQUIN,RICHARD,J **Time** 13:32:06 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG128602
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/13/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/15/2023 14:42

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003195

Agency	FFWCC	**Date**	09/15/2023	**Type**	Agency Inspection
Inspector	MARROQUIN,RICHARD,J	**Time**	16:05:45	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	AG310302
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	04/13/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/15/2023 16:16

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

Agency	FFWCC	**Date**	01/14/2023	**Type**	Agency Inspection
Inspector	LAWRENCE,JOSHUA,S	**Time**	21:25:55	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202108D	**0.20 Lot #**	202106B	**0.08 Gas Lot #**	AG129201
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/22/2023	**Expiration**	10/19/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/19/2023 11:17

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

Agency FFWCC **Date** 02/09/2023 **Type** Department Inspection
Inspector NICODEMO,PHIL, **Time** 11:09:15 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.077
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.077
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.077
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.077
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.077
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 02/09/2023 11:16

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

Agency	FFWCC	**Date**	03/04/2023	**Type**	Agency Inspection
Inspector	LAWRENCE,JOSHUA,S	**Time**	14:52:45	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	AG129201
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	10/19/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/04/2023 18:52

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

Agency FFWCC **Date** 04/14/2023 **Type** Agency Inspection
Inspector LAWRENCE,JOSHUA,S **Time** 14:05:26 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG129201
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/19/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/14/2023 15:47

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

Agency FFWCC **Date** 05/22/2023 **Type** Agency Inspection
Inspector LAWRENCE,JOSHUA,S **Time** 18:14:38 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG310302
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/13/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/22/2023 19:16

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

Agency	FFWCC	**Date**	06/20/2023	**Type**	Agency Inspection
Inspector	LAWRENE,JOSHUA,S	**Time**	11:15:39	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	AG310302
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	04/13/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/20/2023 11:56

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

Agency FFWCC **Date** 07/24/2023 **Type** Agency Inspection
Inspector LAWRENCE,JOSHUA,S **Time** 11:28:06 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG310302
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/13/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # (Repeat) 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # (Repeat) 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # (Repeat) 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

20: RFI Detect.RADIO REMOVED FROM OFFICE

Data Download Date/Time 07/24/2023 13:58

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

Agency	FFWCC	**Date**	08/18/2023	**Type**	Agency Inspection
Inspector	LAWRENCE,JOSHUA,S	**Time**	23:02:32	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	AG310302
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	04/13/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/21/2023 16:13

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

Agency FFWCC **Date** 09/19/2023 **Type** Agency Inspection
Inspector LAWRENCE,JOSHUA,S **Time** 15:34:03 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG310302
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/13/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.191
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.193
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

A F / M A:FAILED DUE TO TOO LITTLE MA. CORRECTED/PASSED

Data Download Date/Time 09/20/2023 10:46

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

Agency	FFWCC	**Date**	10/29/2023	**Type**	Agency Inspection
Inspector	LAWRENCE,JOSHUA,S	**Time**	20:16:27	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	AG310302
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	04/13/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

PASSED

Data Download Date/Time 10/30/2023 16:58

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

Agency	FFWCC	**Date**	11/10/2023	**Type**	Agency Inspection
Inspector	LAWRNCE,JOSHUA,S	**Time**	22:14:54	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	AG310302
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	04/13/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/11/2023 15:02

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

Agency	FFWCC	**Date**	12/29/2023	**Type**	Agency Inspection
Inspector	LAWRENCE,JOSHUA,S	**Time**	15:10:48	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	AG310302
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	04/13/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/29/2023 16:50

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

Agency	FFWCC	**Date**	01/23/2023	**Type**	Agency Inspection
Inspector	BASFORD,NICOLE,M	**Time**	17:48:57	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202108D	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	1349789
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/15/2024	**Expiration**	07/15/2023

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 02/03/2023 16:14

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

Agency	FFWCC	**Date**	07/14/2023	**Type**	Agency Inspection
Inspector	SOTO,ISRAEL,	**Time**	07:17:10	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Data Download Date/Time 07/14/2023 10:38

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

Agency FFWCC **Date** 07/14/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 09:32:14 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** 06723080A5
Expiration 01/11/2024 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 04/05/2025

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 07/14/2023 10:38

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

Agency FFWCC **Date** 08/15/2023 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 16:50:59 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402635303
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/25/2026

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/15/2023 17:58

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

Agency FFWCC **Date** 09/29/2023 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 19:23:09 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402635303
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/25/2026

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/29/2023 20:37

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

Agency	FFWCC	**Date**	10/02/2023	**Type**	Agency Inspection
Inspector	BASFORD,NICOLE,M	**Time**	17:39:04	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	402635303
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	01/25/2026

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/02/2023 18:46

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

Agency FFWCC **Date** 11/26/2023 **Type** Agency Inspection
Inspector BASFORD,NICOLE,M **Time** 13:22:11 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402635303
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/25/2026

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/26/2023 14:30

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

Agency	FFWCC	**Date**	12/22/2023	**Type**	Agency Inspection
Inspector	BASFORD,NICOLE,M	**Time**	16:25:12	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	402635303
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	01/25/2026

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 12/22/2023 17:33

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

Agency MIAMI BEACH PD **Date** 01/12/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 11:21:08 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1355293
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/29/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

AI PRICE #958 PRESENT

Data Download Date/Time 01/12/2023 14:24

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

Agency	MIAMI BEACH PD	**Date**	02/23/2023	**Type**	Agency Inspection
Inspector	SOCARRAS,PEDRO,A	**Time**	10:08:35	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202108D	**0.20 Lot #**	202106B	**0.08 Gas Lot #**	1355293
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/22/2023	**Expiration**	07/29/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI MILLAN #141 PRESENT

Data Download Date/Time 02/23/2023 14:00

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

Agency MIAMI BEACH PD **Date** 03/30/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 11:34:24 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1355293
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/29/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI MILLAN #141 PRESENT

Data Download Date/Time 03/30/2023 13:40

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

Agency	MIAMI BEACH PD	**Date**	04/11/2023	**Type**	Agency Inspection
Inspector	SOCARRAS,PEDRO,A	**Time**	09:38:16	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	1355293
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	07/29/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

AI MILLAN #141 PRESENT

Data Download Date/Time 04/11/2023 13:36

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

Agency	MIAMI BEACH PD	**Date**	05/25/2023	**Type**	Agency Inspection
Inspector	SOCARRAS,PEDRO,A	**Time**	10:06:25	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	1355293
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	07/29/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI MILLAN #141 PRESENT

Data Download Date/Time 05/25/2023 10:43

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

Agency MIAMI BEACH PD **Date** 06/15/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 09:25:25 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1355293
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/29/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

AI MILLAN #141 PRESENT

Data Download Date/Time 06/15/2023 10:49

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

Agency MIAMI BEACH PD **Date** 07/20/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 11:20:53 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

AI PRICE #958 PRESENT

Data Download Date/Time 07/20/2023 14:03

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

Agency MIAMI BEACH PD **Date** 08/24/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 10:36:59 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/24/2023 11:18

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

Agency MIAMI BEACH PD **Date** 09/21/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 10:22:56 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

AI PRICE #958 PRESENT

Data Download Date/Time 09/21/2023 12:21

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

Agency MIAMI BEACH PD **Date** 10/19/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 10:57:31 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

AI MILLAN #141 PRESENT

Data Download Date/Time 10/19/2023 13:38

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

Agency MIAMI BEACH PD **Date** 10/25/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:11:39 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.195
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.196
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003224

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.196
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.196
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 10/25/2023 13:20

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO **Date** 01/16/2023 **Type** Agency Inspection
Inspector NIXON,BARRY, **Time** 10:09:41 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: No Sample Provided.

Data Download Date/Time 01/16/2023 10:22

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO **Date** 02/07/2023 **Type** Agency Inspection
Inspector NIXON,BARRY, **Time** 03:40:10 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.045
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.045
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/07/2023 13:16

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO **Date** 02/07/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 12:57:29 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.081
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.081
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/07/2023 13:16

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO **Date** 02/07/2023 **Type** Agency Inspection
Inspector NIXON,BARRY, **Time** 18:54:39 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/07/2023 19:02

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO **Date** 03/27/2023 **Type** Agency Inspection
Inspector NIXON,BARRY, **Time** 13:17:29 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/27/2023 13:28

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO **Date** 04/29/2023 **Type** Agency Inspection
Inspector NIXON,BARRY, **Time** 16:00:12 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/08/2023 10:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO **Date** 05/08/2023 **Type** Agency Inspection
Inspector NIXON,BARRY, **Time** 10:29:54 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 11821080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/08/2023 10:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO **Date** 06/30/2023 **Type** Agency Inspection
Inspector NIXON,BARRY, **Time** 18:34:57 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/30/2023 18:50

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO **Date** 07/11/2023 **Type** Agency Inspection
Inspector NIXON,BARRY, **Time** 15:26:27 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/11/2023 15:41

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO **Date** 08/28/2023 **Type** Agency Inspection
Inspector NIXON,BARRY, **Time** 14:02:04 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/28/2023 14:09

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO **Date** 09/16/2023 **Type** Agency Inspection
Inspector NIXON,BARRY, **Time** 10:53:10 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/16/2023 11:01

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO **Date** 10/13/2023 **Type** Agency Inspection
Inspector NIXON,BARRY, **Time** 16:18:41 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/13/2023 16:27

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO **Date** 11/27/2023 **Type** Agency Inspection
Inspector NIXON,BARRY, **Time** 11:32:55 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #**
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.157
0.20 g/210L Test # (Repeat) 1	0.154
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.157
0.20 g/210L Test # (Repeat) 2	0.155
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.173
0.20 g/210L Test # (Repeat) 3	0.172
Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:TAKEN OUT OF SERVICE.

Data Download Date/Time 11/27/2023 11:42

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO **Date** 12/06/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 14:47:11 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration 03/29/2025 **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Average Standard Deviation** **Compliance** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	No
0.08 g/210L Test: OK	No
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test # 1	0.000
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.050
0.05 g/210L Test # (Repeat) 1	0.049
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	INT
0.05 g/210L Test # (Repeat) 2	0.050
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # (Repeat) 3	0.049
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
Alcohol Free Test # (Repeat) 4	0.000
0.05 g/210L Test # (Repeat) 4	0.049
Interferent Detect Test # 4	INT
Alcohol Free Test # (Repeat) 5	0.000
0.05 g/210L Test # (Repeat) 5	0.049
Interferent Detect Test # 5	INT
Alcohol Free Test # (Repeat) 6	0.000
0.05 g/210L Test # (Repeat) 6	0.049
Interferent Detect Test # 6	INT
Alcohol Free Test # (Repeat) 7	0.000
0.05 g/210L Test # (Repeat) 7	0.049
Interferent Detect Test # 7	INT
Alcohol Free Test # (Repeat) 8	0.000
0.05 g/210L Test # (Repeat) 8	INT
Interferent Detect Test # 8	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Alcohol Free Test # (Repeat) 9	0.000
Interferent Detect Test # 9	INT
Alcohol Free Test # (Repeat) 10	0.000
Interferent Detect Test # 10	INT

Remarks:

00: Ambient Fail CLEARED AREA AROUND INSTRUMENT. 05: Interferent Detect ENSURED SIM HAD NO LEAKS, Interferen. Non -compliance: WILL REPEAT OPT BENCH CAL ADJ.

Data Download Date/Time 12/06/2023 14:56

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Agency PASCO COUNTY SO **Date** 12/11/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 11:33:58 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG229803
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.196
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003347

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.196
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 12/11/2023 11:42

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI **Date** 01/31/2023 **Type** Agency Inspection
Inspector LORENZO,YOSNIEL, **Time** 22:14:33 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402310313
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/15/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

.

Data Download Date/Time 02/01/2023 18:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI **Date** 02/01/2023 **Type** Agency Inspection
Inspector LORENZO,YOSNIEL, **Time** 01:28:18 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402310313
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/15/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/01/2023 18:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI **Date** 03/31/2023 **Type** Agency Inspection
Inspector LORENZO,YOSNIEL, **Time** 22:37:04 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402310313
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/15/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # (Repeat) 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # (Repeat) 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # (Repeat) 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

20: RFI Detect TROOPER NEAR ON PHONE.

Data Download Date/Time 03/31/2023 22:45

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI **Date** 04/01/2023 **Type** Agency Inspection
Inspector LORENZO,YOSNIEL, **Time** 01:22:18 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202112A **0.20 Lot #** 202106B **0.08 Gas Lot #** 402310313
Expiration 12/07/2023 **Expiration** 12/07/2023 **Expiration** 06/22/2023 **Expiration** 01/15/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	No
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

A F / M A: Range ExceededTOO MUCH MOUTH ALCOHOL. 00: RFI Detect NEARBY PHONE .

Data Download Date/Time 04/01/2023 01:29

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI **Date** 04/01/2023 **Type** Agency Inspection
Inspector LORENZO,YOSNIEL, **Time** 02:14:51 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402310313
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/15/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

.

Data Download Date/Time 04/01/2023 02:21

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI **Date** 05/31/2023 **Type** Agency Inspection
Inspector LORENZO,YOSNIEL, **Time** 23:25:18 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 402310313
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 01/15/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: RFI Detect.RADIO ON NEAR

Data Download Date/Time 05/31/2023 23:43

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI **Date** 07/01/2023 **Type** Agency Inspection
Inspector LORENZO,YOSNIEL, **Time** 00:10:25 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402310313
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/15/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # (Repeat) 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # (Repeat) 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # (Repeat) 3	INT

Remarks:

A F / M A: RFI DetectRFI Detect. Int Det: RFI Detect RADIO NEAR BY .

Data Download Date/Time 07/01/2023 00:17

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI **Date** 08/01/2023 **Type** Agency Inspection
Inspector LORENZO,YOSNIEL, **Time** 00:08:58 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402310313
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 01/15/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

A F / M A: RFI Detect.PHONE NEAR CLEARED

Data Download Date/Time 08/01/2023 00:17

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI **Date** 08/31/2023 **Type** Agency Inspection
Inspector LORENZO,YOSNIEL, **Time** 23:30:57 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402310313
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/15/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # (Repeat) 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # (Repeat) 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # (Repeat) 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

05: RFI Detect TROOPER PHONE.

Data Download Date/Time 08/31/2023 23:40

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FHP TROOP E MIAMI **Date** 09/30/2023 **Type** Agency Inspection
Inspector LORENZO,YOSNIEL, **Time** 19:41:36 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402310313
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/15/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

A F / M A: RFI DetectRFI Detect.TROOPER RADIO NEAR BY

Data Download Date/Time 09/30/2023 19:53

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FL HIGHWAY PATROL **Date** 10/04/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 12:59:09 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # (Repeat) 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # (Repeat) 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # (Repeat) 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # (Repeat) 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # (Repeat) 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # (Repeat) 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # (Repeat) 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # (Repeat) 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # (Repeat) 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # (Repeat) 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev) (Repeat)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

20: RFI Detect. CELL PHONE BROUGHT TOO CLOSE..

Data Download Date/Time 10/04/2023 13:08

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FL HIGHWAY PATROL **Date** 10/31/2023 **Type** Agency Inspection
Inspector LORENZO,YOSNIEL, **Time** 14:57:50 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402310313
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/15/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/31/2023 15:07

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FL HIGHWAY PATROL **Date** 11/29/2023 **Type** Agency Inspection
Inspector BRANDT,FABIAN,E **Time** 19:49:03 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402310313
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/15/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

TROOPERS RADIO.

Data Download Date/Time 11/29/2023 20:10

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

Agency FL HIGHWAY PATROL **Date** 12/27/2023 **Type** Agency Inspection
Inspector BRANDT,FABIAN,E **Time** 17:52:01 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402310313
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/15/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Interferent Detect Test # (Repeat) 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # (Repeat) 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # (Repeat) 3	INT

Remarks:

POLICE RADIO.

Data Download Date/Time 12/27/2023 18:02

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency HILLSBOROUGH CO SO **Date** 08/14/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 12:55:17 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** 06723080A5
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 04/05/2025

Number of Simulators 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.081
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.082
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.082
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.052
0.08 g/210L Test # 10	0.082
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0014
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 08/14/2023 13:02

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency	HILLSBOROUGH CO SO	**Date**	09/05/2023	**Type**	Agency Inspection
Inspector	KNOX,STEPHANIE,R	**Time**	09:22:34	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	08223080A4
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT FOR INSPECTION.

Data Download Date/Time 09/05/2023 09:30

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency HILLSBOROUGH CO SO **Date** 10/02/2023 **Type** Agency Inspection
Inspector KNOX,STEPHANIE,R **Time** 13:07:52 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT FOR INSPECTION.

Data Download Date/Time 10/02/2023 16:22

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency	HILLSBOROUGH CO SO	**Date**	11/08/2023	**Type**	Agency Inspection
Inspector	KNOX,STEPHANIE,R	**Time**	10:45:08	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	08223080A4
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

AI B. KROME #263389 PRESENT FOR INSPECTION.

Data Download Date/Time 11/08/2023 11:08

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

Agency HILLSBOROUGH CO SO **Date** 12/01/2023 **Type** Agency Inspection
Inspector KNOX,STEPHANIE,R **Time** 11:43:56 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 03/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

AI B. KROME #263389 PRESENT FOR INSPECTION.

Data Download Date/Time 12/01/2023 11:57

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency HILLSBOROUGH CO SO **Date** 08/14/2023 **Type** Department Inspection
Inspector SIDDOWAY,BENJAMIN,W **Time** 14:15:46 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** 06723080A5
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 04/05/2025

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.082
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.082
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.082
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.082
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.082
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.082
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 08/14/2023 14:23

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency	HILLSBOROUGH CO SO	**Date**	09/22/2023	**Type**	Agency Inspection
Inspector	KNOX,STEPHANIE,R	**Time**	08:35:19	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	08223080A4
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT FOR INSPECTION.

Data Download Date/Time 09/22/2023 08:45

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency	HILLSBOROUGH CO SO	**Date**	10/02/2023	**Type**	Agency Inspection
Inspector	KNOX,STEPHANIE,R	**Time**	11:03:31	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	08223080A4
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT FOR INSPECTION.

Data Download Date/Time 10/02/2023 11:35

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency HILLSBOROUGH CO SO **Date** 11/08/2023 **Type** Agency Inspection
Inspector KNOX,STEPHANIE,R **Time** 13:50:51 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.075
0.20 g/210L Test # 1	0.191
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.075
0.20 g/210L Test # 2	0.192
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.192
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

AI B. KROME #263389 PRESENT FOR INSPECTION.

Data Download Date/Time 11/08/2023 14:18

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

Agency	HILLSBOROUGH CO SO	**Date**	12/01/2023	**Type**	Agency Inspection
Inspector	KNOX,STEPHANIE,R	**Time**	14:15:32	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202307A	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	08223080A4
Expiration	03/28/2025	**Expiration**	07/11/2025	**Expiration**	06/15/2024	**Expiration**	06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.190
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.190
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.075
		0.20 g/210L Test # 3	0.190
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: No Sample Provided.NO AF IN TIME LIMIT.

Data Download Date/Time 12/01/2023 14:26

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003404

Agency	FL HIGHWAY PATROL	**Date**	12/19/2023	**Type**	Agency Inspection
Inspector	GUTSCHOW,TAYLOR,D	**Time**	09:12:58	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Data Download Date/Time 12/19/2023 13:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003404

Agency FL HIGHWAY PATROL **Date** 12/19/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:23:36 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** 01923080A3
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 02/05/2025

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.082
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.082
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003404

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.082
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.082
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.082
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.082
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.082
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 12/19/2023 13:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

Agency FHP TROOP K **Date** 01/16/2023 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 16:58:23 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1346146
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 07/07/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/16/2023 17:06

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

Agency FHP TROOP K **Date** 03/09/2023 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 17:47:58 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1346146
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 07/07/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/09/2023 17:54

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

Agency	FHP TROOP K	**Date**	04/06/2023	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	21:53:37	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202108D	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	1346146
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/15/2024	**Expiration**	07/07/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/07/2023 13:54

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

Agency FHP TROOP K **Date** 05/05/2023 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 02:16:43 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1346146
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/07/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/05/2023 02:23

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

Agency	FHP TROOP K	**Date**	07/26/2023	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	23:53:23	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	15922080A5
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/27/2023 00:05

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

Agency FHP TROOP K **Date** 08/08/2023 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 23:06:36 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A5
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/08/2023 23:14

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

Agency FHP TROOP K **Date** 09/08/2023 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 21:57:29 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A5
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

A F / M A:NOT ENOUGH MOUTH ALCOHOL.

Data Download Date/Time 09/08/2023 22:04

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

Agency	FHP TROOP K	**Date**	10/19/2023	**Type**	Agency Inspection
Inspector	THOMAS,CHRISTOPHER,L	**Time**	22:44:39	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	1432080A2
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/19/2023 22:52

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

Agency FHP TROOP K **Date** 11/16/2023 **Type** Agency Inspection
Inspector THOMAS,CHRISTOPHER,L **Time** 22:44:06 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 14323080A2
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/16/2023 22:53

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

Agency FHP TROOP K **Date** 12/07/2023 **Type** Department Inspection
Inspector SIDDOWAY,BENJAMIN,W **Time** 13:37:14 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG325603
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 09/13/2025

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.052
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.081
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003405

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

A F / M A: No Sample Provided.

Data Download Date/Time 12/07/2023 13:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL **Date** 03/06/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 09:36:52 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # (Repeat) 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # (Repeat) 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # (Repeat) 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # (Repeat) 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # (Repeat) 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # (Repeat) 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # (Repeat) 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # (Repeat) 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # (Repeat) 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # (Repeat) 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Int Det: Improper Sample HAD HAND OVER BREATH TUBE.

Data Download Date/Time 03/06/2023 09:37

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency	FL HIGHWAY PATROL	**Date**	04/12/2023	**Type**	Agency Inspection
Inspector	HENING,FRED,C	**Time**	11:32:16	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	33922080A1
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	01/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 04/12/2023 11:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL **Date** 06/07/2023 **Type** Agency Inspection
Inspector HENING,FRED,C **Time** 19:09:07 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 06/08/2023 12:04

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL **Date** 07/10/2023 **Type** Agency Inspection
Inspector HENING,FRED,C **Time** 12:00:35 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

00: RFI Detect PHONE TOO CLOSE.

Data Download Date/Time 07/10/2023 12:08

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL **Date** 08/29/2023 **Type** Agency Inspection
Inspector HENING,FRED,C **Time** 21:40:03 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 08/29/2023 21:50

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency	FL HIGHWAY PATROL	**Date**	09/21/2023	**Type**	Agency Inspection
Inspector	HENING,FRED,C	**Time**	12:04:19	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	33922080A1
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	01/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 09/21/2023 12:13

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency	FL HIGHWAY PATROL	**Date**	10/19/2023	**Type**	Agency Inspection
Inspector	HENING,FRED,C	**Time**	11:50:41	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	33922080A1
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	01/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.084
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.084
Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 10/19/2023 11:59

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency FL HIGHWAY PATROL **Date** 11/29/2023 **Type** Agency Inspection
Inspector HENING,FRED,C **Time** 11:18:37 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # (Repeat) 1	0.079
0.20 g/210L Test # 1	0.192
0.08 g/210L Dry Gas Std Test # 1	0.084
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # (Repeat) 2	0.080
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.085
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # (Repeat) 3	0.080
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.084
Interferent Detect Test # 3	INT

Remarks:

08: RFI Detect PHONE TOO CLOSE.

Data Download Date/Time 11/29/2023 11:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

Agency	FL HIGHWAY PATROL	**Date**	12/20/2023	**Type**	Agency Inspection
Inspector	HENING,FRED,C	**Time**	11:28:29	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	33922080A1
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	01/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.084
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.085
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.084
Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 12/20/2023 11:39

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency	FL HIGHWAY PATROL	**Date**	01/30/2023	**Type**	Agency Inspection
Inspector	MURPHY,JAVON,D	**Time**	11:30:38	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

N/ANon-compliance: .

Data Download Date/Time 01/30/2023 12:45

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 01/30/2023 **Type** Agency Inspection
Inspector MURPHY,JAVON,D **Time** 12:36:30 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.084
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.084
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.084
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/30/2023 12:45

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 02/27/2023 **Type** Agency Inspection
Inspector MURPHY,JAVON,D **Time** 12:51:35 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.084
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.084
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.084
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/27/2023 13:12

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency	FL HIGHWAY PATROL	**Date**	04/11/2023	**Type**	Agency Inspection
Inspector	MURPHY,JAVON,D	**Time**	18:55:52	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202108D	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	1433249
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/15/2024	**Expiration**	01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/11/2023 19:11

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 05/25/2023 **Type** Agency Inspection
Inspector HILLIARD,KARI,A **Time** 12:17:39 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 05/25/2023 12:48

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 06/06/2023 **Type** Department Inspection
Inspector NICODEMO,PHIL, **Time** 15:27:01 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** 06723080A5
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 04/05/2025

Number of Simulators 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 06/06/2023 15:32

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 06/30/2023 **Type** Agency Inspection
Inspector MURPHY,JAVON,D **Time** 18:58:38 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/01/2023 17:25

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 07/29/2023 **Type** Agency Inspection
Inspector MURPHY,JAVON,D **Time** 16:28:45 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.192
0.08 g/210L Dry Gas Std Test # 1	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 07/29/2023 17:53

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 08/31/2023 **Type** Agency Inspection
Inspector MURPHY,JAVON,D **Time** 12:37:10 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.192
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/31/2023 15:38

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency FL HIGHWAY PATROL **Date** 10/14/2023 **Type** Agency Inspection
Inspector MURPHY,JAVON,D **Time** 01:38:13 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 10/14/2023 03:11

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

Agency	FL HIGHWAY PATROL	**Date**	12/02/2023	**Type**	Agency Inspection
Inspector	MURPHY,JAVON,D	**Time**	19:54:59	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202307A	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	1433249
Expiration	03/28/2025	**Expiration**	07/11/2025	**Expiration**	06/15/2024	**Expiration**	01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/03/2023 21:00

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI **Date** 01/31/2023 **Type** Agency Inspection
Inspector LORENZO,YOSNIEL, **Time** 22:34:09 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1356262
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/03/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/01/2023 18:28

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI **Date** 02/01/2023 **Type** Agency Inspection
Inspector LORENZO,YOSNIEL, **Time** 01:47:26 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1356262
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/03/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/01/2023 18:28

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI **Date** 02/17/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 16:57:32 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/17/2023 17:04

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI **Date** 05/31/2023 **Type** Agency Inspection
Inspector LORENZO,YOSNIEL, **Time** 23:49:20 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 402678400
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 02/24/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

.

Data Download Date/Time 06/01/2023 00:04

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI **Date** 07/01/2023 **Type** Agency Inspection
Inspector LORENZO,YOSNIEL, **Time** 00:38:03 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402678400
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/24/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

.

Data Download Date/Time 07/01/2023 00:47

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI **Date** 08/01/2023 **Type** Agency Inspection
Inspector LORENZO,YOSNIEL, **Time** 00:22:55 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402678400
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 02/24/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

.

Data Download Date/Time 08/01/2023 00:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency	FHP TROOP E MIAMI	**Date**	08/31/2023	**Type**	Agency Inspection
Inspector	LORENZO,YOSNIEL,	**Time**	23:47:55	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	402678400
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	02/24/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

.

Data Download Date/Time 08/31/2023 23:56

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI **Date** 09/30/2023 **Type** Agency Inspection
Inspector LORENZO,YOSNIEL, **Time** 19:45:45 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402678400
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/24/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

.

Data Download Date/Time 09/30/2023 19:58

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI **Date** 10/31/2023 **Type** Agency Inspection
Inspector LORENZO,YOSNIEL, **Time** 15:08:50 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402678400
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/24/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/31/2023 15:17

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency FHP TROOP E MIAMI **Date** 11/29/2023 **Type** Agency Inspection
Inspector BRANDT,FABIAN,E **Time** 20:00:37 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402678400
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/24/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.080
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
Interferent Detect Test # 3	INT

Remarks:

NOT PROPERLY INSTALLED.

Data Download Date/Time 11/29/2023 20:11

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

Agency	FHP TROOP E MIAMI	**Date**	12/27/2023	**Type**	Agency Inspection
Inspector	BRANDT,FABIAN,E	**Time**	17:43:25	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	402678400
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	02/24/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/27/2023 17:54

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

Agency	FL HIGHWAY PATROL	**Date**	06/08/2023	**Type**	Agency Inspection
Inspector	GUTSCHOW,TAYLOR,D	**Time**	13:34:32	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Data Download Date/Time 06/09/2023 14:58

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

Agency FL HIGHWAY PATROL **Date** 06/09/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 14:47:35 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 06/09/2023 14:58

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency	FL HIGHWAY PATROL	**Date**	01/24/2023	**Type**	Agency Inspection
Inspector	BUDAY,JOSHUA,M	**Time**	11:58:08	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202108D	**0.20 Lot #**	202106B	**0.08 Gas Lot #**	212654
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/22/2023	**Expiration**	01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

UNIT CLEAN/FUNCTIONAL AT TIME OF INSPECTION.

Data Download Date/Time 01/24/2023 12:03

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency FL HIGHWAY PATROL **Date** 02/28/2023 **Type** Agency Inspection
Inspector HITCHCOCK,WILLIAM,D **Time** 11:19:11 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 212654
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.191
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/28/2023 11:25

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency FL HIGHWAY PATROL **Date** 03/31/2023 **Type** Agency Inspection
Inspector HITCHCOCK,WILLIAM,D **Time** 11:19:08 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 212654
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.192
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/31/2023 13:14

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency	FL HIGHWAY PATROL	**Date**	04/28/2023	**Type**	Agency Inspection
Inspector	BUDAY,JOSHUA,M	**Time**	11:09:15	**Software**	8100.27

0.05 Lot #	20212A	**0.08 Lot #**	202108D	**0.20 Lot #**	202106B	**0.08 Gas Lot #**	212654
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/22/2023	**Expiration**	01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.190
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.191
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

UNIT CLEAN/FUNCTIONAL AT TIME OF INSPECTION.

Data Download Date/Time 04/28/2023 11:21

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency FL HIGHWAY PATROL **Date** 05/04/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 15:13:23 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 05/04/2023 15:22

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

Agency FL HIGHWAY PATROL **Date** 12/28/2023 **Type** Agency Inspection
Inspector BUDAY,JOSHUA,M **Time** 09:32:49 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202307C **0.08 Gas Lot #** 402872586
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 07/12/2025 **Expiration** 10/12/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.191
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	0.000
Interferent Detect Test # (Repeat) 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.191
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	0.000
Interferent Detect Test # (Repeat) 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det:CHECK SIM SEAL AND TIGHTENED..ALL TESTS OK.

Data Download Date/Time 12/28/2023 09:40

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 01/31/2023 **Type** Agency Inspection
Inspector GONZALEZ,CHRISTOPHER,M **Time** 01:54:51 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED, NO CRACKS. IN COMPLIANCE.

Data Download Date/Time 02/01/2023 12:46

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 02/28/2023 **Type** Agency Inspection
Inspector DOBOSIEWICZ,ANTHONY,N **Time** 13:21:40 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Time-Date changed. A F / M A: RFI Detect.

Data Download Date/Time 02/28/2023 14:39

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 03/30/2023 **Type** Agency Inspection
Inspector GONZALEZ,CHRISTOPHER,M **Time** 16:05:58 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

INSPECTED CASE NO CRACKS, IN COMPLIANCE

Data Download Date/Time 03/31/2023 16:12

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 04/28/2023 **Type** Agency Inspection
Inspector GONZALEZ,CHRISTOPHER,M **Time** 17:48:32 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

CHECKED CASE AND WIPED DOWN. IN COMPLIANCE.

Data Download Date/Time 05/03/2023 01:46

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 05/29/2023 **Type** Agency Inspection
Inspector DOBOSIEWICZ,ANTHONY,N **Time** 03:20:00 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 05/29/2023 03:54

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency	FL HIGHWAY PATROL	**Date**	06/22/2023	**Type**	Agency Inspection
Inspector	GONZALEZ,CHRISTOPHER,M	**Time**	18:02:00	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	1433249
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED AND WIPED DOWN

Data Download Date/Time 06/22/2023 18:09

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 07/21/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 11:15:22 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** 06723080A5
Expiration 01/11/2024 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 04/05/2025

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 07/21/2023 11:20

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 08/24/2023 **Type** Agency Inspection
Inspector DOBOSIEWICZ,ANTHONY,N **Time** 23:51:32 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # (Repeat) 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # (Repeat) 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # (Repeat) 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

08: RFI Detect/ TROOPER WITH RADIO.

Data Download Date/Time 08/25/2023 00:28

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 09/05/2023 **Type** Agency Inspection
Inspector DOBOSIEWICZ,ANTHONY,N **Time** 10:28:00 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206B **0.08 Gas Lot #** 1433249
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/14/2024 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED NO CRACKS/ IN COMPLIANCE

Data Download Date/Time 09/05/2023 10:10

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency	FL HIGHWAY PATROL	**Date**	10/30/2023	**Type**	Agency Inspection
Inspector	GONZALEZ,CHRISTOPHER,M	**Time**	17:33:35	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	1433249
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED OK.

Data Download Date/Time 10/30/2023 17:43

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 11/29/2023 **Type** Agency Inspection
Inspector DOBOSIEWICZ,ANTHONY,N **Time** 04:47:59 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED NO CRACKS/ IN COMPLIANCE

Data Download Date/Time 11/29/2023 04:58

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

Agency FL HIGHWAY PATROL **Date** 12/30/2023 **Type** Agency Inspection
Inspector DOBOSIEWICZ,ANTHONY,N **Time** 02:47:05 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433249
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.08 g/210L Test # (Repeat) 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	
0.08 g/210L Test # (Repeat) 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # (Repeat) 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

08: RFI Detect/ TROOPER WITH CELL PHONE WALKED IN.

Data Download Date/Time 12/30/2023 03:00

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency FL HIGHWAY PATROL **Date** 03/15/2023 **Type** Agency Inspection
Inspector LAMONT,GREGORY,S **Time** 14:55:54 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433248
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/06/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/15/2023 16:12

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency FL HIGHWAY PATROL **Date** 03/20/2023 **Type** Agency Inspection
Inspector LAMONT,GREGORY,S **Time** 13:43:30 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433248
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/06/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/04/2023 08:34

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency	FL HIGHWAY PATROL	**Date**	05/03/2023	**Type**	Agency Inspection
Inspector	LAMONT,GREGORY,S	**Time**	13:58:43	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	1433248
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	01/06/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/04/2023 08:34

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency FL HIGHWAY PATROL **Date** 05/26/2023 **Type** Department Inspection
Inspector NICODEMO,PHIL, **Time** 15:37:20 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** 06723080A5
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 04/05/2025

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.081
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 05/26/2023 15:43

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency FL HIGHWAY PATROL **Date** 08/18/2023 **Type** Agency Inspection
Inspector LAMONT,GREGORY,S **Time** 10:48:38 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433248
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/06/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/18/2023 13:13

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency FL HIGHWAY PATROL **Date** 09/08/2023 **Type** Agency Inspection
Inspector LAMONT,GREGORY,S **Time** 09:50:39 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433248
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/06/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/08/2023 11:40

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency	FL HIGHWAY PATROL	**Date**	10/17/2023	**Type**	Agency Inspection
Inspector	LAMONT,GREGORY,S	**Time**	12:45:28	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	1433248
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	01/06/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/17/2023 12:55

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency FL HIGHWAY PATROL **Date** 11/20/2023 **Type** Agency Inspection
Inspector HERNANDEZ,ALAIN, **Time** 10:33:21 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1433248
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/06/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/20/2023 11:14

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

Agency	FL HIGHWAY PATROL	**Date**	12/21/2023	**Type**	Agency Inspection
Inspector	HERNANDEZ,ALAIN,	**Time**	11:20:19	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	1433248
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	01/06/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/21/2023 11:41

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL **Date** 01/06/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 09:01:56 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG113403
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 05/14/2023

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 01/06/2023 10:10

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL **Date** 01/27/2023 **Type** Agency Inspection
Inspector MARCUM,JOSEPH,K **Time** 10:08:29 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402357111
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.075
0.20 g/210L Test # 1	0.192
0.20 g/210L Test # (Repeat) 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.075
0.20 g/210L Test # 2	0.190
0.20 g/210L Test # (Repeat) 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.188
0.20 g/210L Test # (Repeat) 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

AMB FAIL- TURNED ON FAN .20- TIGHTENED SEAL

Data Download Date/Time 01/27/2023 16:15

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency	FL HIGHWAY PATROL	**Date**	02/16/2023	**Type**	Agency Inspection
Inspector	BARGE,SUSAN,M	**Time**	11:47:01	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: RFI Detect, RFI Detect UNKNOWN. Non-compliance:RFI TWICE.

Data Download Date/Time 02/16/2023 12:56

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL **Date** 02/21/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 10:20:10 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.077
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.047
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.077
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.077
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.047
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.046
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0007
0.08 g/210L Test (Std Dev)	0.0008
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/21/2023 11:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL **Date** 03/30/2023 **Type** Agency Inspection
Inspector DRANE,RICHARD,B **Time** 10:25:21 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 03/30/2023 14:55

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL **Date** 04/21/2023 **Type** Agency Inspection
Inspector DRANE,RICHARD,B **Time** 13:49:46 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.193
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 04/21/2023 16:49

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency	FL HIGHWAY PATROL	**Date**	05/16/2023	**Type**	Agency Inspection
Inspector	WALLEN,JUSTIN,R	**Time**	12:32:16	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	402357111
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 05/16/2023 14:51

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL **Date** 06/14/2023 **Type** Agency Inspection
Inspector WALLEN,JUSTIN,R **Time** 11:42:44 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 06/14/2023 13:15

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL **Date** 07/19/2023 **Type** Agency Inspection
Inspector WALLEN,JUSTIN,R **Time** 12:51:15 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

N/A

Data Download Date/Time 07/19/2023 15:49

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL **Date** 08/28/2023 **Type** Agency Inspection
Inspector DRANE,RICHARD,B **Time** 11:34:17 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.127
0.20 g/210L Test # (Repeat) 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.133
0.20 g/210L Test # (Repeat) 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.194
0.20 g/210L Test # (Repeat) 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance.

Data Download Date/Time 08/28/2023 14:02

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL **Date** 09/26/2023 **Type** Agency Inspection
Inspector DRANE,RICHARD,B **Time** 07:31:21 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
Expiration 12/07/2023 **Expiration** 05/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.190
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 09/25/2023 10:08

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL **Date** 10/21/2023 **Type** Agency Inspection
Inspector WALLEN,JUSTIN,R **Time** 22:11:08 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

NA

Data Download Date/Time 10/24/2023 10:59

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL **Date** 11/21/2023 **Type** Agency Inspection
Inspector WALLEN,JUSTIN,R **Time** 10:10:30 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 202206B
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/14/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.081
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 11/21/2023 11:58

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

Agency FL HIGHWAY PATROL **Date** 12/04/2023 **Type** Agency Inspection
Inspector DRANE,RICHARD,B **Time** 09:08:52 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402357111
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/10/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.184
0.20 g/210L Test # (Repeat) 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.182
0.20 g/210L Test # (Repeat) 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.181
0.20 g/210L Test # (Repeat) 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance.

Data Download Date/Time 12/04/2023 10:25

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 01/13/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 13:34:04 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG113403
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 05/14/2023

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 01/13/2023 13:40

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 01/27/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 14:09:56 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402328118
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/20/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/27/2023 14:48

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 02/24/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 14:11:27 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402328118
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/20/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.081
Interferent Detect Test # 3	INT

Remarks:

08:Cntrl Out Tol DID NOT CONNECT DRY GAS HOSE,RE-CONNECTED . RE-TESTED/NOT RE-STARTED

Data Download Date/Time 02/24/2023 14:16

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 03/24/2023 **Type** Agency Inspection
Inspector LYONS/DANIEL',DANIEL/OWEN, **Time** 11:48:17 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402328118
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/20/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/24/2023 11:57

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 04/25/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 14:02:11 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402328118
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/20/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/25/2023 14:48

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 05/19/2023 **Type** Agency Inspection
Inspector LYOONS,DANIEL,E **Time** 13:07:16 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402328118
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/20/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/19/2023 13:23

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 06/21/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 14:20:26 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402328118
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/20/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/21/2023 14:34

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 07/17/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 13:26:20 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402328118
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 01/20/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/17/2023 13:42

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 08/29/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 08:54:19 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402328114
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/20/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/29/2023 09:03

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 09/12/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 17:09:38 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402328118
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/20/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/12/2023 17:30

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 10/23/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 15:38:28 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402328118
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/20/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/23/2023 15:49

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 11/22/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 15:53:51 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402328118
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/20/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/22/2023 16:14

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency OSCEOLA COUNTY S.O. **Date** 12/21/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 12:13:29 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402328118
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/20/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/21/2023 12:25

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 01/03/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 13:30:11 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402330433
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/03/2023 13:38

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 01/18/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 11:18:52 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402330433
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 05/22/2023 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/18/2023 12:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 01/27/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 14:12:22 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402330433
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/27/2023 14:45

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 02/13/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL,X **Time** 09:05:47 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.202
0.08 g/210L Dry Gas Std Test # 4	0.082
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.081
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.082
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 02/13/2023 09:10

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 02/24/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 14:21:35 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402330433
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.084
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.084
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.084
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/24/2023 14:29

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 03/24/2023 **Type** Agency Inspection
Inspector LYONS/GAYLE,DANIEL/OWEN,E **Time** 10:56:09 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402330433
Expiration 12/07/2023 **Expiration** 08/17/2021 **Expiration** 06/22/2023 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/24/2023 11:03

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 04/25/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 14:35:16 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402330433
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/25/2023 14:52

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 05/19/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 13:18:23 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402330433
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/19/2023 13:29

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 06/21/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 14:51:10 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402330433
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	No
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

AF/MA:MOUTH ALCOHOL NOT RINSED FROM PREVIOUS INSTRUMENT .

Data Download Date/Time 06/21/2023 15:53

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 06/21/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 15:43:27 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402330433
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.084
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.084
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/21/2023 15:53

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 07/17/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 14:00:58 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402330433
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/17/2023 14:08

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 08/29/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 08:24:51 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402330433
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/29/2023 08:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 09/12/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 17:40:39 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402330433
Expiration 03/28/2025 **Expiration** 06/14/2022 **Expiration** 06/15/2024 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/12/2023 18:24

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 10/23/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 16:05:48 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402330433
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.084
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.084
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.084
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/23/2023 16:19

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 11/22/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 15:34:25 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402330433
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.084
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/22/2023 15:48

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency OSCEOLA COUNTY S.O. **Date** 12/21/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 12:21:04 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402330433
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.084
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/21/2023 12:32

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency	LAKELAND PD	**Date**	08/07/2023	**Type**	Agency Inspection
Inspector	GUTSCHOW,TAYLOR,D	**Time**	13:44:20	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Data Download Date/Time 08/08/2023 14:11

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD **Date** 08/08/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 14:00:54 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 08/08/2023 14:11

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD **Date** 08/21/2023 **Type** Agency Inspection
Inspector ALMEIDA,CAMILO,J **Time** 10:09:31 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06723080A5
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/21/2023 10:15

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD **Date** 09/25/2023 **Type** Agency Inspection
Inspector ALMEIDA,CAMILO,J **Time** 09:16:11 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06723080A5
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/25/2023 09:39

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD **Date** 10/30/2023 **Type** Agency Inspection
Inspector ALMEIDA,CAMILO,J **Time** 11:01:13 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 14323080A3
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

NO REASON

Data Download Date/Time 10/30/2023 11:11

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD **Date** 11/28/2023 **Type** Agency Inspection
Inspector ALMEIDA,CAMILO,J **Time** 08:55:44 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 14323080A3
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/28/2023 13:26

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency LAKELAND PD **Date** 12/18/2023 **Type** Agency Inspection
Inspector ALMEIDA,CAMILO,J **Time** 10:31:20 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 14323080A3
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 09/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/18/2023 14:35

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 01/03/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 13:48:21 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AF/MA;SUBJ PROVIDING M/A RAN OUT OF BREATH, RETESTED

Data Download Date/Time 01/03/2023 13:55

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 01/09/2023 **Type** Department Inspection
Inspector NICODEMO,PHIL, **Time** 13:36:57 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.202
0.08 g/210L Dry Gas Std Test # 4	0.077
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.077
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.076
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0009

Remarks:

Data Download Date/Time 01/09/2023 13:41

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 01/18/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 11:33:21 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/18/2023 12:29

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 01/27/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 13:58:25 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/27/2023 14:40

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 02/24/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 13:28:25 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/24/2023 14:10

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 03/24/2023 **Type** Agency Inspection
Inspector LYONS/GAYLE,DANIEL/OWEN,E **Time** 11:42:14 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/24/2023 11:49

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 04/25/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 14:49:17 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/25/2023 14:57

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 05/19/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 13:01:25 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/19/2023 13:19

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 06/21/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 14:13:36 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/21/2023 14:30

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 07/17/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 13:34:41 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402397742
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2022 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/17/2023 13:48

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 08/29/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 08:48:38 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402397742
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/29/2023 08:58

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 09/12/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 17:28:03 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402397742
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/12/2023 18:09

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 10/23/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 15:50:40 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402397742
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/23/2023 16:50

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 11/22/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 15:59:18 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402397742
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/22/2023 16:24

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency OSCEOLA COUNTY S.O. **Date** 12/21/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 12:03:39 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402397742
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/21/2023 12:14

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency FORT WALTON BEACH PD **Date** 01/29/2023 **Type** Agency Inspection
Inspector CRAWFORD,BENJAMIN,D **Time** 06:31:28 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 211143
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/02/2023 00:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency FORT WALTON BEACH PD **Date** 02/28/2023 **Type** Agency Inspection
Inspector CRAWFORD,BENJAMIN,D **Time** 00:30:01 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration 12/07/2023 **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test # 1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test # 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # (Repeat) 2	0.044
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test # 3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test # (Repeat) 3	0.044
0.20 g/210L Test: OK	No	Interferent Detect Test # 3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

Data Download Date/Time 03/05/2023 01:58

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency FORT WALTON BEACH PD **Date** 02/28/2023 **Type** Agency Inspection
Inspector CRAWFORD,BENJAMIN,D **Time** 01:09:21 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 211143
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/05/2023 01:58

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency FORT WALTON BEACH PD **Date** 03/30/2023 **Type** Agency Inspection
Inspector CRAWFORD,BENJAMIN,D **Time** 17:44:26 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 211143
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/30/2023 18:53

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency FORT WALTON BEACH PD **Date** 04/30/2023 **Type** Agency Inspection
Inspector CRAWFORD,BENJAMIN,D **Time** 17:53:27 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 211143
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/05/2023 02:25

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency FORT WALTON BEACH PD **Date** 05/30/2023 **Type** Agency Inspection
Inspector CRAWFORD,BENJAMIN,D **Time** 07:16:39 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 211122
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2023 08:24

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency FORT WALTON BEACH PD **Date** 06/30/2023 **Type** Agency Inspection
Inspector CRAWFORD,BENJAMIN,D **Time** 12:59:49 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 211122
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/30/2023 14:07

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency FORT WALTON BEACH PD **Date** 07/31/2023 **Type** Agency Inspection
Inspector CRAWFORD,BENJAMIN,D **Time** 11:56:20 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 211122
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/31/2023 13:04

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency FORT WALTON BEACH PD **Date** 08/23/2023 **Type** Agency Inspection
Inspector CRAWFORD,BENJAMIN,D **Time** 19:28:37 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 211122
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.000
0.05 g/210L Test # (Repeat) 1	0.046
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.173
0.20 g/210L Test # (Repeat) 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.000
0.05 g/210L Test # (Repeat) 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.180
0.20 g/210L Test # (Repeat) 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.000
0.05 g/210L Test # (Repeat) 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.182
0.20 g/210L Test # (Repeat) 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance. 20: Control Outside Tolerance.

Data Download Date/Time 08/23/2023 20:37

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency FORT WALTON BEACH PD **Date** 09/29/2023 **Type** Agency Inspection
Inspector CRAWFORD,BENJAMIN,D **Time** 09:19:46 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 211122
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/29/2023 10:29

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency FORT WALTON BEACH PD **Date** 10/25/2023 **Type** Agency Inspection
Inspector CRAWFORD,BENJAMIN,D **Time** 09:22:23 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 211122
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/25/2023 10:35

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency FORT WALTON BEACH PD **Date** 11/30/2023 **Type** Agency Inspection
Inspector CRAWFORD,BENJAMIN,D **Time** 07:42:06 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 231294
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.084
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.084
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/30/2023 08:55

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency FORT WALTON BEACH PD **Date** 12/22/2023 **Type** Department Inspection
Inspector SIDDOWAY,BENJAMIN,W **Time** 12:03:46 **Software** 8100.27

0.05 Lot # 202303K **Expiration** 03/29/2025
0.08 Lot # 202303L **Expiration** 03/29/2025
0.20 Lot # 202304C **Expiration** 04/05/2025
0.08 Gas Lot # AG310901 **Expiration** 04/19/2025

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.196
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 12/22/2023 13:14

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency FORT WALTON BEACH PD **Date** 12/24/2023 **Type** Agency Inspection
Inspector CRAWFORD,BENJAMIN,D **Time** 21:20:31 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 231294
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/24/2023 22:29

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

Agency	GULF BREEZE PD	**Date**	03/07/2023	**Type**	Agency Inspection
Inspector	SOTO,ISRAEL,	**Time**	07:24:22	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Data Download Date/Time 03/07/2023 10:59

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

Agency GULF BREEZE PD **Date** 03/07/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 09:44:27 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.075
0.20 g/210L Test # 4	0.196
0.08 g/210L Dry Gas Std Test # 4	0.076
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.047
0.08 g/210L Test # 5	0.076
0.20 g/210L Test # 5	0.196
0.08 g/210L Dry Gas Std Test # 5	0.076
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.047
0.08 g/210L Test # 6	0.076
0.20 g/210L Test # 6	0.196
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.046
0.08 g/210L Test # 7	0.076
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.047
0.08 g/210L Test # 8	0.076
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.076
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.047
0.08 g/210L Test # 9	0.076
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.076
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.047
0.08 g/210L Test # 10	0.076
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.076
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 03/07/2023 10:59

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

Agency GULF BREEZE PD **Date** 05/02/2023 **Type** Agency Inspection
Inspector ROBERSON,THOMAS,C **Time** 00:55:37 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1356262
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/03/2023

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.074
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.074
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.075
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.FRESH CYCLINDER. RERAN TEST.

Data Download Date/Time 05/02/2023 21:17

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

Agency	GULF BREEZE PD	**Date**	06/05/2023	**Type**	Agency Inspection
Inspector	ROBERSON,THOMAS,C	**Time**	00:41:38	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	402697419
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	03/20/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/14/2023 01:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

Agency	GULF BREEZE PD	**Date**	07/05/2023	**Type**	Agency Inspection
Inspector	ROBERSON,THOMAS,C	**Time**	23:55:49	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	402697419
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	03/20/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/07/2023 00:20

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

Agency GULF BREEZE PD **Date** 08/06/2023 **Type** Agency Inspection
Inspector ROBERSON,THOMAS,C **Time** 06:10:32 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402697419
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/20/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.045
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/06/2023 19:17

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

Agency GULF BREEZE PD **Date** 09/17/2023 **Type** Agency Inspection
Inspector ROBERSON,THOMAS,C **Time** 22:45:49 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402697419
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/20/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.044
0.05 g/210L Test # (Repeat) 1	0.045
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.190
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.044
0.05 g/210L Test # (Repeat) 2	0.045
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.045
0.05 g/210L Test # (Repeat) 3	0.045
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

CHECKED CONNECTIONS. RERAN TEST.

Data Download Date/Time 09/18/2023 01:26

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

Agency GULF BREEZE PD **Date** 10/01/2023 **Type** Agency Inspection
Inspector ROBERSON,THOMAS,C **Time** 01:43:08 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402367487
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/28/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/01/2023 03:51

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 01/30/2023 **Type** Agency Inspection
Inspector NICHOLS,MICHAEL,J **Time** 15:37:14 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** O325341-9
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 11/18/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/30/2023 16:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 02/24/2023 **Type** Agency Inspection
Inspector NICHOLS,MICHAEL,J **Time** 13:30:33 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** O325341-9
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 11/18/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

A F / M A: Range Exceeded. TOO MUCH MA SOLUTION FOR MA TEST

Data Download Date/Time 02/24/2023 14:37

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 03/17/2023 **Type** Agency Inspection
Inspector NICHOLS,MICHAEL,J **Time** 14:37:04 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** O325341-9
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/18/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/17/2023 15:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 04/28/2023 **Type** Agency Inspection
Inspector NICHOLS,MICHAEL,J **Time** 15:10:50 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** O325341-4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/18/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/28/2023 16:19

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 05/19/2023 **Type** Agency Inspection
Inspector NICHOLS,MICHAEL,J **Time** 10:42:29 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** O325341-4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/18/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/19/2023 11:49

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 06/04/2023 **Type** Agency Inspection
Inspector NICHOLS,MICHAEL,J **Time** 14:29:36 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** O325341-4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/18/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/05/2023 08:46

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 07/14/2023 **Type** Agency Inspection
Inspector NICHOLS,MICHAEL,J **Time** 11:16:07 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** O325341-4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/18/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

NOT ENOUGH MA SOLUTION ON TONGUE.

Data Download Date/Time 07/14/2023 12:23

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 08/23/2023 **Type** Agency Inspection
Inspector NICHOLS,MICHAEL,J **Time** 14:55:55 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** O325341-4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/18/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/23/2023 16:05

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 09/28/2023 **Type** Agency Inspection
Inspector NICHOLS,MICHAEL,J **Time** 14:34:20 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** O325341-4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/18/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/28/2023 15:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 10/30/2023 **Type** Agency Inspection
Inspector NICHOLS,MICHAEL,J **Time** 11:27:56 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/22/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/30/2023 12:42

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 11/21/2023 **Type** Agency Inspection
Inspector NICHOLS,MICHAEL,J **Time** 13:53:54 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/22/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/21/2023 15:35

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 11/21/2023 **Type** Agency Inspection
Inspector NICHOLS,MICHAEL,J **Time** 18:00:49 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/22/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/21/2023 19:09

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency OKALOOSA COUNTY SO **Date** 12/05/2023 **Type** Agency Inspection
Inspector NICHOLS,MICHAEL,J **Time** 15:14:44 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402623828
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/22/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.084
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.084
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/05/2023 16:28

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 01/17/2023 **Type** Agency Inspection
Inspector DUNLAP,GABRIEL,C **Time** 17:33:33 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402590345
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 11/04/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/17/2023 18:39

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 02/27/2023 **Type** Agency Inspection
Inspector DUNLAP,GABRIEL,C **Time** 05:27:31 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402590345
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 11/04/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/02/2023 19:00

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 03/31/2023 **Type** Agency Inspection
Inspector DUNLAP,GABRIEL,C **Time** 04:29:19 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402590345
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 11/04/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.084
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.084
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.084
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/10/2023 13:17

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency	CRESTVIEW PD	**Date**	04/28/2023	**Type**	Agency Inspection
Inspector	DUNLAP,GABRIEL,C	**Time**	03:05:18	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	402590345
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	11/04/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/28/2023 04:12

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 05/30/2023 **Type** Agency Inspection
Inspector DUNLAP,GABRIEL,C **Time** 09:50:04 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402590345
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/04/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/30/2023 10:59

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 07/01/2023 **Type** Agency Inspection
Inspector DUNLAP,GABRIEL,C **Time** 10:11:28 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402590345
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/04/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/01/2023 11:18

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 07/30/2023 **Type** Agency Inspection
Inspector DUNLAP,GABRIEL,C **Time** 08:16:23 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402590345
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/04/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.084
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.084
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/07/2023 18:15

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 08/31/2023 **Type** Agency Inspection
Inspector DUNLAP,GABRIEL,C **Time** 21:38:13 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402590345
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/04/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.084
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.084
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.084
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/31/2023 22:46

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 09/29/2023 **Type** Agency Inspection
Inspector DUNLAP,GABRIEL,C **Time** 19:30:03 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402590345
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/04/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/29/2023 20:40

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 10/31/2023 **Type** Agency Inspection
Inspector DUNLAP,GABRIEL,C **Time** 10:56:01 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 402590345
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 11/04/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.191
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.084
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.192
0.08 g/210L Dry Gas Std Test # 3	0.084
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/31/2023 12:04

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 12/10/2023 **Type** Agency Inspection
Inspector DUNLAP,GABRIEL,C **Time** 17:06:40 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 402590345
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 11/04/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.107
0.20 g/210L Test # (Repeat) 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.20 g/210L Test # (Repeat) 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.20 g/210L Test # (Repeat) 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance.

Data Download Date/Time 12/10/2023 18:15

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 12/11/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 10:56:44 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG229803
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.082
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.082
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.082
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.082
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.082
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.082
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.052
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.082
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 12/11/2023 12:05

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency CRESTVIEW PD **Date** 12/11/2023 **Type** Agency Inspection
Inspector DUNLAP,GABRIEL,C **Time** 14:38:12 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 402590345
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 11/04/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/11/2023 15:47

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 01/28/2023 **Type** Agency Inspection
Inspector BRENCKLE,DEP,T **Time** 23:06:04 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 15922080A4
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.079
Interferent Detect Test # 3	INT

Remarks:

08: RFI Detect.

Data Download Date/Time 02/01/2023 12:37

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 02/24/2023 **Type** Agency Inspection
Inspector BRENCKLE,TIMOTHY,M **Time** 20:15:20 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 15922080A4
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 02/24/2023 21:23

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 03/26/2023 **Type** Agency Inspection
Inspector BRENCKLE,TIMOTHY,M **Time** 22:12:57 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 15922080A4
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

CHANGED TIME

Data Download Date/Time 03/30/2023 10:20

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 04/09/2023 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 10:06:59 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 15922080A4
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 04/13/2023 07:28

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 05/04/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 14:18:06 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 05/04/2023 14:37

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 05/12/2023 **Type** Agency Inspection
Inspector BRENCKLE,TIMOTHY,M **Time** 12:14:04 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 15922080A4
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 05/16/2023 18:40

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 06/23/2023 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 10:51:05 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 06/23/2023 11:00

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 07/27/2023 **Type** Agency Inspection
Inspector SANDERS,DEPUTY,J **Time** 08:44:47 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

PASSED INSP AT JAIL

Data Download Date/Time 07/27/2023 10:18

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 08/29/2023 **Type** Agency Inspection
Inspector SANDERS,DEPUTY,J **Time** 10:45:18 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.075
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.076
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.074
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.076
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.074
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.076
Interferent Detect Test # 3	INT

Remarks:

PASSED INSP AT N JAIL

Data Download Date/Time 08/29/2023 10:52

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency SARASOTA COUNTY SO **Date** 09/25/2023 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 09:36:57 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 09/25/2023 09:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency	SARASOTA COUNTY SO	**Date**	10/18/2023	**Type**	Agency Inspection
Inspector	SANDERS,DEPUTY,J	**Time**	10:36:42	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	15922080A4
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

INSPECTION PASSED AT N JAIL

Data Download Date/Time 10/18/2023 10:46

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency	SARASOTA COUNTY SO	**Date**	11/30/2023	**Type**	Agency Inspection
Inspector	SANDERS,DEPUTY,J	**Time**	11:45:00	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	15922080A4
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

PASSED INSPEC AT JAIL

Data Download Date/Time 11/30/2023 11:57

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

Agency	SARASOTA COUNTY SO	**Date**	12/19/2023	**Type**	Agency Inspection
Inspector	SANDERS,DEPUTY,J	**Time**	10:22:31	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	15922080A4
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT JAIL

Data Download Date/Time 12/19/2023 10:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD **Date** 02/23/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:42:36 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0008
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/23/2023 13:47

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD **Date** 03/01/2023 **Type** Agency Inspection
Inspector DE JESUS,RAY,N **Time** 20:11:17 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/01/2023 20:19

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD **Date** 04/05/2023 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 19:15:21 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 29122080A1
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 12/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/05/2023 19:20

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD **Date** 05/10/2023 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 20:58:38 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 29122080A1 **0.08 Gas Lot #** 29122080A1
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 12/05/2024 **Expiration** 12/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

00: RFI Detect-CLEANING CREW IN THE HOLDING CELLS.

Data Download Date/Time 06/16/2023 17:27

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD **Date** 06/21/2023 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 22:31:02 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 12/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/21/2023 22:37

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD **Date** 07/03/2023 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 22:04:14 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.192
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/03/2023 22:10

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD **Date** 08/03/2023 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 21:19:36 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/08/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.192
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.075
0.20 g/210L Test # 2	0.190
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.192
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/03/2023 21:27

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD **Date** 09/28/2023 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 23:04:02 **Software** 8100.27

0.05 Lot # 20112A **0.08 Lot #** 20226B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

00: RFI Detect-DETECTIVES IN THE HOLDING CELL.

Data Download Date/Time 09/28/2023 23:14

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency	NORTH MIAMI PD	**Date**	10/14/2023	**Type**	Agency Inspection
Inspector	DE JESUS,RAY,	**Time**	17:23:11	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	29122080A1
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	12/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/14/2023 17:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD **Date** 11/24/2023 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 18:57:51 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/24/2023 19:07

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

Agency NORTH MIAMI PD **Date** 12/02/2023 **Type** Agency Inspection
Inspector DE JESUS,RAY, **Time** 22:52:26 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/02/2023 23:36

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO **Date** 01/10/2023 **Type** Agency Inspection
Inspector DENDY,HALEY,M **Time** 09:37:34 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 35521080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 03/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI S. KNOX #261252 PRESENT FOR INSPECTION.

Data Download Date/Time 01/10/2023 09:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO **Date** 02/07/2023 **Type** Agency Inspection
Inspector KNOX,STEPHANIE,R **Time** 12:50:33 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

AI H.DENDY #260868 PRESENT FOR INSPECTION.

Data Download Date/Time 02/07/2023 13:21

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO **Date** 03/02/2023 **Type** Agency Inspection
Inspector KNOX,STEPHANIE,R **Time** 08:37:27 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

AI H. DENDY #260868 PRESENT FOR INSPECTION.

Data Download Date/Time 03/02/2023 08:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO **Date** 04/10/2023 **Type** Agency Inspection
Inspector KNOX,STEPHANIE,R **Time** 11:21:05 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/14/2024 **Expiration** 03/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT FOR INSPECTION.

Data Download Date/Time 04/10/2023 13:38

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency	HILLSBOROUGH CO SO	**Date**	05/02/2023	**Type**	Agency Inspection
Inspector	KNOX,STEPHANIE,R	**Time**	09:15:38	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C\	**0.08 Gas Lot #**	35521080A2
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	03/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT FOR INSPECTION.

Data Download Date/Time 05/02/2023 10:09

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO **Date** 06/05/2023 **Type** Agency Inspection
Inspector KNOX,STEPHANIE,R **Time** 11:29:11 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 35521080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT FOR INSPECTION.

Data Download Date/Time 06/05/2023 11:49

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO **Date** 07/05/2023 **Type** Agency Inspection
Inspector KNOX,STEPHANIE,R **Time** 09:50:33 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

AI B. KROME #263389 PRESENT FOR INSPECTION.

Data Download Date/Time 07/05/2023 10:01

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO **Date** 08/01/2023 **Type** Agency Inspection
Inspector KNOX,STEPHANIE,R **Time** 09:52:35 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT FOR INSPECTION.

Data Download Date/Time 08/01/2023 10:08

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO **Date** 09/05/2023 **Type** Agency Inspection
Inspector KNOX,STEPHANIE,R **Time** 09:33:34 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT FOR INSPECTION.

Data Download Date/Time 09/05/2023 09:49

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO **Date** 09/12/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 11:02:21 **Software** 8100.27

0.05 Lot #	202303K	**0.08 Lot #**	202303L	**0.20 Lot #**	202304C	**0.08 Gas Lot #**	06723080A5
Expiration	03/29/2025	**Expiration**	03/29/2025	**Expiration**	04/05/2025	**Expiration**	04/05/2025

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0007
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 09/12/2023 11:09

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO **Date** 09/22/2023 **Type** Agency Inspection
Inspector KNOX,STEPHANIE,R **Time** 08:40:24 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 29122080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT FOR INSPECTION.

Data Download Date/Time 09/22/2023 08:50

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO **Date** 10/02/2023 **Type** Agency Inspection
Inspector KNOX,STEPHANIE,R **Time** 13:01:52 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

AI J. DAIL #254236 PRESENT FOR INSPECTION.

Data Download Date/Time 10/02/2023 13:16

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency HILLSBOROUGH CO SO **Date** 11/08/2023 **Type** Agency Inspection
Inspector KNOX,STEPHANIE,R **Time** 10:53:36 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

A F / M A: NOT ENOUGH MA. AI B. KROME #263389 PRESENT

Data Download Date/Time 11/08/2023 11:13

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

Agency	HILLSBOROUGH CO SO	**Date**	12/01/2023	**Type**	Agency Inspection
Inspector	KNOX,STEPHANIE,R	**Time**	11:49:35	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202307A	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	08223080A4
Expiration	03/28/2025	**Expiration**	07/11/2025	**Expiration**	06/15/2024	**Expiration**	06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

AI B. KROME #263389 PRESENT FOR INSPECTION.

Data Download Date/Time 12/01/2023 12:02

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency FISH & WILDLIFE CC **Date** 01/23/2023 **Type** Agency Inspection
Inspector CARTAYA,GUILLERMO, **Time** 11:25:19 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 402477284
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 01/23/2023 12:17

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency	FISH & WILDLIFE CC	**Date**	04/24/2023	**Type**	Agency Inspection
Inspector	GUTSCHOW,TAYLOR,D	**Time**	11:20:52	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. BYPASSED TO OPERATE INSTRUMENT.

Data Download Date/Time 04/26/2023 13:12

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency FISH & WILDLIFE CC **Date** 04/26/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 12:55:56 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.204
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.204
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.203
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.203
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.203
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.203
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 04/26/2023 13:13

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency FFWCC **Date** 08/10/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 12:56:58 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.204
0.08 g/210L Dry Gas Std Test # 1	0.076
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.203
0.08 g/210L Dry Gas Std Test # 4	0.076
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.203
0.08 g/210L Dry Gas Std Test # 5	0.076
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.076
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.076
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.204
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.203
0.08 g/210L Dry Gas Std Test # 9	0.076
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.076
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 08/10/2023 13:05

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency FFWCC **Date** 08/31/2023 **Type** Agency Inspection
Inspector CONROY,JESSICA,P **Time** 14:30:26 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/31/2023 14:41

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency FFWCC **Date** 09/25/2023 **Type** Agency Inspection
Inspector SAAVEDRA,OSCAR, **Time** 14:31:49 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/25/2023 14:40

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency FFWCC **Date** 10/24/2023 **Type** Agency Inspection
Inspector SAAVEDRA,OSCAR, **Time** 13:53:06 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402477284
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.203
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.204
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

A F / M A:ALCOHOL WAS DETECTED IN BREATH.

Data Download Date/Time 10/24/2023 14:02

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency	FFWCC	**Date**	11/27/2023	**Type**	Agency Inspection
Inspector	SAAVEDRA,OSCAR,	**Time**	17:32:11	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	4024772840
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.203
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/27/2023 16:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

Agency	FFWCC	**Date**	12/28/2023	**Type**	Agency Inspection
Inspector	CONROY,JESSICA,P	**Time**	15:17:36	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202307A	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	402477284
Expiration	03/28/2025	**Expiration**	07/11/2025	**Expiration**	06/15/2024	**Expiration**	06/27/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/28/2023 14:25

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO **Date** 01/14/2023 **Type** Agency Inspection
Inspector COLLITO,BRAD,R **Time** 14:16:11 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A5
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/14/2023 14:24

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency	HERNANDO COUNTY SO	**Date**	02/11/2023	**Type**	Agency Inspection
Inspector	COLLITO,BRAD,R	**Time**	11:20:31	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	INT
Date and/or Time Adjusted	No	Alcohol Free Test # (Repeat) 1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test # (Repeat) 2	0.000
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Interferent Detect, Ambient Fail. Non-compliance:INT AN AMB FAILURE.

Data Download Date/Time 02/11/2023 11:36

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency	HERNANDO COUNTY SO	**Date**	02/11/2023	**Type**	Agency Inspection
Inspector	COLLITO,BRAD,R	**Time**	11:34:51	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	INT
Date and/or Time Adjusted	No	Alcohol Free Test # (Repeat) 1	0.000
Minimum Sample Volume Check: OK		Alcohol Free Test # (Repeat) 2	0.000
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Interferent Detect, Ambient Fail. Non-compliance:INT AN AMB FAILURE. A F / M A: Ambient Fail. Non-compliance: .

Data Download Date/Time 02/16/2023 14:53

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO **Date** 06/22/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 12:06:13 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.202
0.08 g/210L Dry Gas Std Test # 4	0.076
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.076
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 06/22/2023 12:13

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO **Date** 07/19/2023 **Type** Agency Inspection
Inspector BAKER,PATRICK,M **Time** 18:45:37 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

PASSED INSPECTION

Data Download Date/Time 07/19/2023 18:56

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO **Date** 08/13/2023 **Type** Agency Inspection
Inspector BAKER,PATRICK,M **Time** 19:57:41 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/13/2023 20:06

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO **Date** 09/29/2023 **Type** Agency Inspection
Inspector BAKER,PATRICK,M **Time** 19:19:40 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A3
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/29/2023 19:29

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO **Date** 10/30/2023 **Type** Agency Inspection
Inspector BAKER,PATRICK,M **Time** 20:05:30 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A3
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Non-compliance:DIAGNOSTIC FAILURE.

Data Download Date/Time 10/30/2023 20:23

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

Agency HERNANDO COUNTY SO **Date** 10/30/2023 **Type** Agency Inspection
Inspector BAKER,PATRICK,M **Time** 21:29:49 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A3
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

TEMPERATURE REGULATOR FAILURE CAUSING DIAGNISTIC FAILURENon-compliance: .

Data Download Date/Time 10/30/2023 21:40

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO **Date** 01/21/2023 **Type** Agency Inspection
Inspector DEANE,M, **Time** 01:21:05 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG205902
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/28/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/23/2023 09:20

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO **Date** 02/10/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 13:49:35 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.077
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.077
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.077
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.077
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.077
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.077
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.077
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 02/10/2023 13:55

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO **Date** 02/18/2023 **Type** Agency Inspection
Inspector DEANE,M, **Time** 17:42:16 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG205902
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/28/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/21/2023 09:42

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO **Date** 03/22/2023 **Type** Agency Inspection
Inspector DEANE,M, **Time** 11:00:25 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG205902
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/28/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/23/2023 21:02

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO **Date** 04/19/2023 **Type** Agency Inspection
Inspector DEANE,M, **Time** 11:23:29 **Software** 8100.27

0.05 Lot # 202112A' **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG205902
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/28/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/19/2023 11:35

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO **Date** 05/10/2023 **Type** Agency Inspection
Inspector DEANE,M, **Time** 11:32:03 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG205902
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/28/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/10/2023 11:46

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO **Date** 06/14/2023 **Type** Agency Inspection
Inspector DEANE,M, **Time** 10:07:31 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG229803
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/14/2023 10:16

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO **Date** 07/05/2023 **Type** Agency Inspection
Inspector DEANE,M, **Time** 13:54:58 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG229803
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/05/2023 14:04

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO **Date** 08/16/2023 **Type** Agency Inspection
Inspector DEANE,M, **Time** 06:17:10 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG229803
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/16/2023 06:28

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO **Date** 09/06/2023 **Type** Agency Inspection
Inspector DEANE,M, **Time** 11:43:05 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG229803
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/06/2023 11:52

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO **Date** 10/18/2023 **Type** Agency Inspection
Inspector DEANE,M, **Time** 12:42:59 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG229803
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/18/2023 12:52

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO **Date** 11/15/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 11:00:41 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG310901
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 04/19/2025

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.082
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.082
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.082
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.082
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.083
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.082
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.082
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 11/15/2023 11:11

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO **Date** 11/29/2023 **Type** Agency Inspection
Inspector DEANE,M, **Time** 20:16:49 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG229803
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.083
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance REG HOSE NOT HOOKED UP.

Data Download Date/Time 11/29/2023 20:27

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

Agency PINELLAS COUNTY SO **Date** 12/20/2023 **Type** Agency Inspection
Inspector DEANE,M, **Time** 10:14:11 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** AG229803
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/20/2023 10:27

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency	PERRY POLICE DEPT	**Date**	01/26/2023	**Type**	Agency Inspection
Inspector	HENING,FRED,C	**Time**	11:58:23	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202108D	**0.20 Lot #**	202106B	**0.08 Gas Lot #**	34520080A1
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/22/2023	**Expiration**	02/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 01/26/2023 12:11

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT **Date** 02/09/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 10:22:40 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Data Download Date/Time 02/09/2023 10:27

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency	PERRY POLICE DEPT	**Date**	02/14/2023	**Type**	Agency Inspection
Inspector	HENING,FRED,C	**Time**	13:51:48	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202108D	**0.20 Lot #**	202106B	**0.08 Gas Lot #**	33321080A4
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/22/2023	**Expiration**	01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 02/14/2023 14:13

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT **Date** 03/23/2023 **Type** Agency Inspection
Inspector HENING,FRED,C **Time** 13:02:54 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33922080A1
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 03/23/2023 13:56

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT **Date** 04/11/2023 **Type** Agency Inspection
Inspector HENING,FRED,C **Time** 12:18:44 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.203
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.204
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.206
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 04/11/2023 12:27

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT **Date** 05/23/2023 **Type** Agency Inspection
Inspector HENING,FRED,C **Time** 13:05:29 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 05/23/2023 13:22

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency	PERRY POLICE DEPT	**Date**	06/13/2023	**Type**	Agency Inspection
Inspector	HENING,FRED,C	**Time**	13:12:59	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	33922080A1
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	01/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 06/13/2023 13:36

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT **Date** 07/04/2023 **Type** Agency Inspection
Inspector HENING,FRED,C **Time** 10:53:47 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 07/04/2023 11:03

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT **Date** 08/28/2023 **Type** Agency Inspection
Inspector HENING,FRED,C **Time** 16:06:40 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.204
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.204
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.204
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 08/28/2023 16:17

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT **Date** 09/25/2023 **Type** Agency Inspection
Inspector HENING,FRED,C **Time** 12:09:56 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance INLET TUBE NOT SECURE.

Data Download Date/Time 09/25/2023 13:03

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT **Date** 10/18/2023 **Type** Agency Inspection
Inspector HENING,FRED,C **Time** 10:40:46 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A4
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 10/18/2023 10:50

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency	PERRY POLICE DEPT	**Date**	11/28/2023	**Type**	Agency Inspection
Inspector	HENING,FRED,C	**Time**	12:52:13	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	33321080A4
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 11/28/2023 13:03

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

Agency PERRY POLICE DEPT **Date** 12/14/2023 **Type** Agency Inspection
Inspector HENING,FRED,C **Time** 11:54:56 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33922080A1
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.205
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.205
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.206
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 12/14/2023 12:07

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD **Date** 01/17/2023 **Type** Agency Inspection
Inspector MCCULLOUGH,SEAN,J **Time** 09:08:53 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 08121080A1
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 05/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.052
0.08 g/210L Test # 1	0.082
0.20 g/210L Test # 1	0.203
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.083
0.20 g/210L Test # 2	0.204
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.053
0.08 g/210L Test # 3	0.083
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/17/2023 12:39

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD **Date** 02/23/2023 **Type** Agency Inspection
Inspector MCCULLOUGH,SEAN,J **Time** 09:28:55 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.204
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.204
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.204
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/23/2023 10:43

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD **Date** 03/29/2023 **Type** Agency Inspection
Inspector MCCULLOUGH,SEAN,J **Time** 08:44:34 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.204
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 03/29/2023 09:28

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD **Date** 04/04/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 10:47:37 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.203
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.203
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.203
0.08 g/210L Dry Gas Std Test # 5	0.081
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.052
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.203
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.204
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.052
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0008
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 04/04/2023 10:54

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency	ST PETERSBURG PD	**Date**	04/27/2023	**Type**	Agency Inspection
Inspector	MCCULLOUGH,SEAN,J	**Time**	13:50:12	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	08622080A1
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	06/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 04/27/2023 15:36

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD **Date** 05/30/2023 **Type** Agency Inspection
Inspector MCCULLOUGH,SEAN,J **Time** 10:30:01 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #**
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	No
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.20 g/210L Test # (Repeat) 1	0.207
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.052
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.206
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.211
Interferent Detect Test # 3	INT

Remarks:

20: Purge Fail, Purge Fail. Non-compliance:PURGE FAIL.

Data Download Date/Time 05/30/2023 11:34

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD **Date** 10/23/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 13:52:18 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG229803
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.200
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 10/23/2023 14:01

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD **Date** 11/29/2023 **Type** Agency Inspection
Inspector LENAHAN,RYAN,P **Time** 16:03:44 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 06023080A1
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 11/30/2023 09:08

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

Agency ST PETERSBURG PD **Date** 12/28/2023 **Type** Agency Inspection
Inspector MCCULLOUGH,SEAN,J **Time** 16:42:39 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206B **0.08 Gas Lot #** 06023080A1
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/14/2024 **Expiration** 04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/03/2024 08:22

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency PORT ORANGE PD **Date** 01/31/2023 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 17:07:32 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402328118
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/25/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/09/2023 10:08

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency PORT ORANGE PD **Date** 02/28/2023 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 13:55:30 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402328118
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/25/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/28/2023 14:03

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency PORT ORANGE PD **Date** 03/31/2023 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 13:13:10 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402328118
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/25/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/05/2023 10:58

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency PORT ORANGE PD **Date** 04/28/2023 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 14:17:13 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402328118
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/25/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/01/2023 11:56

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency PORT ORANGE PD **Date** 05/31/2023 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 17:20:11 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402328118
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/25/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/02/2023 15:27

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency	PORT ORANGE PD	**Date**	06/30/2023	**Type**	Agency Inspection
Inspector	HARRISON,WILLIAM,C	**Time**	17:29:39	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	402328118
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	01/25/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

A F / M A: No Sample Provided.

Data Download Date/Time 06/30/2023 17:36

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency PORT ORANGE PD **Date** 07/31/2023 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 11:25:52 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402328118
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/25/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/31/2023 11:36

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency PORT ORANGE PD **Date** 08/25/2023 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 15:29:30 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402328118
Expiration 03/28/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/25/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/25/2023 15:37

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency PORT ORANGE PD **Date** 09/29/2023 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 15:47:03 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402328118
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/25/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.075
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.076
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/29/2023 15:56

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency PORT ORANGE PD **Date** 10/31/2023 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 15:27:57 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402328118
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/25/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.192
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.193
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.193
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/31/2023 15:38

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency PORT ORANGE PD **Date** 11/30/2023 **Type** Agency Inspection
Inspector HARRISON,WILLIAM,C **Time** 11:43:26 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402328118
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/25/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.044
0.05 g/210L Test # (Repeat) 1	0.046
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.192
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.045
0.05 g/210L Test # (Repeat) 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.045
0.05 g/210L Test # (Repeat) 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

05: Control Outside TEMP LOW IN ROOM INLET TUBE INSULATED .

Data Download Date/Time 11/30/2023 11:55

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency	PORT ORANGE PD	**Date**	12/06/2023	**Type**	Agency Inspection
Inspector	HARRISON,WILLIAM,C	**Time**	15:27:00	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202307A	**0.20 Lot #**	202307C	**0.08 Gas Lot #**	402328118
Expiration	03/28/2025	**Expiration**	07/11/2025	**Expiration**	07/12/2025	**Expiration**	01/25/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.191
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/06/2023 15:38

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency PORT ORANGE PD **Date** 12/14/2023 **Type** Department Inspection
Inspector SIDDOWAY,BENJAMIN,W **Time** 13:58:20 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** AG325603
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 09/13/2025

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.194
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.194
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.194
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.195
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.194
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.194
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.194
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.194
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 12/14/2023 14:06

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005404

Agency POLK CSO **Date** 01/10/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 15:10:56 **Software** 8100.27

0.05 Lot #	202201C	**0.08 Lot #**	202201D	**0.20 Lot #**	202201E	**0.08 Gas Lot #**	00521080A2
Expiration	01/11/2024	**Expiration**	01/18/2024	**Expiration**	01/18/2024	**Expiration**	02/05/2023

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.203
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.203
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005404

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.203
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.203
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 01/10/2023 15:23

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency OCALA POLICE DEPT **Date** 07/17/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 09:30:07 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** 06723080A5
Expiration 01/11/2024 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 04/05/2025

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.047
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Data Download Date/Time 07/17/2023 09:35

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency OCALA POLICE DEPT **Date** 08/19/2023 **Type** Agency Inspection
Inspector JOEDICKE,WILLIAM,R **Time** 13:02:54 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33522080A5
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/25/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/08/2023 13:39

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency OCALA POLICE DEPT **Date** 09/08/2023 **Type** Agency Inspection
Inspector JOEDICKE,WILLIAM,R **Time** 12:19:41 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33522080A5
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

NONE

Data Download Date/Time 09/08/2023 13:39

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

Agency OCALA POLICE DEPT **Date** 12/11/2023 **Type** Agency Inspection
Inspector JOEDICKE,WILLIAM,R **Time** 14:58:12 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33522080A5
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

A F / M A: Range Exceeded TOO MUCH MA.

Data Download Date/Time 01/11/2024 11:46

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD **Date** 01/25/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 14:18:58 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/25/2023 14:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD **Date** 02/14/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 09:14:38 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/14/2023 09:26

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD **Date** 03/28/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 09:50:37 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/29/2023 10:39

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD **Date** 04/18/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 08:53:05 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/18/2023 09:02

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD **Date** 05/16/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 11:42:38 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/16/2023 12:10

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD **Date** 06/21/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 12:38:24 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/21/2023 12:50

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD **Date** 07/12/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 08:30:03 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/12/2023 08:52

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD **Date** 08/18/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 08:03:17 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/18/2023 08:17

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD **Date** 09/19/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 09:10:46 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
Interferent Detect Test # 3	INT

Remarks:

DID NOT PLUG CONTROL TUBE IN..

Data Download Date/Time 09/19/2023 09:27

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD **Date** 09/21/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 14:41:47 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.077
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.077
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.077
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.077
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 09/21/2023 14:49

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD **Date** 10/20/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 12:02:24 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/20/2023 12:13

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD **Date** 11/10/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 08:35:45 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/10/2023 08:47

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

Agency FORT MYERS PD **Date** 12/21/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 13:57:38 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/21/2023 14:18

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 01/17/2023 **Type** Agency Inspection
Inspector BALDWIN,DEAN, **Time** 16:41:38 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 25721080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2023

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/17/2023 16:49

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 02/16/2023 **Type** Agency Inspection
Inspector BALDWIN,DEAN, **Time** 16:34:08 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 25721080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2023

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/16/2023 16:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 03/20/2023 **Type** Agency Inspection
Inspector BALDWIN,DEAN, **Time** 13:47:32 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 25721080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2023

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

NO REMARKS

Data Download Date/Time 03/22/2023 14:30

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 04/11/2023 **Type** Agency Inspection
Inspector BALDWIN,DEAN, **Time** 16:21:27 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 25721080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2023

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/11/2023 16:28

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 05/23/2023 **Type** Agency Inspection
Inspector BALDWIN,DEAN, **Time** 17:10:52 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 25721080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2023

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/23/2023 17:17

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 06/14/2023 **Type** Agency Inspection
Inspector BALDWIN,DEAN, **Time** 14:06:15 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 25721080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2023

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/14/2023 14:16

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 06/23/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 09:53:13 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.078
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Data Download Date/Time 06/23/2023 09:59

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 07/13/2023 **Type** Agency Inspection
Inspector BALDWIN,DEAN, **Time** 14:10:28 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 25721080A2
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2023

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

IN SERVICE

Data Download Date/Time 07/13/2023 14:26

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 08/19/2023 **Type** Agency Inspection
Inspector BALDWIN,DEAN, **Time** 06:43:45 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 25721080A2
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2023

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/19/2023 06:53

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 09/18/2023 **Type** Agency Inspection
Inspector BALDWIN,DEAN, **Time** 16:23:47 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 25721080A2
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2023

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/24/2023 11:06

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 10/16/2023 **Type** Agency Inspection
Inspector BALDWIN,DEAN, **Time** 14:11:45 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 25721080A2
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2023

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/16/2023 14:20

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency DADE CITY PD **Date** 11/30/2023 **Type** Agency Inspection
Inspector BALDWIN,DEAN, **Time** 16:11:08 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 24923080A3
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 11/05/2025

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/30/2023 16:21

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005477

Agency	DADE CITY PD	**Date**	12/25/2023	**Type**	Agency Inspection
Inspector	DEL VALLE,M,	**Time**	10:37:14	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202307A	**0.20 Lot #**	202307C	**0.08 Gas Lot #**	24923080A3
Expiration	03/28/2025	**Expiration**	07/11/2025	**Expiration**	07/12/2025	**Expiration**	10/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

NO VISIBLE DAMAGE TO INTOX

Data Download Date/Time 12/25/2023 10:46

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency DAYTONA BEACH PD **Date** 05/17/2023 **Type** Agency Inspection
Inspector DUNNE,JUSTIN,R **Time** 11:25:10 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1475154
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/17/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/17/2023 11:50

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency DAYTONA BEACH PD **Date** 06/14/2023 **Type** Agency Inspection
Inspector DUNNE,JUSTIN,R **Time** 10:46:28 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1475154
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/17/2024

Number of Simulators 5 **Compliance** No

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	No
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	No

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.044
0.05 g/210L Test # (Repeat) 1	0.046
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	0.000
Interferent Detect Test # (Repeat) 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.045
0.05 g/210L Test # (Repeat) 2	0.046
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	0.000
Interferent Detect Test # (Repeat) 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.045
0.05 g/210L Test # (Repeat) 3	0.046
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
Interferent Detect Test # 3	0.000
Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det:NO ACETONE IN SIMULATOR. 05: Control Outside ToleranceNOT WARM ENOUGH. 08: RFI Detect. Non-compliance:RFI.

Data Download Date/Time 06/19/2023 19:12

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency	DAYTONA BEACH PD	**Date**	06/14/2023	**Type**	Agency Inspection
Inspector	DUNNE,JUSTIN,R	**Time**	11:37:45	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	1475154
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	03/17/2024

Number of Simulators 4 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.045
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.046
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/19/2023 19:12

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency DAYTONA BEACH PD **Date** 07/31/2023 **Type** Agency Inspection
Inspector DUNNE,JUSTIN,R **Time** 11:27:21 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 1475154
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 03/17/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/01/2023 11:53

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency DAYTONA BEACH PD **Date** 08/23/2023 **Type** Agency Inspection
Inspector DUNNE,JUSTIN,R **Time** 09:19:09 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1475154
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 03/17/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/23/2023 09:27

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency	DAYTONA BEACH PD	**Date**	09/15/2023	**Type**	Agency Inspection
Inspector	DUNNE,JUSTIN,R	**Time**	11:32:04	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	1475154
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	03/17/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/15/2023 11:59

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency	DAYTONA BEACH PD	**Date**	09/20/2023	**Type**	Department Inspection
Inspector	SIDDOWAY,BENJAMIN,W	**Time**	15:20:11	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators 5 **Average Standard Deviation** **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK	Yes		
Barometric Pressure Sensor Check: OK	Yes		
Alcohol Free Subject Test: 0.000	Yes		
Mouth Alcohol Test: Slope Not Met	Yes		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

CHANGED BATTERY, PORT MISALIGNEDNon-compliance:PURGE FAIL, BATTERY CHANGE MISALIGNED PORT .

Data Download Date/Time 09/21/2023 14:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency DAYTONA BEACH PD **Date** 09/21/2023 **Type** Department Inspection
Inspector SIDDOWAY,BENJAMIN,W **Time** 14:36:18 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202304C **0.08 Gas Lot #** 06723080A5
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 04/05/2025 **Expiration** 04/05/2025

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.076
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.075
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.075
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.074
0.08 g/210L Dry Gas Std Test # (Repeat) 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.196
0.08 g/210L Dry Gas Std Test # 5	0.072
0.08 g/210L Dry Gas Std Test # (Repeat) 5	0.078
Interferent Detect Test # 5	INT
Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.069
0.08 g/210L Dry Gas Std Test # (Repeat) 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.196
0.08 g/210L Dry Gas Std Test # 10	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 10	0.078
Interferent Detect Test # 10	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0380
0.08 g/210L Dry Gas Std Test (Std Dev) (Repeat)	0.0004

Remarks:

08: Control Outside Tolerance, TANK RAN EMPTY.

Data Download Date/Time 09/21/2023 14:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency	DAYTONA BEACH PD	**Date**	11/13/2023	**Type**	Agency Inspection
Inspector	DUNNE,JUSTIN,R	**Time**	14:37:55	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	402290677
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	12/07/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/13/2023 14:56

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

Agency DAYTONA BEACH PD **Date** 12/31/2023 **Type** Agency Inspection
Inspector DUNNE,JUSTIN,R **Time** 14:29:55 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402290677
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 12/07/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/31/2023 14:45

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 01/20/2023 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 13:04:55 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 22722080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/20/2023 13:08

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 02/08/2023 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 17:48:46 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 22722080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.043
0.05 g/210L Test # (Repeat) 1	0.048
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.190
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.05 g/210L Test # (Repeat) 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.191
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.05 g/210L Test # (Repeat) 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.191
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance CHECKED TUBING SMALL AIR LEAK.

Data Download Date/Time 02/08/2023 18:00

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 03/13/2023 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 06:49:49 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration 12/07/2023 **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.044
Barometric Pressure Sensor Check: OK		Interferent Detect Test # 1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test # 2	0.045
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # (Repeat) 2	0.044
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 3	0.000
0.05 g/210L Test: OK	No	0.05 g/210L Test # 3	0.044
0.08 g/210L Test: OK	No	0.05 g/210L Test # (Repeat) 3	0.044
0.20 g/210L Test: OK	No	Interferent Detect Test # 3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance, Control Outside Tolerance. Non-compliance:.05 OUT OF TOLERANCE.

Data Download Date/Time 03/13/2023 09:01

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 04/11/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 11:16:11 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.080
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 04/11/2023 11:19

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 05/11/2023 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 17:59:02 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 22722080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 05/11/2023 18:27

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency	FERNANDINA BEACH PD	**Date**	06/27/2023	**Type**	Agency Inspection
Inspector	SMITH,JASON,E	**Time**	21:24:49	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	22722080A1
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	10/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/27/2023 21:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 07/23/2023 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 18:56:49 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 22722080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.189
0.20 g/210L Test # (Repeat) 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.20 g/210L Test # (Repeat) 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.192
0.20 g/210L Test # (Repeat) 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance.

Data Download Date/Time 07/23/2023 19:03

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 08/26/2023 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 04:48:35 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 22722080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.075
0.20 g/210L Test # 1	0.194
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/26/2023 04:55

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 09/25/2023 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 01:02:01 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206B **0.08 Gas Lot #** 22722080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/14/2024 **Expiration** 10/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.046
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.046
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/25/2023 01:09

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 10/17/2023 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 19:11:35 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 22722080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 10/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.084
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.084
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.084
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/17/2023 19:19

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 11/13/2023 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 18:16:02 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 22722080A1
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 10/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.047
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.083
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.083
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.083
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/13/2023 18:25

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency FERNANDINA BEACH PD **Date** 12/18/2023 **Type** Agency Inspection
Inspector SMITH,JASON,E **Time** 16:32:56 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202307A **0.20 Lot #** 202206C **0.08 Gas Lot #** 22722080A1
Expiration 03/28/2025 **Expiration** 07/11/2025 **Expiration** 06/15/2024 **Expiration** 10/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.193
0.08 g/210L Dry Gas Std Test # 1	0.085
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.047
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.085
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.085
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/18/2023 16:41

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 01/05/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 02:11:26 **Software** 8100.27

0.05 Lot # 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 01/05/2023 07:08

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 02/09/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 03:48:44 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 02/09/2023 04:22

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 03/02/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 03:50:25 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 03/02/2023 04:45

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 04/06/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 03:53:28 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 04/06/2023 07:45

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 05/12/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 04:45:34 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.188
0.20 g/210L Test # (Repeat) 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.178
0.20 g/210L Test # (Repeat) 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.179
0.20 g/210L Test # (Repeat) 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance/POSS SIM ISSUE/LOOSE CONNECT .

Data Download Date/Time 05/12/2023 07:16

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 06/09/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 04:25:42 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/09/2023 07:48

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 06/14/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 12:09:25 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Alcohol Free Test # (Repeat) 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.078
Interferent Detect Test # 4	INT
Alcohol Free Test # (Repeat) 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Alcohol Free Test # (Repeat) 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # (Repeat) 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # (Repeat) 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # (Repeat) 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # (Repeat) 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

00: Ambient FaiL. UNKNOWN. AIRED OUT PRIOR TO REPEATING. .

Data Download Date/Time 06/14/2023 12:18

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 07/06/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 02:01:53 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

AI G PACHECO PRESENT

Data Download Date/Time 07/06/2023 02:34

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 08/24/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 01:48:51 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 08/24/2023 02:01

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 09/01/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 02:44:55 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/01/2023 02:55

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 10/05/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 03:37:23 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 10/05/2023 04:12

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 11/03/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 04:38:08 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202106B **0.20 Lot #** 202106C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 11/03/2023 05:27

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

Agency MIAMI DADE PD **Date** 12/01/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 03:53:05 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 12/01/2023 04:07

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 01/05/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 04:39:25 **Software** 8100.27

0.05 Lot # 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 01/05/2023 08:00

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 01/11/2023 **Type** Department Inspection
Inspector REYES-RIVERA,DAVID,E **Time** 13:10:10 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** 00521080A2
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 02/05/2023

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 01/11/2023 13:25

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 02/16/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 03:48:02 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 02/16/2023 07:39

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 03/02/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 03:46:48 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	0.000
Interferent Detect Test # (Repeat) 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Interferent Detect Test # (Repeat) 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT
Interferent Detect Test # (Repeat) 3	INT

Remarks:

Int Det:NOT CONNECTED ON 1ST/AI R MYRTIL PRESENT

Data Download Date/Time 03/02/2023 05:03

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 04/13/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 02:53:53 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 04/13/2023 03:11

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 05/11/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 04:11:45 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AI G PACHECO PRESENT

Data Download Date/Time 05/11/2023 07:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 06/02/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 04:07:15 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 06/02/2023 08:23

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency	MIAMI DADE PD	**Date**	07/06/2023	**Type**	Agency Inspection
Inspector	CLOSIUS,RICHARD,L	**Time**	03:47:53	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	33321080A2
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

AI G PACHECO PRESENT

Data Download Date/Time 07/06/2023 03:58

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 08/17/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 04:00:06 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 08/17/2023 04:14

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 09/01/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 03:01:31 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/01/2023 03:16

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 10/05/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 03:36:01 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 10/05/2023 04:02

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 11/03/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 04:38:45 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202106B **0.20 Lot #** 202106C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 11/03/2023 05:14

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

Agency MIAMI DADE PD **Date** 12/01/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 06:35:04 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 12/13/2023 14:35

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD **Date** 01/05/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 02:17:00 **Software** 8100.27

0.05 Lot # 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 01/05/2023 07:15

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD **Date** 01/17/2023 **Type** Department Inspection
Inspector REYES-RIVERA,DAVID,E **Time** 13:14:00 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** 00521080A2
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 02/05/2023

Number of Simulators 5 **Average Standard Deviation** 0.0001 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.203
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.051
0.08 g/210L Test # 4	0.081
0.20 g/210L Test # 4	0.203
0.08 g/210L Dry Gas Std Test # 4	0.081
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.051
0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.203
0.08 g/210L Dry Gas Std Test # 5	0.081
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.203
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.203
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 01/17/2023 13:48

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD **Date** 02/09/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 03:47:10 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 02/09/2023 04:14

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD **Date** 03/02/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 03:40:57 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.052
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 03/02/2023 05:05

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD **Date** 04/06/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 03:56:25 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 04/06/2023 08:07

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD **Date** 05/11/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 04:18:17 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.075
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

AI G PACHECO PRESENT

Data Download Date/Time 05/11/2023 07:53

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD **Date** 06/16/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 14:35:46 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.202
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 06/16/2023 14:42

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD **Date** 07/06/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 02:10:31 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.082
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

AI G PACHECO PRESENT

Data Download Date/Time 07/06/2023 02:38

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD **Date** 08/17/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 04:02:17 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 08/17/2023 04:21

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD **Date** 09/01/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 02:58:14 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.203
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.204
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/01/2023 03:25

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD **Date** 10/05/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 03:39:16 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 10/05/2023 04:09

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD **Date** 11/03/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 04:41:31 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202106B **0.20 Lot #** 202106C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.076
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 11/03/2023 05:22

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

Agency MIAMI DADE PD **Date** 12/01/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 06:32:25 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 12/01/2023 07:08

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD **Date** 01/05/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 02:28:17 **Software** 8100.27

0.05 Lot # 202101C **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 01/12/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 01/05/2023 08:07

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD **Date** 01/12/2023 **Type** Department Inspection
Inspector REYES-RIVERA,DAVID,E **Time** 08:29:12 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** 00521080A2
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 02/05/2023

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.202
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 01/12/2023 08:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD **Date** 02/16/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 03:50:35 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.054
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.054
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.054
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.078
Interferent Detect Test # 3	INT

Remarks:

08: RFI-POLICE RADIO/AI MYRTIL PRES

Data Download Date/Time 02/16/2023 07:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD **Date** 03/09/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 03:16:14 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.082
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 03/09/2023 03:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD **Date** 04/06/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 04:05:25 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/22/2023 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 04/06/2023 07:38

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD **Date** 10/10/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 14:13:31 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.203
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.203
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.204
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.204
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.204
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.204
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.204
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.204
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Data Download Date/Time 10/10/2023 14:20

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD **Date** 11/03/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 02:05:55 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202106B **0.20 Lot #** 202106C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.205
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.205
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.205
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 11/03/2023 02:25

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

Agency MIAMI DADE PD **Date** 12/01/2023 **Type** Agency Inspection
Inspector CLOSIUS,RICHARD,L **Time** 06:30:34 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A2
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.205
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.205
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.051
0.08 g/210L Test # 3	0.081
0.20 g/210L Test # 3	0.205
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AI R MYRTIL PRESENT

Data Download Date/Time 12/01/2023 06:58

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD **Date** 01/25/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 12:56:51 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.076
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/25/2023 13:06

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD **Date** 02/14/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 10:27:19 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/14/2023 10:57

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD **Date** 03/28/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 11:59:46 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/29/2023 11:09

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD **Date** 04/18/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 10:53:33 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 04/18/2023 11:05

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD **Date** 05/16/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 10:23:02 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.076
Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 05/16/2023 10:36

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD **Date** 06/21/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 14:56:05 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/21/2023 15:07

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD **Date** 07/12/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 10:01:37 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/12/2023 10:12

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD **Date** 08/18/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 10:55:48 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/18/2023 11:06

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD **Date** 09/19/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 10:29:10 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/19/2023 10:41

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD **Date** 09/20/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:04:57 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.077
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.077
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.077
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.077
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.077
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.077
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 09/20/2023 13:12

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD **Date** 10/20/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 13:11:29 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/20/2023 13:20

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD **Date** 11/10/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 11:17:52 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A3
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/10/2023 11:30

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD **Date** 12/21/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 08:30:29 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 33321080A3
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/21/2023 10:11

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005655

Agency FORT MYERS PD **Date** 12/21/2023 **Type** Agency Inspection
Inspector AGUILERA,DANIEL,J **Time** 09:55:49 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 08223080A3
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 06/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/21/2023 10:11

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 01/03/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 13:39:35 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1472328
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/02/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/03/2023 13:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 01/18/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 11:24:04 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1472328
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 04/02/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/18/2023 12:26

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 01/25/2023 **Type** Department Inspection
Inspector NICODEMO,PHIL, **Time** 10:53:18 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.198
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.077
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.077
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.077
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 01/25/2023 10:57

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 02/24/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 14:15:56 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402330433
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/24/2023 14:25

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 03/24/2023 **Type** Agency Inspection
Inspector LYONS/GAYLE,DANIEL/OWEN,E **Time** 11:33:25 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402330433
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/24/2023 11:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 04/25/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 14:54:42 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402330433
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/25/2023 15:05

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 05/19/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 12:52:55 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402330433
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	Yes
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Time-Date changed.

Data Download Date/Time 05/19/2023 13:15

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 06/21/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 14:43:09 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 402330433
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.000
0.20 g/210L Test # (Repeat) 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.000
0.20 g/210L Test # (Repeat) 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.000
0.20 g/210L Test # (Repeat) 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

AF/MA:MOUTH ALCOHOL NOT RINSED FROM PREVIOUS TEST. 0.20 OUTSIDE TOLERANCE-DID NOT CONNECT SIMULATOR, .

Data Download Date/Time 06/21/2023 14:55

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 07/17/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 13:53:06 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 402330433
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/17/2023 14:05

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency	OSCEOLA COUNTY SO	**Date**	08/29/2023	**Type**	Agency Inspection
Inspector	LYONS,DANIEL,E	**Time**	07:34:33	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

DIAG FAIL DURING SET UPNon-compliance:DIAG FAIL DURING SET UP.

Data Download Date/Time 08/29/2023 08:45

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 08/29/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 08:28:39 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402330433
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/29/2023 08:45

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 09/12/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 17:34:39 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402330433
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/12/2023 18:16

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 10/23/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 16:07:07 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402330433
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/23/2023 16:25

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 11/22/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 16:07:59 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402330433
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

AF/MA ABIENT FAIL.

Data Download Date/Time 11/22/2023 16:26

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

Agency OSCEOLA COUNTY SO **Date** 12/21/2023 **Type** Agency Inspection
Inspector LYONS,DANIEL,E **Time** 11:43:06 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 402330433
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 01/24/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/21/2023 11:53

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency SARASOTA COUNTY SO **Date** 01/28/2023 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 11:08:34 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 15421080A1
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 01/29/2023 13:23

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency	SARASOTA COUNTY SO	**Date**	02/24/2023	**Type**	Agency Inspection
Inspector	BRENCKLE,TIMOTHY,M	**Time**	21:35:15	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202108D	**0.20 Lot #**	202106B	**0.08 Gas Lot #**	15922080A4
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/22/2023	**Expiration**	08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 02/28/2023 17:52

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency	SARASOTA COUNTY SO	**Date**	03/26/2023	**Type**	Agency Inspection
Inspector	BRENCKLE,TIMOTHY,M	**Time**	22:26:19	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202108D	**0.20 Lot #**	202106B	**0.08 Gas Lot #**	15922080A4
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/22/2023	**Expiration**	08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

CHANGED TIME

Data Download Date/Time 03/30/2023 10:15

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency SARASOTA COUNTY SO **Date** 04/09/2023 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 09:13:21 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 15922080A4
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.079
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.088
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.079
Interferent Detect Test # 3	INT

Remarks:

CHECK CYLINDER.

Data Download Date/Time 04/13/2023 07:31

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency SARASOTA COUNTY SO **Date** 05/04/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:40:52 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.201
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 05/04/2023 14:11

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency SARASOTA COUNTY SO **Date** 05/12/2023 **Type** Agency Inspection
Inspector BRENCKLE,TIMOTHY,M **Time** 12:20:05 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 15922080A4
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 05/17/2023 17:41

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency SARASOTA COUNTY SO **Date** 06/23/2023 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 11:01:32 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 06/23/2023 11:07

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency SARASOTA COUNTY SO **Date** 07/27/2023 **Type** Agency Inspection
Inspector SANDERS,DEPUTY,J **Time** 08:49:37 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

PASSED INSP AT JAIL

Data Download Date/Time 07/27/2023 10:20

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency	SARASOTA COUNTY SO	**Date**	08/29/2023	**Type**	Agency Inspection
Inspector	SANDERS,DEPUTY,J	**Time**	09:51:54	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	15922080A4
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

PASSED INSP

Data Download Date/Time 08/29/2023 09:58

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency SARASOTA COUNTY SO **Date** 09/25/2023 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 08:30:45 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL\

Data Download Date/Time 09/25/2023 08:38

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency	SARASOTA COUNTY SO	**Date**	10/18/2023	**Type**	Agency Inspection
Inspector	SANDERS,DEPUTY,J	**Time**	09:49:56	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	15922080A4
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 10/18/2023 09:57

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency	SARASOTA COUNTY SO	**Date**	11/30/2023	**Type**	Agency Inspection
Inspector	SANDERS,DEPUTY,J	**Time**	12:27:44	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	15922080A4
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

PASSED INSPECTION AT JAIL

Data Download Date/Time 11/30/2023 12:37

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

Agency SARASOTA COUNTY SO **Date** 12/19/2023 **Type** Agency Inspection
Inspector FELTOVIC,DEP,M **Time** 10:14:40 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A4
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.080
Interferent Detect Test # 3	INT

Remarks:

.08 HOOK UP GAS TUBE.

Data Download Date/Time 12/19/2023 10:23

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

Agency PALM BEACH CO SO **Date** 01/13/2023 **Type** Agency Inspection
Inspector KARLECKE,JASON,C **Time** 12:49:02 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 08622080A1
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 06/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/13/2023 12:58

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

Agency PALM BEACH CO SO **Date** 02/07/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 12:54:03 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG222203
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/10/2024

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.201
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 02/07/2023 13:01

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

Agency PALM BEACH CO SO **Date** 02/10/2023 **Type** Agency Inspection
Inspector KARLECKE,JASON,C **Time** 12:14:19 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 19021080A2
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 09/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/10/2023 12:24

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

Agency PALM BEACH CO SO **Date** 03/10/2023 **Type** Agency Inspection
Inspector KARLECKE,JASON,C **Time** 13:54:29 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A3
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.201
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.203
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/10/2023 14:00

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

Agency	PALM BEACH CO SO	**Date**	04/14/2023	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	15:56:12	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202108D	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	15922080A3
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/15/2024	**Expiration**	08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/14/2023 16:02

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

Agency	PALM BEACH CO SO	**Date**	05/19/2023	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	12:20:19	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202108D	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	29122080A1
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/15/2024	**Expiration**	12/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 05/19/2023 12:27

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

Agency	PALM BEACH CO SO	**Date**	06/23/2023	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	16:21:16	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202108D	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	08622080A1
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/15/2024	**Expiration**	06/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/23/2023 16:29

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

Agency PALM BEACH CO SO **Date** 07/14/2023 **Type** Agency Inspection
Inspector KARLECKE,JASON,C **Time** 13:36:56 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 08622080A1
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 06/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/14/2023 13:44

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

Agency	PALM BEACH CO SO	**Date**	08/11/2023	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	12:38:46	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202108D	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	29122080A1
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/15/2024	**Expiration**	12/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/11/2023 12:46

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

Agency	PALM BEACH CO SO	**Date**	09/15/2023	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	12:25:50	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	08622080A1
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	06/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/15/2023 12:35

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

Agency	PALM BEACH CO SO	**Date**	10/20/2023	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	13:48:01	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	08622080A1
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	06/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/20/2023 13:55

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

Agency	PALM BEACH CO SO	**Date**	11/17/2023	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	12:19:05	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	08622080A1
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	06/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/17/2023 12:27

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

Agency	PALM BEACH CO SO	**Date**	12/08/2023	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	14:41:36	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	06723080A5
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.201
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/08/2023 14:50

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

Agency PALM BEACH CO SO **Date** 01/10/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 14:12:28 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** 00521080A2
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 02/05/2023

Number of Simulators 5 **Average Standard Deviation** 0.0012 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.045
0.08 g/210L Test # 1	0.075
0.20 g/210L Test # 1	0.184
0.20 g/210L Test # (Repeat) 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.194
0.20 g/210L Test # (Repeat) 2	0.203
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.20 g/210L Test # (Repeat) 3	0.204
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.200
0.20 g/210L Test # (Repeat) 4	0.204
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.050

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.202
0.20 g/210L Test # (Repeat) 5	0.203
0.08 g/210L Dry Gas Std Test # 5	0.081
Interferent Detect Test # 5	INT
Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.202
0.20 g/210L Test # (Repeat) 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.202
0.20 g/210L Test # (Repeat) 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.202
0.20 g/210L Test # (Repeat) 8	0.203
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.203
0.20 g/210L Test # (Repeat) 9	0.203
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.203
0.20 g/210L Test # (Repeat) 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

0.05 g/210L Test (Std Dev)	0.0015
0.08 g/210L Test (Std Dev)	0.0015
0.20 g/210L Test (Std Dev)	0.0059
0.20 g/210L Test (Std Dev) (Repeat)	0.0013
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

20: Control Outside Tolerance.

Data Download Date/Time 01/10/2023 14:18

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

Agency	PALM BEACH CO SO	**Date**	01/13/2023	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	15:36:43	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202108D	**0.20 Lot #**	202106B	**0.08 Gas Lot #**	19021080A2
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/22/2023	**Expiration**	09/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/13/2023 15:46

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

Agency	PALM BEACH CO SO	**Date**	02/10/2023	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	13:10:33	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202108D	**0.20 Lot #**	202106B	**0.08 Gas Lot #**	02021080A1
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/22/2023	**Expiration**	03/05/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.080
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/10/2023 13:16

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

Agency	PALM BEACH CO SO	**Date**	03/10/2023	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	14:09:52	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202108D	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	29122080A1
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/15/2024	**Expiration**	12/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.081
0.20 g/210L Test # 1	0.202
0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.081
0.20 g/210L Test # 2	0.202
0.08 g/210L Dry Gas Std Test # 2	0.082
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.202
0.08 g/210L Dry Gas Std Test # 3	0.082
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/10/2023 14:17

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

Agency PALM BEACH CO SO **Date** 06/08/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 12:06:25 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG229803
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 10/25/2024

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.048
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.048
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Data Download Date/Time 06/08/2023 13:34

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

Agency PALM BEACH CO SO **Date** 06/23/2023 **Type** Agency Inspection
Inspector KARLECKE,JASON,C **Time** 15:49:32 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A3
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 06/23/2023 15:56

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

Agency PALM BEACH CO SO **Date** 07/14/2023 **Type** Agency Inspection
Inspector KARLECKE,JASON,C **Time** 16:35:32 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A3
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 07/14/2023 16:42

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

Agency	PALM BEACH CO SO	**Date**	08/11/2023	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	13:22:10	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202108D	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	08622080A1
Expiration	12/07/2023	**Expiration**	08/17/2023	**Expiration**	06/15/2024	**Expiration**	06/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/11/2023 13:32

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

Agency PALM BEACH CO SO **Date** 09/15/2023 **Type** Agency Inspection
Inspector KARLECKE,JASON,C **Time** 12:57:30 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 15922080A3
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 08/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 09/15/2023 13:08

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

Agency	PALM BEACH CO SO	**Date**	10/20/2023	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	17:04:02	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	06723080A5
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.198
0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.081
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.081
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 10/20/2023 17:15

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

Agency	PALM BEACH CO SO	**Date**	11/17/2023	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	15:18:03	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	08622080A1
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	06/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 11/17/2023 15:28

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

Agency	PALM BEACH CO SO	**Date**	12/08/2023	**Type**	Agency Inspection
Inspector	KARLECKE,JASON,C	**Time**	17:11:19	**Software**	8100.27

0.05 Lot #	202303J	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	06723080A5
Expiration	03/28/2025	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	04/05/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 12/08/2023 17:20

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

Agency FHP TROOP C **Date** 01/24/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 14:42:08 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** 00521080A2
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 02/05/2023

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.076
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test # 2	0.076
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.076
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.077
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.076
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.077
0.20 g/210L Test # 5	0.199
0.08 g/210L Dry Gas Std Test # 5	0.076
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.077
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.076
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.076
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.077
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.077
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.077
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0009

Remarks:

Data Download Date/Time 01/24/2023 14:46

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

Agency FHP TROOP C **Date** 02/28/2023 **Type** Agency Inspection
Inspector DOBOSIEWICZ,ANTHONY,N **Time** 13:29:54 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 2402498227
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/22/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.080
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.077
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.077
Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED NO CRACKS/ IN COMPLIANCE

Data Download Date/Time 02/28/2023 15:18

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

Agency FHP TROOP C **Date** 03/30/2023 **Type** Agency Inspection
Inspector GONZALEZ,CHRISTOPHER,M **Time** 16:06:09 **Software** 8100.27

0.05 Lot # 20112A **0.08 Lot #** 202206B **0.20 Lot #** 202106B **0.08 Gas Lot #** 2402498227
Expiration 12/07/2023 **Expiration** 06/14/2022 **Expiration** 06/22/2023 **Expiration** 07/22/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.080
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

INSPECTED CASE NO CRACKS, IN COMPLIANCE

Data Download Date/Time 03/31/2023 15:45

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

Agency	FHP TROOP C	**Date**	04/28/2023	**Type**	Agency Inspection
Inspector	GONZALEZ,CHRISTOPHER,M	**Time**	18:14:26	**Software**	8100.27

0.05 Lot #	202112A	**0.08 Lot #**	202206B	**0.20 Lot #**	202206C	**0.08 Gas Lot #**	2402498227
Expiration	12/07/2023	**Expiration**	06/14/2024	**Expiration**	06/15/2024	**Expiration**	07/22/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

CHECKED CASE AND WIPED DOWN. IN COMPLIANCE.

Data Download Date/Time 05/03/2023 02:35

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

Agency FHP TROOP C **Date** 05/29/2023 **Type** Agency Inspection
Inspector DOBOSIEWICZ,ANTHONY,N **Time** 03:42:51 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 2402498227
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/22/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.201
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.198
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED NO CRACKS IN COMPLIANCE

Data Download Date/Time 05/29/2023 04:27

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

Agency FHP TROOP C **Date** 06/22/2023 **Type** Agency Inspection
Inspector GONZALEZ,CHRISTOPHER,M **Time** 17:40:33 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 2402498227
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/22/2025

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

CASE CHECKED AND WIPED DOWN

Data Download Date/Time 06/22/2023 17:52

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

Agency FHP TROOP C **Date** 07/28/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 14:56:18 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** 06723080A5
Expiration 01/11/2024 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 04/05/2025

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.199
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.048
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.199
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 07/28/2023 15:00

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

Agency	FHP TROOP C	**Date**	08/01/2023	**Type**	Agency Inspection
Inspector	SOTO,ISRAEL,	**Time**	11:17:12	**Software**	8100.27

0.05 Lot #	**0.08 Lot #**	**0.20 Lot #**	**0.08 Gas Lot #**
Expiration	**Expiration**	**Expiration**	**Expiration**

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

Data Download Date/Time 08/01/2023 14:15

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

Agency FHP TROOP C **Date** 08/01/2023 **Type** Department Inspection
Inspector SOTO,ISRAEL, **Time** 14:11:05 **Software** 8100.27

0.05 Lot # 202303K **0.08 Lot #** 202303L **0.20 Lot #** 202201E **0.08 Gas Lot #** 06723080A5
Expiration 03/29/2025 **Expiration** 03/29/2025 **Expiration** 01/18/2024 **Expiration** 04/05/2025

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.007
Alcohol Free Test # (Repeat) 1	0.000
0.05 g/210L Test # 1	0.051
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.200
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
Alcohol Free Test # (Repeat) 2	0.000
0.05 g/210L Test # 2	0.051
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
Alcohol Free Test # (Repeat) 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.200
0.08 g/210L Dry Gas Std Test # 3	0.080
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
Alcohol Free Test # (Repeat) 4	0.000
0.05 g/210L Test # 4	0.050
0.08 g/210L Test # 4	0.079
0.20 g/210L Test # 4	0.199
0.08 g/210L Dry Gas Std Test # 4	0.080
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
Alcohol Free Test # (Repeat) 5	0.000

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

0.05 g/210L Test # 5	0.050
0.08 g/210L Test # 5	0.079
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.080
Interferent Detect Test # 5	INT
Alcohol Free Test # 6	0.000
Alcohol Free Test # (Repeat) 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
Alcohol Free Test # (Repeat) 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
Alcohol Free Test # (Repeat) 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
Alcohol Free Test # (Repeat) 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
Alcohol Free Test # (Repeat) 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

00: Control Outside Tolerance CHECKED SIM AND CLEARED AREA .

Data Download Date/Time 08/01/2023 14:15

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

Agency US COAST GUARD MIAMI **Date** 01/12/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 11:25:54 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1355293
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/29/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI PRICE #958 PRESENT

Data Download Date/Time 01/12/2023 14:12

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

Agency US COAST GUARD MIAMI **Date** 02/23/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 10:10:12 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1355293
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/29/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI MILLAN #141 PRESENT

Data Download Date/Time 02/23/2023 13:49

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

Agency US COAST GUARD MIAMI **Date** 03/30/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 11:41:12 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202108D **0.20 Lot #** 202106B **0.08 Gas Lot #** 1355293
Expiration 12/07/2023 **Expiration** 08/17/2023 **Expiration** 06/22/2023 **Expiration** 07/29/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

AI MILLAN #141 PRESENT

Data Download Date/Time 03/30/2023 13:33

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

Agency US COAST GUARD MIAMI **Date** 04/11/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 09:45:44 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1355293
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/29/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.080
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.047
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.195
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AI MILLAN #141 PRESENT

Data Download Date/Time 04/11/2023 13:49

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

Agency US COAST GUARD MIAMI **Date** 04/26/2023 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 12:06:49 **Software** 8100.27

0.05 Lot # 202201C **0.08 Lot #** 202201D **0.20 Lot #** 202201E **0.08 Gas Lot #** AG223802
Expiration 01/11/2024 **Expiration** 01/18/2024 **Expiration** 01/18/2024 **Expiration** 08/26/2024

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	Yes
Barometric Pressure Sensor Check: OK	Yes
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.050
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.050
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT
Alcohol Free Test # 4	0.000
0.05 g/210L Test # 4	0.049
0.08 g/210L Test # 4	0.078
0.20 g/210L Test # 4	0.197
0.08 g/210L Dry Gas Std Test # 4	0.079
Interferent Detect Test # 4	INT
Alcohol Free Test # 5	0.000
0.05 g/210L Test # 5	0.049
0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.197
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 04/26/2023 12:45

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

Agency US COAST GUARD MIAMI **Date** 05/25/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 10:07:40 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1355293
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/29/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.197
0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.079
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

AI MILLAN #141 PRESENT

Data Download Date/Time 05/25/2023 10:52

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

Agency US COAST GUARD MIAMI **Date** 06/15/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 09:26:38 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 1355293
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 07/29/2023

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.048
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AI MILLAN #141 PRESENT

Data Download Date/Time 06/15/2023 10:42

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

Agency US COAST GUARD MIAMI **Date** 07/20/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 11:17:01 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 02422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI PRICE #958 PRESENT

Data Download Date/Time 07/20/2023 14:06

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

Agency US COAST GUARD MIAMI **Date** 08/24/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 10:34:10 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 02422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.077
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.077
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.077
0.20 g/210L Test # 3	0.197
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 08/24/2023 11:13

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

Agency US COAST GUARD MIAMI **Date** 09/21/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 10:25:40 **Software** 8100.27

0.05 Lot # 202112A **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 02422080A1
Expiration 12/07/2023 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI PRICE #958 PRESENT

Data Download Date/Time 09/21/2023 11:45

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

Agency US COAST GUARD MIAMI **Date** 10/19/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 11:07:15 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 02422080A1
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

AI MILLAN #141 PRESENT

Data Download Date/Time 10/19/2023 13:51

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

Agency US COAST GUARD MIAMI **Date** 11/16/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 10:22:04 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 02422080A1
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.079
0.20 g/210L Test # 1	0.195
0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.048
0.08 g/210L Test # 2	0.078
0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test # 2	0.079
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.050
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.079
Interferent Detect Test # 3	INT

Remarks:

AI K. MILLAN #141 PRESENT

Data Download Date/Time 11/16/2023 11:29

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

Agency US COAST GUARD MIAMI **Date** 12/27/2023 **Type** Agency Inspection
Inspector SOCARRAS,PEDRO,A **Time** 11:57:40 **Software** 8100.27

0.05 Lot # 202303J **0.08 Lot #** 202206B **0.20 Lot #** 202206C **0.08 Gas Lot #** 02422080A1
Expiration 03/28/2025 **Expiration** 06/14/2024 **Expiration** 06/15/2024 **Expiration** 02/05/2024

Number of Simulators 5 **Compliance** Yes

Check or Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes
Date and/or Time Adjusted	No
Minimum Sample Volume Check: OK	
Barometric Pressure Sensor Check: OK	
Alcohol Free Subject Test: 0.000	Yes
Mouth Alcohol Test: Slope Not Met	Yes
Interferent Detect Test: Interferent Detect	Yes
Diagnostic Check (Post-Inspection): OK	Yes
Alcohol Free Test: OK	Yes
0.05 g/210L Test: OK	Yes
0.08 g/210L Test: OK	Yes
0.20 g/210L Test: OK	Yes
0.08 g/210L Dry Gas Std Test: OK	Yes

Test	Result
Alcohol Free Test # 1	0.000
0.05 g/210L Test # 1	0.049
0.08 g/210L Test # 1	0.078
0.20 g/210L Test # 1	0.196
0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test # 1	INT
Alcohol Free Test # 2	0.000
0.05 g/210L Test # 2	0.049
0.08 g/210L Test # 2	0.079
0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test # 2	0.078
Interferent Detect Test # 2	INT
Alcohol Free Test # 3	0.000
0.05 g/210L Test # 3	0.049
0.08 g/210L Test # 3	0.078
0.20 g/210L Test # 3	0.196
0.08 g/210L Dry Gas Std Test # 3	0.078
Interferent Detect Test # 3	INT

Remarks:

AI MILLAN #141 PRESENT

Data Download Date/Time 01/03/2024 09:37

Printed: 24 January 2024

COBRA ©1997-2014 CMI, Inc.